

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: JIGUAN MAURICIO

DATE/TIME OF TEST: 10/01/2023 02:23:53

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.208	02:19
Breath Volume	2.710	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.205	02:22
Breath Volume	2.753	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,DARIN,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LOWER, TIFFANY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NOSTRANT

DATE/TIME OF TEST: 10/02/2023 03:47:01

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.179	03:41
Breath Volume	1.886	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.172	03:45
Breath Volume	2.125	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL, GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PEREZ, RICKY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: UNGER

DATE/TIME OF TEST: 10/05/2023 12:44:27

Diagnostics Check	OK	12:38
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:39
Subject Sample # 1	0.167	12:40
Breath Volume	1.546	12:40
Air Blank	0.000	12:40
Air Blank	0.000	12:42
Subject Sample # 2	0.157	12:42
Breath Volume	2.042	12:42
Air Blank	0.000	12:43
Control Test	0.081	12:43
Air Blank	0.000	12:44
Diagnostics Check	OK	12:44

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTY,EMILY,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WHITTY,EMILY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: HOLLIFIELD

DATE/TIME OF TEST: 10/05/2023 15:49:20

Diagnostics Check	OK	15:43
Air Blank	0.000	15:43
Control Test	0.081	15:44
Air Blank	0.000	15:44
Subject Sample # 1	0.000	15:45
Breath Volume	1.554	15:45
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 2	0.000	15:47
Breath Volume	1.429	15:47
Air Blank	0.000	15:48
Control Test	0.080	15:48
Air Blank	0.000	15:49
Diagnostics Check	OK	15:49

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CRUZ MELCHOR

DATE/TIME OF TEST: 10/08/2023 00:28:21

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.081	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.189	00:23
Breath Volume	2.042	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.185	00:26
Breath Volume	2.296	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: FOWLES

DATE/TIME OF TEST: 10/08/2023 01:01:03

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:56
Breath Volume	1.351	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	00:59
Breath Volume	2.601	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:01

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YOCUM,C
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: TORRES-VELASCO

DATE/TIME OF TEST: 10/08/2023 05:45:31

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:41	
Control Test	0.081	05:42	
Air Blank	0.000	05:42	
Subject Sample	# 1	05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Diagnostics Check	OK	05:45	

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: YOC

DATE/TIME OF TEST: 10/08/2023 21:42:30

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.150	21:37
Breath Volume	1.484	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.150	21:40
Breath Volume	1.214	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 10/09/2023 02:18:07

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.204	02:13
Breath Volume	2.578	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.202	02:16
Breath Volume	2.570	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DIEGO-MIGUEL

DATE/TIME OF TEST: 10/14/2023 00:34:15

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.207	00:27	
Breath Volume	1.968	00:27	
Air Blank	0.000	00:28	
Air Blank	0.000	00:30	
Subject Sample	# 2	00:32	Subject Test Refused
Breath Volume	0.210	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KINSEY

DATE/TIME OF TEST: 10/15/2023 00:17:05

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.290	00:12
Breath Volume	2.160	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.292	00:15
Breath Volume	2.726	00:15
Air Blank	0.000	00:16
Control Test	0.078	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STILTNER,JARRED,T
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KOGER,AARON
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: BUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: KROEGER

DATE/TIME OF TEST: 10/15/2023 21:53:07

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.081	21:48
Breath Volume	2.433	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.085	21:51
Breath Volume	2.136	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 10/18/2023 02:33:12

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.183	02:28
Breath Volume	2.542	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.184	02:31
Breath Volume	1.687	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: CAAL POP

DATE/TIME OF TEST: 10/18/2023 22:56:08

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.161	22:51
Breath Volume	2.218	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:53
Subject Sample # 2	0.159	22:54
Breath Volume	1.886	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 10/19/2023 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: AJEATAS

DATE/TIME OF TEST: 10/21/2023 04:28:15

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.119	04:23
Breath Volume	2.632	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.121	04:26
Breath Volume	2.984	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/23/2023 13:57:58

Diagnostics Check	OK	13:49
Air Blank	0.000	13:50
Control Test	0.080	13:50
Air Blank	0.000	13:51
Subject Sample # 1	0.222	13:53
Breath Volume	1.503	13:53
Air Blank	0.000	13:53
Air Blank	0.000	13:55
Subject Sample # 2	0.232	13:56
Breath Volume	3.601	13:56
Air Blank	0.000	13:57
Control Test	0.079	13:57
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ARCHELUS

DATE/TIME OF TEST: 10/24/2023 06:39:50

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.117	06:35
Breath Volume	1.421	06:35
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.114	06:38
Breath Volume	2.070	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: OBER

DATE/TIME OF TEST: 10/27/2023 23:56:55

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.130	23:52
Breath Volume	2.363	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.131	23:55
Breath Volume	1.800	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: BRANN

DATE/TIME OF TEST: 10/28/2023 08:04:27

Diagnostics Check	OK	08:01	
Air Blank	0.000	08:01	
Control Test	0.080	08:02	
Air Blank	0.000	08:02	
Subject Sample	# 1	08:03	Subject Test Refused
Breath Volume	0.000	08:03	
Air Blank	0.000	08:03	
Control Test	0.080	08:03	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 10/29/2023 01:15:08

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.150	01:10
Breath Volume	1.144	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.152	01:13
Breath Volume	1.253	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 11/16/2023 20:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOERING,C,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MOERING,C
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MALMROS

DATE/TIME OF TEST: 10/06/2023 22:18:38

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.077	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.094	22:14
Breath Volume	2.695	22:14
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.095	22:17
Breath Volume	3.253	22:17
Air Blank	0.000	22:17
Control Test	0.077	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 10/25/2023 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEY-SHADE,W,T
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: COFFEY-SHADE,W
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LATTIN

DATE/TIME OF TEST: 10/19/2023 19:01:33

Diagnostics Check	OK	18:49	
Air Blank	0.000	18:49	
Control Test	0.079	18:49	
Air Blank	0.000	18:50	
Subject Sample	# 1 0.131	18:53	Volume Not Met
Breath Volume	0.996	18:53	
Air Blank	0.000	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 2 0.130	18:56	
Breath Volume	1.503	18:56	
Air Blank	0.000	18:57	
Air Blank	0.000	18:59	
Subject Sample	# 3 0.146	18:59	
Breath Volume	1.585	18:59	
Air Blank	0.000	19:00	
Control Test	0.073	19:00	Control Outside Tolerance
Air Blank	0.000	19:01	

Data Download Date/Time: 10/25/2023 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEY-SHADE,W,T
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: COFFEY-SHADE,W
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LATTIN

DATE/TIME OF TEST: 10/19/2023 19:19:49

Diagnostics Check	OK	19:13
Air Blank	0.000	19:14
Control Test	0.078	19:14
Air Blank	0.000	19:14
Subject Sample # 1	0.142	19:15
Breath Volume	2.238	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:17
Subject Sample # 2	0.144	19:18
Breath Volume	1.921	19:18
Air Blank	0.000	19:18
Control Test	0.076	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 10/25/2023 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 09/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 10/21/2023 01:40:00

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.126	01:35
Breath Volume	1.148	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:37
Subject Sample # 2	0.120	01:38
Breath Volume	1.878	01:38
Air Blank	0.000	01:39
Control Test	0.077	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/25/2023 07:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIALOBRZESKI,M
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: LEBAR

DATE/TIME OF TEST: 10/02/2023 18:25:53

Diagnostics Check	OK	18:14	
Air Blank	0.000	18:14	
Control Test	0.077	18:15	
Air Blank	0.000	18:15	
Subject Sample	# 1 0.357	18:18	Volume Not Met
Breath Volume	0.378	18:18	
Air Blank	0.000	18:19	
Air Blank	0.000	18:21	
Subject Sample	# 2	18:24	No Sample Provided
Breath Volume	0.000	18:24	
Air Blank	0.000	18:24	
Control Test	0.073	18:25	Control Outside Tolerance
Air Blank	0.000	18:25	

Data Download Date/Time: 10/17/2023 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 10/15/2023 04:21:43

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.167	04:16
Breath Volume	1.421	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.170	04:20
Breath Volume	2.425	04:20
Air Blank	0.000	04:20
Control Test	0.075	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 10/17/2023 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YANICK,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WOOTEN,HANNAH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 10/29/2023 01:39:32

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.168	01:34
Breath Volume	2.175	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.167	01:37
Breath Volume	2.371	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/30/2023 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY, MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: KIRKSEY

DATE/TIME OF TEST: 10/29/2023 07:04:45

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.080	06:57
Air Blank	0.000	06:57
Subject Sample # 1	0.225	06:59
Breath Volume	1.195	06:59
Air Blank	0.000	07:00
Air Blank	0.000	07:01
Subject Sample # 2	0.219	07:03
Breath Volume	1.542	07:03
Air Blank	0.000	07:03
Control Test	0.077	07:04
Air Blank	0.000	07:04
Diagnostics Check	OK	07:04

Data Download Date/Time: 11/30/2023 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ABOY

DATE/TIME OF TEST: 10/21/2023 02:08:05

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.082	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.119	02:02
Breath Volume	2.523	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.116	02:06
Breath Volume	3.425	02:06
Air Blank	0.000	02:07
Control Test	0.078	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/30/2023 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RAMIREZ GARCIA,HERNAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: CANO JIMENEZ

DATE/TIME OF TEST: 10/22/2023 20:56:46

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.083	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.206	20:51
Breath Volume	2.812	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.211	20:54
Breath Volume	2.707	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 11/30/2023 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MOWERY

DATE/TIME OF TEST: 10/01/2023 03:15:23

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.127	03:10
Breath Volume	1.171	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.126	03:13
Breath Volume	1.585	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 10/05/2023 02:02:06

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.077	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.182	01:57
Breath Volume	2.183	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.172	02:00
Breath Volume	2.367	02:00
Air Blank	0.000	02:01
Control Test	0.075	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NOBORE,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: TINAYNAN

DATE/TIME OF TEST: 10/05/2023 03:19:42

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.077	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.120	03:14
Breath Volume	2.449	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.117	03:18
Breath Volume	3.191	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/06/2023 21:26:46

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.078	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:22	No Sample Provided
Breath Volume	0.000	21:22	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2	21:25	Subject Test Refused
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Control Test	0.078	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCQUEEN,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/06/2023 23:17:15

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.078	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:16	
Control Test	0.078	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 10/07/2023 05:15:37

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.150	05:10
Breath Volume	1.148	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.169	05:13
Breath Volume	2.343	05:13
Air Blank	0.000	05:14
Control Test	0.078	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 10/07/2023 22:04:41

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.078	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.064	21:59
Breath Volume	3.203	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.051	22:03
Breath Volume	1.382	22:03
Air Blank	0.000	22:03
Control Test	0.076	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCQUEEN,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: WARDEN

DATE/TIME OF TEST: 10/07/2023 23:34:31

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.079	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKCLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JOHNSEN

DATE/TIME OF TEST: 10/13/2023 03:10:43

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.238	03:05
Breath Volume	2.621	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.232	03:09
Breath Volume	1.878	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/14/2023 21:30:47

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.078	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.000	21:26
Breath Volume	2.925	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.000	21:29
Breath Volume	3.023	21:29
Air Blank	0.000	21:29
Control Test	0.078	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: BRACKINS

DATE/TIME OF TEST: 10/15/2023 03:51:23

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.145	03:46
Breath Volume	1.882	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.140	03:49
Breath Volume	1.109	03:49
Air Blank	0.000	03:50
Control Test	0.077	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARIN,M,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DEMAY

DATE/TIME OF TEST: 10/17/2023 21:02:54

Diagnostics Check	OK	20:59	
Air Blank	0.000	21:00	
Control Test	0.078	21:00	
Air Blank	0.000	21:01	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.079	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: WELSHANS

DATE/TIME OF TEST: 10/18/2023 13:34:02

Diagnostics Check	OK	13:24
Air Blank	0.000	13:25
Control Test	0.078	13:25
Air Blank	0.000	13:26
Subject Sample # 1	0.114	13:27
Breath Volume	1.242	13:27
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.114	13:32
Breath Volume	1.285	13:32
Air Blank	0.000	13:33
Control Test	0.079	13:33
Air Blank	0.000	13:33
Diagnostics Check	OK	13:34

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNANT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 09/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: FEAGANS

DATE/TIME OF TEST: 10/22/2023 08:54:06

Diagnostics Check	OK	08:41	
Air Blank	0.000	08:42	
Control Test	0.078	08:42	
Air Blank	0.000	08:42	
Subject Sample	# 1 0.275	08:46	Volume Not Met
Breath Volume	0.496	08:46	
Air Blank	0.000	08:46	
Air Blank	0.000	08:48	
Subject Sample	# 2 0.281	08:49	
Breath Volume	2.296	08:49	
Air Blank	0.000	08:49	
Air Blank	0.000	08:51	
Subject Sample	# 3 0.264	08:52	
Breath Volume	2.558	08:52	
Air Blank	0.000	08:53	
Control Test	0.076	08:53	
Air Blank	0.000	08:54	
Diagnostics Check	OK	08:54	

Data Download Date/Time: 10/23/2023 07:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,JOHN,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SNOW,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 10/23/2023 12:33:00

Diagnostics Check	OK	12:21	
Air Blank	0.000	12:22	
Control Test	0.079	12:22	
Air Blank	0.000	12:22	
Subject Sample	# 1	12:26	No Sample Provided
Breath Volume	0.000	12:26	
Air Blank	0.000	12:26	
Air Blank	0.000	12:28	
Subject Sample	# 2	12:31	No Sample Provided
Breath Volume	0.000	12:31	
Air Blank	0.000	12:32	
Control Test	0.078	12:32	
Air Blank	0.000	12:32	
Diagnostics Check	OK	12:33	

Data Download Date/Time: 11/26/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 10/24/2023 21:03:01

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.078	20:52	
Air Blank	0.000	20:52	
Subject Sample	# 1 0.158	20:56	Volume Not Met
Breath Volume	0.128	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:02	
Control Test	0.077	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:03	

Data Download Date/Time: 11/26/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDED RRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HOUK

DATE/TIME OF TEST: 10/26/2023 02:33:32

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:29	No Sample Provided
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample	# 2	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.078	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 11/26/2023 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSLEY,J,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WITTREICH

DATE/TIME OF TEST: 10/28/2023 03:33:33

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.078	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:29	No Sample Provided
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 11/26/2023 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 10/29/2023 06:24:55

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.078	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.128	06:19
Breath Volume	2.007	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.131	06:23
Breath Volume	2.351	06:23
Air Blank	0.000	06:24
Control Test	0.078	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 11/26/2023 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVIS, MILJEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 10/20/2023 20:16:55

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.000	20:12
Breath Volume	2.847	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.000	20:15
Breath Volume	3.128	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PEREZ-ORDONEZ

DATE/TIME OF TEST: 10/21/2023 01:17:31

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.172	01:12
Breath Volume	1.714	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.166	01:15
Breath Volume	1.648	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: MERAZ CABRERA

DATE/TIME OF TEST: 10/21/2023 18:01:31

Diagnostics Check	OK	17:54
Air Blank	0.000	17:55
Control Test	0.081	17:55
Air Blank	0.000	17:56
Subject Sample # 1	0.110	17:57
Breath Volume	1.312	17:57
Air Blank	0.000	17:57
Air Blank	0.000	17:59
Subject Sample # 2	0.112	17:59
Breath Volume	1.500	17:59
Air Blank	0.000	18:00
Control Test	0.080	18:00
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OLSEN,JORDAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: IGLESIAS

DATE/TIME OF TEST: 10/21/2023 21:55:31

Diagnostics Check	OK	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.166	21:50
Breath Volume	2.781	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.165	21:53
Breath Volume	2.953	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PEREZ ESCALANTE

DATE/TIME OF TEST: 10/21/2023 22:26:01

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.079	22:12	
Air Blank	0.000	22:13	
Subject Sample # 1	0.286	22:14	
Breath Volume	1.425	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample # 2	0.242	22:20	Volume Not Met
Breath Volume	0.304	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:22	
Subject Sample # 3		22:24	Subject Test Refused
Breath Volume	0.324	22:24	
Air Blank	0.000	22:25	
Control Test	0.075	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: TAMAYAC BOLIVAR

DATE/TIME OF TEST: 10/22/2023 01:41:32

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.160	01:37
Breath Volume	3.855	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.145	01:39
Breath Volume	4.066	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRATEREAUX,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 10/22/2023 06:54:40

Diagnostics Check	OK	06:48
Air Blank	0.000	06:48
Control Test	0.080	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.157	06:50
Breath Volume	2.511	06:50
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.158	06:53
Breath Volume	2.855	06:53
Air Blank	0.000	06:53
Control Test	0.079	06:54
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 10/23/2023 00:26:46

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.078	00:22
Breath Volume	1.656	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.077	00:25
Breath Volume	1.738	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DE LA CRUZ JAVIER

DATE/TIME OF TEST: 10/23/2023 04:13:23

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.000	04:08
Breath Volume	1.273	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.000	04:12
Breath Volume	1.246	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DORIVAL,ADLER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: DEJANUS

DATE/TIME OF TEST: 10/23/2023 06:39:16

Diagnostics Check	OK	06:32
Air Blank	0.000	06:33
Control Test	0.080	06:33
Air Blank	0.000	06:34
Subject Sample # 1	0.176	06:34
Breath Volume	1.664	06:34
Air Blank	0.000	06:35
Air Blank	0.000	06:37
Subject Sample # 2	0.180	06:37
Breath Volume	1.816	06:37
Air Blank	0.000	06:38
Control Test	0.081	06:38
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CEDILLO DE LA CRUZ

DATE/TIME OF TEST: 10/26/2023 02:23:41

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.030	02:19
Breath Volume	2.093	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.033	02:22
Breath Volume	2.820	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON, WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/26/2023 23:43:02

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.178	23:37
Breath Volume	2.210	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.179	23:41
Breath Volume	4.046	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SURIEL

DATE/TIME OF TEST: 10/28/2023 01:42:57

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.189	01:38
Breath Volume	3.917	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.178	01:41
Breath Volume	1.453	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SABBATINO

DATE/TIME OF TEST: 10/29/2023 01:01:57

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Control Test	0.080	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MAGANA

DATE/TIME OF TEST: 10/29/2023 05:08:35

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.155	05:03
Breath Volume	1.312	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.170	05:07
Breath Volume	2.539	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MAYHOOK

DATE/TIME OF TEST: 10/29/2023 20:00:25

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.080	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.152	19:54
Breath Volume	1.246	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.148	19:58
Breath Volume	1.351	19:58
Air Blank	0.000	19:59
Control Test	0.079	19:59
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/29/2023 20:27:58

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.020	20:23
Breath Volume	2.769	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.017	20:26
Breath Volume	1.648	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MARCOS GONZALEZ

DATE/TIME OF TEST: 10/29/2023 21:32:29

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.205	21:27
Breath Volume	1.984	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.198	21:30
Breath Volume	1.179	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: SCHEMBRI

DATE/TIME OF TEST: 10/30/2023 19:26:25

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.080	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:25	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RINI,LAWRENCE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: MORRA

DATE/TIME OF TEST: 10/31/2023 16:50:05

Diagnostics Check	OK	16:38	
Air Blank	0.000	16:38	
Control Test	0.080	16:39	
Air Blank	0.000	16:39	
Subject Sample	# 1 0.092	16:42	Volume Not Met
Breath Volume	0.812	16:42	
Air Blank	0.000	16:43	
Air Blank	0.000	16:45	
Subject Sample	# 2 0.095	16:45	
Breath Volume	1.207	16:45	
Air Blank	0.000	16:46	
Air Blank	0.000	16:48	
Subject Sample	# 3 0.096	16:48	
Breath Volume	1.199	16:48	
Air Blank	0.000	16:49	
Control Test	0.079	16:49	
Air Blank	0.000	16:50	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 11/17/2023 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: QUINTANA REYES

DATE/TIME OF TEST: 10/01/2023 00:54:37

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.163	00:50
Breath Volume	1.570	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.164	00:53
Breath Volume	1.503	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GABRIEL HERNANDEZ

DATE/TIME OF TEST: 10/01/2023 02:45:39

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.194	02:41
Breath Volume	1.125	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.186	02:44
Breath Volume	2.199	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MERCEDES DE JESUS

DATE/TIME OF TEST: 10/01/2023 06:05:47

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.080	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.168	06:01
Breath Volume	1.492	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.169	06:04
Breath Volume	2.105	06:04
Air Blank	0.000	06:04
Control Test	0.080	06:05
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: PETTIS

DATE/TIME OF TEST: 10/01/2023 22:45:58

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.173	22:41
Breath Volume	1.472	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.170	22:44
Breath Volume	1.343	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOVETT,JAMES
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MARTINEZ JR

DATE/TIME OF TEST: 10/02/2023 02:00:48

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.257	01:56
Breath Volume	3.824	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.239	01:59
Breath Volume	2.593	01:59
Air Blank	0.000	01:59
Control Test	0.075	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLAKE, XAVIAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:19

SUBJECT'S LAST NAME: MENENDEZ JR

DATE/TIME OF TEST: 10/02/2023 10:54:03

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.081	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.131	10:48
Breath Volume	1.632	10:48
Air Blank	0.000	10:49
Air Blank	0.000	10:51
Subject Sample # 2	0.131	10:52
Breath Volume	1.285	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:53
Diagnostics Check	OK	10:54

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: EL YACOUBI

DATE/TIME OF TEST: 10/04/2023 05:05:44

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.144	05:01
Breath Volume	1.500	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.154	05:04
Breath Volume	1.261	05:04
Air Blank	0.000	05:04
Control Test	0.081	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: PAIZ-LOPEZ

DATE/TIME OF TEST: 10/06/2023 00:26:30

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.237	00:21
Breath Volume	2.976	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.237	00:24
Breath Volume	3.171	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VASQUEZ,FELIPE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 10/06/2023 03:54:44

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.233	03:50
Breath Volume	2.292	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.231	03:53
Breath Volume	3.339	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: GOREN

DATE/TIME OF TEST: 10/06/2023 20:17:41

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.081	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:10	No Sample Provided
Breath Volume	0.000	20:10	
Air Blank	0.000	20:11	
Air Blank	0.000	20:13	
Subject Sample	# 2	20:16	No Sample Provided
Breath Volume	0.000	20:16	
Air Blank	0.000	20:16	
Control Test	0.081	20:17	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: GOREN

DATE/TIME OF TEST: 10/06/2023 20:38:39

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.082	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.132	20:33
Breath Volume	1.753	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.125	20:37
Breath Volume	1.878	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: GALINDO

DATE/TIME OF TEST: 10/07/2023 02:01:55

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.143	01:57
Breath Volume	1.585	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.146	02:00
Breath Volume	2.660	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BORJA ORROSTIETA

DATE/TIME OF TEST: 10/07/2023 02:58:40

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.131	02:54
Breath Volume	2.886	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.134	02:57
Breath Volume	2.746	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRATEREAUX,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: MARTINEZ GARCIA

DATE/TIME OF TEST: 10/07/2023 08:16:19

Diagnostics Check	OK	08:09
Air Blank	0.000	08:10
Control Test	0.080	08:10
Air Blank	0.000	08:10
Subject Sample # 1	0.131	08:11
Breath Volume	3.656	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.133	08:14
Breath Volume	4.089	08:14
Air Blank	0.000	08:15
Control Test	0.081	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ANDRES RAMIREZ

DATE/TIME OF TEST: 10/08/2023 02:05:14

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.209	02:00
Breath Volume	2.960	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.194	02:03
Breath Volume	1.371	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NALERIO,LUIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/08/2023 03:31:34

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.169	03:27
Breath Volume	1.671	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.161	03:30
Breath Volume	1.292	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: DUQUE NUNEZ

DATE/TIME OF TEST: 10/08/2023 06:06:07

Diagnostics Check	OK	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:00
Subject Sample # 1	0.148	06:01
Breath Volume	1.207	06:01
Air Blank	0.000	06:02
Air Blank	0.000	06:03
Subject Sample # 2	0.145	06:04
Breath Volume	1.773	06:04
Air Blank	0.000	06:05
Control Test	0.081	06:05
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAMILTON,CORY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HARNED

DATE/TIME OF TEST: 10/08/2023 21:50:43

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.082	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.139	21:44
Breath Volume	1.953	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.132	21:49
Breath Volume	1.523	21:49
Air Blank	0.000	21:49
Control Test	0.081	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/08/2023 23:05:17

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.258	23:00
Breath Volume	1.960	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.259	23:03
Breath Volume	3.902	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BARRIOS GONZALEZ

DATE/TIME OF TEST: 10/08/2023 23:52:33

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.227	23:47
Breath Volume	2.632	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.208	23:50
Breath Volume	1.933	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REID,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PABLO VELASQUEZ

DATE/TIME OF TEST: 10/09/2023 01:17:44

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.208	01:12
Breath Volume	3.761	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.204	01:16
Breath Volume	3.054	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,KERWIN
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: HALICKI

DATE/TIME OF TEST: 10/09/2023 15:32:45

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.080	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.176	15:28
Breath Volume	3.980	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.174	15:31
Breath Volume	4.691	15:31
Air Blank	0.000	15:31
Control Test	0.080	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SCULLY DOOLEY

DATE/TIME OF TEST: 10/10/2023 04:03:07

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.212	03:56
Breath Volume	1.292	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.213	04:01
Breath Volume	1.449	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY, CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GAETA

DATE/TIME OF TEST: 10/10/2023 22:58:11

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.265	22:53
Breath Volume	1.128	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.264	22:56
Breath Volume	1.574	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: JIMENEZ PEREZ

DATE/TIME OF TEST: 10/10/2023 23:46:02

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.081	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.218	23:41
Breath Volume	2.574	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.220	23:44
Breath Volume	1.847	23:44
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:46

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 10/11/2023 05:18:10

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.171	05:13
Breath Volume	1.281	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.179	05:16
Breath Volume	1.425	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BADOLATI, NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: PETRON

DATE/TIME OF TEST: 10/11/2023 21:47:15

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.165	21:40	Volume Not Met
Breath Volume	0.906	21:40	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2 0.169	21:45	Volume Not Met
Breath Volume	0.035	21:45	
Air Blank	0.000	21:46	
Control Test	0.079	21:46	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: ZIEMINSKI

DATE/TIME OF TEST: 10/13/2023 20:44:45

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.149	20:40
Breath Volume	1.250	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.149	20:43
Breath Volume	1.644	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 10/14/2023 22:34:50

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.192	22:30
Breath Volume	1.812	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.196	22:33
Breath Volume	2.062	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ANDRES RAYMUNDO

DATE/TIME OF TEST: 10/15/2023 03:55:43

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.144	03:50
Breath Volume	1.480	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.149	03:54
Breath Volume	2.035	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROWE,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: MENDOZA SANCHEZ

DATE/TIME OF TEST: 10/15/2023 09:45:17

Diagnostics Check	OK	09:38
Air Blank	0.000	09:38
Control Test	0.080	09:39
Air Blank	0.000	09:39
Subject Sample # 1	0.219	09:40
Breath Volume	2.546	09:40
Air Blank	0.000	09:41
Air Blank	0.000	09:42
Subject Sample # 2	0.214	09:43
Breath Volume	3.828	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Diagnostics Check	OK	09:45

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: D'ANGELO IV

DATE/TIME OF TEST: 10/15/2023 18:48:43

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.100	18:44
Breath Volume	2.472	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.097	18:47
Breath Volume	2.566	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/15/2023 20:10:21

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.266	20:05
Breath Volume	1.734	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.273	20:08
Breath Volume	2.500	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 10/16/2023 03:29:08

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.204	03:24
Breath Volume	1.375	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.205	03:27
Breath Volume	1.562	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MELENDEZ ESCOBAR

DATE/TIME OF TEST: 10/17/2023 22:44:22

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:35
Subject Sample # 1	0.230	22:36
Breath Volume	2.921	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.204	22:39
Breath Volume	1.660	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 3	0.205	22:42
Breath Volume	1.988	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 10/18/2023 00:26:31

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.175	00:22
Breath Volume	1.535	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.178	00:24
Breath Volume	1.769	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 10/20/2023 03:01:18

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.155	02:56
Breath Volume	2.144	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.156	02:59
Breath Volume	2.394	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 10/20/2023 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HIGHTOWER

DATE/TIME OF TEST: 10/21/2023 06:00:49

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.000	05:55
Breath Volume	1.765	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.000	05:59
Breath Volume	1.339	05:59
Air Blank	0.000	05:59
Control Test	0.081	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ, ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RICARDO ENRIQUE

DATE/TIME OF TEST: 10/22/2023 00:36:38

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.209	00:28
Breath Volume	2.136	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.231	00:32
Breath Volume	2.105	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 3	0.218	00:35
Breath Volume	1.882	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MELGAR ALCANTARA

DATE/TIME OF TEST: 10/22/2023 01:35:45

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.230	01:31
Breath Volume	2.406	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.225	01:34
Breath Volume	2.140	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANCHEZ, JULIAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: RENEUS

DATE/TIME OF TEST: 10/22/2023 05:51:30

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.081	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.162	05:47
Breath Volume	1.964	05:47
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.162	05:49
Breath Volume	1.500	05:49
Air Blank	0.000	05:50
Control Test	0.081	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LANDA, BRISA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: SEWELL

DATE/TIME OF TEST: 10/22/2023 07:42:01

Diagnostics Check	OK	07:35
Air Blank	0.000	07:36
Control Test	0.081	07:36
Air Blank	0.000	07:37
Subject Sample # 1	0.194	07:37
Breath Volume	2.347	07:37
Air Blank	0.000	07:38
Air Blank	0.000	07:40
Subject Sample # 2	0.196	07:40
Breath Volume	2.597	07:40
Air Blank	0.000	07:41
Control Test	0.080	07:41
Air Blank	0.000	07:41
Diagnostics Check	OK	07:42

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DERBY,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 10/22/2023 11:28:26

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.081	11:21
Air Blank	0.000	11:21
Subject Sample # 1	0.044	11:22
Breath Volume	1.722	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.041	11:26
Breath Volume	1.343	11:26
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:28
Diagnostics Check	OK	11:28

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ROMAN VEGA

DATE/TIME OF TEST: 10/23/2023 02:27:35

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.155	02:23
Breath Volume	1.617	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.154	02:26
Breath Volume	1.601	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO, DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/23/2023 08:18:14

Diagnostics Check	OK	08:08
Air Blank	0.000	08:09
Control Test	0.081	08:09
Air Blank	0.000	08:10
Subject Sample # 1	0.166	08:12
Breath Volume	2.160	08:12
Air Blank	0.000	08:13
Air Blank	0.000	08:15
Subject Sample # 2	0.160	08:16
Breath Volume	1.960	08:16
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CORNETTE

DATE/TIME OF TEST: 10/26/2023 23:34:05

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.081	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:29
Breath Volume	3.007	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:32
Breath Volume	3.574	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HEARN,DEREK
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GIRON LOPEZ

DATE/TIME OF TEST: 10/27/2023 00:23:27

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.232	00:18
Breath Volume	1.542	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.248	00:21
Breath Volume	1.511	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ONG,ANDY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: RAMIREZ HERNANDEZ

DATE/TIME OF TEST: 10/27/2023 05:36:05

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.180	05:31
Breath Volume	2.765	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.173	05:34
Breath Volume	2.539	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: HALPERIN

DATE/TIME OF TEST: 10/27/2023 23:51:41

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.159	23:46
Breath Volume	1.257	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.161	23:50
Breath Volume	1.125	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: RANTINELLA

DATE/TIME OF TEST: 10/28/2023 00:12:18

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.000	00:08
Breath Volume	3.273	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.000	00:10
Breath Volume	3.164	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/28/2023 01:52:17

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.197	01:47
Breath Volume	3.285	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.199	01:50
Breath Volume	3.117	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROBLERO PEREZ

DATE/TIME OF TEST: 10/28/2023 04:34:28

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.231	04:29
Breath Volume	1.238	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.229	04:32
Breath Volume	2.062	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: TAMAYO ELIZONDO

DATE/TIME OF TEST: 10/28/2023 23:22:02

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.293	23:17
Breath Volume	2.968	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.298	23:20
Breath Volume	2.527	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:22

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RENTERIA,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: OMEIR

DATE/TIME OF TEST: 10/29/2023 00:47:39

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.075	00:41
Breath Volume	1.519	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.076	00:46
Breath Volume	2.144	00:46
Air Blank	0.000	00:46
Control Test	0.081	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 10/29/2023 01:28:07

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.114	01:23
Breath Volume	2.023	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.117	01:26
Breath Volume	2.425	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CARDENAS CRESPO

DATE/TIME OF TEST: 10/29/2023 02:21:52

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.181	02:17
Breath Volume	2.644	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.180	02:20
Breath Volume	2.605	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ, ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 10/30/2023 00:08:31

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.243	00:03
Breath Volume	1.363	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.230	00:06
Breath Volume	1.484	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GONZALEZ OTERO

DATE/TIME OF TEST: 10/30/2023 04:03:30

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.205	03:53
Breath Volume	1.773	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.227	03:56
Breath Volume	3.621	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 3	0.223	04:01
Breath Volume	3.042	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 10/30/2023 23:18:47

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.080	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.161	23:14
Breath Volume	3.402	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.162	23:17
Breath Volume	3.550	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: CARDENAS SANCHEZ

DATE/TIME OF TEST: 10/31/2023 01:26:58

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.202	01:22
Breath Volume	1.925	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.206	01:25
Breath Volume	1.582	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 10/31/2023 04:33:18

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.137	04:28
Breath Volume	1.246	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.139	04:31
Breath Volume	1.570	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 11/17/2023 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BURMAA

DATE/TIME OF TEST: 10/29/2023 04:37:21

Diagnostics Check	OK	04:36	
Air Blank		04:36	Ambient Fail
Air Blank	0.000	04:37	

Data Download Date/Time: 12/05/2023 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BURMAA

DATE/TIME OF TEST: 10/29/2023 04:40:15

Diagnostics Check	OK	04:39	
Air Blank		04:39	Ambient Fail
Air Blank	0.000	04:40	

Data Download Date/Time: 12/05/2023 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FHP TROOP F

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLAKE,DAVID,A

BREATH TEST OPERATOR AGENCY: FHP Troop F

ARRESTING OFFICER: HASH,CONNOR

ARRESTING OFFICER AGENCY: FHP Troop F

LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 3024023571

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 04:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BURMAA

DATE/TIME OF TEST: 10/29/2023 04:43:56

Diagnostics Check	OK	04:43	
Air Blank		04:43	Ambient Fail
Air Blank	0.000	04:43	

Data Download Date/Time: 12/05/2023 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 10/03/2023 22:30:41

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.079	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.155	22:25
Breath Volume	1.656	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:27
Subject Sample # 2	0.162	22:29
Breath Volume	2.863	22:29
Air Blank	0.000	22:29
Control Test	0.078	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 10/19/2023 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MORETTI

DATE/TIME OF TEST: 10/08/2023 22:55:36

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.082	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.179	22:50
Breath Volume	1.429	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.194	22:53
Breath Volume	2.140	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/19/2023 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ISENHOUR

DATE/TIME OF TEST: 10/20/2023 00:05:21

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.050	00:00
Breath Volume	3.789	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.051	00:03
Breath Volume	3.347	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 11/17/2023 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: KUCHEROVA KOUZINE

DATE/TIME OF TEST: 10/21/2023 03:14:41

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.236	03:10
Breath Volume	1.453	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.237	03:13
Breath Volume	1.375	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 11/17/2023 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BREEDLOVE

DATE/TIME OF TEST: 10/22/2023 01:41:34

Diagnostics Check	OK	01:40	
Air Blank		01:41	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 11/17/2023 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BREEDLOVE

DATE/TIME OF TEST: 10/22/2023 01:49:31

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.083	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.151	01:45
Breath Volume	1.941	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.148	01:47
Breath Volume	2.054	01:47
Air Blank	0.000	01:48
Control Test	0.082	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 11/17/2023 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 10/26/2023 01:48:53

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.060	01:44
Breath Volume	2.109	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.059	01:47
Breath Volume	1.730	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 11/17/2023 19:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DAMON,RUSSELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HIGHAM

DATE/TIME OF TEST: 10/20/2023 23:52:20

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.083	23:41	
Air Blank	0.000	23:41	
Subject Sample	# 1 0.186	23:44	Volume Not Met
Breath Volume	0.117	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2 0.192	23:50	Volume Not Met
Breath Volume	0.289	23:50	
Air Blank	0.000	23:51	
Control Test	0.081	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 11/16/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/21/2023 02:26:23

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.128	02:21
Breath Volume	3.539	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.126	02:24
Breath Volume	3.269	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/16/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SOWARDS,JACOB
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 10/08/2023 00:24:26

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SCHAEFFER,WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: RICHEY

DATE/TIME OF TEST: 10/14/2023 03:20:27

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: WARMACK

DATE/TIME OF TEST: 10/15/2023 07:06:37

Diagnostics Check	OK	06:56	
Air Blank	0.000	06:57	
Control Test	0.078	06:57	
Air Blank	0.000	06:58	
Subject Sample	# 1	07:01	No Sample Provided
Breath Volume	0.000	07:01	
Air Blank	0.000	07:01	
Air Blank	0.000	07:03	
Subject Sample	# 2	07:05	Subject Test Refused
Breath Volume	0.000	07:05	
Air Blank	0.000	07:05	
Control Test	0.078	07:05	
Air Blank	0.000	07:06	
Diagnostics Check	OK	07:06	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/21/2023 00:07:31

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.078	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.128	00:02
Breath Volume	2.218	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.123	00:05
Breath Volume	1.941	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING,EMMA,G
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SPAULDING,EMMA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 10/22/2023 16:45:41

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:40	
Control Test	0.079	16:40	
Air Blank	0.000	16:41	
Subject Sample	# 1 0.202	16:44	Volume Not Met
Breath Volume	0.031	16:44	
Air Blank		16:45	Purge Fail
Air Blank	0.000	16:45	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING,EMMA,G
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SPAULDING,EMMA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 10/22/2023 16:50:30

Diagnostics Check	OK	16:49	
Air Blank	0.000	16:49	
Control Test	0.074	16:49	Control Outside Tolerance
Air Blank	0.000	16:50	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPAULDING,EMMA,G
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SPAULDING,EMMA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 10/22/2023 17:04:17

Diagnostics Check	OK	16:52	
Air Blank	0.000	16:52	
Control Test	0.079	16:53	
Air Blank	0.000	16:53	
Subject Sample	# 1 0.158	16:56	Volume Not Met
Breath Volume	0.730	16:56	
Air Blank	0.000	16:57	
Air Blank	0.000	16:59	
Subject Sample	# 2 0.164	17:02	Volume Not Met
Breath Volume	0.175	17:02	
Air Blank	0.000	17:03	
Control Test	0.074	17:03	Control Outside Tolerance
Air Blank	0.000	17:04	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUYTON,JOHN,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29821080A3
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MAST

DATE/TIME OF TEST: 10/23/2023 02:58:11

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.078	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:53	No Sample Provided
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Air Blank	0.000	02:56	
Subject Sample	# 2	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 10/23/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,CHARLES,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MITCHELL,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: NEEL

DATE/TIME OF TEST: 10/06/2023 06:59:22

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.080	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.223	06:54
Breath Volume	1.351	06:54
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.220	06:57
Breath Volume	2.226	06:57
Air Blank	0.000	06:58
Control Test	0.080	06:58
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 10/11/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSMAN,MATTHEW,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HAUSMAN,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/07/2023 03:22:02

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 10/11/2023 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT, COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 07:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: FERRERA

DATE/TIME OF TEST: 10/28/2023 07:49:40

Diagnostics Check	OK	07:43
Air Blank	0.000	07:43
Control Test	0.080	07:43
Air Blank	0.000	07:44
Subject Sample # 1	0.097	07:45
Breath Volume	3.222	07:45
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.102	07:48
Breath Volume	2.687	07:48
Air Blank	0.000	07:48
Control Test	0.079	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 11/21/2023 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: NAVELSKI

DATE/TIME OF TEST: 10/29/2023 04:13:19

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.149	04:08
Breath Volume	1.828	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.154	04:11
Breath Volume	2.285	04:11
Air Blank	0.000	04:12
Control Test	0.076	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/21/2023 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEJARZA,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LEJARZA,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 10/04/2023 21:37:28

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.079	21:35	
Air Blank	0.000	21:35	
Subject Sample	# 1	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.080	21:36	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 10/27/2023 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEJARZA,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LEJARZA,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: SENTI MIRANDA

DATE/TIME OF TEST: 10/05/2023 01:08:58

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.198	01:04
Breath Volume	1.699	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.198	01:07
Breath Volume	1.570	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 10/27/2023 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 03:14
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 09/01/2023	OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HERNANDEZ ROMAN	DATE/TIME OF TEST: 10/08/2023 04:13:56
--------------------------------------	--

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.079	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUZ,ASHLEY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LARA,JOSHUA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CALERO

DATE/TIME OF TEST: 10/15/2023 01:10:17

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOSA,DAYNER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RUBIO,H
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 10/15/2023 06:04:06

Diagnostics Check	OK	05:55	
Air Blank	0.000	05:56	
Control Test	0.079	05:56	
Air Blank	0.000	05:56	
Subject Sample	# 1	06:00	No Sample Provided
Breath Volume	0.000	06:00	
Air Blank	0.000	06:00	
Air Blank	0.000	06:02	
Subject Sample	# 2	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:03	
Control Test	0.079	06:03	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 01:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DELGADO FUENTES

DATE/TIME OF TEST: 10/22/2023 02:55:00

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.124	02:50
Breath Volume	2.539	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.120	02:53
Breath Volume	1.417	02:53
Air Blank	0.000	02:54
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,M,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RIOS RIVA

DATE/TIME OF TEST: 10/22/2023 04:08:10

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.080	04:03
Breath Volume	2.445	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.080	04:06
Breath Volume	2.265	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: WELSCHINGER,MICKAEL,L	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: WELSCHINGER,MICKAEL	ARREST TIME: 01:35
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 09/01/2023	OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ORTEZ GODOY	DATE/TIME OF TEST: 10/27/2023	02:45:32
----------------------------------	-------------------------------	----------

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.105	02:35	Volume Not Met
Breath Volume	0.640	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 2 0.119	02:39	
Breath Volume	2.179	02:39	
Air Blank	0.000	02:40	
Air Blank	0.000	02:41	
Subject Sample	# 3 0.127	02:43	
Breath Volume	1.117	02:43	
Air Blank	0.000	02:44	
Control Test	0.075	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 10/27/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,ERIC,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,ERIC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 10/28/2023 22:16:13

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.079	22:11	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:14	RFI Detect
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 11/27/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,ERIC,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,ERIC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 10/28/2023 22:30:20

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:18	
Control Test	0.081	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1 0.308	22:22	Volume Not Met
Breath Volume	0.371	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2 0.341	22:28	Volume Not Met
Breath Volume	1.187	22:28	
Air Blank	0.000	22:29	
Control Test	0.077	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 11/27/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,ERIC,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,DANELIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/29/2023 02:22:36

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1 0.126	02:15	Volume Not Met
Breath Volume	0.039	02:15	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2 0.123	02:20	Volume Not Met
Breath Volume	0.074	02:20	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 11/27/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPSHUCK,CHRISTINA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: KAPSHUCK,CHRISTINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 10/31/2023 00:19:58

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.078	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 11/27/2023 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 989917
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A	DRY GAS EXPIRY DATE: 07/25/2021
BREATH TEST OPERATOR AGENCY: Miami PD	VIOLATION CODE: DUI
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO	ARREST TIME: 02:24
ARRESTING OFFICER AGENCY: Miami PD	GENDER: Male
LAST AGENCY INSPECTION: 10/27/2023	OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BLANDON CRUZ	DATE/TIME OF TEST: 10/29/2023 03:21:26
-----------------------------------	--

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.244	03:16
Breath Volume	1.632	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.246	03:19
Breath Volume	1.914	03:19
Air Blank	0.000	03:20
Control Test	0.076	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 11/27/2023 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRERO-MARTINEZ,KEVEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MINCEY

DATE/TIME OF TEST: 10/01/2023 01:58:37

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.054	01:53
Breath Volume	1.117	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.056	01:57
Breath Volume	1.167	01:57
Air Blank	0.000	01:57
Control Test	0.080	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 10/04/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: KORMENDI

DATE/TIME OF TEST: 10/03/2023 21:14:19

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.082	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1 0.208	21:07	Volume Not Met
Breath Volume	0.167	21:07	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2	21:12	No Sample Provided
Breath Volume	0.000	21:12	
Air Blank	0.000	21:13	
Control Test	0.078	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 10/04/2023 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: TSANG

DATE/TIME OF TEST: 10/03/2023 22:55:57

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.080	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.123	22:51
Breath Volume	1.453	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.118	22:54
Breath Volume	2.585	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 10/04/2023 16:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGE,DANIEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: EDGE,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: FUZY

DATE/TIME OF TEST: 10/13/2023 22:30:22

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.080	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:29	
Control Test	0.079	22:29	
Air Blank	0.000	22:30	
Diagnostics Check	OK	22:30	

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HARRILL,ANDERSEN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 11:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: PENA AQUINO

DATE/TIME OF TEST: 10/15/2023 12:26:53

Diagnostics Check	OK	12:20
Air Blank	0.000	12:21
Control Test	0.080	12:21
Air Blank	0.000	12:21
Subject Sample # 1	0.176	12:22
Breath Volume	2.515	12:22
Air Blank	0.000	12:23
Air Blank	0.000	12:24
Subject Sample # 2	0.170	12:25
Breath Volume	2.457	12:25
Air Blank	0.000	12:25
Control Test	0.081	12:26
Air Blank	0.000	12:26
Diagnostics Check	OK	12:26

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,JONATHAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 10/22/2023 03:44:43

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.080	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ, JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DIEGO GASPAR

DATE/TIME OF TEST: 10/24/2023 00:25:06

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.161	00:20
Breath Volume	2.429	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.161	00:23
Breath Volume	1.597	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JAMIE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: TORBIO

DATE/TIME OF TEST: 10/28/2023 17:19:16

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:10	
Control Test	0.080	17:10	
Air Blank	0.000	17:11	
Subject Sample	# 1	17:14	No Sample Provided
Breath Volume	0.000	17:14	
Air Blank	0.000	17:14	
Air Blank	0.000	17:16	
Subject Sample	# 2	17:17	Subject Test Refused
Breath Volume	0.000	17:17	
Air Blank	0.000	17:18	
Control Test	0.080	17:18	
Air Blank	0.000	17:19	
Diagnostics Check	OK	17:19	

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JAMIE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: SIMON MENDOZA

DATE/TIME OF TEST: 10/28/2023 23:45:52

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.081	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.079	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,J
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/29/2023 03:49:40

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.119	03:45
Breath Volume	2.710	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.115	03:48
Breath Volume	1.710	03:48
Air Blank	0.000	03:48
Control Test	0.078	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 11/09/2023 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: BLOSSMAN

DATE/TIME OF TEST: 10/05/2023 22:34:52

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.078	22:32	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:33	
Control Test	0.079	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/09/2023 15:21:59

Diagnostics Check	OK	15:18	
Air Blank	0.000	15:19	
Control Test	0.081	15:19	
Air Blank	0.000	15:20	
Subject Sample	# 1	15:20	RFI Detect
Breath Volume	0.000	15:20	
Air Blank	0.000	15:21	
Control Test	0.082	15:21	
Air Blank	0.000	15:21	
Diagnostics Check	OK	15:21	

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/09/2023 15:30:46

Diagnostics Check	OK	15:23
Air Blank	0.000	15:24
Control Test	0.080	15:24
Air Blank	0.000	15:25
Subject Sample # 1	0.219	15:25
Breath Volume	1.718	15:25
Air Blank	0.000	15:26
Air Blank	0.000	15:28
Subject Sample # 2	0.226	15:29
Breath Volume	1.882	15:29
Air Blank	0.000	15:29
Control Test	0.080	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 10/09/2023 20:15:24

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:13	
Control Test	0.080	20:13	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:14	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: REMBISZ

DATE/TIME OF TEST: 10/18/2023 21:52:56

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:38	
Air Blank	0.000	21:38	
Subject Sample	# 1 0.107	21:41	Volume Not Met
Breath Volume	0.101	21:41	
Air Blank	0.000	21:42	
Air Blank	0.000	21:44	
Subject Sample	# 2 0.135	21:45	
Breath Volume	1.808	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 3 0.120	21:51	Volume Not Met
Breath Volume	0.855	21:51	
Air Blank	0.000	21:51	
Control Test	0.077	21:52	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: REMBISZ

DATE/TIME OF TEST: 10/18/2023 22:02:42

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:55	
Control Test	0.079	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1 0.112	21:59	Volume Not Met
Breath Volume	0.847	21:59	
Air Blank	0.000	22:00	
Air Blank	0.000	22:01	
Subject Sample	# 2	22:01	Subject Test Refused
Breath Volume	0.000	22:01	
Air Blank	0.000	22:02	
Control Test	0.075	22:02	RFI Detect
Air Blank		22:02	RFI Detect

Data Download Date/Time: 10/24/2023 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: JONES TORRES

DATE/TIME OF TEST: 10/27/2023 02:18:39

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.079	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.078	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 11/21/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 10/28/2023 03:19:37

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.155	03:14
Breath Volume	1.347	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.158	03:17
Breath Volume	3.066	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 11/21/2023 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ROCHA

DATE/TIME OF TEST: 10/16/2023 05:21:13

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.082	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.183	05:16
Breath Volume	2.050	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:18
Subject Sample # 2	0.183	05:19
Breath Volume	2.687	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/31/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 09/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BIANCHI

DATE/TIME OF TEST: 10/22/2023 01:12:28

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.183	01:07
Breath Volume	3.328	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.181	01:10
Breath Volume	3.187	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 10/31/2023 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: STREICHER

DATE/TIME OF TEST: 10/15/2023 03:26:46

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.000	03:26	Control Outside Tolerance
Air Blank	0.000	03:26	

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: STREICHER

DATE/TIME OF TEST: 10/15/2023 03:37:32

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.169	03:30
Breath Volume	1.484	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.156	03:35
Breath Volume	1.515	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPIC,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ESCALANTE HERNANDEZ

DATE/TIME OF TEST: 10/26/2023 21:09:17

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.077	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.178	21:02
Breath Volume	1.894	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.171	21:07
Breath Volume	1.859	21:07
Air Blank	0.000	21:08
Control Test	0.078	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MEDINA RAMIREZ

DATE/TIME OF TEST: 10/27/2023 00:53:43

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.077	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.142	00:48
Breath Volume	3.035	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.133	00:52
Breath Volume	3.121	00:52
Air Blank	0.000	00:52
Control Test	0.076	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MCCRATER

DATE/TIME OF TEST: 10/28/2023 01:45:31

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.077	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.254	01:40
Breath Volume	2.003	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.244	01:43
Breath Volume	1.566	01:43
Air Blank	0.000	01:44
Control Test	0.077	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ODOR,ROBERT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: LORMIL

DATE/TIME OF TEST: 10/29/2023 02:13:38

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.263	02:08
Breath Volume	1.210	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.256	02:11
Breath Volume	1.621	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 11/08/2023 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MEJIA GARCIA

DATE/TIME OF TEST: 10/29/2023 02:58:34

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.153	02:53
Breath Volume	1.855	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.157	02:56
Breath Volume	1.371	02:56
Air Blank	0.000	02:57
Control Test	0.077	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 11/08/2023 13:03