

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,ERIC,G
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: ROMAN,CARLOS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/01/2023 16:17:03

Diagnostics Check	OK	16:10
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Subject Sample # 1	0.266	16:12
Breath Volume	3.003	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.267	16:15
Breath Volume	2.925	16:15
Air Blank	0.000	16:16
Control Test	0.079	16:16
Air Blank	0.000	16:16
Diagnostics Check	OK	16:17

Data Download Date/Time: 11/19/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: DIAZ,YEENIS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: DE LEON SALAS

DATE/TIME OF TEST: 11/04/2023 06:11:03

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.081	05:58	
Air Blank	0.000	05:58	
Subject Sample	# 1	06:01	No Sample Provided
Breath Volume	0.000	06:01	
Air Blank	0.000	06:02	
Air Blank	0.000	06:04	
Subject Sample	# 2	0.256	06:05
Breath Volume	1.109	06:05	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample	# 3	0.239	06:09
Breath Volume	1.480	06:09	
Air Blank	0.000	06:10	
Control Test	0.082	06:10	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:11	

Data Download Date/Time: 11/19/2023 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 11/11/2023 21:49:41

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.000	21:39	Volume Not Met
Breath Volume	0.148	21:39	
Air Blank	0.000	21:39	
Air Blank	0.000	21:41	
Subject Sample	# 2 0.170	21:42	No .020 agreement
Breath Volume	1.226	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 3 0.192	21:47	No .020 agreement
Breath Volume	1.515	21:47	
Air Blank	0.000	21:48	
Control Test	0.083	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 11/19/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 11/11/2023 22:22:55

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.175	22:18
Breath Volume	1.851	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample # 2	0.171	22:21
Breath Volume	1.285	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 11/19/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: LEDDY

DATE/TIME OF TEST: 11/17/2023 12:45:20

Diagnostics Check	OK	12:33	
Air Blank	0.000	12:34	
Control Test	0.079	12:34	
Air Blank	0.000	12:35	
Subject Sample	# 1	12:38	No Sample Provided
Breath Volume	0.000	12:38	
Air Blank	0.000	12:38	
Air Blank	0.000	12:40	
Subject Sample	# 2	12:43	No Sample Provided
Breath Volume	0.000	12:43	
Air Blank	0.000	12:44	
Control Test	0.080	12:44	
Air Blank	0.000	12:45	
Diagnostics Check	OK	12:45	

Data Download Date/Time: 11/19/2023 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: DIAZ,Y
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HEMKE-FISH

DATE/TIME OF TEST: 11/19/2023 22:38:14

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.255	22:30
Breath Volume	3.820	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:32
Subject Sample # 2	0.280	22:33
Breath Volume	3.898	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:35
Subject Sample # 3	0.272	22:36
Breath Volume	2.500	22:36
Air Blank	0.000	22:37
Control Test	0.077	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 12/16/2023 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MURPHY,KEVIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/26/2023 22:46:29

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.181	22:41
Breath Volume	1.402	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.171	22:44
Breath Volume	1.863	22:44
Air Blank	0.000	22:45
Control Test	0.078	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 12/16/2023 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 11/17/2023 03:21:19

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.214	03:14
Breath Volume	2.699	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.190	03:16
Breath Volume	1.523	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 3	0.205	03:19
Breath Volume	3.765	03:19
Air Blank	0.000	03:20
Control Test	0.077	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 12/03/2023 20:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SORRELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/18/2023 03:35:24

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 12/03/2023 20:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 11/22/2023 22:02:23

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.103	21:57
Breath Volume	1.886	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.100	22:00
Breath Volume	1.960	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 12/03/2023 20:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BUDGIS III

DATE/TIME OF TEST: 11/03/2023 23:01:19

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.079	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.079	23:00	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 11/08/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 11/04/2023 00:54:51

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank		00:54	RFI Detect
Air Blank	0.000	00:54	

Data Download Date/Time: 11/08/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 11/04/2023 00:59:36

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:57	
Control Test	0.079	00:57	
Air Blank	0.000	00:57	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 11/08/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ABUNDIZ-MAJIA

DATE/TIME OF TEST: 11/09/2023 03:01:15

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.201	02:56
Breath Volume	3.089	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.194	02:59
Breath Volume	1.597	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: BAUDUIN

DATE/TIME OF TEST: 11/10/2023 22:35:34

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.078	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:34	Subject Test Refused
Breath Volume	0.000	22:34	
Air Blank	0.000	22:34	
Control Test	0.078	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: DUPREY

DATE/TIME OF TEST: 11/16/2023 22:36:30

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:32
Breath Volume	1.746	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:35
Breath Volume	1.449	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KNUDSON

DATE/TIME OF TEST: 11/17/2023 02:52:21

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.179	02:47
Breath Volume	1.234	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.181	02:50
Breath Volume	1.679	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: REYES-NONTHE

DATE/TIME OF TEST: 11/18/2023 02:28:05

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.078	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 11/19/2023 00:34:17

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.130	00:29
Breath Volume	2.535	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.126	00:32
Breath Volume	1.789	00:32
Air Blank	0.000	00:33
Control Test	0.079	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PANZO CONTRERAS

DATE/TIME OF TEST: 11/19/2023 00:56:04

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.129	00:49
Breath Volume	1.222	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.124	00:54
Breath Volume	1.449	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:56

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 11/19/2023 01:44:28

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 12/20/2023 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DUNLAVY

DATE/TIME OF TEST: 11/19/2023 04:04:57

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: MERCIER

DATE/TIME OF TEST: 11/22/2023 22:05:51

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:54	
Control Test	0.078	21:55	
Air Blank	0.000	21:55	
Subject Sample	# 1 0.143	21:58	Volume Not Met
Breath Volume	0.519	21:58	
Air Blank	0.000	21:59	
Air Blank	0.000	22:01	
Subject Sample	# 2 0.132	22:04	Volume Not Met
Breath Volume	0.312	22:04	
Air Blank	0.000	22:04	
Control Test	0.079	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: TECHLER

DATE/TIME OF TEST: 11/22/2023 22:16:51

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.184	22:12
Breath Volume	1.535	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.170	22:15
Breath Volume	1.324	22:15
Air Blank	0.000	22:15
Control Test	0.078	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ISLEY

DATE/TIME OF TEST: 11/23/2023 02:13:36

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 11/23/2023 03:49:36

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.163	03:45
Breath Volume	1.812	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.163	03:48
Breath Volume	1.578	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HILLIARD

DATE/TIME OF TEST: 11/23/2023 23:05:50

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.108	23:01
Breath Volume	1.398	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.105	23:04
Breath Volume	1.203	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CIHLAR

DATE/TIME OF TEST: 11/24/2023 01:23:40

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: STRINE

DATE/TIME OF TEST: 11/24/2023 20:18:58

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.079	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.107	20:14
Breath Volume	1.441	20:14
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.108	20:17
Breath Volume	1.753	20:17
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: FONTANEZ

DATE/TIME OF TEST: 11/24/2023 21:35:47

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.079	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Control Test	0.079	21:35	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: HAUSER

DATE/TIME OF TEST: 11/25/2023 02:04:00

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.204	01:59
Breath Volume	1.425	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.205	02:02
Breath Volume	1.500	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG210103
DRY GAS EXPIRY DATE: 04/11/2024
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ROMERO-ERICKSON

DATE/TIME OF TEST: 11/26/2023 02:42:17

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.078	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 12/20/2023 14:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARRISH,J,L
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: PAVLAKOS,P
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 11/04/2023 15:00:29

Diagnostics Check	OK	14:54
Air Blank	0.000	14:54
Control Test	0.080	14:55
Air Blank	0.000	14:55
Subject Sample # 1	0.000	14:56
Breath Volume	2.484	14:56
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.000	14:59
Breath Volume	1.898	14:59
Air Blank	0.000	14:59
Control Test	0.079	14:59
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

Data Download Date/Time: 11/06/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: PINSKER, JOSEPH
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HERNANDEZ RUIZ

DATE/TIME OF TEST: 11/06/2023 00:54:36

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.100	00:50
Breath Volume	2.804	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.103	00:53
Breath Volume	2.390	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 11/06/2023 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARRISH,J,
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: COSTELLO,K
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 11/10/2023 15:59:48

Diagnostics Check	OK	15:53
Air Blank	0.000	15:53
Control Test	0.080	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.000	15:55
Breath Volume	3.433	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.000	15:58
Breath Volume	3.156	15:58
Air Blank	0.000	15:58
Control Test	0.079	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 12/28/2023 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: BOILS

DATE/TIME OF TEST: 11/12/2023 00:19:56

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.226	00:15
Breath Volume	2.976	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.219	00:18
Breath Volume	1.984	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 12/28/2023 15:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: HAYCOCK,CODY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HERNANDEZ-SANCHEZ

DATE/TIME OF TEST: 11/12/2023 01:14:39

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.189	01:10
Breath Volume	1.644	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.192	01:13
Breath Volume	1.925	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: HAYCOCK,CODY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: VAN LEAR

DATE/TIME OF TEST: 11/12/2023 20:37:34

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.029	20:33
Breath Volume	1.218	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.027	20:36
Breath Volume	1.183	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MATHEWS,ARNOLD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 11/14/2023 12:52:12

Diagnostics Check	OK	12:45
Air Blank	0.000	12:46
Control Test	0.081	12:46
Air Blank	0.000	12:47
Subject Sample # 1	0.349	12:47
Breath Volume	1.496	12:47
Air Blank	0.000	12:48
Air Blank	0.000	12:49
Subject Sample # 2	0.340	12:50
Breath Volume	2.789	12:50
Air Blank	0.000	12:51
Control Test	0.079	12:51
Air Blank	0.000	12:52
Diagnostics Check	OK	12:52

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 11/16/2023 01:09:02

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.152	01:04
Breath Volume	2.339	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.146	01:07
Breath Volume	2.152	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELIONE, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PAVLAKOS, P
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:29

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 11/17/2023 15:37:28

Diagnostics Check	OK	15:30
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:32
Subject Sample # 1	0.220	15:32
Breath Volume	3.152	15:32
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.222	15:35
Breath Volume	3.093	15:35
Air Blank	0.000	15:36
Control Test	0.078	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGDON,TODD,S
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HODGDON,TODD
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 11/19/2023 01:10:21

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1 0.127	00:59	Volume Not Met
Breath Volume	0.140	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2 0.133	01:03	
Breath Volume	1.980	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 3 0.120	01:08	
Breath Volume	1.429	01:08	
Air Blank	0.000	01:09	
Control Test	0.080	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGDON,TODD,S
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HODGDON,TODD
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/19/2023 03:13:01

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.146	03:08
Breath Volume	2.546	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.147	03:11
Breath Volume	1.910	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGDON,TODD,S
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: COON,DAVID
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GONEZ

DATE/TIME OF TEST: 11/19/2023 04:01:59

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.123	03:57
Breath Volume	1.539	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.120	04:00
Breath Volume	1.269	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: FEARNOW,S
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 11/21/2023 20:49:51

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:44
Air Blank	0.000	20:44
Subject Sample # 1	0.193	20:45
Breath Volume	2.312	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.185	20:48
Breath Volume	1.535	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: WEBER,DYLAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 11/27/2023 17:40:59

Diagnostics Check	OK	17:34
Air Blank	0.000	17:35
Control Test	0.081	17:35
Air Blank	0.000	17:35
Subject Sample # 1	0.166	17:36
Breath Volume	2.910	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:38
Subject Sample # 2	0.158	17:39
Breath Volume	2.054	17:39
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:40
Diagnostics Check	OK	17:40

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,B,
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LEONARD,B
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: LAURIA

DATE/TIME OF TEST: 11/28/2023 22:23:11

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.141	22:18
Breath Volume	1.785	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.145	22:21
Breath Volume	1.789	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAU,BRITTNEY,D
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MATHEWS,ARNOLD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 11/30/2023 13:12:30

Diagnostics Check	OK	13:06
Air Blank	0.000	13:06
Control Test	0.081	13:06
Air Blank	0.000	13:07
Subject Sample # 1	0.122	13:07
Breath Volume	3.000	13:07
Air Blank	0.000	13:08
Air Blank	0.000	13:10
Subject Sample # 2	0.111	13:10
Breath Volume	2.683	13:10
Air Blank	0.000	13:11
Control Test	0.078	13:11
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

Data Download Date/Time: 12/14/2023 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 19:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 11/02/2023 19:47:05

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.073	19:42
Breath Volume	2.140	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.074	19:45
Breath Volume	1.933	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:47

Data Download Date/Time: 11/28/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 11/09/2023 18:13:53

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.078	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.194	18:08
Breath Volume	1.804	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.202	18:12
Breath Volume	1.949	18:12
Air Blank	0.000	18:12
Control Test	0.078	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 11/28/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 16:55

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:44

SUBJECT'S LAST NAME: LOWER

DATE/TIME OF TEST: 11/15/2023 17:20:24

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:06	
Control Test	0.080	17:06	
Air Blank	0.000	17:06	
Subject Sample	# 1 0.178	17:09	
Breath Volume	2.023	17:09	
Air Blank	0.000	17:10	
Air Blank	0.000	17:12	
Subject Sample	# 2 0.177	17:15	Volume Not Met
Breath Volume	0.789	17:15	
Air Blank	0.000	17:16	
Air Blank	0.000	17:17	
Subject Sample	# 3 0.174	17:18	
Breath Volume	1.839	17:18	
Air Blank	0.000	17:19	
Control Test	0.076	17:19	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:20	

Data Download Date/Time: 11/28/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 18:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 11/24/2023 19:03:34

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.079	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.219	18:58
Breath Volume	2.164	18:58
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.204	19:01
Breath Volume	2.457	19:01
Air Blank	0.000	19:02
Control Test	0.079	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 11/28/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MARKUNAS,ANDREW
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MALONEY

DATE/TIME OF TEST: 11/05/2023 02:29:18

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.305	02:22
Breath Volume	1.601	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.296	02:27
Breath Volume	1.785	02:27
Air Blank	0.000	02:28
Control Test	0.078	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP TROOP D
ARRESTING OFFICER: LOPEZ,GEORGE
ARRESTING OFFICER AGENCY: FHP TROOP D
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 11/05/2023 07:07:46

Diagnostics Check	OK	07:01
Air Blank	0.000	07:01
Control Test	0.082	07:01
Air Blank	0.000	07:02
Subject Sample # 1	0.134	07:03
Breath Volume	1.843	07:03
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample # 2	0.141	07:06
Breath Volume	2.347	07:06
Air Blank	0.000	07:06
Control Test	0.081	07:07
Air Blank	0.000	07:07
Diagnostics Check	OK	07:07

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SCHUCK,KYLE
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCCASLIN

DATE/TIME OF TEST: 11/06/2023 03:06:28

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.000	03:01
Breath Volume	3.531	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:05
Breath Volume	4.863	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FRASER,JERROD
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 11/09/2023 19:29:54

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.080	19:24
Air Blank	0.000	19:24
Subject Sample # 1	0.168	19:25
Breath Volume	1.757	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:27
Subject Sample # 2	0.169	19:28
Breath Volume	1.800	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,F
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25721080A2
DRY GAS EXPIRY DATE: 11/05/2023
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: NEIDRINGHAUS

DATE/TIME OF TEST: 11/10/2023 19:05:30

Diagnostics Check	OK	18:59
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	19:00
Subject Sample # 1	0.146	19:01
Breath Volume	3.000	19:01
Air Blank	0.000	19:01
Air Blank	0.000	19:03
Subject Sample # 2	0.144	19:03
Breath Volume	2.839	19:03
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SCHUCK,KYLE
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CINTOGABRIEL

DATE/TIME OF TEST: 11/18/2023 02:12:03

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.167	02:07
Breath Volume	3.515	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.168	02:10
Breath Volume	3.800	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:12

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HULL,PAUL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SHASTRI

DATE/TIME OF TEST: 11/18/2023 05:01:12

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.082	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.009	04:56
Breath Volume	1.828	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.007	04:59
Breath Volume	1.625	04:59
Air Blank	0.000	05:00
Control Test	0.081	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SANCHEZ,HIBER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 11/19/2023 03:20:48

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.082	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1 0.116	03:13	Volume Not Met
Breath Volume	0.593	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 2 0.131	03:16	
Breath Volume	2.500	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 3 0.124	03:19	
Breath Volume	2.570	03:19	
Air Blank	0.000	03:19	
Control Test	0.079	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MARKUNAS,ANDREW
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: TRIM

DATE/TIME OF TEST: 11/24/2023 00:10:24

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:54	
Control Test	0.082	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 3 0.144	23:58	Volume Not Met
Breath Volume	0.070	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.151	00:02	
Breath Volume	1.765	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.123	00:08	Volume Not Met
Breath Volume	0.578	00:08	
Air Blank	0.000	00:09	
Control Test	0.079	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HULL JR,PAUL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: FUNK

DATE/TIME OF TEST: 11/28/2023 05:27:47

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.082	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.000	05:23
Breath Volume	2.238	05:23
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.000	05:26
Breath Volume	1.886	05:26
Air Blank	0.000	05:26
Control Test	0.082	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: DETTMAN

DATE/TIME OF TEST: 11/29/2023 05:18:22

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.083	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.246	05:13
Breath Volume	2.414	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.243	05:16
Breath Volume	1.593	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 11/30/2023 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TAYLOR,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 11/01/2023 05:37:58

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.077	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.078	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 11/01/2023 23:58:59

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.077	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.182	23:54
Breath Volume	1.589	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.186	23:57
Breath Volume	1.312	23:57
Air Blank	0.000	23:58
Control Test	0.077	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SCHEIDER

DATE/TIME OF TEST: 11/03/2023 23:05:00

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.077	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:04	
Control Test	0.077	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 11/04/2023 01:00:57

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.078	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.078	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CALIOR

DATE/TIME OF TEST: 11/04/2023 02:36:16

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.077	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.205	02:31
Breath Volume	2.785	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.206	02:34
Breath Volume	3.140	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DE MELO MORAES

DATE/TIME OF TEST: 11/04/2023 03:49:49

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:47	
Control Test	0.077	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.077	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 11/05/2023 03:44:14

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.142	03:39
Breath Volume	2.234	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.145	03:42
Breath Volume	1.984	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 11/05/2023 04:56:55

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.078	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.181	04:52
Breath Volume	2.074	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.176	04:55
Breath Volume	2.542	04:55
Air Blank	0.000	04:55
Control Test	0.077	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 11/05/2023 21:37:13

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.193	21:32
Breath Volume	1.839	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.195	21:35
Breath Volume	1.972	21:35
Air Blank	0.000	21:36
Control Test	0.077	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 11/17/2023 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 11/06/2023 00:50:38

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.078	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.059	00:45
Breath Volume	2.554	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.064	00:49
Breath Volume	2.812	00:49
Air Blank	0.000	00:49
Control Test	0.077	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 11/17/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/06/2023 02:30:12

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.077	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.077	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 11/17/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TILTON

DATE/TIME OF TEST: 11/10/2023 02:29:47

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.077	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.187	02:25
Breath Volume	1.972	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.185	02:28
Breath Volume	1.992	02:28
Air Blank	0.000	02:28
Control Test	0.077	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/17/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/10/2023 06:14:27

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.077	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.140	06:09
Breath Volume	3.855	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.132	06:12
Breath Volume	3.578	06:12
Air Blank	0.000	06:13
Control Test	0.076	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 11/17/2023 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BRANAMAN

DATE/TIME OF TEST: 11/13/2023 23:12:26

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:10	
Control Test	0.077	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:11	
Control Test	0.078	23:11	
Air Blank	0.000	23:12	
Diagnostics Check	OK	23:12	

Data Download Date/Time: 11/17/2023 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LAUREANO

DATE/TIME OF TEST: 11/04/2023 03:15:37

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 11/26/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: GOMES VIEIRA

DATE/TIME OF TEST: 11/07/2023 04:22:05

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.231	04:17
Breath Volume	3.281	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.236	04:20
Breath Volume	1.746	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 11/26/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,DAVID,J
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: SOKOLOHORSKY,ANDREW
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: DANNAH

DATE/TIME OF TEST: 11/11/2023 05:20:39

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.206	05:16
Breath Volume	2.496	05:16
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.198	05:19
Breath Volume	2.093	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 11/26/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: JENNINGS,LUKE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: S415200761110

DATE/TIME OF TEST: 11/14/2023 02:27:59

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.239	02:23
Breath Volume	2.988	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.240	02:26
Breath Volume	3.132	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/26/2023 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CARTY

DATE/TIME OF TEST: 11/27/2023 19:37:11

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Control Test	0.079	19:36	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 12/27/2023 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/26/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ANIL

DATE/TIME OF TEST: 11/30/2023 02:24:37

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.150	02:20
Breath Volume	3.160	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.142	02:23
Breath Volume	2.167	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/27/2023 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,Y
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: MEJIA VASQUEZ

DATE/TIME OF TEST: 11/16/2023 20:10:36

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:08	
Control Test	0.082	20:08	
Air Blank	0.000	20:09	
Subject Sample	# 1	20:09	Subject Test Refused
Breath Volume	0.000	20:09	
Air Blank	0.000	20:09	
Control Test	0.082	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/17/2023 00:46:34

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.083	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.051	00:41
Breath Volume	1.484	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.055	00:44
Breath Volume	1.476	00:44
Air Blank	0.000	00:45
Control Test	0.083	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CURRINGTON

DATE/TIME OF TEST: 11/17/2023 03:39:21

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.110	03:34
Breath Volume	1.386	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.107	03:37
Breath Volume	1.492	03:37
Air Blank	0.000	03:38
Control Test	0.083	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: DIAZ RAMIREZ

DATE/TIME OF TEST: 11/18/2023 01:50:22

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.082	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.247	01:45
Breath Volume	1.621	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.245	01:48
Breath Volume	1.843	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: NOBLE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/18/2023 05:58:49

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.083	05:53
Air Blank	0.000	05:53
Subject Sample # 1	0.108	05:54
Breath Volume	2.023	05:54
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 2	0.112	05:57
Breath Volume	1.644	05:57
Air Blank	0.000	05:57
Control Test	0.082	05:58
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 11/18/2023 18:34:07

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:31	
Control Test	0.083	18:31	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:33	
Control Test	0.083	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 11/18/2023 22:32:16

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.083	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.190	22:27
Breath Volume	2.054	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.172	22:30
Breath Volume	1.906	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CRUZ ROSALES

DATE/TIME OF TEST: 11/19/2023 04:15:33

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.083	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.196	04:10
Breath Volume	1.160	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.186	04:13
Breath Volume	1.281	04:13
Air Blank	0.000	04:14
Control Test	0.083	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 12/16/2023 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: POOLE,Z
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SHULTZ

DATE/TIME OF TEST: 11/19/2023 23:03:31

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.083	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.191	22:58
Breath Volume	1.335	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.195	23:01
Breath Volume	1.636	23:01
Air Blank	0.000	23:02
Control Test	0.082	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: IVY

DATE/TIME OF TEST: 11/20/2023 01:40:03

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.082	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1 0.146	01:27	
Breath Volume	1.234	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 2 0.116	01:33	Volume Not Met
Breath Volume	0.457	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 3 0.134	01:38	
Breath Volume	1.386	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 11/20/2023 02:24:07

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.247	02:19
Breath Volume	2.949	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.243	02:22
Breath Volume	2.382	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JIMINEZ GUERRERO

DATE/TIME OF TEST: 11/20/2023 04:07:53

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.044	04:03
Breath Volume	2.347	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.045	04:06
Breath Volume	1.417	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/20/2023 17:38:44

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:36	
Control Test	0.083	17:36	
Air Blank	0.000	17:37	
Subject Sample	# 1	17:37	Subject Test Refused
Breath Volume	0.000	17:37	
Air Blank	0.000	17:37	
Control Test	0.083	17:38	
Air Blank	0.000	17:38	
Diagnostics Check	OK	17:38	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PENA HERRERA

DATE/TIME OF TEST: 11/21/2023 04:59:24

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.084	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1 0.058	04:50	Volume Not Met
Breath Volume	0.324	04:50	
Air Blank	0.000	04:50	
Air Blank	0.000	04:52	
Subject Sample	# 2 0.056	04:53	
Breath Volume	1.859	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 3 0.055	04:57	
Breath Volume	1.773	04:57	
Air Blank	0.000	04:58	
Control Test	0.083	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: FUHLER

DATE/TIME OF TEST: 11/21/2023 23:56:12

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.083	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.117	23:51
Breath Volume	1.925	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.115	23:54
Breath Volume	1.378	23:54
Air Blank	0.000	23:55
Control Test	0.083	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 11/22/2023 02:06:36

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.084	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.209	02:01
Breath Volume	1.496	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.203	02:04
Breath Volume	1.484	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MEISSNER

DATE/TIME OF TEST: 11/23/2023 02:15:54

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.084	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.127	02:10
Breath Volume	1.292	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.118	02:14
Breath Volume	1.441	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SANCHEZ CALMO

DATE/TIME OF TEST: 11/23/2023 21:43:17

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.085	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.230	21:38
Breath Volume	2.722	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.221	21:41
Breath Volume	2.500	21:41
Air Blank	0.000	21:42
Control Test	0.083	21:42
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: AGUILAR DOMINGO

DATE/TIME OF TEST: 11/24/2023 00:29:14

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.085	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.085	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A4
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JELOUANE

DATE/TIME OF TEST: 11/24/2023 03:57:49

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.085	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.109	03:52
Breath Volume	1.500	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.109	03:56
Breath Volume	1.503	03:56
Air Blank	0.000	03:56
Control Test	0.084	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MCKEIVER

DATE/TIME OF TEST: 11/25/2023 03:11:27

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.082	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.034	03:07
Breath Volume	1.113	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.033	03:09
Breath Volume	1.570	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: LOPEX CONTERAS

DATE/TIME OF TEST: 11/25/2023 19:44:13

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.268	19:39
Breath Volume	2.945	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:41
Subject Sample # 2	0.266	19:42
Breath Volume	2.449	19:42
Air Blank	0.000	19:43
Control Test	0.081	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: FUENTES ALVAREZ

DATE/TIME OF TEST: 11/25/2023 23:50:56

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.082	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.202	23:46
Breath Volume	1.542	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.210	23:49
Breath Volume	1.617	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: UZOMA HUMES,U
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/26/2023 19:50:39

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.082	19:48	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:49	
Control Test	0.082	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LUFF

DATE/TIME OF TEST: 11/26/2023 23:35:15

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.082	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.124	23:30
Breath Volume	2.285	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.126	23:33
Breath Volume	1.992	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 11/27/2023 22:58:06

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.143	22:53
Breath Volume	1.464	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.140	22:56
Breath Volume	1.707	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 11/28/2023 00:21:00

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.082	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.101	00:16
Breath Volume	1.480	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.090	00:19
Breath Volume	1.136	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCHRUM,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: M650786893100

DATE/TIME OF TEST: 11/28/2023 15:30:37

Diagnostics Check	OK	15:27	
Air Blank	0.000	15:28	
Control Test	0.082	15:28	
Air Blank	0.000	15:29	
Subject Sample	# 1	15:29	Subject Test Refused
Breath Volume	0.000	15:29	
Air Blank	0.000	15:29	
Control Test	0.082	15:30	
Air Blank	0.000	15:30	
Diagnostics Check	OK	15:30	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GODINEZ DIAZ

DATE/TIME OF TEST: 11/30/2023 00:53:50

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.083	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.124	00:48
Breath Volume	2.425	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.122	00:52
Breath Volume	1.996	00:52
Air Blank	0.000	00:52
Control Test	0.083	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/30/2023 03:48:18

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.083	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.082	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTELT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 11/30/2023 06:51:50

Diagnostics Check	OK	06:39	
Air Blank	0.000	06:39	
Control Test	0.082	06:39	
Air Blank	0.000	06:40	
Subject Sample	# 1 0.167	06:43	Volume Not Met
Breath Volume	1.066	06:43	
Air Blank	0.000	06:44	
Air Blank	0.000	06:46	
Subject Sample	# 2 0.170	06:46	
Breath Volume	1.414	06:46	
Air Blank	0.000	06:47	
Air Blank	0.000	06:49	
Subject Sample	# 3 0.179	06:50	
Breath Volume	1.335	06:50	
Air Blank	0.000	06:50	
Control Test	0.081	06:51	
Air Blank	0.000	06:51	
Diagnostics Check	OK	06:51	

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: GARCIA BAMACA

DATE/TIME OF TEST: 11/30/2023 23:34:02

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.093	23:29
Breath Volume	1.734	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.091	23:32
Breath Volume	1.695	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:34

Data Download Date/Time: 12/16/2023 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

INSTRUMENT REGISTERED TO: LAKE ALFRED PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GETTLE,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Lake Alfred PD
ARRESTING OFFICER: KING,PATRICK
ARRESTING OFFICER AGENCY: Lake Alfred PD
LAST AGENCY INSPECTION: 10/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 11/11/2023 22:41:29

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.124	22:36
Breath Volume	1.921	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.130	22:39
Breath Volume	1.894	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 11/30/2023 23:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE, MAKINA, A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: JENNINGS, LUKE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SCHIFFMILLER

DATE/TIME OF TEST: 11/04/2023 00:18:33

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.319	00:14
Breath Volume	2.242	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.310	00:16
Breath Volume	1.628	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 11/28/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAYE, MAKINA, A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: JENNINGS, LUKE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 11/23/2023 01:37:40

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.140	01:33
Breath Volume	4.632	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.144	01:36
Breath Volume	2.917	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 11/28/2023 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MILLIAN,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 11/04/2023 21:29:28

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.178	21:24
Breath Volume	1.875	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.175	21:27
Breath Volume	1.937	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKES,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LAWSON,H
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/07/2023 04:01:00

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.169	03:56
Breath Volume	2.492	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.167	03:59
Breath Volume	1.113	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 11/08/2023 23:32:50

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.028	23:28
Breath Volume	2.531	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:30
Subject Sample # 2	0.022	23:31
Breath Volume	2.664	23:31
Air Blank	0.000	23:31
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 11/09/2023 03:10:31

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEFFERSON,ALI,S
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: EAST,ZACH
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: MORALES-COLON

DATE/TIME OF TEST: 11/10/2023 00:36:04

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:31
Breath Volume	2.835	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.000	00:34
Breath Volume	2.449	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUTCHERSON,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/15/2023 17:28:33

Diagnostics Check	OK	17:25	
Air Blank	0.000	17:26	
Control Test	0.080	17:26	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:27	Subject Test Refused
Breath Volume	0.000	17:27	
Air Blank	0.000	17:27	
Control Test	0.080	17:27	
Air Blank	0.000	17:28	
Diagnostics Check	OK	17:28	

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BURNETTE,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PENICHE MEDINA

DATE/TIME OF TEST: 11/18/2023 03:58:02

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.112	03:53
Breath Volume	3.289	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.112	03:56
Breath Volume	2.984	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTES,JOSEAS,I
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BARNES,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 11/18/2023 22:22:12

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.208	22:16
Breath Volume	1.105	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.203	22:20
Breath Volume	1.859	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CREWS,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 11/19/2023 07:28:25

Diagnostics Check	OK	07:22
Air Blank	0.000	07:22
Control Test	0.079	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.217	07:24
Breath Volume	2.199	07:24
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.214	07:26
Breath Volume	2.546	07:26
Air Blank	0.000	07:27
Control Test	0.080	07:27
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMPTON,J,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 11/21/2023 18:07:10

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.079	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.015	18:02
Breath Volume	1.746	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.015	18:05
Breath Volume	3.882	18:05
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 11/21/2023 20:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: CANOSA,TIMOTHY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497672
DRY GAS EXPIRY DATE: 05/20/2024
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MELHADO

DATE/TIME OF TEST: 11/03/2023 01:17:08

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.148	01:11
Breath Volume	1.972	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.154	01:15
Breath Volume	1.750	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 11/20/2023 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/11/2023 02:18:56

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.147	02:14
Breath Volume	2.722	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.146	02:17
Breath Volume	2.238	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 11/29/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MARCANO FIGUEROA

DATE/TIME OF TEST: 11/24/2023 03:56:57

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.201	03:52
Breath Volume	1.316	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.209	03:55
Breath Volume	1.121	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 11/29/2023 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 11/25/2023 01:40:04

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.166	01:35
Breath Volume	2.019	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.173	01:38
Breath Volume	1.292	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/29/2023 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 11/29/2023 00:38:50

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.116	00:34
Breath Volume	1.902	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.113	00:37
Breath Volume	1.687	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 11/29/2023 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOTHARD, RONALD, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: GOTHARD, RONALD
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: EPPERSON

DATE/TIME OF TEST: 11/15/2023 01:09:19

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.077	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.186	01:04
Breath Volume	2.101	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.181	01:07
Breath Volume	1.304	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 12/18/2023 16:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DEVITO,KELLY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GODINEZ MARTIN

DATE/TIME OF TEST: 11/12/2023 06:15:35

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.077	06:09
Air Blank	0.000	06:10
Subject Sample # 1	0.206	06:11
Breath Volume	2.121	06:11
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.214	06:14
Breath Volume	2.394	06:14
Air Blank	0.000	06:14
Control Test	0.076	06:14
Air Blank	0.000	06:15
Diagnostics Check	OK	06:15

Data Download Date/Time: 12/13/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DEQUARTO,NICHOLAS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: DIAZ REYES

DATE/TIME OF TEST: 11/19/2023 06:00:15

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.077	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.120	05:55
Breath Volume	1.750	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.123	05:58
Breath Volume	1.777	05:58
Air Blank	0.000	05:59
Control Test	0.077	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 12/13/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REBOLLEDO, OSCAR, J
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DIAZ, CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 11/24/2023 23:54:58

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.077	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.246	23:46	Volume Not Met
Breath Volume	0.753	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 2 0.244	23:49	
Breath Volume	1.539	23:49	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 3 0.247	23:53	
Breath Volume	1.613	23:53	
Air Blank	0.000	23:54	
Control Test	0.071	23:54	Control Outside Tolerance
Air Blank	0.000	23:54	

Data Download Date/Time: 12/13/2023 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WOOD,NICHOLAS
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: RISOLDI

DATE/TIME OF TEST: 11/05/2023 04:08:04

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.197	04:03
Breath Volume	2.890	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.196	04:06
Breath Volume	2.703	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/29/2023 07:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/01/2023 02:48:25

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.139	02:43
Breath Volume	1.144	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.131	02:46
Breath Volume	1.183	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/29/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MONTHE AMBROCIO

DATE/TIME OF TEST: 11/03/2023 01:23:11

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.116	01:18
Breath Volume	1.566	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.116	01:21
Breath Volume	1.527	01:21
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 11/29/2023 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VANDEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: PEREZ MONCAYO

DATE/TIME OF TEST: 11/04/2023 05:53:10

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.153	05:48
Breath Volume	2.210	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.153	05:51
Breath Volume	2.261	05:51
Air Blank	0.000	05:52
Control Test	0.078	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: BURNSIDE

DATE/TIME OF TEST: 11/04/2023 20:44:19

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.079	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.078	20:37	Volume Not Met
Breath Volume	0.191	20:37	
Air Blank	0.000	20:37	
Air Blank	0.000	20:39	
Subject Sample	# 2 0.067	20:42	Volume Not Met
Breath Volume	0.183	20:42	
Air Blank	0.000	20:43	
Control Test	0.079	20:43	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: BURNSIDE

DATE/TIME OF TEST: 11/04/2023 21:02:37

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.078	20:48	
Air Blank	0.000	20:49	
Subject Sample	# 1 0.067	20:49	
Breath Volume	1.179	20:49	
Air Blank	0.000	20:50	
Air Blank	0.000	20:52	
Subject Sample	# 2 0.062	20:55	Volume Not Met
Breath Volume	0.511	20:55	
Air Blank	0.000	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 3 0.074	21:01	
Breath Volume	1.371	21:01	
Air Blank	0.000	21:01	
Control Test	0.079	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SPELMAN,WILLIAM
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 11/06/2023 20:09:00

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.162	20:04
Breath Volume	1.152	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:06
Subject Sample # 2	0.160	20:07
Breath Volume	1.464	20:07
Air Blank	0.000	20:08
Control Test	0.078	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:09

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DARAG

DATE/TIME OF TEST: 11/11/2023 03:55:44

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.078	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:48	No Sample Provided
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample	# 2	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.078	03:55	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: ARCINIEGAS

DATE/TIME OF TEST: 11/11/2023 20:24:14

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.078	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:20
Breath Volume	3.101	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	2.382	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS, STEWART, A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: PHILLIPS, STEWART
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/11/2023 22:01:44

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.078	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.162	21:55
Breath Volume	1.917	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.152	21:59
Breath Volume	2.488	21:59
Air Blank	0.000	22:00
Control Test	0.078	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: JANSSEN

DATE/TIME OF TEST: 11/15/2023 22:22:28

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.141	22:17
Breath Volume	1.371	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.157	22:20
Breath Volume	2.503	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 11/17/2023 00:03:43

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.142	23:59
Breath Volume	2.105	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.135	00:02
Breath Volume	2.015	00:02
Air Blank	0.000	00:02
Control Test	0.078	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RAY,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ANGELES

DATE/TIME OF TEST: 11/17/2023 00:16:34

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.078	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.190	00:09
Breath Volume	1.214	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.185	00:14
Breath Volume	1.492	00:14
Air Blank	0.000	00:15
Control Test	0.078	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SZWAKOP,MONIKA
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ROJO

DATE/TIME OF TEST: 11/17/2023 03:02:36

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.114	02:58
Breath Volume	1.851	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.110	03:01
Breath Volume	1.382	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 11/17/2023 20:41:21

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.077	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.210	20:36
Breath Volume	1.296	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.229	20:39
Breath Volume	1.644	20:39
Air Blank	0.000	20:40
Control Test	0.077	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: BOYNTON

DATE/TIME OF TEST: 11/17/2023 20:49:32

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.078	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.041	20:45
Breath Volume	2.566	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.042	20:48
Breath Volume	2.179	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: ESPINO BUEXJUAJI

DATE/TIME OF TEST: 11/18/2023 00:44:29

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.137	00:39
Breath Volume	1.531	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.139	00:42
Breath Volume	1.273	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: KANYHA

DATE/TIME OF TEST: 11/23/2023 10:52:46

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.078	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.124	10:48
Breath Volume	1.238	10:48
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.125	10:51
Breath Volume	1.398	10:51
Air Blank	0.000	10:51
Control Test	0.079	10:52
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 11/29/2023 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 11/24/2023 00:44:44

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.185	00:40
Breath Volume	2.316	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.172	00:43
Breath Volume	1.386	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 11/29/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SHAUGHNESSY

DATE/TIME OF TEST: 11/25/2023 16:00:08

Diagnostics Check	OK	15:54
Air Blank	0.000	15:54
Control Test	0.079	15:54
Air Blank	0.000	15:55
Subject Sample # 1	0.000	15:55
Breath Volume	3.093	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:58
Subject Sample # 2	0.000	15:58
Breath Volume	2.753	15:58
Air Blank	0.000	15:59
Control Test	0.079	15:59
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 11/29/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,OFFICER,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: LEBEL

DATE/TIME OF TEST: 11/25/2023 17:09:28

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.078	17:03
Air Blank	0.000	17:03
Subject Sample # 1	0.285	17:04
Breath Volume	1.621	17:04
Air Blank	0.000	17:05
Air Blank	0.000	17:07
Subject Sample # 2	0.284	17:07
Breath Volume	1.441	17:07
Air Blank	0.000	17:08
Control Test	0.078	17:08
Air Blank	0.000	17:09
Diagnostics Check	OK	17:09

Data Download Date/Time: 11/29/2023 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ARTEAGA ZUNIGA

DATE/TIME OF TEST: 11/01/2023 00:06:44

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.082	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.069	00:01
Breath Volume	1.390	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:03
Subject Sample # 2	0.071	00:05
Breath Volume	1.964	00:05
Air Blank	0.000	00:05
Control Test	0.082	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MATUTE GUERRERO

DATE/TIME OF TEST: 11/01/2023 03:32:02

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.143	03:27
Breath Volume	2.683	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.127	03:30
Breath Volume	2.335	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:32

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: BETTMAN

DATE/TIME OF TEST: 11/01/2023 06:20:59

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.082	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.034	06:16
Breath Volume	1.691	06:16
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.036	06:19
Breath Volume	1.539	06:19
Air Blank	0.000	06:20
Control Test	0.083	06:20
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ECHEVERRIAVIDAL,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/02/2023 23:25:42

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:23	
Control Test	0.082	23:23	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Control Test	0.082	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BARTOLON MORALES

DATE/TIME OF TEST: 11/03/2023 00:51:46

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.084	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.275	00:47
Breath Volume	1.855	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.279	00:50
Breath Volume	1.308	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/03/2023 03:39:28

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:38	Subject Test Refused
Breath Volume	0.000	03:38	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:39	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: DELEO,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GEBRETSADIK

DATE/TIME OF TEST: 11/04/2023 03:36:39

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.083	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.082	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAGOMENI,G,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/05/2023 03:24:35

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.083	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DE SOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 11/08/2023 02:21:46

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.142	02:16
Breath Volume	1.554	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.155	02:20
Breath Volume	1.378	02:20
Air Blank	0.000	02:20
Control Test	0.083	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 11/17/2023 12:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 11/09/2023 05:07:32

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.072	05:02
Breath Volume	1.148	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.071	05:05
Breath Volume	1.242	05:05
Air Blank	0.000	05:06
Control Test	0.084	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 11/10/2023 01:46:53

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.083	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:39	No Sample Provided
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.083	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HERMSDORFER

DATE/TIME OF TEST: 11/10/2023 03:18:34

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.083	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 11/10/2023 21:28:38

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.082	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.136	21:24
Breath Volume	1.816	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.134	21:26
Breath Volume	1.742	21:26
Air Blank	0.000	21:27
Control Test	0.082	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CROUSE

DATE/TIME OF TEST: 11/10/2023 23:33:37

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.082	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.139	23:28
Breath Volume	2.125	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.141	23:31
Breath Volume	2.738	23:31
Air Blank	0.000	23:32
Control Test	0.082	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,TIMOTHY
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: VALDIVIA DAVILA

DATE/TIME OF TEST: 11/11/2023 02:27:02

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.082	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:22
Breath Volume	2.921	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.000	02:25
Breath Volume	3.347	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/11/2023 19:25:01

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:22	
Control Test	0.082	19:22	
Air Blank	0.000	19:23	
Subject Sample	# 1	19:23	Subject Test Refused
Breath Volume	0.000	19:23	
Air Blank	0.000	19:24	
Control Test	0.082	19:24	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SHERVEY

DATE/TIME OF TEST: 11/11/2023 20:53:46

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.082	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.100	20:49
Breath Volume	1.230	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.115	20:52
Breath Volume	1.457	20:52
Air Blank	0.000	20:52
Control Test	0.082	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/11/2023 22:33:55

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.083	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.093	22:29
Breath Volume	2.847	22:29
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.094	22:32
Breath Volume	3.667	22:32
Air Blank	0.000	22:32
Control Test	0.082	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HENLY

DATE/TIME OF TEST: 11/12/2023 00:49:12

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.115	00:44
Breath Volume	1.925	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.113	00:47
Breath Volume	2.078	00:47
Air Blank	0.000	00:48
Control Test	0.083	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOMERO

DATE/TIME OF TEST: 11/12/2023 01:56:18

Diagnostics Check	OK	01:53	
Air Blank	0.000	01:53	
Control Test	0.082	01:54	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:55	
Control Test	0.081	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CANALES GARCIA

DATE/TIME OF TEST: 11/12/2023 04:15:48

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.082	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.071	04:11
Breath Volume	2.125	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.073	04:14
Breath Volume	2.269	04:14
Air Blank	0.000	04:14
Control Test	0.083	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C.W
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 11/12/2023 05:54:17

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:51	
Control Test	0.082	05:51	
Air Blank	0.000	05:52	
Subject Sample	# 1	05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:53	
Control Test	0.082	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOMM,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 11/12/2023 06:40:08

Diagnostics Check	OK	06:33
Air Blank	0.000	06:33
Control Test	0.082	06:34
Air Blank	0.000	06:34
Subject Sample # 1	0.119	06:35
Breath Volume	2.042	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.119	06:38
Breath Volume	1.730	06:38
Air Blank	0.000	06:39
Control Test	0.080	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/12/2023 22:18:09

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.230	22:13
Breath Volume	2.398	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:15
Subject Sample # 2	0.228	22:16
Breath Volume	1.984	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HASTY

DATE/TIME OF TEST: 11/12/2023 23:20:14

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.081	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1 0.137	23:15	
Breath Volume	1.437	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 2 0.137	23:18	
Breath Volume	1.562	23:18	
Air Blank		23:19	Ambient Fail
Air Blank	0.000	23:20	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 11/13/2023 02:02:15

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.081	02:00	
Air Blank	0.000	02:00	
Subject Sample # 1	0.027	02:01	
Breath Volume	1.148	02:01	
Air Blank		02:01	Ambient Fail
Air Blank	0.000	02:02	

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,JAMES,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: PUHACH

DATE/TIME OF TEST: 11/15/2023 18:14:46

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.082	18:06
Air Blank	0.000	18:06
Subject Sample # 1	0.149	18:09
Breath Volume	1.128	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:11
Subject Sample # 2	0.133	18:13
Breath Volume	1.281	18:13
Air Blank	0.000	18:13
Control Test	0.079	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 11/17/2023 12:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 11/09/2023 10:35:27

Diagnostics Check	OK	10:28
Air Blank	0.000	10:29
Control Test	0.079	10:29
Air Blank	0.000	10:29
Subject Sample # 1	0.000	10:30
Breath Volume	2.851	10:30
Air Blank	0.000	10:31
Air Blank	0.000	10:33
Subject Sample # 2	0.000	10:34
Breath Volume	2.152	10:34
Air Blank	0.000	10:34
Control Test	0.078	10:34
Air Blank	0.000	10:35
Diagnostics Check	OK	10:35

Data Download Date/Time: 11/26/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/11/2023 02:28:17

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.139	02:23
Breath Volume	1.156	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.143	02:26
Breath Volume	1.593	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 11/26/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,J,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GESOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 11/18/2023 02:19:56

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.175	02:14
Breath Volume	1.468	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.183	02:18
Breath Volume	1.695	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/26/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/23/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: BERRIAN

DATE/TIME OF TEST: 11/22/2023 02:20:59

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.209	02:16
Breath Volume	2.121	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.194	02:19
Breath Volume	2.250	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 11/26/2023 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,DEBBIE,L
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COLON,JAILENE
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 10/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: LOVEDAY

DATE/TIME OF TEST: 11/10/2023 03:19:31

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.147	03:15
Breath Volume	1.691	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.151	03:17
Breath Volume	1.449	03:17
Air Blank	0.000	03:18
Control Test	0.077	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 11/28/2023 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AMATRUDA,CHRIS,G

BREATH TEST OPERATOR AGENCY: Dade City PD

ARRESTING OFFICER: SERRANO,VICENTE

ARRESTING OFFICER AGENCY: Dade City PD

LAST AGENCY INSPECTION: 10/16/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A3

DRY GAS EXPIRY DATE: 09/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HOYT

DATE/TIME OF TEST: 11/20/2023 19:27:52

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.203	19:23
Breath Volume	1.566	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.201	19:26
Breath Volume	2.410	19:26
Air Blank	0.000	19:26
Control Test	0.078	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 11/28/2023 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: MIDOLO,MICHAEL
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 11/29/2023 14:33:52

Diagnostics Check	OK	14:27
Air Blank	0.000	14:28
Control Test	0.079	14:28
Air Blank	0.000	14:28
Subject Sample # 1	0.000	14:29
Breath Volume	1.546	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.000	14:32
Breath Volume	1.953	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:33

Data Download Date/Time: 12/25/2023 11:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SCHROCK

DATE/TIME OF TEST: 11/30/2023 23:49:02

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.000	23:48	Control Outside Tolerance
Air Blank	0.000	23:49	

Data Download Date/Time: 12/18/2023 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENCZYNSKI,JONATHAN,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 11/01/2023 05:18:42

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.183	05:14
Breath Volume	1.882	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.177	05:17
Breath Volume	2.324	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 11/30/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 11/01/2023 23:12:37

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.401	23:08
Breath Volume	1.867	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample # 2	0.403	23:11
Breath Volume	1.878	23:11
Air Blank	0.000	23:11
Control Test	0.076	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 11/30/2023 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 11/02/2023 00:45:08

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.080	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 11/30/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: PORT

DATE/TIME OF TEST: 11/02/2023 02:22:53

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 11/30/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 11/02/2023 23:39:52

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:37	
Air Blank	0.000	23:37	
Subject Sample # 1	0.112	23:38	
Breath Volume	2.300	23:38	
Air Blank		23:38	RFI Detect
Air Blank	0.000	23:38	
Control Test	0.079	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 11/30/2023 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CROCKETT

DATE/TIME OF TEST: 11/02/2023 23:47:28

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Subject Sample # 1	0.110	23:43
Breath Volume	2.011	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.112	23:45
Breath Volume	1.542	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 11/03/2023 22:07:37

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.080	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.082	22:03
Breath Volume	2.828	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.087	22:06
Breath Volume	3.347	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: STEINHEISER

DATE/TIME OF TEST: 11/04/2023 00:37:44

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.167	00:33
Breath Volume	3.191	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.169	00:36
Breath Volume	3.335	00:36
Air Blank	0.000	00:36
Control Test	0.078	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/04/2023 21:49:00

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.078	21:44
Breath Volume	1.984	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.076	21:47
Breath Volume	3.246	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:49

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/04/2023 23:15:21

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.025	23:11
Breath Volume	1.164	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.023	23:13
Breath Volume	2.105	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KLOEMPKE

DATE/TIME OF TEST: 11/04/2023 23:55:04

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.078	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: CICOLETTI

DATE/TIME OF TEST: 11/05/2023 01:45:27

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.109	01:41
Breath Volume	2.156	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.109	01:43
Breath Volume	2.511	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON, CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: FULTON

DATE/TIME OF TEST: 11/07/2023 22:50:09

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:46	No Sample Provided
Breath Volume	0.000	22:46	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 2	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.080	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NATZEL

DATE/TIME OF TEST: 11/08/2023 02:17:38

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.126	02:13
Breath Volume	2.070	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.123	02:16
Breath Volume	2.148	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: ARLINE

DATE/TIME OF TEST: 11/08/2023 11:46:05

Diagnostics Check	OK	11:38
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.000	11:41
Breath Volume	1.441	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:44
Breath Volume	1.542	11:44
Air Blank	0.000	11:45
Control Test	0.079	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 11/09/2023 06:30:23

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.080	06:21
Air Blank	0.000	06:21
Subject Sample # 1	0.318	06:23
Breath Volume	2.785	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.344	06:25
Breath Volume	2.550	06:25
Air Blank	0.000	06:26
Air Blank	0.000	06:28
Subject Sample # 3	0.326	06:28
Breath Volume	1.656	06:28
Air Blank	0.000	06:29
Control Test	0.079	06:29
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 11/09/2023 23:09:42

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GOMEZ RAMIREZ

DATE/TIME OF TEST: 11/10/2023 01:56:19

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:45	No Sample Provided
Breath Volume	0.000	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2	0.277	No .020 agreement
Breath Volume	2.750	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 3	0.252	No .020 agreement
Breath Volume	1.359	01:54	
Air Blank	0.000	01:55	
Control Test	0.078	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GOMEZ RAMIREZ

DATE/TIME OF TEST: 11/10/2023 02:06:12

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.269	02:01
Breath Volume	1.589	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.287	02:04
Breath Volume	1.632	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BAVA

DATE/TIME OF TEST: 11/10/2023 22:05:13

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.146	22:00
Breath Volume	3.367	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.142	22:03
Breath Volume	2.878	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ALBERT,JR

DATE/TIME OF TEST: 11/10/2023 22:44:22

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.089	22:39
Breath Volume	1.925	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.083	22:42
Breath Volume	1.609	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 11/11/2023 01:00:24

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 11/11/2023 01:47:13

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.163	01:42
Breath Volume	2.437	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.159	01:45
Breath Volume	1.976	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 11/11/2023 02:52:07

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.193	02:47
Breath Volume	2.214	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.183	02:50
Breath Volume	1.601	02:50
Air Blank	0.000	02:51
Control Test	0.076	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PETERING,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: MCSWAIN

DATE/TIME OF TEST: 11/11/2023 05:52:19

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.080	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.202	05:46
Breath Volume	1.511	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.197	05:50
Breath Volume	2.152	05:50
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Diagnostics Check	OK	05:52

Data Download Date/Time: 11/30/2023 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,N,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: LAMONT

DATE/TIME OF TEST: 11/11/2023 17:57:01

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.080	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.079	17:51
Breath Volume	2.062	17:51
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.074	17:55
Breath Volume	2.191	17:55
Air Blank	0.000	17:56
Control Test	0.080	17:56
Air Blank	0.000	17:56
Diagnostics Check	OK	17:57

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 11/13/2023 02:22:31

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:15	No Sample Provided
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 2	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BARRETT,P
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: DENISTY

DATE/TIME OF TEST: 11/13/2023 19:33:00

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:22	
Control Test	0.080	19:22	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:26	No Sample Provided
Breath Volume	0.000	19:26	
Air Blank	0.000	19:26	
Air Blank	0.000	19:28	
Subject Sample	# 2	19:31	No Sample Provided
Breath Volume	0.000	19:31	
Air Blank	0.000	19:32	
Control Test	0.079	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH, KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART, C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 11/14/2023 01:21:55

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.225	01:17
Breath Volume	2.226	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.213	01:20
Breath Volume	2.109	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENCZYNSKI,JONATHAN,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: GLAZENER

DATE/TIME OF TEST: 11/14/2023 07:22:07

Diagnostics Check	OK	07:15
Air Blank	0.000	07:15
Control Test	0.081	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.000	07:17
Breath Volume	1.367	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:20
Subject Sample # 2	0.000	07:20
Breath Volume	1.492	07:20
Air Blank	0.000	07:21
Control Test	0.080	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 11/14/2023 14:06:20

Diagnostics Check	OK	13:59
Air Blank	0.000	14:00
Control Test	0.080	14:00
Air Blank	0.000	14:00
Subject Sample # 1	0.000	14:02
Breath Volume	1.531	14:02
Air Blank	0.000	14:02
Air Blank	0.000	14:04
Subject Sample # 2	0.000	14:04
Breath Volume	1.738	14:04
Air Blank	0.000	14:05
Control Test	0.079	14:05
Air Blank	0.000	14:06
Diagnostics Check	OK	14:06

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:36

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 11/15/2023 10:04:55

Diagnostics Check	OK	09:57
Air Blank	0.000	09:58
Control Test	0.080	09:58
Air Blank	0.000	09:58
Subject Sample # 1	0.000	09:59
Breath Volume	1.328	09:59
Air Blank	0.000	09:59
Air Blank	0.000	10:02
Subject Sample # 2	0.000	10:03
Breath Volume	1.347	10:03
Air Blank	0.000	10:03
Control Test	0.080	10:04
Air Blank	0.000	10:04
Diagnostics Check	OK	10:04

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,N,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 11/15/2023 11:11:33

Diagnostics Check	OK	11:04
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:05
Subject Sample # 1	0.010	11:07
Breath Volume	2.292	11:07
Air Blank	0.000	11:07
Air Blank	0.000	11:09
Subject Sample # 2	0.009	11:10
Breath Volume	2.734	11:10
Air Blank	0.000	11:10
Control Test	0.079	11:10
Air Blank	0.000	11:11
Diagnostics Check	OK	11:11

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 11/16/2023 00:09:46

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Subject Sample # 1	0.195	00:00
Breath Volume	3.140	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.167	00:05
Breath Volume	2.832	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 3	0.198	00:08
Breath Volume	1.453	00:08
Air Blank	0.000	00:08
Control Test	0.078	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 11/17/2023 02:04:42

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.156	02:00
Breath Volume	1.226	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.149	02:03
Breath Volume	1.800	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/17/2023 16:19:39

Diagnostics Check	OK	16:13
Air Blank	0.000	16:13
Control Test	0.080	16:14
Air Blank	0.000	16:14
Subject Sample # 1	0.126	16:15
Breath Volume	2.253	16:15
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.123	16:18
Breath Volume	2.113	16:18
Air Blank	0.000	16:18
Control Test	0.079	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: BELLO

DATE/TIME OF TEST: 11/17/2023 20:55:46

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.080	20:45	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:48	No Sample Provided
Breath Volume	0.000	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:54	No Sample Provided
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.080	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 11/17/2023 22:16:18

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:08	
Control Test	0.080	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:12	No Sample Provided
Breath Volume	0.000	22:12	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 11/17/2023 23:39:47

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.187	23:34
Breath Volume	2.234	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.196	23:38
Breath Volume	1.726	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU, A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 11/18/2023 01:33:25

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.079	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Air Blank	0.000	01:31	
Subject Sample	# 2	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU, A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SALZER

DATE/TIME OF TEST: 11/18/2023 03:40:11

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.288	03:35
Breath Volume	3.304	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.280	03:38
Breath Volume	3.199	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/18/2023 14:54:14

Diagnostics Check	OK	14:51	
Air Blank	0.000	14:51	
Control Test	0.081	14:51	
Air Blank	0.000	14:52	
Subject Sample	# 1	14:52	Subject Test Refused
Breath Volume	0.000	14:52	
Air Blank	0.000	14:53	
Control Test	0.079	14:53	
Air Blank	0.000	14:54	
Diagnostics Check	OK	14:54	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 11/18/2023 23:25:10

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.081	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:20	No Sample Provided
Breath Volume	0.000	23:20	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 2	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.080	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN, MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/19/2023 01:40:49

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.157	01:35
Breath Volume	1.835	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.161	01:39
Breath Volume	1.445	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: VACHON

DATE/TIME OF TEST: 11/19/2023 03:30:54

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.000	03:26
Breath Volume	4.710	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.000	03:29
Breath Volume	4.054	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/30/2023 09:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/19/2023 19:39:07

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.140	19:34
Breath Volume	3.164	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.139	19:37
Breath Volume	3.285	19:37
Air Blank	0.000	19:38
Control Test	0.079	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTENSEN,S,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ANDREW

DATE/TIME OF TEST: 11/20/2023 01:54:01

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:43	
Control Test	0.080	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:47	No Sample Provided
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,AUSTIN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 11/20/2023 03:29:13

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:27	
Air Blank	0.000	03:27	
Subject Sample # 1	0.219	03:28	
Breath Volume	2.207	03:28	
Air Blank		03:28	RFI Detect
Air Blank		03:28	RFI Detect
Air Blank	0.000	03:29	

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,AUSTIN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 11/20/2023 03:38:07

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.219	03:33
Breath Volume	2.402	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.220	03:36
Breath Volume	2.218	03:36
Air Blank	0.000	03:37
Control Test	0.078	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 11/21/2023 12:14:53

Diagnostics Check	OK	12:08
Air Blank	0.000	12:09
Control Test	0.080	12:09
Air Blank	0.000	12:09
Subject Sample # 1	0.000	12:10
Breath Volume	3.609	12:10
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.000	12:13
Breath Volume	3.019	12:13
Air Blank	0.000	12:13
Control Test	0.080	12:14
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LENAHAN,RYAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: GIBB

DATE/TIME OF TEST: 11/21/2023 15:05:51

Diagnostics Check	OK	14:57	
Air Blank	0.000	14:57	
Control Test	0.079	14:58	
Air Blank	0.000	14:58	
Subject Sample	# 1	15:01	No Sample Provided
Breath Volume	0.000	15:01	
Air Blank	0.000	15:02	
Air Blank	0.000	15:04	
Subject Sample	# 2	15:04	Subject Test Refused
Breath Volume	0.000	15:04	
Air Blank	0.000	15:04	
Control Test	0.079	15:05	
Air Blank	0.000	15:05	
Diagnostics Check	OK	15:05	

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 11/22/2023 01:02:53

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.080	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/23/2023 01:10:54

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.000	01:06
Breath Volume	2.804	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.000	01:09
Breath Volume	2.585	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 11/23/2023 01:44:24

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:40	No Sample Provided
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.079	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 11/23/2023 03:16:05

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.145	03:11
Breath Volume	1.601	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.154	03:14
Breath Volume	2.140	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON, CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 11/23/2023 22:05:12

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:59
Air Blank	0.000	21:59
Subject Sample # 1	0.204	22:00
Breath Volume	2.054	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.200	22:03
Breath Volume	1.980	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 11/24/2023 23:52:55

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.000	23:48
Breath Volume	1.894	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.000	23:51
Breath Volume	1.378	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/25/2023 02:11:44

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.136	02:07
Breath Volume	2.742	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.137	02:10
Breath Volume	3.000	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ROMANO,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/25/2023 05:28:39

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.208	05:23
Breath Volume	2.542	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.209	05:27
Breath Volume	2.292	05:27
Air Blank	0.000	05:27
Control Test	0.078	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 11/25/2023 09:39:46

Diagnostics Check	OK	09:33
Air Blank	0.000	09:33
Control Test	0.080	09:34
Air Blank	0.000	09:34
Subject Sample # 1	0.162	09:35
Breath Volume	2.679	09:35
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.162	09:38
Breath Volume	3.640	09:38
Air Blank	0.000	09:38
Control Test	0.079	09:39
Air Blank	0.000	09:39
Diagnostics Check	OK	09:39

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTOCH,JEFFREY,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: REMBERT

DATE/TIME OF TEST: 11/28/2023 08:26:53

Diagnostics Check	OK	08:20
Air Blank	0.000	08:21
Control Test	0.081	08:21
Air Blank	0.000	08:22
Subject Sample # 1	0.000	08:22
Breath Volume	4.199	08:22
Air Blank	0.000	08:23
Air Blank	0.000	08:25
Subject Sample # 2	0.000	08:25
Breath Volume	3.949	08:25
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:26
Diagnostics Check	OK	08:26

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/28/2023 14:19:16

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.082	14:13
Air Blank	0.000	14:14
Subject Sample # 1	0.000	14:14
Breath Volume	1.863	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.000	14:17
Breath Volume	1.816	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:19
Diagnostics Check	OK	14:19

Data Download Date/Time: 11/30/2023 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 11/29/2023 23:11:00

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.081	23:00	
Air Blank	0.000	23:00	
Subject Sample	# 1 0.105	23:03	Volume Not Met
Breath Volume	0.285	23:03	
Air Blank	0.000	23:04	
Air Blank	0.000	23:06	
Subject Sample	# 2 0.098	23:09	Volume Not Met
Breath Volume	0.035	23:09	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 11/30/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 11/29/2023 23:25:58

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.079	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1 0.102	23:17	Volume Not Met
Breath Volume	0.785	23:17	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.107	23:21	
Breath Volume	1.320	23:21	
Air Blank	0.000	23:21	
Air Blank	0.000	23:23	
Subject Sample	# 3 0.112	23:24	
Breath Volume	1.210	23:24	
Air Blank	0.000	23:25	
Control Test	0.080	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 11/30/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: SCHATZLE

DATE/TIME OF TEST: 11/30/2023 03:57:27

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 11/30/2023 09:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 11/02/2023 01:40:09

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.175	01:35
Breath Volume	3.500	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.182	01:38
Breath Volume	2.964	01:38
Air Blank	0.000	01:39
Control Test	0.080	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/03/2023 01:55:03

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.108	01:50
Breath Volume	1.589	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.112	01:53
Breath Volume	1.753	01:53
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:54
Diagnostics Check	OK	01:55

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: COLINDRES MENCIA

DATE/TIME OF TEST: 11/03/2023 03:22:26

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.081	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 11/04/2023 02:17:15

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.081	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 11/05/2023 00:43:00

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.237	00:38
Breath Volume	2.097	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.237	00:41
Breath Volume	2.785	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:42
Diagnostics Check	OK	00:43

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RASHEED

DATE/TIME OF TEST: 11/05/2023 02:51:23

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.206	02:46
Breath Volume	3.738	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.194	02:49
Breath Volume	2.093	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 11/07/2023 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATZ,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: WEBBER,JAYSON
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: THOMAN

DATE/TIME OF TEST: 11/09/2023 10:09:55

Diagnostics Check	OK	10:03	
Air Blank	0.000	10:04	
Control Test	0.081	10:04	
Air Blank	0.000	10:05	
Subject Sample # 1	0.000	10:05	
Breath Volume	1.757	10:05	
Air Blank	0.000	10:06	
Air Blank	0.000	10:08	
Subject Sample # 2	0.000	10:09	
Breath Volume	2.156	10:09	
Air Blank		10:09	RFI Detect
Air Blank	0.000	10:09	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 11/10/2023 04:08:26

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.081	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: OBEREMPT

DATE/TIME OF TEST: 11/10/2023 22:52:54

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:42	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:45	No Sample Provided
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 2	22:51	No Sample Provided
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: OBEREMPT

DATE/TIME OF TEST: 11/10/2023 23:19:00

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2	23:17	No Sample Provided
Breath Volume	0.000	23:17	
Air Blank	0.000	23:18	
Control Test	0.081	23:18	
Air Blank	0.000	23:18	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: OBEREMPT

DATE/TIME OF TEST: 11/10/2023 23:45:36

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.115	23:41
Breath Volume	2.230	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.101	23:44
Breath Volume	1.355	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 20:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HIPOLITO GONZALEZ

DATE/TIME OF TEST: 11/11/2023 21:03:10

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.169	20:58
Breath Volume	2.203	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.154	21:01
Breath Volume	1.687	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CERMAK,OFC,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: ANDREWS,KEVYN

ARRESTING OFFICER AGENCY: Indian Shores PD

LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:43

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/12/2023 00:46:47

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.103	00:41
Breath Volume	1.382	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.096	00:45
Breath Volume	1.382	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CLARK,RYAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: LA FUENTE CUELLAR

DATE/TIME OF TEST: 11/14/2023 01:54:05

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	Subject Test Refused
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Control Test	0.080	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: CONCEPCION, MARINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: NILSEN

DATE/TIME OF TEST: 11/15/2023 17:15:51

Diagnostics Check	OK	17:08
Air Blank	0.000	17:08
Control Test	0.080	17:08
Air Blank	0.000	17:09
Subject Sample # 1	0.174	17:11
Breath Volume	1.699	17:11
Air Blank	0.000	17:12
Air Blank	0.000	17:13
Subject Sample # 2	0.166	17:14
Breath Volume	1.355	17:14
Air Blank	0.000	17:14
Control Test	0.080	17:15
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CAPEHART

DATE/TIME OF TEST: 11/17/2023 02:09:06

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS, KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: UPCHURCH, JERRY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15922080A4

DRY GAS EXPIRY DATE: 08/05/2024

VIOLATION CODE: DUI

ARREST TIME: 23:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: RICKELMANN

DATE/TIME OF TEST: 11/22/2023 00:55:44

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.310	00:50
Breath Volume	1.531	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.304	00:54
Breath Volume	1.808	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: RIVERO

DATE/TIME OF TEST: 11/24/2023 02:08:05

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.143	02:02
Breath Volume	1.695	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.144	02:06
Breath Volume	1.976	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:08

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GOPIE

DATE/TIME OF TEST: 11/24/2023 03:17:29

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.092	03:12
Breath Volume	1.441	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.088	03:16
Breath Volume	1.390	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GRIGGS,JACK
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BURROWS

DATE/TIME OF TEST: 11/24/2023 23:41:00

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.081	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Control Test	0.081	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RAMOS ALTAMIRANO

DATE/TIME OF TEST: 11/26/2023 02:58:41

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.190	02:54
Breath Volume	3.316	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.187	02:57
Breath Volume	2.671	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALAMEDA,JORGE,L
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SCHWEIGERT,ANDREW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 11/28/2023 16:33:03

Diagnostics Check	OK	16:27
Air Blank	0.000	16:27
Control Test	0.080	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.000	16:28
Breath Volume	3.320	16:28
Air Blank	0.000	16:29
Air Blank	0.000	16:31
Subject Sample # 2	0.000	16:31
Breath Volume	3.328	16:31
Air Blank	0.000	16:32
Control Test	0.080	16:32
Air Blank	0.000	16:32
Diagnostics Check	OK	16:33

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: RALPH

DATE/TIME OF TEST: 11/29/2023 22:50:24

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.107	22:45
Breath Volume	2.421	22:45
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.109	22:48
Breath Volume	2.121	22:48
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/07/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A4
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: BREBAN LOPEZ

DATE/TIME OF TEST: 11/30/2023 20:27:22

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.081	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.110	20:22
Breath Volume	1.878	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.109	20:25
Breath Volume	1.500	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/25/2023 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CASTRO,JOHN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: FOLMAR

DATE/TIME OF TEST: 11/04/2023 07:47:25

Diagnostics Check	OK	07:40
Air Blank	0.000	07:41
Control Test	0.081	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.129	07:42
Breath Volume	1.468	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.129	07:45
Breath Volume	2.023	07:45
Air Blank	0.000	07:46
Control Test	0.080	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 11/14/2023 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTALBANO,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COMBS,JESSE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: HINCAPIE-VEGA

DATE/TIME OF TEST: 11/04/2023 08:23:45

Diagnostics Check	OK	08:17
Air Blank	0.000	08:17
Control Test	0.080	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.102	08:19
Breath Volume	3.035	08:19
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample # 2	0.106	08:22
Breath Volume	2.152	08:22
Air Blank	0.000	08:22
Control Test	0.080	08:23
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

Data Download Date/Time: 11/14/2023 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 01:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/05/2023 02:47:37

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.155	02:41
Breath Volume	1.101	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.151	02:46
Breath Volume	1.734	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/14/2023 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PROPST,JOHN,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: PROPST,JOHN

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 20:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 11/08/2023 21:26:38

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:19	No Sample Provided
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 11/14/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PROPST,JOHN,B

BREATH TEST OPERATOR AGENCY: Auburndale PD

ARRESTING OFFICER: PROPST,JOHN

ARRESTING OFFICER AGENCY: Auburndale PD

LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HOUE

DATE/TIME OF TEST: 11/09/2023 00:28:15

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.044	00:23
Breath Volume	2.343	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.045	00:26
Breath Volume	2.367	00:26
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 11/14/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: GRISSOM,TIMOTHY

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 01:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/10/2023 02:59:35

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.081	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	02:58	No Sample Provided
Breath Volume	0.000	02:58	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 11/14/2023 10:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: WELDEN

DATE/TIME OF TEST: 11/15/2023 20:46:53

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.176	20:42
Breath Volume	2.582	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.179	20:45
Breath Volume	2.382	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/16/2023 19:52:59

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.194	19:48
Breath Volume	1.339	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.187	19:51
Breath Volume	1.722	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 19:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 11/16/2023 20:32:45

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.080	20:22	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:25	No Sample Provided
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Air Blank	0.000	20:28	
Subject Sample	# 2	20:31	No Sample Provided
Breath Volume	0.000	20:31	
Air Blank	0.000	20:31	
Control Test	0.080	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FUORE,ROBERT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: PROCICK

DATE/TIME OF TEST: 11/16/2023 22:27:07

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.137	22:19	Volume Not Met
Breath Volume	0.046	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2 0.110	22:25	Volume Not Met
Breath Volume	0.519	22:25	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: GRISSOM,ARTHUR

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 11/19/2023 19:29:25

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:18	
Control Test	0.080	19:18	
Air Blank	0.000	19:19	
Subject Sample	# 1 0.134	19:22	Volume Not Met
Breath Volume	0.179	19:22	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2 0.109	19:27	Volume Not Met
Breath Volume	0.535	19:27	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:29	
Diagnostics Check	OK	19:29	

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: POPP

DATE/TIME OF TEST: 11/21/2023 19:18:36

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.000	19:14
Breath Volume	1.601	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.000	19:17
Breath Volume	2.367	19:17
Air Blank	0.000	19:17
Control Test	0.080	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 12/05/2023 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 19:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/21/2023 20:47:37

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.081	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:38	No Sample Provided
Breath Volume	0.000	20:38	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 2	20:42	
Breath Volume	1.644	20:42	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 3	20:46	
Breath Volume	1.718	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:47	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EVANS,KALEIGH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GARCIA JR

DATE/TIME OF TEST: 11/21/2023 23:26:29

Diagnostics Check	OK	23:15	
Air Blank	0.000	23:15	
Control Test	0.081	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 1	23:19	No Sample Provided
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample	# 2	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PROPST,JOHN,B
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: PROPST,JOHN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 11/22/2023 23:51:26

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.189	23:45
Breath Volume	1.183	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.184	23:49
Breath Volume	1.199	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 17:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: JEREZ AMEY

DATE/TIME OF TEST: 11/23/2023 18:53:57

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.080	18:46
Air Blank	0.000	18:47
Subject Sample # 1	0.237	18:48
Breath Volume	1.746	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.254	18:52
Breath Volume	1.535	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: GRISSOM,TIMOTHY

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 22:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: CROVO

DATE/TIME OF TEST: 11/23/2023 23:12:54

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.061	23:07
Breath Volume	2.054	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.061	23:11
Breath Volume	2.390	23:11
Air Blank	0.000	23:11
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GILES,DANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 11/25/2023 04:39:31

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.080	04:34
Breath Volume	3.375	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.076	04:37
Breath Volume	3.082	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 05:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 11/25/2023 06:12:03

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:06	
Control Test	0.080	06:06	
Air Blank	0.000	06:07	
Subject Sample	# 1 0.264	06:07	
Breath Volume	1.304	06:07	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 2	06:10	Slope Not Met
Breath Volume	1.128	06:10	
Air Blank	0.000	06:11	
Control Test	0.079	06:11	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,J

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 05:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 11/25/2023 06:20:43

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.079	06:14
Air Blank	0.000	06:14
Subject Sample # 1	0.250	06:15
Breath Volume	1.664	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.235	06:19
Breath Volume	1.671	06:19
Air Blank	0.000	06:19
Control Test	0.078	06:20
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RENDON

DATE/TIME OF TEST: 11/26/2023 03:31:28

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.023	03:27
Breath Volume	3.734	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.020	03:29
Breath Volume	3.574	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELONY,ANATUS,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: BELONY,ANATUS
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/26/2023 03:49:17

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.146	03:44
Breath Volume	1.386	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.143	03:47
Breath Volume	1.914	03:47
Air Blank	0.000	03:48
Control Test	0.077	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELONY,ANATUS,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: AGOSTINO,VINCENT
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 11/26/2023 04:12:26

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 2	04:10	Volume Not Met
Breath Volume	0.179	04:10	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: ALCALDE,ALEXANDER

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: OSTRANDER

DATE/TIME OF TEST: 11/26/2023 04:29:22

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.202	04:24
Breath Volume	1.652	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.199	04:27
Breath Volume	2.578	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: REGISTER,JUSTIN,D

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: REGISTER,JUSTIN

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 03:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/26/2023 04:43:21

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.137	04:38
Breath Volume	1.210	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.147	04:41
Breath Volume	1.191	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 12/05/2023 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 11/27/2023 01:12:47

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.197	01:04
Breath Volume	1.148	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.176	01:07
Breath Volume	1.558	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 3	0.197	01:11
Breath Volume	1.523	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 12/05/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COLON,JOLNETH,B

BREATH TEST OPERATOR AGENCY: Polk County SO

ARRESTING OFFICER: COLON,JOLNETH

ARRESTING OFFICER AGENCY: Polk County SO

LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A2

DRY GAS EXPIRY DATE: 05/05/2025

VIOLATION CODE: DUI

ARREST TIME: 01:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: TKACIK

DATE/TIME OF TEST: 11/28/2023 02:58:47

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.220	02:54
Breath Volume	2.964	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.214	02:57
Breath Volume	1.937	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/05/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: LANCE

DATE/TIME OF TEST: 11/29/2023 16:39:15

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.080	16:33
Air Blank	0.000	16:34
Subject Sample # 1	0.088	16:34
Breath Volume	1.363	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.091	16:37
Breath Volume	1.972	16:37
Air Blank	0.000	16:38
Control Test	0.081	16:38
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 12/05/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: RIGUAL RODRIGUEZ

DATE/TIME OF TEST: 11/29/2023 23:01:08

Diagnostics Check	OK	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.162	22:56
Breath Volume	1.101	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.162	22:59
Breath Volume	1.441	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 12/05/2023 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARQUHARSON,KEVIN,D
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BROWN,EDDIE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/31/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 223083
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: VEVE

DATE/TIME OF TEST: 11/12/2023 00:29:34

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.171	00:24
Breath Volume	2.589	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.161	00:27
Breath Volume	2.855	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 11/30/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 223083
BREATH TEST OPERATOR: MIOT,XAVIER,	DRY GAS EXPIRY DATE: 04/05/2025
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: GOODLETT,GIOVANNI	ARREST TIME: 23:07
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Female
LAST AGENCY INSPECTION: 10/31/2023	OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SANCHEZ	DATE/TIME OF TEST: 11/17/2023	00:38:22
------------------------------	-------------------------------	----------

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.000	00:33
Breath Volume	2.113	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.000	00:36
Breath Volume	1.144	00:36
Air Blank	0.000	00:37
Control Test	0.077	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 11/30/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 223083
BREATH TEST OPERATOR: KARNS,BRANDON,M	DRY GAS EXPIRY DATE: 04/05/2025
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: CAPOTA,J	ARREST TIME: 22:00
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Male
LAST AGENCY INSPECTION: 10/31/2023	OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: COLLINS	DATE/TIME OF TEST: 11/19/2023	22:59:26
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.078	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1 0.175	22:50	Volume Not Met
Breath Volume	0.324	22:50	
Air Blank	0.000	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 2 0.168	22:54	
Breath Volume	1.378	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 3 0.186	22:57	
Breath Volume	2.101	22:57	
Air Blank	0.000	22:58	
Control Test	0.073	22:58	Control Outside Tolerance
Air Blank	0.000	22:59	

Data Download Date/Time: 11/30/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 223083
BREATH TEST OPERATOR: MIOT,XAVIER,	DRY GAS EXPIRY DATE: 04/05/2025
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: FARQUHARSON,KEVIN	ARREST TIME: 21:17
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Male
LAST AGENCY INSPECTION: 10/31/2023	OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: RIVERA	DATE/TIME OF TEST: 11/23/2023	22:04:33
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.078	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.219	21:59
Breath Volume	2.722	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:01
Subject Sample # 2	0.216	22:02
Breath Volume	2.453	22:02
Air Blank	0.000	22:03
Control Test	0.076	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 11/30/2023 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HUMES,U
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 11/03/2023 02:11:03

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.122	02:06
Breath Volume	2.570	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.113	02:09
Breath Volume	1.402	02:09
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: MENDOZA MIRELES

DATE/TIME OF TEST: 11/03/2023 06:28:13

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.082	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.109	06:23
Breath Volume	2.132	06:23
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.111	06:26
Breath Volume	2.886	06:26
Air Blank	0.000	06:27
Control Test	0.081	06:27
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COMAYAGUA,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: DOSES

DATE/TIME OF TEST: 11/04/2023 00:50:06

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.083	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.126	00:45
Breath Volume	1.496	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.134	00:48
Breath Volume	1.718	00:48
Air Blank	0.000	00:49
Control Test	0.082	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/04/2023 03:12:05

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.083	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.200	03:07
Breath Volume	2.714	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.188	03:10
Breath Volume	1.683	03:10
Air Blank	0.000	03:11
Control Test	0.082	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL, EDWARD, A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN, C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 11/05/2023 01:34:22

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.083	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.053	01:29
Breath Volume	1.285	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.059	01:32
Breath Volume	1.531	01:32
Air Blank	0.000	01:33
Control Test	0.082	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,Y
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GORDILLO RINCON

DATE/TIME OF TEST: 11/05/2023 02:02:26

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.083	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.109	01:58
Breath Volume	2.792	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.108	02:00
Breath Volume	1.976	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GROCE

DATE/TIME OF TEST: 11/05/2023 03:14:26

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.082	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.239	03:10
Breath Volume	2.410	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.225	03:12
Breath Volume	2.062	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 11/05/2023 04:47:23

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.083	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.211	04:42
Breath Volume	2.753	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.207	04:45
Breath Volume	2.992	04:45
Air Blank	0.000	04:46
Control Test	0.082	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/05/2023 05:30:09

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.082	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.083	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIMMONS,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 11/06/2023 01:00:02

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	
Control Test	0.082	00:59	
Air Blank	0.000	00:59	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: HIKE

DATE/TIME OF TEST: 11/08/2023 06:01:49

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.083	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:00	Subject Test Refused
Breath Volume	0.000	06:00	
Air Blank	0.000	06:00	
Control Test	0.082	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CRUZ SALAZAR

DATE/TIME OF TEST: 11/09/2023 00:14:36

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:12	
Control Test	0.082	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.082	00:14	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J.R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 11/12/2023 01:55:50

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.082	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.082	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J.R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 11/12/2023 04:12:03

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.179	04:07
Breath Volume	4.578	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.173	04:10
Breath Volume	4.230	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BATEH,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HASTY

DATE/TIME OF TEST: 11/13/2023 00:01:43

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.083	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.139	23:56
Breath Volume	1.250	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.127	00:00
Breath Volume	1.691	00:00
Air Blank	0.000	00:00
Control Test	0.082	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 11/13/2023 02:11:53

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.027	02:07
Breath Volume	2.433	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:09
Subject Sample # 2	0.025	02:10
Breath Volume	2.136	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/13/2023 17:18:58

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.082	17:13
Air Blank	0.000	17:13
Subject Sample # 1	0.000	17:14
Breath Volume	1.851	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.000	17:17
Breath Volume	1.562	17:17
Air Blank	0.000	17:18
Control Test	0.082	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: PHELPS

DATE/TIME OF TEST: 11/16/2023 05:18:23

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.082	05:16	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:17	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:18	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON, CONNER, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SALDANA

DATE/TIME OF TEST: 11/17/2023 03:38:30

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.082	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 11/17/2023 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: BELTON

DATE/TIME OF TEST: 11/18/2023 19:35:46

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.082	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.126	19:31
Breath Volume	2.300	19:31
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.121	19:34
Breath Volume	2.062	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:35
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: SHAR

DATE/TIME OF TEST: 11/18/2023 23:25:45

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.196	23:21
Breath Volume	2.359	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.190	23:24
Breath Volume	2.285	23:24
Air Blank	0.000	23:24
Control Test	0.081	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CONSTANTS

DATE/TIME OF TEST: 11/19/2023 00:37:24

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.029	00:32
Breath Volume	1.968	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.028	00:35
Breath Volume	1.917	00:35
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 11/19/2023 04:02:02

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.082	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.056	03:56
Breath Volume	1.453	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.056	04:00
Breath Volume	2.269	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 11/19/2023 04:55:00

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.082	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.093	04:50
Breath Volume	3.132	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.095	04:53
Breath Volume	3.207	04:53
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:55

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: SELVA-NEGRON

DATE/TIME OF TEST: 11/19/2023 21:15:10

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.082	21:13	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:14	
Control Test	0.081	21:14	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:15	

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DIAZ RODAS

DATE/TIME OF TEST: 11/19/2023 22:18:16

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.173	22:13
Breath Volume	2.324	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.167	22:16
Breath Volume	1.761	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 12/15/2023 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: REYES MARTINEZ

DATE/TIME OF TEST: 11/20/2023 01:42:27

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.099	01:37
Breath Volume	1.343	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.097	01:40
Breath Volume	1.117	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: ONDREY

DATE/TIME OF TEST: 11/20/2023 21:49:40

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.082	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.144	21:44
Breath Volume	1.148	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.141	21:48
Breath Volume	2.007	21:48
Air Blank	0.000	21:48
Control Test	0.082	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,E,V
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GETZ,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: ALDERMAN

DATE/TIME OF TEST: 11/21/2023 07:44:08

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.081	07:38
Air Blank	0.000	07:39
Subject Sample # 1	0.086	07:39
Breath Volume	2.937	07:39
Air Blank	0.000	07:40
Air Blank	0.000	07:42
Subject Sample # 2	0.088	07:42
Breath Volume	2.597	07:42
Air Blank	0.000	07:43
Control Test	0.081	07:43
Air Blank	0.000	07:44
Diagnostics Check	OK	07:44

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TYSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 11/23/2023 03:01:04

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.082	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.166	02:56
Breath Volume	2.058	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.163	02:59
Breath Volume	1.792	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 11/23/2023 06:01:57

Diagnostics Check	OK	05:59	
Air Blank	0.000	05:59	
Control Test	0.081	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:00	Subject Test Refused
Breath Volume	0.000	06:00	
Air Blank	0.000	06:01	
Control Test	0.082	06:01	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 11/24/2023 00:43:50

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.161	00:39
Breath Volume	1.101	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.162	00:42
Breath Volume	1.976	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 11/24/2023 05:03:44

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:50	
Control Test	0.082	04:50	
Air Blank	0.000	04:51	
Subject Sample # 1	0.184	04:53	
Breath Volume	1.496	04:53	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample # 2	0.163	04:58	Volume Not Met
Breath Volume	0.617	04:58	
Air Blank	0.000	04:59	
Air Blank	0.000	05:01	
Subject Sample # 3	0.171	05:02	
Breath Volume	1.488	05:02	
Air Blank	0.000	05:02	
Control Test	0.081	05:03	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 11/25/2023 02:39:55

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.083	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.080	02:35
Breath Volume	1.480	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.079	02:38
Breath Volume	1.609	02:38
Air Blank	0.000	02:39
Control Test	0.082	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RAYMUNDO JIMENEZ

DATE/TIME OF TEST: 11/25/2023 04:26:27

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.083	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.166	04:22
Breath Volume	1.613	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.167	04:24
Breath Volume	2.128	04:24
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: DIGGINS

DATE/TIME OF TEST: 11/27/2023 03:16:14

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.135	03:11
Breath Volume	1.691	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:13
Subject Sample # 2	0.130	03:14
Breath Volume	1.128	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LANHAM

DATE/TIME OF TEST: 11/27/2023 22:54:11

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:51	
Control Test	0.082	22:52	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:53	
Control Test	0.082	22:53	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 11/28/2023 04:16:58

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.082	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.171	04:12
Breath Volume	2.292	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.158	04:15
Breath Volume	1.261	04:15
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ROSTON

DATE/TIME OF TEST: 11/29/2023 03:34:07

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.083	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.323	03:29
Breath Volume	3.929	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.329	03:32
Breath Volume	3.250	03:32
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/30/2023 03:22:54

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.083	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.083	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 11/30/2023 04:37:12

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.083	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.083	04:36	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KRAMER,S
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: MCCULLOCH

DATE/TIME OF TEST: 11/30/2023 23:39:41

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.082	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:38	
Control Test	0.083	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 12/15/2023 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WILLIAMS, DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WHYTE

DATE/TIME OF TEST: 11/17/2023 23:02:29

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.145	22:58
Breath Volume	1.949	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.135	23:00
Breath Volume	1.730	23:00
Air Blank	0.000	23:01
Control Test	0.080	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 11/30/2023 19:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HELTON

DATE/TIME OF TEST: 11/05/2023 02:32:21

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.180	02:27
Breath Volume	1.734	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.175	02:30
Breath Volume	2.144	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MACIEL VILLANUEVA

DATE/TIME OF TEST: 11/19/2023 01:14:49

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.081	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.128	01:10
Breath Volume	1.652	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:12
Subject Sample # 2	0.126	01:13
Breath Volume	1.335	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 11/19/2023 03:15:21

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.102	03:10
Breath Volume	3.347	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.100	03:13
Breath Volume	2.789	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOMAGALA

DATE/TIME OF TEST: 11/19/2023 04:23:15

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.143	04:18
Breath Volume	2.203	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.147	04:21
Breath Volume	1.714	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 11/23/2023 02:14:01

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Subject Sample # 1	0.050	02:10	
Breath Volume	1.765	02:10	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample # 2	0.052	02:13	
Breath Volume	2.535	02:13	
Air Blank		02:13	RFI Detect
Air Blank	0.000	02:14	

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 11/23/2023 02:21:30

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.049	02:17
Breath Volume	1.312	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.048	02:20
Breath Volume	1.781	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 11/29/2023 05:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BUONICONTI,ALEXI
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GENDRON

DATE/TIME OF TEST: 11/01/2023 01:46:35

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	1.820	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:45
Breath Volume	2.730	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/02/2023 04:57:47

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:56	
Control Test	0.080	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: STAHL, ASHLEY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: ST JOHN

DATE/TIME OF TEST: 11/05/2023 18:50:06

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.080	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.165	18:45
Breath Volume	2.347	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:48
Subject Sample # 2	0.166	18:48
Breath Volume	1.546	18:48
Air Blank	0.000	18:49
Control Test	0.079	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 11/08/2023 15:59:36

Diagnostics Check	OK	15:52
Air Blank	0.000	15:53
Control Test	0.080	15:53
Air Blank	0.000	15:54
Subject Sample # 1	0.000	15:54
Breath Volume	1.230	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.000	15:58
Breath Volume	1.343	15:58
Air Blank	0.000	15:58
Control Test	0.080	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: POWER

DATE/TIME OF TEST: 11/15/2023 03:19:32

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.079	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1 0.126	03:12	Volume Not Met
Breath Volume	0.304	03:12	
Air Blank	0.000	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 2 0.124	03:17	Volume Not Met
Breath Volume	0.597	03:17	
Air Blank	0.000	03:18	
Control Test	0.077	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ADRIANNE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: HANSOTTE

DATE/TIME OF TEST: 11/15/2023 14:13:04

Diagnostics Check	OK	14:06
Air Blank	0.000	14:07
Control Test	0.080	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:08
Breath Volume	2.777	14:08
Air Blank	0.000	14:09
Air Blank	0.000	14:11
Subject Sample # 2	0.000	14:11
Breath Volume	2.601	14:11
Air Blank	0.000	14:12
Control Test	0.080	14:12
Air Blank	0.000	14:12
Diagnostics Check	OK	14:13

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: GOMEZ-HERNANDEZ

DATE/TIME OF TEST: 11/16/2023 22:32:53

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.076	22:28
Breath Volume	1.109	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.073	22:31
Breath Volume	1.351	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 11/18/2023 04:18:07

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.078	04:10	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.078	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 11/18/2023 19:46:41

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.079	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.000	19:42
Breath Volume	3.406	19:42
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.000	19:45
Breath Volume	4.031	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUNNER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RUNNER,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: BENITZE

DATE/TIME OF TEST: 11/19/2023 08:53:57

Diagnostics Check	OK	08:46
Air Blank	0.000	08:47
Control Test	0.078	08:47
Air Blank	0.000	08:48
Subject Sample # 1	0.127	08:49
Breath Volume	2.406	08:49
Air Blank	0.000	08:49
Air Blank	0.000	08:51
Subject Sample # 2	0.124	08:52
Breath Volume	2.117	08:52
Air Blank	0.000	08:53
Control Test	0.079	08:53
Air Blank	0.000	08:53
Diagnostics Check	OK	08:53

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONNELL,JACOB,B
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CONNELL,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: LARRY

DATE/TIME OF TEST: 11/20/2023 13:34:31

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.079	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.000	13:30
Breath Volume	2.066	13:30
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.000	13:33
Breath Volume	1.945	13:33
Air Blank	0.000	13:33
Control Test	0.081	13:33
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,JEFFERSON,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CHEATHAM,TRAVIS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: WAICZOREK

DATE/TIME OF TEST: 11/21/2023 20:54:23

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.275	20:50
Breath Volume	2.273	20:50
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.285	20:52
Breath Volume	2.480	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 11/22/2023 18:13:07

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.078	18:07
Air Blank	0.000	18:07
Subject Sample # 1	0.311	18:08
Breath Volume	2.738	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.324	18:11
Breath Volume	2.679	18:11
Air Blank	0.000	18:12
Control Test	0.076	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COLLINS,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 11/24/2023 03:22:55

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.145	03:17
Breath Volume	1.753	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.148	03:21
Breath Volume	1.378	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,KEVIN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HAMILTON,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DUME

DATE/TIME OF TEST: 11/24/2023 03:47:53

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.147	03:43
Breath Volume	3.273	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.149	03:46
Breath Volume	2.617	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A2
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: KEARNEY

DATE/TIME OF TEST: 11/27/2023 00:01:30

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.078	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:00	Interferent Detect
Breath Volume	3.500	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 11/27/2023 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 212654
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 11/10/2023 00:28:43

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 11/22/2023 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILASECA,HENNESY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/03/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/21/2023 11:29:36

Diagnostics Check	OK	11:23
Air Blank	0.000	11:24
Control Test	0.080	11:24
Air Blank	0.000	11:25
Subject Sample # 1	0.000	11:25
Breath Volume	2.183	11:25
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.000	11:28
Breath Volume	2.246	11:28
Air Blank	0.000	11:28
Control Test	0.080	11:29
Air Blank	0.000	11:29
Diagnostics Check	OK	11:29

Data Download Date/Time: 11/22/2023 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAGA,JASMINE,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SLAGA,JASMINE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: DOZIER

DATE/TIME OF TEST: 11/28/2023 17:20:16

Diagnostics Check	OK	17:08	
Air Blank	0.000	17:09	
Control Test	0.081	17:09	
Air Blank	0.000	17:09	
Subject Sample	# 1 0.249	17:13	Volume Not Met
Breath Volume	0.593	17:13	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample	# 2 0.256	17:18	Volume Not Met
Breath Volume	0.085	17:18	
Air Blank	0.000	17:19	
Control Test	0.076	17:19	
Air Blank	0.000	17:20	
Diagnostics Check	OK	17:20	

Data Download Date/Time: 12/28/2023 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SPRAGUE,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1185592
DRY GAS EXPIRY DATE: 01/04/2024
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 11/30/2023 21:43:48

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.104	21:38
Breath Volume	1.765	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.096	21:42
Breath Volume	1.328	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 12/28/2023 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 11/01/2023 20:45:58

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.148	20:41
Breath Volume	2.347	20:41
Air Blank	0.000	20:42
Air Blank	0.000	20:43
Subject Sample # 2	0.150	20:44
Breath Volume	2.148	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: RADCLIFFE

DATE/TIME OF TEST: 11/01/2023 21:16:49

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.080	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:09	No Sample Provided
Breath Volume	0.000	21:09	
Air Blank	0.000	21:10	
Air Blank	0.000	21:12	
Subject Sample	# 2	21:15	No Sample Provided
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:16	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,OMARIA
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GREEN,OMARI
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: CONTRERAS CORDERO

DATE/TIME OF TEST: 11/01/2023 22:24:14

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.045	22:19
Breath Volume	1.609	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.046	22:22
Breath Volume	1.566	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 11/01/2023 23:33:27

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.202	23:29
Breath Volume	1.207	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.195	23:31
Breath Volume	1.390	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: FRANCISCO ROSALEZ

DATE/TIME OF TEST: 11/02/2023 05:03:42

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.165	04:59
Breath Volume	1.152	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.184	05:02
Breath Volume	1.226	05:02
Air Blank	0.000	05:02
Control Test	0.078	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GADACZ,MATTHEW
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: FERNANDEZ-PEREZ

DATE/TIME OF TEST: 11/03/2023 01:16:52

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.133	01:12
Breath Volume	1.121	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.144	01:15
Breath Volume	1.250	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, EMILY, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWSON, ASHLEY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SAINT-VAL

DATE/TIME OF TEST: 11/04/2023 00:32:36

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:28
Breath Volume	2.195	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.550	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PEREZ VALDES

DATE/TIME OF TEST: 11/05/2023 02:17:15

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.165	02:12
Breath Volume	2.589	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.167	02:15
Breath Volume	3.226	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS, TYREE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NASON, TAYLOR
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRAXTON

DATE/TIME OF TEST: 11/05/2023 03:12:42

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.119	03:08
Breath Volume	1.914	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.119	03:11
Breath Volume	1.656	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: LAVERDE MORENO

DATE/TIME OF TEST: 11/05/2023 14:02:06

Diagnostics Check	OK	14:01	
Air Blank		14:01	RFI Detect
Air Blank	0.000	14:02	

Data Download Date/Time: 11/14/2023 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:24

SUBJECT'S LAST NAME: LAVERDE MORENO

DATE/TIME OF TEST: 11/05/2023 14:17:43

Diagnostics Check	OK	14:12	
Air Blank	0.000	14:12	
Control Test	0.080	14:13	
Air Blank	0.000	14:13	
Subject Sample	# 1	14:16	Subject Test Refused
Breath Volume	0.000	14:16	
Air Blank	0.000	14:16	
Control Test	0.080	14:17	
Air Blank	0.000	14:17	
Diagnostics Check	OK	14:17	

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, EMILY, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ACOSTA, YOSUET
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/06/2023 01:19:46

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.155	01:15
Breath Volume	2.292	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.152	01:18
Breath Volume	1.738	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/07/2023 15:00:52

Diagnostics Check	OK	14:49	
Air Blank	0.000	14:49	
Control Test	0.079	14:50	
Air Blank	0.000	14:50	
Subject Sample	# 1 0.207	14:53	Volume Not Met
Breath Volume	0.097	14:53	
Air Blank	0.000	14:54	
Air Blank	0.000	14:56	
Subject Sample	# 2 0.198	14:59	Volume Not Met
Breath Volume	0.222	14:59	
Air Blank	0.000	14:59	
Control Test	0.078	15:00	
Air Blank	0.000	15:00	
Diagnostics Check	OK	15:00	

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/07/2023 15:11:45

Diagnostics Check	OK	15:03
Air Blank	0.000	15:04
Control Test	0.078	15:04
Air Blank	0.000	15:05
Subject Sample # 1	0.193	15:06
Breath Volume	1.222	15:06
Air Blank	0.000	15:07
Air Blank	0.000	15:08
Subject Sample # 2	0.189	15:10
Breath Volume	1.363	15:10
Air Blank	0.000	15:10
Control Test	0.078	15:11
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GREEN,JOSHUA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/10/2023 22:43:21

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.079	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.159	22:38
Breath Volume	1.109	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.144	22:41
Breath Volume	1.449	22:41
Air Blank	0.000	22:42
Control Test	0.079	22:42
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: SCHULER

DATE/TIME OF TEST: 11/11/2023 01:08:07

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.163	01:03
Breath Volume	2.226	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.157	01:06
Breath Volume	2.125	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BARTOLO

DATE/TIME OF TEST: 11/11/2023 04:08:04

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.203	04:03
Breath Volume	1.347	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.204	04:06
Breath Volume	1.796	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:08

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 10/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26422080A1
DRY GAS EXPIRY DATE: 11/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: POUGH

DATE/TIME OF TEST: 11/11/2023 22:00:20

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.080	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 11/14/2023 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUABA,JASON,T
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: GUABA,JASON
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOWMAN

DATE/TIME OF TEST: 11/15/2023 03:37:51

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 12/05/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWSON,ASHLEY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 11/18/2023 01:37:10

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.169	01:32
Breath Volume	1.683	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.168	01:35
Breath Volume	1.785	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/05/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,RANDY,O
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WILSON,RANDY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SGRO

DATE/TIME OF TEST: 11/19/2023 03:22:27

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.152	03:18
Breath Volume	2.664	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.143	03:21
Breath Volume	3.351	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/05/2023 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS, TYREE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NASON, TAYLOR
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GAMBA-GAMBOA

DATE/TIME OF TEST: 11/19/2023 04:47:04

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.047	04:42
Breath Volume	1.839	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.044	04:45
Breath Volume	1.968	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, KIARA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MELENDEZ ESPINO

DATE/TIME OF TEST: 11/21/2023 03:05:42

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.032	03:01
Breath Volume	1.164	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.031	03:04
Breath Volume	1.410	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SERRANO-RIVERA

DATE/TIME OF TEST: 11/21/2023 03:41:09

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TURNER, LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ADDISON

DATE/TIME OF TEST: 11/21/2023 06:28:36

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:17	
Control Test	0.081	06:17	
Air Blank	0.000	06:18	
Subject Sample	# 1	06:21	No Sample Provided
Breath Volume	0.000	06:21	
Air Blank	0.000	06:21	
Air Blank	0.000	06:23	
Subject Sample	# 2	06:27	Volume Not Met
Breath Volume	0.050	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, EMILY, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: QUATTLEBAUM, VICTORIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/23/2023 03:59:05

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.196	03:54
Breath Volume	2.808	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.199	03:57
Breath Volume	2.804	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLALTA,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MCCOSCO,LANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: VOSS

DATE/TIME OF TEST: 11/23/2023 08:37:08

Diagnostics Check	OK	08:31
Air Blank	0.000	08:31
Control Test	0.081	08:31
Air Blank	0.000	08:32
Subject Sample # 1	0.000	08:32
Breath Volume	1.492	08:32
Air Blank	0.000	08:33
Air Blank	0.000	08:35
Subject Sample # 2	0.000	08:35
Breath Volume	1.937	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,A,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 11/24/2023 02:48:56

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.000	02:44
Breath Volume	2.636	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.000	02:47
Breath Volume	2.203	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 12/05/2023 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,KIARA,F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BAGGETT,TYLER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DOCKERY

DATE/TIME OF TEST: 11/26/2023 02:47:59

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1 0.173	02:36	
Breath Volume	1.175	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.167	02:42	Volume Not Met
Breath Volume	0.164	02:42	
Air Blank	0.000	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 3 0.159	02:46	
Breath Volume	1.117	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 12/05/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELONY,ANATUS,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: BELONY,ANATUS
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 8223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/26/2023 03:46:06

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1 0.123	03:39	Volume Not Met
Breath Volume	0.628	03:39	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample	# 2	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 12/05/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RAMOS, CARLOS
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 8223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: OSORNO LOPEZ

DATE/TIME OF TEST: 11/26/2023 18:28:10

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.081	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.150	18:22
Breath Volume	1.625	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.149	18:26
Breath Volume	1.980	18:26
Air Blank	0.000	18:27
Control Test	0.080	18:27
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 12/05/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILLS, MARK, A
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: SILLS, MARK
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 8223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: FRANSON

DATE/TIME OF TEST: 11/27/2023 13:08:03

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.080	13:01
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:03
Breath Volume	1.609	13:03
Air Blank	0.000	13:03
Air Blank	0.000	13:05
Subject Sample # 2	0.000	13:06
Breath Volume	1.320	13:06
Air Blank	0.000	13:07
Control Test	0.080	13:07
Air Blank	0.000	13:07
Diagnostics Check	OK	13:08

Data Download Date/Time: 12/05/2023 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTZER,ETHAN,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GUTHRIE,ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 11/15/2023 02:14:55

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.076	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.240	02:10
Breath Volume	2.667	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.239	02:13
Breath Volume	2.675	02:13
Air Blank	0.000	02:14
Control Test	0.076	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 11/16/2023 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,NICKOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212513
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 11/02/2023 14:58:54

Diagnostics Check	OK	14:50	
Air Blank	0.000	14:50	
Control Test	0.080	14:51	
Air Blank	0.000	14:51	
Subject Sample	# 1 0.258	14:54	Volume Not Met
Breath Volume	0.078	14:54	
Air Blank	0.000	14:55	
Air Blank	0.000	14:57	
Subject Sample	# 2	14:57	Subject Test Refused
Breath Volume	0.000	14:57	
Air Blank	0.000	14:58	
Control Test	0.079	14:58	
Air Blank	0.000	14:58	
Diagnostics Check	OK	14:58	

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEILY,STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: KIRKPATRICK

DATE/TIME OF TEST: 11/03/2023 22:52:45

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.148	22:48
Breath Volume	1.453	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.145	22:51
Breath Volume	1.398	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CORACORA

DATE/TIME OF TEST: 11/05/2023 01:47:00

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.180	01:42
Breath Volume	1.437	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.180	01:45
Breath Volume	2.492	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:47

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: BONIN

DATE/TIME OF TEST: 11/05/2023 16:34:48

Diagnostics Check	OK	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Subject Sample # 1	0.000	16:30
Breath Volume	2.226	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.000	16:33
Breath Volume	2.105	16:33
Air Blank	0.000	16:33
Control Test	0.080	16:34
Air Blank	0.000	16:34
Diagnostics Check	OK	16:34

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GOMEZ, RAMON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GETCHIS

DATE/TIME OF TEST: 11/10/2023 10:31:40

Diagnostics Check	OK	10:25
Air Blank	0.000	10:26
Control Test	0.080	10:26
Air Blank	0.000	10:26
Subject Sample # 1	0.000	10:27
Breath Volume	1.500	10:27
Air Blank	0.000	10:27
Air Blank	0.000	10:29
Subject Sample # 2	0.000	10:30
Breath Volume	1.996	10:30
Air Blank	0.000	10:30
Control Test	0.080	10:31
Air Blank	0.000	10:31
Diagnostics Check	OK	10:31

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 11/11/2023 01:14:36

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.150	01:10
Breath Volume	1.812	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.147	01:13
Breath Volume	1.222	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/16/2023 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,NICKOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/24/2023 15:49:28

Diagnostics Check	OK	15:37	
Air Blank	0.000	15:38	
Control Test	0.079	15:38	
Air Blank	0.000	15:38	
Subject Sample	# 1 0.200	15:41	Volume Not Met
Breath Volume	0.179	15:41	
Air Blank	0.000	15:42	
Air Blank	0.000	15:44	
Subject Sample	# 2 0.201	15:44	
Breath Volume	3.023	15:44	
Air Blank	0.000	15:45	
Air Blank	0.000	15:47	
Subject Sample	# 3 0.210	15:47	
Breath Volume	3.105	15:47	
Air Blank	0.000	15:48	
Control Test	0.080	15:48	
Air Blank	0.000	15:49	
Diagnostics Check	OK	15:49	

Data Download Date/Time: 12/05/2023 10:05