

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: BARRENECHE,YVETTE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ARCENTALES

DATE/TIME OF TEST: 12/04/2023 02:25:46

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 12/12/2023 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ORTEGA,ALEXANDER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: LOVO

DATE/TIME OF TEST: 12/05/2023 15:32:11

Diagnostics Check	OK	15:23	
Air Blank	0.000	15:24	
Control Test	0.081	15:24	
Air Blank	0.000	15:25	
Subject Sample	# 1	15:28	No Sample Provided
Breath Volume	0.000	15:28	
Air Blank	0.000	15:28	
Air Blank	0.000	15:30	
Subject Sample	# 2	15:30	Subject Test Refused
Breath Volume	0.000	15:30	
Air Blank	0.000	15:31	
Control Test	0.081	15:31	
Air Blank	0.000	15:32	
Diagnostics Check	OK	15:32	

Data Download Date/Time: 12/12/2023 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEMAN,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ELLIOTT,BRYANT
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 12/17/2023 03:28:21

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.164	03:23
Breath Volume	3.152	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.161	03:26
Breath Volume	2.242	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/11/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENCY,ALAIN,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CLARKE,MICHEAL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/23/2023 03:18:36

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 01/11/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOSO,ANGEL,E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 212641
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 12/31/2023 07:41:19

Diagnostics Check	OK	07:32	
Air Blank	0.000	07:33	
Control Test	0.081	07:33	
Air Blank	0.000	07:34	
Subject Sample	# 1 0.160	07:37	Volume Not Met
Breath Volume	0.089	07:37	
Air Blank	0.000	07:38	
Air Blank	0.000	07:39	
Subject Sample	# 2	07:39	Subject Test Refused
Breath Volume	0.000	07:39	
Air Blank	0.000	07:40	
Control Test	0.077	07:40	
Air Blank	0.000	07:41	
Diagnostics Check	OK	07:41	

Data Download Date/Time: 01/11/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCLJOSE,F
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: TUCCLJOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/02/2023 22:23:05

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:13	
Control Test	0.078	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:17	No Sample Provided
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:22	
Control Test	0.080	22:22	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGOS, GUILLERMO, E
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PALACIO, LIVAN
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/09/2023 22:59:24

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:55	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VILLAMIL,NICHOLAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CUERVO

DATE/TIME OF TEST: 12/15/2023 02:24:30

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.080	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KOVARY

DATE/TIME OF TEST: 12/18/2023 21:13:20

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.079	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:11	Subject Test Refused
Breath Volume	0.000	21:11	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE, LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: PRUDOT

DATE/TIME OF TEST: 12/21/2023 21:48:09

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:47	
Control Test	0.079	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 12/27/2023 00:24:43

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.136	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: HACKWORTH,EARL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 12/28/2023 23:14:23

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: VILLAMIL,NICHOLAS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/30/2023 04:05:24

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 01/03/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKEWELL,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: BAKEWELL,ZACHARY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: CALERO

DATE/TIME OF TEST: 12/04/2023 06:57:55

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.080	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.188	06:52
Breath Volume	1.445	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.201	06:56
Breath Volume	1.152	06:56
Air Blank	0.000	06:56
Control Test	0.080	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANJUL,MAXIMO,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: FANJUL,MAXIMO
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GUILLAMA

DATE/TIME OF TEST: 12/04/2023 22:58:22

Diagnostics Check	OK	22:57	
Air Blank		22:57	Ambient Fail
Air Blank	0.000	22:58	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANJUL,MAXIMO,D
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: FANJUL,MAXIMO
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GUILLAMA

DATE/TIME OF TEST: 12/04/2023 23:07:29

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.215	23:02
Breath Volume	1.925	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.216	23:05
Breath Volume	2.363	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIN,SHELSEA,
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: MARIN,SHELSEA
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: GILBERTO

DATE/TIME OF TEST: 12/05/2023 21:50:02

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:39	
Control Test	0.079	21:40	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:43	No Sample Provided
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 2	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.079	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: LOPES-VIEIRA

DATE/TIME OF TEST: 12/11/2023 09:38:20

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.079	09:32
Air Blank	0.000	09:33
Subject Sample # 1	0.219	09:33
Breath Volume	1.519	09:33
Air Blank	0.000	09:34
Air Blank	0.000	09:36
Subject Sample # 2	0.211	09:36
Breath Volume	2.195	09:36
Air Blank	0.000	09:37
Control Test	0.080	09:37
Air Blank	0.000	09:38
Diagnostics Check	OK	09:38

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANJUL,MAXIMO,
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: FANJUL,MAXIMO
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ESPITIA

DATE/TIME OF TEST: 12/18/2023 01:42:59

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1 0.156	01:35	Volume Not Met
Breath Volume	0.054	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2 0.141	01:41	Volume Not Met
Breath Volume	0.386	01:41	
Air Blank	0.000	01:42	
Control Test	0.078	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANJUL,MAXIMO,
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: FANJUL,MAXIMO
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ESPITIA

DATE/TIME OF TEST: 12/18/2023 01:57:29

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1 0.144	01:50	Volume Not Met
Breath Volume	0.367	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 2 0.115	01:55	Volume Not Met
Breath Volume	0.777	01:55	
Air Blank	0.000	01:56	
Control Test	0.077	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGO,TOMAS,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REGO,TOMAS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: AROCHA ROA

DATE/TIME OF TEST: 12/20/2023 06:26:12

Diagnostics Check	OK	06:25	
Air Blank		06:25	RFI Detect
Air Blank	0.000	06:26	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGO,TOMAS,
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: REGO,TOMAS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: AROCHA ROA

DATE/TIME OF TEST: 12/20/2023 06:58:22

Diagnostics Check	OK	06:46	
Air Blank	0.000	06:47	
Control Test	0.079	06:47	
Air Blank	0.000	06:48	
Subject Sample	# 1 0.130	06:51	Volume Not Met
Breath Volume	0.121	06:51	
Air Blank	0.000	06:51	
Air Blank	0.000	06:53	
Subject Sample	# 2 0.154	06:56	Volume Not Met
Breath Volume	0.015	06:56	
Air Blank	0.000	06:57	
Control Test	0.080	06:57	
Air Blank	0.000	06:58	
Diagnostics Check	OK	06:58	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGO,TOMAS,
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: REGO,TOMAS
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: AROCHA ROA

DATE/TIME OF TEST: 12/20/2023 07:20:52

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:17	
Control Test	0.078	07:17	
Air Blank	0.000	07:17	
Subject Sample	# 1	07:19	Subject Test Refused
Breath Volume	0.089	07:19	
Air Blank	0.000	07:19	
Control Test	0.079	07:20	
Air Blank	0.000	07:20	
Diagnostics Check	OK	07:20	

Data Download Date/Time: 01/11/2024 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SALGADO RODRIGUEZ

DATE/TIME OF TEST: 12/09/2023 00:36:38

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:23	
Control Test	0.079	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1 0.116	00:27	Volume Not Met
Breath Volume	0.675	00:27	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2 0.133	00:30	
Breath Volume	1.613	00:30	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample	# 3 0.113	00:35	
Breath Volume	1.558	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 01/04/2024 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TORIBIO

DATE/TIME OF TEST: 12/16/2023 01:31:46

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.195	01:24
Breath Volume	1.589	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.218	01:27
Breath Volume	1.937	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 3	0.211	01:30
Breath Volume	2.078	01:30
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/04/2024 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SHILON

DATE/TIME OF TEST: 12/17/2023 23:26:33

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.076	23:21	
Air Blank	0.000	23:21	
Subject Sample # 1	0.187	23:22	
Breath Volume	1.363	23:22	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample # 2	0.176	23:25	
Breath Volume	1.566	23:25	
Air Blank	0.000	23:25	
Control Test	0.073	23:25	Control Outside Tolerance
Air Blank	0.000	23:26	

Data Download Date/Time: 01/04/2024 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SHILON

DATE/TIME OF TEST: 12/17/2023 23:31:33

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.073	23:30	Control Outside Tolerance
Air Blank	0.000	23:31	

Data Download Date/Time: 01/04/2024 04:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/27/2023 18:12:10

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.078	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.000	18:07
Breath Volume	2.480	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 2	0.000	18:10
Breath Volume	2.453	18:10
Air Blank	0.000	18:11
Control Test	0.080	18:11
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

Data Download Date/Time: 01/04/2024 04:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 12/29/2023 00:46:43

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.078	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.077	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 01/04/2024 04:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,SHAUNETTA,J
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: NOWACK,ANGELICA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RODRIGUEZ TERUEL

DATE/TIME OF TEST: 12/08/2023 00:59:36

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:58	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTROV,VANESSA,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OSTROV,VANESSA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: FAITH

DATE/TIME OF TEST: 12/14/2023 02:01:06

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.162	01:56
Breath Volume	1.472	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.154	01:59
Breath Volume	1.277	01:59
Air Blank	0.000	02:00
Control Test	0.075	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: RUIZ,DANIEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: RAMIREZ-DOMINGO

DATE/TIME OF TEST: 12/17/2023 03:31:42

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.224	03:26
Breath Volume	1.390	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.223	03:29
Breath Volume	1.363	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODUM,ICE,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DORCENT,GEDSON
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MATTIS

DATE/TIME OF TEST: 12/17/2023 06:08:43

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:05	
Control Test	0.079	06:05	
Air Blank	0.000	06:05	
Subject Sample	# 1 0.179	06:06	
Breath Volume	3.242	06:06	
Air Blank		06:07	RFI Detect
Air Blank	0.000	06:07	
Control Test	0.078	06:08	
Air Blank	0.000	06:08	
Diagnostics Check	OK	06:08	

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ODUM,ICE,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DORCENT,GEDSON
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MATTIS

DATE/TIME OF TEST: 12/17/2023 06:17:26

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.077	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.186	06:12
Breath Volume	2.695	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.188	06:15
Breath Volume	2.675	06:15
Air Blank	0.000	06:16
Control Test	0.077	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER,HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 12/18/2023 01:28:51

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.079	01:27	
Air Blank		01:28	RFI Detect
Air Blank	0.000	01:28	

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER,HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 12/18/2023 01:34:26

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: RUIZ,DANIEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 11/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MENDOZA-PASTRANA

DATE/TIME OF TEST: 12/24/2023 04:41:41

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.131	04:37
Breath Volume	3.050	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.130	04:40
Breath Volume	2.734	04:40
Air Blank	0.000	04:40
Control Test	0.077	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 12/29/2023 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: BUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/02/2023 17:34:15

Diagnostics Check	OK	17:31	
Air Blank	0.000	17:31	
Control Test	0.080	17:32	
Air Blank	0.000	17:32	
Subject Sample	# 1	17:32	Subject Test Refused
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Control Test	0.080	17:33	
Air Blank	0.000	17:34	
Diagnostics Check	OK	17:34	

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FARRINGTON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/03/2023 04:33:03

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERY,DIEDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 12/03/2023 23:09:55

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAMAN,KELLEY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BERNER

DATE/TIME OF TEST: 12/04/2023 01:35:59

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.184	01:31
Breath Volume	1.906	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.182	01:34
Breath Volume	1.429	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,ROBERT,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MACK,JAVON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/04/2023 01:51:51

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.184	01:47
Breath Volume	1.878	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.186	01:50
Breath Volume	1.746	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ABOUZALAM

DATE/TIME OF TEST: 12/04/2023 02:21:52

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.080	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 12/20/2023 09:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZSIDO,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GATICA

DATE/TIME OF TEST: 12/05/2023 02:23:22

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.086	02:18
Breath Volume	1.445	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.088	02:21
Breath Volume	1.910	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: NELEPOVITZ

DATE/TIME OF TEST: 12/05/2023 03:12:10

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.046	03:07
Breath Volume	3.542	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.044	03:10
Breath Volume	3.902	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,L,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROSS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 12/06/2023 02:47:41

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.081	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSTER,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SULAIMAN,RAMIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: UITHOVEN

DATE/TIME OF TEST: 12/07/2023 13:33:26

Diagnostics Check	OK	13:26
Air Blank	0.000	13:27
Control Test	0.081	13:27
Air Blank	0.000	13:28
Subject Sample # 1	0.383	13:29
Breath Volume	1.199	13:29
Air Blank	0.000	13:29
Air Blank	0.000	13:31
Subject Sample # 2	0.385	13:31
Breath Volume	1.910	13:31
Air Blank	0.000	13:32
Control Test	0.078	13:32
Air Blank	0.000	13:33
Diagnostics Check	OK	13:33

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 12/08/2023 01:20:30

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: JAEGER

DATE/TIME OF TEST: 12/08/2023 09:40:44

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.080	09:32
Air Blank	0.000	09:33
Subject Sample # 1	0.098	09:35
Breath Volume	3.339	09:35
Air Blank	0.000	09:36
Air Blank	0.000	09:38
Subject Sample # 2	0.096	09:39
Breath Volume	2.398	09:39
Air Blank	0.000	09:39
Control Test	0.079	09:40
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VILLALPANDO

DATE/TIME OF TEST: 12/08/2023 23:44:01

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.124	23:39
Breath Volume	2.398	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.117	23:42
Breath Volume	1.789	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATA-ROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: THIES-IGLESIAS

DATE/TIME OF TEST: 12/09/2023 01:19:23

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.161	01:13
Breath Volume	2.000	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.152	01:17
Breath Volume	1.234	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTOR,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/09/2023 04:42:31

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.163	04:37
Breath Volume	2.253	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.163	04:41
Breath Volume	1.789	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: DOMINGUEZ-MARTINEZ

DATE/TIME OF TEST: 12/09/2023 05:57:14

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:55	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:55	Subject Test Refused
Breath Volume	0.000	05:55	
Air Blank	0.000	05:56	
Control Test	0.080	05:56	
Air Blank	0.000	05:57	
Diagnostics Check	OK	05:57	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRADY,J
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 12/09/2023 21:54:15

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.236	21:49
Breath Volume	1.648	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.222	21:52
Breath Volume	1.601	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 12/10/2023 00:02:36

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.348	23:58
Breath Volume	1.406	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.339	00:01
Breath Volume	1.621	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAPPIELLO,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SAVRANAKIS

DATE/TIME OF TEST: 12/11/2023 22:34:00

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: NIERENGARTEN

DATE/TIME OF TEST: 12/12/2023 01:16:46

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNOR,BRENNAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DEYO

DATE/TIME OF TEST: 12/12/2023 02:35:21

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.221	02:30
Breath Volume	1.714	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.216	02:33
Breath Volume	2.175	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 12/13/2023 03:45:23

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAMPION,KRISTIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: SPRAGUE

DATE/TIME OF TEST: 12/13/2023 16:47:07

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:44	
Control Test	0.080	16:45	
Air Blank	0.000	16:45	
Subject Sample	# 1	16:45	Subject Test Refused
Breath Volume	0.000	16:45	
Air Blank	0.000	16:46	
Control Test	0.080	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: PRUNA JR

DATE/TIME OF TEST: 12/14/2023 00:13:02

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.000	00:08
Breath Volume	1.261	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:11
Breath Volume	1.304	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:13

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NONTHE MARTINEZ

DATE/TIME OF TEST: 12/15/2023 03:30:56

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.081	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: TRUNK

DATE/TIME OF TEST: 12/15/2023 21:09:02

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.000	21:04
Breath Volume	1.226	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:07
Breath Volume	1.671	21:07
Air Blank	0.000	21:08
Control Test	0.081	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:09

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: KREUTER

DATE/TIME OF TEST: 12/16/2023 04:24:14

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.103	04:19
Breath Volume	1.464	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.107	04:22
Breath Volume	1.175	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/16/2023 05:25:57

Diagnostics Check	OK	05:23	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:24	
Subject Sample	# 1	05:24	Subject Test Refused
Breath Volume	0.000	05:24	
Air Blank	0.000	05:25	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/16/2023 22:37:35

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.137	22:33
Breath Volume	2.628	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.135	22:36
Breath Volume	1.687	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: EARLEY

DATE/TIME OF TEST: 12/16/2023 23:52:04

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAMAN,KELLY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/16/2023 23:59:44

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Control Test	0.080	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 12/17/2023 01:28:44

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.146	01:23
Breath Volume	1.414	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.153	01:27
Breath Volume	1.281	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BREUER

DATE/TIME OF TEST: 12/17/2023 19:46:17

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.080	19:40
Air Blank	0.000	19:41
Subject Sample # 1	0.046	19:41
Breath Volume	1.503	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:44
Subject Sample # 2	0.046	19:44
Breath Volume	1.328	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/17/2023 23:46:26

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/18/2023 05:06:19

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.080	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 12/20/2023 00:49:06

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.080	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.265	00:44
Breath Volume	3.234	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.256	00:47
Breath Volume	2.101	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 12/20/2023 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: TRESH

DATE/TIME OF TEST: 12/20/2023 23:10:29

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.000	23:05
Breath Volume	1.539	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:09
Breath Volume	1.691	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 12/21/2023 04:21:07

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.080	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 12/21/2023 21:04:27

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.080	21:02	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:03	Subject Test Refused
Breath Volume	0.000	21:03	
Air Blank	0.000	21:03	
Control Test	0.080	21:03	
Air Blank	0.000	21:04	
Diagnostics Check	OK	21:04	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 12/22/2023 05:48:39

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.000	05:44
Breath Volume	1.238	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.000	05:47
Breath Volume	1.492	05:47
Air Blank	0.000	05:47
Control Test	0.081	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/22/2023 23:14:40

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	2.031	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	1.839	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/23/2023 03:10:46

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,T,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPENCER,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/23/2023 03:41:59

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	No Sample Provided
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:40	Subject Test Refused
Breath Volume	0.000	03:40	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILSON,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: POPPLETON

DATE/TIME OF TEST: 12/23/2023 23:52:16

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.080	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:45	No Sample Provided
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2	23:50	No Sample Provided
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GEOGAN

DATE/TIME OF TEST: 12/24/2023 02:39:06

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOPPE,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: DEL VALLE

DATE/TIME OF TEST: 12/25/2023 00:02:58

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 2 0.144	23:50	
Breath Volume	1.105	23:50	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 3 0.132	23:55	Volume Not Met
Breath Volume	0.265	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1 0.123	00:01	Volume Not Met
Breath Volume	0.363	00:01	
Air Blank	0.000	00:02	
Control Test	0.079	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIDY,DEPUTY,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOPPE,A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: DEL VALLE

DATE/TIME OF TEST: 12/25/2023 00:15:46

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.133	00:09
Breath Volume	2.089	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.130	00:14
Breath Volume	1.488	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: DANNEO

DATE/TIME OF TEST: 12/26/2023 19:32:01

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.000	19:27
Breath Volume	1.593	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.000	19:30
Breath Volume	1.140	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:32

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DELLECHIAIE,DAKOTA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/27/2023 03:10:16

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.081	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAVER,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: GUARINO

DATE/TIME OF TEST: 12/27/2023 22:48:41

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.000	22:44
Breath Volume	1.476	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:47
Breath Volume	1.699	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 12/28/2023 04:23:54

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:23	
Control Test	0.079	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CORNELL

DATE/TIME OF TEST: 12/29/2023 01:40:56

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.152	01:36
Breath Volume	2.851	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.148	01:39
Breath Volume	2.699	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: COSGROVE

DATE/TIME OF TEST: 12/29/2023 12:42:03

Diagnostics Check	OK	12:35
Air Blank	0.000	12:36
Control Test	0.080	12:36
Air Blank	0.000	12:37
Subject Sample # 1	0.244	12:37
Breath Volume	2.023	12:37
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.259	12:40
Breath Volume	1.472	12:40
Air Blank	0.000	12:41
Control Test	0.078	12:41
Air Blank	0.000	12:41
Diagnostics Check	OK	12:42

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: BRADHAM

DATE/TIME OF TEST: 12/29/2023 19:13:42

Diagnostics Check	OK	19:10	
Air Blank	0.000	19:11	
Control Test	0.080	19:11	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:12	
Control Test	0.081	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/29/2023 20:56:13

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.081	20:51
Breath Volume	3.085	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.080	20:54
Breath Volume	1.843	20:54
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEON GONZALEZ

DATE/TIME OF TEST: 12/30/2023 03:25:48

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.156	03:19
Breath Volume	2.273	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.167	03:24
Breath Volume	1.464	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 12/30/2023 07:54:03

Diagnostics Check	OK	07:46
Air Blank	0.000	07:46
Control Test	0.080	07:47
Air Blank	0.000	07:47
Subject Sample # 1	0.000	07:49
Breath Volume	1.664	07:49
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.000	07:52
Breath Volume	1.242	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:53
Diagnostics Check	OK	07:54

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLE,STEPHAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLE,STEPHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 12/31/2023 02:56:21

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:49	No Sample Provided
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2	02:54	No Sample Provided
Breath Volume	0.000	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: HEMINGWAY

DATE/TIME OF TEST: 12/31/2023 03:41:24

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.121	03:36
Breath Volume	1.699	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.120	03:39
Breath Volume	1.621	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 12/31/2023 05:59:55

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:49	
Control Test	0.080	05:50	
Air Blank	0.000	05:50	
Subject Sample	# 1	05:53	No Sample Provided
Breath Volume	0.000	05:53	
Air Blank	0.000	05:54	
Air Blank	0.000	05:56	
Subject Sample	# 2	05:58	Subject Test Refused
Breath Volume	0.000	05:58	
Air Blank	0.000	05:59	
Control Test	0.081	05:59	
Air Blank	0.000	05:59	
Diagnostics Check	OK	05:59	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: CRONIN

DATE/TIME OF TEST: 12/31/2023 06:38:54

Diagnostics Check	OK	06:31
Air Blank	0.000	06:32
Control Test	0.080	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.192	06:33
Breath Volume	1.734	06:33
Air Blank	0.000	06:34
Air Blank	0.000	06:35
Subject Sample # 2	0.179	06:37
Breath Volume	1.816	06:37
Air Blank	0.000	06:38
Control Test	0.077	06:38
Air Blank	0.000	06:38
Diagnostics Check	OK	06:38

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORRIS,BRIDGETT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: ROSATI

DATE/TIME OF TEST: 12/31/2023 06:54:07

Diagnostics Check	OK	06:51	
Air Blank	0.000	06:51	
Control Test	0.081	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1	06:52	Subject Test Refused
Breath Volume	0.000	06:52	
Air Blank	0.000	06:53	
Control Test	0.080	06:53	
Air Blank	0.000	06:54	
Diagnostics Check	OK	06:54	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/31/2023 13:07:06

Diagnostics Check	OK	13:00
Air Blank	0.000	13:01
Control Test	0.080	13:01
Air Blank	0.000	13:02
Subject Sample # 1	0.152	13:02
Breath Volume	1.738	13:02
Air Blank	0.000	13:03
Air Blank	0.000	13:04
Subject Sample # 2	0.154	13:05
Breath Volume	1.859	13:05
Air Blank	0.000	13:06
Control Test	0.080	13:06
Air Blank	0.000	13:07
Diagnostics Check	OK	13:07

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COYLE,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: BORGERSEN

DATE/TIME OF TEST: 12/31/2023 17:29:44

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:27	
Control Test	0.080	17:27	
Air Blank	0.000	17:28	
Subject Sample	# 1	17:28	Subject Test Refused
Breath Volume	0.000	17:28	
Air Blank	0.000	17:28	
Control Test	0.080	17:29	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:29	

Data Download Date/Time: 01/17/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN,DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN,DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/02/2023 03:02:15

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: TEASLEY

DATE/TIME OF TEST: 12/07/2023 01:16:14

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.140	01:11
Breath Volume	2.125	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.134	01:14
Breath Volume	1.988	01:14
Air Blank	0.000	01:15
Control Test	0.083	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SCHEELE

DATE/TIME OF TEST: 12/07/2023 02:12:39

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.082	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.083	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/07/2023 04:26:16

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.082	04:24	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.082	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: KOEPKE

DATE/TIME OF TEST: 12/09/2023 02:33:04

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.117	02:28
Breath Volume	1.515	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.117	02:31
Breath Volume	1.347	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:33

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: ABALOS

DATE/TIME OF TEST: 12/09/2023 19:15:06

Diagnostics Check	OK	19:12	
Air Blank	0.000	19:12	
Control Test	0.081	19:12	
Air Blank	0.000	19:13	
Subject Sample	# 1	19:13	Subject Test Refused
Breath Volume	0.000	19:13	
Air Blank	0.000	19:14	
Control Test	0.081	19:14	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:15	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: DENGES

DATE/TIME OF TEST: 12/09/2023 22:25:57

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.029	22:21
Breath Volume	2.195	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.025	22:24
Breath Volume	3.082	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 12/14/2023 14:53:47

Diagnostics Check	OK	14:47
Air Blank	0.000	14:47
Control Test	0.081	14:47
Air Blank	0.000	14:48
Subject Sample # 1	0.169	14:48
Breath Volume	1.250	14:48
Air Blank	0.000	14:49
Air Blank	0.000	14:51
Subject Sample # 2	0.159	14:52
Breath Volume	1.417	14:52
Air Blank	0.000	14:52
Control Test	0.082	14:53
Air Blank	0.000	14:53
Diagnostics Check	OK	14:53

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/14/2023 23:10:13

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.081	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:10	
Diagnostics Check	OK	23:10	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CHILDS

DATE/TIME OF TEST: 12/15/2023 02:21:05

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.015	02:16
Breath Volume	1.164	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.016	02:19
Breath Volume	1.414	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HAMMON

DATE/TIME OF TEST: 12/16/2023 00:53:52

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 12/20/2023 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: KUEBLER

DATE/TIME OF TEST: 12/22/2023 02:19:52

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.082	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 01/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 12/23/2023 01:27:12

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.180	01:22
Breath Volume	2.460	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.183	01:25
Breath Volume	2.273	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/25/2023 23:40:34

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 01/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: DELGAUDIO

DATE/TIME OF TEST: 12/28/2023 00:50:38

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.068	00:46
Breath Volume	1.417	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.062	00:49
Breath Volume	1.378	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HERSEY

DATE/TIME OF TEST: 12/28/2023 23:26:05

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:23	
Control Test	0.081	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Control Test	0.082	23:25	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 01/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/31/2023 23:07:04

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.082	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.031	23:02
Breath Volume	2.210	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.031	23:05
Breath Volume	2.687	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:07

Data Download Date/Time: 01/17/2024 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,L,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,L
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: KLIER

DATE/TIME OF TEST: 12/04/2023 23:28:20

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.081	23:26	
Air Blank	0.000	23:26	
Subject Sample	# 1	23:26	Subject Test Refused
Breath Volume	0.000	23:26	
Air Blank	0.000	23:27	
Control Test	0.081	23:27	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 12/20/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 12/09/2023 01:37:30

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.108	01:32
Breath Volume	1.535	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.108	01:35
Breath Volume	1.667	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/20/2023 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: COWARD

DATE/TIME OF TEST: 12/20/2023 23:06:53

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 01/17/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TIPTON

DATE/TIME OF TEST: 12/24/2023 02:51:50

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.080	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 01/17/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ
VILLANUEVA,ANTONIO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FERNANDEZ
VILLANUEVA,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024

VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male

OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/27/2023 22:06:25

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.215	22:01
Breath Volume	1.761	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.227	22:04
Breath Volume	2.558	22:04
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 01/17/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RAMSEY,DAVID,L

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: CHRISTMAS,FOLIN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 11/10/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: DUI

ARREST TIME: 21:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BOILLAT

DATE/TIME OF TEST: 12/01/2023 22:59:36

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.165	22:55
Breath Volume	2.414	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.170	22:58
Breath Volume	2.164	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 12/29/2023 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,PABLO,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MALLONE,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 12/30/2023 01:10:33

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.105	01:05
Breath Volume	1.125	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.114	01:08
Breath Volume	1.535	01:08
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 01/18/2024 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SPEICH, R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 11/10/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/02/2023 04:13:14

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.190	04:08
Breath Volume	1.210	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.189	04:11
Breath Volume	1.339	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/05/2023 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER JR,S,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JOSEPH,W
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LIGHTHART

DATE/TIME OF TEST: 12/07/2023 01:31:12

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.203	01:26
Breath Volume	4.187	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:28
Subject Sample # 2	0.203	01:29
Breath Volume	3.445	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMANDO,TINA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WALKER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SCHIEFERLE

DATE/TIME OF TEST: 12/07/2023 03:25:42

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.156	03:21
Breath Volume	1.480	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.159	03:24
Breath Volume	2.746	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL, ADAM, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VERREAULT, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: LICONAJUAREZ

DATE/TIME OF TEST: 12/07/2023 23:29:33

Diagnostics Check	OK	23:24	
Air Blank	0.000	23:24	
Control Test	0.079	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.062	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: FERRY

DATE/TIME OF TEST: 12/10/2023 23:15:18

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.128	23:10
Breath Volume	2.578	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.128	23:13
Breath Volume	1.730	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GOODBRED,K
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 12/14/2023 00:04:06

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Subject Sample # 2	0.162	23:59
Breath Volume	1.437	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.161	00:02
Breath Volume	1.246	00:02
Air Blank	0.000	00:03
Control Test	0.078	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 12/16/2023 00:08:28

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.167	00:04
Breath Volume	2.187	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.178	00:06
Breath Volume	3.656	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZTRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,ORELYS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MARTINEZ CARIZALEZ

DATE/TIME OF TEST: 12/17/2023 03:49:37

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.259	03:43
Breath Volume	2.156	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.255	03:47
Breath Volume	2.457	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO,JULIO,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HERNANDEZ,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06021080A2
DRY GAS EXPIRY DATE: 04/05/2023
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CASTIL HERNANDEZ

DATE/TIME OF TEST: 12/18/2023 06:02:07

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.080	05:56
Air Blank	0.000	05:56
Subject Sample # 1	0.076	05:57
Breath Volume	1.734	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample # 2	0.077	06:00
Breath Volume	1.691	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDASH, MATTHEW, B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VALEZ, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 12/19/2023 02:28:11

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.150	02:23
Breath Volume	2.531	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.148	02:26
Breath Volume	2.015	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIFF,DEVIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 12/20/2023 00:32:10

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.225	00:27
Breath Volume	2.238	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.222	00:30
Breath Volume	1.851	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SUDANO,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PASICZNYK

DATE/TIME OF TEST: 12/23/2023 01:08:19

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.092	01:03
Breath Volume	1.562	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.092	01:06
Breath Volume	1.410	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HOSIER

DATE/TIME OF TEST: 12/23/2023 02:59:31

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.097	02:55
Breath Volume	2.984	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.093	02:57
Breath Volume	1.148	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,DEAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PARKS,DEAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: LOPEZCRISOSTOMO

DATE/TIME OF TEST: 12/23/2023 20:09:54

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.079	20:04
Air Blank	0.000	20:04
Subject Sample # 1	0.221	20:05
Breath Volume	1.570	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.233	20:08
Breath Volume	2.449	20:08
Air Blank	0.000	20:08
Control Test	0.079	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 12/28/2023 05:08:04

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.177	05:03
Breath Volume	1.738	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.192	05:06
Breath Volume	1.488	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:08

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUILES,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LOVOS

DATE/TIME OF TEST: 12/29/2023 05:17:10

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.132	05:11
Breath Volume	1.417	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.121	05:15
Breath Volume	1.667	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VELEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: PIEDADE

DATE/TIME OF TEST: 12/30/2023 01:40:39

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.143	01:36
Breath Volume	1.167	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.140	01:39
Breath Volume	1.863	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ TRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ESTEBAN ALVAREZ

DATE/TIME OF TEST: 12/31/2023 23:09:24

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.205	23:01
Breath Volume	1.324	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.182	23:04
Breath Volume	1.890	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 3	0.202	23:07
Breath Volume	1.925	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 01/10/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALLADINO,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BLACKMAN,ANTHONY
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 11/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A4
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/09/2023 20:38:36

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.077	20:33	
Air Blank	0.000	20:33	
Subject Sample	# 1	20:37	No Sample Provided
Breath Volume	0.000	20:37	
Air Blank		20:37	Purge Fail
Air Blank		20:38	Purge Fail

Data Download Date/Time: 12/24/2023 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA, MARIANA, Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SCHIPANSKY, CHAD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 11/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A4
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: DAVITO

DATE/TIME OF TEST: 12/21/2023 21:00:38

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:58	
Control Test	0.078	20:58	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.079	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 12/24/2023 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SCHIPANSKY,CHAD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 12/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A4
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 12/30/2023 18:04:25

Diagnostics Check	OK	17:57
Air Blank	0.000	17:58
Control Test	0.078	17:58
Air Blank	0.000	17:59
Subject Sample # 1	0.037	17:59
Breath Volume	2.230	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.034	18:02
Breath Volume	2.335	18:02
Air Blank	0.000	18:03
Control Test	0.077	18:03
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

Data Download Date/Time: 01/15/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL,HEATHER,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL,HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/02/2023 16:30:03

Diagnostics Check	OK	16:15	
Air Blank	0.000	16:15	
Control Test	0.081	16:16	
Air Blank	0.000	16:16	
Subject Sample	# 1 0.114	16:17	
Breath Volume	4.730	16:17	
Air Blank	0.000	16:18	
Air Blank	0.000	16:19	
Subject Sample	# 2	16:22	No Sample Provided
Breath Volume	0.000	16:22	
Air Blank	0.000	16:23	
Air Blank	0.000	16:25	
Subject Sample	# 3	16:28	No Sample Provided
Breath Volume	0.000	16:28	
Air Blank	0.000	16:29	
Control Test	0.081	16:29	
Air Blank	0.000	16:29	
Diagnostics Check	OK	16:30	

Data Download Date/Time: 12/20/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL,HEATHER,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL,HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/02/2023 16:56:28

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:54	
Control Test	0.082	16:54	
Air Blank	0.000	16:54	
Subject Sample	# 1	16:55	Subject Test Refused
Breath Volume	0.000	16:55	
Air Blank	0.000	16:55	
Control Test	0.081	16:55	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 12/20/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,JACOB,D
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: SCOTT,JACOB
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/03/2023 22:09:48

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.082	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:08	
Control Test	0.082	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 12/20/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,BECKY,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLSON,SCOTT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/10/2023 19:24:38

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:20	
Control Test	0.082	19:20	
Air Blank	0.000	19:21	
Subject Sample	# 1	19:23	Subject Test Refused
Breath Volume	0.000	19:23	
Air Blank	0.000	19:23	
Control Test	0.083	19:24	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 12/20/2023 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BRIAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WARREN,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WALTERMON

DATE/TIME OF TEST: 12/14/2023 02:26:47

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:21
Breath Volume	2.718	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:25
Breath Volume	1.285	02:25
Air Blank	0.000	02:25
Control Test	0.084	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 12/20/2023 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RODRIGUEZ AVILES

DATE/TIME OF TEST: 12/01/2023 03:24:43

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.185	03:18
Breath Volume	1.121	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.181	03:23
Breath Volume	1.113	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 12/01/2023 03:55:20

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.266	03:51
Breath Volume	1.957	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.267	03:53
Breath Volume	2.148	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DESCHRYVER,KYLE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 12/02/2023 02:03:08

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.088	01:58
Breath Volume	1.296	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.087	02:01
Breath Volume	1.332	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CUVILJE

DATE/TIME OF TEST: 12/02/2023 03:43:41

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.078	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.078	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARC,VUKANSON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: CHARACO

DATE/TIME OF TEST: 12/07/2023 02:36:14

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.078	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.172	02:30
Breath Volume	1.312	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.175	02:34
Breath Volume	1.621	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HUSSAIN

DATE/TIME OF TEST: 12/08/2023 02:04:00

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.120	01:59
Breath Volume	3.519	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.112	02:02
Breath Volume	2.656	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OJEDA COLON,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: O'NEILL

DATE/TIME OF TEST: 12/09/2023 03:45:41

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.169	03:41
Breath Volume	1.847	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.166	03:44
Breath Volume	1.187	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: VASQUEZ MORALES

DATE/TIME OF TEST: 12/11/2023 04:18:19

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.239	04:12
Breath Volume	1.855	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.246	04:16
Breath Volume	1.988	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/12/2023 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSO,NICHOLAS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/12/2023 20:52:14

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.165	20:47
Breath Volume	1.703	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.154	20:50
Breath Volume	1.621	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 01/15/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOULTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BOSCAN RINCON

DATE/TIME OF TEST: 12/13/2023 03:42:12

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.185	03:37
Breath Volume	1.605	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.186	03:40
Breath Volume	1.171	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/15/2024 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RADIG,TARAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: GEFFRARD

DATE/TIME OF TEST: 12/13/2023 08:34:33

Diagnostics Check	OK	08:28
Air Blank	0.000	08:28
Control Test	0.078	08:28
Air Blank	0.000	08:29
Subject Sample # 1	0.223	08:30
Breath Volume	1.164	08:30
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.224	08:33
Breath Volume	2.125	08:33
Air Blank	0.000	08:33
Control Test	0.078	08:33
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVER, ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NESTON

DATE/TIME OF TEST: 12/13/2023 22:38:39

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.078	22:33	
Air Blank	0.000	22:33	
Subject Sample # 1	0.357	22:34	
Breath Volume	1.371	22:34	
Air Blank	0.000	22:35	
Air Blank	0.000	22:36	
Subject Sample # 2		22:37	Slope Not Met
Breath Volume	1.109	22:37	
Air Blank	0.000	22:37	
Control Test	0.077	22:38	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVER, ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: NESTON

DATE/TIME OF TEST: 12/13/2023 23:08:48

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.077	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.361	23:04
Breath Volume	2.121	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.351	23:07
Breath Volume	3.101	23:07
Air Blank	0.000	23:07
Control Test	0.078	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMON,JEREMY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: TOURNER

DATE/TIME OF TEST: 12/14/2023 03:33:43

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.078	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MANSO,CRISTOBAL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: NEVAREZ

DATE/TIME OF TEST: 12/14/2023 14:28:37

Diagnostics Check	OK	14:18
Air Blank	0.000	14:19
Control Test	0.077	14:19
Air Blank	0.000	14:20
Subject Sample # 1	0.334	14:21
Breath Volume	1.667	14:21
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.332	14:27
Breath Volume	1.281	14:27
Air Blank	0.000	14:27
Control Test	0.078	14:28
Air Blank	0.000	14:28
Diagnostics Check	OK	14:28

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: RIBEIRO

DATE/TIME OF TEST: 12/15/2023 01:47:39

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.147	01:42
Breath Volume	2.128	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.145	01:46
Breath Volume	1.500	01:46
Air Blank	0.000	01:46
Control Test	0.077	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBSTER, BRAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RATTIGAN

DATE/TIME OF TEST: 12/16/2023 02:19:42

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.078	02:11	
Air Blank	0.000	02:12	
Subject Sample # 1	0.135	02:13	
Breath Volume	1.500	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample # 2		02:18	Slope Not Met
Breath Volume	0.769	02:18	
Air Blank	0.000	02:18	
Control Test	0.078	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBSTER, BRAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RATTIGAN

DATE/TIME OF TEST: 12/16/2023 02:54:44

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.131	02:50
Breath Volume	1.359	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.127	02:53
Breath Volume	1.792	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATOS,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 12/16/2023 05:03:26

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.189	04:58
Breath Volume	2.824	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.174	05:01
Breath Volume	2.316	05:01
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COHEN, JOSHUA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: COLINDRES

DATE/TIME OF TEST: 12/16/2023 08:44:40

Diagnostics Check	OK	08:31	
Air Blank	0.000	08:32	
Control Test	0.078	08:32	
Air Blank	0.000	08:33	
Subject Sample	# 1 0.140	08:36	Volume Not Met
Breath Volume	1.015	08:36	
Air Blank	0.000	08:36	
Air Blank	0.000	08:38	
Subject Sample	# 2 0.142	08:40	
Breath Volume	1.703	08:40	
Air Blank	0.000	08:40	
Air Blank	0.000	08:42	
Subject Sample	# 3 0.133	08:43	
Breath Volume	1.136	08:43	
Air Blank	0.000	08:43	
Control Test	0.077	08:44	
Air Blank	0.000	08:44	
Diagnostics Check	OK	08:44	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CUEVAS,JOSE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: PEREZAROSTEGUI

DATE/TIME OF TEST: 12/16/2023 09:21:25

Diagnostics Check	OK	09:15
Air Blank	0.000	09:15
Control Test	0.078	09:16
Air Blank	0.000	09:16
Subject Sample # 1	0.217	09:17
Breath Volume	3.886	09:17
Air Blank	0.000	09:17
Air Blank	0.000	09:19
Subject Sample # 2	0.216	09:19
Breath Volume	3.492	09:19
Air Blank	0.000	09:20
Control Test	0.078	09:20
Air Blank	0.000	09:21
Diagnostics Check	OK	09:21

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PEREZ SALES

DATE/TIME OF TEST: 12/17/2023 04:31:47

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.216	04:27
Breath Volume	1.753	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.226	04:30
Breath Volume	2.914	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUERRERO,MANUEL
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DELCERRORODRIGUEZ

DATE/TIME OF TEST: 12/17/2023 06:40:16

Diagnostics Check	OK	06:25	
Air Blank	0.000	06:25	
Control Test	0.078	06:26	
Air Blank	0.000	06:26	
Subject Sample	# 1 0.161	06:29	Volume Not Met
Breath Volume	0.488	06:29	
Air Blank	0.000	06:30	
Air Blank	0.000	06:32	
Subject Sample	# 2 0.198	06:33	
Breath Volume	1.148	06:33	
Air Blank	0.000	06:34	
Air Blank	0.000	06:36	
Subject Sample	# 3 0.203	06:38	
Breath Volume	1.558	06:38	
Air Blank	0.000	06:39	
Control Test	0.073	06:39	Control Outside Tolerance
Air Blank	0.000	06:40	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA,NIKOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: QUINTERO JR

DATE/TIME OF TEST: 12/17/2023 07:00:53

Diagnostics Check	OK	06:54
Air Blank	0.000	06:55
Control Test	0.077	06:55
Air Blank	0.000	06:56
Subject Sample # 1	0.120	06:56
Breath Volume	2.226	06:56
Air Blank	0.000	06:57
Air Blank	0.000	06:58
Subject Sample # 2	0.115	06:59
Breath Volume	3.433	06:59
Air Blank	0.000	06:59
Control Test	0.078	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDERO,PATRICIA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 12/19/2023 06:06:46

Diagnostics Check	OK	06:00
Air Blank	0.000	06:00
Control Test	0.078	06:00
Air Blank	0.000	06:01
Subject Sample # 1	0.256	06:02
Breath Volume	1.386	06:02
Air Blank	0.000	06:02
Air Blank	0.000	06:04
Subject Sample # 2	0.273	06:05
Breath Volume	2.097	06:05
Air Blank	0.000	06:05
Control Test	0.078	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:06

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVER,ISAIAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: JEAN PAUL

DATE/TIME OF TEST: 12/20/2023 00:15:15

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1 0.166	00:04	Volume Not Met
Breath Volume	1.093	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:07	
Subject Sample	# 2 0.183	00:10	
Breath Volume	1.789	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 3 0.188	00:13	
Breath Volume	2.078	00:13	
Air Blank	0.000	00:14	
Control Test	0.078	00:14	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVAS,TABITHA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ALCARAZ PULIDO

DATE/TIME OF TEST: 12/20/2023 01:35:11

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.078	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.148	01:30
Breath Volume	1.414	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.140	01:33
Breath Volume	1.652	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: REYES HERNANDEZ

DATE/TIME OF TEST: 12/22/2023 01:45:47

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.181	01:41
Breath Volume	3.574	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.180	01:44
Breath Volume	4.660	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BOTELHO, ISAC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 12/22/2023 05:02:13

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.078	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.181	04:58
Breath Volume	2.250	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.186	05:00
Breath Volume	2.585	05:00
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATOSVALENTIN,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: MINUCHEFALCONES

DATE/TIME OF TEST: 12/22/2023 15:59:18

Diagnostics Check	OK	15:53
Air Blank	0.000	15:53
Control Test	0.078	15:54
Air Blank	0.000	15:54
Subject Sample # 1	0.187	15:55
Breath Volume	2.214	15:55
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.194	15:57
Breath Volume	2.257	15:57
Air Blank	0.000	15:58
Control Test	0.077	15:58
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH,ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SANTELL GONZALEZ

DATE/TIME OF TEST: 12/23/2023 02:25:18

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.167	02:20
Breath Volume	1.175	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.164	02:23
Breath Volume	2.007	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMALL,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: JOR ALVA

DATE/TIME OF TEST: 12/23/2023 03:34:10

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.235	03:29
Breath Volume	2.980	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.235	03:32
Breath Volume	2.488	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLABAUGH, ARIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: LEIVA MADRID

DATE/TIME OF TEST: 12/23/2023 05:08:03

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.137	05:03
Breath Volume	1.703	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.139	05:06
Breath Volume	2.617	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:08

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ PIETRI,VICTOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MENDOZA II

DATE/TIME OF TEST: 12/24/2023 05:31:32

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.078	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.108	05:26
Breath Volume	1.375	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.118	05:30
Breath Volume	2.332	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 01/15/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WACHTEL,BRUCE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: CUMARESENCIAL

DATE/TIME OF TEST: 12/25/2023 07:56:31

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.078	07:46
Air Blank	0.000	07:46
Subject Sample # 1	0.212	07:49
Breath Volume	1.183	07:49
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.212	07:54
Breath Volume	1.394	07:54
Air Blank	0.000	07:55
Control Test	0.078	07:55
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DUFT,CHALON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BUSHEY

DATE/TIME OF TEST: 12/27/2023 13:02:03

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.079	12:56
Air Blank	0.000	12:57
Subject Sample # 1	0.000	12:57
Breath Volume	3.625	12:57
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.000	13:00
Breath Volume	1.691	13:00
Air Blank	0.000	13:01
Control Test	0.079	13:01
Air Blank	0.000	13:01
Diagnostics Check	OK	13:02

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GONCALVES NETO

DATE/TIME OF TEST: 12/28/2023 22:41:53

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.171	22:37
Breath Volume	3.261	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.170	22:40
Breath Volume	2.593	22:40
Air Blank	0.000	22:41
Control Test	0.078	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIESENFELT,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SEADER

DATE/TIME OF TEST: 12/30/2023 03:49:54

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.136	03:45
Breath Volume	2.195	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.131	03:48
Breath Volume	1.171	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/30/2023 04:17:10

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.179	04:12
Breath Volume	2.500	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.175	04:15
Breath Volume	2.808	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEONARD,WAKENS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 12/30/2023 11:58:14

Diagnostics Check	OK	11:52
Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53
Subject Sample # 1	0.270	11:53
Breath Volume	1.847	11:53
Air Blank	0.000	11:54
Air Blank	0.000	11:56
Subject Sample # 2	0.276	11:56
Breath Volume	2.023	11:56
Air Blank	0.000	11:57
Control Test	0.078	11:57
Air Blank	0.000	11:58
Diagnostics Check	OK	11:58

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/31/2023 08:48:16

Diagnostics Check	OK	08:34	
Air Blank	0.000	08:34	
Control Test	0.079	08:35	
Air Blank	0.000	08:35	
Subject Sample	# 1 0.171	08:37	
Breath Volume	1.343	08:37	
Air Blank	0.000	08:38	
Air Blank	0.000	08:40	
Subject Sample	# 2 0.152	08:43	Volume Not Met
Breath Volume	0.558	08:43	
Air Blank	0.000	08:44	
Air Blank	0.000	08:45	
Subject Sample	# 3 0.164	08:46	
Breath Volume	1.312	08:46	
Air Blank	0.000	08:47	
Control Test	0.078	08:47	
Air Blank	0.000	08:48	
Diagnostics Check	OK	08:48	

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CERVELLERA,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A3
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 12/31/2023 10:28:37

Diagnostics Check	OK	10:22
Air Blank	0.000	10:22
Control Test	0.079	10:22
Air Blank	0.000	10:23
Subject Sample # 1	0.132	10:23
Breath Volume	2.867	10:23
Air Blank	0.000	10:24
Air Blank	0.000	10:26
Subject Sample # 2	0.134	10:27
Breath Volume	2.578	10:27
Air Blank	0.000	10:27
Control Test	0.078	10:28
Air Blank	0.000	10:28
Diagnostics Check	OK	10:28

Data Download Date/Time: 01/15/2024 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/01/2023 00:16:04

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.083	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:11
Breath Volume	1.289	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:14
Breath Volume	1.937	00:14
Air Blank	0.000	00:15
Control Test	0.082	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/21/2023 11:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GONZALEZ RENDON

DATE/TIME OF TEST: 12/05/2023 00:43:35

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.083	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.172	00:38
Breath Volume	1.789	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.173	00:42
Breath Volume	1.984	00:42
Air Blank	0.000	00:42
Control Test	0.083	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 12/21/2023 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/05/2023 02:12:57

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.083	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.012	02:08
Breath Volume	2.660	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.012	02:11
Breath Volume	2.261	02:11
Air Blank	0.000	02:12
Control Test	0.083	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SESSIONS

DATE/TIME OF TEST: 12/05/2023 23:04:37

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.083	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.093	23:00
Breath Volume	1.945	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.092	23:03
Breath Volume	1.718	23:03
Air Blank	0.000	23:03
Control Test	0.083	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 12/07/2023 03:08:18

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.083	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.098	03:03
Breath Volume	1.554	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.091	03:06
Breath Volume	1.105	03:06
Air Blank	0.000	03:07
Control Test	0.084	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ESCOBAR GUARDADO

DATE/TIME OF TEST: 12/08/2023 22:04:39

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.083	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.050	21:59
Breath Volume	1.585	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.048	22:03
Breath Volume	1.207	22:03
Air Blank	0.000	22:03
Control Test	0.083	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GUADALUPE,GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CASTRO-LAROTTA

DATE/TIME OF TEST: 12/09/2023 05:02:21

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.083	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1 0.115	04:55	Volume Not Met
Breath Volume	0.234	04:55	
Air Blank	0.000	04:55	
Air Blank	0.000	04:57	
Subject Sample	# 2 0.126	05:00	Volume Not Met
Breath Volume	0.097	05:00	
Air Blank	0.000	05:01	
Control Test	0.082	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYBIRAL,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CONTRAMAESTRE,ANDRES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/10/2023 14:05:58

Diagnostics Check	OK	13:59
Air Blank	0.000	13:59
Control Test	0.083	13:59
Air Blank	0.000	14:00
Subject Sample # 1	0.199	14:01
Breath Volume	1.515	14:01
Air Blank	0.000	14:02
Air Blank	0.000	14:03
Subject Sample # 2	0.219	14:04
Breath Volume	3.093	14:04
Air Blank	0.000	14:05
Control Test	0.081	14:05
Air Blank	0.000	14:05
Diagnostics Check	OK	14:05

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TOALA-CONSTANTE

DATE/TIME OF TEST: 12/10/2023 22:06:56

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.083	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.167	22:02
Breath Volume	1.917	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.167	22:05
Breath Volume	1.511	22:05
Air Blank	0.000	22:06
Control Test	0.083	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,ELIEZER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SHEHADEH ALICEA

DATE/TIME OF TEST: 12/11/2023 01:43:11

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.083	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.044	01:38
Breath Volume	3.253	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.043	01:41
Breath Volume	3.562	01:41
Air Blank	0.000	01:42
Control Test	0.083	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/14/2023 02:29:49

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.083	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.179	02:25
Breath Volume	3.535	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.164	02:28
Breath Volume	2.507	02:28
Air Blank	0.000	02:28
Control Test	0.082	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: KISER

DATE/TIME OF TEST: 12/14/2023 20:04:12

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.083	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.250	19:59
Breath Volume	4.843	19:59
Air Blank	0.000	20:00
Air Blank	0.000	20:01
Subject Sample # 2	0.245	20:02
Breath Volume	2.917	20:02
Air Blank	0.000	20:03
Control Test	0.082	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GUADALUPE,GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CASTROS PERERA

DATE/TIME OF TEST: 12/15/2023 03:24:31

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.084	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.171	03:19
Breath Volume	2.378	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:21
Subject Sample # 2	0.173	03:22
Breath Volume	1.656	03:22
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,JASON,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VICKERS,SETH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/16/2023 07:41:28

Diagnostics Check	OK	07:40	
Air Blank		07:41	RFI Detect
Air Blank	0.000	07:41	

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,JASON,H
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VICKERS,SETH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/16/2023 08:10:34

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.083	08:04
Air Blank	0.000	08:05
Subject Sample # 1	0.000	08:06
Breath Volume	1.625	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.000	08:09
Breath Volume	1.804	08:09
Air Blank	0.000	08:09
Control Test	0.081	08:09
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCANLAN,JAMES,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESPIET,ROSALYN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: BOOKS

DATE/TIME OF TEST: 12/18/2023 18:20:25

Diagnostics Check	OK	18:12
Air Blank	0.000	18:12
Control Test	0.083	18:12
Air Blank	0.000	18:13
Subject Sample # 1	0.251	18:15
Breath Volume	1.953	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.263	18:18
Breath Volume	2.109	18:18
Air Blank	0.000	18:19
Control Test	0.079	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,JAVIER,O
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 12/19/2023 03:06:14

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.082	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.000	03:00
Breath Volume	2.390	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.000	03:04
Breath Volume	2.636	03:04
Air Blank	0.000	03:05
Control Test	0.084	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENWALL,ROBERT,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ROSARIO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: CORTEZMILAN

DATE/TIME OF TEST: 12/20/2023 20:07:49

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.083	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.279	20:03
Breath Volume	1.218	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.267	20:06
Breath Volume	3.566	20:06
Air Blank	0.000	20:06
Control Test	0.082	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 12/21/2023 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RIVERA
NAVARRO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CZECH

DATE/TIME OF TEST: 12/23/2023 05:07:04

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.083	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.171	05:02
Breath Volume	3.046	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.156	05:05
Breath Volume	2.167	05:05
Air Blank	0.000	05:06
Control Test	0.082	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:07

Data Download Date/Time: 01/03/2024 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,JAVIER,O
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 12/24/2023 20:43:50

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:41	
Control Test	0.083	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:42	RFI Detect
Breath Volume	0.000	20:42	
Air Blank	0.000	20:42	
Control Test	0.084	20:43	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 01/03/2024 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDERO,JAVIER,O
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,RAUL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SANTIAGO JR.

DATE/TIME OF TEST: 12/24/2023 20:52:48

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.083	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.018	20:48
Breath Volume	1.875	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.018	20:51
Breath Volume	1.511	20:51
Air Blank	0.000	20:51
Control Test	0.083	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 01/03/2024 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GUADALUPE, GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/25/2023 05:37:33

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.084	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.192	05:33
Breath Volume	1.371	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.202	05:35
Breath Volume	1.359	05:35
Air Blank	0.000	05:36
Control Test	0.082	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 01/03/2024 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAREY,VENESSA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BUITRAGO

DATE/TIME OF TEST: 12/29/2023 07:16:59

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.083	07:11
Air Blank	0.000	07:12
Subject Sample # 1	0.116	07:12
Breath Volume	2.886	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:14
Subject Sample # 2	0.118	07:15
Breath Volume	3.050	07:15
Air Blank	0.000	07:16
Control Test	0.081	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 01/03/2024 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 12/31/2023 23:33:18

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.083	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.000	23:28
Breath Volume	1.683	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:31
Breath Volume	2.070	23:31
Air Blank	0.000	23:32
Control Test	0.083	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/03/2024 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GRISCOM

DATE/TIME OF TEST: 12/01/2023 19:38:06

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:37	
Control Test	0.000	19:37	RFI Detect
Air Blank		19:38	RFI Detect

Data Download Date/Time: 12/25/2023 13:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: GRISCOM

DATE/TIME OF TEST: 12/01/2023 19:46:49

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.078	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.073	19:42
Breath Volume	1.234	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.074	19:45
Breath Volume	1.183	19:45
Air Blank	0.000	19:45
Control Test	0.078	19:46
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CURBELO,A.
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: COUGHLIN

DATE/TIME OF TEST: 12/02/2023 03:48:27

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.109	03:43
Breath Volume	1.957	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.109	03:46
Breath Volume	2.039	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JASPER,ANDREW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:16

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/02/2023 09:42:28

Diagnostics Check	OK	09:36
Air Blank	0.000	09:36
Control Test	0.076	09:37
Air Blank	0.000	09:37
Subject Sample # 1	0.245	09:38
Breath Volume	1.734	09:38
Air Blank	0.000	09:38
Air Blank	0.000	09:40
Subject Sample # 2	0.237	09:40
Breath Volume	1.339	09:40
Air Blank	0.000	09:41
Control Test	0.077	09:41
Air Blank	0.000	09:42
Diagnostics Check	OK	09:42

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BENNU

DATE/TIME OF TEST: 12/06/2023 20:09:28

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.079	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.161	20:04
Breath Volume	1.441	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:06
Subject Sample # 2	0.161	20:07
Breath Volume	2.027	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 12/16/2023 01:52:09

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.075	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1 0.161	01:44	Volume Not Met
Breath Volume	0.562	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample	# 2 0.169	01:50	Volume Not Met
Breath Volume	0.691	01:50	
Air Blank	0.000	01:51	
Control Test	0.078	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: STAHL, A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BURPEE

DATE/TIME OF TEST: 12/19/2023 22:08:06

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.119	22:03
Breath Volume	2.601	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.115	22:06
Breath Volume	2.652	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COLLINS,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ZINGE

DATE/TIME OF TEST: 12/21/2023 23:06:35

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.000	23:01
Breath Volume	1.667	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.000	23:05
Breath Volume	1.527	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN, MARK, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HADLEY, R
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: 3 MCPHAIL

DATE/TIME OF TEST: 12/24/2023 03:28:26

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.195	03:23
Breath Volume	1.390	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.184	03:26
Breath Volume	1.742	03:26
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/25/2023 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 12/01/2023 19:42:35

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:34	
Control Test	0.080	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1 0.132	19:38	Volume Not Met
Breath Volume	0.015	19:38	
Air Blank	0.000	19:39	
Air Blank	0.000	19:40	
Subject Sample	# 2	19:41	Subject Test Refused
Breath Volume	0.000	19:41	
Air Blank	0.000	19:41	
Control Test	0.078	19:41	
Air Blank	0.000	19:42	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 12/21/2023 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NEWELL,JEFF
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: DELACEY

DATE/TIME OF TEST: 12/02/2023 03:55:52

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.122	03:49
Breath Volume	1.527	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.121	03:54
Breath Volume	1.929	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/21/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CINTRON CONCEPCION

DATE/TIME OF TEST: 12/02/2023 04:12:03

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1 0.100	04:04	Volume Not Met
Breath Volume	0.746	04:04	
Air Blank	0.000	04:04	
Air Blank	0.000	04:06	
Subject Sample	# 2 0.116	04:07	
Breath Volume	1.722	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 3 0.117	04:10	
Breath Volume	1.609	04:10	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 12/21/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: SRODTMAN,BARBARA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: PEDRO ANTONIO

DATE/TIME OF TEST: 12/02/2023 20:39:40

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.080	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1 0.168	20:32	Volume Not Met
Breath Volume	0.667	20:32	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 2 0.158	20:38	Volume Not Met
Breath Volume	0.273	20:38	
Air Blank	0.000	20:38	
Control Test	0.079	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 12/21/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUILES-DAVILA,DARIEL,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 11/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/03/2023 01:36:24

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.185	01:31
Breath Volume	1.859	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.188	01:34
Breath Volume	1.726	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 12/21/2023 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NEWELL,JEFF
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/25/2023 21:57:09

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.185	21:52
Breath Volume	3.523	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.179	21:55
Breath Volume	3.394	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ELDER,MICHEAL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: TAFOLLA RODRIGUEZ

DATE/TIME OF TEST: 12/26/2023 00:20:04

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.198	00:15
Breath Volume	2.011	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.196	00:18
Breath Volume	1.414	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:20

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ZULUAGA,SANTIAGO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 12/28/2023 07:00:10

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:53
Subject Sample # 1	0.174	06:55
Breath Volume	3.253	06:55
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.175	06:58
Breath Volume	2.929	06:58
Air Blank	0.000	06:59
Control Test	0.079	06:59
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CADY,PETER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GARZON TUERO

DATE/TIME OF TEST: 12/29/2023 22:30:28

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.080	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.150	22:25
Breath Volume	2.488	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.149	22:28
Breath Volume	2.027	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TAYLOR,JAKE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: BRICENO CHAUSTRE

DATE/TIME OF TEST: 12/30/2023 03:56:49

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.087	03:52
Breath Volume	1.578	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.086	03:55
Breath Volume	1.445	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 12/31/2023 01:34:34

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.081	01:24	
Air Blank	0.000	01:24	
Subject Sample	# 1 0.195	01:27	Volume Not Met
Breath Volume	0.015	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:32	Subject Test Refused
Breath Volume	0.156	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 01/03/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MARTE BLANDINO

DATE/TIME OF TEST: 12/05/2023 03:51:40

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.000	03:47
Breath Volume	3.527	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.000	03:50
Breath Volume	2.679	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WILLETT

DATE/TIME OF TEST: 12/08/2023 21:53:06

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.079	21:42	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:46	No Sample Provided
Breath Volume	0.000	21:46	
Air Blank	0.000	21:46	
Air Blank	0.000	21:48	
Subject Sample	# 2	21:51	No Sample Provided
Breath Volume	0.000	21:51	
Air Blank	0.000	21:52	
Control Test	0.079	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STANLEY, WILLIAM
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/09/2023 07:25:36

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.079	07:17
Air Blank	0.000	07:17
Subject Sample # 1	0.067	07:20
Breath Volume	2.687	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample # 2	0.064	07:24
Breath Volume	1.789	07:24
Air Blank	0.000	07:24
Control Test	0.079	07:25
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PAYNE, TAYLOR
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: ZUK

DATE/TIME OF TEST: 12/09/2023 15:14:55

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.078	15:09
Air Blank	0.000	15:10
Subject Sample # 1	0.167	15:10
Breath Volume	1.753	15:10
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.157	15:13
Breath Volume	1.457	15:13
Air Blank	0.000	15:14
Control Test	0.079	15:14
Air Blank	0.000	15:14
Diagnostics Check	OK	15:14

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TATE,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SITTHISAY

DATE/TIME OF TEST: 12/14/2023 00:31:36

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.078	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.145	00:26
Breath Volume	1.875	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.147	00:30
Breath Volume	1.914	00:30
Air Blank	0.000	00:30
Control Test	0.077	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WETH

DATE/TIME OF TEST: 12/14/2023 02:32:02

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.050	02:27
Breath Volume	3.507	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.048	02:30
Breath Volume	3.398	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:32

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSEN, IAN, N
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HITCHCOCK, LUCAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/16/2023 02:01:21

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.184	01:56
Breath Volume	1.605	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.186	01:59
Breath Volume	1.964	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CIUPAK

DATE/TIME OF TEST: 12/22/2023 22:16:27

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.078	22:11
Air Blank	0.000	22:11
Subject Sample # 1	0.259	22:11
Breath Volume	1.667	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.261	22:14
Breath Volume	2.425	22:14
Air Blank	0.000	22:15
Control Test	0.077	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/23/2023 02:56:16

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.077	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.172	02:50
Breath Volume	1.261	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.170	02:54
Breath Volume	1.988	02:54
Air Blank	0.000	02:55
Control Test	0.077	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STANLEY, WILLIAM
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 12/23/2023 18:27:17

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.077	18:21
Air Blank	0.000	18:21
Subject Sample # 1	0.000	18:22
Breath Volume	1.417	18:22
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.000	18:25
Breath Volume	1.902	18:25
Air Blank	0.000	18:26
Control Test	0.077	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 042
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/24/2023 02:56:59

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.169	02:52
Breath Volume	2.511	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.166	02:55
Breath Volume	2.660	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,JAVIER,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: BOGNER

DATE/TIME OF TEST: 12/24/2023 16:47:03

Diagnostics Check	OK	16:35	
Air Blank	0.000	16:36	
Control Test	0.077	16:36	
Air Blank	0.000	16:37	
Subject Sample	# 1 0.303	16:40	
Breath Volume	1.769	16:40	
Air Blank	0.000	16:40	
Air Blank	0.000	16:42	
Subject Sample	# 2 0.313	16:45	
Breath Volume	2.007	16:45	
Air Blank	0.000	16:46	
Control Test	0.074	16:46	Control Outside Tolerance
Air Blank	0.000	16:47	

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,JAVIER,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:13

SUBJECT'S LAST NAME: BOGNER

DATE/TIME OF TEST: 12/24/2023 17:00:20

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.076	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.293	16:55
Breath Volume	1.503	16:55
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample # 2	0.298	16:58
Breath Volume	1.636	16:58
Air Blank	0.000	16:59
Control Test	0.077	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 01/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSEN, IAN, N
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ZARIV, ALI
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2
DRY GAS EXPIRY DATE: 12/12/2033
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BEEBE

DATE/TIME OF TEST: 12/29/2023 19:50:15

Diagnostics Check	OK	19:38	
Air Blank	0.000	19:38	
Control Test	0.077	19:39	
Air Blank	0.000	19:39	
Subject Sample	# 1 0.151	19:42	Volume Not Met
Breath Volume	0.144	19:42	
Air Blank	0.000	19:43	
Air Blank	0.000	19:45	
Subject Sample	# 2 0.149	19:45	
Breath Volume	1.332	19:45	
Air Blank	0.000	19:46	
Air Blank	0.000	19:48	
Subject Sample	# 3 0.152	19:48	
Breath Volume	2.679	19:48	
Air Blank	0.000	19:49	
Control Test	0.075	19:49	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:50	

Data Download Date/Time: 01/07/2024 19:47