

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HAWRYLUK

DATE/TIME OF TEST: 12/02/2023 03:52:58

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.135	03:48
Breath Volume	1.597	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.138	03:51
Breath Volume	2.628	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KILPATRICK,ALEX
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: SCHUMMER

DATE/TIME OF TEST: 12/02/2023 07:40:33

Diagnostics Check	OK	07:37	
Air Blank	0.000	07:38	
Control Test	0.078	07:38	
Air Blank	0.000	07:39	
Subject Sample	# 1	07:39	Subject Test Refused
Breath Volume	0.000	07:39	
Air Blank	0.000	07:39	
Control Test	0.078	07:39	
Air Blank	0.000	07:40	
Diagnostics Check	OK	07:40	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,MAXWELL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: EMLEY

DATE/TIME OF TEST: 12/08/2023 16:17:36

Diagnostics Check	OK	16:16	
Air Blank		16:17	RFI Detect
Air Blank	0.000	16:17	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,MAXWELL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: EMLEY

DATE/TIME OF TEST: 12/08/2023 16:22:23

Diagnostics Check	OK	16:19	
Air Blank	0.000	16:20	
Control Test	0.079	16:20	
Air Blank	0.000	16:20	
Subject Sample	# 1	16:21	Subject Test Refused
Breath Volume	0.000	16:21	
Air Blank	0.000	16:21	
Control Test	0.079	16:21	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,SHANICE,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/09/2023 01:42:40

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.078	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,J,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GORDON,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: TODD

DATE/TIME OF TEST: 12/09/2023 03:47:02

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.120	03:42
Breath Volume	3.144	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.123	03:45
Breath Volume	2.308	03:45
Air Blank	0.000	03:46
Control Test	0.077	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/10/2023 05:15:57

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:08	
Control Test	0.077	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.185	05:11	Volume Not Met
Breath Volume	0.015	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:14	
Subject Sample	# 2	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:15	
Control Test	0.076	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,JAXON,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,JAXON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 12/11/2023 01:32:32

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:25	No Sample Provided
Breath Volume	0.000	01:25	
Air Blank	0.000	01:26	
Air Blank	0.000	01:28	
Subject Sample	# 2	01:31	No Sample Provided
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.078	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PERKINS,KAYLEIGH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PICCIARELLI

DATE/TIME OF TEST: 12/12/2023 02:48:17

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.079	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PRENTISS,HAYLEIGH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: VIERA

DATE/TIME OF TEST: 12/14/2023 23:33:59

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.078	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DORON

DATE/TIME OF TEST: 12/17/2023 03:56:30

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.204	03:51
Breath Volume	3.296	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.205	03:54
Breath Volume	2.996	03:54
Air Blank	0.000	03:55
Control Test	0.078	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 12/19/2023 00:41:01

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.105	00:36
Breath Volume	1.421	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.095	00:39
Breath Volume	1.429	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,JAXON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARLOTTA,VINCENZO
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GERRY

DATE/TIME OF TEST: 12/20/2023 00:52:47

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.143	00:48
Breath Volume	1.921	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.143	00:51
Breath Volume	1.292	00:51
Air Blank	0.000	00:51
Control Test	0.078	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BISHOP,EVAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HOYT

DATE/TIME OF TEST: 12/22/2023 01:17:37

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.211	01:12
Breath Volume	2.503	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.208	01:16
Breath Volume	2.390	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KOTOWSKI,PHILIP
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MANHART-FOSS

DATE/TIME OF TEST: 12/22/2023 04:11:10

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.078	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.130	04:06
Breath Volume	1.910	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.128	04:09
Breath Volume	1.757	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/22/2023 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MANKA,KYLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/26/2023 21:01:42

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.077	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1 0.261	20:56	Volume Not Met
Breath Volume	0.414	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	21:00	Subject Test Refused
Breath Volume	0.566	21:00	
Air Blank	0.000	21:00	
Control Test	0.076	21:01	
Air Blank		21:01	RFI Detect
Air Blank		21:01	RFI Detect

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 12/26/2023 21:25:25

Diagnostics Check	OK	21:18
Air Blank	0.000	21:19
Control Test	0.078	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.088	21:20
Breath Volume	3.585	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.085	21:23
Breath Volume	2.101	21:23
Air Blank	0.000	21:24
Control Test	0.078	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTLEY, BRANDEN, C
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RAY, KEATON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: LAUMAN

DATE/TIME OF TEST: 12/29/2023 16:04:29

Diagnostics Check	OK	15:58
Air Blank	0.000	15:58
Control Test	0.079	15:59
Air Blank	0.000	15:59
Subject Sample # 1	0.160	16:00
Breath Volume	2.089	16:00
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 2	0.164	16:02
Breath Volume	1.910	16:02
Air Blank	0.000	16:03
Control Test	0.079	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: HOLCOMB

DATE/TIME OF TEST: 12/29/2023 17:46:42

Diagnostics Check	OK	17:40
Air Blank	0.000	17:41
Control Test	0.079	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.272	17:42
Breath Volume	1.109	17:42
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.277	17:45
Breath Volume	2.296	17:45
Air Blank	0.000	17:45
Control Test	0.079	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 12/30/2023 01:54:33

Diagnostics Check	OK	01:53	
Air Blank		01:54	RFI Detect
Air Blank	0.000	01:54	

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06022080A3
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 12/30/2023 02:03:48

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1 0.150	02:00	
Breath Volume	1.378	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2 0.149	02:02	
Breath Volume	1.562	02:02	
Air Blank		02:03	RFI Detect
Air Blank	0.000	02:03	

Data Download Date/Time: 01/22/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMAS,JEFFREY
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LACAYO

DATE/TIME OF TEST: 12/01/2023 03:11:13

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.220	03:06
Breath Volume	1.242	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.215	03:09
Breath Volume	1.398	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/13/2023 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 12/01/2023 04:46:53

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:33	
Air Blank	0.000	04:33	
Subject Sample	# 1 0.155	04:36	
Breath Volume	1.222	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 2 0.147	04:42	Volume Not Met
Breath Volume	0.449	04:42	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample	# 3 0.163	04:45	
Breath Volume	1.855	04:45	
Air Blank	0.000	04:45	
Control Test	0.079	04:46	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 12/13/2023 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MENDIETA ACERENZA

DATE/TIME OF TEST: 12/02/2023 23:38:32

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.151	23:34
Breath Volume	2.742	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.152	23:37
Breath Volume	2.324	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 12/13/2023 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/04/2023 22:10:30

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.080	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:09	Slope Not Met
Breath Volume	1.027	22:09	
Air Blank	0.000	22:09	
Control Test	0.081	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 12/13/2023 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/04/2023 22:41:17

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.248	22:36
Breath Volume	2.785	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.253	22:39
Breath Volume	3.558	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 12/13/2023 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBNETT,STEPHANIE,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/10/2023 00:46:59

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:40	No Sample Provided
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBNETT,STEPHANIE,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SCOTT,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 12/10/2023 01:19:13

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.263	01:14
Breath Volume	4.191	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.246	01:17
Breath Volume	3.511	01:17
Air Blank	0.000	01:18
Control Test	0.078	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BARTLING

DATE/TIME OF TEST: 12/10/2023 01:36:37

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.078	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: REBSTOCK

DATE/TIME OF TEST: 12/17/2023 02:51:26

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.128	02:46
Breath Volume	1.226	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.117	02:49
Breath Volume	1.207	02:49
Air Blank	0.000	02:50
Control Test	0.077	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,CARLOS,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HARRIGAN,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KOUVARAS

DATE/TIME OF TEST: 12/18/2023 02:23:02

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.110	02:18
Breath Volume	1.363	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.109	02:21
Breath Volume	1.402	02:21
Air Blank	0.000	02:22
Control Test	0.077	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,HANS,H
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARLSON,HANS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/18/2023 09:45:55

Diagnostics Check	OK	09:39
Air Blank	0.000	09:39
Control Test	0.080	09:40
Air Blank	0.000	09:40
Subject Sample # 1	0.149	09:41
Breath Volume	2.582	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:43
Subject Sample # 2	0.143	09:44
Breath Volume	2.527	09:44
Air Blank	0.000	09:44
Control Test	0.077	09:45
Air Blank	0.000	09:45
Diagnostics Check	OK	09:45

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: TINNEN

DATE/TIME OF TEST: 12/21/2023 03:56:20

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.078	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Control Test	0.078	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 01/03/2024 08:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BILLS

DATE/TIME OF TEST: 12/03/2023 00:35:00

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.082	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.221	00:30
Breath Volume	3.570	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.213	00:33
Breath Volume	3.445	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:35

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 12/05/2023 02:28:24

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.083	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.049	02:24
Breath Volume	1.574	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.049	02:26
Breath Volume	1.230	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MIKAELIAN, MIKAEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 12/05/2023 20:11:03

Diagnostics Check	OK	20:04
Air Blank	0.000	20:05
Control Test	0.083	20:05
Air Blank	0.000	20:06
Subject Sample # 1	0.158	20:06
Breath Volume	1.457	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 2	0.155	20:09
Breath Volume	1.660	20:09
Air Blank	0.000	20:10
Control Test	0.082	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:11

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN, D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 12/08/2023 04:31:11

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.220	04:26
Breath Volume	1.433	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.219	04:29
Breath Volume	2.023	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,SHAE,A
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: CIERZNIIEWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BAUTISTA

DATE/TIME OF TEST: 12/09/2023 01:50:22

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.082	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.183	01:45
Breath Volume	2.964	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.176	01:48
Breath Volume	3.171	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, SHAE, A
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: KHALIL, FAYTHE
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PERNAS

DATE/TIME OF TEST: 12/09/2023 02:51:42

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.221	02:47
Breath Volume	1.972	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.215	02:50
Breath Volume	2.734	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILLS,JOHN,T
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARTIN,JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/12/2023 05:17:13

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.082	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.000	05:12
Breath Volume	3.703	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.000	05:15
Breath Volume	3.468	05:15
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 12/13/2023 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DIAZ, DIANA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 12/16/2023 20:33:28

Diagnostics Check	OK	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.057	20:28
Breath Volume	1.531	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.055	20:31
Breath Volume	1.113	20:31
Air Blank	0.000	20:32
Control Test	0.082	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 01/18/2024 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,ROBERT,A
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/21/2023 01:58:20

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.083	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.163	01:53
Breath Volume	2.332	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.167	01:56
Breath Volume	2.445	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/18/2024 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH,TYLER,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MIKAELIAN,MIKAEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/26/2023 18:39:41

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:38	
Control Test	0.082	18:39	RFI Detect
Air Blank	0.000	18:39	

Data Download Date/Time: 01/18/2024 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 12/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MOREAU

DATE/TIME OF TEST: 12/27/2023 03:25:26

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.083	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.029	03:20
Breath Volume	1.468	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.029	03:23
Breath Volume	1.867	03:23
Air Blank	0.000	03:24
Control Test	0.083	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 01/18/2024 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,KEVIN,R
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: MASHBURN,TONY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 12/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/31/2023 07:19:31

Diagnostics Check	OK	07:13
Air Blank	0.000	07:13
Control Test	0.082	07:14
Air Blank	0.000	07:14
Subject Sample # 1	0.089	07:15
Breath Volume	1.171	07:15
Air Blank	0.000	07:15
Air Blank	0.000	07:17
Subject Sample # 2	0.076	07:18
Breath Volume	1.148	07:18
Air Blank	0.000	07:18
Control Test	0.082	07:18
Air Blank	0.000	07:19
Diagnostics Check	OK	07:19

Data Download Date/Time: 01/18/2024 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-WILSON,TREVON,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: PONAS,CLIFFORD
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/20/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: GARAS

DATE/TIME OF TEST: 12/14/2023 21:01:15

Diagnostics Check	OK	21:00	
Air Blank		21:00	RFI Detect
Air Blank	0.000	21:01	

Data Download Date/Time: 12/21/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-WILSON,TREVON,L
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: PONAS,CLIFFORD
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: GARAS

DATE/TIME OF TEST: 12/14/2023 21:20:57

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
Control Test	0.083	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:11	No Sample Provided
Breath Volume	0.000	21:11	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample	# 2	0.282	21:16
Breath Volume	1.757	21:16	
Air Blank	0.000	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 3	0.293	21:19
Breath Volume	2.812	21:19	
Air Blank	0.000	21:20	
Control Test	0.081	21:20	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 12/21/2023 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MACAVOY,JASON

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 12/03/2023 00:54:08

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.229	00:49
Breath Volume	2.617	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.212	00:52
Breath Volume	2.023	00:52
Air Blank	0.000	00:53
Control Test	0.078	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 12/20/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUFTS,KEVIN,S
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: CHAN,KAM
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: SORTO

DATE/TIME OF TEST: 12/03/2023 10:41:40

Diagnostics Check	OK	10:34
Air Blank	0.000	10:34
Control Test	0.081	10:35
Air Blank	0.000	10:35
Subject Sample # 1	0.230	10:36
Breath Volume	2.097	10:36
Air Blank	0.000	10:37
Air Blank	0.000	10:39
Subject Sample # 2	0.221	10:39
Breath Volume	2.019	10:39
Air Blank	0.000	10:40
Control Test	0.079	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 12/20/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 23:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: FELSKE

DATE/TIME OF TEST: 12/08/2023 00:28:41

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.081	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.082	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 12/20/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,C,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KRUSE,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 23:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: DUSS

DATE/TIME OF TEST: 12/14/2023 00:39:13

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.000	00:38	Control Outside Tolerance
Air Blank	0.000	00:39	

Data Download Date/Time: 12/20/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,C,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: KRUSE,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: DUSS

DATE/TIME OF TEST: 12/14/2023 00:49:57

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.106	00:45
Breath Volume	1.230	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.110	00:48
Breath Volume	1.703	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 12/20/2023 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BEENER,KYLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 23:05

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 12/15/2023 23:44:10

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.082	23:42	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.082	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 12/20/2023 13:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: STERNBERG

DATE/TIME OF TEST: 12/02/2023 19:20:14

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.080	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.227	19:15
Breath Volume	1.500	19:15
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.222	19:18
Breath Volume	1.277	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 12/04/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HERNANDEZ-SARMIENTO

DATE/TIME OF TEST: 12/03/2023 03:55:05

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:43	
Control Test	0.083	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1 0.139	03:47	Volume Not Met
Breath Volume	0.453	03:47	
Air Blank	0.000	03:47	
Air Blank	0.000	03:49	
Subject Sample	# 2 0.145	03:50	
Breath Volume	2.179	03:50	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 3 0.150	03:53	
Breath Volume	2.468	03:53	
Air Blank	0.000	03:54	
Control Test	0.079	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 12/04/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIMIRO,B,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ADDINGTON,C
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28021080A2
DRY GAS EXPIRY DATE: 12/05/2023
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: PLEASANTS

DATE/TIME OF TEST: 12/03/2023 23:10:38

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.257	23:05
Breath Volume	2.519	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.266	23:09
Breath Volume	2.015	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 12/04/2023 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: CARDENAS ANDINO

DATE/TIME OF TEST: 12/02/2023 01:21:31

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 12/18/2023 15:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GUIONNAUD

DATE/TIME OF TEST: 12/08/2023 01:55:11

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.082	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.192	01:46
Breath Volume	1.511	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.167	01:49
Breath Volume	2.335	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 3	0.202	01:53
Breath Volume	3.445	01:53
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 12/29/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FLEURMOND

DATE/TIME OF TEST: 12/10/2023 01:28:58

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.082	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1 0.165	01:18	
Breath Volume	1.210	01:18	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2 0.129	01:24	Volume Not Met
Breath Volume	0.527	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.151	01:27	
Breath Volume	1.500	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 12/29/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 12/23/2023 02:46:56

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.082	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.164	02:39
Breath Volume	1.140	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.158	02:45
Breath Volume	1.578	02:45
Air Blank	0.000	02:46
Control Test	0.082	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 12/29/2023 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/08/2023 08:18:39

Diagnostics Check	OK	08:11
Air Blank	0.000	08:12
Control Test	0.081	08:12
Air Blank	0.000	08:12
Subject Sample # 1	0.060	08:13
Breath Volume	2.289	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.060	08:17
Breath Volume	1.937	08:17
Air Blank	0.000	08:17
Control Test	0.081	08:18
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: BRODIE

DATE/TIME OF TEST: 12/09/2023 12:01:13

Diagnostics Check	OK	11:54
Air Blank	0.000	11:54
Control Test	0.081	11:54
Air Blank	0.000	11:55
Subject Sample # 1	0.000	11:56
Breath Volume	1.421	11:56
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.000	11:59
Breath Volume	2.621	11:59
Air Blank	0.000	12:00
Control Test	0.081	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COLWELL,NICOLAS
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CEDILLO-TORRES

DATE/TIME OF TEST: 12/10/2023 03:08:39

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.150	03:04
Breath Volume	1.277	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.150	03:07
Breath Volume	2.121	03:07
Air Blank	0.000	03:07
Control Test	0.081	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: SURA MARTINEZ

DATE/TIME OF TEST: 12/11/2023 13:01:57

Diagnostics Check	OK	12:55
Air Blank	0.000	12:56
Control Test	0.082	12:56
Air Blank	0.000	12:57
Subject Sample # 1	0.000	12:57
Breath Volume	3.167	12:57
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.000	13:00
Breath Volume	2.519	13:00
Air Blank	0.000	13:01
Control Test	0.083	13:01
Air Blank	0.000	13:01
Diagnostics Check	OK	13:01

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,ROBERT,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MCKEEBY,WADE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:03

SUBJECT'S LAST NAME: BLEDSOE

DATE/TIME OF TEST: 12/20/2023 11:29:38

Diagnostics Check	OK	11:23
Air Blank	0.000	11:23
Control Test	0.082	11:24
Air Blank	0.000	11:24
Subject Sample # 1	0.000	11:25
Breath Volume	1.472	11:25
Air Blank	0.000	11:25
Air Blank	0.000	11:27
Subject Sample # 2	0.000	11:28
Breath Volume	2.378	11:28
Air Blank	0.000	11:28
Control Test	0.082	11:29
Air Blank	0.000	11:29
Diagnostics Check	OK	11:29

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/23/2023 20:47:15

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.082	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.081	20:42
Breath Volume	1.105	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.078	20:45
Breath Volume	1.480	20:45
Air Blank	0.000	20:46
Control Test	0.081	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: STEVEN BROWN

DATE/TIME OF TEST: 12/25/2023 02:56:18

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.210	02:51
Breath Volume	2.746	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.215	02:54
Breath Volume	1.910	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BENNETT,BRANDY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SALAS BARONI

DATE/TIME OF TEST: 12/28/2023 02:49:08

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.082	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Sequence Aborted
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MONROE,JOEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 12/28/2023 20:25:26

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.082	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.000	20:21
Breath Volume	1.343	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.000	20:24
Breath Volume	2.089	20:24
Air Blank	0.000	20:24
Control Test	0.082	20:24
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LAWLEY,MASON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MAJORS

DATE/TIME OF TEST: 12/30/2023 03:32:10

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.000	03:27
Breath Volume	2.214	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:30
Breath Volume	1.425	03:30
Air Blank	0.000	03:31
Control Test	0.082	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/05/2024 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 12/03/2023 11:06:27

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.080	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.000	11:02
Breath Volume	4.140	11:02
Air Blank	0.000	11:02
Air Blank	0.000	11:04
Subject Sample # 2	0.000	11:05
Breath Volume	2.414	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:05
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 12/12/2023 02:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 12/09/2023 02:42:16

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.204	02:37
Breath Volume	2.300	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.197	02:40
Breath Volume	1.710	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/12/2023 02:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: GROOMS,SAMUEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/13/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: CHINOS

DATE/TIME OF TEST: 12/09/2023 03:22:16

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.139	03:17
Breath Volume	3.921	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.133	03:20
Breath Volume	3.277	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/12/2023 02:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 12/12/2023 12:17:53

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.080	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.000	12:13
Breath Volume	1.308	12:13
Air Blank	0.000	12:13
Air Blank	0.000	12:15
Subject Sample # 2	0.000	12:16
Breath Volume	1.140	12:16
Air Blank	0.000	12:16
Control Test	0.080	12:17
Air Blank	0.000	12:17
Diagnostics Check	OK	12:17

Data Download Date/Time: 01/15/2024 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MANSON,TIMMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 12/13/2023 21:15:21

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.131	21:10
Breath Volume	1.593	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:13
Subject Sample # 2	0.125	21:13
Breath Volume	2.191	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 01/15/2024 00:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAUGA,JOHN,E
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: GROOMS,SAMUEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/16/2023 23:41:52

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.197	23:37
Breath Volume	1.914	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.196	23:40
Breath Volume	1.285	23:40
Air Blank	0.000	23:40
Control Test	0.078	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 01/15/2024 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY,TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: SPEIGHTS

DATE/TIME OF TEST: 12/17/2023 15:47:21

Diagnostics Check	OK	15:40
Air Blank	0.000	15:40
Control Test	0.078	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.315	15:42
Breath Volume	1.445	15:42
Air Blank	0.000	15:43
Air Blank	0.000	15:44
Subject Sample # 2	0.324	15:45
Breath Volume	1.359	15:45
Air Blank	0.000	15:46
Control Test	0.078	15:46
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 01/15/2024 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LEE,NICHOLAS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/23/2023 01:28:33

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.116	01:23
Breath Volume	2.593	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.117	01:27
Breath Volume	2.558	01:27
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 01/15/2024 00:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: LOUIS-PIERRE,RUBEN
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 12/02/2023 01:12:28

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.177	01:07
Breath Volume	1.464	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.173	01:10
Breath Volume	2.453	01:10
Air Blank	0.000	01:11
Control Test	0.078	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 12/19/2023 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS-PIERRE,RUBEN,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: IZMA,KENNY
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/02/2023 22:54:02

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.079	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.272	22:43	
Breath Volume	1.164	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2 0.220	22:49	Volume Not Met
Breath Volume	1.015	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 3	22:53	Sequence Aborted
Breath Volume	0.000	22:53	
Air Blank	0.000	22:54	

Data Download Date/Time: 12/19/2023 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: ESCOBAR CASTANEDA

DATE/TIME OF TEST: 12/04/2023 21:52:35

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.079	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:52	No Sample Provided
Breath Volume	0.000	21:52	
Air Blank		21:52	Sequence Aborted
Air Blank	0.000	21:52	

Data Download Date/Time: 12/19/2023 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BIGGS

DATE/TIME OF TEST: 12/01/2023 20:07:36

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.123	20:02
Breath Volume	2.070	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.124	20:06
Breath Volume	2.230	20:06
Air Blank	0.000	20:06
Control Test	0.080	20:07
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOULIHAN,RILEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MILLETT

DATE/TIME OF TEST: 12/02/2023 01:32:08

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.170	01:27
Breath Volume	2.515	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.181	01:30
Breath Volume	2.601	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: SKELLY

DATE/TIME OF TEST: 12/02/2023 02:16:51

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.000	02:12
Breath Volume	3.367	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.000	02:15
Breath Volume	3.257	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 12/02/2023 03:03:10

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.129	02:58
Breath Volume	2.070	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.129	03:01
Breath Volume	2.246	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DENNIS,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GIRON GONZALEZ

DATE/TIME OF TEST: 12/02/2023 06:04:49

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.080	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:03	Subject Test Refused
Breath Volume	0.000	06:03	
Air Blank	0.000	06:03	
Control Test	0.080	06:04	
Air Blank	0.000	06:04	
Diagnostics Check	OK	06:04	

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MUNDELLA,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CHINCHILLA

DATE/TIME OF TEST: 12/03/2023 01:33:50

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SAHAGIAN,JACOB
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: TOMA SAMBRANO

DATE/TIME OF TEST: 12/03/2023 04:10:04

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Subject Sample	# 1 0.243	04:09	
Breath Volume	1.683	04:09	
Air Blank		04:09	RFI Detect
Air Blank		04:09	RFI Detect
Air Blank		04:10	RFI Detect

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SAHAGIAN,JACOB
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: TOMAS SAMBRANO

DATE/TIME OF TEST: 12/03/2023 04:19:33

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.246	04:14
Breath Volume	1.546	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.229	04:18
Breath Volume	2.171	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GOMEZ NARANJO

DATE/TIME OF TEST: 12/04/2023 03:46:04

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.122	03:41
Breath Volume	1.261	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.133	03:44
Breath Volume	2.585	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:46

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MORALES,CHRIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 12/06/2023 01:13:10

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.209	01:08
Breath Volume	1.894	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.199	01:11
Breath Volume	1.328	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/06/2023 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 12/06/2023 20:17:51

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.081	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1 0.000	20:10	Volume Not Met
Breath Volume	0.113	20:10	
Air Blank	0.000	20:11	
Air Blank	0.000	20:13	
Subject Sample	# 2 0.000	20:16	Volume Not Met
Breath Volume	0.800	20:16	
Air Blank	0.000	20:16	
Control Test	0.081	20:17	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIDGES,JARRETT,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGES,JARRETT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: RABIDOUX

DATE/TIME OF TEST: 12/07/2023 19:31:14

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.080	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.149	19:26
Breath Volume	2.105	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.165	19:29
Breath Volume	2.593	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: MILLIKEN

DATE/TIME OF TEST: 12/08/2023 17:59:02

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.080	17:53
Air Blank	0.000	17:53
Subject Sample # 1	0.051	17:54
Breath Volume	1.253	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.054	17:57
Breath Volume	1.542	17:57
Air Blank	0.000	17:58
Control Test	0.080	17:58
Air Blank	0.000	17:58
Diagnostics Check	OK	17:59

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DARIUS,DONALSON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: BINUELO CAZUELA

DATE/TIME OF TEST: 12/08/2023 22:04:23

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.162	21:59
Breath Volume	2.949	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.164	22:02
Breath Volume	3.105	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 12/08/2023 22:56:14

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.271	22:51
Breath Volume	1.789	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.258	22:54
Breath Volume	2.265	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,ROHAN,G
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SELLERS,MATTHEW
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/09/2023 01:56:12

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.140	01:51
Breath Volume	1.464	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.138	01:54
Breath Volume	1.585	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,JOSHUA,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NORRIS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 12/09/2023 16:43:36

Diagnostics Check	OK	16:36
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:37
Subject Sample # 1	0.111	16:39
Breath Volume	2.246	16:39
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.107	16:42
Breath Volume	1.593	16:42
Air Blank	0.000	16:42
Control Test	0.080	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:43

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,DONALD,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HRONCICH,NIKO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 12/10/2023 01:36:58

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1 0.219	01:34	
Breath Volume	3.960	01:34	
Air Blank		01:35	RFI Detect
Air Blank	0.000	01:36	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,DONALD,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HRONCICH,NIKO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 12/10/2023 01:40:49

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank		01:40	RFI Detect
Air Blank	0.000	01:40	

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 12/12/2023 00:28:13

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.264	00:19
Breath Volume	2.640	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.239	00:23
Breath Volume	1.730	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 3	0.238	00:26
Breath Volume	1.437	00:26
Air Blank	0.000	00:27
Control Test	0.077	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: BRANSCOME

DATE/TIME OF TEST: 12/12/2023 20:35:25

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.079	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.178	20:29
Breath Volume	1.269	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:31
Subject Sample # 2	0.172	20:33
Breath Volume	1.816	20:33
Air Blank	0.000	20:34
Control Test	0.078	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VULICEVIC,R
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 12/12/2023 23:33:30

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.350	23:29
Breath Volume	1.917	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.335	23:31
Breath Volume	1.253	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: RANGEL MEDRANO

DATE/TIME OF TEST: 12/13/2023 19:44:23

Diagnostics Check	OK	19:38
Air Blank	0.000	19:38
Control Test	0.080	19:39
Air Blank	0.000	19:39
Subject Sample # 1	0.122	19:40
Breath Volume	1.308	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.132	19:42
Breath Volume	1.386	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: PANGBURN

DATE/TIME OF TEST: 12/14/2023 02:04:38

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.000	02:00
Breath Volume	1.914	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:03
Breath Volume	1.175	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINEBERGER,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 12/14/2023 04:12:16

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.229	04:07
Breath Volume	1.789	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.234	04:10
Breath Volume	1.484	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MOLISH

DATE/TIME OF TEST: 12/14/2023 20:22:19

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.079	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.000	20:17
Breath Volume	1.519	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.000	20:20
Breath Volume	1.378	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SAHAGIAN,JACOB
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: WHITMARSH

DATE/TIME OF TEST: 12/15/2023 19:37:30

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.080	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.202	19:29
Breath Volume	1.347	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.226	19:32
Breath Volume	1.761	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 3	0.219	19:35
Breath Volume	2.308	19:35
Air Blank	0.000	19:36
Control Test	0.078	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BEHLING

DATE/TIME OF TEST: 12/16/2023 01:09:58

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.080	01:02	
Air Blank	0.000	01:03	
Subject Sample # 1	0.000	01:05	
Breath Volume	1.718	01:05	
Air Blank	0.000	01:06	
Air Blank	0.000	01:08	
Subject Sample # 2		01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PESOLA,CODY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 12/16/2023 02:52:01

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.219	02:47
Breath Volume	2.207	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.229	02:50
Breath Volume	2.097	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: SANTIAGO SANTIAGO

DATE/TIME OF TEST: 12/16/2023 03:15:47

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1 0.140	03:05	
Breath Volume	1.640	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 2 0.135	03:11	Volume Not Met
Breath Volume	0.285	03:11	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 3 0.140	03:14	
Breath Volume	1.332	03:14	
Air Blank	0.000	03:14	
Control Test	0.077	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DELGATTI,B
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 12/16/2023 03:52:30

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.159	03:48
Breath Volume	2.800	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.158	03:50
Breath Volume	2.910	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/24/2024 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MALLONE,NICHOLAS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/16/2023 04:55:59

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.171	04:51
Breath Volume	1.531	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.169	04:54
Breath Volume	1.843	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 12/16/2023 21:57:02

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.150	21:52
Breath Volume	2.082	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 2	0.154	21:55
Breath Volume	2.078	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:57

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHANTHAVONG,PHOURATSA
MAY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male

OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PEREZ PELENZUELA

DATE/TIME OF TEST: 12/17/2023 01:44:12

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.187	01:39
Breath Volume	2.890	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.178	01:42
Breath Volume	1.445	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DE ANDA

DATE/TIME OF TEST: 12/17/2023 02:21:06

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.251	02:16
Breath Volume	1.921	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.247	02:19
Breath Volume	2.234	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,J,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LEDOUX

DATE/TIME OF TEST: 12/18/2023 00:34:04

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1	00:32	Subject Test Refused
Breath Volume	0.175	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/19/2023 00:20:18

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.056	00:15
Breath Volume	1.101	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.060	00:18
Breath Volume	1.132	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA, STEVENS, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GOMEZ, GABRIEL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: OXLAJ GONZALEZ

DATE/TIME OF TEST: 12/19/2023 05:06:07

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.117	05:01
Breath Volume	1.613	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.117	05:04
Breath Volume	1.339	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: OSEGUERA RIOS

DATE/TIME OF TEST: 12/20/2023 03:14:21

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.110	03:09
Breath Volume	1.417	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.113	03:12
Breath Volume	1.378	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAHLKE,M,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: PAHLKE,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BADILLO

DATE/TIME OF TEST: 12/21/2023 00:19:57

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.161	00:15
Breath Volume	2.648	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.159	00:18
Breath Volume	2.964	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/21/2023 21:21:40

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.080	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1 0.000	21:14	Volume Not Met
Breath Volume	0.957	21:14	
Air Blank	0.000	21:15	
Air Blank	0.000	21:17	
Subject Sample	# 2 0.000	21:20	Volume Not Met
Breath Volume	0.605	21:20	
Air Blank	0.000	21:20	
Control Test	0.079	21:21	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BUNTING

DATE/TIME OF TEST: 12/22/2023 02:39:58

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:39	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CIEKLINSKI,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CHAVOEN

DATE/TIME OF TEST: 12/22/2023 05:11:43

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.029	05:07
Breath Volume	2.582	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.029	05:10
Breath Volume	2.371	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA, STEVENS, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LINEBERGER, RYAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MENDOZA REBOLLOSO

DATE/TIME OF TEST: 12/22/2023 23:47:36

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.061	23:42
Breath Volume	1.367	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.059	23:46
Breath Volume	2.308	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA, STEVENS, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: AMENDOLA, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TEEMS

DATE/TIME OF TEST: 12/23/2023 00:20:00

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.188	00:15
Breath Volume	3.976	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.186	00:18
Breath Volume	2.054	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:20

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CLEMENTS, TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: VALLE ZAPATA

DATE/TIME OF TEST: 12/23/2023 04:22:58

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.179	04:18
Breath Volume	1.867	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.192	04:21
Breath Volume	2.265	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PRIETO, CARLOS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MORENO GONZALEZ

DATE/TIME OF TEST: 12/23/2023 04:32:41

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.148	04:28
Breath Volume	2.527	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.145	04:31
Breath Volume	3.078	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,DONALD,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STANKOSKY,DONALD
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/23/2023 05:32:52

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.080	05:27
Air Blank	0.000	05:27
Subject Sample # 1	0.115	05:28
Breath Volume	3.667	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.101	05:31
Breath Volume	2.781	05:31
Air Blank	0.000	05:31
Control Test	0.079	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GOMEZ, G
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 12/23/2023 18:10:57

Diagnostics Check	OK	18:03
Air Blank	0.000	18:04
Control Test	0.081	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.000	18:06
Breath Volume	2.367	18:06
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.000	18:09
Breath Volume	1.511	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 12/24/2023 19:58:48

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank	0.000	19:56	
Subject Sample	# 1	19:57	Subject Test Refused
Breath Volume	0.000	19:57	
Air Blank	0.000	19:57	
Control Test	0.080	19:58	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: LUME

DATE/TIME OF TEST: 12/25/2023 04:56:34

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.079	04:45	
Air Blank	0.000	04:45	
Subject Sample	# 1 0.104	04:48	Volume Not Met
Breath Volume	0.710	04:48	
Air Blank	0.000	04:49	
Air Blank	0.000	04:51	
Subject Sample	# 2 0.093	04:51	
Breath Volume	1.300	04:51	
Air Blank	0.000	04:52	
Air Blank	0.000	04:54	
Subject Sample	# 3 0.098	04:55	
Breath Volume	1.281	04:55	
Air Blank	0.000	04:55	
Control Test	0.078	04:55	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MYNAUGH,JOSEPH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: OSPINA RAIGOSA

DATE/TIME OF TEST: 12/25/2023 05:49:00

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.209	05:44
Breath Volume	1.429	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.211	05:47
Breath Volume	1.742	05:47
Air Blank	0.000	05:48
Control Test	0.079	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:49

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HAGENKOTTER

DATE/TIME OF TEST: 12/27/2023 02:23:49

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.170	02:19
Breath Volume	3.039	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.162	02:22
Breath Volume	3.117	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 01/24/2024 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: THORNBURG

DATE/TIME OF TEST: 12/28/2023 23:42:39

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.196	23:38
Breath Volume	1.425	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.197	23:41
Breath Volume	1.468	23:41
Air Blank	0.000	23:41
Control Test	0.078	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COKER,JUSTIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SOLIS AGUSTIN

DATE/TIME OF TEST: 12/29/2023 00:17:03

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.157	00:12
Breath Volume	1.972	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.170	00:15
Breath Volume	2.578	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING, JONATHAN, L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MCMILLION, IAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CHAMBROT

DATE/TIME OF TEST: 12/29/2023 01:50:36

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.187	01:43
Breath Volume	2.941	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.164	01:45
Breath Volume	1.128	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 3	0.169	01:49
Breath Volume	1.781	01:49
Air Blank	0.000	01:49
Control Test	0.076	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 12/29/2023 02:33:53

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.170	02:29
Breath Volume	1.648	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.168	02:32
Breath Volume	2.171	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VEGA, JAHMIL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 12/29/2023 03:10:58

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.126	03:06
Breath Volume	1.738	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.118	03:09
Breath Volume	1.734	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHLENCAMP,HOLLY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: PENCE

DATE/TIME OF TEST: 12/29/2023 23:19:53

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.132	23:15
Breath Volume	1.933	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.135	23:18
Breath Volume	1.613	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHLENCAMP,HOLLY,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOWERS,KAITLIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: OCONNELL

DATE/TIME OF TEST: 12/29/2023 23:37:25

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.196	23:32
Breath Volume	1.554	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.200	23:35
Breath Volume	3.500	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RHYNES

DATE/TIME OF TEST: 12/30/2023 02:11:35

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.115	02:07
Breath Volume	1.320	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.113	02:10
Breath Volume	1.417	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOJ PEREZ

DATE/TIME OF TEST: 12/30/2023 03:57:56

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.192	03:53
Breath Volume	3.015	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.198	03:56
Breath Volume	3.046	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CHANTHAVONG,PHOURATSA
MAY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/31/2023 00:44:34

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.220	00:39
Breath Volume	2.273	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.228	00:43
Breath Volume	2.812	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KNAPP,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PABELLON

DATE/TIME OF TEST: 12/31/2023 03:48:25

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.085	03:43
Breath Volume	1.843	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.088	03:46
Breath Volume	1.410	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REID,JILLIAN,F
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: BORNSCHEIN

DATE/TIME OF TEST: 12/31/2023 05:07:06

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.079	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:05	Subject Test Refused
Breath Volume	0.199	05:05	
Air Blank	0.000	05:06	
Control Test	0.079	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/31/2023 05:31:19

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.124	05:26
Breath Volume	2.000	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.128	05:29
Breath Volume	1.769	05:29
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCO, ANTHONY, P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN, TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 12/31/2023 20:32:26

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.178	20:27
Breath Volume	3.199	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.161	20:30
Breath Volume	2.144	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCPHERSON,JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DAHLMAN

DATE/TIME OF TEST: 12/31/2023 20:58:14

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.079	20:57	
Air Blank		20:57	RFI Detect
Air Blank	0.000	20:58	

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCPHERSON,JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DAHLMAN

DATE/TIME OF TEST: 12/31/2023 21:06:21

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.184	21:01
Breath Volume	1.621	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.178	21:04
Breath Volume	1.859	21:04
Air Blank	0.000	21:05
Control Test	0.079	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 01/24/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGAN,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: LAWSON,MADISON
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/19/2023 03:44:41

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.082	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.236	03:39
Breath Volume	2.007	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.232	03:43
Breath Volume	1.968	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/04/2024 12:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLARD,DANIEL,C
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: FISHER,CHRISTOPHER
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: REYNOSA PEREZ

DATE/TIME OF TEST: 12/01/2023 22:33:11

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.239	22:28
Breath Volume	1.218	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.235	22:31
Breath Volume	1.980	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SLAZAS

DATE/TIME OF TEST: 12/03/2023 20:27:20

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.169	20:22
Breath Volume	1.976	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.172	20:25
Breath Volume	3.414	20:25
Air Blank	0.000	20:26
Control Test	0.078	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LOPEZ SONTAY

DATE/TIME OF TEST: 12/06/2023 20:47:47

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.204	20:43
Breath Volume	1.945	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.210	20:46
Breath Volume	1.871	20:46
Air Blank	0.000	20:46
Control Test	0.080	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: SUPIJEVA,ZUKHRA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: COHENS

DATE/TIME OF TEST: 12/06/2023 23:33:28

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.079	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	0.227	23:31
Breath Volume	0.437	23:31	Volume Not Met
Air Blank	0.000	23:32	
Control Test	0.079	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: SUPIJEVA,ZUKHRA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: COHENS

DATE/TIME OF TEST: 12/06/2023 23:43:31

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.242	23:37
Breath Volume	1.996	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.251	23:41
Breath Volume	2.585	23:41
Air Blank	0.000	23:42
Control Test	0.077	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TOUSSAINT

DATE/TIME OF TEST: 12/11/2023 18:49:33

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.146	18:44
Breath Volume	1.941	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.150	18:47
Breath Volume	2.253	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: SOLERO

DATE/TIME OF TEST: 12/15/2023 19:46:34

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Subject Sample # 1	0.202	19:41
Breath Volume	1.539	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.214	19:44
Breath Volume	1.601	19:44
Air Blank	0.000	19:45
Control Test	0.078	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,JAMESLOO,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: BOGAERT,CASSANDRA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: KOLEK

DATE/TIME OF TEST: 12/16/2023 02:03:33

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.231	01:59
Breath Volume	2.003	01:59
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.237	02:01
Breath Volume	1.769	02:01
Air Blank	0.000	02:02
Control Test	0.077	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: YOUNG,CHARLES
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: TURNQUIST

DATE/TIME OF TEST: 12/16/2023 21:34:02

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Control Test	0.078	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 12/19/2023 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 02422080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 12/20/2023 21:52:13

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.405	21:44
Breath Volume	1.199	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.370	21:47
Breath Volume	1.105	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 3	0.378	21:50
Breath Volume	1.183	21:50
Air Blank	0.000	21:51
Control Test	0.077	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 01/17/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,LAUREN,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: CHAVIS,H
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CHEVARRIA MEDINA

DATE/TIME OF TEST: 12/24/2023 01:02:40

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.079	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.195	00:58
Breath Volume	2.468	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.193	01:01
Breath Volume	1.847	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 01/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTH,L,A
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: RICHETTI,K
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A3
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: FIGUEROA ALVARADO

DATE/TIME OF TEST: 12/24/2023 19:33:32

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.079	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.157	19:29
Breath Volume	2.609	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.147	19:31
Breath Volume	2.121	19:31
Air Blank	0.000	19:32
Control Test	0.078	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 01/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: IGNOFFO,ZACHARY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: SUTHERLAND

DATE/TIME OF TEST: 12/19/2023 21:39:11

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.167	21:34
Breath Volume	1.597	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample # 2	0.166	21:37
Breath Volume	2.238	21:37
Air Blank	0.000	21:38
Control Test	0.078	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 01/10/2024 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: AGATA

DATE/TIME OF TEST: 12/19/2023 23:01:34

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.171	22:57
Breath Volume	1.632	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.173	22:59
Breath Volume	1.886	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 01/10/2024 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BILLUE,MOLLY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SPINK

DATE/TIME OF TEST: 12/27/2023 23:07:44

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.086	23:03
Breath Volume	1.425	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.080	23:06
Breath Volume	1.878	23:06
Air Blank	0.000	23:06
Control Test	0.077	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 12/29/2023 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 12/09/2023 22:47:11

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.123	22:42
Breath Volume	2.070	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.119	22:45
Breath Volume	1.472	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 12/21/2023 07:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: DAMON,RUSSELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: AUPLANT

DATE/TIME OF TEST: 12/26/2023 03:02:03

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.149	02:57
Breath Volume	2.878	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.160	03:00
Breath Volume	2.085	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

Data Download Date/Time: 01/14/2024 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HITCHCOCK,LUCAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: OLSEN

DATE/TIME OF TEST: 12/02/2023 01:39:34

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.082	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	3.304	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	3.144	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/15/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BATCH,JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: TUREY

DATE/TIME OF TEST: 12/03/2023 04:01:05

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.172	03:56
Breath Volume	1.972	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.175	03:59
Breath Volume	2.164	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/15/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SCHULZE,JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CROW

DATE/TIME OF TEST: 12/08/2023 03:19:06

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.195	03:14
Breath Volume	3.113	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.186	03:17
Breath Volume	2.761	03:17
Air Blank	0.000	03:18
Control Test	0.077	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 12/15/2023 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIAS,HUNTER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GALLOS

DATE/TIME OF TEST: 12/24/2023 00:55:14

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.082	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.075	00:50
Breath Volume	2.253	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.075	00:53
Breath Volume	1.792	00:53
Air Blank	0.000	00:54
Control Test	0.082	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 01/25/2024 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LOPEZ MAZARIEGOS

DATE/TIME OF TEST: 12/01/2023 00:38:37

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.166	00:34
Breath Volume	3.394	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.161	00:36
Breath Volume	3.066	00:36
Air Blank	0.000	00:37
Control Test	0.081	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 12/12/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PETERSON, THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 12/01/2023 03:40:57

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.121	03:36
Breath Volume	2.382	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.120	03:39
Breath Volume	1.328	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 12/12/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDERO, PATRICIA
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MIRANDA AGUILAR

DATE/TIME OF TEST: 12/01/2023 23:31:22

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.081	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.179	23:26
Breath Volume	3.257	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.178	23:29
Breath Volume	3.339	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 12/12/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: OSWALT

DATE/TIME OF TEST: 12/02/2023 01:34:35

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.083	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.169	01:30
Breath Volume	1.335	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.171	01:32
Breath Volume	1.953	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 12/12/2023 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LONEY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROMAN LEON

DATE/TIME OF TEST: 12/03/2023 03:14:34

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.082	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.161	03:10
Breath Volume	2.097	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.157	03:12
Breath Volume	1.546	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 12/12/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANCEL, KAREN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: DUBOURG

DATE/TIME OF TEST: 12/06/2023 01:31:08

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.083	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.141	01:26
Breath Volume	1.597	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.139	01:29
Breath Volume	1.839	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 12/12/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: KIEL

DATE/TIME OF TEST: 12/07/2023 00:21:53

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.083	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.184	00:16
Breath Volume	1.390	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.189	00:20
Breath Volume	2.425	00:20
Air Blank	0.000	00:20
Control Test	0.083	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/12/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: AGUILAR AGUILAR

DATE/TIME OF TEST: 12/09/2023 01:08:45

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.082	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.194	01:04
Breath Volume	2.648	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.182	01:07
Breath Volume	2.628	01:07
Air Blank	0.000	01:07
Control Test	0.081	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/12/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATTINGLY, JONATHAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: LUNA LOPEZ

DATE/TIME OF TEST: 12/10/2023 23:41:29

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.082	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.175	23:36
Breath Volume	2.179	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.187	23:39
Breath Volume	4.160	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 12/12/2023 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DA FONSECA,ARTHUR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 12/17/2023 04:08:29

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.100	04:00	Volume Not Met
Breath Volume	1.066	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 2 0.101	04:03	
Breath Volume	1.156	04:03	
Air Blank	0.000	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 3 0.099	04:06	
Breath Volume	1.109	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRULE,JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: OCHOA CASTILLO

DATE/TIME OF TEST: 12/17/2023 05:41:06

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.081	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1 0.153	05:33	Volume Not Met
Breath Volume	0.199	05:33	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample	# 2 0.148	05:39	Volume Not Met
Breath Volume	0.015	05:39	
Air Blank	0.000	05:40	
Control Test	0.079	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUERRERO,MANUEL
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DELCERRORODRIGUEZ

DATE/TIME OF TEST: 12/17/2023 07:02:36

Diagnostics Check	OK	06:49	
Air Blank	0.000	06:49	
Control Test	0.081	06:50	
Air Blank	0.000	06:50	
Subject Sample	# 1 0.191	06:52	
Breath Volume	1.343	06:52	
Air Blank	0.000	06:53	
Air Blank	0.000	06:55	
Subject Sample	# 2	06:58	No Sample Provided
Breath Volume	0.000	06:58	
Air Blank	0.000	06:58	
Air Blank	0.000	07:00	
Subject Sample	# 3	07:01	Subject Test Refused
Breath Volume	0.000	07:01	
Air Blank	0.000	07:01	
Control Test	0.081	07:01	
Air Blank	0.000	07:02	
Diagnostics Check	OK	07:02	

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMPSON,KYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BLACKLEDGE

DATE/TIME OF TEST: 12/19/2023 01:26:55

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.083	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.207	01:22
Breath Volume	1.710	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.210	01:25
Breath Volume	1.148	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 12/23/2023 02:48:57

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.109	02:44
Breath Volume	1.277	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.121	02:47
Breath Volume	1.207	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHARLES,DYON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: ROSALES GALLEGOS

DATE/TIME OF TEST: 12/24/2023 04:01:05

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.202	03:56
Breath Volume	2.496	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.201	03:59
Breath Volume	2.000	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:01

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: MARIMON

DATE/TIME OF TEST: 12/31/2023 05:04:29

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.083	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.153	04:59
Breath Volume	1.402	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.152	05:02
Breath Volume	1.265	05:02
Air Blank	0.000	05:03
Control Test	0.082	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 01/15/2024 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WHITE III

DATE/TIME OF TEST: 12/15/2023 02:01:49

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.089	01:57
Breath Volume	2.421	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.090	02:00
Breath Volume	2.109	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHARLES,DYON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: CIESLIK

DATE/TIME OF TEST: 12/16/2023 05:06:07

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.076	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.145	05:01
Breath Volume	1.527	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.156	05:04
Breath Volume	1.738	05:04
Air Blank	0.000	05:05
Control Test	0.076	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: QUINTANA BARRETO

DATE/TIME OF TEST: 12/18/2023 22:42:26

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.078	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.000	22:38
Breath Volume	2.324	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.000	22:41
Breath Volume	3.046	22:41
Air Blank	0.000	22:41
Control Test	0.078	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMON, JEREMY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 12/24/2023 23:45:36

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.077	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.259	23:38
Breath Volume	3.234	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.265	23:44
Breath Volume	3.347	23:44
Air Blank	0.000	23:44
Control Test	0.075	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAMANNA, MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: SALOME ECHEVARRIA

DATE/TIME OF TEST: 12/25/2023 22:19:42

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.093	22:15
Breath Volume	2.250	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.100	22:18
Breath Volume	2.152	22:18
Air Blank	0.000	22:18
Control Test	0.077	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: VAN DER BERG

DATE/TIME OF TEST: 12/30/2023 01:15:11

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.078	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.160	01:10
Breath Volume	1.324	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.151	01:13
Breath Volume	1.804	01:13
Air Blank	0.000	01:14
Control Test	0.076	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/12/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 12/31/2023 02:24:46

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.202	02:20
Breath Volume	3.445	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.194	02:23
Breath Volume	2.773	02:23
Air Blank	0.000	02:23
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 01/15/2024 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOPEZ,GEORGE,O

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: LOPEZ,GEORGE

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A2

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 19:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/10/2023 20:15:10

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.182	20:10
Breath Volume	1.449	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.185	20:13
Breath Volume	2.476	20:13
Air Blank	0.000	20:14
Control Test	0.080	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 01/12/2024 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WIGGINS, TYLER, D

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: WIGGINS, TYLER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A2

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 05:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: LANZA

DATE/TIME OF TEST: 12/24/2023 06:47:54

Diagnostics Check	OK	06:47	
Air Blank		06:47	RFI Detect
Air Blank	0.000	06:47	

Data Download Date/Time: 01/12/2024 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WIGGINS, TYLER, D

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: WIGGINS, TYLER

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 12/06/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A2

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 05:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: LANZA

DATE/TIME OF TEST: 12/24/2023 06:56:55

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.081	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.160	06:52
Breath Volume	1.160	06:52
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.169	06:55
Breath Volume	1.730	06:55
Air Blank	0.000	06:56
Control Test	0.081	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 01/12/2024 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOREHEAD,AVERY,F
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON,CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 12/03/2023 03:38:17

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.156	03:33
Breath Volume	2.367	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.160	03:36
Breath Volume	2.214	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/19/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BOLDT,PAIGE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:17

SUBJECT'S LAST NAME: FLORES MORALES

DATE/TIME OF TEST: 12/10/2023 10:43:17

Diagnostics Check	OK	10:38	
Air Blank	0.000	10:39	
Control Test	0.079	10:39	
Air Blank	0.000	10:40	
Subject Sample	# 1 0.255	10:42	
Breath Volume	1.593	10:42	
Air Blank		10:42	Ambient Fail
Air Blank	0.000	10:43	

Data Download Date/Time: 12/19/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BOLDT,PAIGE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/16/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:17

SUBJECT'S LAST NAME: FLORES MORALES

DATE/TIME OF TEST: 12/10/2023 10:52:51

Diagnostics Check	OK	10:46
Air Blank	0.000	10:46
Control Test	0.079	10:47
Air Blank	0.000	10:47
Subject Sample # 1	0.256	10:48
Breath Volume	1.683	10:48
Air Blank	0.000	10:49
Air Blank	0.000	10:50
Subject Sample # 2	0.252	10:51
Breath Volume	1.718	10:51
Air Blank	0.000	10:51
Control Test	0.079	10:52
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 12/19/2023 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LEBRETON, QUINN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 12/24/2023 00:03:04

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.166	23:58
Breath Volume	1.640	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.173	00:01
Breath Volume	1.625	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:03

Data Download Date/Time: 01/15/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SCHALLON, CHRISTIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 12/24/2023 00:28:38

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.241	00:23
Breath Volume	2.214	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.236	00:27
Breath Volume	1.699	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/15/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BOLDT, PAIGE
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: METZLER

DATE/TIME OF TEST: 12/28/2023 21:38:55

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.078	21:33	
Air Blank	0.000	21:33	
Subject Sample	# 1 0.102	21:34	
Breath Volume	1.574	21:34	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 2 0.107	21:37	
Breath Volume	3.554	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:38	
Air Blank		21:38	RFI Detect
Air Blank	0.000	21:38	

Data Download Date/Time: 01/15/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: STEPP,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/01/2023 00:53:49

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.079	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.154	00:49
Breath Volume	3.050	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.150	00:52
Breath Volume	2.816	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUMONT, TRAVIS, H

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: DUMONT, TRAVIS

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 04:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/03/2023 04:36:20

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.125	04:30
Breath Volume	3.972	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.113	04:34
Breath Volume	2.914	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: DALTON,JAKE
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 12/08/2023 08:24:01

Diagnostics Check	OK	08:18
Air Blank	0.000	08:18
Control Test	0.081	08:18
Air Blank	0.000	08:19
Subject Sample # 1	0.000	08:19
Breath Volume	1.960	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:22
Subject Sample # 2	0.000	08:22
Breath Volume	1.699	08:22
Air Blank	0.000	08:23
Control Test	0.080	08:23
Air Blank	0.000	08:23
Diagnostics Check	OK	08:24

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNETT,JAMES,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BARNETT,JAMES

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GUGEL

DATE/TIME OF TEST: 12/10/2023 01:58:09

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.094	01:53
Breath Volume	1.148	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.095	01:56
Breath Volume	1.390	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 02:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/10/2023 03:07:07

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.168	03:02
Breath Volume	1.851	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.169	03:05
Breath Volume	1.773	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: FRANZE,NICHOLAS
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: LIFORD

DATE/TIME OF TEST: 12/18/2023 09:40:58

Diagnostics Check	OK	09:29	
Air Blank	0.000	09:30	
Control Test	0.080	09:30	
Air Blank	0.000	09:30	
Subject Sample	# 1	09:34	No Sample Provided
Breath Volume	0.000	09:34	
Air Blank	0.000	09:34	
Air Blank	0.000	09:36	
Subject Sample	# 2	09:39	No Sample Provided
Breath Volume	0.000	09:39	
Air Blank	0.000	09:40	
Control Test	0.081	09:40	
Air Blank	0.000	09:40	
Diagnostics Check	OK	09:40	

Data Download Date/Time: 12/27/2023 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 12/01/2023 00:28:34

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.079	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.079	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 12/01/2023 04:00:04

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.074	03:55
Breath Volume	1.507	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.074	03:58
Breath Volume	1.187	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: IBRISEVIC

DATE/TIME OF TEST: 12/02/2023 01:55:34

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.077	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, LEVI, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE, LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/02/2023 02:58:21

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.078	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/10/2023 01:36:42

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.079	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CASTELLUCI

DATE/TIME OF TEST: 12/10/2023 03:25:53

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SCIARA

DATE/TIME OF TEST: 12/14/2023 23:07:25

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 12/20/2023 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,LEVI,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,LEVI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 12/20/2023 20:48:12

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.023	20:43
Breath Volume	3.031	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.022	20:46
Breath Volume	2.765	20:46
Air Blank	0.000	20:47
Control Test	0.078	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 01/17/2024 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: LUKASZ

DATE/TIME OF TEST: 12/22/2023 00:29:23

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.135	00:24
Breath Volume	2.285	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.131	00:27
Breath Volume	2.183	00:27
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 01/17/2024 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACIA,STEVEN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRAHAM,DEANGELO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG205902
DRY GAS EXPIRY DATE: 02/28/2024
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 12/31/2023 18:55:06

Diagnostics Check	OK	18:47
Air Blank	0.000	18:48
Control Test	0.079	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.179	18:49
Breath Volume	1.238	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.170	18:53
Breath Volume	1.648	18:53
Air Blank	0.000	18:54
Control Test	0.077	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 01/17/2024 11:16