

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BARNETT,JAMES

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 28021080A3

DRY GAS EXPIRY DATE: 12/05/2023

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/14/2024 02:34:36

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.077	02:29
Breath Volume	2.019	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.066	02:32
Breath Volume	1.417	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 01/28/2024 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HILL,CORNELIUS
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 01/01/2024 00:30:37

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.184	00:26
Breath Volume	2.000	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.176	00:29
Breath Volume	1.695	00:29
Air Blank	0.000	00:29
Control Test	0.077	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 01/18/2024 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESSLER,GREYSON,L
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: PEREZ,SAMANTHA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 01/02/2024 23:16:36

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.081	23:14	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:16	Sequence Aborted
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	

Data Download Date/Time: 01/18/2024 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESSLER,GREYSON,L
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/03/2024 01:30:06

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.081	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	2.156	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.000	01:28
Breath Volume	2.203	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 01/18/2024 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 01/08/2024 01:11:25

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.172	01:05
Breath Volume	1.250	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.185	01:09
Breath Volume	1.796	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 01/18/2024 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/12/2024 23:04:06

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:03	
Control Test	0.078	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 01/18/2024 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: WELLS,R
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 01/18/2024 17:43:16

Diagnostics Check	OK	17:37
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.216	17:38
Breath Volume	1.460	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.210	17:41
Breath Volume	1.484	17:41
Air Blank	0.000	17:42
Control Test	0.078	17:42
Air Blank	0.000	17:43
Diagnostics Check	OK	17:43

Data Download Date/Time: 02/20/2024 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/20/2024 00:13:41

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:11	
Control Test	0.076	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:12	
Control Test	0.076	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 02/20/2024 15:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: KAPLAN

DATE/TIME OF TEST: 01/21/2024 02:01:01

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.077	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 02/20/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZZI,SHANNON,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: VASQUEZ HERNANDEZ

DATE/TIME OF TEST: 01/23/2024 01:03:10

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.225	00:58
Breath Volume	1.125	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.213	01:01
Breath Volume	1.968	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 02/20/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 01/26/2024 03:58:12

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.078	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 02/20/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: SOULE-WEAVER

DATE/TIME OF TEST: 01/27/2024 19:54:17

Diagnostics Check	OK	19:47
Air Blank	0.000	19:47
Control Test	0.078	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.124	19:49
Breath Volume	1.105	19:49
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.131	19:52
Breath Volume	1.316	19:52
Air Blank	0.000	19:53
Control Test	0.078	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 02/20/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MATHOSIAN,M
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: NHON

DATE/TIME OF TEST: 01/29/2024 01:31:35

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:24	
Subject Sample	# 1	01:27	No Sample Provided
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.079	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 02/20/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CPL. CONANT,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: CASTELLANO CRUZ

DATE/TIME OF TEST: 01/01/2024 04:22:24

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.229	04:17
Breath Volume	2.675	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.223	04:20
Breath Volume	1.550	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 01/18/2024 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: JAGIELSKI

DATE/TIME OF TEST: 01/01/2024 05:51:09

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:45
Subject Sample # 1	0.177	05:46
Breath Volume	1.246	05:46
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.166	05:49
Breath Volume	1.292	05:49
Air Blank	0.000	05:50
Control Test	0.080	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 01/18/2024 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEPUTY,SHERIFF
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 01/02/2024 17:53:18

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.081	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.000	17:48
Breath Volume	1.644	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.000	17:51
Breath Volume	1.804	17:51
Air Blank	0.000	17:52
Control Test	0.081	17:52
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 01/18/2024 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/04/2024 13:44:07

Diagnostics Check	OK	13:38
Air Blank	0.000	13:38
Control Test	0.080	13:38
Air Blank	0.000	13:39
Subject Sample # 1	0.000	13:39
Breath Volume	2.425	13:39
Air Blank	0.000	13:40
Air Blank	0.000	13:42
Subject Sample # 2	0.000	13:42
Breath Volume	2.015	13:42
Air Blank	0.000	13:43
Control Test	0.081	13:43
Air Blank	0.000	13:44
Diagnostics Check	OK	13:44

Data Download Date/Time: 01/18/2024 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,C,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,CDFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 01/06/2024 23:04:21

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.188	22:57	Volume Not Met
Breath Volume	0.207	22:57	
Air Blank	0.000	22:57	
Air Blank	0.000	22:59	
Subject Sample	# 2 0.175	23:02	Volume Not Met
Breath Volume	0.492	23:02	
Air Blank	0.000	23:03	
Control Test	0.078	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 01/18/2024 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARRO,M
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BEVERIDGE

DATE/TIME OF TEST: 01/07/2024 03:46:08

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample # 1	0.183	03:41	
Breath Volume	1.914	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample # 2	0.182	03:44	
Breath Volume	1.187	03:44	
Air Blank	0.000	03:45	
Control Test	0.074	03:45	Control Outside Tolerance
Air Blank	0.000	03:46	

Data Download Date/Time: 01/18/2024 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Punta Gorda PD
ARRESTING OFFICER: SMITH,C
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PINAULT

DATE/TIME OF TEST: 01/12/2024 22:02:25

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.330	21:57
Breath Volume	1.347	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.338	22:00
Breath Volume	2.417	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 01/18/2024 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, STEVEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CARROLL, STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/13/2024 03:25:39

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.081	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1 0.152	03:21	Volume Not Met
Breath Volume	0.062	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.079	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 01/18/2024 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/15/2024 02:52:29

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.070	02:47
Breath Volume	1.835	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.069	02:51
Breath Volume	1.753	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/18/2024 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/15/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 01/16/2024 00:24:30

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.109	00:20
Breath Volume	1.472	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.117	00:23
Breath Volume	2.390	00:23
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 01/18/2024 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL, STEVEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CARROLL, STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WIRKUS JOHNSON

DATE/TIME OF TEST: 01/18/2024 22:37:52

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.240	22:33
Breath Volume	1.980	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:35
Subject Sample # 2	0.239	22:36
Breath Volume	2.171	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 02/25/2024 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/19/2024 20:31:08

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.080	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.116	20:26
Breath Volume	2.210	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.117	20:29
Breath Volume	2.511	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 02/25/2024 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WYMER,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ROHLING

DATE/TIME OF TEST: 01/19/2024 23:41:57

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.187	23:37
Breath Volume	2.046	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.174	23:40
Breath Volume	1.242	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 02/25/2024 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWARTZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/20/2024 00:10:36

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:06
Breath Volume	1.750	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:09
Breath Volume	1.574	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 02/25/2024 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BALSINGER,DANE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MILES-BRIGGS

DATE/TIME OF TEST: 01/21/2024 02:14:52

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.082	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.189	02:10
Breath Volume	1.527	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.188	02:13
Breath Volume	1.214	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 02/25/2024 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARCHMAN,MARCHMAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/21/2024 18:20:23

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.081	18:14
Air Blank	0.000	18:15
Subject Sample # 1	0.221	18:16
Breath Volume	2.187	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.219	18:18
Breath Volume	1.882	18:18
Air Blank	0.000	18:19
Control Test	0.081	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 02/25/2024 20:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ACEBO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: COLGAN

DATE/TIME OF TEST: 01/22/2024 01:15:22

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.081	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.208	01:10
Breath Volume	2.281	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.216	01:13
Breath Volume	2.699	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 02/25/2024 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARCH,MARCH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 01/22/2024 02:25:17

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.082	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.051	02:20
Breath Volume	1.953	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.051	02:23
Breath Volume	1.812	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/25/2024 20:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOPPLERBOUGH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 09:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GOFF

DATE/TIME OF TEST: 01/22/2024 10:03:20

Diagnostics Check	OK	09:57
Air Blank	0.000	09:57
Control Test	0.082	09:57
Air Blank	0.000	09:58
Subject Sample # 1	0.209	09:58
Breath Volume	2.191	09:58
Air Blank	0.000	09:59
Air Blank	0.000	10:01
Subject Sample # 2	0.194	10:01
Breath Volume	2.347	10:01
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:03
Diagnostics Check	OK	10:03

Data Download Date/Time: 02/25/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CD,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CID,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BIERNAT

DATE/TIME OF TEST: 01/25/2024 20:26:04

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.081	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.014	20:21
Breath Volume	1.140	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:23
Subject Sample # 2	0.013	20:24
Breath Volume	1.667	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:26

Data Download Date/Time: 02/25/2024 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAZZARESE,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,C
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MEDINA DE LA CRUZ

DATE/TIME OF TEST: 01/27/2024 21:01:21

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.082	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.184	20:49	
Breath Volume	1.363	20:49	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2 0.146	20:54	Volume Not Met
Breath Volume	0.191	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample	# 3 0.204	20:59	
Breath Volume	1.117	20:59	
Air Blank	0.000	21:00	
Control Test	0.080	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 02/25/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHWOCHO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 01/28/2024 13:44:30

Diagnostics Check	OK	13:38
Air Blank	0.000	13:38
Control Test	0.081	13:39
Air Blank	0.000	13:39
Subject Sample # 1	0.000	13:39
Breath Volume	3.316	13:39
Air Blank	0.000	13:40
Air Blank	0.000	13:42
Subject Sample # 2	0.000	13:43
Breath Volume	3.269	13:43
Air Blank	0.000	13:43
Control Test	0.081	13:43
Air Blank	0.000	13:44
Diagnostics Check	OK	13:44

Data Download Date/Time: 02/25/2024 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: DITOTA

DATE/TIME OF TEST: 01/28/2024 14:04:25

Diagnostics Check	OK	13:57
Air Blank	0.000	13:57
Control Test	0.082	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.000	14:00
Breath Volume	1.121	14:00
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.000	14:03
Breath Volume	1.796	14:03
Air Blank	0.000	14:03
Control Test	0.081	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 02/25/2024 20:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEONARD,ALEYA,D
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TEMS,NICHLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: TINGLEY

DATE/TIME OF TEST: 01/30/2024 22:30:59

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.173	22:26
Breath Volume	1.570	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.179	22:29
Breath Volume	1.832	22:29
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 02/25/2024 20:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GUILARTE-DOMINGUEZ

DATE/TIME OF TEST: 01/01/2024 03:20:47

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.000	03:16
Breath Volume	1.593	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.000	03:19
Breath Volume	1.171	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 01/15/2024 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: VALENCIA VIXTHA

DATE/TIME OF TEST: 01/01/2024 05:01:24

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.170	04:53
Breath Volume	3.097	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.147	04:56
Breath Volume	2.085	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 3	0.162	04:59
Breath Volume	4.406	04:59
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 01/15/2024 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STYLES,LAUREN,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REID,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 01/01/2024 18:38:19

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	
Air Blank	0.000	18:24	
Subject Sample	# 1 0.167	18:28	Volume Not Met
Breath Volume	0.476	18:28	
Air Blank	0.000	18:28	
Air Blank	0.000	18:30	
Subject Sample	# 2 0.186	18:31	
Breath Volume	1.289	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample	# 3 0.199	18:36	
Breath Volume	1.691	18:36	
Air Blank	0.000	18:37	
Control Test	0.077	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REICH,OWEN,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,OWEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: VLASYUK

DATE/TIME OF TEST: 01/03/2024 22:15:01

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	2.476	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	2.933	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:15

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 01/04/2024 21:51:10

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.079	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:46
Breath Volume	2.042	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:49
Breath Volume	1.824	21:49
Air Blank	0.000	21:50
Control Test	0.079	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: GRIFFITHS

DATE/TIME OF TEST: 01/06/2024 02:55:09

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1 0.043	02:47	Volume Not Met
Breath Volume	1.050	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2 0.047	02:50	
Breath Volume	1.117	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 3 0.044	02:53	
Breath Volume	1.144	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/06/2024 20:03:30

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.151	19:58
Breath Volume	3.253	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.159	20:01
Breath Volume	3.429	20:01
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: RADDAS

DATE/TIME OF TEST: 01/13/2024 08:04:46

Diagnostics Check	OK	07:57
Air Blank	0.000	07:57
Control Test	0.080	07:58
Air Blank	0.000	07:58
Subject Sample # 1	0.141	07:59
Breath Volume	4.500	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.131	08:03
Breath Volume	3.046	08:03
Air Blank	0.000	08:03
Control Test	0.077	08:04
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: FERNANDEZ CUEVAS

DATE/TIME OF TEST: 01/14/2024 03:39:57

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.045	03:35
Breath Volume	1.359	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.048	03:38
Breath Volume	1.378	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: BRIGHTMAN

DATE/TIME OF TEST: 01/14/2024 18:33:56

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:22	
Control Test	0.080	18:22	
Air Blank	0.000	18:23	
Subject Sample	# 1 0.137	18:26	Volume Not Met
Breath Volume	0.148	18:26	
Air Blank	0.000	18:27	
Air Blank	0.000	18:29	
Subject Sample	# 2 0.145	18:32	Volume Not Met
Breath Volume	0.035	18:32	
Air Blank	0.000	18:33	
Control Test	0.077	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/29/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: BRIGHTMAN

DATE/TIME OF TEST: 01/14/2024 19:07:25

Diagnostics Check	OK	18:56
Air Blank	0.000	18:57
Control Test	0.077	18:57
Air Blank	0.000	18:58
Subject Sample # 1	0.129	19:00
Breath Volume	1.515	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.132	19:05
Breath Volume	1.449	19:05
Air Blank	0.000	19:06
Control Test	0.076	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 01/15/2024 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,X
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: DESPANZA

DATE/TIME OF TEST: 01/17/2024 22:49:36

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.080	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: RUIZ-CORNEJO

DATE/TIME OF TEST: 01/18/2024 15:02:52

Diagnostics Check	OK	14:56
Air Blank	0.000	14:57
Control Test	0.080	14:57
Air Blank	0.000	14:57
Subject Sample # 1	0.000	14:58
Breath Volume	2.914	14:58
Air Blank	0.000	14:59
Air Blank	0.000	15:01
Subject Sample # 2	0.000	15:01
Breath Volume	2.843	15:01
Air Blank	0.000	15:01
Control Test	0.079	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 01/20/2024 03:35:11

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.000	03:30
Breath Volume	2.511	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.000	03:33
Breath Volume	2.906	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,LANE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SCALF

DATE/TIME OF TEST: 01/20/2024 05:12:11

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 1 0.055	05:05	Volume Not Met
Breath Volume	0.601	05:05	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 2	05:10	No Sample Provided
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Control Test	0.076	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: SIMON-HUIZACHE

DATE/TIME OF TEST: 01/20/2024 06:18:53

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.165	06:14
Breath Volume	1.281	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:16
Subject Sample # 2	0.164	06:17
Breath Volume	2.371	06:17
Air Blank	0.000	06:17
Control Test	0.078	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: ROCHA MIRANDA

DATE/TIME OF TEST: 01/20/2024 18:55:11

Diagnostics Check	OK	18:47
Air Blank	0.000	18:48
Control Test	0.080	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.000	18:50
Breath Volume	2.554	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.000	18:53
Breath Volume	1.300	18:53
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:55
Diagnostics Check	OK	18:55

Data Download Date/Time: 02/29/2024 18:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/21/2024 04:00:58

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.166	03:56
Breath Volume	2.500	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.167	03:59
Breath Volume	2.585	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: TREJO

DATE/TIME OF TEST: 01/21/2024 07:47:10

Diagnostics Check	OK	07:40
Air Blank	0.000	07:41
Control Test	0.080	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.061	07:42
Breath Volume	2.117	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.060	07:45
Breath Volume	2.187	07:45
Air Blank	0.000	07:46
Control Test	0.079	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASCKO,J,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LASCKO,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/21/2024 15:08:09

Diagnostics Check	OK	15:01
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:03
Subject Sample # 1	0.222	15:03
Breath Volume	3.832	15:03
Air Blank	0.000	15:04
Air Blank	0.000	15:06
Subject Sample # 2	0.215	15:06
Breath Volume	2.960	15:06
Air Blank	0.000	15:07
Control Test	0.077	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/26/2024 02:05:59

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.182	02:01
Breath Volume	2.503	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.188	02:04
Breath Volume	2.316	02:04
Air Blank	0.000	02:05
Control Test	0.077	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/27/2024 03:49:56

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank		03:49	RFI Detect
Air Blank	0.000	03:49	

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 01/27/2024 03:57:49

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.172	03:53
Breath Volume	1.406	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.173	03:56
Breath Volume	1.519	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 01/28/2024 00:19:16

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.139	00:14
Breath Volume	3.574	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.131	00:17
Breath Volume	2.015	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/29/2024 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 01/05/2024 20:25:54

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.079	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.242	20:21
Breath Volume	2.652	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.244	20:24
Breath Volume	2.304	20:24
Air Blank	0.000	20:24
Control Test	0.079	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SANCHEZ ESCOBAR

DATE/TIME OF TEST: 01/06/2024 04:17:39

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.222	04:12
Breath Volume	2.003	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.232	04:15
Breath Volume	2.511	04:15
Air Blank	0.000	04:16
Control Test	0.077	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/06/2024 22:49:14

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1 0.143	22:44	
Breath Volume	1.707	22:44	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Control Test	0.079	22:48	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 01/06/2024 22:57:52

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.142	22:53
Breath Volume	1.945	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.140	22:56
Breath Volume	2.609	22:56
Air Blank	0.000	22:56
Control Test	0.077	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LUNA SANCHEZ

DATE/TIME OF TEST: 01/07/2024 05:29:17

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.077	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:27	Subject Test Refused
Breath Volume	0.000	05:27	
Air Blank	0.000	05:28	
Control Test	0.077	05:28	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:29	

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,BARBARA,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 01/07/2024 16:03:00

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:00	
Control Test	0.079	16:00	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:01	Subject Test Refused
Breath Volume	0.000	16:01	
Air Blank	0.000	16:02	
Control Test	0.079	16:02	
Air Blank	0.000	16:02	
Diagnostics Check	OK	16:03	

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,BARBARA,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 01/07/2024 16:12:56

Diagnostics Check	OK	16:10	
Air Blank	0.000	16:10	
Control Test	0.080	16:10	
Air Blank	0.000	16:11	
Subject Sample	# 1	16:11	Subject Test Refused
Breath Volume	0.000	16:11	
Air Blank	0.000	16:11	
Control Test	0.079	16:12	
Air Blank	0.000	16:12	
Diagnostics Check	OK	16:12	

Data Download Date/Time: 01/15/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VENEGAS CORTES

DATE/TIME OF TEST: 01/08/2024 04:17:26

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.126	04:12
Breath Volume	2.507	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.125	04:15
Breath Volume	2.277	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 01/15/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/08/2024 06:47:25

Diagnostics Check	OK	06:40
Air Blank	0.000	06:41
Control Test	0.079	06:41
Air Blank	0.000	06:42
Subject Sample # 1	0.133	06:42
Breath Volume	3.003	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.136	06:45
Breath Volume	3.207	06:45
Air Blank	0.000	06:46
Control Test	0.077	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 01/15/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MUNIZ

DATE/TIME OF TEST: 01/11/2024 02:33:50

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.196	02:28
Breath Volume	1.222	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.205	02:32
Breath Volume	2.425	02:32
Air Blank	0.000	02:32
Control Test	0.075	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/15/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 01/11/2024 22:46:56

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.080	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.155	22:42
Breath Volume	1.367	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.162	22:45
Breath Volume	1.175	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 01/29/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/13/2024 00:33:32

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.161	00:28
Breath Volume	1.457	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.166	00:32
Breath Volume	1.171	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 01/29/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 01/25/2024 00:50:12

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.240	00:45
Breath Volume	1.992	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.242	00:48
Breath Volume	1.792	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/29/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: FRANZONI

DATE/TIME OF TEST: 01/30/2024 21:34:50

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.176	21:27
Breath Volume	1.488	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:29
Subject Sample # 2	0.199	21:30
Breath Volume	1.941	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:32
Subject Sample # 3	0.193	21:33
Breath Volume	2.363	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 02/26/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 01/31/2024 16:41:08

Diagnostics Check	OK	16:34
Air Blank	0.000	16:35
Control Test	0.081	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.164	16:36
Breath Volume	3.203	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.161	16:39
Breath Volume	3.972	16:39
Air Blank	0.000	16:40
Control Test	0.082	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 02/26/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: RATEL

DATE/TIME OF TEST: 01/04/2024 20:43:12

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.180	20:38
Breath Volume	1.937	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.185	20:41
Breath Volume	1.425	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 01/26/2024 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JAMES, ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: NERBER

DATE/TIME OF TEST: 01/18/2024 02:30:22

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.171	02:25
Breath Volume	2.496	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.172	02:28
Breath Volume	2.058	02:28
Air Blank	0.000	02:29
Control Test	0.078	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 01/26/2024 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LANDON,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: BLISSETT,DEVON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 01/19/2024 00:04:57

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.077	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.171	23:52	
Breath Volume	2.300	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 3 0.122	23:57	Volume Not Met
Breath Volume	0.332	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:03	No Sample Provided
Breath Volume	0.000	00:03	
Air Blank	0.000	00:04	
Control Test	0.075	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 01/26/2024 17:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANDRE,
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33321080A2
DRY GAS EXPIRY DATE: 01/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 01/22/2024 02:04:49

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.192	02:00
Breath Volume	1.531	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.180	02:03
Breath Volume	1.394	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/26/2024 17:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ALFARO,DANIEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MARREROREYES

DATE/TIME OF TEST: 01/26/2024 05:16:26

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.079	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1 0.151	05:09	Volume Not Met
Breath Volume	0.488	05:09	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2 0.114	05:14	Volume Not Met
Breath Volume	0.449	05:14	
Air Blank	0.000	05:15	
Control Test	0.077	05:15	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 02/08/2024 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTILUS,LEBEDOPH,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: OTILUS,LEBEDOPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FORMIGA VERDAGUER

DATE/TIME OF TEST: 01/26/2024 18:27:34

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:25	
Control Test	0.079	18:25	
Air Blank	0.000	18:25	
Subject Sample	# 1	18:26	RFI Detect
Breath Volume	0.000	18:26	
Air Blank	0.000	18:26	
Control Test	0.077	18:26	
Air Blank	0.000	18:27	
Diagnostics Check	OK	18:27	

Data Download Date/Time: 01/30/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTILUS,LEBEDOPH,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: OTILUS,LEBEDOPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FORMIGA VERDAGUER

DATE/TIME OF TEST: 01/26/2024 18:41:05

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:30	
Control Test	0.078	18:30	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:34	No Sample Provided
Breath Volume	0.000	18:34	
Air Blank	0.000	18:34	
Air Blank	0.000	18:36	
Subject Sample	# 2	18:39	No Sample Provided
Breath Volume	0.000	18:39	
Air Blank	0.000	18:40	
Control Test	0.078	18:40	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 01/30/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OTILUS,LEBEDOPH,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: OTILUS,LEBEDOPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1141545
DRY GAS EXPIRY DATE: 04/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FORMIGA VERDAGUER

DATE/TIME OF TEST: 01/26/2024 18:57:19

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.079	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.277	18:52
Breath Volume	1.992	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.269	18:55
Breath Volume	2.050	18:55
Air Blank	0.000	18:56
Control Test	0.076	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 01/30/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 12/22/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 01/01/2024 03:42:53

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.092	03:38
Breath Volume	2.265	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.091	03:41
Breath Volume	2.242	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/29/2024 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/19/2024 00:21:01

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 01/25/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LOAISIGAOBANDO

DATE/TIME OF TEST: 01/21/2024 00:26:06

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.171	00:21
Breath Volume	1.300	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.168	00:24
Breath Volume	1.921	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/25/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEDROZAMARTINEZ

DATE/TIME OF TEST: 01/27/2024 02:03:44

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1 0.143	02:00	
Breath Volume	3.007	02:00	
Air Blank	0.000	02:00	
Air Blank		02:02	RFI Detect
Air Blank	0.000	02:02	
Control Test	0.081	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 02/29/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEDROSAMARTINEZ

DATE/TIME OF TEST: 01/27/2024 02:12:41

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.133	02:08
Breath Volume	1.335	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.120	02:11
Breath Volume	1.140	02:11
Air Blank	0.000	02:11
Control Test	0.079	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 02/29/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MOUMNE

DATE/TIME OF TEST: 01/06/2024 02:06:25

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.082	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.135	02:01
Breath Volume	1.992	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.135	02:04
Breath Volume	1.730	02:04
Air Blank	0.000	02:05
Control Test	0.083	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BARD

DATE/TIME OF TEST: 01/06/2024 02:56:19

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.100	02:51
Breath Volume	2.105	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.100	02:54
Breath Volume	1.933	02:54
Air Blank	0.000	02:55
Control Test	0.083	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MILEA

DATE/TIME OF TEST: 01/06/2024 04:10:41

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.019	04:06
Breath Volume	2.695	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.018	04:09
Breath Volume	2.695	04:09
Air Blank	0.000	04:09
Control Test	0.083	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: RODRIGUEZ MOTA FERNA

DATE/TIME OF TEST: 01/07/2024 05:42:44

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.082	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.097	05:37
Breath Volume	2.316	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.096	05:41
Breath Volume	2.527	05:41
Air Blank	0.000	05:41
Control Test	0.081	05:42
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/07/2024 21:32:49

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.082	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.168	21:28
Breath Volume	1.570	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:30
Subject Sample # 2	0.148	21:31
Breath Volume	1.304	21:31
Air Blank	0.000	21:31
Control Test	0.082	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 01/08/2024 04:16:37

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.084	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.138	04:12
Breath Volume	2.941	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.135	04:15
Breath Volume	2.593	04:15
Air Blank	0.000	04:15
Control Test	0.083	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 02/02/2024 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: WESTLEY JR

DATE/TIME OF TEST: 01/09/2024 05:47:51

Diagnostics Check	OK	05:34	
Air Blank	0.000	05:35	
Control Test	0.083	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:39	No Sample Provided
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Air Blank	0.000	05:41	
Subject Sample	# 2	0.154 05:42	
Breath Volume	1.328	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:44	
Subject Sample	# 3	0.168 05:46	
Breath Volume	2.011	05:46	
Air Blank	0.000	05:46	
Control Test	0.082	05:47	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:47	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ZELLEM

DATE/TIME OF TEST: 01/13/2024 01:28:08

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.082	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:23
Breath Volume	2.550	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:26
Breath Volume	2.023	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 01/13/2024 02:42:17

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.082	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.138	02:37
Breath Volume	3.019	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.138	02:40
Breath Volume	1.824	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: FIGUEREDO

DATE/TIME OF TEST: 01/13/2024 03:18:04

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.123	03:13
Breath Volume	2.656	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.130	03:16
Breath Volume	3.109	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PEREZ-PAZ

DATE/TIME OF TEST: 01/14/2024 01:25:33

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.082	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.163	01:20
Breath Volume	1.203	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.166	01:23
Breath Volume	1.523	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SILVA COBAS

DATE/TIME OF TEST: 01/14/2024 01:41:44

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.082	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.116	01:37
Breath Volume	2.695	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.112	01:40
Breath Volume	2.003	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 01/14/2024 03:09:22

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.083	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1 0.174	02:59	Volume Not Met
Breath Volume	0.238	02:59	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.153	03:04	No .020 agreement
Breath Volume	1.941	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 3 0.183	03:07	No .020 agreement
Breath Volume	2.585	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ROSALES,F
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZ LOPEZ

DATE/TIME OF TEST: 01/14/2024 03:17:38

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.178	03:13
Breath Volume	2.390	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.172	03:16
Breath Volume	2.246	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: DAVIS III

DATE/TIME OF TEST: 01/14/2024 03:58:44

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.081	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:57	No Sample Provided
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.082	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: CANDELARIA

DATE/TIME OF TEST: 01/14/2024 07:57:27

Diagnostics Check	OK	07:50
Air Blank	0.000	07:51
Control Test	0.083	07:51
Air Blank	0.000	07:52
Subject Sample # 1	0.130	07:52
Breath Volume	3.253	07:52
Air Blank	0.000	07:53
Air Blank	0.000	07:55
Subject Sample # 2	0.130	07:55
Breath Volume	2.871	07:55
Air Blank	0.000	07:56
Control Test	0.082	07:56
Air Blank	0.000	07:57
Diagnostics Check	OK	07:57

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 01/16/2024 01:43:11

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.082	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.154	01:38
Breath Volume	1.347	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.162	01:41
Breath Volume	1.550	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 01/20/2024 00:54:44

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.082	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.061	00:50
Breath Volume	2.003	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.057	00:53
Breath Volume	1.296	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: VIERA

DATE/TIME OF TEST: 01/20/2024 04:09:30

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.082	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.046	04:05
Breath Volume	2.378	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.046	04:07
Breath Volume	2.515	04:07
Air Blank	0.000	04:08
Control Test	0.082	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LOPEZ MENDEZ

DATE/TIME OF TEST: 01/20/2024 04:32:41

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.082	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.116	04:25	Volume Not Met
Breath Volume	0.183	04:25	
Air Blank	0.000	04:26	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: LOPEZ MENDEZ

DATE/TIME OF TEST: 01/20/2024 04:58:35

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.082	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.138	04:48	
Breath Volume	1.476	04:48	
Air Blank	0.000	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 2 0.128	04:53	Volume Not Met
Breath Volume	0.410	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 3 0.139	04:56	
Breath Volume	1.183	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FIEBIG

DATE/TIME OF TEST: 01/21/2024 03:50:26

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.083	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.255	03:45
Breath Volume	1.460	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.253	03:48
Breath Volume	1.488	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/21/2024 04:16:10

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.083	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LYLES

DATE/TIME OF TEST: 01/22/2024 22:16:04

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.083	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.216	22:10
Breath Volume	1.375	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.220	22:14
Breath Volume	1.835	22:14
Air Blank	0.000	22:15
Control Test	0.082	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: RAPOSO

DATE/TIME OF TEST: 01/22/2024 22:39:48

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.082	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.031	22:35
Breath Volume	1.363	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.030	22:38
Breath Volume	1.253	22:38
Air Blank	0.000	22:38
Control Test	0.082	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/25/2024 04:26:24

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.211	04:21
Breath Volume	1.457	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.218	04:24
Breath Volume	1.898	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FUNES RODRIGUEZ

DATE/TIME OF TEST: 01/26/2024 02:06:12

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.082	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.115	02:01
Breath Volume	2.121	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.114	02:04
Breath Volume	1.832	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MERCURIO

DATE/TIME OF TEST: 01/26/2024 03:40:54

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MCGUIRE V

DATE/TIME OF TEST: 01/27/2024 02:22:58

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.100	02:18
Breath Volume	1.109	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.100	02:21
Breath Volume	1.316	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DOLIN

DATE/TIME OF TEST: 01/28/2024 00:06:47

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.082	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.151	00:02
Breath Volume	2.097	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.139	00:05
Breath Volume	1.503	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/28/2024 02:18:24

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.141	02:13
Breath Volume	3.011	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.141	02:16
Breath Volume	3.253	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LOPEZ BELLO

DATE/TIME OF TEST: 01/28/2024 04:33:44

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Control Test	0.081	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FERNANDEZ TENESACA

DATE/TIME OF TEST: 01/28/2024 21:23:22

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.082	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.195	21:18
Breath Volume	1.582	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.204	21:21
Breath Volume	1.648	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/29/2024 03:46:10

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.181	03:41
Breath Volume	2.496	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:43
Subject Sample # 2	0.173	03:44
Breath Volume	2.460	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 02/02/2024 08:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: VASQUEZ SANCHEZ

DATE/TIME OF TEST: 01/29/2024 04:26:46

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2	0.214	04:19
Breath Volume	1.699	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:21	
Subject Sample	# 3	0.177	04:25
Breath Volume	0.519	04:25	Volume Not Met
Air Blank	0.000	04:25	
Control Test	0.081	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 02/02/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUARTE,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 01/29/2024 14:56:18

Diagnostics Check	OK	14:41	
Air Blank	0.000	14:41	
Control Test	0.083	14:41	
Air Blank	0.000	14:42	
Subject Sample	# 1 0.238	14:45	
Breath Volume	2.074	14:45	
Air Blank	0.000	14:46	
Air Blank	0.000	14:47	
Subject Sample	# 2	14:51	No Sample Provided
Breath Volume	0.000	14:51	
Air Blank	0.000	14:51	
Air Blank	0.000	14:53	
Subject Sample	# 3	14:54	Subject Test Refused
Breath Volume	0.000	14:54	
Air Blank	0.000	14:55	
Control Test	0.082	14:55	
Air Blank	0.000	14:56	
Diagnostics Check	OK	14:56	

Data Download Date/Time: 02/02/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 01/30/2024 20:41:51

Diagnostics Check	OK	20:33
Air Blank	0.000	20:34
Control Test	0.083	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.271	20:36
Breath Volume	1.472	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.271	20:40
Breath Volume	1.726	20:40
Air Blank	0.000	20:40
Control Test	0.081	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 02/02/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 01/01/2024 01:13:29

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.083	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.102	01:08
Breath Volume	2.460	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.102	01:11
Breath Volume	2.214	01:11
Air Blank	0.000	01:12
Control Test	0.083	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SAMALA

DATE/TIME OF TEST: 01/01/2024 02:18:23

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.084	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.127	02:14
Breath Volume	1.226	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.132	02:16
Breath Volume	1.500	02:16
Air Blank	0.000	02:17
Control Test	0.083	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/01/2024 03:50:00

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.084	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.112	03:45
Breath Volume	2.210	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.109	03:48
Breath Volume	2.359	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:50

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BERG,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: HAYES JR

DATE/TIME OF TEST: 01/01/2024 04:44:58

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.083	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.165	04:40
Breath Volume	2.621	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.168	04:43
Breath Volume	2.621	04:43
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/01/2024 05:02:45

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.083	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.165	04:58
Breath Volume	1.734	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.178	05:01
Breath Volume	1.566	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 01/01/2024 05:37:53

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.083	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.108	05:33
Breath Volume	1.390	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.110	05:36
Breath Volume	2.492	05:36
Air Blank	0.000	05:36
Control Test	0.083	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,KATIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: POWE

DATE/TIME OF TEST: 01/01/2024 20:44:44

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.083	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.153	20:40
Breath Volume	1.949	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.148	20:43
Breath Volume	1.796	20:43
Air Blank	0.000	20:43
Control Test	0.082	20:44
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: TODOROVICH

DATE/TIME OF TEST: 01/02/2024 02:48:39

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.082	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.082	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/01/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DINARDO II

DATE/TIME OF TEST: 01/02/2024 03:53:52

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.083	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Control Test	0.083	03:53	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 01/03/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKLEY,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: LOCKLEY,GABRIEL
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/27/2024 20:51:56

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
		20:51	RFI Detect
Air Blank		20:51	RFI Detect

Data Download Date/Time: 02/19/2024 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,CHARMAYNE,
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: LOCKLEY,GABRIEL
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/27/2024 20:57:48

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:57	
Control Test	0.082	20:57	
Air Blank		20:57	RFI Detect
Air Blank		20:57	RFI Detect

Data Download Date/Time: 02/19/2024 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKLEY,GABRIEL,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: LOCKLEY,GABRIEL
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/27/2024 21:13:44

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.082	21:03	
Air Blank	0.000	21:04	
Subject Sample	# 1 0.207	21:07	Volume Not Met
Breath Volume	0.210	21:07	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2 0.195	21:12	Volume Not Met
Breath Volume	0.171	21:12	
Air Blank		21:13	Improper Sample
Air Blank	0.000	21:13	

Data Download Date/Time: 02/19/2024 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,CHARMAYNE,
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: LOCKLEY,GABRIEL
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/27/2024 21:28:34

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.079	21:18	
Air Blank	0.000	21:18	
Subject Sample	# 1 0.193	21:19	
Breath Volume	1.265	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2 0.188	21:25	Volume Not Met
Breath Volume	0.019	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 3	21:28	Sequence Aborted
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	

Data Download Date/Time: 02/19/2024 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CULVERHOUSE

DATE/TIME OF TEST: 01/28/2024 03:35:29

Diagnostics Check	OK	03:34	
Air Blank		03:35	RFI Detect
Air Blank	0.000	03:35	

Data Download Date/Time: 01/28/2024 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/02/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CULVERHOUSE

DATE/TIME OF TEST: 01/28/2024 03:43:56

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.205	03:39
Breath Volume	2.671	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.214	03:42
Breath Volume	2.890	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/28/2024 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANDT,FABIAN,E
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: RISCO,DARIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BERNARDO

DATE/TIME OF TEST: 01/03/2024 21:27:45

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:24	
Air Blank	0.000	21:25	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.081	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FERRO,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/06/2024 11:15:29

Diagnostics Check	OK	11:02	
Air Blank	0.000	11:03	
Control Test	0.081	11:03	
Air Blank	0.000	11:03	
Subject Sample	# 1 0.203	11:07	Slope Not Level
Breath Volume	1.625	11:07	
Air Blank	0.000	11:07	
Air Blank	0.000	11:09	
Subject Sample	# 2 0.210	11:10	
Breath Volume	1.710	11:10	
Air Blank	0.000	11:11	
Air Blank	0.000	11:13	
Subject Sample	# 3 0.213	11:13	
Breath Volume	1.714	11:13	
Air Blank	0.000	11:14	
Control Test	0.079	11:14	
Air Blank	0.000	11:15	
Diagnostics Check	OK	11:15	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,JACOB,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: POST,JACOB
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 01/09/2024 01:19:45

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.076	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.078	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 01/13/2024 07:45:45

Diagnostics Check	OK	07:39
Air Blank	0.000	07:39
Control Test	0.080	07:40
Air Blank	0.000	07:40
Subject Sample # 1	0.196	07:41
Breath Volume	1.804	07:41
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.191	07:44
Breath Volume	2.136	07:44
Air Blank	0.000	07:44
Control Test	0.079	07:45
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ BERITAN, LUIS, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LUNA, ANGEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: CAVINESS

DATE/TIME OF TEST: 01/13/2024 20:29:22

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.081	20:24	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,LUIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: LOPEZ,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 01/14/2024 13:54:12

Diagnostics Check	OK	13:47
Air Blank	0.000	13:47
Control Test	0.081	13:48
Air Blank	0.000	13:48
Subject Sample # 1	0.170	13:49
Breath Volume	2.335	13:49
Air Blank	0.000	13:50
Air Blank	0.000	13:52
Subject Sample # 2	0.172	13:52
Breath Volume	3.058	13:52
Air Blank	0.000	13:53
Control Test	0.080	13:53
Air Blank	0.000	13:54
Diagnostics Check	OK	13:54

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIDRON,JULIO,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CIDRON,JULIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TESILLOS

DATE/TIME OF TEST: 01/15/2024 00:27:23

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1 0.159	00:20	Volume Not Met
Breath Volume	0.246	00:20	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 2 0.183	00:25	Volume Not Met
Breath Volume	0.218	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TESILLOS, TEODORO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CIDRON, JULIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: TESILLOS

DATE/TIME OF TEST: 01/15/2024 00:37:13

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.178	00:32
Breath Volume	1.976	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.170	00:35
Breath Volume	2.164	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEMORIN,DEVITCH,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PORTUGUES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: NEMORIN

DATE/TIME OF TEST: 01/19/2024 04:32:56

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSALES,ANIBAL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROSALES,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SILVA MACIAS

DATE/TIME OF TEST: 01/21/2024 04:13:56

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.155	04:09
Breath Volume	1.710	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.156	04:12
Breath Volume	2.132	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WROY,WALNER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: DEAVILA

DATE/TIME OF TEST: 01/24/2024 16:46:45

Diagnostics Check	OK	16:45	
Air Blank	0.000	16:46	
		16:46	RFI Detect
Air Blank	0.000	16:46	

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WROY,WALNER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: DEAVILA

DATE/TIME OF TEST: 01/24/2024 16:54:46

Diagnostics Check	OK	16:48
Air Blank	0.000	16:49
Control Test	0.080	16:49
Air Blank	0.000	16:50
Subject Sample # 1	0.085	16:50
Breath Volume	2.285	16:50
Air Blank	0.000	16:51
Air Blank	0.000	16:52
Subject Sample # 2	0.079	16:53
Breath Volume	1.257	16:53
Air Blank	0.000	16:53
Control Test	0.081	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: OCHOA,GUILLERMO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/27/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1356262
DRY GAS EXPIRY DATE: 08/03/2023
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: MONEGRO

DATE/TIME OF TEST: 01/25/2024 09:41:40

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.213	09:37
Breath Volume	1.746	09:37
Air Blank	0.000	09:37
Air Blank	0.000	09:39
Subject Sample # 2	0.200	09:40
Breath Volume	2.035	09:40
Air Blank	0.000	09:40
Control Test	0.082	09:41
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 01/25/2024 13:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/30/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433249
DRY GAS EXPIRY DATE: 01/08/2024
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/14/2024 06:07:17

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.046	06:02
Breath Volume	1.691	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.043	06:05
Breath Volume	1.589	06:05
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 01/31/2024 23:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TOSCANO LUDENA

DATE/TIME OF TEST: 01/04/2024 22:16:57

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.176	22:12
Breath Volume	3.003	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.172	22:15
Breath Volume	1.984	22:15
Air Blank	0.000	22:16
Control Test	0.082	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 02/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SLINGERLAMD

DATE/TIME OF TEST: 01/05/2024 03:02:27

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.098	02:57
Breath Volume	2.761	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.098	03:00
Breath Volume	2.800	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 02/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,EDSON,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CHAVEZ,EDSON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: FLORES CUSTODIO

DATE/TIME OF TEST: 01/14/2024 23:22:55

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.214	23:18
Breath Volume	2.089	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.208	23:21
Breath Volume	1.617	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 02/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/27/2024 04:15:22

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.151	04:10
Breath Volume	1.484	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.159	04:13
Breath Volume	1.789	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 02/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BURGOS MONTANEZ

DATE/TIME OF TEST: 01/28/2024 01:50:32

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.082	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.090	01:43	Volume Not Met
Breath Volume	0.710	01:43	
Air Blank	0.000	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 2 0.110	01:49	Volume Not Met
Breath Volume	0.070	01:49	
Air Blank	0.000	01:49	
Control Test	0.082	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 02/23/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLIPIN,AARON,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: CHAPPEL,O.
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 01/02/2024 22:35:08

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.082	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.000	22:30
Breath Volume	1.144	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.000	22:33
Breath Volume	2.281	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 01/29/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: RUIZ LEDEZMA

DATE/TIME OF TEST: 01/25/2024 16:19:49

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:14
Subject Sample # 1	0.060	16:14
Breath Volume	1.265	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.063	16:18
Breath Volume	1.253	16:18
Air Blank	0.000	16:18
Control Test	0.080	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 01/29/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 01/30/2024 16:38:56

Diagnostics Check	OK	16:32
Air Blank	0.000	16:32
Control Test	0.082	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.217	16:34
Breath Volume	2.003	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.211	16:37
Breath Volume	1.503	16:37
Air Blank	0.000	16:38
Control Test	0.082	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 02/26/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BURGER, WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 01/07/2024 05:54:30

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.079	05:45	
Air Blank	0.000	05:45	
Subject Sample	# 1	05:48	No Sample Provided
Breath Volume	0.000	05:48	
Air Blank	0.000	05:49	
Air Blank	0.000	05:51	
Subject Sample	# 2	05:53	Subject Test Refused
Breath Volume	0.000	05:53	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 01/27/2024 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: MARTIN, ANDREW
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/24/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: NAIDOO

DATE/TIME OF TEST: 01/27/2024 00:47:04

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.139	00:42
Breath Volume	1.906	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.136	00:45
Breath Volume	1.550	00:45
Air Blank	0.000	00:46
Control Test	0.077	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:47

Data Download Date/Time: 01/27/2024 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MINNIFIELD,SAM,M

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: MINNIFIELD,SAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: DUI

ARREST TIME: 01:58

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 01/01/2024 02:42:38

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.084	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.113	02:37
Breath Volume	1.277	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.114	02:41
Breath Volume	1.363	02:41
Air Blank	0.000	02:41
Control Test	0.082	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/01/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: LEWIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 01/14/2024 19:56:02

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.084	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1 0.063	19:47	Volume Not Met
Breath Volume	0.562	19:47	
Air Blank	0.000	19:48	
Air Blank	0.000	19:50	
Subject Sample	# 2 0.063	19:51	
Breath Volume	1.925	19:51	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample	# 3 0.061	19:54	
Breath Volume	1.566	19:54	
Air Blank	0.000	19:55	
Control Test	0.083	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 04/01/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MINNIFIELD,SAM,M

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: MINNIFIELD,SAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: DUI

ARREST TIME: 00:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: BANEY

DATE/TIME OF TEST: 01/21/2024 01:33:16

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.084	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.084	01:28
Breath Volume	2.375	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.088	01:31
Breath Volume	2.746	01:31
Air Blank	0.000	01:32
Control Test	0.085	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/01/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRY,NICHOLAS
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/11/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/29/2024 02:50:35

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.084	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.083	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 04/01/2024 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: RODRIGUEZ SOSA

DATE/TIME OF TEST: 01/07/2024 01:55:15

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.135	01:50
Breath Volume	1.238	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.147	01:53
Breath Volume	1.269	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: VELAZQUEZ GARCIA

DATE/TIME OF TEST: 01/07/2024 02:30:26

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.081	02:26
Breath Volume	2.027	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.076	02:28
Breath Volume	1.214	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RANDALL JR

DATE/TIME OF TEST: 01/12/2024 02:39:31

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.161	02:35
Breath Volume	1.648	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.166	02:38
Breath Volume	2.449	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 01/13/2024 02:21:57

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:17
Breath Volume	2.316	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	1.812	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JIMENEZ REYNALDO

DATE/TIME OF TEST: 01/13/2024 03:23:25

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:19
Breath Volume	2.554	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.000	03:22
Breath Volume	2.281	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 01/13/2024 05:01:43

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.223	04:57
Breath Volume	2.433	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.205	05:00
Breath Volume	2.015	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: CARIAS PENATE

DATE/TIME OF TEST: 01/21/2024 04:28:19

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.082	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.114	04:22
Breath Volume	1.777	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.107	04:26
Breath Volume	1.515	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 01/22/2024 01:59:06

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.082	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.165	01:54
Breath Volume	3.218	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.163	01:57
Breath Volume	3.265	01:57
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 01/26/2024 04:04:20

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.135	03:59
Breath Volume	1.675	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.144	04:02
Breath Volume	1.550	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MOTA BALTAZAR

DATE/TIME OF TEST: 01/28/2024 00:24:10

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.169	00:19
Breath Volume	1.769	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.166	00:22
Breath Volume	1.859	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 02/02/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GONZALEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/01/2024 07:14:08

Diagnostics Check	OK	07:07
Air Blank	0.000	07:07
Control Test	0.083	07:08
Air Blank	0.000	07:08
Subject Sample # 1	0.142	07:09
Breath Volume	1.625	07:09
Air Blank	0.000	07:10
Air Blank	0.000	07:11
Subject Sample # 2	0.141	07:12
Breath Volume	1.808	07:12
Air Blank	0.000	07:13
Control Test	0.083	07:13
Air Blank	0.000	07:14
Diagnostics Check	OK	07:14

Data Download Date/Time: 01/17/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUIRGUIS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:01

SUBJECT'S LAST NAME: BROGAN

DATE/TIME OF TEST: 01/06/2024 12:28:04

Diagnostics Check	OK	12:21
Air Blank	0.000	12:22
Control Test	0.082	12:22
Air Blank	0.000	12:22
Subject Sample # 1	0.000	12:23
Breath Volume	1.621	12:23
Air Blank	0.000	12:24
Air Blank	0.000	12:26
Subject Sample # 2	0.000	12:26
Breath Volume	1.824	12:26
Air Blank	0.000	12:27
Control Test	0.082	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:28

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GWYNNE,KARL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/06/2024 15:38:45

Diagnostics Check	OK	15:35	
Air Blank	0.000	15:36	
Control Test	0.083	15:36	
Air Blank	0.000	15:37	
Subject Sample	# 1	15:37	Subject Test Refused
Breath Volume	0.000	15:37	
Air Blank	0.000	15:37	
Control Test	0.082	15:38	
Air Blank	0.000	15:38	
Diagnostics Check	OK	15:38	

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,B,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZIMMERMAN,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: MUSANTRY

DATE/TIME OF TEST: 01/07/2024 17:43:50

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:35	
Control Test	0.083	17:36	
Air Blank	0.000	17:36	
Subject Sample	# 1	17:39	No Sample Provided
Breath Volume	0.000	17:39	
Air Blank	0.000	17:40	
Air Blank	0.000	17:42	
Subject Sample	# 2	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:42	
Control Test	0.083	17:43	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENTLEY, TYLER, T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNELL, CHASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: BACKERT 3

DATE/TIME OF TEST: 01/08/2024 08:49:52

Diagnostics Check	OK	08:42
Air Blank	0.000	08:43
Control Test	0.084	08:43
Air Blank	0.000	08:44
Subject Sample # 1	0.173	08:45
Breath Volume	3.437	08:45
Air Blank	0.000	08:45
Air Blank	0.000	08:47
Subject Sample # 2	0.172	08:48
Breath Volume	4.023	08:48
Air Blank	0.000	08:48
Control Test	0.082	08:49
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,ANDREW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 01/08/2024 21:11:22

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.084	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.000	21:07
Breath Volume	2.835	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.000	21:10
Breath Volume	1.550	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORGAN,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: DIFAZIO

DATE/TIME OF TEST: 01/09/2024 15:45:32

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:43	
Control Test	0.083	15:43	
Air Blank	0.000	15:43	
Subject Sample	# 1	15:44	Subject Test Refused
Breath Volume	0.000	15:44	
Air Blank	0.000	15:44	
Control Test	0.082	15:44	
Air Blank	0.000	15:45	
Diagnostics Check	OK	15:45	

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DIEBEL,JANAE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/09/2024 17:43:03

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.083	17:37
Air Blank	0.000	17:38
Subject Sample # 1	0.326	17:38
Breath Volume	1.605	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:41
Subject Sample # 2	0.320	17:41
Breath Volume	1.710	17:41
Air Blank	0.000	17:42
Control Test	0.081	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:43

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: RAHMAN

DATE/TIME OF TEST: 01/13/2024 01:49:09

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.083	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.048	01:44
Breath Volume	1.179	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.048	01:47
Breath Volume	1.312	01:47
Air Blank	0.000	01:48
Control Test	0.081	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARQUEZ-CARRILLO, JOSE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 07:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: TREWHELLA

DATE/TIME OF TEST: 01/13/2024 08:55:54

Diagnostics Check	OK	08:52	
Air Blank	0.000	08:53	
Control Test	0.083	08:53	
Air Blank	0.000	08:54	
Subject Sample	# 1	08:54	Subject Test Refused
Breath Volume	0.000	08:54	
Air Blank	0.000	08:54	
Control Test	0.083	08:55	
Air Blank	0.000	08:55	
Diagnostics Check	OK	08:55	

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER, LOREN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: HEMPFLING

DATE/TIME OF TEST: 01/14/2024 07:17:57

Diagnostics Check	OK	07:11
Air Blank	0.000	07:12
Control Test	0.083	07:12
Air Blank	0.000	07:13
Subject Sample # 1	0.113	07:13
Breath Volume	4.019	07:13
Air Blank	0.000	07:14
Air Blank	0.000	07:15
Subject Sample # 2	0.114	07:16
Breath Volume	3.281	07:16
Air Blank	0.000	07:17
Control Test	0.083	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/20/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/14/2024 21:23:39

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.083	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.083	21:23	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 01/17/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 01/20/2024 21:12:39

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:10	
Control Test	0.085	21:10	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:11	Subject Test Refused
Breath Volume	0.000	21:11	
Air Blank	0.000	21:11	
Control Test	0.083	21:12	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 02/21/2024 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBEL,Z,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GAMBEL,Z
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ZINNO

DATE/TIME OF TEST: 01/21/2024 01:10:14

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.084	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.084	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 02/21/2024 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 01/25/2024 01:29:25

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:24
Breath Volume	1.765	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:28
Breath Volume	2.007	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 02/21/2024 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG229803
DRY GAS EXPIRY DATE: 10/25/2024
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: FEASTER III

DATE/TIME OF TEST: 01/25/2024 03:52:02

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.083	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.139	03:47
Breath Volume	1.378	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.139	03:50
Breath Volume	1.105	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/21/2024 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:49

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/26/2024 11:18:10

Diagnostics Check	OK	11:12
Air Blank	0.000	11:12
Control Test	0.080	11:13
Air Blank	0.000	11:13
Subject Sample # 1	0.000	11:14
Breath Volume	1.683	11:14
Air Blank	0.000	11:14
Air Blank	0.000	11:16
Subject Sample # 2	0.000	11:16
Breath Volume	2.117	11:16
Air Blank	0.000	11:17
Control Test	0.078	11:17
Air Blank	0.000	11:18
Diagnostics Check	OK	11:18

Data Download Date/Time: 02/14/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 01/12/2024 01:36:23

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: COX,C
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GIRON MONTENEGRO

DATE/TIME OF TEST: 01/13/2024 03:45:29

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.176	03:40
Breath Volume	1.492	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.178	03:43
Breath Volume	1.722	03:43
Air Blank	0.000	03:44
Control Test	0.077	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RIOS,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RIOS MATUS

DATE/TIME OF TEST: 01/13/2024 04:13:40

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.077	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.180	04:08
Breath Volume	1.792	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.167	04:12
Breath Volume	1.800	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LORFILS

DATE/TIME OF TEST: 01/21/2024 02:02:28

Diagnostics Check	Fail	02:01	Temperature Regulation Test
Air Blank	0.000	02:02	

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: LORFILS

DATE/TIME OF TEST: 01/21/2024 02:11:25

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.158	02:06
Breath Volume	1.332	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.153	02:09
Breath Volume	1.406	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IRIAS MONCADA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PENA LARA

DATE/TIME OF TEST: 01/27/2024 01:26:54

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
		01:26	RFI Detect
Air Blank		01:26	RFI Detect

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IRIAS MONCADA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PENA LARA

DATE/TIME OF TEST: 01/27/2024 01:30:08

Diagnostics Check	OK	01:29	
Air Blank		01:29	RFI Detect
Air Blank		01:30	RFI Detect

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: IRIAS MONCADA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PENA LARA

DATE/TIME OF TEST: 01/27/2024 01:41:10

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.248	01:36
Breath Volume	2.148	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.253	01:39
Breath Volume	1.550	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,F,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: OBANDO SOLAND

DATE/TIME OF TEST: 01/28/2024 08:13:22

Diagnostics Check	OK	08:05
Air Blank	0.000	08:06
Control Test	0.079	08:06
Air Blank	0.000	08:07
Subject Sample # 1	0.169	08:08
Breath Volume	1.539	08:08
Air Blank	0.000	08:08
Air Blank	0.000	08:10
Subject Sample # 2	0.172	08:11
Breath Volume	1.414	08:11
Air Blank	0.000	08:12
Control Test	0.077	08:12
Air Blank	0.000	08:13
Diagnostics Check	OK	08:13

Data Download Date/Time: 02/01/2024 04:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MENDEZ-SIMON

DATE/TIME OF TEST: 01/23/2024 23:31:09

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:25
Subject Sample # 1	0.164	23:26
Breath Volume	2.625	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.161	23:29
Breath Volume	2.613	23:29
Air Blank	0.000	23:30
Control Test	0.077	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 02/01/2024 05:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CARDONA-MORALES

DATE/TIME OF TEST: 01/27/2024 03:46:18

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.203	03:39
Breath Volume	1.648	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.206	03:44
Breath Volume	1.371	03:44
Air Blank	0.000	03:45
Control Test	0.077	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 02/01/2024 05:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 01/28/2024 04:02:18

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.192	03:57
Breath Volume	2.066	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.192	04:00
Breath Volume	2.253	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 02/01/2024 05:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 01/14/2024 02:40:37

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	
Control Test	0.079	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 02/01/2024 04:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 01/21/2024 03:12:05

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.150	03:07
Breath Volume	2.812	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.158	03:10
Breath Volume	2.707	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:12

Data Download Date/Time: 02/01/2024 04:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/27/2024 02:22:58

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.187	02:18
Breath Volume	2.402	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.189	02:21
Breath Volume	1.363	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 02/01/2024 04:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CRUZ-ORTIZ

DATE/TIME OF TEST: 01/28/2024 02:32:28

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 02/01/2024 04:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: CABAN COLON

DATE/TIME OF TEST: 01/03/2024 19:58:20

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:51	
Control Test	0.081	19:52	
Air Blank	0.000	19:52	
Subject Sample	# 1 0.000	19:54	
Breath Volume	2.531	19:54	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample	# 2	19:56	RFI Detect
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Control Test	0.081	19:57	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

Data Download Date/Time: 02/23/2024 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: CABAN COLON

DATE/TIME OF TEST: 01/03/2024 20:07:30

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.000	20:03
Breath Volume	1.988	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.000	20:06
Breath Volume	2.406	20:06
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 02/23/2024 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GRIMALDI,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 01/09/2024 23:29:49

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.081	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Sequence Aborted
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	

Data Download Date/Time: 02/23/2024 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GRIMALDI,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 01/10/2024 00:48:44

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.080	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.156	00:43
Breath Volume	2.722	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.150	00:46
Breath Volume	2.675	00:46
Air Blank	0.000	00:47
Control Test	0.077	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: MARCO

DATE/TIME OF TEST: 01/10/2024 20:48:14

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.081	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.195	20:43
Breath Volume	2.156	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.187	20:46
Breath Volume	2.187	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PULGAR ARAUJO

DATE/TIME OF TEST: 01/11/2024 03:59:05

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.080	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JOHNSON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CORRAO

DATE/TIME OF TEST: 01/11/2024 23:53:39

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.157	23:49	
Breath Volume	1.148	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 2	23:52	RFI Detect
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.079	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: JOHNSON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: CORRAO

DATE/TIME OF TEST: 01/12/2024 00:02:30

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.162	23:57
Breath Volume	1.144	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.182	00:00
Breath Volume	1.398	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,CHRISTIAN,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAREY,VANESSA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:04

SUBJECT'S LAST NAME: FONTANEZ

DATE/TIME OF TEST: 01/13/2024 09:43:52

Diagnostics Check	OK	09:36
Air Blank	0.000	09:36
Control Test	0.081	09:37
Air Blank	0.000	09:37
Subject Sample # 1	0.248	09:38
Breath Volume	1.332	09:38
Air Blank	0.000	09:39
Air Blank	0.000	09:41
Subject Sample # 2	0.232	09:42
Breath Volume	1.292	09:42
Air Blank	0.000	09:42
Control Test	0.077	09:43
Air Blank	0.000	09:43
Diagnostics Check	OK	09:43

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TURNER,RODNEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: POSADA GUAPACHA

DATE/TIME OF TEST: 01/14/2024 04:16:43

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1 0.132	04:10	
Breath Volume	1.507	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:15	Subject Test Refused
Breath Volume	0.937	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILAN,VICTOR,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MILAN,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: BREA SALOMON

DATE/TIME OF TEST: 01/20/2024 09:57:49

Diagnostics Check	OK	09:49
Air Blank	0.000	09:49
Control Test	0.082	09:49
Air Blank	0.000	09:50
Subject Sample # 1	0.163	09:52
Breath Volume	2.070	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:54
Subject Sample # 2	0.161	09:56
Breath Volume	2.500	09:56
Air Blank	0.000	09:56
Control Test	0.080	09:57
Air Blank	0.000	09:57
Diagnostics Check	OK	09:57

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: LOPEZ-TZINA

DATE/TIME OF TEST: 01/20/2024 17:47:30

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.081	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.219	17:42
Breath Volume	1.933	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.219	17:45
Breath Volume	2.296	17:45
Air Blank	0.000	17:46
Control Test	0.081	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCANLAN,JAMES,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: VAZQUEZ-RODRIGUEZ

DATE/TIME OF TEST: 01/21/2024 01:13:47

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.031	01:09
Breath Volume	1.722	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.025	01:12
Breath Volume	2.265	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GARCIA, CORINA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MARTINEZ DIAZ

DATE/TIME OF TEST: 01/21/2024 04:52:20

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.025	04:47
Breath Volume	4.203	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.021	04:50
Breath Volume	3.453	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/26/2024 05:05:51

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.161	05:01
Breath Volume	3.003	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.161	05:04
Breath Volume	2.882	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HELLER,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BERGEY

DATE/TIME OF TEST: 01/26/2024 19:59:02

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.081	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.000	19:54
Breath Volume	3.906	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:57
Subject Sample # 2	0.000	19:57
Breath Volume	3.261	19:57
Air Blank	0.000	19:58
Control Test	0.082	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:59

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HARRIS,CASEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: PEREZ BALLESTER

DATE/TIME OF TEST: 01/27/2024 03:03:16

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.211	02:56
Breath Volume	1.335	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.210	03:01
Breath Volume	1.507	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORR,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SILVA CARIDAD

DATE/TIME OF TEST: 01/28/2024 00:51:57

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:47
Breath Volume	2.808	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	3.074	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NIEVES,JAYLYNN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ASFOUR

DATE/TIME OF TEST: 01/28/2024 23:58:36

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.106	23:51	Volume Not Met
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 2 0.128	23:57	Volume Not Met
Breath Volume	0.015	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NIEVES,JAYLYNN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ASFOUR

DATE/TIME OF TEST: 01/29/2024 00:07:19

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:02	
Control Test	0.080	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.085	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 02/23/2024 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON, RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1472328
DRY GAS EXPIRY DATE: 04/02/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: LILMAN

DATE/TIME OF TEST: 01/29/2024 01:04:27

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.191	00:59
Breath Volume	1.402	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.178	01:02
Breath Volume	1.285	01:02
Air Blank	0.000	01:03
Control Test	0.077	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 02/23/2024 11:17