

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KNOOP,K
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: DELLES

DATE/TIME OF TEST: 01/01/2024 02:07:16

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.124	02:02
Breath Volume	3.195	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.120	02:05
Breath Volume	3.359	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FARFAN MORENO

DATE/TIME OF TEST: 01/02/2024 03:27:14

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Subject Sample # 1	0.099	03:23	
Breath Volume	3.523	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample # 2	0.097	03:26	
Breath Volume	4.832	03:26	
Air Blank		03:26	Improper Sample
Air Blank	0.000	03:27	

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FARFAN MORENO

DATE/TIME OF TEST: 01/02/2024 03:37:04

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.098	03:32
Breath Volume	4.394	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.092	03:35
Breath Volume	3.101	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LOPEZ-LOPEZ

DATE/TIME OF TEST: 01/03/2024 22:52:01

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.246	22:46
Breath Volume	1.273	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.242	22:50
Breath Volume	1.226	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:52

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 01/06/2024 01:40:42

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.203	01:36
Breath Volume	1.605	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.199	01:39
Breath Volume	2.437	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/06/2024 02:53:41

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.207	02:49
Breath Volume	1.593	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.199	02:52
Breath Volume	2.550	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HERRERA TORRES

DATE/TIME OF TEST: 01/07/2024 00:07:24

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Subject Sample # 3	0.244	23:59
Breath Volume	2.136	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.270	00:02
Breath Volume	1.160	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.265	00:05
Breath Volume	1.597	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: YERSHALMI

DATE/TIME OF TEST: 01/07/2024 04:38:01

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.114	04:33
Breath Volume	1.152	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.100	04:36
Breath Volume	1.992	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/07/2024 20:45:17

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.080	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.256	20:40
Breath Volume	1.328	20:40
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.256	20:43
Breath Volume	1.312	20:43
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 01/24/2024 21:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HUGHES, LUCAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 01/11/2024 23:49:05

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:44
Breath Volume	3.195	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:47
Breath Volume	2.570	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVERRIA-BRAVO

DATE/TIME OF TEST: 01/12/2024 01:27:45

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.171	01:22
Breath Volume	2.703	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.180	01:26
Breath Volume	2.992	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 01/12/2024 22:57:09

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.081	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1	22:49	No Sample Provided
Breath Volume	0.000	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2	0.205	22:55
Breath Volume	0.015	22:55	Volume Not Met
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: LAKE

DATE/TIME OF TEST: 01/12/2024 23:35:46

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.186	23:30
Breath Volume	1.796	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.191	23:34
Breath Volume	1.894	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ARENAS MALDONADO

DATE/TIME OF TEST: 01/13/2024 03:35:00

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.127	03:30
Breath Volume	2.941	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.125	03:33
Breath Volume	1.683	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOVEE, TYLER, J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ETELAMAKI, RYAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: SILVA HERNANDEZ

DATE/TIME OF TEST: 01/13/2024 19:11:35

Diagnostics Check	OK	19:07	
Air Blank	0.000	19:08	
Control Test	0.080	19:08	
Air Blank	0.000	19:09	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:10	
Control Test	0.080	19:10	
Air Blank	0.000	19:11	
Diagnostics Check	OK	19:11	

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ANGELES SANCHEZ

DATE/TIME OF TEST: 01/14/2024 01:49:54

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.214	01:45
Breath Volume	1.765	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.221	01:48
Breath Volume	2.015	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: CHANTHAVONG

DATE/TIME OF TEST: 01/14/2024 02:40:33

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.167	02:36
Breath Volume	3.945	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.166	02:39
Breath Volume	3.402	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 01/15/2024 22:24:45

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.041	22:20
Breath Volume	1.351	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.045	22:23
Breath Volume	1.320	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: CHIGUIL TIGUILA

DATE/TIME OF TEST: 01/17/2024 21:23:09

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.299	21:18
Breath Volume	1.691	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.301	21:21
Breath Volume	2.843	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: VIGNAU

DATE/TIME OF TEST: 01/18/2024 21:19:11

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.107	21:14
Breath Volume	1.730	21:14
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.112	21:17
Breath Volume	1.605	21:17
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HIRO

DATE/TIME OF TEST: 01/21/2024 03:31:05

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.212	03:26
Breath Volume	2.117	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.206	03:29
Breath Volume	1.695	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLOD,STEPHEN,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: SPIRITO,GEORGE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/14/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: GOLOD

DATE/TIME OF TEST: 01/22/2024 14:20:14

Diagnostics Check	OK	14:13
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:14
Subject Sample # 1	0.109	14:15
Breath Volume	3.113	14:15
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.107	14:18
Breath Volume	4.328	14:18
Air Blank	0.000	14:19
Control Test	0.080	14:19
Air Blank	0.000	14:20
Diagnostics Check	OK	14:20

Data Download Date/Time: 01/24/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: OVIAN

DATE/TIME OF TEST: 01/27/2024 01:42:05

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:40	
Control Test	0.081	01:40	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:41	Sequence Aborted
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: OVIAN

DATE/TIME OF TEST: 01/27/2024 01:56:25

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.087	01:51
Breath Volume	3.945	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.089	01:54
Breath Volume	3.722	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: RAMIREZ-VINCENT

DATE/TIME OF TEST: 01/27/2024 23:11:19

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.240	23:06
Breath Volume	3.257	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.245	23:09
Breath Volume	3.339	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: NUNEZ NAUN

DATE/TIME OF TEST: 01/28/2024 05:06:55

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.081	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.088	05:02
Breath Volume	1.996	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.088	05:05
Breath Volume	1.359	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DREYER

DATE/TIME OF TEST: 01/29/2024 00:06:02

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.080	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:04	
Subject Sample	# 2	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:05	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:06	

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRYANT,S
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: MONTEMAGNI

DATE/TIME OF TEST: 01/29/2024 22:18:06

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.192	22:13
Breath Volume	2.953	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.192	22:16
Breath Volume	2.937	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 02/27/2024 20:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 01/01/2024 00:04:10

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.083	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.324	23:59
Breath Volume	1.585	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.321	00:02
Breath Volume	1.964	00:02
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HERNANDEZ, LAURA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 01/01/2024 08:30:56

Diagnostics Check	OK	08:27	
Air Blank	0.000	08:28	
Control Test	0.083	08:28	
Air Blank	0.000	08:29	
Subject Sample	# 1	08:29	Subject Test Refused
Breath Volume	0.000	08:29	
Air Blank	0.000	08:29	
Control Test	0.083	08:30	
Air Blank	0.000	08:30	
Diagnostics Check	OK	08:30	

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIENTOS,KENNETH
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CRISTOBAL

DATE/TIME OF TEST: 01/04/2024 01:33:49

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.082	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.169	01:32	Volume Not Met
Breath Volume	0.773	01:32	
Air Blank		01:33	Purge Fail
Air Blank	0.000	01:33	

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIENTOS,KENNETH
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CRISTOBAL

DATE/TIME OF TEST: 01/04/2024 01:44:14

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.077	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.179	01:39
Breath Volume	1.234	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.183	01:42
Breath Volume	1.164	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRAYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BLUMENFELD

DATE/TIME OF TEST: 01/08/2024 00:40:50

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.083	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.000	00:36
Breath Volume	2.894	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:39
Breath Volume	2.671	00:39
Air Blank	0.000	00:39
Control Test	0.083	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BUSBIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 01/14/2024 11:43:46

Diagnostics Check	OK	11:35	
Air Blank	0.000	11:35	
Control Test	0.080	11:35	
Air Blank	0.000	11:36	
Subject Sample	# 1	11:39	No Sample Provided
Breath Volume	0.000	11:39	
Air Blank	0.000	11:40	
Air Blank	0.000	11:42	
Subject Sample	# 2	11:42	Subject Test Refused
Breath Volume	0.000	11:42	
Air Blank	0.000	11:42	
Control Test	0.082	11:43	
Air Blank	0.000	11:43	
Diagnostics Check	OK	11:43	

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SAUCEDA

DATE/TIME OF TEST: 01/14/2024 20:41:35

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.082	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.145	20:36
Breath Volume	1.882	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.138	20:39
Breath Volume	2.398	20:39
Air Blank	0.000	20:40
Control Test	0.082	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GARZA,RAUL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 12/17/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 07:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/22/2024 08:42:36

Diagnostics Check	OK	08:36
Air Blank	0.000	08:37
Control Test	0.083	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.000	08:38
Breath Volume	3.714	08:38
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.000	08:41
Breath Volume	3.644	08:41
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 01/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: JIMENEZ ESCOBEDO

DATE/TIME OF TEST: 01/27/2024 03:09:35

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.083	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.145	03:04
Breath Volume	3.031	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.143	03:07
Breath Volume	2.402	03:07
Air Blank	0.000	03:08
Control Test	0.082	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 03/04/2024 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 01/28/2024 02:24:06

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.083	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.163	02:19
Breath Volume	2.972	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.167	02:22
Breath Volume	2.828	02:22
Air Blank	0.000	02:23
Control Test	0.082	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/04/2024 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COLO ZAMORA

DATE/TIME OF TEST: 01/02/2024 00:56:54

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.256	00:52
Breath Volume	2.121	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.244	00:55
Breath Volume	1.816	00:55
Air Blank	0.000	00:55
Control Test	0.076	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: D'AMICO

DATE/TIME OF TEST: 01/02/2024 01:13:34

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.195	01:09
Breath Volume	1.902	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.198	01:11
Breath Volume	1.214	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LEMAY

DATE/TIME OF TEST: 01/03/2024 01:52:22

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.135	01:47
Breath Volume	1.578	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.123	01:50
Breath Volume	1.937	01:50
Air Blank	0.000	01:51
Control Test	0.075	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BARRETT,CODY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 01/03/2024 12:27:31

Diagnostics Check	OK	12:20
Air Blank	0.000	12:21
Control Test	0.080	12:21
Air Blank	0.000	12:22
Subject Sample # 1	0.000	12:22
Breath Volume	1.542	12:22
Air Blank	0.000	12:23
Air Blank	0.000	12:25
Subject Sample # 2	0.000	12:26
Breath Volume	2.335	12:26
Air Blank	0.000	12:26
Control Test	0.076	12:26
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: JUNCO

DATE/TIME OF TEST: 01/04/2024 10:33:25

Diagnostics Check	OK	10:27
Air Blank	0.000	10:27
Control Test	0.079	10:27
Air Blank	0.000	10:28
Subject Sample # 1	0.269	10:28
Breath Volume	4.675	10:28
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.257	10:31
Breath Volume	4.500	10:31
Air Blank	0.000	10:32
Control Test	0.077	10:32
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SOPER

DATE/TIME OF TEST: 01/04/2024 20:15:17

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.177	20:10
Breath Volume	1.316	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.184	20:13
Breath Volume	1.437	20:13
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELASQUEZ,ARTURO,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VELASQUEZ,ARTURO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:31

SUBJECT'S LAST NAME: PATERSON

DATE/TIME OF TEST: 01/06/2024 11:02:38

Diagnostics Check	OK	10:56
Air Blank	0.000	10:56
Control Test	0.079	10:57
Air Blank	0.000	10:57
Subject Sample # 1	0.130	10:58
Breath Volume	3.898	10:58
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.133	11:00
Breath Volume	2.804	11:00
Air Blank	0.000	11:01
Control Test	0.079	11:02
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALKER,CALVIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/07/2024 18:34:34

Diagnostics Check	OK	18:27
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.282	18:29
Breath Volume	2.312	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.267	18:32
Breath Volume	2.175	18:32
Air Blank	0.000	18:33
Control Test	0.076	18:33
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TINFINA,LOGAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: HERT

DATE/TIME OF TEST: 01/09/2024 14:25:01

Diagnostics Check	OK	14:18
Air Blank	0.000	14:19
Control Test	0.079	14:19
Air Blank	0.000	14:20
Subject Sample # 1	0.000	14:20
Breath Volume	2.972	14:20
Air Blank	0.000	14:21
Air Blank	0.000	14:23
Subject Sample # 2	0.000	14:23
Breath Volume	3.117	14:23
Air Blank	0.000	14:24
Control Test	0.078	14:24
Air Blank	0.000	14:24
Diagnostics Check	OK	14:25

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RESENDIZ RAMIREZ

DATE/TIME OF TEST: 01/10/2024 20:32:29

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.211	20:27
Breath Volume	1.109	20:27
Air Blank	0.000	20:28
Air Blank	0.000	20:29
Subject Sample # 2	0.215	20:30
Breath Volume	2.433	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREYRA JR,ROBERT,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PEREYRA JR,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: TADEO CARMELO

DATE/TIME OF TEST: 01/11/2024 01:13:02

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.188	01:07
Breath Volume	1.214	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.184	01:11
Breath Volume	1.210	01:11
Air Blank	0.000	01:12
Control Test	0.075	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 01/15/2024 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: WOODSTUFF

DATE/TIME OF TEST: 01/01/2024 05:12:49

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.170	05:07
Breath Volume	1.968	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.158	05:11
Breath Volume	1.414	05:11
Air Blank	0.000	05:11
Control Test	0.079	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 01/03/2024 21:05:27

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.079	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.353	21:00
Breath Volume	1.496	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.366	21:03
Breath Volume	2.652	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: IMBODEN

DATE/TIME OF TEST: 01/04/2024 02:55:39

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.059	02:51
Breath Volume	3.890	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.056	02:54
Breath Volume	2.992	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/04/2024 19:26:42

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:18	
Control Test	0.080	19:19	
Air Blank	0.000	19:19	
Subject Sample	# 1	19:22	No Sample Provided
Breath Volume	0.000	19:22	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample	# 2	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,ZN
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 01/04/2024 19:36:37

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.080	19:28	
Air Blank	0.000	19:29	
Subject Sample	# 1 0.187	19:32	Volume Not Met
Breath Volume	0.273	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:34	
Subject Sample	# 2	19:35	Subject Test Refused
Breath Volume	0.000	19:35	
Air Blank	0.000	19:35	
Control Test	0.079	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:36	

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 01/06/2024 03:56:30

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:52	No Sample Provided
Breath Volume	0.000	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 2	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WARD,G B
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: JACKSON JR

DATE/TIME OF TEST: 01/06/2024 04:13:18

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:08	No Sample Provided
Breath Volume	0.000	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample	# 2	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: CICERO

DATE/TIME OF TEST: 01/07/2024 00:12:17

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.059	00:07
Breath Volume	2.593	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.056	00:10
Breath Volume	3.375	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: ROSENBERGER

DATE/TIME OF TEST: 01/07/2024 03:57:06

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.125	03:52
Breath Volume	3.152	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.116	03:55
Breath Volume	3.031	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: DORUILUS

DATE/TIME OF TEST: 01/11/2024 20:19:27

Diagnostics Check	OK	20:12
Air Blank	0.000	20:12
Control Test	0.079	20:13
Air Blank	0.000	20:13
Subject Sample # 1	0.101	20:14
Breath Volume	1.183	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.101	20:17
Breath Volume	1.378	20:17
Air Blank	0.000	20:18
Control Test	0.079	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TORRES,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 01/12/2024 04:14:52

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:06	
Control Test	0.079	04:07	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LENOVER

DATE/TIME OF TEST: 01/13/2024 03:38:58

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.139	03:33
Breath Volume	3.382	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.138	03:37
Breath Volume	2.695	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 01/27/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FAIRCLOTH,D
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BORLAND

DATE/TIME OF TEST: 01/15/2024 00:14:36

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:10	No Sample Provided
Breath Volume	0.000	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2	00:13	Subject Test Refused
Breath Volume	0.000	00:13	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: LIGON

DATE/TIME OF TEST: 01/18/2024 00:23:31

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.080	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	No Sample Provided
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 2	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.080	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/19/2024 00:42:37

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 2	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.080	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CROSBY,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: FISHBURNE

DATE/TIME OF TEST: 01/20/2024 02:46:50

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1	02:42	No Sample Provided
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Air Blank	0.000	02:44	
Subject Sample	# 2	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 01/20/2024 22:53:49

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.177	22:49
Breath Volume	2.843	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.184	22:52
Breath Volume	3.054	22:52
Air Blank	0.000	22:52
Control Test	0.082	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 01/23/2024 01:15:33

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1	01:09	No Sample Provided
Breath Volume	0.000	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: SIEMINSKI

DATE/TIME OF TEST: 01/24/2024 04:15:39

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Air Blank	0.000	04:14	
Subject Sample	# 2	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.079	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: POUGH

DATE/TIME OF TEST: 01/25/2024 04:07:30

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.156	04:02
Breath Volume	2.078	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.155	04:05
Breath Volume	2.101	04:05
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: LEA

DATE/TIME OF TEST: 01/26/2024 19:31:08

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.135	19:26
Breath Volume	2.382	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:28
Subject Sample # 2	0.135	19:29
Breath Volume	1.402	19:29
Air Blank	0.000	19:30
Control Test	0.078	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/25/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: LAKHTAR

DATE/TIME OF TEST: 01/27/2024 01:48:18

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1 0.128	01:41	Volume Not Met
Breath Volume	0.062	01:41	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:46	Subject Test Refused
Breath Volume	0.351	01:46	
Air Blank	0.000	01:47	
Control Test	0.078	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 01/27/2024 11:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: MENDENHALL

DATE/TIME OF TEST: 01/30/2024 12:32:11

Diagnostics Check	OK	12:31	
Air Blank		12:31	RFI Detect
Air Blank	0.000	12:32	

Data Download Date/Time: 02/26/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:08

SUBJECT'S LAST NAME: MENDENHALL

DATE/TIME OF TEST: 01/30/2024 12:44:10

Diagnostics Check	OK	12:35
Air Blank	0.000	12:35
Control Test	0.080	12:35
Air Blank	0.000	12:36
Subject Sample # 1	0.000	12:38
Breath Volume	1.175	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:42
Breath Volume	1.523	12:42
Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44
Diagnostics Check	OK	12:44

Data Download Date/Time: 02/26/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: CRITES

DATE/TIME OF TEST: 01/31/2024 21:25:56

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.081	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.103	21:21
Breath Volume	4.050	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.102	21:24
Breath Volume	3.980	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 02/26/2024 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ZUNIGA CRUZ

DATE/TIME OF TEST: 01/01/2024 02:05:51

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.151	02:01
Breath Volume	1.406	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:03
Subject Sample # 2	0.156	02:04
Breath Volume	1.851	02:04
Air Blank	0.000	02:04
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 01/01/2024 04:08:45

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.172	04:03
Breath Volume	2.152	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.172	04:07
Breath Volume	1.421	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORRIS,ARIANNA
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SCHNABEL

DATE/TIME OF TEST: 01/01/2024 05:04:21

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.219	04:59
Breath Volume	3.074	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.216	05:02
Breath Volume	2.992	05:02
Air Blank	0.000	05:03
Control Test	0.080	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: MAURICIO TOMAS

DATE/TIME OF TEST: 01/01/2024 20:03:53

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.081	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.173	19:59
Breath Volume	1.812	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.176	20:02
Breath Volume	3.097	20:02
Air Blank	0.000	20:02
Control Test	0.079	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BAGDZIUNAS,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: ORLATEI

DATE/TIME OF TEST: 01/01/2024 20:45:15

Diagnostics Check	OK	20:41	
Air Blank	0.000	20:41	
Control Test	0.080	20:41	
Air Blank	0.000	20:42	
Subject Sample	# 1	20:43	Subject Test Refused
Breath Volume	0.000	20:43	
Air Blank	0.000	20:44	
Control Test	0.080	20:44	
Air Blank	0.000	20:45	
Diagnostics Check	OK	20:45	

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ALBIZ ANAYA

DATE/TIME OF TEST: 01/02/2024 00:13:00

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.131	00:08
Breath Volume	2.898	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.126	00:11
Breath Volume	1.582	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MENDEZ CHONA

DATE/TIME OF TEST: 01/02/2024 02:37:34

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.205	02:32
Breath Volume	1.953	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.203	02:35
Breath Volume	2.250	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: SYLVAIN

DATE/TIME OF TEST: 01/02/2024 16:06:30

Diagnostics Check	OK	15:58
Air Blank	0.000	15:58
Control Test	0.080	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.089	16:01
Breath Volume	1.929	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:03
Subject Sample # 2	0.085	16:04
Breath Volume	1.769	16:04
Air Blank	0.000	16:05
Control Test	0.079	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 01/03/2024 03:13:55

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.157	03:08
Breath Volume	1.238	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.158	03:12
Breath Volume	1.128	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: SOBHAN

DATE/TIME OF TEST: 01/04/2024 21:02:58

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.000	20:58
Breath Volume	1.976	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:01
Breath Volume	2.265	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: LONDRE

DATE/TIME OF TEST: 01/04/2024 22:26:05

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.080	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.186	22:21
Breath Volume	1.593	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.196	22:24
Breath Volume	1.460	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BADOLATI, NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: BERK

DATE/TIME OF TEST: 01/04/2024 22:58:53

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.203	22:54
Breath Volume	2.167	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.193	22:57
Breath Volume	2.476	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/12/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BROWN,MARSHON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 01/04/2024 23:26:32

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.167	23:21
Breath Volume	2.175	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.163	23:24
Breath Volume	1.378	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 01/05/2024 06:58:01

Diagnostics Check	OK	06:51
Air Blank	0.000	06:51
Control Test	0.081	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.199	06:53
Breath Volume	2.316	06:53
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.200	06:56
Breath Volume	2.820	06:56
Air Blank	0.000	06:57
Control Test	0.080	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:58

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY, CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: HITCHENS

DATE/TIME OF TEST: 01/05/2024 19:37:31

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:32
Subject Sample # 1	0.255	19:32
Breath Volume	3.781	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.253	19:35
Breath Volume	4.015	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: JUAN FRANCISCO

DATE/TIME OF TEST: 01/05/2024 22:17:29

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.283	22:12
Breath Volume	1.957	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.270	22:15
Breath Volume	1.867	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PEREZ ANTON

DATE/TIME OF TEST: 01/05/2024 23:51:01

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.255	23:46
Breath Volume	3.128	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.239	23:49
Breath Volume	1.480	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: LOZIN

DATE/TIME OF TEST: 01/06/2024 02:51:10

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.246	02:46
Breath Volume	1.152	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.249	02:49
Breath Volume	1.699	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WERTMAN,GREG
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LEBRUN

DATE/TIME OF TEST: 01/06/2024 04:47:35

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.226	04:43
Breath Volume	1.468	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.231	04:46
Breath Volume	1.593	04:46
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SURBER

DATE/TIME OF TEST: 01/07/2024 03:56:47

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.145	03:52
Breath Volume	1.566	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.144	03:55
Breath Volume	1.214	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 01/07/2024 11:09:29

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.080	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.202	11:04
Breath Volume	1.570	11:04
Air Blank	0.000	11:05
Air Blank	0.000	11:07
Subject Sample # 2	0.200	11:07
Breath Volume	1.496	11:07
Air Blank	0.000	11:08
Control Test	0.077	11:08
Air Blank	0.000	11:09
Diagnostics Check	OK	11:09

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DORIVAL, ADLER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/08/2024 03:33:39

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.150	03:29
Breath Volume	2.835	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.142	03:32
Breath Volume	1.847	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJADA,ALBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 01/08/2024 09:03:36

Diagnostics Check	OK	08:57
Air Blank	0.000	08:57
Control Test	0.081	08:57
Air Blank	0.000	08:58
Subject Sample # 1	0.176	08:58
Breath Volume	1.828	08:58
Air Blank	0.000	08:59
Air Blank	0.000	09:01
Subject Sample # 2	0.183	09:02
Breath Volume	2.441	09:02
Air Blank	0.000	09:02
Control Test	0.081	09:02
Air Blank	0.000	09:03
Diagnostics Check	OK	09:03

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FELIPE ACEYTUNO

DATE/TIME OF TEST: 01/08/2024 23:15:08

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.169	23:10
Breath Volume	1.222	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:12
Subject Sample # 2	0.175	23:13
Breath Volume	2.464	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEGRAFFENREID,JADA
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HERRAN

DATE/TIME OF TEST: 01/09/2024 01:10:29

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:58	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.154	01:02	Volume Not Met
Breath Volume	0.382	01:02	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.181	01:05	
Breath Volume	1.109	01:05	
Air Blank	0.000	01:05	
Air Blank	0.000	01:07	
Subject Sample	# 3 0.179	01:08	
Breath Volume	1.660	01:08	
Air Blank	0.000	01:09	
Control Test	0.078	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANCHEZ,JULIAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/09/2024 03:52:02

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.145	03:47
Breath Volume	2.113	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.129	03:50
Breath Volume	1.968	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN, LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: DOWD

DATE/TIME OF TEST: 01/10/2024 13:05:47

Diagnostics Check	OK	12:59
Air Blank	0.000	12:59
Control Test	0.080	13:00
Air Blank	0.000	13:00
Subject Sample # 1	0.072	13:01
Breath Volume	1.886	13:01
Air Blank	0.000	13:01
Air Blank	0.000	13:03
Subject Sample # 2	0.072	13:04
Breath Volume	1.714	13:04
Air Blank	0.000	13:04
Control Test	0.080	13:05
Air Blank	0.000	13:05
Diagnostics Check	OK	13:05

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HADLEY,NICHOLAS
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: PECHONIS

DATE/TIME OF TEST: 01/10/2024 16:16:04

Diagnostics Check	OK	16:09
Air Blank	0.000	16:09
Control Test	0.080	16:10
Air Blank	0.000	16:10
Subject Sample # 1	0.181	16:11
Breath Volume	2.531	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:13
Subject Sample # 2	0.173	16:14
Breath Volume	3.246	16:14
Air Blank	0.000	16:15
Control Test	0.080	16:15
Air Blank	0.000	16:15
Diagnostics Check	OK	16:16

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: LUCAS MENDEZ

DATE/TIME OF TEST: 01/10/2024 21:22:22

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.175	21:17
Breath Volume	1.375	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.187	21:20
Breath Volume	1.230	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: COBO GUZMAN

DATE/TIME OF TEST: 01/11/2024 20:27:36

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.175	20:23
Breath Volume	2.667	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.174	20:26
Breath Volume	2.691	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 01/12/2024 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 01/05/2024 01:35:46

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.082	01:34	
Air Blank		01:35	RFI Detect
Air Blank	0.000	01:35	

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 01/05/2024 01:39:20

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank		01:38	RFI Detect
Air Blank	0.000	01:39	

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: FERRO

DATE/TIME OF TEST: 01/05/2024 01:48:19

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.124	01:43
Breath Volume	1.964	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.122	01:46
Breath Volume	1.234	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 01/13/2024 02:42:15

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.140	02:37
Breath Volume	1.968	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.137	02:40
Breath Volume	1.898	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HOSEY

DATE/TIME OF TEST: 01/18/2024 02:04:51

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.181	01:59
Breath Volume	1.828	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.180	02:03
Breath Volume	1.257	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: RICKETT

DATE/TIME OF TEST: 01/22/2024 20:22:30

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.081	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.113	20:17
Breath Volume	2.382	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.108	20:20
Breath Volume	2.187	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: COFFMAN,DYLAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: LANDGRAF

DATE/TIME OF TEST: 01/23/2024 03:54:58

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.109	03:50
Breath Volume	1.714	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.111	03:53
Breath Volume	1.765	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: COFFMAN,DYLAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 01/23/2024 21:19:55

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.157	21:15
Breath Volume	1.683	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.156	21:18
Breath Volume	1.328	21:18
Air Blank	0.000	21:18
Control Test	0.078	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CANGEMI

DATE/TIME OF TEST: 01/23/2024 23:25:27

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.172	23:20
Breath Volume	2.398	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.165	23:23
Breath Volume	2.230	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 01/26/2024 18:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 01/27/2024 01:13:21

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.089	01:08
Breath Volume	1.503	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.085	01:11
Breath Volume	1.613	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 02/01/2024 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: NEVES

DATE/TIME OF TEST: 01/28/2024 02:39:05

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.154	02:34
Breath Volume	3.191	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.137	02:37
Breath Volume	2.761	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/01/2024 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MESHELLE,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CARROLL,SEAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/05/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/14/2024 21:39:56

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.085	21:33
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:35
Breath Volume	3.136	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:38
Breath Volume	2.125	21:38
Air Blank	0.000	21:39
Control Test	0.081	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 01/30/2024 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHOTT,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 01/07/2024 02:35:21

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.221	02:30
Breath Volume	3.566	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.221	02:33
Breath Volume	3.519	02:33
Air Blank	0.000	02:34
Control Test	0.079	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADER,JAMES,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WEEKES,DEMITRI
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DILLARD

DATE/TIME OF TEST: 01/08/2024 04:33:41

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.182	04:28
Breath Volume	3.332	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.180	04:31
Breath Volume	3.050	04:31
Air Blank	0.000	04:32
Control Test	0.081	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RILEY,BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:07:34

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
		22:07	RFI Detect
Air Blank	0.000	22:07	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TOBY, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RILEY, BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:13:47

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:09	
Control Test	0.080	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1 0.269	22:11	
Breath Volume	2.332	22:11	
Air Blank		22:11	RFI Detect
Air Blank	0.000	22:12	
Control Test	0.079	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TOBY, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RILEY, BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:21:51

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1 0.268	22:17	
Breath Volume	1.878	22:17	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2	22:20	RFI Detect
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.079	22:21	
Air Blank		22:21	RFI Detect
Air Blank	0.000	22:21	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TOBY, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RILEY, BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:30:49

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.079	22:24	
Air Blank	0.000	22:25	
Subject Sample # 1	0.272	22:25	
Breath Volume	2.761	22:25	
Air Blank	0.000	22:27	
Air Blank	0.000	22:28	
Subject Sample # 2	0.258	22:28	
Breath Volume	2.285	22:28	
Air Blank	0.000	22:29	
Control Test	0.074	22:30	Control Outside Tolerance
Air Blank	0.000	22:30	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TOBY,B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RILEY,BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:44:18

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:43	
Control Test	0.072	22:43	Control Outside Tolerance
Air Blank	0.000	22:44	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TOBY, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RALEY, BRITT
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/08/2024 22:57:27

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.255	22:52
Breath Volume	3.371	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.250	22:55
Breath Volume	3.683	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADER,JAMES,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JOHNSON,DUSTIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/13/2024 05:19:42

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.000	05:15
Breath Volume	1.367	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.000	05:18
Breath Volume	2.488	05:18
Air Blank	0.000	05:18
Control Test	0.079	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,AUTIS,E
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GIBSON,AUTIS
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/27/2024 03:48:11

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 01/30/2024 20:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 01/08/2024 04:38:08

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.156	04:32
Breath Volume	1.183	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.159	04:36
Breath Volume	3.425	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/16/2024 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/11/2024 07:16:49

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.082	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.000	07:12
Breath Volume	3.589	07:12
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.000	07:15
Breath Volume	3.820	07:15
Air Blank	0.000	07:15
Control Test	0.081	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 01/16/2024 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: SPONHEIM

DATE/TIME OF TEST: 01/15/2024 20:18:15

Diagnostics Check	OK	20:02	
Air Blank	0.000	20:03	
Control Test	0.081	20:03	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.157	20:06	Volume Not Met
Breath Volume	0.394	20:06	
Air Blank	0.000	20:08	
Air Blank	0.000	20:09	
Subject Sample	# 2 0.156	20:11	
Breath Volume	1.699	20:11	
Air Blank	0.000	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 3 0.145	20:16	
Breath Volume	1.175	20:16	
Air Blank	0.000	20:17	
Control Test	0.074	20:17	Control Outside Tolerance
Air Blank	0.000	20:18	

Data Download Date/Time: 01/16/2024 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,REGINALD,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROGERS,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 09:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: CRUZLICONA

DATE/TIME OF TEST: 01/01/2024 10:51:55

Diagnostics Check	OK	10:45
Air Blank	0.000	10:45
Control Test	0.081	10:46
Air Blank	0.000	10:46
Subject Sample # 1	0.209	10:47
Breath Volume	1.679	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:49
Subject Sample # 2	0.214	10:50
Breath Volume	1.464	10:50
Air Blank	0.000	10:50
Control Test	0.076	10:51
Air Blank	0.000	10:51
Diagnostics Check	OK	10:51

Data Download Date/Time: 01/16/2024 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,MATTHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HALL,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: CARRILLO TRINIDAD

DATE/TIME OF TEST: 01/02/2024 16:56:06

Diagnostics Check	OK	16:44	
Air Blank	0.000	16:45	
Control Test	0.083	16:45	
Air Blank	0.000	16:46	
Subject Sample	# 1 0.076	16:49	Volume Not Met
Breath Volume	0.214	16:49	
Air Blank	0.000	16:49	
Air Blank	0.000	16:51	
Subject Sample	# 2 0.072	16:52	
Breath Volume	2.035	16:52	
Air Blank	0.000	16:52	
Air Blank	0.000	16:54	
Subject Sample	# 3 0.067	16:54	
Breath Volume	2.332	16:54	
Air Blank	0.000	16:55	
Control Test	0.000	16:55	RFI Detect
Air Blank		16:56	RFI Detect

Data Download Date/Time: 01/16/2024 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: CARVER, JAMIE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 01/20/2024 01:18:24

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.084	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.147	01:13
Breath Volume	1.824	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.138	01:16
Breath Volume	1.566	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 02/14/2024 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,CARL,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAIGLE,T
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ELOUADI

DATE/TIME OF TEST: 01/24/2024 05:15:18

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.201	05:10
Breath Volume	3.171	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.192	05:13
Breath Volume	1.781	05:13
Air Blank	0.000	05:14
Control Test	0.077	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 02/14/2024 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: VUITTONET

DATE/TIME OF TEST: 01/27/2024 12:40:45

Diagnostics Check	OK	12:40	
Air Blank		12:40	RFI Detect
Air Blank		12:40	RFI Detect

Data Download Date/Time: 02/14/2024 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: VUITTONET

DATE/TIME OF TEST: 01/27/2024 12:43:43

Diagnostics Check	OK	12:43	
Air Blank		12:43	RFI Detect
Air Blank		12:43	RFI Detect

Data Download Date/Time: 02/14/2024 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTHE,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOOTHE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CUEVAS ALAMILLO

DATE/TIME OF TEST: 01/28/2024 03:09:42

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.082	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.196	03:05
Breath Volume	1.941	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.195	03:08
Breath Volume	1.722	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 02/14/2024 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,ANDREW,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROJAS,ANDREW
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HAYMAN

DATE/TIME OF TEST: 01/01/2024 01:50:23

Diagnostics Check	OK	01:50	
Air Blank		01:50	Improper Sample
Air Blank		01:50	Improper Sample

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,ANDREW,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROJAS,ANDREW
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: HAYMAN

DATE/TIME OF TEST: 01/01/2024 01:53:48

Diagnostics Check	OK	01:53	
Air Blank		01:53	Improper Sample
Air Blank		01:53	Improper Sample

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCOIS,MARC,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FRANCOIS,MARC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 01/01/2024 10:31:49

Diagnostics Check	OK	10:28	
Air Blank	0.000	10:29	
Control Test	0.078	10:29	
Air Blank	0.000	10:30	
Subject Sample	# 1	10:30	Subject Test Refused
Breath Volume	0.000	10:30	
Air Blank	0.000	10:30	
Control Test	0.078	10:31	
Air Blank	0.000	10:31	
Diagnostics Check	OK	10:31	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCE, DOMINUSE-CAIN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FRANCOIS, MARC
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 01/01/2024 11:54:55

Diagnostics Check	OK	11:49	
Air Blank	0.000	11:50	
Control Test	0.079	11:50	
Air Blank	0.000	11:50	
Subject Sample	# 1	11:53	Subject Test Refused
Breath Volume	0.000	11:53	
Air Blank	0.000	11:54	
Control Test	0.079	11:54	
Air Blank	0.000	11:54	
Diagnostics Check	OK	11:54	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEYNE, DAVANO, L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALLEYNE, DAVANO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: AROCHE

DATE/TIME OF TEST: 01/01/2024 23:37:41

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.080	23:35	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:36	
Control Test	0.080	23:37	
Air Blank	0.000	23:37	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GAITOR

DATE/TIME OF TEST: 01/07/2024 01:20:52

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.078	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIETO,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 01/07/2024 02:43:59

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUZ,ASHLEY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 01/07/2024 02:51:52

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.079	02:51	
Air Blank		02:51	RFI Detect
Air Blank	0.000	02:51	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUZ,ASHLEY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,CAMILA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 01/07/2024 02:58:30

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,PENELOPE,I
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,PENELOPE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 01/07/2024 09:26:34

Diagnostics Check	OK	09:22	
Air Blank	0.000	09:23	
Control Test	0.079	09:23	
Air Blank	0.000	09:24	
Subject Sample	# 1	09:25	Subject Test Refused
Breath Volume	0.000	09:25	
Air Blank	0.000	09:25	
Control Test	0.079	09:25	
Air Blank	0.000	09:26	
Diagnostics Check	OK	09:26	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MALDONADO BONILLA

DATE/TIME OF TEST: 01/19/2024 06:04:20

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.079	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.166	05:59
Breath Volume	1.839	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.166	06:02
Breath Volume	1.343	06:02
Air Blank	0.000	06:03
Control Test	0.079	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 00:31

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ALCANTARA

DATE/TIME OF TEST: 01/21/2024 01:23:46

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.076	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.167	01:19
Breath Volume	2.171	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:21
Subject Sample # 2	0.169	01:22
Breath Volume	2.480	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 03:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: DIAZ MONTES

DATE/TIME OF TEST: 01/21/2024 04:43:12

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.079	04:40	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.079	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 04:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 01/21/2024 06:01:51

Diagnostics Check	OK	05:59	
Air Blank	0.000	06:00	
Control Test	0.079	06:00	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:01	Subject Test Refused
Breath Volume	0.000	06:01	
Air Blank		06:01	RFI Detect
Air Blank	0.000	06:01	

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MARTINEZ,REMY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: DUI

ARREST TIME: 11:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: SOMARRIBA

DATE/TIME OF TEST: 01/21/2024 12:52:17

Diagnostics Check	OK	12:43
Air Blank	0.000	12:43
Control Test	0.079	12:44
Air Blank	0.000	12:44
Subject Sample # 1	0.283	12:44
Breath Volume	1.847	12:44
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.255	12:47
Breath Volume	1.683	12:47
Air Blank	0.000	12:48
Air Blank	0.000	12:50
Subject Sample # 3	0.276	12:50
Breath Volume	2.082	12:50
Air Blank	0.000	12:51
Control Test	0.078	12:51
Air Blank	0.000	12:52
Diagnostics Check	OK	12:52

Data Download Date/Time: 01/22/2024 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEJARZA,LUIS,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CAPAZ,TAMMY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: GALLEGO GUTIERREZ

DATE/TIME OF TEST: 01/29/2024 02:00:07

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.079	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.141	01:51	Volume Not Met
Breath Volume	0.808	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 2 0.157	01:54	
Breath Volume	1.441	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 3 0.154	01:58	
Breath Volume	1.707	01:58	
Air Blank	0.000	01:59	
Control Test	0.077	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 02/16/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CANOVAS,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: TORREZ

DATE/TIME OF TEST: 01/07/2024 02:59:59

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.077	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 01/22/2024 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSQUERA,SANTIAGO,D
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MOSQUERA,SANTIAGO
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/18/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 989917
DRY GAS EXPIRY DATE: 07/25/2021
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MOLINA SANCHEZ

DATE/TIME OF TEST: 01/19/2024 00:14:43

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.000	00:10
Breath Volume	2.343	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.000	00:13
Breath Volume	3.222	00:13
Air Blank	0.000	00:13
Control Test	0.077	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 01/22/2024 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1506449
BREATH TEST OPERATOR: PEREZ,ALEXANDER,	DRY GAS EXPIRY DATE: 06/10/2024
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: PEREZ,ALEXANDER	ARREST TIME: 19:06
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 01/22/2024	OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: TORRES	DATE/TIME OF TEST: 01/27/2024	21:19:23
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	21:13
Air Blank	0.000	21:13
Control Test	0.079	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:15
Breath Volume	1.585	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.000	21:18
Breath Volume	3.226	21:18
Air Blank	0.000	21:18
Control Test	0.077	21:18
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 02/16/2024 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PEREZ,ALEXANDER,

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: GUTIERREZ,DANIEL

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 01/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1506449

DRY GAS EXPIRY DATE: 06/10/2024

VIOLATION CODE: BUI

ARREST TIME: 21:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 01/28/2024 22:11:32

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.056	22:06
Breath Volume	1.199	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.052	22:09
Breath Volume	1.238	22:09
Air Blank	0.000	22:10
Control Test	0.077	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 02/16/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 12/21/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 01/03/2024 00:48:09

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.250	00:43
Breath Volume	1.742	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.242	00:46
Breath Volume	1.789	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 01/03/2024 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,ANFERNEE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: STEINES

DATE/TIME OF TEST: 01/04/2024 22:42:46

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Control Test	0.081	22:42	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: RIZZO

DATE/TIME OF TEST: 01/05/2024 02:28:30

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.034	02:23
Breath Volume	2.484	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.034	02:26
Breath Volume	3.042	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLICK, LOREN, D
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: SLICK, LOREN
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 01/05/2024 20:48:16

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.252	20:42
Breath Volume	1.578	20:42
Air Blank	0.000	20:42
Air Blank	0.000	20:44
Subject Sample # 2	0.247	20:46
Breath Volume	2.164	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARDNER,LANDON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:04

SUBJECT'S LAST NAME: NARDOLILLO

DATE/TIME OF TEST: 01/11/2024 13:30:48

Diagnostics Check	OK	13:24
Air Blank	0.000	13:25
Control Test	0.081	13:25
Air Blank	0.000	13:26
Subject Sample # 1	0.000	13:26
Breath Volume	3.765	13:26
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.000	13:29
Breath Volume	2.851	13:29
Air Blank	0.000	13:29
Control Test	0.080	13:30
Air Blank	0.000	13:30
Diagnostics Check	OK	13:30

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENZUELA,ROBERT,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: VALENZUELA,ROBERT
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BRASHARES

DATE/TIME OF TEST: 01/16/2024 01:15:15

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Control Test	0.079	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 01/24/2024 20:39:55

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:38	Subject Test Refused
Breath Volume	0.000	20:38	
Air Blank	0.000	20:38	
Control Test	0.082	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 01/24/2024 22:00:57

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.200	21:54
Breath Volume	1.460	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.198	21:59
Breath Volume	1.843	21:59
Air Blank	0.000	22:00
Control Test	0.079	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GILROY,ZACHARY
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/28/2024 02:18:49

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.128	02:13
Breath Volume	2.179	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.123	02:17
Breath Volume	1.894	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: TURCIOS-BENITES

DATE/TIME OF TEST: 01/29/2024 01:20:40

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 02/22/2024 08:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SEIB,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: PLAZA ESPINDOLA

DATE/TIME OF TEST: 01/01/2024 07:35:34

Diagnostics Check	OK	07:29
Air Blank	0.000	07:29
Control Test	0.080	07:30
Air Blank	0.000	07:30
Subject Sample # 1	0.198	07:31
Breath Volume	1.121	07:31
Air Blank	0.000	07:31
Air Blank	0.000	07:33
Subject Sample # 2	0.194	07:33
Breath Volume	2.625	07:33
Air Blank	0.000	07:34
Control Test	0.079	07:34
Air Blank	0.000	07:35
Diagnostics Check	OK	07:35

Data Download Date/Time: 01/31/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/04/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1355293
DRY GAS EXPIRY DATE: 07/29/2023
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: CASTLE

DATE/TIME OF TEST: 01/19/2024 18:03:01

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:00	
Control Test	0.081	18:00	
Air Blank	0.000	18:01	
Subject Sample	# 1	18:01	Subject Test Refused
Breath Volume	0.000	18:01	
Air Blank	0.000	18:02	
Control Test	0.081	18:02	
Air Blank	0.000	18:02	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 01/31/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,A
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: BLANDON,ERICK
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: AMBROCIO MIRANDA

DATE/TIME OF TEST: 01/02/2024 04:58:36

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:56	
Control Test	0.082	04:56	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.082	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 01/31/2024 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 01/11/2024 03:43:14

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.015	03:41	
Air Blank	0.000	03:42	
Control Test	0.082	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDA,CHELSY,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: MONTALVO

DATE/TIME OF TEST: 01/13/2024 17:47:33

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:42	
Control Test	0.082	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1	17:46	Subject Test Refused
Breath Volume	0.000	17:46	
Air Blank	0.000	17:46	
Control Test	0.082	17:46	
Air Blank	0.000	17:47	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/20/2024 03:23:58

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:10	
Control Test		0.082	03:10	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.074	03:14	Volume Not Met
Breath Volume		0.542	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 2	0.081	03:17	
Breath Volume		1.300	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 3	0.083	03:22	
Breath Volume		1.671	03:22	
Air Blank		0.000	03:23	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,OBEL,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: RANDELLO,ALFONSO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: PADRON AVILES

DATE/TIME OF TEST: 01/24/2024 23:17:29

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1	23:16	Subject Test Refused
Breath Volume	0.000	23:16	
Air Blank	0.000	23:16	
Control Test	0.082	23:16	
Air Blank	0.000	23:17	
Diagnostics Check	OK	23:17	

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUADRA,MARVIN,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCEACHIN,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 01/30/2024 17:08:44

Diagnostics Check	OK	17:05	
Air Blank	0.000	17:06	
Control Test	0.083	17:06	
Air Blank	0.000	17:06	
Subject Sample	# 1	17:07	RFI Detect
Breath Volume	0.000	17:07	
Air Blank	0.000	17:07	
Control Test	0.083	17:08	
Air Blank	0.000	17:08	
Diagnostics Check	OK	17:08	

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUADRA,MARVIN,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: MCEACHIN,SCOTT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/28/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 01/30/2024 17:16:42

Diagnostics Check	OK	17:10
Air Blank	0.000	17:11
Control Test	0.082	17:11
Air Blank	0.000	17:12
Subject Sample # 1	0.000	17:12
Breath Volume	2.609	17:12
Air Blank	0.000	17:12
Air Blank	0.000	17:14
Subject Sample # 2	0.000	17:15
Breath Volume	2.308	17:15
Air Blank	0.000	17:15
Control Test	0.082	17:16
Air Blank	0.000	17:16
Diagnostics Check	OK	17:16

Data Download Date/Time: 01/31/2024 21:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HEPBURN

DATE/TIME OF TEST: 01/19/2024 06:02:14

Diagnostics Check	OK	05:55
Air Blank	0.000	05:56
Control Test	0.078	05:56
Air Blank	0.000	05:57
Subject Sample # 1	0.185	05:57
Breath Volume	2.757	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample # 2	0.174	06:00
Breath Volume	2.816	06:00
Air Blank	0.000	06:01
Control Test	0.077	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 02/07/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: MEDINA HAZ

DATE/TIME OF TEST: 01/20/2024 04:41:05

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.077	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.081	04:35
Breath Volume	2.539	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.071	04:39
Breath Volume	1.386	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/07/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/01/2024 02:26:22

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.297	02:21
Breath Volume	2.269	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.305	02:24
Breath Volume	3.093	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: BJORKLUND

DATE/TIME OF TEST: 01/01/2024 05:51:09

Diagnostics Check	OK	05:48	
Air Blank	0.000	05:48	
Control Test	0.081	05:49	
Air Blank	0.000	05:49	
Subject Sample	# 1	05:49	Subject Test Refused
Breath Volume	0.000	05:49	
Air Blank	0.000	05:50	
Control Test	0.080	05:50	
Air Blank	0.000	05:51	
Diagnostics Check	OK	05:51	

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HARRIS,JAMES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/03/2024 19:12:02

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.356	19:07
Breath Volume	2.484	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:09
Subject Sample # 2	0.363	19:10
Breath Volume	2.527	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:12

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PEREZ VASQUEZ

DATE/TIME OF TEST: 01/06/2024 03:16:27

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.232	03:11
Breath Volume	2.726	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.231	03:14
Breath Volume	2.636	03:14
Air Blank	0.000	03:15
Control Test	0.076	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ORTIZ ORTIZ

DATE/TIME OF TEST: 01/07/2024 01:44:43

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.213	01:39
Breath Volume	1.910	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.212	01:43
Breath Volume	1.878	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCGILLICUDDY, STEVEN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 01/09/2024 19:04:40

Diagnostics Check	OK	18:58
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.040	19:00
Breath Volume	1.292	19:00
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.039	19:03
Breath Volume	1.285	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:04
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANDERSON,SHANE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: MCCUTCHAN

DATE/TIME OF TEST: 01/09/2024 20:02:42

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.080	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.175	19:58
Breath Volume	1.515	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.188	20:01
Breath Volume	1.914	20:01
Air Blank	0.000	20:01
Control Test	0.079	20:02
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/08/2023

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/11/2024 23:52:58

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.222	23:48
Breath Volume	1.476	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.216	23:51
Breath Volume	1.238	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 01/12/2024 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 01/12/2024 21:11:09

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.000	21:06
Breath Volume	1.253	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.000	21:09
Breath Volume	1.136	21:09
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANCHEZ, JULIAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ORTIZ PEREZ

DATE/TIME OF TEST: 01/12/2024 23:45:08

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.226	23:40
Breath Volume	2.414	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.214	23:43
Breath Volume	2.421	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JAIME
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/13/2024 00:32:09

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.079	00:27
Breath Volume	1.562	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.079	00:30
Breath Volume	3.664	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR, NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BACHLI

DATE/TIME OF TEST: 01/13/2024 03:35:14

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.203	03:30
Breath Volume	1.460	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:32
Subject Sample # 2	0.208	03:33
Breath Volume	1.585	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER, MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MATEO MATEO

DATE/TIME OF TEST: 01/13/2024 03:52:46

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.078	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.184	03:48
Breath Volume	1.453	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.187	03:51
Breath Volume	1.566	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RICCI,MATTHEW
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/13/2024 23:35:58

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.158	23:27
Breath Volume	1.253	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.182	23:31
Breath Volume	2.757	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 3	0.177	23:34
Breath Volume	2.472	23:34
Air Blank	0.000	23:35
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANCHEZ, JULIAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: RODRIGUEZ MEJIA

DATE/TIME OF TEST: 01/13/2024 23:55:00

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.155	23:50
Breath Volume	2.218	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.156	23:53
Breath Volume	2.890	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:55

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ZELFINE

DATE/TIME OF TEST: 01/14/2024 23:02:04

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1 0.098	22:54	Volume Not Met
Breath Volume	0.914	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.102	23:00	Volume Not Met
Breath Volume	0.019	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALES-LOPEZ,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GARCIA VASQUEZ

DATE/TIME OF TEST: 01/15/2024 01:41:55

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.205	01:37
Breath Volume	2.132	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.200	01:40
Breath Volume	1.859	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 01/15/2024 02:14:58

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.041	02:10
Breath Volume	1.546	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.043	02:13
Breath Volume	2.316	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ESTEBAN SILVESTRE

DATE/TIME OF TEST: 01/16/2024 00:04:27

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.125	00:00
Breath Volume	3.031	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.125	00:02
Breath Volume	2.671	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: CROMETIE

DATE/TIME OF TEST: 01/16/2024 03:16:28

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BARTOLOME JUAN

DATE/TIME OF TEST: 01/17/2024 23:59:24

Diagnostics Check	OK	23:52
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.265	23:54
Breath Volume	3.265	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.252	23:57
Breath Volume	2.691	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CALEL SONTAY

DATE/TIME OF TEST: 01/19/2024 22:21:49

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.129	22:17
Breath Volume	1.843	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.126	22:20
Breath Volume	2.464	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BRYANT,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PEREZ RAMIREZ

DATE/TIME OF TEST: 01/19/2024 23:52:20

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.200	23:47
Breath Volume	1.105	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.201	23:50
Breath Volume	1.722	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 01/20/2024 20:10:20

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.159	20:05
Breath Volume	1.441	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.145	20:08
Breath Volume	1.492	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOODY,MARIA,R
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BARNEY

DATE/TIME OF TEST: 01/20/2024 21:52:55

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.000	21:48
Breath Volume	1.578	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.000	21:51
Breath Volume	1.667	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANEGO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 01/21/2024 16:37:43

Diagnostics Check	OK	16:24	
Air Blank	0.000	16:24	
Control Test	0.080	16:25	
Air Blank	0.000	16:25	
Subject Sample	# 1 0.069	16:28	Volume Not Met
Breath Volume	0.097	16:28	
Air Blank	0.000	16:29	
Air Blank	0.000	16:31	
Subject Sample	# 2 0.076	16:33	
Breath Volume	1.144	16:33	
Air Blank	0.000	16:33	
Air Blank	0.000	16:35	
Subject Sample	# 3 0.084	16:36	
Breath Volume	1.343	16:36	
Air Blank	0.000	16:36	
Control Test	0.079	16:37	
Air Blank	0.000	16:37	
Diagnostics Check	OK	16:37	

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: GARCIA MATEO

DATE/TIME OF TEST: 01/21/2024 20:18:21

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.184	20:13
Breath Volume	1.207	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.181	20:16
Breath Volume	1.453	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLAKE, XAVIER
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 01/21/2024 22:20:45

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.122	22:16
Breath Volume	2.734	22:16
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.118	22:19
Breath Volume	2.574	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BUNSIE

DATE/TIME OF TEST: 01/23/2024 00:54:16

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.182	00:44
Breath Volume	1.773	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.161	00:49
Breath Volume	1.199	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 3	0.177	00:52
Breath Volume	1.328	00:52
Air Blank	0.000	00:53
Control Test	0.076	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 01/24/2024 07:24:26

Diagnostics Check	OK	07:08	
Air Blank	0.000	07:08	
Control Test	0.079	07:08	
Air Blank	0.000	07:09	
Subject Sample	# 1 0.097	07:11	
Breath Volume	2.835	07:11	
Air Blank	0.000	07:12	
Air Blank	0.000	07:14	
Subject Sample	# 2	07:17	No Sample Provided
Breath Volume	0.000	07:17	
Air Blank	0.000	07:17	
Air Blank	0.000	07:19	
Subject Sample	# 3	07:23	No Sample Provided
Breath Volume	0.000	07:23	
Air Blank	0.000	07:23	
Control Test	0.079	07:23	
Air Blank	0.000	07:24	
Diagnostics Check	OK	07:24	

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANNIGAN, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: LAMOUREUX

DATE/TIME OF TEST: 01/24/2024 13:10:13

Diagnostics Check	OK	13:03
Air Blank	0.000	13:04
Control Test	0.079	13:04
Air Blank	0.000	13:04
Subject Sample # 1	0.189	13:05
Breath Volume	1.558	13:05
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.191	13:08
Breath Volume	2.164	13:08
Air Blank	0.000	13:09
Control Test	0.078	13:09
Air Blank	0.000	13:10
Diagnostics Check	OK	13:10

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 01/25/2024 00:14:07

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.117	00:09
Breath Volume	1.125	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.124	00:12
Breath Volume	2.585	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/25/2024 00:31:09

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.080	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.212	00:26
Breath Volume	4.214	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.214	00:29
Breath Volume	4.183	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: AL SALEH

DATE/TIME OF TEST: 01/25/2024 03:01:03

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.000	02:56
Breath Volume	2.679	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.000	02:59
Breath Volume	2.757	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: SANTIZO

DATE/TIME OF TEST: 01/25/2024 09:26:45

Diagnostics Check	OK	09:19
Air Blank	0.000	09:20
Control Test	0.080	09:20
Air Blank	0.000	09:20
Subject Sample # 1	0.179	09:21
Breath Volume	1.957	09:21
Air Blank	0.000	09:22
Air Blank	0.000	09:23
Subject Sample # 2	0.171	09:25
Breath Volume	1.667	09:25
Air Blank	0.000	09:25
Control Test	0.078	09:26
Air Blank	0.000	09:26
Diagnostics Check	OK	09:26

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CIARRONE,JONATHAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BARRERA GOMEZ

DATE/TIME OF TEST: 01/25/2024 22:57:47

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.135	22:52
Breath Volume	1.687	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.132	22:56
Breath Volume	1.390	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 01/26/2024 02:46:24

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.168	02:41
Breath Volume	2.515	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.156	02:44
Breath Volume	1.492	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ACEVEDO, MANUEL
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/26/2024 11:03:18

Diagnostics Check	OK	10:57
Air Blank	0.000	10:57
Control Test	0.079	10:57
Air Blank	0.000	10:58
Subject Sample # 1	0.073	10:58
Breath Volume	1.687	10:58
Air Blank	0.000	10:59
Air Blank	0.000	11:01
Subject Sample # 2	0.068	11:01
Breath Volume	1.347	11:01
Air Blank	0.000	11:02
Control Test	0.079	11:02
Air Blank	0.000	11:03
Diagnostics Check	OK	11:03

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH, MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/27/2024 23:54:24

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.154	23:49
Breath Volume	1.382	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.166	23:52
Breath Volume	2.324	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: D ANDREA, ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: VELASQUEZ CHILEL

DATE/TIME OF TEST: 01/28/2024 18:59:31

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:53
Subject Sample # 1	0.273	18:54
Breath Volume	2.117	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.270	18:57
Breath Volume	2.281	18:57
Air Blank	0.000	18:58
Control Test	0.077	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRATEREAUX,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: ROBLERO MAZARIEGOS

DATE/TIME OF TEST: 01/28/2024 22:34:26

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1 0.219	22:22	
Breath Volume	1.316	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 2 0.185	22:28	Volume Not Met
Breath Volume	0.722	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:30	
Subject Sample	# 3 0.202	22:32	
Breath Volume	1.132	22:32	
Air Blank	0.000	22:33	
Control Test	0.077	22:33	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WERTMAN,GREG
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PHILLIPPI

DATE/TIME OF TEST: 01/29/2024 20:42:17

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.342	20:37
Breath Volume	1.738	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.338	20:40
Breath Volume	2.371	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WILLIAMS, JEREMY
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: EASTMAN

DATE/TIME OF TEST: 01/30/2024 09:44:17

Diagnostics Check	OK	09:37
Air Blank	0.000	09:37
Control Test	0.080	09:38
Air Blank	0.000	09:38
Subject Sample # 1	0.094	09:39
Breath Volume	1.902	09:39
Air Blank	0.000	09:40
Air Blank	0.000	09:42
Subject Sample # 2	0.094	09:42
Breath Volume	1.804	09:42
Air Blank	0.000	09:43
Control Test	0.079	09:43
Air Blank	0.000	09:44
Diagnostics Check	OK	09:44

Data Download Date/Time: 02/16/2024 16:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 01/13/2024 02:21:08

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.275	02:16
Breath Volume	3.183	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.269	02:19
Breath Volume	3.832	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AMADON,WILL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: VAGANAY

DATE/TIME OF TEST: 01/13/2024 03:42:57

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.206	03:38
Breath Volume	2.195	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.207	03:41
Breath Volume	1.496	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: ROBLERO MORALES

DATE/TIME OF TEST: 01/13/2024 23:46:38

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.257	23:41
Breath Volume	2.390	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.251	23:44
Breath Volume	2.300	23:44
Air Blank	0.000	23:45
Control Test	0.077	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GIBBS,EMMA
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 01/14/2024 05:06:07

Diagnostics Check	OK	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.187	05:01
Breath Volume	1.726	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:03
Subject Sample # 2	0.191	05:04
Breath Volume	2.058	05:04
Air Blank	0.000	05:05
Control Test	0.078	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ZAVALA AQUINO

DATE/TIME OF TEST: 01/15/2024 01:21:31

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.078	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.176	01:15
Breath Volume	1.386	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.179	01:19
Breath Volume	1.621	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE, RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: VOZZOLO

DATE/TIME OF TEST: 01/16/2024 00:40:22

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.000	00:36
Breath Volume	1.703	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.000	00:38
Breath Volume	1.324	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANCHIPOLOVSKY,DIMITRI
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: GOSCINSKI

DATE/TIME OF TEST: 01/19/2024 18:02:19

Diagnostics Check	OK	17:55
Air Blank	0.000	17:56
Control Test	0.079	17:56
Air Blank	0.000	17:57
Subject Sample # 1	0.155	17:57
Breath Volume	1.777	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.153	18:00
Breath Volume	2.039	18:00
Air Blank	0.000	18:01
Control Test	0.078	18:01
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: WIMBERLEY

DATE/TIME OF TEST: 01/19/2024 21:54:26

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.147	21:49
Breath Volume	1.183	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.145	21:52
Breath Volume	1.812	21:52
Air Blank	0.000	21:53
Control Test	0.078	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAYOR,STEVEN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GROUT

DATE/TIME OF TEST: 01/20/2024 01:46:59

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.078	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1 0.117	01:40	Volume Not Met
Breath Volume	0.523	01:40	
Air Blank	0.000	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 2	01:45	Subject Test Refused
Breath Volume	0.988	01:45	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: KOMAR JR

DATE/TIME OF TEST: 01/20/2024 20:24:21

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.035	20:19
Breath Volume	1.238	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.033	20:22
Breath Volume	1.535	20:22
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CARDILLO

DATE/TIME OF TEST: 01/26/2024 03:58:18

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.000	03:53
Breath Volume	2.023	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.007	03:56
Breath Volume	3.003	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MILLARD

DATE/TIME OF TEST: 01/26/2024 20:47:21

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.231	20:42
Breath Volume	2.753	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.237	20:45
Breath Volume	2.910	20:45
Air Blank	0.000	20:46
Control Test	0.078	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 02/16/2024 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TORRES,FELIX
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: PERDOMO GUTIERREZ

DATE/TIME OF TEST: 01/28/2024 06:12:01

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.080	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.185	06:07
Breath Volume	3.171	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample # 2	0.184	06:10
Breath Volume	3.351	06:10
Air Blank	0.000	06:11
Control Test	0.079	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:12

Data Download Date/Time: 02/16/2024 13:40