

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANTIZ HERNANDEZ

DATE/TIME OF TEST: 02/01/2024 01:57:40

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.165	01:52
Breath Volume	1.765	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.163	01:55
Breath Volume	2.128	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 02/01/2024 04:26:40

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.014	04:21
Breath Volume	3.007	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.014	04:25
Breath Volume	2.613	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURHAM, C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 02/01/2024 21:16:23

Diagnostics Check	OK	21:09
Air Blank	0.000	21:09
Control Test	0.081	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.131	21:11
Breath Volume	2.468	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.132	21:14
Breath Volume	2.152	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HALL II

DATE/TIME OF TEST: 02/02/2024 00:44:59

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.295	00:40
Breath Volume	2.566	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 2	0.302	00:43
Breath Volume	2.554	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CEDRECHI

DATE/TIME OF TEST: 02/02/2024 04:33:54

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.105	04:29
Breath Volume	1.488	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.103	04:32
Breath Volume	1.492	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: DILILLO JR

DATE/TIME OF TEST: 02/02/2024 04:54:01

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.102	04:49
Breath Volume	2.070	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.099	04:52
Breath Volume	2.695	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:54

Data Download Date/Time: 02/02/2024 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERRERA, A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MELENDEZ ADORNO

DATE/TIME OF TEST: 02/03/2024 05:37:26

Diagnostics Check	OK	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.000	05:33
Breath Volume	2.066	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.000	05:36
Breath Volume	2.910	05:36
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 03/06/2024 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 02/04/2024 04:43:07

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.081	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.112	04:38
Breath Volume	2.386	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.101	04:41
Breath Volume	1.199	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 03/06/2024 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: ROJAS LEYVA

DATE/TIME OF TEST: 02/04/2024 05:09:49

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.119	05:05
Breath Volume	1.722	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.119	05:08
Breath Volume	1.464	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 03/06/2024 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/05/2024 05:24:02

Diagnostics Check	OK	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.141	05:19
Breath Volume	2.789	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:21
Subject Sample # 2	0.140	05:22
Breath Volume	3.328	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:24

Data Download Date/Time: 03/06/2024 14:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 02/07/2024 03:55:15

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.082	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.105	03:50
Breath Volume	1.804	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.108	03:53
Breath Volume	1.500	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: NINIS

DATE/TIME OF TEST: 02/08/2024 02:45:53

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.148	02:41
Breath Volume	1.910	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.133	02:44
Breath Volume	1.652	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOWNE, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DOTSON

DATE/TIME OF TEST: 02/09/2024 00:35:00

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.148	00:30
Breath Volume	1.562	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.152	00:33
Breath Volume	1.398	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:35

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 02/09/2024 02:44:02

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.200	02:36	Volume Not Met
Breath Volume	1.003	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 2 0.176	02:42	Volume Not Met
Breath Volume	0.640	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PINEDA CHAVARRIA

DATE/TIME OF TEST: 02/09/2024 21:38:08

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.081	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.165	21:33
Breath Volume	1.851	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.157	21:36
Breath Volume	1.160	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 02/09/2024 23:51:12

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1 0.137	23:43	Volume Not Met
Breath Volume	1.031	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2 0.158	23:46	
Breath Volume	3.566	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 3 0.158	23:49	
Breath Volume	1.531	23:49	
Air Blank	0.000	23:50	
Control Test	0.080	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: OVINCY

DATE/TIME OF TEST: 02/10/2024 04:11:28

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.153	04:06
Breath Volume	1.699	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.142	04:09
Breath Volume	1.574	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 02/10/2024 20:03:52

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.081	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.166	19:58
Breath Volume	2.460	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.149	20:02
Breath Volume	1.558	20:02
Air Blank	0.000	20:02
Control Test	0.081	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: PINKSTON

DATE/TIME OF TEST: 02/10/2024 22:05:34

Diagnostics Check	OK	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.203	22:00
Breath Volume	1.316	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.206	22:03
Breath Volume	1.609	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/10/2024 23:01:27

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.249	22:56
Breath Volume	1.707	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.249	22:59
Breath Volume	2.285	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 02/11/2024 01:22:59

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.081	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1 0.104	01:14	Volume Not Met
Breath Volume	0.039	01:14	
Air Blank	0.000	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 2 0.118	01:17	
Breath Volume	1.375	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 3 0.121	01:21	
Breath Volume	1.707	01:21	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LOPEZ MARTINEZ

DATE/TIME OF TEST: 02/11/2024 03:19:19

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.121	03:14
Breath Volume	2.347	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:16
Subject Sample # 2	0.122	03:17
Breath Volume	2.800	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ILLES

DATE/TIME OF TEST: 02/11/2024 03:35:22

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.135	03:30
Breath Volume	2.757	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.123	03:33
Breath Volume	1.898	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FURNICE

DATE/TIME OF TEST: 02/11/2024 03:48:18

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON, N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: LARK

DATE/TIME OF TEST: 02/11/2024 04:57:59

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.081	04:55	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: NERI TOLENTINO

DATE/TIME OF TEST: 02/11/2024 22:59:31

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.232	22:54
Breath Volume	2.023	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.218	22:57
Breath Volume	1.390	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 02/12/2024 00:01:02

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.169	23:56
Breath Volume	2.597	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.163	23:59
Breath Volume	2.464	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: PEREZ RUIZ

DATE/TIME OF TEST: 02/12/2024 00:26:57

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.160	00:22
Breath Volume	1.164	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.162	00:25
Breath Volume	1.347	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 02/12/2024 01:00:31

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.012	00:56
Breath Volume	1.789	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.011	00:59
Breath Volume	1.777	00:59
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KIRK,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: MARTE

DATE/TIME OF TEST: 02/12/2024 01:44:12

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.200	01:39
Breath Volume	2.929	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.189	01:42
Breath Volume	1.269	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MOLITOR,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/12/2024 01:57:28

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.099	01:53
Breath Volume	2.035	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.100	01:55
Breath Volume	2.250	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/06/2024 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 02/12/2024 03:20:09

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.145	03:15
Breath Volume	1.625	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.141	03:18
Breath Volume	1.773	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 02/12/2024 04:20:20

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.081	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: PEREZ-RODRIGUEZ

DATE/TIME OF TEST: 02/16/2024 00:26:24

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.079	00:21
Breath Volume	2.777	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.081	00:24
Breath Volume	2.417	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/16/2024 01:06:41

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.127	01:02
Breath Volume	1.218	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.122	01:05
Breath Volume	1.140	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: FRIEDE

DATE/TIME OF TEST: 02/16/2024 05:39:58

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:37	
Control Test	0.081	05:37	
Air Blank	0.000	05:37	
Subject Sample	# 1	05:38	Subject Test Refused
Breath Volume	0.000	05:38	
Air Blank	0.000	05:39	
Control Test	0.081	05:39	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:51

SUBJECT'S LAST NAME: SUSHCHIK DATE/TIME OF TEST: 02/16/2024 07:18:56

Diagnostics Check	OK	07:11
Air Blank	0.000	07:11
Control Test	0.081	07:12
Air Blank	0.000	07:12
Subject Sample # 1	0.064	07:14
Breath Volume	1.917	07:14
Air Blank	0.000	07:14
Air Blank	0.000	07:16
Subject Sample # 2	0.055	07:17
Breath Volume	1.292	07:17
Air Blank	0.000	07:18
Control Test	0.080	07:18
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 02/16/2024 22:58:24

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.130	22:53
Breath Volume	1.761	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.128	22:56
Breath Volume	1.851	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAZAQ

DATE/TIME OF TEST: 02/16/2024 23:27:17

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.115	23:22
Breath Volume	3.804	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.106	23:25
Breath Volume	2.921	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: VAZQUEZ PEREZ

DATE/TIME OF TEST: 02/17/2024 01:25:56

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.194	01:21
Breath Volume	2.546	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.195	01:24
Breath Volume	2.621	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: COMM VEHIC
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 02/17/2024 03:23:16

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.094	03:18
Breath Volume	1.816	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.091	03:21
Breath Volume	1.660	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS, L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/17/2024 03:48:21

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.084	03:43
Breath Volume	1.406	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.090	03:46
Breath Volume	1.343	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ORDONEZ FELIPE

DATE/TIME OF TEST: 02/18/2024 02:27:33

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.168	02:23
Breath Volume	1.593	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.169	02:25
Breath Volume	1.835	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 02/19/2024 02:06:26

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.081	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.267	02:01
Breath Volume	2.082	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.262	02:04
Breath Volume	1.800	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: RODRIGUEZ SANTANA

DATE/TIME OF TEST: 02/20/2024 03:07:35

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.082	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.108	03:02
Breath Volume	1.394	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.107	03:05
Breath Volume	1.550	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CHIC TEBALAN

DATE/TIME OF TEST: 02/21/2024 00:45:12

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.204	00:37
Breath Volume	3.234	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.182	00:40
Breath Volume	1.917	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 3	0.195	00:43
Breath Volume	2.054	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 02/21/2024 03:15:54

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.081	03:07	
Air Blank	0.000	03:08	
Subject Sample	# 1 0.163	03:11	Volume Not Met
Breath Volume	0.019	03:11	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 2	03:14	Subject Test Refused
Breath Volume	0.000	03:14	
Air Blank	0.000	03:14	
Control Test	0.079	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: CRUZ LOPEZ

DATE/TIME OF TEST: 02/21/2024 05:29:53

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.081	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.158	05:24
Breath Volume	2.121	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.154	05:28
Breath Volume	2.246	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ALEJO

DATE/TIME OF TEST: 02/21/2024 22:07:42

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:06	
Control Test	0.081	22:07	
Air Blank	0.000	22:07	
Diagnostics Check	OK	22:07	

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: PEDRO RAMIREZ

DATE/TIME OF TEST: 02/22/2024 00:54:37

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.206	00:50
Breath Volume	2.476	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.200	00:52
Breath Volume	2.820	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FULLERTON

DATE/TIME OF TEST: 02/22/2024 01:04:44

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.140	01:00
Breath Volume	2.511	01:00
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.140	01:03
Breath Volume	2.179	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BLANDING

DATE/TIME OF TEST: 02/22/2024 03:21:49

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:17	No Sample Provided
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 03/06/2024 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: CHARLTON

DATE/TIME OF TEST: 02/23/2024 03:53:28

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.178	03:48
Breath Volume	1.433	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.178	03:51
Breath Volume	1.101	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/23/2024 23:29:39

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.000	23:25
Breath Volume	2.949	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.000	23:28
Breath Volume	3.195	23:28
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: MOLL

DATE/TIME OF TEST: 02/24/2024 01:20:35

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.195	01:15
Breath Volume	2.207	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.198	01:18
Breath Volume	2.562	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THEODORE, P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ANTONIO-MATEO

DATE/TIME OF TEST: 02/24/2024 02:44:00

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.286	02:39
Breath Volume	2.761	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.285	02:42
Breath Volume	1.488	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 02/24/2024 08:01:03

Diagnostics Check	OK	07:53
Air Blank	0.000	07:53
Control Test	0.080	07:54
Air Blank	0.000	07:54
Subject Sample # 1	0.096	07:55
Breath Volume	1.359	07:55
Air Blank	0.000	07:56
Air Blank	0.000	07:58
Subject Sample # 2	0.104	07:59
Breath Volume	2.093	07:59
Air Blank	0.000	08:00
Control Test	0.080	08:00
Air Blank	0.000	08:00
Diagnostics Check	OK	08:01

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CHE BUTZ

DATE/TIME OF TEST: 02/25/2024 00:53:43

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.215	00:48
Breath Volume	1.750	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.208	00:52
Breath Volume	1.546	00:52
Air Blank	0.000	00:52
Control Test	0.078	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THEODORE,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORSE

DATE/TIME OF TEST: 02/25/2024 02:20:46

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.212	02:13
Breath Volume	2.769	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.183	02:16
Breath Volume	1.265	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 3	0.204	02:19
Breath Volume	2.312	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILKINSON,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CAAL-CACAO

DATE/TIME OF TEST: 02/25/2024 03:42:32

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.145	03:37
Breath Volume	2.519	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.151	03:40
Breath Volume	2.753	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 02/25/2024 04:18:44

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.085	04:14
Breath Volume	2.015	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.086	04:17
Breath Volume	1.785	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MORALES MORALES

DATE/TIME OF TEST: 02/25/2024 04:41:49

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.132	04:36
Breath Volume	1.917	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.127	04:40
Breath Volume	2.199	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: NIEVES ORTIZ

DATE/TIME OF TEST: 02/25/2024 05:12:25

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: MAZZOTTA

DATE/TIME OF TEST: 02/25/2024 06:20:36

Diagnostics Check	OK	06:13
Air Blank	0.000	06:14
Control Test	0.080	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.069	06:16
Breath Volume	1.445	06:16
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.063	06:19
Breath Volume	1.222	06:19
Air Blank	0.000	06:19
Control Test	0.080	06:19
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/25/2024 20:34:42

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.081	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1 0.140	20:27	
Breath Volume	3.222	20:27	
Air Blank	0.000	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 2 0.119	20:30	
Breath Volume	1.667	20:30	
Air Blank	0.000	20:31	
Air Blank	0.000	20:33	
Subject Sample	# 3	20:33	Subject Test Refused
Breath Volume	0.000	20:33	
Air Blank	0.000	20:33	
Control Test	0.081	20:34	
Air Blank	0.000	20:34	
Diagnostics Check	OK	20:34	

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/25/2024 20:51:30

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.130	20:46
Breath Volume	3.968	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.127	20:49
Breath Volume	1.886	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: AGUILAR CHINCHILLA

DATE/TIME OF TEST: 02/25/2024 23:04:37

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.145	22:59
Breath Volume	1.730	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.145	23:02
Breath Volume	1.945	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KIRK,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 02/25/2024 23:26:25

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.105	23:21
Breath Volume	2.511	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.098	23:24
Breath Volume	1.492	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 02/26/2024 02:16:45

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/26/2024 02:48:38

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SOREY

DATE/TIME OF TEST: 02/28/2024 23:11:52

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.080	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.203	23:06
Breath Volume	1.429	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.194	23:10
Breath Volume	1.316	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 03/06/2024 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HIRSCHI,MATTHEW
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: LAROCHE

DATE/TIME OF TEST: 02/27/2024 17:44:30

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:43	
Control Test	0.082	17:43	
Air Blank		17:44	RFI Detect
Air Blank	0.000	17:44	

Data Download Date/Time: 03/16/2024 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HIRSCHI,MATTHEW
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: LAROCHE

DATE/TIME OF TEST: 02/27/2024 17:54:11

Diagnostics Check	OK	17:46
Air Blank	0.000	17:47
Control Test	0.081	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.036	17:48
Breath Volume	2.132	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.032	17:52
Breath Volume	1.898	17:52
Air Blank	0.000	17:53
Control Test	0.081	17:53
Air Blank	0.000	17:54
Diagnostics Check	OK	17:54

Data Download Date/Time: 03/16/2024 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BLAKE,PAIGE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JOHNSEN

DATE/TIME OF TEST: 02/28/2024 01:34:05

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.082	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.082	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 03/16/2024 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CARTER, BRYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:41

SUBJECT'S LAST NAME: STAINA

DATE/TIME OF TEST: 02/28/2024 11:08:23

Diagnostics Check	OK	11:01
Air Blank	0.000	11:02
Control Test	0.082	11:02
Air Blank	0.000	11:03
Subject Sample # 1	0.281	11:03
Breath Volume	1.402	11:03
Air Blank	0.000	11:04
Air Blank	0.000	11:06
Subject Sample # 2	0.272	11:06
Breath Volume	1.421	11:06
Air Blank	0.000	11:07
Control Test	0.080	11:07
Air Blank	0.000	11:08
Diagnostics Check	OK	11:08

Data Download Date/Time: 03/16/2024 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINEY,CRISTA,N
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TOWNS,ROBIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: MANTILLA LOPEZ

DATE/TIME OF TEST: 02/29/2024 02:02:02

Diagnostics Check	OK	01:59	
Air Blank	0.000	01:59	
Control Test	0.082	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 03/16/2024 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SANGER,MOOKIE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/01/2024 21:33:58

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.079	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Control Test	0.079	21:33	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 02/16/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASAS,JAVIER,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/02/2024 16:35:05

Diagnostics Check	OK	16:28
Air Blank	0.000	16:29
Control Test	0.079	16:29
Air Blank	0.000	16:29
Subject Sample # 1	0.134	16:30
Breath Volume	1.980	16:30
Air Blank	0.000	16:31
Air Blank	0.000	16:32
Subject Sample # 2	0.133	16:33
Breath Volume	2.070	16:33
Air Blank	0.000	16:34
Control Test	0.080	16:34
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 02/16/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIMONEY,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SANGER,MOOKIE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: ASHE

DATE/TIME OF TEST: 02/02/2024 20:24:56

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:20	
Control Test	0.080	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.035	20:23	
Air Blank	0.000	20:24	
Control Test	0.081	20:24	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 02/16/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SANGER,MOOKIE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 01/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ANTONIO LEON

DATE/TIME OF TEST: 02/09/2024 23:58:50

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.179	23:53
Breath Volume	2.953	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.180	23:57
Breath Volume	3.148	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 02/16/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: QUEBRADA

DATE/TIME OF TEST: 02/18/2024 03:08:29

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 03/26/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: LAUBHEIMER

DATE/TIME OF TEST: 02/25/2024 19:26:36

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.079	19:24	
Air Blank	0.000	19:25	
Subject Sample	# 1	19:25	Subject Test Refused
Breath Volume	0.000	19:25	
Air Blank	0.000	19:25	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 03/26/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MOREIRA MELO

DATE/TIME OF TEST: 02/26/2024 04:00:59

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.166	03:55
Breath Volume	1.945	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.164	03:59
Breath Volume	1.558	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 03/26/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TURNER,JACOB
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BRANCHDOR

DATE/TIME OF TEST: 02/29/2024 01:40:10

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 03/26/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 02/07/2024 03:02:01

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.207	02:49	No .020 agreement
Breath Volume	1.394	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.188	02:55	Volume Not Met
Breath Volume	0.519	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 3 0.183	03:00	No .020 agreement
Breath Volume	1.453	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 02/12/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 02/07/2024 03:15:24

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.187	03:10
Breath Volume	1.675	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.197	03:13
Breath Volume	1.410	03:13
Air Blank	0.000	03:14
Control Test	0.076	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 02/12/2024 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIEGLER,KYLE,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BRINTON

DATE/TIME OF TEST: 02/09/2024 22:41:17

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1 0.244	22:33	No .020 agreement
Breath Volume	2.343	22:33	
Air Blank	0.000	22:33	
Air Blank	0.000	22:35	
Subject Sample	# 2 0.287	22:36	No .020 agreement
Breath Volume	2.324	22:36	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample	# 3 0.266	22:39	No .020 agreement
Breath Volume	2.226	22:39	
Air Blank	0.000	22:40	
Control Test	0.083	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 02/12/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ROTOLO

DATE/TIME OF TEST: 02/11/2024 00:09:33

Diagnostics Check	OK	00:08	
Air Blank		00:09	RFI Detect
Air Blank	0.000	00:09	

Data Download Date/Time: 02/12/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40258895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: ROTOLO

DATE/TIME OF TEST: 02/11/2024 00:17:30

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.000	00:12
Breath Volume	3.808	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.000	00:16
Breath Volume	3.058	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 02/12/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICODEMO,PHIL,
BREATH TEST OPERATOR AGENCY: FDLE
ARRESTING OFFICER: NICODEMO,PHIL
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 02/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: LICENSE

DATE/TIME OF TEST: 02/22/2024 12:09:46

Diagnostics Check	OK	12:04
Air Blank	0.000	12:04
Control Test	0.080	12:04
Air Blank	0.000	12:05
Subject Sample # 1	0.000	12:05
Breath Volume	2.402	12:05
Air Blank	0.000	12:06
Air Blank	0.000	12:08
Subject Sample # 2	0.000	12:08
Breath Volume	2.152	12:08
Air Blank	0.000	12:08
Control Test	0.080	12:09
Air Blank	0.000	12:09
Diagnostics Check	OK	12:09

Data Download Date/Time: 02/22/2024 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,QUINN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: MORIARTY,QUINN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HARTMANN

DATE/TIME OF TEST: 02/02/2024 20:46:25

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.080	20:44	
Air Blank	0.000	20:44	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:45	
Control Test	0.079	20:45	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:46	

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 02/03/2024 03:24:25

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.134	03:19
Breath Volume	2.058	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.134	03:22
Breath Volume	1.742	03:22
Air Blank	0.000	03:23
Control Test	0.077	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUSSEAU,DMITRY,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/03/2024 20:40:52

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:38	
Control Test	0.079	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:39	
Control Test	0.079	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:40	

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUSSEAU,DMITRY,A
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 02/05/2024 01:40:02

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:33	No Sample Provided
Breath Volume	0.000	01:33	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/06/2024 02:45:39

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.131	02:40
Breath Volume	1.515	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.130	02:43
Breath Volume	1.285	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 02/07/2024 17:42:47

Diagnostics Check	OK	17:36
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:37
Subject Sample # 1	0.089	17:38
Breath Volume	2.476	17:38
Air Blank	0.000	17:39
Air Blank	0.000	17:40
Subject Sample # 2	0.084	17:41
Breath Volume	2.089	17:41
Air Blank	0.000	17:41
Control Test	0.077	17:42
Air Blank	0.000	17:42
Diagnostics Check	OK	17:42

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: GALYAMOVA

DATE/TIME OF TEST: 02/09/2024 18:48:03

Diagnostics Check	OK	18:44	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:45	
Subject Sample	# 1	18:46	Subject Test Refused
Breath Volume	0.000	18:46	
Air Blank	0.000	18:47	
Control Test	0.077	18:47	
Air Blank	0.000	18:47	
Diagnostics Check	OK	18:48	

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BRINTON

DATE/TIME OF TEST: 02/09/2024 23:56:45

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.079	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.277	23:52
Breath Volume	1.351	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.271	23:55
Breath Volume	1.898	23:55
Air Blank	0.000	23:55
Control Test	0.078	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASON, BRADEN, M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN, JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: MULCAHY

DATE/TIME OF TEST: 02/10/2024 01:11:37

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.214	01:06
Breath Volume	1.433	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.214	01:09
Breath Volume	1.621	01:09
Air Blank	0.000	01:10
Control Test	0.077	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CASTELLANOS AVELAR

DATE/TIME OF TEST: 02/11/2024 00:25:52

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.172	00:21
Breath Volume	1.613	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.170	00:24
Breath Volume	2.027	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: RODRIGUEZ,MIGUEL
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: SANCHEZ GOMEZ

DATE/TIME OF TEST: 02/11/2024 03:40:36

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.184	03:35
Breath Volume	2.746	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.182	03:38
Breath Volume	3.031	03:38
Air Blank	0.000	03:39
Control Test	0.077	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: DECICCO

DATE/TIME OF TEST: 02/11/2024 03:51:00

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.171	03:46
Breath Volume	1.480	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.167	03:49
Breath Volume	1.562	03:49
Air Blank	0.000	03:50
Control Test	0.077	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:51

Data Download Date/Time: 02/12/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACKWELL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 02/15/2024 00:04:56

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.201	23:58
Breath Volume	2.019	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.209	00:03
Breath Volume	1.320	00:03
Air Blank	0.000	00:03
Control Test	0.078	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 03/11/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 02/26/2024 04:37:00

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.143	04:32
Breath Volume	2.648	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.155	04:35
Breath Volume	2.226	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

Data Download Date/Time: 03/11/2024 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: NOVAK,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: O'DEA

DATE/TIME OF TEST: 02/29/2024 02:37:17

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.173	02:32
Breath Volume	2.664	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.153	02:35
Breath Volume	2.480	02:35
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/11/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MCMAHON

DATE/TIME OF TEST: 02/04/2024 02:17:34

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.205	02:12
Breath Volume	1.199	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.214	02:15
Breath Volume	2.085	02:15
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 02/04/2024 21:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 02/17/2024 02:32:23

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.083	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.120	02:27
Breath Volume	2.507	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.116	02:30
Breath Volume	2.292	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 02/21/2024 23:26:45

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.083	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.171	23:22
Breath Volume	3.828	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.162	23:25
Breath Volume	4.097	23:25
Air Blank	0.000	23:25
Control Test	0.082	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DIMARIA

DATE/TIME OF TEST: 02/27/2024 01:56:29

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:53	
Subject Sample # 1	0.155	01:54	
Breath Volume	2.425	01:54	
Air Blank		01:55	RFI Detect
Air Blank	0.000	01:55	
Control Test	0.083	01:55	
Air Blank		01:56	RFI Detect
Air Blank	0.000	01:56	

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DIMARIA

DATE/TIME OF TEST: 02/27/2024 02:07:47

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.083	01:59	
Air Blank	0.000	01:59	
Subject Sample # 1	0.170	02:00	
Breath Volume	1.886	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample # 2	0.141	02:03	
Breath Volume	1.585	02:03	
Air Blank	0.000	02:04	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	
Control Test	0.082	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DIMARIA

DATE/TIME OF TEST: 02/27/2024 02:13:02

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.083	02:10	
Air Blank	0.000	02:10	
Subject Sample	# 1 0.164	02:11	
Breath Volume	2.644	02:11	
Air Blank		02:11	RFI Detect
Air Blank	0.000	02:12	
Control Test	0.083	02:12	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: DIMARIA

DATE/TIME OF TEST: 02/27/2024 02:37:02

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.082	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.153	02:32
Breath Volume	2.421	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.151	02:35
Breath Volume	2.878	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/03/2024 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HAUGHEY

DATE/TIME OF TEST: 02/03/2024 05:13:34

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:11	
Control Test	0.079	05:11	
Air Blank	0.000	05:12	
Subject Sample	# 1	05:12	Subject Test Refused
Breath Volume	0.000	05:12	
Air Blank	0.000	05:12	
Control Test	0.079	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 02/13/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: FERRIGINE,JOSEPH
ARRESTING OFFICER AGENCY: Longboat Key PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LETTORALE JR

DATE/TIME OF TEST: 02/04/2024 22:43:46

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.078	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:42	
Control Test	0.078	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 02/13/2024 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETTS,JOSHUA,
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: MATHIS,MIKE
ARRESTING OFFICER AGENCY: Longboat Key PD
LAST AGENCY INSPECTION: 02/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: AGUILAR QUIROZ

DATE/TIME OF TEST: 02/17/2024 04:20:34

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:19	
Control Test	0.079	04:19	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 03/16/2024 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MORADEL

DATE/TIME OF TEST: 02/23/2024 22:59:23

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Subject Sample	# 1 0.089	22:49	Volume Not Met
Breath Volume	0.117	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 2 0.086	22:53	
Breath Volume	1.207	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 3 0.088	22:57	
Breath Volume	1.921	22:57	
Air Blank	0.000	22:58	
Control Test	0.074	22:58	Control Outside Tolerance
Air Blank	0.000	22:59	

Data Download Date/Time: 03/11/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MORADEL

DATE/TIME OF TEST: 02/23/2024 23:09:15

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.078	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.091	23:04
Breath Volume	1.378	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:06
Subject Sample # 2	0.087	23:07
Breath Volume	1.554	23:07
Air Blank	0.000	23:08
Control Test	0.076	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 03/11/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 02/24/2024 01:32:05

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.125	01:27
Breath Volume	2.636	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.122	01:30
Breath Volume	2.437	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE, GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE, GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: NEUMANN

DATE/TIME OF TEST: 02/24/2024 03:13:07

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.130	03:08
Breath Volume	1.484	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.130	03:11
Breath Volume	1.835	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GRIECO

DATE/TIME OF TEST: 02/26/2024 02:05:51

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.162	02:00	Volume Not Met
Breath Volume	0.015	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2	02:04	Subject Test Refused
Breath Volume	0.171	02:04	
Air Blank	0.000	02:04	
Control Test	0.076	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 02/26/2024 04:11:15

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.209	04:06
Breath Volume	2.203	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.199	04:09
Breath Volume	2.347	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: MAGOWAN

DATE/TIME OF TEST: 02/28/2024 18:44:53

Diagnostics Check	OK	18:38
Air Blank	0.000	18:39
Control Test	0.080	18:39
Air Blank	0.000	18:40
Subject Sample # 1	0.072	18:40
Breath Volume	1.562	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.068	18:43
Breath Volume	1.898	18:43
Air Blank	0.000	18:43
Control Test	0.077	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 02/29/2024 00:19:11

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.044	00:14
Breath Volume	4.085	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.044	00:17
Breath Volume	4.097	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 03/11/2024 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 02/11/2024 22:02:26

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:01	
Control Test	0.077	22:01	RFI Detect
Air Blank	0.000	22:02	

Data Download Date/Time: 03/05/2024 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,JR,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ELMORE,TT
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 02/11/2024 22:10:31

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.093	22:06
Breath Volume	2.652	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.092	22:08
Breath Volume	1.976	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 03/05/2024 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/27/2024 02:35:15

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.154	02:30
Breath Volume	2.773	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.151	02:33
Breath Volume	1.972	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 03/05/2024 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 02/06/2024 01:29:56

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.248	01:25
Breath Volume	2.398	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.233	01:28
Breath Volume	1.714	01:28
Air Blank	0.000	01:28
Control Test	0.078	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 02/12/2024 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VAZQUEZ PAIS

DATE/TIME OF TEST: 02/07/2024 02:58:33

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank		02:58	RFI Detect
Air Blank	0.000	02:58	

Data Download Date/Time: 02/12/2024 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: VAZQUEZ PAIS

DATE/TIME OF TEST: 02/07/2024 03:07:11

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.146	03:02
Breath Volume	1.601	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.136	03:05
Breath Volume	1.718	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 02/12/2024 09:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024023571
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: ADMIN
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/27/2024 00:49:21

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:45
Breath Volume	3.718	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	3.585	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 03/07/2024 04:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GARCIA FRANCO

DATE/TIME OF TEST: 02/04/2024 01:31:59

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.197	01:27
Breath Volume	2.546	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.203	01:30
Breath Volume	2.453	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 02/07/2024 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DESROSIER

DATE/TIME OF TEST: 02/04/2024 04:34:14

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.300	04:28
Breath Volume	1.382	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.294	04:32
Breath Volume	1.394	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 02/07/2024 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BROMLEY

DATE/TIME OF TEST: 02/05/2024 06:32:08

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:26	
Control Test	0.081	06:27	
Air Blank	0.000	06:27	
Subject Sample	# 1 0.169	06:27	
Breath Volume	1.414	06:27	
Air Blank	0.000	06:28	
Air Blank	0.000	06:30	
Subject Sample	# 2 0.164	06:30	
Breath Volume	1.539	06:30	
Air Blank	0.000	06:31	
		06:31	RFI Detect
Air Blank	0.000	06:32	

Data Download Date/Time: 02/07/2024 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/06/2024 03:58:37

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.050	03:54
Breath Volume	1.390	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.049	03:57
Breath Volume	2.640	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 02/07/2024 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: LAWS

DATE/TIME OF TEST: 02/02/2024 00:51:27

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.083	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.124	00:46
Breath Volume	1.812	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.121	00:49
Breath Volume	2.035	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/28/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/15/2024 01:35:06

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.147	01:30
Breath Volume	1.539	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.143	01:33
Breath Volume	1.406	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 02/28/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GUTSHELL,C
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: REAGAN

DATE/TIME OF TEST: 02/16/2024 22:19:01

Diagnostics Check	OK	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.077	22:13
Breath Volume	2.175	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.079	22:17
Breath Volume	2.078	22:17
Air Blank	0.000	22:18
Control Test	0.084	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:19

Data Download Date/Time: 02/28/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: MUNROE

DATE/TIME OF TEST: 02/23/2024 21:50:08

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.082	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1 0.000	21:43	Volume Not Met
Breath Volume	0.703	21:43	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 2 0.000	21:48	Volume Not Met
Breath Volume	0.277	21:48	
Air Blank	0.000	21:49	
Control Test	0.081	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 02/28/2024 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WARREN,Z
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: REDMON

DATE/TIME OF TEST: 02/24/2024 23:44:03

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.082	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.133	23:39
Breath Volume	2.421	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.147	23:42
Breath Volume	1.730	23:42
Air Blank	0.000	23:43
Control Test	0.083	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

Data Download Date/Time: 02/28/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 02/13/2024 01:45:59

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.132	01:41
Breath Volume	1.777	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.133	01:44
Breath Volume	1.847	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 02/27/2024 21:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MEHU

DATE/TIME OF TEST: 02/24/2024 02:40:21

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.184	02:35
Breath Volume	1.140	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.180	02:38
Breath Volume	1.347	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 02/27/2024 21:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PEREYRA

DATE/TIME OF TEST: 02/24/2024 22:59:49

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.149	22:55
Breath Volume	1.863	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.150	22:58
Breath Volume	2.765	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 02/27/2024 21:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUC,NORBENS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/09/2024 02:19:56

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.197	02:15
Breath Volume	2.753	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.193	02:18
Breath Volume	2.136	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 02/09/2024 05:26:35

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.139	05:21
Breath Volume	2.125	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.137	05:24
Breath Volume	1.644	05:24
Air Blank	0.000	05:25
Control Test	0.081	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KAPLAN,SAMANTHA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: METAYER

DATE/TIME OF TEST: 02/09/2024 06:18:01

Diagnostics Check	OK	06:11
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.108	06:13
Breath Volume	1.593	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.101	06:16
Breath Volume	1.167	06:16
Air Blank	0.000	06:17
Control Test	0.081	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:18

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMERENCIANO,MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: REYES ARGUETA

DATE/TIME OF TEST: 02/11/2024 22:02:10

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.089	21:57
Breath Volume	1.914	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.082	22:00
Breath Volume	1.257	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/18/2024 02:20:50

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.221	02:16
Breath Volume	2.566	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.218	02:19
Breath Volume	1.652	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 02/20/2024 02:42:32

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.152	02:38
Breath Volume	1.703	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.156	02:40
Breath Volume	1.531	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: REINOSO OSPINA

DATE/TIME OF TEST: 02/20/2024 04:42:19

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:27	
Control Test	0.081	04:27	
Air Blank	0.000	04:28	
Subject Sample # 1	0.183	04:30	
Breath Volume	1.222	04:30	
Air Blank	0.000	04:30	
Air Blank	0.000	04:32	
Subject Sample # 2	0.130	04:35	Volume Not Met
Breath Volume	0.175	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample # 3	0.171	04:40	
Breath Volume	1.156	04:40	
Air Blank	0.000	04:41	
Control Test	0.079	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 02/26/2024 05:12:33

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.082	05:01	
Air Blank	0.000	05:02	
Subject Sample	# 1 0.215	05:05	Volume Not Met
Breath Volume	0.148	05:05	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 2 0.290	05:08	
Breath Volume	1.753	05:08	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 3 0.284	05:10	
Breath Volume	1.125	05:10	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 03/06/2024 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JERZY

DATE/TIME OF TEST: 02/09/2024 02:18:45

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.185	02:13
Breath Volume	1.613	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.181	02:17
Breath Volume	1.343	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 02/27/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MADERO

DATE/TIME OF TEST: 02/17/2024 01:12:25

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.170	01:07
Breath Volume	1.687	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.168	01:10
Breath Volume	1.457	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/27/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CESAR

DATE/TIME OF TEST: 02/21/2024 04:29:35

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.206	04:25
Breath Volume	1.167	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.198	04:27
Breath Volume	1.710	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 02/27/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALLEN,JUSTIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: NO DL

DATE/TIME OF TEST: 02/03/2024 20:35:44

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.202	20:30
Breath Volume	2.824	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.195	20:34
Breath Volume	2.078	20:34
Air Blank	0.000	20:34
Control Test	0.079	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 03/08/2024 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,VERONICA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MARTINEZ CARDONA

DATE/TIME OF TEST: 02/18/2024 00:16:13

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.109	00:11	Volume Not Met
Breath Volume	0.601	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.079	00:15	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 03/08/2024 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007078

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND, WILLIAM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SPEICH, R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 02/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/21/2024 04:53:43

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:48
Subject Sample # 1	0.000	04:48
Breath Volume	1.824	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.000	04:52
Breath Volume	1.539	04:52
Air Blank	0.000	04:52
Control Test	0.082	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 03/08/2024 18:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KING,E
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 02/09/2024 04:28:40

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.145	04:23
Breath Volume	1.808	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.138	04:27
Breath Volume	1.402	04:27
Air Blank	0.000	04:27
Control Test	0.080	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 02/25/2024 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VETRI,OFC
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 02/13/2024 20:37:51

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.080	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.000	20:33
Breath Volume	1.191	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.000	20:36
Breath Volume	1.425	20:36
Air Blank	0.000	20:36
Control Test	0.080	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 02/25/2024 19:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08622080A1
DRY GAS EXPIRY DATE: 06/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 02/19/2024 01:12:59

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.120	01:05
Breath Volume	1.847	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.111	01:11
Breath Volume	1.820	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/25/2024 19:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMPARO,AURIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: MATIAS MAZARIEGOS

DATE/TIME OF TEST: 02/09/2024 19:42:33

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.209	19:37
Breath Volume	2.898	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.212	19:40
Breath Volume	2.914	19:40
Air Blank	0.000	19:41
Control Test	0.081	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARNECO,TAYLOR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: POLEO SEVERINO

DATE/TIME OF TEST: 02/10/2024 02:20:29

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.147	02:15
Breath Volume	1.437	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.149	02:18
Breath Volume	1.378	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: CACERES

DATE/TIME OF TEST: 02/10/2024 03:03:17

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.220	02:58
Breath Volume	3.003	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.224	03:01
Breath Volume	2.750	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOOLLEY,KYLE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 02/10/2024 04:05:45

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.081	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 2	04:03	Volume Not Met
Breath Volume	0.308	04:03	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOOLLEY,KYLE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 02/10/2024 04:16:24

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.154	04:11
Breath Volume	1.511	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.153	04:14
Breath Volume	1.199	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMERENCIANO,MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 02/10/2024 16:57:23

Diagnostics Check	OK	16:49
Air Blank	0.000	16:50
Control Test	0.081	16:50
Air Blank	0.000	16:51
Subject Sample # 1	0.000	16:52
Breath Volume	1.730	16:52
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:55
Breath Volume	1.257	16:55
Air Blank	0.000	16:56
Control Test	0.081	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/11/2024 02:23:22

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.000	02:18
Breath Volume	3.578	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.000	02:21
Breath Volume	3.035	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: CORREA MOLINA

DATE/TIME OF TEST: 02/11/2024 09:18:11

Diagnostics Check	OK	09:11
Air Blank	0.000	09:11
Control Test	0.080	09:12
Air Blank	0.000	09:12
Subject Sample # 1	0.137	09:13
Breath Volume	1.464	09:13
Air Blank	0.000	09:14
Air Blank	0.000	09:15
Subject Sample # 2	0.137	09:16
Breath Volume	1.671	09:16
Air Blank	0.000	09:17
Control Test	0.082	09:17
Air Blank	0.000	09:18
Diagnostics Check	OK	09:18

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: ARIN

DATE/TIME OF TEST: 02/12/2024 20:23:56

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.081	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:19
Breath Volume	3.460	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	3.695	20:22
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAJETE,LUIS
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MARIN RUIZ

DATE/TIME OF TEST: 02/13/2024 04:06:49

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.133	04:02
Breath Volume	3.207	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.133	04:05
Breath Volume	2.527	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MATAMOROS VILLANUEVA

DATE/TIME OF TEST: 02/15/2024 01:18:09

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.081	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1 0.128	01:10	Volume Not Met
Breath Volume	0.027	01:10	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample	# 2 0.150	01:16	Volume Not Met
Breath Volume	0.140	01:16	
Air Blank	0.000	01:17	
Control Test	0.080	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONZALEZ,ARIEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 02/17/2024 03:26:28

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.194	03:21
Breath Volume	2.097	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.197	03:24
Breath Volume	2.425	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/17/2024 03:47:20

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.219	03:42
Breath Volume	1.332	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.210	03:45
Breath Volume	1.109	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/06/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SANTIAGO,LEONEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: CAMPOS RIOS

DATE/TIME OF TEST: 02/19/2024 04:27:01

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.165	04:22
Breath Volume	2.945	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.158	04:25
Breath Volume	2.332	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIS,JASHAUNIE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: SANTIAGO RODRIGUEZ

DATE/TIME OF TEST: 02/19/2024 18:10:18

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.080	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.227	18:05
Breath Volume	1.210	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:07
Subject Sample # 2	0.208	18:08
Breath Volume	1.343	18:08
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 02/20/2024 23:40:14

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.082	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RYAN,JASON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 02/24/2024 09:10:56

Diagnostics Check	OK	09:03
Air Blank	0.000	09:04
Control Test	0.081	09:04
Air Blank	0.000	09:05
Subject Sample # 1	0.148	09:05
Breath Volume	1.976	09:05
Air Blank	0.000	09:06
Air Blank	0.000	09:08
Subject Sample # 2	0.153	09:09
Breath Volume	2.562	09:09
Air Blank	0.000	09:09
Control Test	0.082	09:10
Air Blank	0.000	09:10
Diagnostics Check	OK	09:10

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 02/25/2024 04:51:59

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.140	04:47
Breath Volume	1.375	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:49
Subject Sample # 2	0.142	04:50
Breath Volume	1.394	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 02/29/2024 02:19:03

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.184	02:14
Breath Volume	2.777	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.182	02:17
Breath Volume	2.523	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 02/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/29/2024 23:56:58

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.082	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.282	23:52
Breath Volume	2.050	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.281	23:55
Breath Volume	3.585	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 03/06/2024 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SIMMONS,BRYAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 02/01/2024 02:28:31

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.083	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.136	02:23
Breath Volume	2.882	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.144	02:26
Breath Volume	2.714	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 02/29/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,HENDRICK,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COLLINS,CHASE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 08:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 02/02/2024 08:50:16

Diagnostics Check	OK	08:44
Air Blank	0.000	08:44
Control Test	0.082	08:45
Air Blank	0.000	08:45
Subject Sample # 1	0.000	08:46
Breath Volume	1.402	08:46
Air Blank	0.000	08:46
Air Blank	0.000	08:48
Subject Sample # 2	0.000	08:48
Breath Volume	1.125	08:48
Air Blank	0.000	08:49
Control Test	0.082	08:49
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 02/29/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 02/03/2024 03:25:34

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.132	03:20
Breath Volume	1.343	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.140	03:23
Breath Volume	1.558	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HASWELL,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GOMEZ CASTILLO

DATE/TIME OF TEST: 02/04/2024 01:35:52

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.019	01:31
Breath Volume	3.519	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.020	01:34
Breath Volume	3.535	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HASWELL,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 02/04/2024 20:01:35

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.082	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.000	19:57
Breath Volume	2.324	19:57
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.000	20:00
Breath Volume	2.253	20:00
Air Blank	0.000	20:00
Control Test	0.082	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY-IVERSEN, MIYA, K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SIPHER, VICTORIA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 02/08/2024 00:51:50

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.217	00:47
Breath Volume	2.503	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.226	00:50
Breath Volume	3.015	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 23:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 02/09/2024 23:38:31

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.085	23:33
Breath Volume	1.757	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.087	23:36
Breath Volume	2.257	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ALDAY GOMEZ

DATE/TIME OF TEST: 02/10/2024 03:55:01

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.133	03:50
Breath Volume	2.226	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.125	03:53
Breath Volume	1.406	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: TEBBS

DATE/TIME OF TEST: 02/10/2024 19:41:55

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.080	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.161	19:35
Breath Volume	2.738	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.156	19:40
Breath Volume	1.621	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 23:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/11/2024 00:06:02

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:01
Breath Volume	3.738	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:04
Breath Volume	3.542	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WIINANEN,JOSIAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 23:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 02/12/2024 01:00:29

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.082	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:56	No Sample Provided
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Control Test	0.081	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WIIANEN,JOSIAH

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 23:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 02/12/2024 01:14:12

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.121	01:09
Breath Volume	2.718	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.112	01:12
Breath Volume	1.402	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 222152
BREATH TEST OPERATOR: COTTON,PIERCE,A	DRY GAS EXPIRY DATE: 01/27/2026
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: COTTON,PIERCE	ARREST TIME: 02:40
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Male
LAST AGENCY INSPECTION: 01/31/2024	OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: NWADIKE	DATE/TIME OF TEST: 02/16/2024	03:22:56
------------------------------	-------------------------------	----------

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.082	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1 0.129	03:10	
Breath Volume	1.875	03:10	
Air Blank	0.000	03:10	
Air Blank	0.000	03:12	
Subject Sample	# 2 0.095	03:15	Volume Not Met
Breath Volume	0.507	03:15	
Air Blank	0.000	03:16	
Air Blank	0.000	03:18	
Subject Sample	# 3 0.091	03:21	Volume Not Met
Breath Volume	0.535	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: COTTON,PIERCE

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 02:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: NWADIKE

DATE/TIME OF TEST: 02/16/2024 03:55:29

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.119	03:50
Breath Volume	2.003	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.117	03:53
Breath Volume	1.808	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TWIDDY,LINDSEY,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: SIPHER,VICTORIA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 02/22/2024 04:08:50

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.082	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.204	04:04
Breath Volume	2.042	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.199	04:07
Breath Volume	1.683	04:07
Air Blank	0.000	04:07
Control Test	0.082	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 20:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 02/23/2024 21:24:41

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.081	21:19
Air Blank	0.000	21:19
Subject Sample # 1	0.000	21:20
Breath Volume	4.312	21:20
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.000	21:23
Breath Volume	2.730	21:23
Air Blank	0.000	21:23
Control Test	0.079	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 02/24/2024 04:27:14

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.082	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.104	04:22
Breath Volume	3.175	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.105	04:25
Breath Volume	3.253	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: STOLLER

DATE/TIME OF TEST: 02/25/2024 01:01:19

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.000	00:55
Breath Volume	2.289	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.000	00:59
Breath Volume	2.605	00:59
Air Blank	0.000	01:00
Control Test	0.082	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 222152
BREATH TEST OPERATOR: COTTON,PIERCE,A	DRY GAS EXPIRY DATE: 01/27/2026
BREATH TEST OPERATOR AGENCY: Pensacola PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BECKER,JONATHAN	ARREST TIME: 19:28
ARRESTING OFFICER AGENCY: Pensacola PD	GENDER: Female
LAST AGENCY INSPECTION: 01/31/2024	OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: REAVES	DATE/TIME OF TEST: 02/25/2024 20:25:07
-----------------------------	--

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.081	20:14	
Air Blank	0.000	20:14	
Subject Sample	# 1 0.165	20:17	Volume Not Met
Breath Volume	0.722	20:17	
Air Blank	0.000	20:18	
Air Blank	0.000	20:20	
Subject Sample	# 2 0.188	20:23	Volume Not Met
Breath Volume	0.894	20:23	
Air Blank	0.000	20:24	
Control Test	0.078	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BECKER,JONATHAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 02/25/2024 20:39:56

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.079	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1 0.145	20:32	Volume Not Met
Breath Volume	0.125	20:32	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 2 0.166	20:38	Volume Not Met
Breath Volume	0.097	20:38	
Air Blank	0.000	20:38	
Control Test	0.078	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,KARISSA,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: MORALES,KARISSA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 02/27/2024 16:14:44

Diagnostics Check	OK	16:08
Air Blank	0.000	16:08
Control Test	0.082	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.000	16:10
Breath Volume	2.562	16:10
Air Blank	0.000	16:10
Air Blank	0.000	16:12
Subject Sample # 2	0.000	16:13
Breath Volume	3.566	16:13
Air Blank	0.000	16:13
Control Test	0.080	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:14

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: DUI

ARREST TIME: 20:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: PENFOLD

DATE/TIME OF TEST: 02/28/2024 21:36:48

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.000	21:32
Breath Volume	3.281	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:35
Breath Volume	2.476	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENT,ROBERT,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: KENT,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222152
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 02/29/2024 01:53:14

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:48
Breath Volume	1.347	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:51
Breath Volume	2.167	01:51
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 02/29/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ALVAREZ-TORRES

DATE/TIME OF TEST: 02/02/2024 01:01:32

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.106	00:56
Breath Volume	1.871	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.106	00:59
Breath Volume	3.136	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/02/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GARRIGA

DATE/TIME OF TEST: 02/02/2024 02:43:01

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.147	02:38
Breath Volume	2.136	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.138	02:41
Breath Volume	1.851	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/02/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CANAVARROS JANKOSWSK

DATE/TIME OF TEST: 02/02/2024 03:17:10

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.033	03:12
Breath Volume	2.429	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.034	03:15
Breath Volume	3.019	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 02/02/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: VEEDER

DATE/TIME OF TEST: 02/02/2024 03:33:45

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.181	03:29
Breath Volume	3.722	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.187	03:32
Breath Volume	4.097	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 02/02/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: AUNG

DATE/TIME OF TEST: 02/02/2024 04:49:47

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.081	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.132	04:45
Breath Volume	1.757	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.128	04:48
Breath Volume	2.121	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 02/02/2024 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa Airport PD
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: JEFFREY

DATE/TIME OF TEST: 02/04/2024 04:51:48

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1 0.122	04:43	Volume Not Met
Breath Volume	0.691	04:43	
Air Blank	0.000	04:44	
Air Blank	0.000	04:46	
Subject Sample	# 2 0.171	04:46	
Breath Volume	1.574	04:46	
Air Blank	0.000	04:47	
Air Blank	0.000	04:49	
Subject Sample	# 3 0.164	04:50	
Breath Volume	1.164	04:50	
Air Blank	0.000	04:50	
Control Test	0.080	04:51	
Air Blank	0.000	04:51	
Diagnostics Check	OK	04:51	

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: REYES ESPINOSA

DATE/TIME OF TEST: 02/04/2024 05:13:28

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.142	05:08
Breath Volume	2.902	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.136	05:11
Breath Volume	1.800	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: HUNEKE

DATE/TIME OF TEST: 02/06/2024 03:25:23

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.075	03:20
Breath Volume	2.425	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.074	03:23
Breath Volume	1.949	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: MONTERROSA

DATE/TIME OF TEST: 02/09/2024 06:05:41

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Subject Sample # 1	0.173	06:01
Breath Volume	1.554	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.158	06:04
Breath Volume	1.375	06:04
Air Blank	0.000	06:04
Control Test	0.080	06:05
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 02/10/2024 03:53:54

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.158	03:48
Breath Volume	1.503	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.150	03:52
Breath Volume	1.523	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILSON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/10/2024 04:10:30

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.159	04:05
Breath Volume	1.796	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.166	04:08
Breath Volume	2.085	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/10/2024 04:49:59

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.228	04:45
Breath Volume	2.097	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample # 2	0.231	04:48
Breath Volume	2.234	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 03/06/2024 15:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 02/17/2024 22:21:05

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.204	22:16
Breath Volume	1.410	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:18
Subject Sample # 2	0.202	22:19
Breath Volume	1.812	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FORY PEDROZA

DATE/TIME OF TEST: 02/19/2024 21:27:03

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.081	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1 0.110	21:19	Volume Not Met
Breath Volume	0.019	21:19	
Air Blank	0.000	21:20	
Air Blank	0.000	21:22	
Subject Sample	# 2 0.116	21:25	Volume Not Met
Breath Volume	0.597	21:25	
Air Blank	0.000	21:26	
Control Test	0.080	21:26	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FORY PEDROZA

DATE/TIME OF TEST: 02/19/2024 21:49:33

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.114	21:39	Volume Not Met
Breath Volume	0.730	21:39	
Air Blank	0.000	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 2 0.113	21:43	
Breath Volume	1.347	21:43	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 3 0.107	21:47	
Breath Volume	1.246	21:47	
Air Blank	0.000	21:48	
Control Test	0.080	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: TENEBRUSO

DATE/TIME OF TEST: 02/23/2024 02:44:00

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.000	02:39
Breath Volume	4.500	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:42
Breath Volume	4.792	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CALLAZO LOPEZ

DATE/TIME OF TEST: 02/24/2024 01:12:01

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.148	01:07
Breath Volume	1.523	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.152	01:10
Breath Volume	2.257	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 02/24/2024 02:54:38

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.140	02:50
Breath Volume	1.644	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.149	02:53
Breath Volume	1.250	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: CUEVAS GALINDO

DATE/TIME OF TEST: 02/24/2024 03:48:12

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.270	03:37	Volume Not Met
Breath Volume	0.902	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2 0.282	03:40	
Breath Volume	1.390	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 3 0.250	03:46	Volume Not Met
Breath Volume	1.085	03:46	
Air Blank	0.000	03:47	
Control Test	0.078	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/24/2024 05:15:41

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.021	05:11
Breath Volume	1.730	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.017	05:14
Breath Volume	1.902	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: YERKES JR

DATE/TIME OF TEST: 02/29/2024 02:19:40

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.080	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.131	02:13
Breath Volume	1.398	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.119	02:17
Breath Volume	1.523	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 03/06/2024 15:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: MICHALOWSKI

DATE/TIME OF TEST: 02/02/2024 16:25:35

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:23	
Control Test	0.081	16:23	
Air Blank	0.000	16:24	
Subject Sample	# 1	16:24	Subject Test Refused
Breath Volume	0.000	16:24	
Air Blank	0.000	16:24	
Control Test	0.080	16:24	
Air Blank	0.000	16:25	
Diagnostics Check	OK	16:25	

Data Download Date/Time: 02/20/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA',C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: BURGESS JR

DATE/TIME OF TEST: 02/03/2024 04:56:23

Diagnostics Check	OK	04:48
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.210	04:51
Breath Volume	3.816	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.198	04:54
Breath Volume	3.195	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 02/20/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BILLMAN

DATE/TIME OF TEST: 02/04/2024 00:31:36

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.160	00:26
Breath Volume	2.230	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.164	00:29
Breath Volume	2.671	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/20/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BUSCEMI,HUNTER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: TOMCZUK

DATE/TIME OF TEST: 02/13/2024 02:31:31

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.219	02:26
Breath Volume	1.445	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.227	02:29
Breath Volume	1.378	02:29
Air Blank	0.000	02:30
Control Test	0.077	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/20/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI,RYAN,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/15/2024 03:03:05

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.224	02:58
Breath Volume	1.515	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.226	03:01
Breath Volume	1.406	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/20/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 02/24/2024 00:20:12

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:19	
Control Test	0.080	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 03/12/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GINEVSKI

DATE/TIME OF TEST: 02/25/2024 00:44:12

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.108	00:39
Breath Volume	3.757	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.106	00:42
Breath Volume	3.601	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/12/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: CLARENC

DATE/TIME OF TEST: 02/25/2024 07:28:40

Diagnostics Check	OK	07:25	
Air Blank	0.000	07:26	
Control Test	0.081	07:26	
Air Blank	0.000	07:26	
Subject Sample	# 1	07:27	Subject Test Refused
Breath Volume	0.000	07:27	
Air Blank	0.000	07:27	
Control Test	0.081	07:28	
Air Blank	0.000	07:28	
Diagnostics Check	OK	07:28	

Data Download Date/Time: 03/12/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BEERBOWER,MICHAEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BOZARD

DATE/TIME OF TEST: 02/27/2024 00:07:15

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.081	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 03/12/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007168

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESPIAN,MIGUEL,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: DESPIAN,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: BUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: GECHTMAN

DATE/TIME OF TEST: 02/22/2024 18:32:40

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.079	18:29	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.000	18:30	
Air Blank	0.000	18:31	
Control Test	0.080	18:31	
Air Blank		18:32	Purge Fail
Air Blank	0.000	18:32	

Data Download Date/Time: 02/28/2024 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: SCAGGS

DATE/TIME OF TEST: 02/03/2024 04:43:36

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1	04:42	Subject Test Refused
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.081	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 02/20/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ARDREY-GIGUERE,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 02/09/2024 04:58:16

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.190	04:53
Breath Volume	1.609	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.192	04:56
Breath Volume	2.058	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 02/20/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GUERRERO,KEVEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DERUDDER

DATE/TIME OF TEST: 02/09/2024 23:09:26

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.080	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 02/20/2024 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VELGER

DATE/TIME OF TEST: 02/17/2024 04:01:45

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.081	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Control Test	0.080	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 02/20/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO, JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO, ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BALLINGER

DATE/TIME OF TEST: 02/18/2024 00:14:45

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.080	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.079	00:10
Breath Volume	1.824	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.076	00:13
Breath Volume	1.593	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 02/20/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: BARTASIS

DATE/TIME OF TEST: 02/23/2024 14:00:25

Diagnostics Check	OK	13:50
Air Blank	0.000	13:51
Control Test	0.081	13:51
Air Blank	0.000	13:52
Subject Sample # 1	0.205	13:53
Breath Volume	1.277	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.190	13:58
Breath Volume	1.386	13:58
Air Blank	0.000	13:59
Control Test	0.078	13:59
Air Blank	0.000	14:00
Diagnostics Check	OK	14:00

Data Download Date/Time: 03/12/2024 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ERYKA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: DOLBILOV

DATE/TIME OF TEST: 02/24/2024 20:24:21

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.200	20:18
Breath Volume	3.523	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.196	20:22
Breath Volume	1.164	20:22
Air Blank	0.000	20:23
Control Test	0.078	20:23
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 03/12/2024 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 02/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SOLER

DATE/TIME OF TEST: 02/27/2024 22:52:46

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.080	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.080	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 03/12/2024 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 02/10/2024 02:04:06

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank		02:03	RFI Detect
Air Blank		02:04	RFI Detect

Data Download Date/Time: 02/16/2024 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KATRISHEN

DATE/TIME OF TEST: 02/11/2024 03:18:39

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.182	03:13
Breath Volume	1.167	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.188	03:16
Breath Volume	1.492	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 02/16/2024 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/11/2024 19:03:07

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.081	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:58	No Sample Provided
Breath Volume	0.000	18:58	
Air Blank	0.000	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 2	19:01	Subject Test Refused
Breath Volume	0.000	19:01	
Air Blank	0.000	19:02	
Control Test	0.080	19:02	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time: 02/16/2024 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RIDDLE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CORVARI

DATE/TIME OF TEST: 02/16/2024 23:38:27

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.210	23:33
Breath Volume	2.105	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.205	23:36
Breath Volume	2.085	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONDANK, BLAKE, P
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BONDANK, BLAKE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ZUCCARELL

DATE/TIME OF TEST: 02/16/2024 23:51:33

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.143	23:47
Breath Volume	2.996	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.143	23:49
Breath Volume	3.277	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWER,D,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: OWER,D
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/17/2024 03:07:02

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.081	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,HUNTER,L
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ECKERT,HUNTER
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 02/17/2024 03:22:06

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.081	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESANDRE,DAVID,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CHICKRIE

DATE/TIME OF TEST: 02/19/2024 01:59:08

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.163	01:54
Breath Volume	2.175	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.158	01:57
Breath Volume	2.132	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEGUERO,FEDERICO,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PEGUERO,FEDERICO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:18

SUBJECT'S LAST NAME: EASTERDAY

DATE/TIME OF TEST: 02/22/2024 12:47:21

Diagnostics Check	OK	12:40
Air Blank	0.000	12:41
Control Test	0.082	12:41
Air Blank	0.000	12:41
Subject Sample # 1	0.258	12:42
Breath Volume	1.761	12:42
Air Blank	0.000	12:43
Air Blank	0.000	12:45
Subject Sample # 2	0.249	12:45
Breath Volume	2.605	12:45
Air Blank	0.000	12:46
Control Test	0.080	12:46
Air Blank	0.000	12:47
Diagnostics Check	OK	12:47

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLANSON,JACE,W
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SHARRETTE

DATE/TIME OF TEST: 02/24/2024 02:02:42

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.000	01:58
Breath Volume	4.062	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.000	02:01
Breath Volume	4.050	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNIPE,JAKE,C
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: KNIPE,JAKE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: KALKAI

DATE/TIME OF TEST: 02/24/2024 07:14:59

Diagnostics Check	OK	07:08
Air Blank	0.000	07:08
Control Test	0.082	07:08
Air Blank	0.000	07:09
Subject Sample # 1	0.220	07:10
Breath Volume	2.871	07:10
Air Blank	0.000	07:11
Air Blank	0.000	07:12
Subject Sample # 2	0.221	07:12
Breath Volume	2.996	07:12
Air Blank	0.000	07:14
Control Test	0.078	07:14
Air Blank	0.000	07:14
Diagnostics Check	OK	07:14

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLOSIMO,KATHLEEN,S
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COLOSIMO,KATHLEEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BOLANOS APARICIO

DATE/TIME OF TEST: 02/25/2024 21:46:14

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.082	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.133	21:41
Breath Volume	3.316	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:43
Subject Sample # 2	0.130	21:44
Breath Volume	3.382	21:44
Air Blank	0.000	21:45
Control Test	0.078	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREITAS,MARCIA,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BOULTON

DATE/TIME OF TEST: 02/26/2024 22:32:36

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.150	22:27
Breath Volume	2.660	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.144	22:30
Breath Volume	3.320	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,BRANDON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: ROBERTS,BRANDON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 02/26/2024 23:41:31

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.132	23:36
Breath Volume	2.621	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.132	23:39
Breath Volume	2.738	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RICKS

DATE/TIME OF TEST: 02/27/2024 22:59:06

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1 0.226	22:53	
Breath Volume	1.132	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 2 0.226	22:57	
Breath Volume	1.320	22:57	
Air Blank		22:58	Purge Fail
Air Blank	0.000	22:59	

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: KLEINWACHTER

DATE/TIME OF TEST: 02/28/2024 23:51:45

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.125	23:46
Breath Volume	2.953	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.126	23:49
Breath Volume	2.441	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 03/22/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: CIRA

DATE/TIME OF TEST: 02/01/2024 19:03:13

Diagnostics Check	OK	18:57	
Air Blank	0.000	18:57	
Control Test	0.080	18:58	
Air Blank	0.000	18:58	
Subject Sample	# 1 0.159	19:01	Volume Not Met
Breath Volume	0.304	19:01	
Air Blank		19:02	Purge Fail
Air Blank	0.000	19:03	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: CIRA

DATE/TIME OF TEST: 02/01/2024 19:10:28

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:05	
Control Test	0.075	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 1 0.169	19:09	Volume Not Met
Breath Volume	0.113	19:09	
Air Blank		19:10	Purge Fail
Air Blank	0.000	19:10	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 02/02/2024 02:36:35

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.151	02:32
Breath Volume	1.515	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.146	02:34
Breath Volume	1.960	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 02/02/2024 05:12:09

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.145	05:07
Breath Volume	2.886	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.140	05:10
Breath Volume	2.402	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,BRYAN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: MOZELL

DATE/TIME OF TEST: 02/03/2024 01:30:00

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.106	01:25
Breath Volume	1.597	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.114	01:28
Breath Volume	1.609	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:30

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY, RONALD, J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BAILEY, RONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: HEUSTON

DATE/TIME OF TEST: 02/03/2024 20:54:47

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.199	20:49
Breath Volume	4.003	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.198	20:52
Breath Volume	4.273	20:52
Air Blank	0.000	20:53
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCELDFOWNEY,STEPHEN,B
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MCELDFOWNEY,STEPHEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: SWILLEY

DATE/TIME OF TEST: 02/04/2024 20:14:02

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.080	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.196	20:09
Breath Volume	2.179	20:09
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.200	20:12
Breath Volume	3.695	20:12
Air Blank	0.000	20:13
Control Test	0.078	20:13
Air Blank	0.000	20:13
Diagnostics Check	OK	20:14

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FEASEL,KELSEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: S421000526470

DATE/TIME OF TEST: 02/10/2024 00:46:14

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:33	
Subject Sample # 1	0.164	00:35	
Breath Volume	1.515	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample # 2		00:41	No Sample Provided
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Air Blank	0.000	00:43	
Subject Sample # 3	0.165	00:44	
Breath Volume	1.367	00:44	
Air Blank		00:45	Purge Fail
Air Blank	0.000	00:46	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEASEL,KELSEY,D
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: FEASEL,KELSEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SALISBURY

DATE/TIME OF TEST: 02/10/2024 00:59:13

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.077	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.167	00:54
Breath Volume	1.746	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.161	00:57
Breath Volume	1.460	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JESSE,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 02/10/2024 02:45:16

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:40
Breath Volume	3.304	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:43
Breath Volume	4.230	02:43
Air Blank	0.000	02:44
Control Test	0.078	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 02/11/2024 01:14:26

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.153	01:09
Breath Volume	2.312	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.146	01:12
Breath Volume	3.191	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADDOCK,CHARLES,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PADDOCK,CHARLES
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: LAUREL

DATE/TIME OF TEST: 02/11/2024 04:45:35

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.179	04:40
Breath Volume	2.156	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.193	04:43
Breath Volume	1.683	04:43
Air Blank	0.000	04:44
Control Test	0.076	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURAJBALI,KYLE,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SURAJBALI,KYLE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 02/11/2024 20:26:00

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:17	
Control Test	0.080	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1 0.249	20:21	
Breath Volume	2.222	20:21	
Air Blank	0.000	20:22	
Air Blank	0.000	20:23	
Subject Sample	# 2 0.239	20:24	
Breath Volume	2.632	20:24	
Air Blank	0.000	20:25	
Control Test	0.074	20:25	Control Outside Tolerance
Air Blank	0.000	20:26	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESANDRE,DAVID,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SCHNARRS

DATE/TIME OF TEST: 02/13/2024 01:41:35

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	RFI Detect
Breath Volume	3.394	01:39	
Air Blank	0.000	01:40	
Control Test	0.030	01:41	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESANDRE,DAVID,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOWMAN,BRYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SCHNARRS

DATE/TIME OF TEST: 02/13/2024 01:51:07

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.145	01:46
Breath Volume	3.800	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.157	01:49
Breath Volume	2.812	01:49
Air Blank	0.000	01:50
Control Test	0.077	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: RIVAS SIERRA

DATE/TIME OF TEST: 02/14/2024 04:40:02

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.081	04:37	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 02/16/2024 18:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDER,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAYHUE,JOSHUA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/03/2024 07:02:16

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:49	
Control Test	0.083	06:49	
Air Blank	0.000	06:49	
Subject Sample	# 1 0.128	06:52	Volume Not Met
Breath Volume	0.796	06:52	
Air Blank	0.000	06:53	
Air Blank	0.000	06:55	
Subject Sample	# 2 0.130	06:56	
Breath Volume	1.281	06:56	
Air Blank	0.000	06:57	
Air Blank	0.000	06:59	
Subject Sample	# 3 0.122	07:00	
Breath Volume	1.632	07:00	
Air Blank	0.000	07:01	
Control Test	0.080	07:01	
Air Blank	0.000	07:02	
Diagnostics Check	OK	07:02	

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEIDNER,SARAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: CAPERS

DATE/TIME OF TEST: 02/08/2024 02:22:01

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.155	02:17
Breath Volume	2.781	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.157	02:20
Breath Volume	2.410	02:20
Air Blank	0.000	02:21
Control Test	0.084	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:22

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYA,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAYA,ZACHARY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: BARROW

DATE/TIME OF TEST: 02/19/2024 18:14:20

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.082	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.169	18:09
Breath Volume	1.894	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.161	18:12
Breath Volume	1.171	18:12
Air Blank	0.000	18:13
Control Test	0.082	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CAAL CUZ

DATE/TIME OF TEST: 02/24/2024 02:40:00

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.083	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.166	02:34
Breath Volume	1.652	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.178	02:37
Breath Volume	2.187	02:37
Air Blank	0.000	02:38
Control Test	0.083	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DEJOHN,CHRISTIAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/24/2024 21:49:50

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.084	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.084	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: TOMAS RECINOS

DATE/TIME OF TEST: 02/25/2024 03:00:17

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.083	02:54	
Air Blank	0.000	02:55	
Subject Sample # 1	0.211	02:55	
Breath Volume	2.062	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample # 2	0.193	02:58	
Breath Volume	2.562	02:58	
Air Blank	0.000	02:59	
Control Test	0.088	02:59	Control Outside Tolerance
Air Blank	0.000	03:00	

Data Download Date/Time: 02/28/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: SMULLINS

DATE/TIME OF TEST: 02/16/2024 20:19:28

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.078	20:13
Air Blank	0.000	20:14
Subject Sample # 1	0.155	20:14
Breath Volume	2.320	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.156	20:17
Breath Volume	2.292	20:17
Air Blank	0.000	20:18
Control Test	0.078	20:18
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,Z,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MCCOY,T
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 10:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: PEREZ VERDUGO

DATE/TIME OF TEST: 02/17/2024 11:07:10

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.078	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.232	11:02
Breath Volume	3.023	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:04
Subject Sample # 2	0.227	11:05
Breath Volume	1.421	11:05
Air Blank	0.000	11:06
Control Test	0.078	11:06
Air Blank	0.000	11:07
Diagnostics Check	OK	11:07

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,JEFFREY,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CASIMIRO,BRANDON
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/19/2024 17:28:32

Diagnostics Check	OK	17:22
Air Blank	0.000	17:22
Control Test	0.078	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.286	17:24
Breath Volume	1.160	17:24
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.294	17:26
Breath Volume	1.109	17:26
Air Blank	0.000	17:27
Control Test	0.078	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,ROBERT,L
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: PIERRE LOUIS

DATE/TIME OF TEST: 02/25/2024 01:51:34

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	2.035	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	2.128	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BEST

DATE/TIME OF TEST: 02/26/2024 23:21:20

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.194	23:16
Breath Volume	1.593	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.209	23:19
Breath Volume	1.507	23:19
Air Blank	0.000	23:20
Control Test	0.078	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MINOSO,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 02/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07622080A2
DRY GAS EXPIRY DATE: 05/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CONSTANT

DATE/TIME OF TEST: 02/28/2024 02:47:06

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1 0.053	02:41	Volume Not Met
Breath Volume	0.218	02:41	
Air Blank	0.000	02:41	
Air Blank	0.000	02:43	
Subject Sample	# 2	02:45	Subject Test Refused
Breath Volume	0.179	02:45	
Air Blank	0.000	02:46	
Control Test	0.078	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 03/27/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 19:54
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 02/01/2024	OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: WOODLING	DATE/TIME OF TEST: 02/02/2024	20:08:49
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.081	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.000	20:04
Breath Volume	3.289	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.000	20:07
Breath Volume	3.859	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 03/18/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LEATIOTA,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 02/03/2024 00:26:39

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.082	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.077	00:22
Breath Volume	3.042	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.077	00:25
Breath Volume	2.992	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 03/18/2024 15:41

Subject Test Electronic Data

DATE/TIME OF TEST: 02/03/2024 03:11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: FRANCO GOMEZ

DATE/TIME OF TEST: 02/04/2024 01:24:35

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.210	01:20
Breath Volume	3.078	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.209	01:22
Breath Volume	3.199	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/18/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 17:23
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 02/01/2024	OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: BEIGHLEY	DATE/TIME OF TEST: 02/05/2024	18:11:51
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	18:05
Air Blank	0.000	18:06
Control Test	0.081	18:06
Air Blank	0.000	18:06
Subject Sample # 1	0.369	18:07
Breath Volume	1.925	18:07
Air Blank	0.000	18:08
Air Blank	0.000	18:09
Subject Sample # 2	0.376	18:10
Breath Volume	2.054	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 03/18/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: HOWARD,J,G	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: HOWARD,J	ARREST TIME: 20:35
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 02/01/2024	OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: PLOY	DATE/TIME OF TEST: 02/21/2024	20:47:41
---------------------------	-------------------------------	----------

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.081	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.110	20:42
Breath Volume	1.808	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.116	20:45
Breath Volume	1.339	20:45
Air Blank	0.000	20:46
Control Test	0.079	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 03/18/2024 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: PETTIGREW

DATE/TIME OF TEST: 02/01/2024 20:38:28

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.081	20:35	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 02/28/2024 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,JOSEPH,P
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MONLAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: DONAY

DATE/TIME OF TEST: 02/03/2024 20:20:36

Diagnostics Check	OK	20:13
Air Blank	0.000	20:13
Control Test	0.081	20:14
Air Blank	0.000	20:14
Subject Sample # 1	0.217	20:15
Breath Volume	2.699	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.212	20:19
Breath Volume	2.136	20:19
Air Blank	0.000	20:19
Control Test	0.080	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 02/28/2024 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBSTER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEBSTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 10:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 02/07/2024 10:34:36

Diagnostics Check	OK	10:28
Air Blank	0.000	10:29
Control Test	0.081	10:29
Air Blank	0.000	10:30
Subject Sample # 1	0.000	10:30
Breath Volume	2.394	10:30
Air Blank	0.000	10:30
Air Blank	0.000	10:32
Subject Sample # 2	0.000	10:33
Breath Volume	2.257	10:33
Air Blank	0.000	10:33
Control Test	0.081	10:34
Air Blank	0.000	10:34
Diagnostics Check	OK	10:34

Data Download Date/Time: 02/28/2024 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 02/08/2024 21:37:01

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:28	
Control Test	0.081	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:32	No Sample Provided
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Air Blank	0.000	21:35	
Subject Sample	# 2	21:35	Subject Test Refused
Breath Volume	0.000	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:36	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCALIA, ADAM, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SCALIA, ADAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 02/09/2024 22:14:32

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.082	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:10
Breath Volume	2.015	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:13
Breath Volume	1.695	22:13
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 02/10/2024 11:02:43

Diagnostics Check	OK	10:56
Air Blank	0.000	10:57
Control Test	0.081	10:57
Air Blank	0.000	10:57
Subject Sample # 1	0.000	10:58
Breath Volume	1.757	10:58
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.000	11:01
Breath Volume	2.570	11:01
Air Blank	0.000	11:01
Control Test	0.081	11:02
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 02/11/2024 10:18:56

Diagnostics Check	OK	10:12
Air Blank	0.000	10:13
Control Test	0.081	10:13
Air Blank	0.000	10:14
Subject Sample # 1	0.000	10:14
Breath Volume	1.738	10:14
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.000	10:17
Breath Volume	1.460	10:17
Air Blank	0.000	10:18
Control Test	0.081	10:18
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAYHUE,JOSHUA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/15/2024 07:01:51

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.081	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.199	06:57
Breath Volume	3.429	06:57
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.202	07:00
Breath Volume	3.316	07:00
Air Blank	0.000	07:00
Control Test	0.080	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBSTER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEBSTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 02/17/2024 14:48:52

Diagnostics Check	OK	14:43
Air Blank	0.000	14:43
Control Test	0.081	14:43
Air Blank	0.000	14:44
Subject Sample # 1	0.000	14:44
Breath Volume	1.585	14:44
Air Blank	0.000	14:45
Air Blank	0.000	14:47
Subject Sample # 2	0.000	14:47
Breath Volume	2.117	14:47
Air Blank	0.000	14:47
Control Test	0.080	14:48
Air Blank	0.000	14:48
Diagnostics Check	OK	14:48

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAYHUE,JOSHUA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 02/18/2024 02:25:54

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.201	02:20
Breath Volume	1.515	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.197	02:24
Breath Volume	2.191	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 02/20/2024 18:02:21

Diagnostics Check	OK	17:55	
Air Blank	0.000	17:56	
Control Test	0.079	17:56	
Air Blank	0.000	17:57	
Subject Sample # 1	0.000	17:57	
Breath Volume	3.410	17:57	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample # 2	0.000	18:00	
Breath Volume	3.281	18:00	
Air Blank	0.000	18:00	
Control Test	0.083	18:01	
Air Blank	0.000	18:01	
Diagnostics Check	Fail	18:01	Analytical Stability Test
Air Blank	0.000	18:02	

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBSTER,CHRIS,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEBSTER,CHRIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 02/22/2024 10:27:35

Diagnostics Check	OK	10:21
Air Blank	0.000	10:21
Control Test	0.083	10:21
Air Blank	0.000	10:22
Subject Sample # 1	0.000	10:22
Breath Volume	1.546	10:22
Air Blank	0.000	10:23
Air Blank	0.000	10:25
Subject Sample # 2	0.000	10:26
Breath Volume	1.367	10:26
Air Blank	0.000	10:26
Control Test	0.083	10:26
Air Blank	0.000	10:27
Diagnostics Check	OK	10:27

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 02/23/2024 03:12:40

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.164	03:07
Breath Volume	1.109	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.184	03:11
Breath Volume	1.617	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBSTER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WEBSTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 02/26/2024 17:47:05

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.081	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.000	17:42
Breath Volume	2.125	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:45
Breath Volume	2.335	17:45
Air Blank	0.000	17:46
Control Test	0.083	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 02/28/2024 00:35:13

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.199	00:26
Breath Volume	1.160	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.174	00:29
Breath Volume	1.781	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 3	0.182	00:33
Breath Volume	2.355	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 02/28/2024 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 02/28/2024 18:28:51

Diagnostics Check	OK	18:22
Air Blank	0.000	18:23
Control Test	0.082	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.000	18:24
Breath Volume	1.851	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.000	18:27
Breath Volume	1.546	18:27
Air Blank	0.000	18:27
Control Test	0.082	18:28
Air Blank	0.000	18:28
Diagnostics Check	OK	18:28

Data Download Date/Time: 03/29/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LARGEN, JACKSON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 02/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/29/2024 22:38:15

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.085	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.195	22:33
Breath Volume	3.050	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.188	22:36
Breath Volume	2.156	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 03/29/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/16/2024 01:37:53

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.135	01:32
Breath Volume	2.109	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.134	01:36
Breath Volume	2.457	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 03/18/2024 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GACKI

DATE/TIME OF TEST: 02/23/2024 01:05:54

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.078	00:55	
Air Blank	0.000	00:56	
Subject Sample # 1	0.179	00:57	
Breath Volume	3.414	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample # 2	0.149	01:01	
Breath Volume	1.871	01:01	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample # 3	0.135	01:04	
Breath Volume	1.320	01:04	
Air Blank	0.000	01:05	
Control Test	0.073	01:05	Control Outside Tolerance
Air Blank	0.000	01:05	

Data Download Date/Time: 03/18/2024 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: GACKI

DATE/TIME OF TEST: 02/23/2024 01:14:00

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.077	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.166	01:09
Breath Volume	2.980	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.149	01:12
Breath Volume	1.925	01:12
Air Blank	0.000	01:13
Control Test	0.075	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:14

Data Download Date/Time: 03/18/2024 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SNEDEN

DATE/TIME OF TEST: 02/27/2024 23:18:35

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.070	23:14
Breath Volume	3.347	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.071	23:17
Breath Volume	3.472	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 03/18/2024 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 01/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RAYMUNDOCORIO

DATE/TIME OF TEST: 02/05/2024 01:10:11

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.175	01:05
Breath Volume	2.699	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.178	01:08
Breath Volume	3.109	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 02/29/2024 17:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REVOREDO,JULIO
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 02/03/2024 08:42:33

Diagnostics Check	OK	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Subject Sample # 1	0.176	08:37
Breath Volume	1.707	08:37
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.179	08:40
Breath Volume	2.382	08:40
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 02/07/2024 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ALFRED

DATE/TIME OF TEST: 02/04/2024 07:28:24

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.081	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.157	07:23
Breath Volume	1.644	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:26
Subject Sample # 2	0.157	07:26
Breath Volume	1.871	07:26
Air Blank	0.000	07:27
Control Test	0.080	07:27
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 02/07/2024 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REYES,ISAAH
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: TOJIL AMBROCIO

DATE/TIME OF TEST: 02/04/2024 21:18:29

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.274	21:09
Breath Volume	1.558	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.249	21:12
Breath Volume	1.164	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 3	0.261	21:16
Breath Volume	1.343	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 02/07/2024 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 01/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: BROMLEY

DATE/TIME OF TEST: 02/05/2024 06:39:53

Diagnostics Check	OK	06:33
Air Blank	0.000	06:34
Control Test	0.081	06:34
Air Blank	0.000	06:35
Subject Sample # 1	0.170	06:35
Breath Volume	1.800	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:37
Subject Sample # 2	0.158	06:38
Breath Volume	1.320	06:38
Air Blank	0.000	06:38
Control Test	0.080	06:39
Air Blank	0.000	06:39
Diagnostics Check	OK	06:39

Data Download Date/Time: 02/07/2024 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVENS,JAMORIS,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CRAVEN,J
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WETHERBEE

DATE/TIME OF TEST: 02/08/2024 23:13:49

Diagnostics Check	Fail	23:13	Internal Printer Test
Air Blank	0.000	23:13	

Data Download Date/Time: 02/17/2024 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVENS,JAMORIS,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CRAVEN,J
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WETHERBEE

DATE/TIME OF TEST: 02/08/2024 23:20:25

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank		23:19	Improper Sample
Air Blank	0.000	23:20	

Data Download Date/Time: 02/17/2024 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007439

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVENS,JAMORIS,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CRAVEN,J
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 01/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 35521080A2
DRY GAS EXPIRY DATE: 03/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WETHERBEE

DATE/TIME OF TEST: 02/08/2024 23:23:33

Diagnostics Check	OK	23:23	
Air Blank		23:23	Improper Sample
Air Blank		23:23	Improper Sample

Data Download Date/Time: 02/17/2024 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDAY,SETH,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: JOHNSON,JOSHUA
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 02/21/2024 21:15:20

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.082	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.132	21:09
Breath Volume	2.011	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.121	21:13
Breath Volume	1.675	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 03/19/2024 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: Gadsden County SO
ARRESTING OFFICER: MITCHELL,BRANDON
ARRESTING OFFICER AGENCY: Havana PD
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 02/23/2024 22:52:38

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.165	22:48
Breath Volume	1.976	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.168	22:50
Breath Volume	1.839	22:50
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 03/19/2024 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ, RESEAN, R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: CRUZ, RESEAN
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 02/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/24/2024 20:46:40

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.082	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1 0.013	20:45	
Breath Volume	1.734	20:45	
Air Blank		20:46	RFI Detect
Air Blank		20:46	RFI Detect
Air Blank	0.000	20:46	

Data Download Date/Time: 03/19/2024 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: CRUZ, RESEAN, R	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: Midway PD	VIOLATION CODE: DUI
ARRESTING OFFICER: CRUZ, RESEAN	ARREST TIME: 19:24
ARRESTING OFFICER AGENCY: Midway PD	GENDER: Male
LAST AGENCY INSPECTION: 02/12/2024	OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: YOUNG	DATE/TIME OF TEST: 02/24/2024	20:56:00
----------------------------	-------------------------------	----------

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.009	20:51
Breath Volume	1.769	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.009	20:54
Breath Volume	1.648	20:54
Air Blank	0.000	20:55
Control Test	0.079	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:56

Data Download Date/Time: 03/19/2024 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007456

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402623828
BREATH TEST OPERATOR: WHITE,MICAIAH,J	DRY GAS EXPIRY DATE: 10/22/2025
BREATH TEST OPERATOR AGENCY: FHP Troop H	VIOLATION CODE: DUI
ARRESTING OFFICER: DAVIS,JONTERIA	ARREST TIME: 20:26
ARRESTING OFFICER AGENCY: Quincy PD	GENDER: Male
LAST AGENCY INSPECTION: 02/12/2024	OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SERRANO	DATE/TIME OF TEST: 02/27/2024	22:17:14
------------------------------	-------------------------------	----------

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.082	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.015	22:12
Breath Volume	2.152	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.012	22:15
Breath Volume	2.003	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 03/19/2024 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FREITAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MENDEZ MENDEZ

DATE/TIME OF TEST: 02/18/2024 03:24:47

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.082	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1 0.169	03:19	
Breath Volume	2.433	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 2 0.173	03:22	
Breath Volume	1.746	03:22	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	Fail	03:24	Internal Printer Test
Air Blank	0.000	03:24	

Data Download Date/Time: 02/29/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FRIETAS,MARCIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MENDEZ MENDEZ

DATE/TIME OF TEST: 02/18/2024 03:35:18

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.160	03:30
Breath Volume	2.265	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.161	03:33
Breath Volume	2.507	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 02/29/2024 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: WORLIE HORNICK

DATE/TIME OF TEST: 02/04/2024 03:43:52

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.198	03:38
Breath Volume	3.503	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.197	03:41
Breath Volume	3.203	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/06/2024 16:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: REYES CARRANZA

DATE/TIME OF TEST: 02/17/2024 00:23:11

Diagnostics Check	OK	00:21	
Air Blank		00:22	Ambient Fail
Air Blank		00:23	Purge Fail

Data Download Date/Time: 03/06/2024 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: REYES CARRANZA

DATE/TIME OF TEST: 02/17/2024 00:25:46

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.088	00:25	Interferent Detect
Air Blank		00:25	Ambient Fail

Data Download Date/Time: 03/06/2024 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA ARIAS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MARCY

DATE/TIME OF TEST: 02/26/2024 03:05:01

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.056	02:59
Breath Volume	1.652	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 2	0.059	03:03
Breath Volume	2.031	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 03/06/2024 16:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILILLO,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,MICHAEL
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 02/03/2024 00:43:15

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.125	00:38
Breath Volume	1.464	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.125	00:41
Breath Volume	1.691	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 03/06/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CANO BORGES

DATE/TIME OF TEST: 02/03/2024 02:44:20

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.165	02:39
Breath Volume	2.019	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.156	02:42
Breath Volume	1.531	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 03/06/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FRANCISCO PEDRO

DATE/TIME OF TEST: 02/10/2024 03:57:50

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.258	03:49
Breath Volume	1.664	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.285	03:53
Breath Volume	1.519	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 3	0.262	03:56
Breath Volume	1.968	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/06/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILILLO,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,MICHAEL
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STALLINGS

DATE/TIME OF TEST: 02/19/2024 01:36:06

Diagnostics Check	Fail	01:35	Internal Printer Test
Air Blank	0.000	01:36	

Data Download Date/Time: 03/06/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILILLO,MICHAEL,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,MICHAEL
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STALLINGS

DATE/TIME OF TEST: 02/19/2024 01:44:18

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.085	01:39
Breath Volume	1.597	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.087	01:42
Breath Volume	1.492	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/06/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: AMBRIOSIO

DATE/TIME OF TEST: 02/19/2024 05:16:01

Diagnostics Check	OK	05:07
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.071	05:09
Breath Volume	1.585	05:09
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.072	05:14
Breath Volume	1.343	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:16

Data Download Date/Time: 03/06/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: DESJARDIN JR

DATE/TIME OF TEST: 02/21/2024 17:03:04

Diagnostics Check	Fail	17:02	Internal Printer Test
Air Blank	0.000	17:03	

Data Download Date/Time: 03/06/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: DESJARDIN JR

DATE/TIME OF TEST: 02/21/2024 17:11:24

Diagnostics Check	OK	17:04
Air Blank	0.000	17:05
Control Test	0.081	17:05
Air Blank	0.000	17:06
Subject Sample # 1	0.240	17:06
Breath Volume	2.816	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.244	17:09
Breath Volume	2.378	17:09
Air Blank	0.000	17:10
Control Test	0.081	17:10
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 03/06/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: SADLER

DATE/TIME OF TEST: 02/27/2024 20:34:58

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.149	20:29
Breath Volume	1.347	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:31
Subject Sample # 2	0.135	20:33
Breath Volume	1.265	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 03/06/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER, WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LUMPKINS

DATE/TIME OF TEST: 02/10/2024 02:45:54

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Subject Sample # 1	0.129	02:41	
Breath Volume	3.539	02:41	
Air Blank	0.000	02:42	
Air Blank	0.000	02:43	
Subject Sample # 2	0.127	02:44	
Breath Volume	2.800	02:44	
Air Blank	0.000	02:44	
Control Test	0.087	02:45	Control Outside Tolerance
Air Blank	0.000	02:45	

Data Download Date/Time: 03/06/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: GIARD

DATE/TIME OF TEST: 02/04/2024 01:16:37

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.229	01:12
Breath Volume	3.296	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.224	01:14
Breath Volume	3.500	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 03/01/2024 19:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 06723080A5
BREATH TEST OPERATOR: HANEY,MICHAEL,P	DRY GAS EXPIRY DATE: 04/05/2025
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)	VIOLATION CODE: OTHER
ARRESTING OFFICER: HANEY,MICHAEL	ARREST TIME: 12:30
ARRESTING OFFICER AGENCY: FFWCC (Lake City)	GENDER: Male
LAST AGENCY INSPECTION: 01/01/2024	OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: TEST	DATE/TIME OF TEST: 02/01/2024 13:12:48
---------------------------	--

Diagnostics Check	OK	13:09	
Air Blank	0.000	13:10	
Control Test	0.082	13:10	
Air Blank	0.000	13:10	
Subject Sample # 1	0.069	13:11	
Breath Volume	2.066	13:11	
Air Blank		13:12	Ambient Fail
Air Blank		13:12	Purge Fail

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:15:13

Diagnostics Check	OK	13:14	
Air Blank		13:14	RFI Detect
Air Blank	0.000	13:15	

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:17:53

Diagnostics Check	OK	13:16	
Air Blank	0.000	13:16	
Control Test	0.070	13:17	Control Outside Tolerance
Air Blank	0.000	13:17	

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:19:41

Diagnostics Check

Fail

13:19

Internal Printer Test

Air Blank

0.000

13:19

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: HANEY

DATE/TIME OF TEST: 02/01/2024 13:21:35

Diagnostics Check

Fail

13:21

Internal Printer Test

Air Blank

0.000

13:21

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:33:27

Diagnostics Check	OK	13:29
Air Blank	0.000	13:30
Control Test	0.079	13:30
Air Blank	0.000	13:31
Subject Sample # 1	0.097	13:31
Breath Volume	3.527	13:31
Air Blank	0.000	13:32
Air Blank	0.000	13:33

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:37:40

Diagnostics Check	OK	13:34
Air Blank	0.000	13:35
Control Test	0.077	13:35
Air Blank	0.000	13:35
Subject Sample # 1	0.114	13:36
Breath Volume	3.308	13:36
Air Blank	0.000	13:37
Air Blank	0.000	13:37

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:41:59

Diagnostics Check	OK	13:38
Air Blank	0.000	13:39
Control Test	0.080	13:39
Air Blank	0.000	13:40
Subject Sample # 1	0.125	13:40
Breath Volume	3.292	13:40
Air Blank	0.000	13:41
Air Blank	0.000	13:41

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007501

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANEY,MICHAEL,P

BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)

ARRESTING OFFICER: HANEY,MICHAEL

ARRESTING OFFICER AGENCY: FFWCC (Lake City)

LAST AGENCY INSPECTION: 01/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: OTHER

ARREST TIME: 12:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 02/01/2024 13:43:44

Diagnostics Check

Fail

13:43

Internal Printer Test

Air Blank

0.000

13:43

Data Download Date/Time: 02/09/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: COURT ORD
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 02/18/2024 11:44:51

Diagnostics Check	OK	11:39
Air Blank	0.000	11:39
Control Test	0.082	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.000	11:40
Breath Volume	1.613	11:40
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.000	11:43
Breath Volume	1.898	11:43
Air Blank	0.000	11:43
Control Test	0.081	11:44
Air Blank	0.000	11:44
Diagnostics Check	OK	11:44

Data Download Date/Time: 02/26/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 08:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 02/01/2024 09:09:01

Diagnostics Check	OK	09:02
Air Blank	0.000	09:02
Control Test	0.082	09:03
Air Blank	0.000	09:03
Subject Sample # 1	0.000	09:04
Breath Volume	2.738	09:04
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.000	09:07
Breath Volume	2.671	09:07
Air Blank	0.000	09:08
Control Test	0.081	09:08
Air Blank	0.000	09:08
Diagnostics Check	OK	09:09

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COTTO,GABRIEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DISDIER

DATE/TIME OF TEST: 02/03/2024 03:49:24

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.176	03:44
Breath Volume	2.074	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.172	03:47
Breath Volume	1.968	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 02/03/2024 10:19:20

Diagnostics Check	OK	10:12
Air Blank	0.000	10:13
Control Test	0.082	10:13
Air Blank	0.000	10:14
Subject Sample # 1	0.000	10:14
Breath Volume	1.773	10:14
Air Blank	0.000	10:15
Air Blank	0.000	10:17
Subject Sample # 2	0.000	10:17
Breath Volume	1.433	10:17
Air Blank	0.000	10:18
Control Test	0.081	10:18
Air Blank	0.000	10:19
Diagnostics Check	OK	10:19

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 02/06/2024 16:21:52

Diagnostics Check	OK	16:15
Air Blank	0.000	16:15
Control Test	0.082	16:16
Air Blank	0.000	16:16
Subject Sample # 1	0.000	16:17
Breath Volume	2.640	16:17
Air Blank	0.000	16:17
Air Blank	0.000	16:19
Subject Sample # 2	0.000	16:20
Breath Volume	2.937	16:20
Air Blank	0.000	16:20
Control Test	0.082	16:21
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FRAZIER,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/12/2024 04:05:59

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.081	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.168	03:56	Volume Not Met
Breath Volume	0.234	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2 0.181	04:00	
Breath Volume	1.347	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 3 0.177	04:04	
Breath Volume	1.566	04:04	
Air Blank	0.000	04:05	
Control Test	0.079	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ALTMAN,B
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 02/14/2024 16:07:37

Diagnostics Check	OK	16:01
Air Blank	0.000	16:01
Control Test	0.082	16:01
Air Blank	0.000	16:02
Subject Sample # 1	0.000	16:03
Breath Volume	2.437	16:03
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.000	16:06
Breath Volume	3.316	16:06
Air Blank	0.000	16:06
Control Test	0.082	16:07
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 02/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,A,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUTLER,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 02/15/2024 14:35:11

Diagnostics Check	OK	14:29
Air Blank	0.000	14:29
Control Test	0.081	14:29
Air Blank	0.000	14:30
Subject Sample # 1	0.000	14:31
Breath Volume	1.789	14:31
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.000	14:33
Breath Volume	2.000	14:33
Air Blank	0.000	14:34
Control Test	0.080	14:34
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 02/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ROBLERO-PEREZ

DATE/TIME OF TEST: 02/17/2024 03:54:57

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.209	03:50
Breath Volume	1.937	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.215	03:53
Breath Volume	2.375	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 02/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILES,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 01/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RODRIGUEZ-SIERRA

DATE/TIME OF TEST: 02/22/2024 03:08:06

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.214	03:03
Breath Volume	3.355	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.228	03:06
Breath Volume	3.945	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 02/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 02/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: EVANGELISTA

DATE/TIME OF TEST: 02/28/2024 03:23:14

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.081	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 03/18/2024 06:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAQUETTE,AARON,Y
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: WALKER,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: LOPEZ-VELAZQUEZ

DATE/TIME OF TEST: 02/03/2024 00:13:06

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.138	00:08
Breath Volume	1.917	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.138	00:11
Breath Volume	3.375	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/04/2024 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO,RAILY,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SERGEYEV,PAVEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MACARIO OLVERA

DATE/TIME OF TEST: 02/11/2024 00:45:50

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.124	00:41
Breath Volume	3.257	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.125	00:44
Breath Volume	2.546	00:44
Air Blank	0.000	00:44
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 03/04/2024 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 02/17/2024 04:33:42

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:22	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.200	04:23	No .020 agreement
Breath Volume	2.921	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 2 0.172	04:26	No .020 agreement
Breath Volume	1.777	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 3 0.154	04:32	Volume Not Met
Breath Volume	0.058	04:32	
Air Blank	0.000	04:32	
Control Test	0.079	04:33	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 03/04/2024 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 02/17/2024 04:47:04

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.186	04:41
Breath Volume	2.054	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.189	04:45
Breath Volume	2.570	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 03/04/2024 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEMHEUER, MICHAEL, W
BREATH TEST OPERATOR AGENCY: Arcadia PD
ARRESTING OFFICER: LANE, DALTON
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 01/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 02/20/2024 00:12:06

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.200	00:07
Breath Volume	3.218	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:09
Subject Sample # 2	0.206	00:10
Breath Volume	2.933	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/04/2024 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: WHIPPLE

DATE/TIME OF TEST: 02/01/2024 02:26:06

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:21
Breath Volume	3.976	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	3.062	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 02/01/2024 21:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 02/04/2024 10:31:25

Diagnostics Check	OK	10:24
Air Blank	0.000	10:25
Control Test	0.079	10:25
Air Blank	0.000	10:26
Subject Sample # 1	0.311	10:26
Breath Volume	2.890	10:26
Air Blank	0.000	10:27
Air Blank	0.000	10:29
Subject Sample # 2	0.293	10:29
Breath Volume	3.671	10:29
Air Blank	0.000	10:30
Control Test	0.077	10:30
Air Blank	0.000	10:31
Diagnostics Check	OK	10:31

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONTENOT,DOMINIC,M
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 02/07/2024 15:12:00

Diagnostics Check	OK	15:05
Air Blank	0.000	15:06
Control Test	0.081	15:06
Air Blank	0.000	15:07
Subject Sample # 1	0.000	15:07
Breath Volume	1.871	15:07
Air Blank	0.000	15:08
Air Blank	0.000	15:10
Subject Sample # 2	0.000	15:10
Breath Volume	2.078	15:10
Air Blank	0.000	15:11
Control Test	0.080	15:11
Air Blank	0.000	15:11
Diagnostics Check	OK	15:12

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,ANDREW,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MIGLIACCIO,N
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 02/08/2024 03:36:28

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.148	03:32
Breath Volume	1.968	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.149	03:34
Breath Volume	1.605	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/10/2024 01:29:10

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.055	01:24
Breath Volume	1.281	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.054	01:27
Breath Volume	1.136	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 02/10/2024 01:40:29

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.143	01:35
Breath Volume	1.933	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.141	01:38
Breath Volume	1.785	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/11/2024 03:20:33

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.171	03:15
Breath Volume	2.460	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.172	03:18
Breath Volume	2.734	03:18
Air Blank	0.000	03:19
Control Test	0.077	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/11/2024 03:39:31

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.081	03:34
Breath Volume	2.363	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.078	03:37
Breath Volume	1.519	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 02/12/2024 00:30:33

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.171	00:25
Breath Volume	2.171	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.177	00:28
Breath Volume	3.390	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,A,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SHUMAKER,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/13/2024 22:31:19

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.185	22:26
Breath Volume	2.355	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.173	22:29
Breath Volume	3.453	22:29
Air Blank	0.000	22:30
Control Test	0.076	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PERRONI

DATE/TIME OF TEST: 02/14/2024 03:25:03

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:14	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Air Blank	0.000	03:20	
Subject Sample	# 2	03:23	No Sample Provided
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.078	03:24	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUTTICK,EMILY,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KELLY,GRIFFIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/14/2024 13:25:04

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.079	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:20
Breath Volume	1.750	13:20
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:23
Breath Volume	1.242	13:23
Air Blank	0.000	13:24
Control Test	0.079	13:24
Air Blank	0.000	13:24
Diagnostics Check	OK	13:25

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 02/14/2024 19:06:38

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.078	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.000	19:02
Breath Volume	1.417	19:02
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.000	19:05
Breath Volume	1.945	19:05
Air Blank	0.000	19:05
Control Test	0.078	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 02/15/2024 00:08:13

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.079	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.199	00:03
Breath Volume	2.367	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.203	00:06
Breath Volume	2.363	00:06
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRYANT,REUBEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/16/2024 08:47:08

Diagnostics Check	OK	08:41
Air Blank	0.000	08:41
Control Test	0.078	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.000	08:42
Breath Volume	1.941	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.000	08:45
Breath Volume	2.402	08:45
Air Blank	0.000	08:46
Control Test	0.077	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SPEROS,KOBE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: MAREK

DATE/TIME OF TEST: 02/16/2024 19:45:08

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.078	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.082	19:38
Breath Volume	1.140	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.084	19:43
Breath Volume	1.546	19:43
Air Blank	0.000	19:44
Control Test	0.078	19:44
Air Blank	0.000	19:45
Diagnostics Check	OK	19:45

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SLOCUM,TERRENCE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: SOK

DATE/TIME OF TEST: 02/16/2024 20:17:41

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.077	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.225	20:13
Breath Volume	1.457	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.234	20:15
Breath Volume	1.656	20:15
Air Blank	0.000	20:16
Control Test	0.077	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALKER JR

DATE/TIME OF TEST: 02/17/2024 23:21:54

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.079	23:15	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.183	23:17	
Breath Volume	1.136	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 2	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.077	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEYMANS,MORNE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HEYMANS,MORNE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 02/19/2024 18:43:49

Diagnostics Check	OK	18:41	
Air Blank	0.000	18:41	
Control Test	0.079	18:41	
Air Blank	0.000	18:42	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.079	18:43	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: REECE,STEPHEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: KHASHCHUK

DATE/TIME OF TEST: 02/20/2024 02:42:17

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.079	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.178	02:31	
Breath Volume	1.324	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 2 0.123	02:37	Volume Not Met
Breath Volume	0.339	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 3 0.159	02:40	
Breath Volume	2.269	02:40	
Air Blank	0.000	02:41	
Control Test	0.076	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: BOARDMAN

DATE/TIME OF TEST: 02/20/2024 21:59:59

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.182	21:55
Breath Volume	1.644	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:57
Subject Sample # 2	0.177	21:58
Breath Volume	1.527	21:58
Air Blank	0.000	21:59
Control Test	0.077	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/24/2024 01:42:19

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:37
Breath Volume	1.863	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:40
Breath Volume	2.925	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 02/25/2024 04:04:35

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.102	03:59
Breath Volume	1.738	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.108	04:02
Breath Volume	2.351	04:02
Air Blank	0.000	04:03
Control Test	0.077	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 02/26/2024 00:56:52

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.079	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.122	00:52
Breath Volume	1.824	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.125	00:55
Breath Volume	2.195	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,N
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: KOSTENSKI

DATE/TIME OF TEST: 02/29/2024 18:44:10

Diagnostics Check	OK	18:37
Air Blank	0.000	18:38
Control Test	0.078	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:39
Breath Volume	1.660	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.000	18:42
Breath Volume	2.007	18:42
Air Blank	0.000	18:43
Control Test	0.076	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 03/18/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MASTERS,C
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 02/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SMEAD

DATE/TIME OF TEST: 02/29/2024 21:12:34

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.077	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.139	21:07
Breath Volume	2.859	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.126	21:11
Breath Volume	1.449	21:11
Air Blank	0.000	21:11
Control Test	0.078	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 03/18/2024 15:36