

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LABONTE,NICOLE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 00:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/05/2024 00:38:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.080 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.225 | 00:33 |
| Breath Volume      | 3.312 | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.221 | 00:36 |
| Breath Volume      | 2.117 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.079 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Diagnostics Check  | OK    | 00:38 |

---

Data Download Date/Time: 05/01/2024 12:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: ERWIN

DATE/TIME OF TEST: 04/12/2024 16:19:17

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 16:09 |                      |
| Air Blank         | 0.000     | 16:09 |                      |
| Control Test      | 0.080     | 16:10 |                      |
| Air Blank         | 0.000     | 16:10 |                      |
| Subject Sample    | # 1 0.292 | 16:13 | Volume Not Met       |
| Breath Volume     | 0.300     | 16:13 |                      |
| Air Blank         | 0.000     | 16:14 |                      |
| Air Blank         | 0.000     | 16:16 |                      |
| Subject Sample    | # 2       | 16:17 | Subject Test Refused |
| Breath Volume     | 0.019     | 16:17 |                      |
| Air Blank         | 0.000     | 16:17 |                      |
| Control Test      | 0.075     | 16:18 |                      |
| Air Blank         | 0.000     | 16:18 |                      |
| Diagnostics Check | Fail      | 16:18 | Voltage/Current Test |
| Air Blank         | 0.000     | 16:19 |                      |

---

Data Download Date/Time: 05/01/2024 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: DOWNES

DATE/TIME OF TEST: 04/15/2024 18:26:19

---

Diagnostics Check

Fail

18:25

Voltage/Current Test

Air Blank

0.000

18:26

---

Data Download Date/Time: 05/01/2024 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: DOWNES

DATE/TIME OF TEST: 04/15/2024 18:29:14

---

Diagnostics Check

Fail

18:28

Voltage/Current Test

Air Blank

0.000

18:29

---

Data Download Date/Time: 05/01/2024 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: COON,DAVID

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 01:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/17/2024 02:39:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.080 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 1 | 0.113 | 02:35 |
| Breath Volume      | 2.007 | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.111 | 02:38 |
| Breath Volume      | 2.152 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.080 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:39 |

---

Data Download Date/Time: 05/01/2024 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 04/18/2024 00:27:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:24 |                      |
| Air Blank         | 0.000 | 00:24 |                      |
| Control Test      | 0.079 | 00:25 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Subject Sample    | # 1   | 00:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:26 |                      |
| Air Blank         | 0.000 | 00:26 |                      |
| Control Test      | 0.079 | 00:27 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Diagnostics Check | OK    | 00:27 |                      |

---

Data Download Date/Time: 05/01/2024 12:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 04/06/2024 00:28:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:25 |                      |
| Air Blank         | 0.000 | 00:25 |                      |
| Control Test      | 0.079 | 00:26 |                      |
| Air Blank         | 0.000 | 00:26 |                      |
| Subject Sample    | # 1   | 00:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:26 |                      |
| Air Blank         | 0.000 | 00:27 |                      |
| Control Test      | 0.078 | 00:27 |                      |
| Air Blank         | 0.000 | 00:28 |                      |
| Diagnostics Check | OK    | 00:28 |                      |

---

Data Download Date/Time: 04/11/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: PINTO SALINAS

DATE/TIME OF TEST: 04/06/2024 06:03:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:56 |
| Air Blank          | 0.000 | 05:57 |
| Control Test       | 0.079 | 05:57 |
| Air Blank          | 0.000 | 05:58 |
| Subject Sample # 1 | 0.277 | 05:58 |
| Breath Volume      | 3.074 | 05:58 |
| Air Blank          | 0.000 | 05:59 |
| Air Blank          | 0.000 | 06:01 |
| Subject Sample # 2 | 0.271 | 06:01 |
| Breath Volume      | 3.015 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Control Test       | 0.077 | 06:02 |
| Air Blank          | 0.000 | 06:03 |
| Diagnostics Check  | OK    | 06:03 |

---

Data Download Date/Time: 04/11/2024 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ADULTE

DATE/TIME OF TEST: 04/07/2024 04:23:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.080 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 1 | 0.090 | 04:18 |
| Breath Volume      | 1.214 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 2 | 0.089 | 04:22 |
| Breath Volume      | 2.101 | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.078 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Diagnostics Check  | OK    | 04:23 |

---

Data Download Date/Time: 04/11/2024 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROZZI,SHANNON,R  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: FRISK,BENJAMIN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 04/07/2024 21:00:28

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:57 |                      |
| Air Blank         | 0.000 | 20:57 |                      |
| Control Test      | 0.079 | 20:58 |                      |
| Air Blank         | 0.000 | 20:58 |                      |
| Subject Sample    | # 1   | 20:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:59 |                      |
| Air Blank         | 0.000 | 20:59 |                      |
| Control Test      | 0.078 | 20:59 |                      |
| Air Blank         | 0.000 | 21:00 |                      |
| Diagnostics Check | OK    | 21:00 |                      |

---

Data Download Date/Time: 04/11/2024 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RESSLER,GREYSON,L  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: DELAY,A.  
ARRESTING OFFICER AGENCY: Kenneth City PD  
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STORCK

DATE/TIME OF TEST: 04/08/2024 22:02:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:53 |
| Air Blank          | 0.000 | 21:54 |
| Control Test       | 0.079 | 21:54 |
| Air Blank          | 0.000 | 21:54 |
| Subject Sample # 1 | 0.182 | 21:57 |
| Breath Volume      | 1.351 | 21:57 |
| Air Blank          | 0.000 | 21:57 |
| Air Blank          | 0.000 | 21:59 |
| Subject Sample # 2 | 0.179 | 22:00 |
| Breath Volume      | 2.199 | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.078 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Diagnostics Check  | OK    | 22:02 |

---

Data Download Date/Time: 04/11/2024 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RESSLER,GREYSON,L  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: GAINER,JAMES  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 17:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 04/13/2024 18:13:57

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 18:01 |                |
| Air Blank         | 0.000     | 18:01 |                |
| Control Test      | 0.079     | 18:02 |                |
| Air Blank         | 0.000     | 18:02 |                |
| Subject Sample    | # 1 0.000 | 18:05 | Volume Not Met |
| Breath Volume     | 0.140     | 18:05 |                |
| Air Blank         | 0.000     | 18:06 |                |
| Air Blank         | 0.000     | 18:08 |                |
| Subject Sample    | # 2 0.000 | 18:08 |                |
| Breath Volume     | 1.121     | 18:08 |                |
| Air Blank         | 0.000     | 18:09 |                |
| Air Blank         | 0.000     | 18:11 |                |
| Subject Sample    | # 3 0.000 | 18:12 |                |
| Breath Volume     | 1.648     | 18:12 |                |
| Air Blank         | 0.000     | 18:13 |                |
| Control Test      | 0.080     | 18:13 |                |
| Air Blank         | 0.000     | 18:13 |                |
| Diagnostics Check | OK        | 18:13 |                |

---

Data Download Date/Time: 05/27/2024 08:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 04/13/2024 22:32:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Control Test       | 0.078 | 22:27 |
| Air Blank          | 0.000 | 22:28 |
| Subject Sample # 1 | 0.168 | 22:28 |
| Breath Volume      | 1.164 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 2 | 0.162 | 22:31 |
| Breath Volume      | 1.535 | 22:31 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.078 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Diagnostics Check  | OK    | 22:32 |

---

Data Download Date/Time: 05/27/2024 08:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BAXCAJAY MARTINEZ

DATE/TIME OF TEST: 04/14/2024 03:19:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.078 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Subject Sample    | # 1   | 03:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Control Test      | 0.078 | 03:19 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Diagnostics Check | OK    | 03:19 |                      |

---

Data Download Date/Time: 05/27/2024 08:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCORDINO, DOUGLAS, J  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: WASHINGTON, RYAN  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SPALDING

DATE/TIME OF TEST: 04/15/2024 16:25:14

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:22 |                      |
| Air Blank         | 0.000 | 16:22 |                      |
| Control Test      | 0.078 | 16:22 |                      |
| Air Blank         | 0.000 | 16:23 |                      |
| Subject Sample    | # 1   | 16:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:23 |                      |
| Air Blank         | 0.000 | 16:24 |                      |
| Control Test      | 0.078 | 16:24 |                      |
| Air Blank         | 0.000 | 16:25 |                      |
| Diagnostics Check | OK    | 16:25 |                      |

Data Download Date/Time: 05/27/2024 08:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 04/20/2024 03:55:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.078 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.162 | 03:48 |
| Breath Volume      | 1.511 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 2 | 0.163 | 03:53 |
| Breath Volume      | 1.445 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.078 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Diagnostics Check  | OK    | 03:55 |

---

Data Download Date/Time: 05/27/2024 08:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STERLING,M  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: ADAMS,STERLING  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SOUZA

DATE/TIME OF TEST: 04/21/2024 01:25:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.078 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 1 | 0.325 | 01:20 |
| Breath Volume      | 2.875 | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 2 | 0.334 | 01:23 |
| Breath Volume      | 2.367 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.077 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Diagnostics Check  | OK    | 01:25 |

---

Data Download Date/Time: 05/27/2024 08:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROZZI,SHANNON,R  
BREATH TEST OPERATOR AGENCY: Pinellas Park PD  
ARRESTING OFFICER: SHEEHAN,MICHAEL  
ARRESTING OFFICER AGENCY: Pinellas Park PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06023080A1  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WOODRUFF

DATE/TIME OF TEST: 04/22/2024 01:57:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.078 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 1 | 0.152 | 01:52 |
| Breath Volume      | 2.242 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 2 | 0.148 | 01:55 |
| Breath Volume      | 1.933 | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.076 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 05/27/2024 08:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MANO,MARDEE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: ADMIN  
ARREST TIME: 12:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: MCNAB

DATE/TIME OF TEST: 04/01/2024 14:59:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:53 |
| Air Blank          | 0.000 | 14:53 |
| Control Test       | 0.081 | 14:53 |
| Air Blank          | 0.000 | 14:54 |
| Subject Sample # 1 | 0.073 | 14:54 |
| Breath Volume      | 2.683 | 14:54 |
| Air Blank          | 0.000 | 14:55 |
| Air Blank          | 0.000 | 14:57 |
| Subject Sample # 2 | 0.077 | 14:57 |
| Breath Volume      | 3.078 | 14:57 |
| Air Blank          | 0.000 | 14:58 |
| Control Test       | 0.081 | 14:58 |
| Air Blank          | 0.000 | 14:59 |
| Diagnostics Check  | OK    | 14:59 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KIMBERLAIN,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 04/04/2024 01:23:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:17 |
| Air Blank          | 0.000 | 01:17 |
| Control Test       | 0.081 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Subject Sample # 1 | 0.120 | 01:19 |
| Breath Volume      | 3.050 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 2 | 0.118 | 01:22 |
| Breath Volume      | 3.242 | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.080 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Diagnostics Check  | OK    | 01:23 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNN,CDFC,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MORITZ,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: KAZAKOV

DATE/TIME OF TEST: 04/09/2024 02:06:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.081 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 1 | 0.178 | 02:02 |
| Breath Volume      | 1.445 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Air Blank          | 0.000 | 02:04 |
| Subject Sample # 2 | 0.186 | 02:05 |
| Breath Volume      | 1.710 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Control Test       | 0.080 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Diagnostics Check  | OK    | 02:06 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HOUSEMAN,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: ADMIN  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HELSEL

DATE/TIME OF TEST: 04/10/2024 01:07:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.081 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Subject Sample # 1 | 0.077 | 01:03 |
| Breath Volume      | 1.902 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.080 | 01:06 |
| Breath Volume      | 1.929 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Control Test       | 0.082 | 01:06 |
| Air Blank          | 0.000 | 01:07 |
| Diagnostics Check  | OK    | 01:07 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILES,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FRIEND,S  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/10/2024 02:42:19

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:30 |                    |
| Air Blank         | 0.000 | 02:31 |                    |
| Control Test      | 0.081 | 02:31 |                    |
| Air Blank         | 0.000 | 02:32 |                    |
| Subject Sample    | # 1   | 02:35 | No Sample Provided |
| Breath Volume     | 0.000 | 02:35 |                    |
| Air Blank         | 0.000 | 02:35 |                    |
| Air Blank         | 0.000 | 02:37 |                    |
| Subject Sample    | # 2   | 02:40 | No Sample Provided |
| Breath Volume     | 0.000 | 02:40 |                    |
| Air Blank         | 0.000 | 02:41 |                    |
| Control Test      | 0.082 | 02:41 |                    |
| Air Blank         | 0.000 | 02:42 |                    |
| Diagnostics Check | OK    | 02:42 |                    |

---

Data Download Date/Time: 04/29/2024 22:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILES,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FRIEND,S  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: N/A

DATE/TIME OF TEST: 04/10/2024 03:13:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:00 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.081 | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 1 | 0.165 | 03:03 |
| Breath Volume      | 1.105 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 2 | 0.187 | 03:07 |
| Breath Volume      | 1.269 | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Air Blank          | 0.000 | 03:09 |
| Subject Sample # 3 | 0.178 | 03:11 |
| Breath Volume      | 2.214 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.081 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:13 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SCHWOCHO,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: ADMIN  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: HELSEL

DATE/TIME OF TEST: 04/10/2024 10:24:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:18 |
| Air Blank          | 0.000 | 10:19 |
| Control Test       | 0.081 | 10:19 |
| Air Blank          | 0.000 | 10:19 |
| Subject Sample # 1 | 0.000 | 10:20 |
| Breath Volume      | 1.675 | 10:20 |
| Air Blank          | 0.000 | 10:20 |
| Air Blank          | 0.000 | 10:22 |
| Subject Sample # 2 | 0.000 | 10:23 |
| Breath Volume      | 1.894 | 10:23 |
| Air Blank          | 0.000 | 10:23 |
| Control Test       | 0.081 | 10:24 |
| Air Blank          | 0.000 | 10:24 |
| Diagnostics Check  | OK    | 10:24 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANO,MARDEE,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MANO,MARDEE  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: ADMIN  
ARREST TIME: 11:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/10/2024 16:31:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:24 |
| Air Blank          | 0.000 | 16:25 |
| Control Test       | 0.081 | 16:25 |
| Air Blank          | 0.000 | 16:26 |
| Subject Sample # 1 | 0.000 | 16:26 |
| Breath Volume      | 2.933 | 16:26 |
| Air Blank          | 0.000 | 16:27 |
| Air Blank          | 0.000 | 16:29 |
| Subject Sample # 2 | 0.000 | 16:29 |
| Breath Volume      | 2.835 | 16:29 |
| Air Blank          | 0.000 | 16:30 |
| Control Test       | 0.081 | 16:30 |
| Air Blank          | 0.000 | 16:30 |
| Diagnostics Check  | OK    | 16:31 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,CELESTE,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CAMERON,C  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: ADMIN  
ARREST TIME: 12:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/11/2024 13:25:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:19 |
| Air Blank          | 0.000 | 13:20 |
| Control Test       | 0.081 | 13:20 |
| Air Blank          | 0.000 | 13:21 |
| Subject Sample # 1 | 0.000 | 13:21 |
| Breath Volume      | 3.578 | 13:21 |
| Air Blank          | 0.000 | 13:21 |
| Air Blank          | 0.000 | 13:23 |
| Subject Sample # 2 | 0.000 | 13:24 |
| Breath Volume      | 2.253 | 13:24 |
| Air Blank          | 0.000 | 13:24 |
| Control Test       | 0.081 | 13:25 |
| Air Blank          | 0.000 | 13:25 |
| Diagnostics Check  | OK    | 13:25 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARROLL,STEVEN,C  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: CARROLL,STEVEN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/12/2024 00:06:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:03 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Control Test      | 0.082 | 00:04 |                      |
| Air Blank         | 0.000 | 00:04 |                      |
| Subject Sample    | # 1   | 00:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:04 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Control Test      | 0.081 | 00:05 |                      |
| Air Blank         | 0.000 | 00:06 |                      |
| Diagnostics Check | OK    | 00:06 |                      |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COURTNEY,PATRICK,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CARRARO,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 04/12/2024 21:04:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.081 | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Subject Sample # 1 | 0.183 | 20:59 |
| Breath Volume      | 2.261 | 20:59 |
| Air Blank          | 0.000 | 21:00 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 2 | 0.197 | 21:03 |
| Breath Volume      | 1.921 | 21:03 |
| Air Blank          | 0.000 | 21:03 |
| Control Test       | 0.081 | 21:04 |
| Air Blank          | 0.000 | 21:04 |
| Diagnostics Check  | OK    | 21:04 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORD,JAMES,G  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: WYMER,DEP  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: KINGS

DATE/TIME OF TEST: 04/14/2024 01:41:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Control Test       | 0.081 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 1 | 0.168 | 01:36 |
| Breath Volume      | 4.625 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Air Blank          | 0.000 | 01:39 |
| Subject Sample # 2 | 0.170 | 01:40 |
| Breath Volume      | 4.820 | 01:40 |
| Air Blank          | 0.000 | 01:40 |
| Control Test       | 0.081 | 01:41 |
| Air Blank          | 0.000 | 01:41 |
| Diagnostics Check  | OK    | 01:41 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUCKER,DEPUTY,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ANDERSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: MONGE

DATE/TIME OF TEST: 04/14/2024 07:56:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:50 |
| Air Blank          | 0.000 | 07:50 |
| Control Test       | 0.080 | 07:50 |
| Air Blank          | 0.000 | 07:51 |
| Subject Sample # 1 | 0.142 | 07:51 |
| Breath Volume      | 2.312 | 07:51 |
| Air Blank          | 0.000 | 07:52 |
| Air Blank          | 0.000 | 07:54 |
| Subject Sample # 2 | 0.148 | 07:54 |
| Breath Volume      | 3.660 | 07:54 |
| Air Blank          | 0.000 | 07:55 |
| Control Test       | 0.080 | 07:55 |
| Air Blank          | 0.000 | 07:56 |
| Diagnostics Check  | OK    | 07:56 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGESS,JERRALD,P  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 04/19/2024 23:33:32

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 23:32 |            |
| Air Blank         |       | 23:33 | RFI Detect |
| Air Blank         | 0.000 | 23:33 |            |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOENNEMANN,ANDREW,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DAVIDSON,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 04/19/2024 23:39:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:36 |                      |
| Air Blank         | 0.000 | 23:36 |                      |
| Control Test      | 0.081 | 23:37 |                      |
| Air Blank         | 0.000 | 23:37 |                      |
| Subject Sample    | # 1   | 23:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:37 |                      |
| Air Blank         | 0.000 | 23:38 |                      |
| Control Test      | 0.081 | 23:38 |                      |
| Air Blank         | 0.000 | 23:39 |                      |
| Diagnostics Check | OK    | 23:39 |                      |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FERNANDEZ,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RODRGUEZ-ARTILES

DATE/TIME OF TEST: 04/21/2024 02:26:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.081 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 1 | 0.152 | 02:22 |
| Breath Volume      | 2.730 | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 2 | 0.144 | 02:25 |
| Breath Volume      | 2.652 | 02:25 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.081 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Diagnostics Check  | OK    | 02:26 |

---

Data Download Date/Time: 04/29/2024 22:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PENCIC,KENNETH,A  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: PAILLE,D  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: LATTIMORE

DATE/TIME OF TEST: 04/28/2024 12:27:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:20 |
| Air Blank          | 0.000 | 12:21 |
| Control Test       | 0.081 | 12:21 |
| Air Blank          | 0.000 | 12:22 |
| Subject Sample # 1 | 0.017 | 12:23 |
| Breath Volume      | 2.203 | 12:23 |
| Air Blank          | 0.000 | 12:23 |
| Air Blank          | 0.000 | 12:25 |
| Subject Sample # 2 | 0.016 | 12:26 |
| Breath Volume      | 2.105 | 12:26 |
| Air Blank          | 0.000 | 12:26 |
| Control Test       | 0.080 | 12:26 |
| Air Blank          | 0.000 | 12:27 |
| Diagnostics Check  | OK    | 12:27 |

---

Data Download Date/Time: 04/29/2024 22:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BONNICI

DATE/TIME OF TEST: 04/04/2024 02:19:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.079 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Subject Sample # 1 | 0.134 | 02:14 |
| Breath Volume      | 1.535 | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 2 | 0.133 | 02:18 |
| Breath Volume      | 1.996 | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.079 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Diagnostics Check  | OK    | 02:19 |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BOLSTER

DATE/TIME OF TEST: 04/04/2024 22:24:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.079 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 1 | 0.000 | 22:20 |
| Breath Volume      | 2.152 | 22:20 |
| Air Blank          | 0.000 | 22:21 |
| Air Blank          | 0.000 | 22:23 |
| Subject Sample # 2 | 0.000 | 22:23 |
| Breath Volume      | 2.101 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.080 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Diagnostics Check  | OK    | 22:24 |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LASCKO,J,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: LASCKO,J  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BRAIA

DATE/TIME OF TEST: 04/06/2024 15:07:58

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 15:07 |            |
| Air Blank         |       | 15:07 | RFI Detect |
| Air Blank         | 0.000 | 15:07 |            |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LASCKO,J,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: LASCKO,J  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BRAIA

DATE/TIME OF TEST: 04/06/2024 15:21:01

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 15:09 |                           |
| Air Blank         | 0.000     | 15:09 |                           |
| Control Test      | 0.079     | 15:10 |                           |
| Air Blank         | 0.000     | 15:10 |                           |
| Subject Sample    | # 1 0.163 | 15:13 | Volume Not Met            |
| Breath Volume     | 0.179     | 15:13 |                           |
| Air Blank         | 0.000     | 15:14 |                           |
| Air Blank         | 0.000     | 15:16 |                           |
| Subject Sample    | # 2 0.132 | 15:19 | Volume Not Met            |
| Breath Volume     | 0.355     | 15:19 |                           |
| Air Blank         | 0.000     | 15:20 |                           |
| Control Test      | 0.074     | 15:20 | Control Outside Tolerance |
| Air Blank         | 0.000     | 15:21 |                           |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUIS,JUSTIN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: FROST,B  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/06/2024 21:05:57

---

|                   |       |       |                  |
|-------------------|-------|-------|------------------|
| Diagnostics Check | OK    | 21:02 |                  |
| Air Blank         | 0.000 | 21:03 |                  |
| Control Test      | 0.079 | 21:03 |                  |
| Air Blank         | 0.000 | 21:04 |                  |
| Subject Sample    | # 1   | 21:05 | Sequence Aborted |
| Breath Volume     | 0.000 | 21:05 |                  |
| Air Blank         | 0.000 | 21:05 |                  |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,IVAN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HERNANDEZ,IVAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCENDREE

DATE/TIME OF TEST: 04/08/2024 22:49:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.079 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.204 | 22:44 |
| Breath Volume      | 3.734 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Air Blank          | 0.000 | 22:47 |
| Subject Sample # 2 | 0.205 | 22:48 |
| Breath Volume      | 2.535 | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.078 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Diagnostics Check  | OK    | 22:49 |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/12/2024 04:04:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.079 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 1 | 0.119 | 03:59 |
| Breath Volume      | 1.761 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.115 | 04:03 |
| Breath Volume      | 2.226 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.079 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,IVAN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HERNANDEZ,IVAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:24

SUBJECT'S LAST NAME: BECKWITH

DATE/TIME OF TEST: 04/12/2024 15:51:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:45 |
| Air Blank          | 0.000 | 15:45 |
| Control Test       | 0.080 | 15:45 |
| Air Blank          | 0.000 | 15:46 |
| Subject Sample # 1 | 0.118 | 15:46 |
| Breath Volume      | 1.593 | 15:46 |
| Air Blank          | 0.000 | 15:47 |
| Air Blank          | 0.000 | 15:49 |
| Subject Sample # 2 | 0.115 | 15:49 |
| Breath Volume      | 1.523 | 15:49 |
| Air Blank          | 0.000 | 15:50 |
| Control Test       | 0.078 | 15:50 |
| Air Blank          | 0.000 | 15:51 |
| Diagnostics Check  | OK    | 15:51 |

---

Data Download Date/Time: 04/27/2024 18:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 04/13/2024 03:48:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Control Test       | 0.080 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 1 | 0.178 | 03:43 |
| Breath Volume      | 2.781 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 2 | 0.183 | 03:46 |
| Breath Volume      | 2.699 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.079 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 04/27/2024 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STIERS,SETH,T  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KELCHNER,M  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 04/18/2024 01:27:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:20 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.079 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Subject Sample # 1 | 0.113 | 01:22 |
| Breath Volume      | 2.507 | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Subject Sample # 2 | 0.102 | 01:26 |
| Breath Volume      | 1.582 | 01:26 |
| Air Blank          | 0.000 | 01:26 |
| Control Test       | 0.078 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Diagnostics Check  | OK    | 01:27 |

---

Data Download Date/Time: 04/27/2024 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,OFFICER,G  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: GILES,OFFICER  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VIZCAINO

DATE/TIME OF TEST: 04/19/2024 03:35:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:30 |                      |
| Air Blank         | 0.000 | 03:31 |                      |
| Control Test      | 0.079 | 03:31 |                      |
| Air Blank         | 0.000 | 03:31 |                      |
| Subject Sample    | # 1   | 03:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:33 |                      |
| Air Blank         | 0.000 | 03:34 |                      |
| Control Test      | 0.079 | 03:34 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Diagnostics Check | OK    | 03:35 |                      |

---

Data Download Date/Time: 04/27/2024 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,IVAN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: HERNANDEZ,IVAN  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 04/24/2024 00:37:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:29 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.079 | 00:30 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 1 | 0.136 | 00:32 |
| Breath Volume      | 1.808 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 2 | 0.141 | 00:35 |
| Breath Volume      | 2.660 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.080 | 00:36 |
| Air Blank          | 0.000 | 00:37 |
| Diagnostics Check  | OK    | 00:37 |

---

Data Download Date/Time: 04/27/2024 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/26/2024 04:01:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.080 | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 1 | 0.275 | 03:56 |
| Breath Volume      | 2.765 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Subject Sample # 2 | 0.268 | 03:59 |
| Breath Volume      | 2.808 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.076 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 04/27/2024 18:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 02:15:52

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:05 |                    |
| Air Blank         | 0.000 | 02:06 |                    |
| Control Test      | 0.080 | 02:06 |                    |
| Air Blank         | 0.000 | 02:07 |                    |
| Subject Sample    | # 1   | 02:10 | No Sample Provided |
| Breath Volume     | 0.000 | 02:10 |                    |
| Air Blank         | 0.000 | 02:10 |                    |
| Air Blank         | 0.000 | 02:12 |                    |
| Subject Sample    | # 2   | 02:14 |                    |
| Breath Volume     | 1.609 | 02:14 |                    |
| Air Blank         |       | 02:15 | Purge Fail         |
| Air Blank         |       | 02:15 | Purge Fail         |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 02:19:04

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:17 |                           |
| Air Blank         | 0.000 | 02:18 |                           |
| Control Test      | 0.047 | 02:18 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:19 |                           |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 02:25:47

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:24 |                           |
| Air Blank         | 0.000 | 02:24 |                           |
| Control Test      | 0.055 | 02:25 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:25 |                           |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 02:29:19

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:27 |                           |
| Air Blank         | 0.000 | 02:28 |                           |
| Control Test      | 0.061 | 02:28 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:29 |                           |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROLICK,ALEX,M  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROLICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 02:58:30

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 02:50 |            |
| Air Blank          | 0.000 | 02:50 |            |
| Control Test       | 0.078 | 02:51 |            |
| Air Blank          | 0.000 | 02:51 |            |
| Subject Sample # 1 | 0.133 | 02:52 |            |
| Breath Volume      | 1.265 | 02:52 |            |
| Air Blank          | 0.000 | 02:53 |            |
| Air Blank          | 0.000 | 02:55 |            |
| Subject Sample # 2 | 0.157 | 02:56 |            |
| Breath Volume      | 2.328 | 02:56 |            |
| Air Blank          |       | 02:57 | RFI Detect |
| Air Blank          | 0.000 | 02:57 |            |
| Control Test       | 0.077 | 02:57 |            |
| Air Blank          | 0.000 | 02:58 |            |
| Diagnostics Check  | OK    | 02:58 |            |

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUIS,JUSTIN,  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KROICK,ALEX  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/28/2024 03:07:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:01 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.078 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Subject Sample # 1 | 0.144 | 03:03 |
| Breath Volume      | 2.234 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Subject Sample # 2 | 0.152 | 03:06 |
| Breath Volume      | 2.121 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.077 | 03:07 |
| Air Blank          | 0.000 | 03:07 |
| Diagnostics Check  | OK    | 03:07 |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WOODIE,OFC,L  
BREATH TEST OPERATOR AGENCY: Clearwater PD  
ARRESTING OFFICER: KELCHNER,M  
ARRESTING OFFICER AGENCY: Clearwater PD  
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROUAN

DATE/TIME OF TEST: 04/29/2024 01:09:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:00 |                      |
| Air Blank         | 0.000 | 01:01 |                      |
| Control Test      | 0.079 | 01:01 |                      |
| Air Blank         | 0.000 | 01:02 |                      |
| Subject Sample    | # 1   | 01:05 | No Sample Provided   |
| Breath Volume     | 0.000 | 01:05 |                      |
| Air Blank         | 0.000 | 01:05 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Subject Sample    | # 2   | 01:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:07 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Control Test      | 0.079 | 01:08 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Diagnostics Check | OK    | 01:09 |                      |

---

Data Download Date/Time: 05/30/2024 22:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DESUYO,AISON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/01/2024 00:04:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Control Test       | 0.081 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.082 | 23:58 |
| Breath Volume      | 1.355 | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.078 | 00:02 |
| Breath Volume      | 1.304 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Control Test       | 0.080 | 00:03 |
| Air Blank          | 0.000 | 00:04 |
| Diagnostics Check  | OK    | 00:04 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BARRERAS,KIMBERLY  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SIBADA SIBADA

DATE/TIME OF TEST: 04/01/2024 02:21:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Control Test       | 0.080 | 02:15 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 1 | 0.110 | 02:16 |
| Breath Volume      | 1.679 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 2 | 0.112 | 02:19 |
| Breath Volume      | 3.347 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.080 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Diagnostics Check  | OK    | 02:21 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: REID,MICHAEL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TOMAS FRANCISCO

DATE/TIME OF TEST: 04/01/2024 06:06:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:59 |
| Air Blank          | 0.000 | 06:00 |
| Control Test       | 0.080 | 06:00 |
| Air Blank          | 0.000 | 06:00 |
| Subject Sample # 1 | 0.221 | 06:01 |
| Breath Volume      | 2.640 | 06:01 |
| Air Blank          | 0.000 | 06:02 |
| Air Blank          | 0.000 | 06:04 |
| Subject Sample # 2 | 0.211 | 06:05 |
| Breath Volume      | 2.890 | 06:05 |
| Air Blank          | 0.000 | 06:05 |
| Control Test       | 0.080 | 06:06 |
| Air Blank          | 0.000 | 06:06 |
| Diagnostics Check  | OK    | 06:06 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TRENTMAN,THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 04/03/2024 09:13:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:05 |
| Air Blank          | 0.000 | 09:05 |
| Control Test       | 0.081 | 09:05 |
| Air Blank          | 0.000 | 09:06 |
| Subject Sample # 1 | 0.000 | 09:07 |
| Breath Volume      | 1.652 | 09:07 |
| Air Blank          | 0.000 | 09:08 |
| Air Blank          | 0.000 | 09:10 |
| Subject Sample # 2 | 0.000 | 09:11 |
| Breath Volume      | 2.542 | 09:11 |
| Air Blank          | 0.000 | 09:12 |
| Control Test       | 0.081 | 09:12 |
| Air Blank          | 0.000 | 09:12 |
| Diagnostics Check  | OK    | 09:13 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: HOLLIGAN,DANIEL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 12:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: MARTONE

DATE/TIME OF TEST: 04/04/2024 15:01:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:54 |
| Air Blank          | 0.000 | 14:55 |
| Control Test       | 0.080 | 14:55 |
| Air Blank          | 0.000 | 14:56 |
| Subject Sample # 1 | 0.153 | 14:56 |
| Breath Volume      | 2.023 | 14:56 |
| Air Blank          | 0.000 | 14:57 |
| Air Blank          | 0.000 | 14:59 |
| Subject Sample # 2 | 0.158 | 14:59 |
| Breath Volume      | 1.578 | 14:59 |
| Air Blank          | 0.000 | 15:00 |
| Control Test       | 0.079 | 15:00 |
| Air Blank          | 0.000 | 15:01 |
| Diagnostics Check  | OK    | 15:01 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN, RENEE, M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TAPPIN, KEVIN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 04/05/2024 20:57:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Control Test       | 0.080 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Subject Sample # 1 | 0.343 | 20:52 |
| Breath Volume      | 2.289 | 20:52 |
| Air Blank          | 0.000 | 20:52 |
| Air Blank          | 0.000 | 20:54 |
| Subject Sample # 2 | 0.333 | 20:55 |
| Breath Volume      | 2.089 | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Control Test       | 0.079 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Diagnostics Check  | OK    | 20:57 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: THOMAS, MICAH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: CANTY

DATE/TIME OF TEST: 04/06/2024 00:39:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.080 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 1 | 0.112 | 00:34 |
| Breath Volume      | 1.949 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:36 |
| Subject Sample # 2 | 0.113 | 00:37 |
| Breath Volume      | 1.972 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.079 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Diagnostics Check  | OK    | 00:39 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DESUYO, AISON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: SANTOS FLORES

DATE/TIME OF TEST: 04/06/2024 08:32:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:25 |
| Air Blank          | 0.000 | 08:26 |
| Control Test       | 0.080 | 08:26 |
| Air Blank          | 0.000 | 08:26 |
| Subject Sample # 1 | 0.109 | 08:27 |
| Breath Volume      | 1.816 | 08:27 |
| Air Blank          | 0.000 | 08:28 |
| Air Blank          | 0.000 | 08:30 |
| Subject Sample # 2 | 0.117 | 08:30 |
| Breath Volume      | 3.449 | 08:30 |
| Air Blank          | 0.000 | 08:31 |
| Control Test       | 0.080 | 08:31 |
| Air Blank          | 0.000 | 08:32 |
| Diagnostics Check  | OK    | 08:32 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CARVER,CAMERON  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LUCAS MENDEZ

DATE/TIME OF TEST: 04/06/2024 23:47:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.081 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.163 | 23:41 |
| Breath Volume      | 1.664 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 2 | 0.167 | 23:45 |
| Breath Volume      | 2.183 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.080 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Diagnostics Check  | OK    | 23:47 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NOEL,KENNETH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: EUCEDA

DATE/TIME OF TEST: 04/07/2024 01:35:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Control Test       | 0.080 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Subject Sample # 1 | 0.190 | 01:31 |
| Breath Volume      | 2.898 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 2 | 0.195 | 01:34 |
| Breath Volume      | 3.140 | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Control Test       | 0.080 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Diagnostics Check  | OK    | 01:35 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KENNY, SHANE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GRANDA ECHEVERRIA

DATE/TIME OF TEST: 04/07/2024 02:34:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Control Test       | 0.080 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 1 | 0.184 | 02:30 |
| Breath Volume      | 3.820 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Air Blank          | 0.000 | 02:32 |
| Subject Sample # 2 | 0.180 | 02:33 |
| Breath Volume      | 3.589 | 02:33 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.079 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Diagnostics Check  | OK    | 02:34 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: BONVOULOIR

DATE/TIME OF TEST: 04/07/2024 03:27:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Control Test       | 0.080 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.155 | 03:22 |
| Breath Volume      | 2.917 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.155 | 03:25 |
| Breath Volume      | 2.113 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.079 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MELLERS,KEVIN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LASTER

DATE/TIME OF TEST: 04/07/2024 04:51:07

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:42 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Control Test      | 0.079 | 04:43 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Subject Sample    | # 1   | 04:47 | No Sample Provided   |
| Breath Volume     | 0.000 | 04:47 |                      |
| Air Blank         | 0.000 | 04:47 |                      |
| Air Blank         | 0.000 | 04:49 |                      |
| Subject Sample    | # 2   | 04:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:49 |                      |
| Air Blank         | 0.000 | 04:50 |                      |
| Control Test      | 0.080 | 04:50 |                      |
| Air Blank         | 0.000 | 04:51 |                      |
| Diagnostics Check | OK    | 04:51 |                      |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SNELGROVE,DYLAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 04/07/2024 06:13:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:06 |
| Air Blank          | 0.000 | 06:06 |
| Control Test       | 0.080 | 06:07 |
| Air Blank          | 0.000 | 06:07 |
| Subject Sample # 1 | 0.107 | 06:08 |
| Breath Volume      | 1.578 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.109 | 06:11 |
| Breath Volume      | 1.796 | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.080 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Diagnostics Check  | OK    | 06:13 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NOEL,KENNETH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 04/08/2024 00:52:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.080 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 1 | 0.107 | 00:48 |
| Breath Volume      | 1.925 | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 2 | 0.095 | 00:51 |
| Breath Volume      | 1.164 | 00:51 |
| Air Blank          | 0.000 | 00:51 |
| Control Test       | 0.081 | 00:52 |
| Air Blank          | 0.000 | 00:52 |
| Diagnostics Check  | OK    | 00:52 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PICARD, RONNIE  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HARLAN

DATE/TIME OF TEST: 04/08/2024 02:17:34

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:04 |                |
| Air Blank         | 0.000     | 02:05 |                |
| Control Test      | 0.080     | 02:05 |                |
| Air Blank         | 0.000     | 02:06 |                |
| Subject Sample    | # 1 0.181 | 02:09 | Volume Not Met |
| Breath Volume     | 0.945     | 02:09 |                |
| Air Blank         | 0.000     | 02:10 |                |
| Air Blank         | 0.000     | 02:11 |                |
| Subject Sample    | # 2 0.186 | 02:12 |                |
| Breath Volume     | 1.894     | 02:12 |                |
| Air Blank         | 0.000     | 02:13 |                |
| Air Blank         | 0.000     | 02:14 |                |
| Subject Sample    | # 3 0.171 | 02:15 |                |
| Breath Volume     | 2.042     | 02:15 |                |
| Air Blank         | 0.000     | 02:16 |                |
| Control Test      | 0.078     | 02:16 |                |
| Air Blank         | 0.000     | 02:17 |                |
| Diagnostics Check | OK        | 02:17 |                |

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAOGLU,YASEMIN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 04/08/2024 03:47:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.080 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.290 | 03:39 |
| Breath Volume      | 1.585 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 2 | 0.256 | 03:42 |
| Breath Volume      | 1.347 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 3 | 0.269 | 03:45 |
| Breath Volume      | 2.230 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.079 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:47 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MACK,CHEYENNE  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: CAPPIELLO

DATE/TIME OF TEST: 04/08/2024 12:52:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:45 |
| Air Blank          | 0.000 | 12:46 |
| Control Test       | 0.080 | 12:46 |
| Air Blank          | 0.000 | 12:47 |
| Subject Sample # 1 | 0.000 | 12:47 |
| Breath Volume      | 1.160 | 12:47 |
| Air Blank          | 0.000 | 12:48 |
| Air Blank          | 0.000 | 12:50 |
| Subject Sample # 2 | 0.000 | 12:51 |
| Breath Volume      | 2.269 | 12:51 |
| Air Blank          | 0.000 | 12:51 |
| Control Test       | 0.081 | 12:52 |
| Air Blank          | 0.000 | 12:52 |
| Diagnostics Check  | OK    | 12:52 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MILORD,JHAMIL  
ARRESTING OFFICER AGENCY: North Palm Beach  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: OTERO ZIRION

DATE/TIME OF TEST: 04/09/2024 02:14:57

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 02:00 |                      |
| Air Blank         | 0.000     | 02:00 |                      |
| Control Test      | 0.081     | 02:01 |                      |
| Air Blank         | 0.000     | 02:01 |                      |
| Subject Sample    | # 1 0.170 | 02:04 | Volume Not Met       |
| Breath Volume     | 0.000     | 02:04 |                      |
| Air Blank         | 0.000     | 02:05 |                      |
| Air Blank         | 0.000     | 02:07 |                      |
| Subject Sample    | # 2 0.168 | 02:07 |                      |
| Breath Volume     | 1.140     | 02:07 |                      |
| Air Blank         | 0.000     | 02:08 |                      |
| Air Blank         | 0.000     | 02:10 |                      |
| Subject Sample    | # 3       | 02:13 | Subject Test Refused |
| Breath Volume     | 0.042     | 02:13 |                      |
| Air Blank         | 0.000     | 02:13 |                      |
| Control Test      | 0.079     | 02:14 |                      |
| Air Blank         | 0.000     | 02:14 |                      |
| Diagnostics Check | OK        | 02:14 |                      |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DENIER, MICHAEL  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MONTEJO DOMINGO

DATE/TIME OF TEST: 04/09/2024 04:02:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.080 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 1 | 0.176 | 03:57 |
| Breath Volume      | 1.574 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 2 | 0.165 | 04:01 |
| Breath Volume      | 1.179 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.079 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Diagnostics Check  | OK    | 04:02 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAOGLU,YASEMIN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KERCE

DATE/TIME OF TEST: 04/10/2024 03:45:34

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:33 |                |
| Air Blank         | 0.000     | 03:33 |                |
| Control Test      | 0.080     | 03:34 |                |
| Air Blank         | 0.000     | 03:34 |                |
| Subject Sample    | # 1 0.215 | 03:35 |                |
| Breath Volume     | 1.203     | 03:35 |                |
| Air Blank         | 0.000     | 03:35 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Subject Sample    | # 2 0.174 | 03:40 | Volume Not Met |
| Breath Volume     | 0.296     | 03:40 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Air Blank         | 0.000     | 03:43 |                |
| Subject Sample    | # 3 0.199 | 03:43 |                |
| Breath Volume     | 1.343     | 03:43 |                |
| Air Blank         | 0.000     | 03:44 |                |
| Control Test      | 0.079     | 03:44 |                |
| Air Blank         | 0.000     | 03:45 |                |
| Diagnostics Check | OK        | 03:45 |                |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HEISING

DATE/TIME OF TEST: 04/11/2024 23:43:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Control Test       | 0.080 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Subject Sample # 1 | 0.083 | 23:39 |
| Breath Volume      | 2.492 | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Air Blank          | 0.000 | 23:41 |
| Subject Sample # 2 | 0.087 | 23:42 |
| Breath Volume      | 3.117 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Control Test       | 0.080 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Diagnostics Check  | OK    | 23:43 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GONZALEZ,ROBERTO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 09:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: MELGAR MELGAR

DATE/TIME OF TEST: 04/12/2024 10:44:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 10:34 |                      |
| Air Blank         | 0.000 | 10:35 |                      |
| Control Test      | 0.080 | 10:35 |                      |
| Air Blank         | 0.000 | 10:36 |                      |
| Subject Sample    | # 1   | 10:39 | No Sample Provided   |
| Breath Volume     | 0.000 | 10:39 |                      |
| Air Blank         | 0.000 | 10:40 |                      |
| Air Blank         | 0.000 | 10:41 |                      |
| Subject Sample    | # 2   | 10:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 10:43 |                      |
| Air Blank         | 0.000 | 10:44 |                      |
| Control Test      | 0.079 | 10:44 |                      |
| Air Blank         | 0.000 | 10:44 |                      |
| Diagnostics Check | OK    | 10:44 |                      |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: TRENTMAN,THOMAS  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:38

SUBJECT'S LAST NAME: PALACIOS ESCAMILLA

DATE/TIME OF TEST: 04/12/2024 12:23:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:16 |
| Air Blank          | 0.000 | 12:16 |
| Control Test       | 0.080 | 12:17 |
| Air Blank          | 0.000 | 12:17 |
| Subject Sample # 1 | 0.249 | 12:18 |
| Breath Volume      | 1.402 | 12:18 |
| Air Blank          | 0.000 | 12:19 |
| Air Blank          | 0.000 | 12:21 |
| Subject Sample # 2 | 0.260 | 12:21 |
| Breath Volume      | 3.125 | 12:21 |
| Air Blank          | 0.000 | 12:22 |
| Control Test       | 0.080 | 12:22 |
| Air Blank          | 0.000 | 12:23 |
| Diagnostics Check  | OK    | 12:23 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SNELGROVE,DYLAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: JUCHNIEWICZ

DATE/TIME OF TEST: 04/13/2024 01:28:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.079 | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Subject Sample # 1 | 0.096 | 01:24 |
| Breath Volume      | 2.203 | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 2 | 0.092 | 01:27 |
| Breath Volume      | 1.875 | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.079 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Diagnostics Check  | OK    | 01:28 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 04/13/2024 03:14:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Control Test       | 0.080 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 1 | 0.165 | 03:08 |
| Breath Volume      | 1.375 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Air Blank          | 0.000 | 03:11 |
| Subject Sample # 2 | 0.177 | 03:12 |
| Breath Volume      | 2.023 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.080 | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Diagnostics Check  | OK    | 03:14 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: NOEL,KENNETH  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 04/13/2024 23:49:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.080 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 1 | 0.231 | 23:45 |
| Breath Volume      | 2.652 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 2 | 0.220 | 23:48 |
| Breath Volume      | 1.875 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Control Test       | 0.079 | 23:49 |
| Air Blank          | 0.000 | 23:49 |
| Diagnostics Check  | OK    | 23:49 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: ABBOTT,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ESPINOZA TAPIA

DATE/TIME OF TEST: 04/14/2024 01:58:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.080 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 1 | 0.146 | 01:52 |
| Breath Volume      | 1.171 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 2 | 0.143 | 01:56 |
| Breath Volume      | 1.589 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.079 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Diagnostics Check  | OK    | 01:58 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT,ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 04/14/2024 03:22:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Control Test       | 0.080 | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 1 | 0.161 | 03:18 |
| Breath Volume      | 1.421 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.143 | 03:21 |
| Breath Volume      | 1.199 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.079 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: EDWARDS,DERRICK  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: URIBE

DATE/TIME OF TEST: 04/14/2024 04:50:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Control Test       | 0.079 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 1 | 0.163 | 04:45 |
| Breath Volume      | 2.035 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Air Blank          | 0.000 | 04:48 |
| Subject Sample # 2 | 0.155 | 04:48 |
| Breath Volume      | 1.738 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Control Test       | 0.080 | 04:49 |
| Air Blank          | 0.000 | 04:50 |
| Diagnostics Check  | OK    | 04:50 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SOTO,MARC  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 04/15/2024 03:40:07

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:28 |                    |
| Air Blank         | 0.000 | 03:28 |                    |
| Control Test      | 0.081 | 03:28 |                    |
| Air Blank         | 0.000 | 03:29 |                    |
| Subject Sample    | # 1   | 03:32 | No Sample Provided |
| Breath Volume     | 0.000 | 03:32 |                    |
| Air Blank         | 0.000 | 03:33 |                    |
| Air Blank         | 0.000 | 03:35 |                    |
| Subject Sample    | # 2   | 0.247 | 03:35              |
| Breath Volume     | 1.156 | 03:35 |                    |
| Air Blank         | 0.000 | 03:36 |                    |
| Air Blank         | 0.000 | 03:38 |                    |
| Subject Sample    | # 3   | 0.248 | 03:38              |
| Breath Volume     | 1.191 | 03:38 |                    |
| Air Blank         | 0.000 | 03:39 |                    |
| Control Test      | 0.080 | 03:39 |                    |
| Air Blank         | 0.000 | 03:40 |                    |
| Diagnostics Check | OK    | 03:40 |                    |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WASHINGTON,GABRIELLE  
ARRESTING OFFICER AGENCY: Delray Beach PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 04/15/2024 21:16:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:09 |
| Air Blank          | 0.000 | 21:10 |
| Control Test       | 0.081 | 21:10 |
| Air Blank          | 0.000 | 21:10 |
| Subject Sample # 1 | 0.229 | 21:11 |
| Breath Volume      | 2.500 | 21:11 |
| Air Blank          | 0.000 | 21:12 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 2 | 0.230 | 21:15 |
| Breath Volume      | 2.664 | 21:15 |
| Air Blank          | 0.000 | 21:15 |
| Control Test       | 0.079 | 21:16 |
| Air Blank          | 0.000 | 21:16 |
| Diagnostics Check  | OK    | 21:16 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: GALLOWAY,SEAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 04/16/2024 21:52:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.080 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 1 | 0.082 | 21:48 |
| Breath Volume      | 2.414 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 2 | 0.079 | 21:51 |
| Breath Volume      | 2.468 | 21:51 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.080 | 21:52 |
| Air Blank          | 0.000 | 21:52 |
| Diagnostics Check  | OK    | 21:52 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DENIER,MICHAEL  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RAMOS RAMIREZ

DATE/TIME OF TEST: 04/17/2024 05:15:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Control Test       | 0.081 | 05:08 |
| Air Blank          | 0.000 | 05:09 |
| Subject Sample # 1 | 0.158 | 05:10 |
| Breath Volume      | 2.757 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Air Blank          | 0.000 | 05:13 |
| Subject Sample # 2 | 0.165 | 05:14 |
| Breath Volume      | 1.828 | 05:14 |
| Air Blank          | 0.000 | 05:14 |
| Control Test       | 0.080 | 05:15 |
| Air Blank          | 0.000 | 05:15 |
| Diagnostics Check  | OK    | 05:15 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: MCGILLICUDDY,STEVEN  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 04/18/2024 15:40:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:33 |
| Air Blank          | 0.000 | 15:34 |
| Control Test       | 0.080 | 15:34 |
| Air Blank          | 0.000 | 15:35 |
| Subject Sample # 1 | 0.261 | 15:35 |
| Breath Volume      | 3.164 | 15:35 |
| Air Blank          | 0.000 | 15:36 |
| Air Blank          | 0.000 | 15:38 |
| Subject Sample # 2 | 0.259 | 15:38 |
| Breath Volume      | 2.953 | 15:38 |
| Air Blank          | 0.000 | 15:39 |
| Control Test       | 0.080 | 15:39 |
| Air Blank          | 0.000 | 15:40 |
| Diagnostics Check  | OK    | 15:40 |

---

Data Download Date/Time: 04/19/2024 14:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: AULD

DATE/TIME OF TEST: 04/02/2024 00:48:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.080 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.071 | 00:44 |
| Breath Volume      | 2.171 | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 2 | 0.069 | 00:47 |
| Breath Volume      | 2.070 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Control Test       | 0.080 | 00:48 |
| Air Blank          | 0.000 | 00:48 |
| Diagnostics Check  | OK    | 00:48 |

---

Data Download Date/Time: 04/08/2024 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: HYATT

DATE/TIME OF TEST: 04/02/2024 23:25:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:13 |                      |
| Air Blank         | 0.000 | 23:13 |                      |
| Control Test      | 0.079 | 23:14 |                      |
| Air Blank         | 0.000 | 23:14 |                      |
| Subject Sample    | # 1   | 23:17 | No Sample Provided   |
| Breath Volume     | 0.000 | 23:17 |                      |
| Air Blank         | 0.000 | 23:18 |                      |
| Air Blank         | 0.000 | 23:20 |                      |
| Subject Sample    | # 2   | 0.197 | 23:23 Volume Not Met |
| Breath Volume     | 0.394 | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Control Test      | 0.080 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Diagnostics Check | OK    | 23:25 |                      |

---

Data Download Date/Time: 04/08/2024 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRADEN,JASON,M  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: PHIPPS

DATE/TIME OF TEST: 04/05/2024 23:29:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:22 |
| Air Blank          | 0.000 | 23:22 |
| Control Test       | 0.080 | 23:23 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 1 | 0.230 | 23:24 |
| Breath Volume      | 1.128 | 23:24 |
| Air Blank          | 0.000 | 23:24 |
| Air Blank          | 0.000 | 23:26 |
| Subject Sample # 2 | 0.231 | 23:27 |
| Breath Volume      | 1.828 | 23:27 |
| Air Blank          | 0.000 | 23:28 |
| Control Test       | 0.080 | 23:28 |
| Air Blank          | 0.000 | 23:28 |
| Diagnostics Check  | OK    | 23:29 |

---

Data Download Date/Time: 04/08/2024 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BRADEN,JASON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HAGELE

DATE/TIME OF TEST: 04/06/2024 22:37:35

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:25 |                |
| Air Blank         | 0.000     | 22:26 |                |
| Control Test      | 0.080     | 22:26 |                |
| Air Blank         | 0.000     | 22:26 |                |
| Subject Sample    | # 1 0.163 | 22:30 | Volume Not Met |
| Breath Volume     | 0.132     | 22:30 |                |
| Air Blank         | 0.000     | 22:30 |                |
| Air Blank         | 0.000     | 22:32 |                |
| Subject Sample    | # 2 0.181 | 22:35 | Volume Not Met |
| Breath Volume     | 0.515     | 22:35 |                |
| Air Blank         | 0.000     | 22:36 |                |
| Control Test      | 0.079     | 22:36 |                |
| Air Blank         | 0.000     | 22:37 |                |
| Diagnostics Check | OK        | 22:37 |                |

---

Data Download Date/Time: 04/08/2024 16:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: BISCHOFF

DATE/TIME OF TEST: 04/11/2024 02:02:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:55 |
| Air Blank          | 0.000 | 01:55 |
| Control Test       | 0.079 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Subject Sample # 1 | 0.150 | 01:57 |
| Breath Volume      | 2.933 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 2 | 0.147 | 02:00 |
| Breath Volume      | 2.707 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.078 | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Diagnostics Check  | OK    | 02:02 |

---

Data Download Date/Time: 05/13/2024 09:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: NOVAK,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: SCHMOTZER

DATE/TIME OF TEST: 04/16/2024 19:02:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Control Test       | 0.079 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Subject Sample # 1 | 0.193 | 18:56 |
| Breath Volume      | 1.800 | 18:56 |
| Air Blank          | 0.000 | 18:57 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.195 | 19:00 |
| Breath Volume      | 1.800 | 19:00 |
| Air Blank          | 0.000 | 19:01 |
| Control Test       | 0.077 | 19:01 |
| Air Blank          | 0.000 | 19:01 |
| Diagnostics Check  | OK    | 19:02 |

---

Data Download Date/Time: 05/13/2024 09:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOVAK,CHRISTOPHER,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: BRADEN,JASON  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: PEREZ GOMEZ

DATE/TIME OF TEST: 04/20/2024 00:03:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.079 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.186 | 23:58 |
| Breath Volume      | 1.171 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Air Blank          | 0.000 | 00:00 |
| Subject Sample # 1 | 0.186 | 00:01 |
| Breath Volume      | 1.824 | 00:01 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.078 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Diagnostics Check  | OK    | 00:03 |

---

Data Download Date/Time: 05/13/2024 09:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/05/2024 21:14:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:08 |
| Air Blank          | 0.000 | 21:09 |
| Control Test       | 0.081 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Subject Sample # 1 | 0.105 | 21:10 |
| Breath Volume      | 1.855 | 21:10 |
| Air Blank          | 0.000 | 21:10 |
| Air Blank          | 0.000 | 21:12 |
| Subject Sample # 2 | 0.102 | 21:13 |
| Breath Volume      | 1.519 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Control Test       | 0.080 | 21:14 |
| Air Blank          | 0.000 | 21:14 |
| Diagnostics Check  | OK    | 21:14 |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: FORTIN

DATE/TIME OF TEST: 04/05/2024 22:34:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.081 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 1 | 0.000 | 22:30 |
| Breath Volume      | 2.128 | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 2 | 0.000 | 22:33 |
| Breath Volume      | 2.320 | 22:33 |
| Air Blank          | 0.000 | 22:33 |
| Control Test       | 0.081 | 22:34 |
| Air Blank          | 0.000 | 22:34 |
| Diagnostics Check  | OK    | 22:34 |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 04/06/2024 02:24:06

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:22 |            |
| Air Blank         | 0.000 | 02:23 |            |
| Control Test      | 0.081 | 02:23 |            |
| Air Blank         |       | 02:23 | RFI Detect |
| Air Blank         |       | 02:24 | RFI Detect |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 04/06/2024 02:28:17

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:26 |            |
| Air Blank         | 0.000 | 02:26 |            |
| Control Test      | 0.081 | 02:26 |            |
| Air Blank         | 0.000 | 02:27 |            |
| Subject Sample    | # 1   | 02:27 | RFI Detect |
| Breath Volume     | 0.000 | 02:27 |            |
| Air Blank         |       | 02:27 | RFI Detect |
| Air Blank         | 0.000 | 02:28 |            |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 04/06/2024 02:35:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.081 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.149 | 02:31 |
| Breath Volume      | 1.679 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.142 | 02:34 |
| Breath Volume      | 1.316 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.080 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 04/14/2024 23:07:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Control Test       | 0.081 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Subject Sample # 1 | 0.153 | 23:03 |
| Breath Volume      | 2.859 | 23:03 |
| Air Blank          | 0.000 | 23:03 |
| Air Blank          | 0.000 | 23:05 |
| Subject Sample # 2 | 0.161 | 23:05 |
| Breath Volume      | 3.339 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Control Test       | 0.080 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Diagnostics Check  | OK    | 23:07 |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: WEATHERFORD

DATE/TIME OF TEST: 04/20/2024 02:18:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.080 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 1 | 0.098 | 02:13 |
| Breath Volume      | 1.953 | 02:13 |
| Air Blank          | 0.000 | 02:14 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 2 | 0.103 | 02:16 |
| Breath Volume      | 2.738 | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.080 | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Diagnostics Check  | OK    | 02:18 |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: DE LEON PEREZ

DATE/TIME OF TEST: 04/21/2024 20:54:59

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 20:53 |            |
| Air Blank         | 0.000 | 20:54 |            |
|                   |       | 20:54 | RFI Detect |
| Air Blank         | 0.000 | 20:54 |            |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 04/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: DE LEON PEREZ

DATE/TIME OF TEST: 04/21/2024 21:05:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:55 |                |
| Air Blank         | 0.000     | 20:56 |                |
| Control Test      | 0.079     | 20:56 |                |
| Air Blank         | 0.000     | 20:57 |                |
| Subject Sample    | # 1 0.290 | 21:00 | Volume Not Met |
| Breath Volume     | 1.039     | 21:00 |                |
| Air Blank         | 0.000     | 21:00 |                |
| Air Blank         | 0.000     | 21:02 |                |
| Subject Sample    | # 2       | 21:04 | RFI Detect     |
| Breath Volume     | 0.000     | 21:04 |                |
| Air Blank         |           | 21:04 | RFI Detect     |
| Air Blank         | 0.000     | 21:05 |                |

---

Data Download Date/Time: 05/03/2024 11:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ISHIKAWA,JORDAN,L  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: ROONEY,ZACHARY  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A4  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: BERROWS

DATE/TIME OF TEST: 04/04/2024 22:29:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.081 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 1 | 0.150 | 22:25 |
| Breath Volume      | 1.398 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 2 | 0.158 | 22:28 |
| Breath Volume      | 1.726 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.080 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Diagnostics Check  | OK    | 22:29 |

---

Data Download Date/Time: 04/29/2024 11:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: DIAMOND,SCOTT  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A4  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: GLOR

DATE/TIME OF TEST: 04/07/2024 17:21:12

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 17:19 |            |
| Air Blank         | 0.000 | 17:20 |            |
| Control Test      | 0.080 | 17:20 |            |
| Air Blank         |       | 17:21 | RFI Detect |
| Air Blank         |       | 17:21 | RFI Detect |

---

Data Download Date/Time: 04/29/2024 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: DIAMOND,SCOTT  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A4  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: GLOR

DATE/TIME OF TEST: 04/07/2024 17:36:14

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 17:22 |                |
| Air Blank         | 0.000     | 17:23 |                |
| Control Test      | 0.081     | 17:23 |                |
| Air Blank         | 0.000     | 17:24 |                |
| Subject Sample    | # 1 0.184 | 17:27 | Volume Not Met |
| Breath Volume     | 0.605     | 17:27 |                |
| Air Blank         | 0.000     | 17:28 |                |
| Air Blank         | 0.000     | 17:29 |                |
| Subject Sample    | # 2 0.189 | 17:31 |                |
| Breath Volume     | 2.105     | 17:31 |                |
| Air Blank         | 0.000     | 17:32 |                |
| Air Blank         | 0.000     | 17:34 |                |
| Subject Sample    | # 3 0.183 | 17:34 |                |
| Breath Volume     | 1.371     | 17:34 |                |
| Air Blank         | 0.000     | 17:35 |                |
| Control Test      | 0.075     | 17:35 |                |
| Air Blank         | 0.000     | 17:36 |                |
| Diagnostics Check | OK        | 17:36 |                |

---

Data Download Date/Time: 04/29/2024 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAY,ANTHONY,D  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: RAY,ANTHONY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A4  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 04/13/2024 00:42:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.081 | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 1 | 0.120 | 00:37 |
| Breath Volume      | 2.460 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Subject Sample # 2 | 0.118 | 00:40 |
| Breath Volume      | 2.218 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.078 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Diagnostics Check  | OK    | 00:42 |

---

Data Download Date/Time: 04/29/2024 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES, LANDON, A  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: MELENDEZ, TIMOTHY  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 03/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A4  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 04/19/2024 23:48:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Control Test       | 0.080 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 1 | 0.232 | 23:40 |
| Breath Volume      | 2.464 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 2 | 0.209 | 23:43 |
| Breath Volume      | 1.375 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 3 | 0.215 | 23:47 |
| Breath Volume      | 1.867 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.078 | 23:48 |
| Air Blank          | 0.000 | 23:48 |
| Diagnostics Check  | OK    | 23:48 |

---

Data Download Date/Time: 04/29/2024 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAGOS,GUILLERMO,E  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BLANCO,JONATHAN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: OBANDO

DATE/TIME OF TEST: 04/04/2024 20:32:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Control Test       | 0.080 | 20:25 |
| Air Blank          | 0.000 | 20:26 |
| Subject Sample # 1 | 0.294 | 20:27 |
| Breath Volume      | 2.410 | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Air Blank          | 0.000 | 20:30 |
| Subject Sample # 2 | 0.281 | 20:31 |
| Breath Volume      | 1.484 | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.078 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Diagnostics Check  | OK    | 20:32 |

---

Data Download Date/Time: 04/26/2024 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: HAUTRIVE,SPENCER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LOPEZ-CHINCHILLA

DATE/TIME OF TEST: 04/05/2024 02:17:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.079 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 1 | 0.000 | 02:13 |
| Breath Volume      | 1.675 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 2 | 0.000 | 02:16 |
| Breath Volume      | 1.457 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.079 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Diagnostics Check  | OK    | 02:17 |

---

Data Download Date/Time: 04/26/2024 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MAURE CASCARET,RENATO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ECHEVARRIA CRUZ

DATE/TIME OF TEST: 04/08/2024 01:11:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.079 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Subject Sample # 1 | 0.000 | 01:07 |
| Breath Volume      | 1.835 | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 2 | 0.000 | 01:10 |
| Breath Volume      | 1.609 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.079 | 01:11 |
| Air Blank          | 0.000 | 01:11 |
| Diagnostics Check  | OK    | 01:11 |

---

Data Download Date/Time: 04/26/2024 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,BRANDON,L  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MILLER,BRANDON  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 04/14/2024 04:04:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.080 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Subject Sample # 1 | 0.255 | 04:00 |
| Breath Volume      | 1.605 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Subject Sample # 2 | 0.263 | 04:03 |
| Breath Volume      | 1.300 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Control Test       | 0.079 | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Diagnostics Check  | OK    | 04:04 |

---

Data Download Date/Time: 04/26/2024 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CREWS,CURTIS,L  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MAURE CASCARET,RENATO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 04/22/2024 00:03:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:00 |                      |
| Air Blank         | 0.000 | 00:00 |                      |
| Control Test      | 0.079 | 00:00 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Subject Sample    | # 1   | 00:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:02 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Control Test      | 0.079 | 00:03 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Diagnostics Check | OK    | 00:03 |                      |

---

Data Download Date/Time: 04/26/2024 13:55

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,DAVID,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: DIAZ,DAVID  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/19/2024 21:51:04

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 21:49 |                           |
| Air Blank         | 0.000 | 21:50 |                           |
| Control Test      | 0.073 | 21:50 | Control Outside Tolerance |
| Air Blank         | 0.000 | 21:51 |                           |

---

Data Download Date/Time: 05/14/2024 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIAZ,DAVID,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: DIAZ,DAVID  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 04/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402370890  
DRY GAS EXPIRY DATE: 05/09/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/19/2024 21:55:04

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 21:53 |                           |
| Air Blank         | 0.000 | 21:54 |                           |
| Control Test      | 0.073 | 21:54 | Control Outside Tolerance |
| Air Blank         | 0.000 | 21:55 |                           |

---

Data Download Date/Time: 05/14/2024 11:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUAREZ,BILL,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: LAUREANO,LUZ  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: SAMPLE

DATE/TIME OF TEST: 04/06/2024 22:37:16

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 22:30 |                           |
| Air Blank         | 0.000     | 22:30 |                           |
| Control Test      | 0.078     | 22:31 |                           |
| Air Blank         | 0.000     | 22:31 |                           |
| Subject Sample    | # 1 0.225 | 22:32 |                           |
| Breath Volume     | 2.812     | 22:32 |                           |
| Air Blank         | 0.000     | 22:33 |                           |
| Air Blank         | 0.000     | 22:35 |                           |
| Subject Sample    | # 2 0.231 | 22:35 |                           |
| Breath Volume     | 2.652     | 22:35 |                           |
| Air Blank         | 0.000     | 22:36 |                           |
| Control Test      | 0.074     | 22:36 | Control Outside Tolerance |
| Air Blank         | 0.000     | 22:37 |                           |

---

Data Download Date/Time: 04/29/2024 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,KENDRICK,J  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: SAINVIL,CAMILLE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1141545  
DRY GAS EXPIRY DATE: 04/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 04/07/2024 09:29:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:25 |                      |
| Air Blank         | 0.000 | 09:26 |                      |
| Control Test      | 0.078 | 09:26 |                      |
| Air Blank         | 0.000 | 09:27 |                      |
| Subject Sample    | # 1   | 09:28 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:28 |                      |
| Air Blank         | 0.000 | 09:28 |                      |
| Control Test      | 0.078 | 09:28 |                      |
| Air Blank         | 0.000 | 09:29 |                      |
| Diagnostics Check | OK    | 09:29 |                      |

---

Data Download Date/Time: 04/29/2024 12:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: BRUTUS,JEFF  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403005538  
DRY GAS EXPIRY DATE: 03/25/2027  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: BRAVE

DATE/TIME OF TEST: 04/13/2024 22:06:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:03 |                      |
| Air Blank         | 0.000 | 22:03 |                      |
| Control Test      | 0.081 | 22:04 |                      |
| Air Blank         | 0.000 | 22:04 |                      |
| Subject Sample    | # 1   | 22:05 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:05 |                      |
| Air Blank         | 0.000 | 22:05 |                      |
| Control Test      | 0.080 | 22:06 |                      |
| Air Blank         | 0.000 | 22:06 |                      |
| Diagnostics Check | OK    | 22:06 |                      |

---

Data Download Date/Time: 04/29/2024 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEAN,ADRIAN,H  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: DEAN,ADRIAN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 403005538  
DRY GAS EXPIRY DATE: 03/25/2027  
VIOLATION CODE: DUI  
ARREST TIME: 08:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: SAINTSUME

DATE/TIME OF TEST: 04/14/2024 09:00:16

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:55 |                      |
| Air Blank         | 0.000 | 08:55 |                      |
| Control Test      | 0.080 | 08:55 |                      |
| Air Blank         | 0.000 | 08:56 |                      |
| Subject Sample    | # 1   | 08:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:58 |                      |
| Air Blank         | 0.000 | 08:59 |                      |
| Control Test      | 0.079 | 08:59 |                      |
| Air Blank         | 0.000 | 09:00 |                      |
| Diagnostics Check | OK    | 09:00 |                      |

Data Download Date/Time: 04/29/2024 12:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 03/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403005538  
DRY GAS EXPIRY DATE: 03/25/2027  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SARRIA

DATE/TIME OF TEST: 04/24/2024 03:56:00

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:44 |                      |
| Air Blank         | 0.000 | 03:45 |                      |
| Control Test      | 0.081 | 03:45 |                      |
| Air Blank         | 0.000 | 03:45 |                      |
| Subject Sample    | # 1   | 03:49 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:49 |                      |
| Air Blank         | 0.000 | 03:49 |                      |
| Air Blank         | 0.000 | 03:51 |                      |
| Subject Sample    | # 2   | 03:54 | Subject Test Refused |
| Breath Volume     | 0.742 | 03:54 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Control Test      | 0.081 | 03:55 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Diagnostics Check | OK    | 03:56 |                      |

Data Download Date/Time: 04/29/2024 12:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, GERALD,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRISON, GERALD  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A2  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/06/2024 00:05:22

---

|                    |       |       |                |
|--------------------|-------|-------|----------------|
| Diagnostics Check  | OK    | 23:55 |                |
| Air Blank          | 0.000 | 23:56 |                |
| Control Test       | 0.081 | 23:56 |                |
| Air Blank          | 0.000 | 23:57 |                |
| Subject Sample # 2 | 0.261 | 23:58 |                |
| Breath Volume      | 1.335 | 23:58 |                |
| Air Blank          | 0.000 | 23:59 |                |
| Air Blank          | 0.000 | 00:00 |                |
| Subject Sample # 1 | 0.234 | 00:04 | Volume Not Met |
| Breath Volume      | 0.171 | 00:04 |                |
| Air Blank          |       | 00:04 | Purge Fail     |
| Air Blank          | 0.000 | 00:05 |                |

---

Data Download Date/Time: 04/30/2024 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,GERALD,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRISON,GERALD  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A2  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/06/2024 00:08:13

---

|                   |       |       |                       |
|-------------------|-------|-------|-----------------------|
| Diagnostics Check | Fail  | 00:07 | Internal Printer Test |
| Air Blank         | 0.000 | 00:08 |                       |

---

Data Download Date/Time: 04/30/2024 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON,GERALD,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRISON,GERALD  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A2  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/06/2024 00:10:42

---

|                   |       |       |                       |
|-------------------|-------|-------|-----------------------|
| Diagnostics Check | Fail  | 00:10 | Internal Printer Test |
| Air Blank         | 0.000 | 00:10 |                       |

---

Data Download Date/Time: 04/30/2024 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRISON, GERALD,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: HARRISON, GERALD  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A2  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 04/06/2024 00:13:52

---

|                   |       |       |                       |
|-------------------|-------|-------|-----------------------|
| Diagnostics Check | Fail  | 00:13 | Internal Printer Test |
| Air Blank         | 0.000 | 00:13 |                       |

---

Data Download Date/Time: 04/30/2024 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARROQUIN,RICHARD,J  
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)  
ARRESTING OFFICER: ALBRITTON,ROY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG310302  
DRY GAS EXPIRY DATE: 04/13/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/26/2024 19:08:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:05 |                      |
| Air Blank         | 0.000 | 19:06 |                      |
| Control Test      | 0.079 | 19:06 |                      |
| Air Blank         | 0.000 | 19:07 |                      |
| Subject Sample    | # 1   | 19:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:07 |                      |
| Air Blank         | 0.000 | 19:07 |                      |
| Control Test      | 0.078 | 19:08 |                      |
| Air Blank         | 0.000 | 19:08 |                      |
| Diagnostics Check | OK    | 19:08 |                      |

---

Data Download Date/Time: 04/27/2024 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: STEFANICK

DATE/TIME OF TEST: 04/10/2024 22:56:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Control Test       | 0.082 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Subject Sample # 1 | 0.229 | 22:51 |
| Breath Volume      | 1.394 | 22:51 |
| Air Blank          | 0.000 | 22:52 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 2 | 0.228 | 22:55 |
| Breath Volume      | 1.644 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.080 | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Diagnostics Check  | OK    | 22:56 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MAZO

DATE/TIME OF TEST: 04/11/2024 02:33:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.082 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 1 | 0.052 | 02:28 |
| Breath Volume      | 1.710 | 02:28 |
| Air Blank          | 0.000 | 02:29 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 2 | 0.053 | 02:32 |
| Breath Volume      | 1.832 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.082 | 02:33 |
| Air Blank          | 0.000 | 02:33 |
| Diagnostics Check  | OK    | 02:33 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FLINT,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HENSEL

DATE/TIME OF TEST: 04/12/2024 02:47:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.081 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.166 | 02:43 |
| Breath Volume      | 1.148 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.179 | 02:46 |
| Breath Volume      | 1.496 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.081 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMOND,JENNY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KUS

DATE/TIME OF TEST: 04/12/2024 03:25:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.081 | 03:20 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 1 | 0.040 | 03:21 |
| Breath Volume      | 1.976 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 2 | 0.043 | 03:24 |
| Breath Volume      | 2.164 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.081 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Diagnostics Check  | OK    | 03:25 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRUEH,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/12/2024 04:13:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.081 | 04:07 |
| Air Blank          | 0.000 | 04:07 |
| Subject Sample # 1 | 0.198 | 04:09 |
| Breath Volume      | 2.406 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 2 | 0.196 | 04:12 |
| Breath Volume      | 1.320 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.080 | 04:13 |
| Air Blank          | 0.000 | 04:13 |
| Diagnostics Check  | OK    | 04:13 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMOND,JENNY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KOPPE,K  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 18:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 04/12/2024 19:44:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:38 |
| Air Blank          | 0.000 | 19:38 |
| Control Test       | 0.082 | 19:39 |
| Air Blank          | 0.000 | 19:39 |
| Subject Sample # 1 | 0.000 | 19:40 |
| Breath Volume      | 1.835 | 19:40 |
| Air Blank          | 0.000 | 19:40 |
| Air Blank          | 0.000 | 19:42 |
| Subject Sample # 2 | 0.000 | 19:43 |
| Breath Volume      | 1.625 | 19:43 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.082 | 19:44 |
| Air Blank          | 0.000 | 19:44 |
| Diagnostics Check  | OK    | 19:44 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LAMBETA, J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 04/12/2024 22:59:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:50 |
| Air Blank          | 0.000 | 22:50 |
| Control Test       | 0.083 | 22:50 |
| Air Blank          | 0.000 | 22:51 |
| Subject Sample # 1 | 0.138 | 22:53 |
| Breath Volume      | 1.429 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 2 | 0.135 | 22:57 |
| Breath Volume      | 1.968 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Control Test       | 0.082 | 22:58 |
| Air Blank          | 0.000 | 22:58 |
| Diagnostics Check  | OK    | 22:59 |

---

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: YAKUBOV

DATE/TIME OF TEST: 04/13/2024 02:56:58

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Control Test      | 0.082 | 02:49 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Subject Sample    | # 1   | 02:52 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:52 |                      |
| Air Blank         | 0.000 | 02:53 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Subject Sample    | # 2   | 02:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:55 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Control Test      | 0.082 | 02:56 |                      |
| Air Blank         | 0.000 | 02:56 |                      |
| Diagnostics Check | OK    | 02:56 |                      |

Data Download Date/Time: 05/08/2024 15:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PEPEN, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/13/2024 04:00:44

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:48 |                |
| Air Blank         | 0.000     | 03:49 |                |
| Control Test      | 0.082     | 03:49 |                |
| Air Blank         | 0.000     | 03:50 |                |
| Subject Sample    | # 1 0.116 | 03:53 | Volume Not Met |
| Breath Volume     | 0.753     | 03:53 |                |
| Air Blank         | 0.000     | 03:54 |                |
| Air Blank         | 0.000     | 03:55 |                |
| Subject Sample    | # 2 0.119 | 03:59 | Volume Not Met |
| Breath Volume     | 0.679     | 03:59 |                |
| Air Blank         | 0.000     | 03:59 |                |
| Control Test      | 0.080     | 04:00 |                |
| Air Blank         | 0.000     | 04:00 |                |
| Diagnostics Check | OK        | 04:00 |                |

Data Download Date/Time: 05/08/2024 15:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PEPEN, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/13/2024 04:17:03

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 04:01 |                |
| Air Blank         | 0.000     | 04:02 |                |
| Control Test      | 0.080     | 04:02 |                |
| Air Blank         | 0.000     | 04:03 |                |
| Subject Sample    | # 1 0.121 | 04:06 | Volume Not Met |
| Breath Volume     | 0.425     | 04:06 |                |
| Air Blank         | 0.000     | 04:07 |                |
| Air Blank         | 0.000     | 04:08 |                |
| Subject Sample    | # 2 0.131 | 04:10 |                |
| Breath Volume     | 1.156     | 04:10 |                |
| Air Blank         | 0.000     | 04:10 |                |
| Air Blank         | 0.000     | 04:12 |                |
| Subject Sample    | # 3 0.127 | 04:15 |                |
| Breath Volume     | 1.335     | 04:15 |                |
| Air Blank         | 0.000     | 04:16 |                |
| Control Test      | 0.080     | 04:16 |                |
| Air Blank         | 0.000     | 04:16 |                |
| Diagnostics Check | OK        | 04:17 |                |

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SARRASIN,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SIMPSON III

DATE/TIME OF TEST: 04/13/2024 05:12:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:08 |                      |
| Air Blank         | 0.000 | 05:08 |                      |
| Control Test      | 0.081 | 05:08 |                      |
| Air Blank         | 0.000 | 05:09 |                      |
| Subject Sample    | # 1   | 05:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:10 |                      |
| Air Blank         | 0.000 | 05:11 |                      |
| Control Test      | 0.081 | 05:11 |                      |
| Air Blank         | 0.000 | 05:12 |                      |
| Diagnostics Check | OK    | 05:12 |                      |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE, B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 04:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: EVONOSKY

DATE/TIME OF TEST: 04/13/2024 05:37:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:31 |
| Air Blank          | 0.000 | 05:31 |
| Control Test       | 0.081 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Subject Sample # 1 | 0.111 | 05:33 |
| Breath Volume      | 2.769 | 05:33 |
| Air Blank          | 0.000 | 05:33 |
| Air Blank          | 0.000 | 05:35 |
| Subject Sample # 2 | 0.117 | 05:36 |
| Breath Volume      | 2.703 | 05:36 |
| Air Blank          | 0.000 | 05:36 |
| Control Test       | 0.081 | 05:37 |
| Air Blank          | 0.000 | 05:37 |
| Diagnostics Check  | OK    | 05:37 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 04/14/2024 01:51:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:43 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.082 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 1 | 0.184 | 01:46 |
| Breath Volume      | 2.714 | 01:46 |
| Air Blank          | 0.000 | 01:47 |
| Air Blank          | 0.000 | 01:49 |
| Subject Sample # 2 | 0.184 | 01:49 |
| Breath Volume      | 2.449 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Control Test       | 0.082 | 01:50 |
| Air Blank          | 0.000 | 01:51 |
| Diagnostics Check  | OK    | 01:51 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: NYLAND

DATE/TIME OF TEST: 04/14/2024 03:09:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.082 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 1 | 0.125 | 03:04 |
| Breath Volume      | 4.683 | 03:04 |
| Air Blank          | 0.000 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 2 | 0.123 | 03:07 |
| Breath Volume      | 4.832 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Control Test       | 0.082 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Diagnostics Check  | OK    | 03:09 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SARRASIN, R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: COLMENAREZ MORILLO

DATE/TIME OF TEST: 04/14/2024 04:57:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.082 | 04:52 |
| Air Blank          | 0.000 | 04:52 |
| Subject Sample # 1 | 0.110 | 04:53 |
| Breath Volume      | 2.675 | 04:53 |
| Air Blank          | 0.000 | 04:53 |
| Air Blank          | 0.000 | 04:55 |
| Subject Sample # 2 | 0.115 | 04:56 |
| Breath Volume      | 2.125 | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Control Test       | 0.082 | 04:57 |
| Air Blank          | 0.000 | 04:57 |
| Diagnostics Check  | OK    | 04:57 |

---

Data Download Date/Time: 05/08/2024 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAAR, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 04/15/2024 02:09:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.082 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 1 | 0.195 | 02:04 |
| Breath Volume      | 1.929 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 2 | 0.191 | 02:07 |
| Breath Volume      | 1.777 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.082 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Diagnostics Check  | OK    | 02:09 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GARCIA,J  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 04/15/2024 23:40:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Control Test       | 0.082 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Subject Sample # 1 | 0.000 | 23:36 |
| Breath Volume      | 1.925 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 2 | 0.000 | 23:39 |
| Breath Volume      | 2.007 | 23:39 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.082 | 23:40 |
| Air Blank          | 0.000 | 23:40 |
| Diagnostics Check  | OK    | 23:40 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: FERNANDES

DATE/TIME OF TEST: 04/16/2024 01:50:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Control Test       | 0.083 | 01:44 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 1 | 0.171 | 01:45 |
| Breath Volume      | 1.429 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 2 | 0.169 | 01:48 |
| Breath Volume      | 1.824 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Control Test       | 0.082 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Diagnostics Check  | OK    | 01:50 |

---

Data Download Date/Time: 05/08/2024 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ABREU III

DATE/TIME OF TEST: 04/17/2024 00:25:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Control Test       | 0.082 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 1 | 0.008 | 00:20 |
| Breath Volume      | 2.031 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 2 | 0.008 | 00:23 |
| Breath Volume      | 1.324 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Control Test       | 0.082 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:25 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: URCADEZ

DATE/TIME OF TEST: 04/17/2024 03:32:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.083 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.120 | 03:28 |
| Breath Volume      | 2.710 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.118 | 03:31 |
| Breath Volume      | 2.652 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.082 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: BALE-ADAMS

DATE/TIME OF TEST: 04/17/2024 03:59:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.083 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Subject Sample # 1 | 0.177 | 03:54 |
| Breath Volume      | 3.113 | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 2 | 0.172 | 03:57 |
| Breath Volume      | 2.593 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.081 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KISS,L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: GREGORIO

DATE/TIME OF TEST: 04/17/2024 04:16:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.081 | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Subject Sample # 1 | 0.122 | 04:12 |
| Breath Volume      | 2.554 | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 2 | 0.119 | 04:14 |
| Breath Volume      | 3.050 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.081 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LIGGINS, A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FERENCZY-ZUMPANO

DATE/TIME OF TEST: 04/19/2024 02:50:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:42 |
| Air Blank          | 0.000 | 02:42 |
| Control Test       | 0.082 | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Subject Sample # 1 | 0.000 | 02:45 |
| Breath Volume      | 1.496 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Air Blank          | 0.000 | 02:48 |
| Subject Sample # 2 | 0.000 | 02:48 |
| Breath Volume      | 1.660 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Control Test       | 0.081 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Diagnostics Check  | OK    | 02:50 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 04/19/2024 23:13:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:09 |                      |
| Air Blank         | 0.000 | 23:09 |                      |
| Control Test      | 0.082 | 23:10 |                      |
| Air Blank         | 0.000 | 23:10 |                      |
| Subject Sample    | # 1   | 23:11 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:11 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Control Test      | 0.081 | 23:12 |                      |
| Air Blank         | 0.000 | 23:13 |                      |
| Diagnostics Check | OK    | 23:13 |                      |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ROSALES, F  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEREZ CITALAN

DATE/TIME OF TEST: 04/20/2024 02:25:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.083 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Subject Sample # 1 | 0.145 | 02:20 |
| Breath Volume      | 1.511 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 2 | 0.145 | 02:23 |
| Breath Volume      | 2.417 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Control Test       | 0.081 | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Diagnostics Check  | OK    | 02:25 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TRUJILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: OPONT

DATE/TIME OF TEST: 04/20/2024 03:52:06

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:40 |                |
| Air Blank         | 0.000     | 03:40 |                |
| Control Test      | 0.082     | 03:41 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Subject Sample    | # 1 0.000 | 03:44 | Volume Not Met |
| Breath Volume     | 0.066     | 03:44 |                |
| Air Blank         | 0.000     | 03:45 |                |
| Air Blank         | 0.000     | 03:47 |                |
| Subject Sample    | # 2 0.312 | 03:50 | Volume Not Met |
| Breath Volume     | 0.089     | 03:50 |                |
| Air Blank         | 0.000     | 03:51 |                |
| Control Test      | 0.079     | 03:51 |                |
| Air Blank         | 0.000     | 03:52 |                |
| Diagnostics Check | OK        | 03:52 |                |

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KISS, L  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: QUIGLEY

DATE/TIME OF TEST: 04/20/2024 05:01:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:54 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.081 | 04:55 |
| Air Blank          | 0.000 | 04:55 |
| Subject Sample # 1 | 0.136 | 04:56 |
| Breath Volume      | 1.730 | 04:56 |
| Air Blank          | 0.000 | 04:57 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 2 | 0.133 | 04:59 |
| Breath Volume      | 1.222 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Control Test       | 0.081 | 05:00 |
| Air Blank          | 0.000 | 05:00 |
| Diagnostics Check  | OK    | 05:01 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GARCIA ABREU

DATE/TIME OF TEST: 04/21/2024 00:24:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:17 |
| Air Blank          | 0.000 | 00:17 |
| Control Test       | 0.082 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Subject Sample # 1 | 0.127 | 00:20 |
| Breath Volume      | 1.445 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Air Blank          | 0.000 | 00:22 |
| Subject Sample # 2 | 0.138 | 00:23 |
| Breath Volume      | 1.574 | 00:23 |
| Air Blank          | 0.000 | 00:23 |
| Control Test       | 0.082 | 00:24 |
| Air Blank          | 0.000 | 00:24 |
| Diagnostics Check  | OK    | 00:24 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: DIAZ GARCIA

DATE/TIME OF TEST: 04/21/2024 01:16:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Control Test       | 0.082 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 1 | 0.173 | 01:11 |
| Breath Volume      | 2.304 | 01:11 |
| Air Blank          | 0.000 | 01:12 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 2 | 0.175 | 01:15 |
| Breath Volume      | 2.066 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Control Test       | 0.081 | 01:16 |
| Air Blank          | 0.000 | 01:16 |
| Diagnostics Check  | OK    | 01:16 |

---

Data Download Date/Time: 05/08/2024 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAILEY,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 01:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 04/21/2024 02:43:54

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 02:31 |                    |
| Air Blank         | 0.000     | 02:31 |                    |
| Control Test      | 0.082     | 02:31 |                    |
| Air Blank         | 0.000     | 02:32 |                    |
| Subject Sample    | # 1 0.187 | 02:33 |                    |
| Breath Volume     | 2.070     | 02:33 |                    |
| Air Blank         | 0.000     | 02:34 |                    |
| Air Blank         | 0.000     | 02:35 |                    |
| Subject Sample    | # 2       | 02:39 | No Sample Provided |
| Breath Volume     | 0.000     | 02:39 |                    |
| Air Blank         | 0.000     | 02:39 |                    |
| Air Blank         | 0.000     | 02:41 |                    |
| Subject Sample    | # 3 0.192 | 02:42 |                    |
| Breath Volume     | 1.386     | 02:42 |                    |
| Air Blank         | 0.000     | 02:42 |                    |
| Control Test      | 0.081     | 02:43 |                    |
| Air Blank         | 0.000     | 02:43 |                    |
| Diagnostics Check | OK        | 02:43 |                    |

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DE LA CRUZ MORALES

DATE/TIME OF TEST: 04/21/2024 04:02:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.082 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 1 | 0.086 | 03:57 |
| Breath Volume      | 1.707 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 2 | 0.079 | 04:00 |
| Breath Volume      | 1.621 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.081 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Diagnostics Check  | OK    | 04:02 |

---

Data Download Date/Time: 05/08/2024 15:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MOREJON,A  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/21/2024 04:31:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.082 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.080 | 04:26 |
| Breath Volume      | 1.671 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.079 | 04:30 |
| Breath Volume      | 1.738 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.081 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMOND,JENNY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KOLANOWSKI,D  
ARRESTING OFFICER AGENCY: Plant City PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 04/21/2024 18:47:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:40 |
| Air Blank          | 0.000 | 18:41 |
| Control Test       | 0.081 | 18:41 |
| Air Blank          | 0.000 | 18:42 |
| Subject Sample # 1 | 0.212 | 18:42 |
| Breath Volume      | 1.644 | 18:42 |
| Air Blank          | 0.000 | 18:43 |
| Air Blank          | 0.000 | 18:45 |
| Subject Sample # 2 | 0.213 | 18:45 |
| Breath Volume      | 1.539 | 18:45 |
| Air Blank          | 0.000 | 18:46 |
| Control Test       | 0.081 | 18:46 |
| Air Blank          | 0.000 | 18:47 |
| Diagnostics Check  | OK    | 18:47 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMOND,JENNY,J  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HILL,B  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 08:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: DEVINNEY

DATE/TIME OF TEST: 04/22/2024 09:41:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:35 |
| Air Blank          | 0.000 | 09:35 |
| Control Test       | 0.082 | 09:35 |
| Air Blank          | 0.000 | 09:36 |
| Subject Sample # 1 | 0.048 | 09:37 |
| Breath Volume      | 1.687 | 09:37 |
| Air Blank          | 0.000 | 09:37 |
| Air Blank          | 0.000 | 09:39 |
| Subject Sample # 2 | 0.048 | 09:40 |
| Breath Volume      | 1.804 | 09:40 |
| Air Blank          | 0.000 | 09:40 |
| Control Test       | 0.081 | 09:41 |
| Air Blank          | 0.000 | 09:41 |
| Diagnostics Check  | OK    | 09:41 |

---

Data Download Date/Time: 05/08/2024 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON, D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 04/24/2024 03:38:25

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:23 |                |
| Air Blank         | 0.000     | 03:24 |                |
| Control Test      | 0.083     | 03:24 |                |
| Air Blank         | 0.000     | 03:24 |                |
| Subject Sample    | # 1 0.154 | 03:28 | Volume Not Met |
| Breath Volume     | 0.246     | 03:28 |                |
| Air Blank         | 0.000     | 03:28 |                |
| Air Blank         | 0.000     | 03:30 |                |
| Subject Sample    | # 2 0.163 | 03:31 |                |
| Breath Volume     | 1.164     | 03:31 |                |
| Air Blank         | 0.000     | 03:32 |                |
| Air Blank         | 0.000     | 03:34 |                |
| Subject Sample    | # 3 0.167 | 03:36 |                |
| Breath Volume     | 1.363     | 03:36 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Control Test      | 0.082     | 03:37 |                |
| Air Blank         | 0.000     | 03:38 |                |
| Diagnostics Check | OK        | 03:38 |                |

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TRUJILLO, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: PHILOGENE

DATE/TIME OF TEST: 04/25/2024 04:02:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:54 |
| Air Blank          | 0.000 | 03:55 |
| Control Test       | 0.082 | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Subject Sample # 1 | 0.150 | 03:57 |
| Breath Volume      | 1.722 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.146 | 04:00 |
| Breath Volume      | 1.554 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.081 | 04:01 |
| Air Blank          | 0.000 | 04:02 |
| Diagnostics Check  | OK    | 04:02 |

---

Data Download Date/Time: 05/08/2024 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: SNAGGS,S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BAKIS

DATE/TIME OF TEST: 04/25/2024 23:42:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:35 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.082 | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Subject Sample # 1 | 0.217 | 23:37 |
| Breath Volume      | 1.332 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 2 | 0.209 | 23:40 |
| Breath Volume      | 1.382 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.082 | 23:41 |
| Air Blank          | 0.000 | 23:42 |
| Diagnostics Check  | OK    | 23:42 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MALAZARTE

DATE/TIME OF TEST: 04/26/2024 04:32:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Control Test       | 0.083 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Subject Sample # 1 | 0.113 | 04:27 |
| Breath Volume      | 1.386 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.120 | 04:30 |
| Breath Volume      | 2.097 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Control Test       | 0.082 | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Diagnostics Check  | OK    | 04:32 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TELLEZ-ALFARO,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: RIVERA GARCIA

DATE/TIME OF TEST: 04/26/2024 05:56:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:50 |
| Air Blank          | 0.000 | 05:50 |
| Control Test       | 0.082 | 05:51 |
| Air Blank          | 0.000 | 05:51 |
| Subject Sample # 1 | 0.000 | 05:52 |
| Breath Volume      | 3.792 | 05:52 |
| Air Blank          | 0.000 | 05:52 |
| Air Blank          | 0.000 | 05:54 |
| Subject Sample # 2 | 0.000 | 05:55 |
| Breath Volume      | 3.285 | 05:55 |
| Air Blank          | 0.000 | 05:55 |
| Control Test       | 0.081 | 05:55 |
| Air Blank          | 0.000 | 05:56 |
| Diagnostics Check  | OK    | 05:56 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,JESSICA,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 04/27/2024 02:39:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.085 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 1 | 0.082 | 02:34 |
| Breath Volume      | 1.832 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.083 | 02:37 |
| Breath Volume      | 1.914 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.082 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:39 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JENKINS,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GARCIA PESCADOR

DATE/TIME OF TEST: 04/27/2024 03:58:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.082 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.226 | 03:54 |
| Breath Volume      | 1.886 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.231 | 03:57 |
| Breath Volume      | 1.339 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.081 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SEMA JURE

DATE/TIME OF TEST: 04/27/2024 04:19:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Control Test       | 0.082 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Subject Sample # 1 | 0.000 | 04:15 |
| Breath Volume      | 4.011 | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 2 | 0.000 | 04:17 |
| Breath Volume      | 2.625 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.081 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Diagnostics Check  | OK    | 04:19 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KROME,BRITTNEY,M  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LIGGINS,A  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/27/2024 04:36:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Control Test       | 0.082 | 04:31 |
| Air Blank          | 0.000 | 04:31 |
| Subject Sample # 1 | 0.197 | 04:32 |
| Breath Volume      | 1.855 | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 2 | 0.194 | 04:35 |
| Breath Volume      | 1.789 | 04:35 |
| Air Blank          | 0.000 | 04:35 |
| Control Test       | 0.082 | 04:36 |
| Air Blank          | 0.000 | 04:36 |
| Diagnostics Check  | OK    | 04:36 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BAAR,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: STAUFER

DATE/TIME OF TEST: 04/27/2024 21:05:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:59 |
| Air Blank          | 0.000 | 20:59 |
| Control Test       | 0.082 | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Subject Sample # 1 | 0.089 | 21:01 |
| Breath Volume      | 2.195 | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Air Blank          | 0.000 | 21:03 |
| Subject Sample # 2 | 0.093 | 21:04 |
| Breath Volume      | 1.289 | 21:04 |
| Air Blank          | 0.000 | 21:04 |
| Control Test       | 0.082 | 21:05 |
| Air Blank          | 0.000 | 21:05 |
| Diagnostics Check  | OK    | 21:05 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BADEN, T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CARITHERS

DATE/TIME OF TEST: 04/28/2024 03:36:15

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:33 |                      |
| Air Blank         | 0.000 | 03:33 |                      |
| Control Test      | 0.082 | 03:33 |                      |
| Air Blank         | 0.000 | 03:34 |                      |
| Subject Sample    | # 1   | 03:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:34 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Control Test      | 0.082 | 03:35 |                      |
| Air Blank         | 0.000 | 03:36 |                      |
| Diagnostics Check | OK    | 03:36 |                      |

Data Download Date/Time: 05/08/2024 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Macdill Air Force Ba  
ARRESTING OFFICER: ULAKOVIC,Z  
ARRESTING OFFICER AGENCY: Macdill Air Force Ba  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MOLINE

DATE/TIME OF TEST: 04/28/2024 04:08:26

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:05 |                      |
| Air Blank         | 0.000 | 04:05 |                      |
| Control Test      | 0.082 | 04:05 |                      |
| Air Blank         | 0.000 | 04:06 |                      |
| Subject Sample    | # 1   | 04:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:07 |                      |
| Air Blank         | 0.000 | 04:07 |                      |
| Control Test      | 0.082 | 04:07 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Diagnostics Check | OK    | 04:08 |                      |

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: EARLYWINE,G  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 05:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 04/28/2024 06:01:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:54 |
| Air Blank          | 0.000 | 05:54 |
| Control Test       | 0.082 | 05:55 |
| Air Blank          | 0.000 | 05:55 |
| Subject Sample # 1 | 0.112 | 05:56 |
| Breath Volume      | 2.464 | 05:56 |
| Air Blank          | 0.000 | 05:57 |
| Air Blank          | 0.000 | 05:59 |
| Subject Sample # 2 | 0.112 | 05:59 |
| Breath Volume      | 2.269 | 05:59 |
| Air Blank          | 0.000 | 06:00 |
| Control Test       | 0.082 | 06:00 |
| Air Blank          | 0.000 | 06:01 |
| Diagnostics Check  | OK    | 06:01 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HENLEY,SAMANTHA  
ARRESTING OFFICER AGENCY: Tampa Airport PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: PINTO

DATE/TIME OF TEST: 04/28/2024 15:56:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:50 |
| Air Blank          | 0.000 | 15:50 |
| Control Test       | 0.082 | 15:50 |
| Air Blank          | 0.000 | 15:51 |
| Subject Sample # 1 | 0.191 | 15:51 |
| Breath Volume      | 1.148 | 15:51 |
| Air Blank          | 0.000 | 15:52 |
| Air Blank          | 0.000 | 15:54 |
| Subject Sample # 2 | 0.190 | 15:54 |
| Breath Volume      | 1.902 | 15:54 |
| Air Blank          | 0.000 | 15:55 |
| Control Test       | 0.081 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Diagnostics Check  | OK    | 15:56 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BURKE,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MOYAO SALAS

DATE/TIME OF TEST: 04/29/2024 03:30:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.082 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.078 | 03:25 |
| Breath Volume      | 1.953 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 2 | 0.080 | 03:28 |
| Breath Volume      | 1.953 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.082 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Diagnostics Check  | OK    | 03:30 |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER,MELONIE,E  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRUEH,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 06:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:46

SUBJECT'S LAST NAME: VANTIL

DATE/TIME OF TEST: 04/29/2024 07:13:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:07 |
| Air Blank          | 0.000 | 07:08 |
| Control Test       | 0.082 | 07:08 |
| Air Blank          | 0.000 | 07:08 |
| Subject Sample # 1 | 0.174 | 07:09 |
| Breath Volume      | 1.769 | 07:09 |
| Air Blank          | 0.000 | 07:10 |
| Air Blank          | 0.000 | 07:11 |
| Subject Sample # 2 | 0.176 | 07:12 |
| Breath Volume      | 1.960 | 07:12 |
| Air Blank          | 0.000 | 07:13 |
| Control Test       | 0.082 | 07:13 |
| Air Blank          | 0.000 | 07:13 |
| Diagnostics Check  | OK    | 07:13 |

---

Data Download Date/Time: 05/08/2024 15:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KRUG, Z  
ARRESTING OFFICER AGENCY: Temple Terrace PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 04/30/2024 03:25:54

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:10 |                |
| Air Blank         | 0.000     | 03:11 |                |
| Control Test      | 0.082     | 03:11 |                |
| Air Blank         | 0.000     | 03:12 |                |
| Subject Sample    | # 1 0.124 | 03:15 | Volume Not Met |
| Breath Volume     | 0.050     | 03:15 |                |
| Air Blank         | 0.000     | 03:15 |                |
| Air Blank         | 0.000     | 03:17 |                |
| Subject Sample    | # 2 0.126 | 03:20 |                |
| Breath Volume     | 1.546     | 03:20 |                |
| Air Blank         | 0.000     | 03:20 |                |
| Air Blank         | 0.000     | 03:22 |                |
| Subject Sample    | # 3 0.121 | 03:24 |                |
| Breath Volume     | 1.187     | 03:24 |                |
| Air Blank         | 0.000     | 03:24 |                |
| Control Test      | 0.080     | 03:25 |                |
| Air Blank         | 0.000     | 03:25 |                |
| Diagnostics Check | OK        | 03:25 |                |

---

Data Download Date/Time: 05/08/2024 15:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ROVIROSA, ARMANDO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 04/14/2024 01:01:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.082 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 1 | 0.000 | 00:57 |
| Breath Volume      | 2.976 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.000 | 01:00 |
| Breath Volume      | 3.765 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.081 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Diagnostics Check  | OK    | 01:01 |

---

Data Download Date/Time: 04/20/2024 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROVIROSA,ARMANDO,R  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ROVIROSA,ARMANDO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 03/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LANTHIER

DATE/TIME OF TEST: 04/20/2024 02:51:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.081 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 1 | 0.154 | 02:42 |
| Breath Volume      | 1.429 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.128 | 02:46 |
| Breath Volume      | 1.597 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 3 | 0.136 | 02:49 |
| Breath Volume      | 1.753 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.078 | 02:50 |
| Air Blank          | 0.000 | 02:51 |
| Diagnostics Check  | OK    | 02:51 |

---

Data Download Date/Time: 04/20/2024 16:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RITTER,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Havana PD  
ARRESTING OFFICER: BLOUNT,CAMERON  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 03/13/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/09/2024 23:45:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.082 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.109 | 23:40 |
| Breath Volume      | 1.738 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 2 | 0.115 | 23:43 |
| Breath Volume      | 1.773 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Control Test       | 0.082 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Diagnostics Check  | OK    | 23:45 |

---

Data Download Date/Time: 04/10/2024 12:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEWITT,JARED,S  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: SHERIDAN,COLTON  
ARRESTING OFFICER AGENCY: Franklin County SO  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: KUNZER

DATE/TIME OF TEST: 04/19/2024 23:49:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Control Test       | 0.081 | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Subject Sample # 1 | 0.217 | 23:44 |
| Breath Volume      | 2.375 | 23:44 |
| Air Blank          | 0.000 | 23:45 |
| Air Blank          | 0.000 | 23:47 |
| Subject Sample # 2 | 0.228 | 23:48 |
| Breath Volume      | 2.277 | 23:48 |
| Air Blank          | 0.000 | 23:48 |
| Control Test       | 0.080 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Diagnostics Check  | OK    | 23:49 |

---

Data Download Date/Time: 05/15/2024 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHIVER,ROBERT,D  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: DAVIS,TIMOTHY  
ARRESTING OFFICER AGENCY: Apalachicola PD  
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/20/2024 22:06:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Control Test       | 0.082 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 1 | 0.208 | 22:01 |
| Breath Volume      | 1.246 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Air Blank          | 0.000 | 22:04 |
| Subject Sample # 2 | 0.221 | 22:04 |
| Breath Volume      | 2.636 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Control Test       | 0.080 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Diagnostics Check  | OK    | 22:06 |

---

Data Download Date/Time: 05/15/2024 11:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLAGA,JASMINE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RICKERSON,KENTON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/06/2024 03:41:16

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:37 |            |
| Air Blank         | 0.000     | 03:38 |            |
| Control Test      | 0.081     | 03:38 |            |
| Air Blank         | 0.000     | 03:39 |            |
| Subject Sample    | # 1 0.171 | 03:39 |            |
| Breath Volume     | 3.621     | 03:39 |            |
| Air Blank         |           | 03:40 | Purge Fail |
| Air Blank         | 0.000     | 03:41 |            |

---

Data Download Date/Time: 04/25/2024 11:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLAGA,JASMINE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RICKERSON,KENTON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/06/2024 03:48:52

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:44 |                    |
| Air Blank         | 0.000 | 03:45 |                    |
| Control Test      | 0.081 | 03:45 |                    |
| Air Blank         | 0.000 | 03:45 |                    |
| Subject Sample    | # 1   | 03:47 | Interferent Detect |
| Breath Volume     | 3.023 | 03:47 |                    |
| Air Blank         |       | 03:48 | Purge Fail         |
| Air Blank         |       | 03:48 | Purge Fail         |

---

Data Download Date/Time: 04/25/2024 11:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLAGA,JASMINE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RICKERSON,KENTON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/06/2024 03:52:51

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:51 |            |
| Air Blank         |       | 03:52 | RFI Detect |
| Air Blank         | 0.000 | 03:52 |            |

---

Data Download Date/Time: 04/25/2024 11:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SLAGA,JASMINE,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RICKERSON,KENTON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/06/2024 04:01:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.081 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 1 | 0.173 | 03:57 |
| Breath Volume      | 1.277 | 03:57 |
| Air Blank          | 0.000 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Subject Sample # 2 | 0.171 | 04:00 |
| Breath Volume      | 2.351 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.083 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Diagnostics Check  | OK    | 04:01 |

---

Data Download Date/Time: 04/25/2024 11:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: BROWN,ZECHARIAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402872586  
DRY GAS EXPIRY DATE: 10/12/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MACKENSON

DATE/TIME OF TEST: 04/07/2024 22:41:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.082 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 1 | 0.049 | 22:36 |
| Breath Volume      | 1.582 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 2 | 0.048 | 22:39 |
| Breath Volume      | 1.847 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.083 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Diagnostics Check  | OK    | 22:41 |

---

Data Download Date/Time: 04/25/2024 11:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: Lakeland PD  
ARRESTING OFFICER: RIGGALL,NICHOLAS  
ARRESTING OFFICER AGENCY: Lakeland PD  
LAST AGENCY INSPECTION: 03/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A3  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HOWEY

DATE/TIME OF TEST: 04/04/2024 22:45:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Control Test       | 0.081 | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.170 | 22:40 |
| Breath Volume      | 1.855 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Air Blank          | 0.000 | 22:42 |
| Subject Sample # 2 | 0.156 | 22:43 |
| Breath Volume      | 1.726 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Control Test       | 0.081 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Diagnostics Check  | OK    | 22:45 |

---

Data Download Date/Time: 04/05/2024 13:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGER, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BURGER, WILLIAM  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231294  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: HOCHSTETLER

DATE/TIME OF TEST: 04/06/2024 04:53:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:46 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.079 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.170 | 04:48 |
| Breath Volume      | 4.535 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Air Blank          | 0.000 | 04:50 |
| Subject Sample # 2 | 0.170 | 04:51 |
| Breath Volume      | 3.617 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Control Test       | 0.077 | 04:52 |
| Air Blank          | 0.000 | 04:53 |
| Diagnostics Check  | OK    | 04:53 |

---

Data Download Date/Time: 04/30/2024 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURGER, WILLIAM, T  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BURGER, WILLIAM  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231294  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/08/2024 01:44:48

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:41 |                      |
| Air Blank         | 0.000 | 01:42 |                      |
| Control Test      | 0.080 | 01:42 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Subject Sample    | # 1   | 01:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:43 |                      |
| Air Blank         | 0.000 | 01:43 |                      |
| Control Test      | 0.080 | 01:44 |                      |
| Air Blank         | 0.000 | 01:44 |                      |
| Diagnostics Check | OK    | 01:44 |                      |

Data Download Date/Time: 04/30/2024 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY, MARCUS, W  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: MONTGOMERY, MARCUS  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231294  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 04/15/2024 17:55:43

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:50 |                      |
| Air Blank         | 0.000 | 17:51 |                      |
| Control Test      | 0.079 | 17:51 |                      |
| Air Blank         | 0.000 | 17:52 |                      |
| Subject Sample    | # 1   | 17:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:54 |                      |
| Air Blank         | 0.000 | 17:54 |                      |
| Control Test      | 0.079 | 17:55 |                      |
| Air Blank         | 0.000 | 17:55 |                      |
| Diagnostics Check | OK    | 17:55 |                      |

---

Data Download Date/Time: 04/30/2024 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIECK,DAVID,J  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: BIECK,DAVID  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231294  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 04/20/2024 20:43:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:37 |
| Air Blank          | 0.000 | 20:38 |
| Control Test       | 0.079 | 20:38 |
| Air Blank          | 0.000 | 20:38 |
| Subject Sample # 1 | 0.179 | 20:39 |
| Breath Volume      | 3.476 | 20:39 |
| Air Blank          | 0.000 | 20:40 |
| Air Blank          | 0.000 | 20:41 |
| Subject Sample # 2 | 0.178 | 20:42 |
| Breath Volume      | 3.519 | 20:42 |
| Air Blank          | 0.000 | 20:42 |
| Control Test       | 0.079 | 20:43 |
| Air Blank          | 0.000 | 20:43 |
| Diagnostics Check  | OK    | 20:43 |

---

Data Download Date/Time: 04/30/2024 12:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS, TYLER, J  
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD  
ARRESTING OFFICER: RIVERA, RITCHIE  
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD  
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 231294  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 04/27/2024 02:35:41

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:32 |                      |
| Air Blank         | 0.000 | 02:33 |                      |
| Control Test      | 0.080 | 02:33 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Subject Sample    | # 1   | 02:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:34 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Control Test      | 0.079 | 02:35 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Diagnostics Check | OK    | 02:35 |                      |

Data Download Date/Time: 04/30/2024 12:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LEWIS,CHRIS,L

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: LEWIS,CHRIS

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: SYS CHECK

ARREST TIME: 14:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 04/01/2024 15:46:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:40 |
| Air Blank          | 0.000 | 15:41 |
| Control Test       | 0.082 | 15:41 |
| Air Blank          | 0.000 | 15:42 |
| Subject Sample # 1 | 0.000 | 15:42 |
| Breath Volume      | 3.562 | 15:42 |
| Air Blank          | 0.000 | 15:42 |
| Air Blank          | 0.000 | 15:44 |
| Subject Sample # 2 | 0.000 | 15:45 |
| Breath Volume      | 3.332 | 15:45 |
| Air Blank          | 0.000 | 15:45 |
| Control Test       | 0.083 | 15:46 |
| Air Blank          | 0.000 | 15:46 |
| Diagnostics Check  | OK    | 15:46 |

---

Data Download Date/Time: 05/01/2024 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Crestview PD  
ARRESTING OFFICER: ENCINIAS,HAILEY  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402590345  
DRY GAS EXPIRY DATE: 11/04/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/14/2024 05:05:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.083 | 05:00 |
| Air Blank          | 0.000 | 05:00 |
| Subject Sample # 1 | 0.200 | 05:01 |
| Breath Volume      | 1.625 | 05:01 |
| Air Blank          | 0.000 | 05:01 |
| Air Blank          | 0.000 | 05:03 |
| Subject Sample # 2 | 0.208 | 05:04 |
| Breath Volume      | 2.800 | 05:04 |
| Air Blank          | 0.000 | 05:04 |
| Control Test       | 0.083 | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Diagnostics Check  | OK    | 05:05 |

---

Data Download Date/Time: 05/01/2024 10:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Crestview PD  
ARRESTING OFFICER: DEAN,TYLER  
ARRESTING OFFICER AGENCY: Crestview PD  
LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402590345  
DRY GAS EXPIRY DATE: 11/04/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: PUGA

DATE/TIME OF TEST: 04/24/2024 23:02:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Control Test       | 0.083 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 1 | 0.061 | 22:57 |
| Breath Volume      | 2.257 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Air Blank          | 0.000 | 23:00 |
| Subject Sample # 2 | 0.065 | 23:00 |
| Breath Volume      | 2.472 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.083 | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Diagnostics Check  | OK    | 23:02 |

---

Data Download Date/Time: 05/01/2024 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: ALVA-ARIAS,K  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BUCARDO GUTIERREZ

DATE/TIME OF TEST: 04/06/2024 01:53:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:46 |
| Air Blank          | 0.000 | 01:46 |
| Control Test       | 0.081 | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Subject Sample # 1 | 0.227 | 01:48 |
| Breath Volume      | 3.371 | 01:48 |
| Air Blank          | 0.000 | 01:49 |
| Air Blank          | 0.000 | 01:50 |
| Subject Sample # 2 | 0.228 | 01:51 |
| Breath Volume      | 2.816 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.081 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Diagnostics Check  | OK    | 01:53 |

---

Data Download Date/Time: 04/10/2024 14:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: FRUEH,D  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: AVILES BATISTA

DATE/TIME OF TEST: 04/06/2024 04:31:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:24 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.081 | 04:25 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 1 | 0.095 | 04:26 |
| Breath Volume      | 1.593 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Air Blank          | 0.000 | 04:29 |
| Subject Sample # 2 | 0.094 | 04:29 |
| Breath Volume      | 2.328 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Control Test       | 0.080 | 04:30 |
| Air Blank          | 0.000 | 04:31 |
| Diagnostics Check  | OK    | 04:31 |

---

Data Download Date/Time: 04/10/2024 14:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: DIBIASE, J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MARTINEZ BAUTISTA

DATE/TIME OF TEST: 04/07/2024 02:10:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:04 |
| Air Blank          | 0.000 | 02:04 |
| Control Test       | 0.081 | 02:05 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 1 | 0.194 | 02:06 |
| Breath Volume      | 1.457 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 2 | 0.190 | 02:09 |
| Breath Volume      | 1.253 | 02:09 |
| Air Blank          | 0.000 | 02:09 |
| Control Test       | 0.081 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Diagnostics Check  | OK    | 02:10 |

---

Data Download Date/Time: 04/10/2024 14:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY,G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ORTEGA DIAZ

DATE/TIME OF TEST: 04/07/2024 04:50:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:44 |
| Air Blank          | 0.000 | 04:44 |
| Control Test       | 0.081 | 04:45 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 1 | 0.166 | 04:46 |
| Breath Volume      | 2.757 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Subject Sample # 2 | 0.174 | 04:49 |
| Breath Volume      | 1.910 | 04:49 |
| Air Blank          | 0.000 | 04:50 |
| Control Test       | 0.081 | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Diagnostics Check  | OK    | 04:50 |

---

Data Download Date/Time: 04/10/2024 14:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: TEMPLE, ELIZABETH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: DITINNO

DATE/TIME OF TEST: 04/04/2024 02:38:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Control Test       | 0.083 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Subject Sample # 1 | 0.000 | 02:34 |
| Breath Volume      | 2.058 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 2 | 0.000 | 02:37 |
| Breath Volume      | 2.230 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.084 | 02:38 |
| Air Blank          | 0.000 | 02:38 |
| Diagnostics Check  | OK    | 02:38 |

---

Data Download Date/Time: 04/23/2024 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEANE,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ZAPATA,F  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/07/2024 01:09:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.084 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 1 | 0.157 | 01:03 |
| Breath Volume      | 1.523 | 01:03 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:05 |
| Subject Sample # 2 | 0.164 | 01:07 |
| Breath Volume      | 2.351 | 01:07 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.083 | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Diagnostics Check  | OK    | 01:09 |

---

Data Download Date/Time: 04/23/2024 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILES,CHRISTOPHER,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: PLUMB,JASON  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 17:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 04/08/2024 18:54:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:48 |
| Air Blank          | 0.000 | 18:49 |
| Control Test       | 0.083 | 18:49 |
| Air Blank          | 0.000 | 18:49 |
| Subject Sample # 1 | 0.160 | 18:50 |
| Breath Volume      | 2.761 | 18:50 |
| Air Blank          | 0.000 | 18:51 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 2 | 0.162 | 18:53 |
| Breath Volume      | 3.199 | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.082 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Diagnostics Check  | OK    | 18:54 |

---

Data Download Date/Time: 04/23/2024 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JEFFREY,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: NEAL,LOUIS  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 14:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/09/2024 15:55:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:52 |                      |
| Air Blank         | 0.000 | 15:52 |                      |
| Control Test      | 0.084 | 15:53 |                      |
| Air Blank         | 0.000 | 15:53 |                      |
| Subject Sample    | # 1   | 15:53 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:53 |                      |
| Air Blank         | 0.000 | 15:54 |                      |
| Control Test      | 0.083 | 15:54 |                      |
| Air Blank         | 0.000 | 15:55 |                      |
| Diagnostics Check | OK    | 15:55 |                      |

---

Data Download Date/Time: 04/23/2024 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: ROSS-JOHNSON,GARRETT  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG229803  
DRY GAS EXPIRY DATE: 10/25/2024  
VIOLATION CODE: DUI  
ARREST TIME: 16:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DOMINGUEZ-DOMINGUEZ

DATE/TIME OF TEST: 04/10/2024 17:44:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.083 | 17:39 |
| Air Blank          | 0.000 | 17:39 |
| Subject Sample # 1 | 0.000 | 17:40 |
| Breath Volume      | 4.625 | 17:40 |
| Air Blank          | 0.000 | 17:40 |
| Air Blank          | 0.000 | 17:42 |
| Subject Sample # 2 | 0.000 | 17:43 |
| Breath Volume      | 5.273 | 17:43 |
| Air Blank          | 0.000 | 17:43 |
| Control Test       | 0.083 | 17:43 |
| Air Blank          | 0.000 | 17:44 |
| Diagnostics Check  | OK    | 17:44 |

---

Data Download Date/Time: 04/23/2024 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: SHOULTA,KATHERINE  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG329101  
DRY GAS EXPIRY DATE: 10/18/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: MAGED

DATE/TIME OF TEST: 04/14/2024 05:27:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:21 |
| Air Blank          | 0.000 | 05:21 |
| Control Test       | 0.084 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 1 | 0.047 | 05:23 |
| Breath Volume      | 2.117 | 05:23 |
| Air Blank          | 0.000 | 05:23 |
| Air Blank          | 0.000 | 05:25 |
| Subject Sample # 2 | 0.045 | 05:25 |
| Breath Volume      | 2.167 | 05:25 |
| Air Blank          | 0.000 | 05:26 |
| Control Test       | 0.083 | 05:26 |
| Air Blank          | 0.000 | 05:27 |
| Diagnostics Check  | OK    | 05:27 |

---

Data Download Date/Time: 04/23/2024 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JEFFREY,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: GUIRGUIS,JOSEPH  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG329101  
DRY GAS EXPIRY DATE: 10/18/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: STONER

DATE/TIME OF TEST: 04/14/2024 15:59:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 15:56 |                      |
| Air Blank         | 0.000 | 15:56 |                      |
| Control Test      | 0.084 | 15:57 |                      |
| Air Blank         | 0.000 | 15:57 |                      |
| Subject Sample    | # 1   | 15:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 15:57 |                      |
| Air Blank         | 0.000 | 15:58 |                      |
| Control Test      | 0.084 | 15:58 |                      |
| Air Blank         | 0.000 | 15:59 |                      |
| Diagnostics Check | OK    | 15:59 |                      |

---

Data Download Date/Time: 04/23/2024 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NELSON,JEFFREY,L  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: FOGLER-TORTORA,JENNIFER  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG329101  
DRY GAS EXPIRY DATE: 10/18/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: WARDWELL

DATE/TIME OF TEST: 04/16/2024 16:55:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:53 |                      |
| Air Blank         | 0.000 | 16:53 |                      |
| Control Test      | 0.085 | 16:53 |                      |
| Air Blank         | 0.000 | 16:54 |                      |
| Subject Sample    | # 1   | 16:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:54 |                      |
| Air Blank         | 0.000 | 16:54 |                      |
| Control Test      | 0.084 | 16:55 |                      |
| Air Blank         | 0.000 | 16:55 |                      |
| Diagnostics Check | OK    | 16:55 |                      |

---

Data Download Date/Time: 04/23/2024 10:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RICE,JOHN,D  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: WIMBERLEY,EDWARD  
ARRESTING OFFICER AGENCY: Taylor County SO  
LAST AGENCY INSPECTION: 04/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A2  
DRY GAS EXPIRY DATE: 05/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 04/27/2024 00:22:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.080 | 00:16 |
| Air Blank          | 0.000 | 00:16 |
| Subject Sample # 1 | 0.288 | 00:17 |
| Breath Volume      | 2.222 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Air Blank          | 0.000 | 00:20 |
| Subject Sample # 2 | 0.289 | 00:20 |
| Breath Volume      | 2.144 | 00:20 |
| Air Blank          | 0.000 | 00:21 |
| Control Test       | 0.079 | 00:21 |
| Air Blank          | 0.000 | 00:21 |
| Diagnostics Check  | OK    | 00:22 |

---

Data Download Date/Time: 05/28/2024 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAFFIN,TRACE,D  
BREATH TEST OPERATOR AGENCY: Perry PD  
ARRESTING OFFICER: NORRIS,JONATHAN  
ARRESTING OFFICER AGENCY: Perry PD  
LAST AGENCY INSPECTION: 04/15/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A2  
DRY GAS EXPIRY DATE: 05/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 04/29/2024 09:42:29

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 09:27 |                |
| Air Blank         | 0.000     | 09:28 |                |
| Control Test      | 0.081     | 09:28 |                |
| Air Blank         | 0.000     | 09:29 |                |
| Subject Sample    | # 1 0.000 | 09:29 |                |
| Breath Volume     | 1.742     | 09:29 |                |
| Air Blank         | 0.000     | 09:30 |                |
| Air Blank         | 0.000     | 09:32 |                |
| Subject Sample    | # 2 0.000 | 09:35 | Volume Not Met |
| Breath Volume     | 0.425     | 09:35 |                |
| Air Blank         | 0.000     | 09:35 |                |
| Air Blank         | 0.000     | 09:37 |                |
| Subject Sample    | # 3 0.000 | 09:41 | Volume Not Met |
| Breath Volume     | 1.015     | 09:41 |                |
| Air Blank         | 0.000     | 09:41 |                |
| Control Test      | 0.079     | 09:41 |                |
| Air Blank         | 0.000     | 09:42 |                |
| Diagnostics Check | OK        | 09:42 |                |

---

Data Download Date/Time: 05/28/2024 13:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBNETT,S,R  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: BOWERS,E  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 04/03/2024 03:57:03

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 03:45 |                    |
| Air Blank         | 0.000 | 03:46 |                    |
| Control Test      | 0.078 | 03:46 |                    |
| Air Blank         | 0.000 | 03:46 |                    |
| Subject Sample    | # 1   | 03:50 | No Sample Provided |
| Breath Volume     | 0.000 | 03:50 |                    |
| Air Blank         | 0.000 | 03:50 |                    |
| Air Blank         | 0.000 | 03:52 |                    |
| Subject Sample    | # 2   | 03:55 | No Sample Provided |
| Breath Volume     | 0.000 | 03:55 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Control Test      | 0.078 | 03:56 |                    |
| Air Blank         | 0.000 | 03:56 |                    |
| Diagnostics Check | OK    | 03:57 |                    |

Data Download Date/Time: 04/04/2024 10:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HALL,DUSTIN,L  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: MCKINNON,CHRISTIAN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 19:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: CATALDO

DATE/TIME OF TEST: 04/03/2024 20:33:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:27 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.077 | 20:28 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 1 | 0.000 | 20:29 |
| Breath Volume      | 2.316 | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 2 | 0.000 | 20:32 |
| Breath Volume      | 1.480 | 20:32 |
| Air Blank          | 0.000 | 20:32 |
| Control Test       | 0.077 | 20:33 |
| Air Blank          | 0.000 | 20:33 |
| Diagnostics Check  | OK    | 20:33 |

---

Data Download Date/Time: 04/04/2024 10:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,SAMANTHA,E  
BREATH TEST OPERATOR AGENCY: St. Petersburg PD  
ARRESTING OFFICER: HALL,DUSTIN  
ARRESTING OFFICER AGENCY: St. Petersburg PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08622080A1  
DRY GAS EXPIRY DATE: 06/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: RAWLS

DATE/TIME OF TEST: 04/03/2024 22:11:11

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:02 |                      |
| Air Blank         | 0.000 | 22:03 |                      |
| Control Test      | 0.077 | 22:03 |                      |
| Air Blank         | 0.000 | 22:03 |                      |
| Subject Sample    | # 1   | 22:07 | No Sample Provided   |
| Breath Volume     | 0.000 | 22:07 |                      |
| Air Blank         | 0.000 | 22:07 |                      |
| Air Blank         | 0.000 | 22:09 |                      |
| Subject Sample    | # 2   | 22:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:09 |                      |
| Air Blank         | 0.000 | 22:10 |                      |
| Control Test      | 0.078 | 22:10 |                      |
| Air Blank         | 0.000 | 22:11 |                      |
| Diagnostics Check | OK    | 22:11 |                      |

Data Download Date/Time: 04/04/2024 10:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEL VALLE,M,  
BREATH TEST OPERATOR AGENCY: Dade City PD  
ARRESTING OFFICER: SERRANO,V  
ARRESTING OFFICER AGENCY: Dade City PD  
LAST AGENCY INSPECTION: 04/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A3  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 04/22/2024 00:27:49

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 00:21 |            |
| Air Blank          | 0.000 | 00:21 |            |
| Control Test       | 0.081 | 00:22 |            |
| Air Blank          | 0.000 | 00:22 |            |
| Subject Sample # 1 | 0.161 | 00:23 |            |
| Breath Volume      | 2.160 | 00:23 |            |
| Air Blank          | 0.000 | 00:23 |            |
| Air Blank          | 0.000 | 00:25 |            |
| Subject Sample # 2 | 0.139 | 00:26 |            |
| Breath Volume      | 2.710 | 00:26 |            |
| Air Blank          |       | 00:27 | Purge Fail |
| Air Blank          | 0.000 | 00:27 |            |

Data Download Date/Time: 05/24/2024 03:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEL VALLE,M,  
BREATH TEST OPERATOR AGENCY: Dade City PD  
ARRESTING OFFICER: SERRANO,V  
ARRESTING OFFICER AGENCY: Dade City PD  
LAST AGENCY INSPECTION: 04/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A3  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 04/22/2024 00:32:09

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 00:30 |                           |
| Air Blank         | 0.000 | 00:31 |                           |
| Control Test      | 0.074 | 00:31 | Control Outside Tolerance |
| Air Blank         | 0.000 | 00:32 |                           |

---

Data Download Date/Time: 05/24/2024 03:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

INSTRUMENT REGISTERED TO: DADE CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEL VALLE,M,  
BREATH TEST OPERATOR AGENCY: Dade City PD  
ARRESTING OFFICER: SERRANO,V  
ARRESTING OFFICER AGENCY: Dade City PD  
LAST AGENCY INSPECTION: 04/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24923080A3  
DRY GAS EXPIRY DATE: 10/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 04/22/2024 00:40:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Control Test       | 0.077 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 1 | 0.150 | 00:35 |
| Breath Volume      | 3.277 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 2 | 0.149 | 00:38 |
| Breath Volume      | 3.195 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.077 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Diagnostics Check  | OK    | 00:40 |

---

Data Download Date/Time: 05/24/2024 03:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUINTANA,JORGE,F  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DELGADO FLORES

DATE/TIME OF TEST: 04/06/2024 01:38:14

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:35 |                      |
| Air Blank         | 0.000 | 01:35 |                      |
| Control Test      | 0.080 | 01:36 |                      |
| Air Blank         | 0.000 | 01:36 |                      |
| Subject Sample    | # 1   | 01:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:36 |                      |
| Air Blank         | 0.000 | 01:37 |                      |
| Control Test      | 0.079 | 01:37 |                      |
| Air Blank         | 0.000 | 01:38 |                      |
| Diagnostics Check | OK    | 01:38 |                      |

---

Data Download Date/Time: 04/11/2024 04:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,DAEL,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: QUINTANA,JORGE  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FERNANDEZ LEON

DATE/TIME OF TEST: 04/07/2024 00:11:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:08 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Control Test      | 0.079 | 00:09 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Subject Sample    | # 1   | 00:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:10 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Control Test      | 0.079 | 00:10 |                      |
| Air Blank         | 0.000 | 00:11 |                      |
| Diagnostics Check | OK    | 00:11 |                      |

---

Data Download Date/Time: 04/11/2024 04:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: HERRERA-MARTINEZ

DATE/TIME OF TEST: 04/12/2024 01:03:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.078 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.199 | 00:58 |
| Breath Volume      | 2.410 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.194 | 01:01 |
| Breath Volume      | 4.210 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.078 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 05/02/2024 03:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,JOSEPH,  
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD  
ARRESTING OFFICER: AZRAK,JOSEPH  
ARRESTING OFFICER AGENCY: MIAMI-DADE PD  
LAST AGENCY INSPECTION: 04/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: MENDIZBAL-DELGADO

DATE/TIME OF TEST: 04/21/2024 03:08:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.077 | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Subject Sample # 1 | 0.162 | 03:03 |
| Breath Volume      | 1.355 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Air Blank          | 0.000 | 03:06 |
| Subject Sample # 2 | 0.162 | 03:06 |
| Breath Volume      | 1.933 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Control Test       | 0.077 | 03:07 |
| Air Blank          | 0.000 | 03:08 |
| Diagnostics Check  | OK    | 03:08 |

---

Data Download Date/Time: 05/02/2024 03:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LEGROS

DATE/TIME OF TEST: 04/20/2024 00:45:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Control Test       | 0.080 | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 1 | 0.063 | 00:40 |
| Breath Volume      | 2.773 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Air Blank          | 0.000 | 00:42 |
| Subject Sample # 2 | 0.066 | 00:43 |
| Breath Volume      | 2.675 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.079 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Diagnostics Check  | OK    | 00:45 |

---

Data Download Date/Time: 05/02/2024 03:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: MARTIN,YSABELLA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/20/2024 04:52:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:49 |                      |
| Air Blank         | 0.000 | 04:50 |                      |
| Control Test      | 0.080 | 04:50 |                      |
| Air Blank         | 0.000 | 04:51 |                      |
| Subject Sample    | # 1   | 04:51 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:51 |                      |
| Air Blank         | 0.000 | 04:51 |                      |
| Control Test      | 0.079 | 04:52 |                      |
| Air Blank         | 0.000 | 04:52 |                      |
| Diagnostics Check | OK    | 04:52 |                      |

---

Data Download Date/Time: 05/02/2024 03:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080A2  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 04/21/2024 05:47:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:44 |                      |
| Air Blank         | 0.000 | 05:44 |                      |
| Control Test      | 0.079 | 05:45 |                      |
| Air Blank         | 0.000 | 05:45 |                      |
| Subject Sample    | # 1   | 05:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:46 |                      |
| Air Blank         | 0.000 | 05:46 |                      |
| Control Test      | 0.079 | 05:46 |                      |
| Air Blank         | 0.000 | 05:47 |                      |
| Diagnostics Check | OK    | 05:47 |                      |

---

Data Download Date/Time: 05/02/2024 03:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSS,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: GOSS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: RHULE

DATE/TIME OF TEST: 04/20/2024 17:08:56

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 17:01 |            |
| Air Blank          | 0.000 | 17:02 |            |
| Control Test       | 0.078 | 17:02 |            |
| Air Blank          | 0.000 | 17:03 |            |
| Subject Sample # 1 | 0.210 | 17:04 |            |
| Breath Volume      | 1.402 | 17:04 |            |
| Air Blank          | 0.000 | 17:05 |            |
| Air Blank          | 0.000 | 17:06 |            |
| Subject Sample # 2 | 0.197 | 17:07 |            |
| Breath Volume      | 1.480 | 17:07 |            |
| Air Blank          |       | 17:08 | Purge Fail |
| Air Blank          | 0.000 | 17:08 |            |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSS,MICHAEL,P  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: GOSS,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: RHULE

DATE/TIME OF TEST: 04/20/2024 17:19:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:12 |
| Air Blank          | 0.000 | 17:13 |
| Control Test       | 0.077 | 17:13 |
| Air Blank          | 0.000 | 17:14 |
| Subject Sample # 1 | 0.188 | 17:15 |
| Breath Volume      | 1.257 | 17:15 |
| Air Blank          | 0.000 | 17:16 |
| Air Blank          | 0.000 | 17:17 |
| Subject Sample # 2 | 0.192 | 17:18 |
| Breath Volume      | 1.289 | 17:18 |
| Air Blank          | 0.000 | 17:18 |
| Control Test       | 0.076 | 17:19 |
| Air Blank          | 0.000 | 17:19 |
| Diagnostics Check  | OK    | 17:19 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: LUTZ, D  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 04/20/2024 23:52:41

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:46 |                      |
| Air Blank         | 0.000 | 23:47 |                      |
| Control Test      | 0.079 | 23:47 |                      |
| Air Blank         | 0.000 | 23:48 |                      |
| Subject Sample    | # 1   | 23:51 | Subject Test Refused |
| Breath Volume     | 0.058 | 23:51 |                      |
| Air Blank         |       | 23:51 | Purge Fail           |
| Air Blank         |       | 23:52 | Purge Fail           |

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: PEREZ, ROBERTO  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 04/25/2024 03:38:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.079 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.281 | 03:33 |
| Breath Volume      | 1.890 | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.274 | 03:36 |
| Breath Volume      | 1.707 | 03:36 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.077 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JORGE,ADRIAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JORGE,ADRIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 04/25/2024 03:49:55

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:46 |            |
| Air Blank         | 0.000     | 03:46 |            |
| Control Test      | 0.077     | 03:47 |            |
| Air Blank         | 0.000     | 03:47 |            |
| Subject Sample    | # 1 0.000 | 03:48 |            |
| Breath Volume     | 2.105     | 03:48 |            |
| Air Blank         |           | 03:48 | RFI Detect |
| Air Blank         | 0.000     | 03:48 |            |
| Control Test      | 0.078     | 03:49 |            |
| Air Blank         | 0.000     | 03:49 |            |
| Diagnostics Check | OK        | 03:49 |            |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JORGE,ADRIAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: JORGE,ADRIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 04/25/2024 03:58:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Control Test       | 0.078 | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Subject Sample # 1 | 0.000 | 03:54 |
| Breath Volume      | 2.226 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Air Blank          | 0.000 | 03:56 |
| Subject Sample # 2 | 0.000 | 03:57 |
| Breath Volume      | 1.777 | 03:57 |
| Air Blank          | 0.000 | 03:57 |
| Control Test       | 0.079 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Diagnostics Check  | OK    | 03:58 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINDEMANN, ANTHONY, M  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: ALLEN, TASHA  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/25/2024 04:12:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.078 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 1 | 0.013 | 04:07 |
| Breath Volume      | 3.765 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 2 | 0.009 | 04:10 |
| Breath Volume      | 1.488 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.078 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Diagnostics Check  | OK    | 04:12 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MANIOS,ALEXANDER,W  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: MANIOS,ALEXANDER  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: SIRK

DATE/TIME OF TEST: 04/25/2024 16:10:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:03 |
| Air Blank          | 0.000 | 16:03 |
| Control Test       | 0.079 | 16:04 |
| Air Blank          | 0.000 | 16:04 |
| Subject Sample # 1 | 0.179 | 16:05 |
| Breath Volume      | 4.093 | 16:05 |
| Air Blank          | 0.000 | 16:06 |
| Air Blank          | 0.000 | 16:07 |
| Subject Sample # 2 | 0.168 | 16:08 |
| Breath Volume      | 3.156 | 16:08 |
| Air Blank          | 0.000 | 16:09 |
| Control Test       | 0.077 | 16:09 |
| Air Blank          | 0.000 | 16:10 |
| Diagnostics Check  | OK    | 16:10 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: RICE,BRET  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:31

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/28/2024 10:57:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:51 |
| Air Blank          | 0.000 | 10:52 |
| Control Test       | 0.080 | 10:52 |
| Air Blank          | 0.000 | 10:52 |
| Subject Sample # 1 | 0.000 | 10:53 |
| Breath Volume      | 2.417 | 10:53 |
| Air Blank          | 0.000 | 10:53 |
| Air Blank          | 0.000 | 10:55 |
| Subject Sample # 2 | 0.000 | 10:56 |
| Breath Volume      | 2.335 | 10:56 |
| Air Blank          | 0.000 | 10:56 |
| Control Test       | 0.078 | 10:57 |
| Air Blank          | 0.000 | 10:57 |
| Diagnostics Check  | OK    | 10:57 |

---

Data Download Date/Time: 05/07/2024 12:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: YOCHUM,CRAIG  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A3  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: RUPP

DATE/TIME OF TEST: 04/06/2024 22:30:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Control Test       | 0.078 | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Subject Sample # 1 | 0.135 | 22:26 |
| Breath Volume      | 2.097 | 22:26 |
| Air Blank          | 0.000 | 22:26 |
| Air Blank          | 0.000 | 22:28 |
| Subject Sample # 2 | 0.129 | 22:29 |
| Breath Volume      | 1.242 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.078 | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Diagnostics Check  | OK    | 22:30 |

---

Data Download Date/Time: 04/19/2024 14:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: PICARD,RONNIE  
ARRESTING OFFICER AGENCY: Jupiter PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 15922080A3  
DRY GAS EXPIRY DATE: 08/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: WEISSING

DATE/TIME OF TEST: 04/07/2024 00:44:19

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:35 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Control Test      | 0.078 | 00:36 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Subject Sample    | # 1   | 00:40 | No Sample Provided   |
| Breath Volume     | 0.000 | 00:40 |                      |
| Air Blank         | 0.000 | 00:40 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Subject Sample    | # 2   | 00:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:42 |                      |
| Air Blank         | 0.000 | 00:43 |                      |
| Control Test      | 0.077 | 00:43 |                      |
| Air Blank         | 0.000 | 00:44 |                      |
| Diagnostics Check | OK    | 00:44 |                      |

Data Download Date/Time: 04/19/2024 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POUND,PARIS,D  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AGAOGLU,YASEMIN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 29122080A1  
DRY GAS EXPIRY DATE: 12/05/2024  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CHACON PAIZ

DATE/TIME OF TEST: 04/19/2024 23:54:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:48 |
| Air Blank          | 0.000 | 23:48 |
| Control Test       | 0.079 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 1 | 0.140 | 23:49 |
| Breath Volume      | 1.503 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 2 | 0.141 | 23:52 |
| Breath Volume      | 1.875 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.080 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Diagnostics Check  | OK    | 23:54 |

---

Data Download Date/Time: 05/17/2024 13:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: REYNA LEYVA

DATE/TIME OF TEST: 04/01/2024 00:49:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.080 | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 1 | 0.089 | 00:43 |
| Breath Volume      | 1.503 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 2 | 0.106 | 00:47 |
| Breath Volume      | 2.976 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.079 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:49 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,THOMASINA,E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: IVANOVIC,MATTHEW  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: RIGGLEMAN

DATE/TIME OF TEST: 04/01/2024 02:32:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.080 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 1 | 0.132 | 02:27 |
| Breath Volume      | 1.687 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 2 | 0.125 | 02:30 |
| Breath Volume      | 2.312 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DESUYO, AISON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: BALLINGER

DATE/TIME OF TEST: 04/02/2024 00:27:56

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:13 |                |
| Air Blank         | 0.000     | 00:14 |                |
| Control Test      | 0.080     | 00:14 |                |
| Air Blank         | 0.000     | 00:15 |                |
| Subject Sample    | # 1 0.135 | 00:18 | Volume Not Met |
| Breath Volume     | 0.484     | 00:18 |                |
| Air Blank         | 0.000     | 00:19 |                |
| Air Blank         | 0.000     | 00:20 |                |
| Subject Sample    | # 2 0.174 | 00:21 |                |
| Breath Volume     | 1.480     | 00:21 |                |
| Air Blank         | 0.000     | 00:22 |                |
| Air Blank         | 0.000     | 00:24 |                |
| Subject Sample    | # 3 0.174 | 00:26 |                |
| Breath Volume     | 1.921     | 00:26 |                |
| Air Blank         | 0.000     | 00:27 |                |
| Control Test      | 0.079     | 00:27 |                |
| Air Blank         | 0.000     | 00:27 |                |
| Diagnostics Check | OK        | 00:27 |                |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARI,L  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DESUYO,AISON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: SANCHEZ MAZARIEGOS

DATE/TIME OF TEST: 04/05/2024 22:37:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Control Test       | 0.080 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Subject Sample # 1 | 0.136 | 22:33 |
| Breath Volume      | 1.179 | 22:33 |
| Air Blank          | 0.000 | 22:33 |
| Air Blank          | 0.000 | 22:35 |
| Subject Sample # 2 | 0.129 | 22:36 |
| Breath Volume      | 1.449 | 22:36 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.080 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Diagnostics Check  | OK    | 22:37 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WHITE,EDWARD  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CARCAMO

DATE/TIME OF TEST: 04/06/2024 02:42:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:35 |
| Air Blank          | 0.000 | 02:36 |
| Control Test       | 0.080 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 1 | 0.170 | 02:37 |
| Breath Volume      | 2.113 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Air Blank          | 0.000 | 02:40 |
| Subject Sample # 2 | 0.154 | 02:40 |
| Breath Volume      | 1.308 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.079 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Diagnostics Check  | OK    | 02:42 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SNELGROVE,DYLAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/06/2024 04:38:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.080 | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Subject Sample # 1 | 0.151 | 04:34 |
| Breath Volume      | 2.320 | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Air Blank          | 0.000 | 04:36 |
| Subject Sample # 2 | 0.141 | 04:37 |
| Breath Volume      | 1.699 | 04:37 |
| Air Blank          | 0.000 | 04:37 |
| Control Test       | 0.079 | 04:38 |
| Air Blank          | 0.000 | 04:38 |
| Diagnostics Check  | OK    | 04:38 |

---

Data Download Date/Time: 04/19/2024 17:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN,RENEE,M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CARTER,RUBEN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: CAREY

DATE/TIME OF TEST: 04/06/2024 15:27:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:17 |
| Air Blank          | 0.000 | 15:17 |
| Control Test       | 0.079 | 15:18 |
| Air Blank          | 0.000 | 15:18 |
| Subject Sample # 1 | 0.244 | 15:19 |
| Breath Volume      | 1.875 | 15:19 |
| Air Blank          | 0.000 | 15:20 |
| Air Blank          | 0.000 | 15:21 |
| Subject Sample # 2 | 0.217 | 15:22 |
| Breath Volume      | 1.273 | 15:22 |
| Air Blank          | 0.000 | 15:23 |
| Air Blank          | 0.000 | 15:24 |
| Subject Sample # 3 | 0.232 | 15:25 |
| Breath Volume      | 2.519 | 15:25 |
| Air Blank          | 0.000 | 15:26 |
| Control Test       | 0.079 | 15:26 |
| Air Blank          | 0.000 | 15:27 |
| Diagnostics Check  | OK    | 15:27 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAGIN, RENEE, M  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DANDREA, ADAM  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: SEBASTIAN

DATE/TIME OF TEST: 04/06/2024 15:46:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:39 |
| Air Blank          | 0.000 | 15:39 |
| Control Test       | 0.079 | 15:40 |
| Air Blank          | 0.000 | 15:40 |
| Subject Sample # 1 | 0.261 | 15:41 |
| Breath Volume      | 1.414 | 15:41 |
| Air Blank          | 0.000 | 15:41 |
| Air Blank          | 0.000 | 15:43 |
| Subject Sample # 2 | 0.259 | 15:44 |
| Breath Volume      | 2.625 | 15:44 |
| Air Blank          | 0.000 | 15:45 |
| Control Test       | 0.079 | 15:45 |
| Air Blank          | 0.000 | 15:46 |
| Diagnostics Check  | OK    | 15:46 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: RICCI,MATTHEW  
ARRESTING OFFICER AGENCY: Juno Beach PD  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ANZALONE

DATE/TIME OF TEST: 04/07/2024 03:36:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.080 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.098 | 03:32 |
| Breath Volume      | 1.398 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Air Blank          | 0.000 | 03:34 |
| Subject Sample # 2 | 0.089 | 03:35 |
| Breath Volume      | 1.527 | 03:35 |
| Air Blank          | 0.000 | 03:35 |
| Control Test       | 0.078 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Diagnostics Check  | OK    | 03:36 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT, ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ORTIZ PEREZ

DATE/TIME OF TEST: 04/08/2024 00:52:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.081 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 1 | 0.289 | 00:47 |
| Breath Volume      | 2.207 | 00:47 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 2 | 0.282 | 00:50 |
| Breath Volume      | 2.437 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Control Test       | 0.079 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Diagnostics Check  | OK    | 00:52 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,MADELIN,  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KAHN,JUSTIN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 04/09/2024 03:37:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.081 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.163 | 03:33 |
| Breath Volume      | 2.187 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.158 | 03:36 |
| Breath Volume      | 2.386 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.079 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SENTMANAT, ADOLFO  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 04/10/2024 00:03:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Control Test       | 0.080 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Subject Sample # 2 | 0.147 | 23:58 |
| Breath Volume      | 2.355 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Air Blank          | 0.000 | 00:01 |
| Subject Sample # 1 | 0.146 | 00:02 |
| Breath Volume      | 2.710 | 00:02 |
| Air Blank          | 0.000 | 00:02 |
| Control Test       | 0.079 | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Diagnostics Check  | OK    | 00:03 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,JOSHUA,J  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: LOVETT,JAMES  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: VORDENBERG

DATE/TIME OF TEST: 04/10/2024 03:03:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:57 |
| Air Blank          | 0.000 | 02:57 |
| Control Test       | 0.080 | 02:58 |
| Air Blank          | 0.000 | 02:58 |
| Subject Sample # 1 | 0.174 | 02:59 |
| Breath Volume      | 2.320 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Air Blank          | 0.000 | 03:01 |
| Subject Sample # 2 | 0.173 | 03:02 |
| Breath Volume      | 1.929 | 03:02 |
| Air Blank          | 0.000 | 03:02 |
| Control Test       | 0.079 | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Diagnostics Check  | OK    | 03:03 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WARD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VICENTE LOPEZ

DATE/TIME OF TEST: 04/11/2024 02:27:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:20 |
| Air Blank          | 0.000 | 02:20 |
| Control Test       | 0.080 | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 1 | 0.107 | 02:22 |
| Breath Volume      | 1.726 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 2 | 0.119 | 02:25 |
| Breath Volume      | 3.406 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.080 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Diagnostics Check  | OK    | 02:27 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: CISSON,JACKIE  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 04/13/2024 00:28:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:22 |
| Air Blank          | 0.000 | 00:22 |
| Control Test       | 0.080 | 00:22 |
| Air Blank          | 0.000 | 00:23 |
| Subject Sample # 1 | 0.149 | 00:23 |
| Breath Volume      | 2.066 | 00:23 |
| Air Blank          | 0.000 | 00:24 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 2 | 0.154 | 00:26 |
| Breath Volume      | 2.886 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Control Test       | 0.080 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Diagnostics Check  | OK    | 00:28 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: RENTERIA,ELIZABETH  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 04/13/2024 02:45:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.081 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.011 | 02:41 |
| Breath Volume      | 1.343 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.011 | 02:44 |
| Breath Volume      | 1.785 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.081 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Diagnostics Check  | OK    | 02:45 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LEVITUS

DATE/TIME OF TEST: 04/13/2024 04:15:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.080 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 1 | 0.246 | 04:10 |
| Breath Volume      | 1.992 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 2 | 0.237 | 04:13 |
| Breath Volume      | 2.402 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Control Test       | 0.080 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Diagnostics Check  | OK    | 04:15 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON,WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 04/13/2024 22:27:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.080 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Subject Sample # 1 | 0.218 | 22:22 |
| Breath Volume      | 2.496 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Air Blank          | 0.000 | 22:25 |
| Subject Sample # 2 | 0.212 | 22:25 |
| Breath Volume      | 1.328 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Control Test       | 0.079 | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Diagnostics Check  | OK    | 22:27 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARI,L  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: DESUYO,AISON  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/14/2024 01:37:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.084 | 01:33 |
| Breath Volume      | 1.144 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.084 | 01:36 |
| Breath Volume      | 1.570 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.080 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: WALTHER, WYATT  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: VICENTE CHAVEZ

DATE/TIME OF TEST: 04/14/2024 02:14:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Control Test       | 0.080 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Subject Sample # 1 | 0.307 | 02:10 |
| Breath Volume      | 3.234 | 02:10 |
| Air Blank          | 0.000 | 02:10 |
| Air Blank          | 0.000 | 02:12 |
| Subject Sample # 2 | 0.306 | 02:13 |
| Breath Volume      | 3.722 | 02:13 |
| Air Blank          | 0.000 | 02:13 |
| Control Test       | 0.079 | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Diagnostics Check  | OK    | 02:14 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON, WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 04/14/2024 04:06:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.080 | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 1 | 0.108 | 04:01 |
| Breath Volume      | 2.781 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Air Blank          | 0.000 | 04:03 |
| Subject Sample # 2 | 0.111 | 04:04 |
| Breath Volume      | 2.589 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.079 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Diagnostics Check  | OK    | 04:06 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: BENNETT, JONATHAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: CHUN CALEL

DATE/TIME OF TEST: 04/14/2024 15:28:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:22 |
| Air Blank          | 0.000 | 15:22 |
| Control Test       | 0.081 | 15:22 |
| Air Blank          | 0.000 | 15:23 |
| Subject Sample # 1 | 0.137 | 15:23 |
| Breath Volume      | 1.679 | 15:23 |
| Air Blank          | 0.000 | 15:24 |
| Air Blank          | 0.000 | 15:26 |
| Subject Sample # 2 | 0.136 | 15:26 |
| Breath Volume      | 2.261 | 15:26 |
| Air Blank          | 0.000 | 15:27 |
| Control Test       | 0.079 | 15:27 |
| Air Blank          | 0.000 | 15:28 |
| Diagnostics Check  | OK    | 15:28 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEAHEY, THOMAS, H  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: AMADON, WILL  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: PEREZ JIMENEZ

DATE/TIME OF TEST: 04/14/2024 19:55:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Control Test       | 0.080 | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Subject Sample # 1 | 0.197 | 19:50 |
| Breath Volume      | 2.761 | 19:50 |
| Air Blank          | 0.000 | 19:51 |
| Air Blank          | 0.000 | 19:53 |
| Subject Sample # 2 | 0.198 | 19:53 |
| Breath Volume      | 2.839 | 19:53 |
| Air Blank          | 0.000 | 19:54 |
| Control Test       | 0.080 | 19:54 |
| Air Blank          | 0.000 | 19:55 |
| Diagnostics Check  | OK    | 19:55 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: KAHN, JUSTIN  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LAPARRA PEREZ

DATE/TIME OF TEST: 04/15/2024 00:51:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.081 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Subject Sample # 1 | 0.175 | 00:46 |
| Breath Volume      | 2.242 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.176 | 00:49 |
| Breath Volume      | 2.847 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.079 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:51 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, THOMASINA, E  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: SNELGROVE, DYLAN  
ARRESTING OFFICER AGENCY: Palm Beach County SO  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LAFONTANT

DATE/TIME OF TEST: 04/17/2024 03:27:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.080 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.013 | 03:23 |
| Breath Volume      | 1.312 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.015 | 03:26 |
| Breath Volume      | 2.195 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.080 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 04/19/2024 17:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CSO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'NEAL,SHARIL  
BREATH TEST OPERATOR AGENCY: Palm Beach County SO  
ARRESTING OFFICER: FLINK,ANDREW  
ARRESTING OFFICER AGENCY: Palm Beach Gardens  
LAST AGENCY INSPECTION: 03/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06723080A5  
DRY GAS EXPIRY DATE: 04/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LOPES

DATE/TIME OF TEST: 04/19/2024 03:02:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Control Test       | 0.081 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 1 | 0.076 | 02:57 |
| Breath Volume      | 2.703 | 02:57 |
| Air Blank          | 0.000 | 02:58 |
| Air Blank          | 0.000 | 03:00 |
| Subject Sample # 2 | 0.074 | 03:00 |
| Breath Volume      | 2.218 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Control Test       | 0.079 | 03:01 |
| Air Blank          | 0.000 | 03:02 |
| Diagnostics Check  | OK    | 03:02 |

---

Data Download Date/Time: 04/19/2024 17:33