

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: VALERIANO MEDRANO

DATE/TIME OF TEST: 04/14/2024 21:17:35

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.175	21:12
Breath Volume	1.722	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.181	21:15
Breath Volume	1.972	21:15
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GENES ARAUJO

DATE/TIME OF TEST: 04/15/2024 02:55:21

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.176	02:50
Breath Volume	2.675	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.174	02:53
Breath Volume	2.468	02:53
Air Blank	0.000	02:54
Control Test	0.081	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: ROELLE

DATE/TIME OF TEST: 04/16/2024 18:10:42

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.082	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.070	18:06
Breath Volume	1.722	18:06
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.074	18:09
Breath Volume	2.015	18:09
Air Blank	0.000	18:09
Control Test	0.080	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROUSSEAU,DMITRY,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: ROUSSEAU,DMITRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CARBAJAL

DATE/TIME OF TEST: 04/19/2024 01:09:39

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.075	01:05
Breath Volume	2.238	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.079	01:07
Breath Volume	2.292	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ROMERO GARCIA

DATE/TIME OF TEST: 04/20/2024 00:52:40

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.132	00:47
Breath Volume	1.289	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.129	00:51
Breath Volume	1.707	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 04/20/2024 19:40:36

Diagnostics Check	OK	19:34
Air Blank	0.000	19:34
Control Test	0.081	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.109	19:36
Breath Volume	1.449	19:36
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.115	19:39
Breath Volume	1.703	19:39
Air Blank	0.000	19:39
Control Test	0.081	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/23/2024 01:35:43

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.050	01:35	Control Outside Tolerance
Air Blank	0.000	01:35	

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: HEIN

DATE/TIME OF TEST: 04/24/2024 19:03:33

Diagnostics Check	OK	19:00	
Air Blank	0.000	19:01	
Control Test	0.081	19:01	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:02	Subject Test Refused
Breath Volume	0.000	19:02	
Air Blank	0.000	19:02	
Control Test	0.081	19:02	
Air Blank	0.000	19:03	
Diagnostics Check	OK	19:03	

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ROBBLOS LOPEZ

DATE/TIME OF TEST: 04/26/2024 23:56:52

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.093	23:52
Breath Volume	1.792	23:52
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.094	23:55
Breath Volume	2.312	23:55
Air Blank	0.000	23:55
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 04/27/2024 04:35:37

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:28	No Sample Provided
Breath Volume	0.000	04:28	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.081	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SOLANO PEREZ

DATE/TIME OF TEST: 04/28/2024 03:29:44

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.153	03:24
Breath Volume	1.429	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.144	03:28
Breath Volume	1.230	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ELDRIDGE

DATE/TIME OF TEST: 04/28/2024 22:02:56

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.081	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.137	21:58
Breath Volume	1.992	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.145	22:01
Breath Volume	1.906	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PINNA

DATE/TIME OF TEST: 04/29/2024 02:37:39

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.241	02:32	
Breath Volume	1.410	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.077	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 05/13/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CRUZ PEREZ

DATE/TIME OF TEST: 04/06/2024 01:58:36

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.191	01:53
Breath Volume	2.656	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.194	01:56
Breath Volume	2.347	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 04/08/2024 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 04/06/2024 23:22:36

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.082	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.093	23:17
Breath Volume	1.312	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.101	23:20
Breath Volume	1.925	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 04/08/2024 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 04/07/2024 01:25:30

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1 0.189	01:17	Volume Not Met
Breath Volume	0.808	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 2 0.205	01:20	
Breath Volume	1.675	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 3 0.190	01:23	
Breath Volume	1.585	01:23	
Air Blank	0.000	01:24	
Control Test	0.077	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 04/08/2024 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/10/2024 22:24:22

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.080	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.092	22:19
Breath Volume	1.578	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.090	22:22
Breath Volume	1.488	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: KRIGE II

DATE/TIME OF TEST: 04/11/2024 05:29:33

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.053	05:24
Breath Volume	1.378	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.059	05:28
Breath Volume	1.593	05:28
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 04/12/2024 04:08:18

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.102	04:03
Breath Volume	1.128	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.102	04:06
Breath Volume	1.753	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/13/2024 01:41:46

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.136	01:36
Breath Volume	1.628	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.133	01:40
Breath Volume	1.371	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/13/2024 03:42:12

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.118	03:37
Breath Volume	3.140	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.115	03:40
Breath Volume	2.312	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: JIMENEZ MEJIA

DATE/TIME OF TEST: 04/13/2024 04:40:32

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.140	04:35
Breath Volume	1.792	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.143	04:38
Breath Volume	1.496	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: ECHEVARRIA GUEVARA

DATE/TIME OF TEST: 04/13/2024 12:58:45

Diagnostics Check	OK	12:44	
Air Blank	0.000	12:44	
Control Test	0.081	12:45	
Air Blank	0.000	12:45	
Subject Sample	# 1 0.059	12:46	
Breath Volume	1.324	12:46	
Air Blank	0.000	12:46	
Air Blank	0.000	12:48	
Subject Sample	# 2	12:51	No Sample Provided
Breath Volume	0.000	12:51	
Air Blank	0.000	12:52	
Air Blank	0.000	12:54	
Subject Sample	# 3	12:57	No Sample Provided
Breath Volume	0.000	12:57	
Air Blank	0.000	12:57	
Control Test	0.081	12:58	
Air Blank	0.000	12:58	
Diagnostics Check	OK	12:58	

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,Z
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/13/2024 22:52:57

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:43	
Control Test	0.081	22:44	
Air Blank	0.000	22:44	
Subject Sample	# 1	22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 2	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ROSALES RUIZ

DATE/TIME OF TEST: 04/14/2024 00:25:33

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.153	00:21
Breath Volume	1.691	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.152	00:23
Breath Volume	2.050	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CABEBE

DATE/TIME OF TEST: 04/14/2024 01:59:32

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HENLEY,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LUU

DATE/TIME OF TEST: 04/14/2024 04:57:36

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.184	04:53
Breath Volume	1.351	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.182	04:55
Breath Volume	2.296	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: BOU SAMRA

DATE/TIME OF TEST: 04/14/2024 05:44:42

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.081	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:37	No Sample Provided
Breath Volume	0.000	05:37	
Air Blank	0.000	05:38	
Air Blank	0.000	05:40	
Subject Sample	# 2	05:43	No Sample Provided
Breath Volume	0.000	05:43	
Air Blank	0.000	05:43	
Control Test	0.081	05:44	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: CALA CASTRO

DATE/TIME OF TEST: 04/14/2024 06:50:54

Diagnostics Check	OK	06:41
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:42
Subject Sample # 1	0.150	06:45
Breath Volume	1.175	06:45
Air Blank	0.000	06:46
Air Blank	0.000	06:47
Subject Sample # 2	0.151	06:49
Breath Volume	1.980	06:49
Air Blank	0.000	06:49
Control Test	0.079	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: METZ

DATE/TIME OF TEST: 04/15/2024 02:30:39

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.169	02:26
Breath Volume	2.394	02:26
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.169	02:29
Breath Volume	1.996	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAISSE,A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: OLIVERA GUADALUPE

DATE/TIME OF TEST: 04/15/2024 07:08:21

Diagnostics Check	OK	06:53	
Air Blank	0.000	06:54	
Control Test	0.081	06:54	
Air Blank	0.000	06:54	
Subject Sample	# 1 0.087	06:58	Volume Not Met
Breath Volume	0.593	06:58	
Air Blank	0.000	06:58	
Air Blank	0.000	07:00	
Subject Sample	# 2 0.100	07:01	
Breath Volume	1.132	07:01	
Air Blank	0.000	07:01	
Air Blank	0.000	07:03	
Subject Sample	# 3 0.101	07:06	Volume Not Met
Breath Volume	0.703	07:06	
Air Blank	0.000	07:07	
Control Test	0.081	07:07	
Air Blank	0.000	07:08	
Diagnostics Check	OK	07:08	

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAISSE,A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: OLIVERA GUADALUPE

DATE/TIME OF TEST: 04/15/2024 07:18:26

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.079	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.096	07:12
Breath Volume	1.222	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.094	07:16
Breath Volume	1.410	07:16
Air Blank	0.000	07:17
Control Test	0.080	07:17
Air Blank	0.000	07:18
Diagnostics Check	OK	07:18

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/15/2024 22:26:06

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.174	22:21
Breath Volume	1.976	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.175	22:24
Breath Volume	1.988	22:24
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 04/16/2024 23:43:38

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.081	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.053	23:39
Breath Volume	2.210	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.057	23:42
Breath Volume	2.539	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ALDENDERFER

DATE/TIME OF TEST: 04/17/2024 02:41:52

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.111	02:37
Breath Volume	2.253	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.100	02:40
Breath Volume	2.097	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: CASTRO SANCHEZ

DATE/TIME OF TEST: 04/17/2024 05:10:42

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.081	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.245	05:06
Breath Volume	1.542	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.241	05:09
Breath Volume	1.953	05:09
Air Blank	0.000	05:09
Control Test	0.081	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 05/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/18/2024 23:23:43

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:18
Breath Volume	3.601	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:22
Breath Volume	3.070	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: WHITLOCK,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: WAITE

DATE/TIME OF TEST: 04/19/2024 02:57:32

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.162	02:53
Breath Volume	2.679	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.153	02:55
Breath Volume	1.804	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: COBOS FRIAS

DATE/TIME OF TEST: 04/19/2024 04:00:55

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.145	03:56
Breath Volume	2.898	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.132	03:59
Breath Volume	2.191	03:59
Air Blank	0.000	03:59
Control Test	0.080	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GOETZ

DATE/TIME OF TEST: 04/19/2024 04:46:27

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.153	04:41
Breath Volume	1.574	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.169	04:44
Breath Volume	1.800	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 04/20/2024 02:43:28

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.179	02:38
Breath Volume	1.843	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.170	02:41
Breath Volume	1.855	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MEZA JIMENEZ

DATE/TIME OF TEST: 04/20/2024 03:22:21

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.145	03:17
Breath Volume	3.394	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.140	03:20
Breath Volume	1.535	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS, K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/20/2024 03:39:10

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.145	03:34
Breath Volume	2.085	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.148	03:37
Breath Volume	1.761	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: HEREDIA VICTORIA

DATE/TIME OF TEST: 04/20/2024 05:16:15

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.165	05:10
Breath Volume	1.535	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.164	05:14
Breath Volume	1.703	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/20/2024 22:48:43

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.079	22:43
Breath Volume	1.613	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.074	22:47
Breath Volume	1.191	22:47
Air Blank	0.000	22:47
Control Test	0.080	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DOWE

DATE/TIME OF TEST: 04/21/2024 01:12:17

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.083	01:07
Breath Volume	3.722	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.084	01:10
Breath Volume	3.367	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MARCELUS

DATE/TIME OF TEST: 04/21/2024 02:21:42

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.074	02:17
Breath Volume	1.765	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.075	02:20
Breath Volume	1.656	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 04/21/2024 02:31:49

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.177	02:27
Breath Volume	3.742	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.181	02:30
Breath Volume	2.539	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/21/2024 04:37:04

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.129	04:31
Breath Volume	1.117	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.131	04:35
Breath Volume	1.960	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:36
Diagnostics Check	OK	04:37

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/21/2024 21:51:25

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.117	21:46
Breath Volume	1.781	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.122	21:49
Breath Volume	2.312	21:49
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: ALZATE

DATE/TIME OF TEST: 04/21/2024 22:11:04

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.158	22:06
Breath Volume	1.132	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.165	22:09
Breath Volume	1.328	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:11

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: HERNANDEZ JIMENEZ

DATE/TIME OF TEST: 04/22/2024 02:54:50

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:49
Air Blank	0.000	02:49
Subject Sample # 1	0.154	02:50
Breath Volume	1.179	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.148	02:53
Breath Volume	1.890	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: OKEEFE

DATE/TIME OF TEST: 04/22/2024 19:22:10

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.080	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.318	19:17
Breath Volume	3.109	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:19
Subject Sample # 2	0.333	19:20
Breath Volume	3.257	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: EMANUEL

DATE/TIME OF TEST: 04/23/2024 00:53:13

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.094	00:48
Breath Volume	1.398	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.101	00:51
Breath Volume	2.367	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Plant City PD
ARRESTING OFFICER: SHIRELING,S
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ROCAMORA

DATE/TIME OF TEST: 04/23/2024 23:11:14

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.081	23:02
Air Blank	0.000	23:02
Subject Sample # 1	0.192	23:03
Breath Volume	1.472	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:05
Subject Sample # 2	0.166	23:06
Breath Volume	1.183	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 3	0.188	23:09
Breath Volume	2.675	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAAR, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/25/2024 00:43:30

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.081	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.268	00:38
Breath Volume	1.570	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.266	00:41
Breath Volume	1.480	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MAHAMUD

DATE/TIME OF TEST: 04/26/2024 00:11:07

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.203	00:05
Breath Volume	1.851	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.192	00:09
Breath Volume	1.589	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 04/26/2024 01:37:33

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.142	01:31
Breath Volume	1.933	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.140	01:35
Breath Volume	1.273	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 05/08/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: VALCINE

DATE/TIME OF TEST: 04/26/2024 02:06:28

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.119	02:01
Breath Volume	1.679	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.119	02:04
Breath Volume	1.839	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MACEDO

DATE/TIME OF TEST: 04/26/2024 02:40:33

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.125	02:36
Breath Volume	1.816	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.122	02:38
Breath Volume	2.062	02:38
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: HIDALGO MALDONADO

DATE/TIME OF TEST: 04/26/2024 23:31:07

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.193	23:26
Breath Volume	1.496	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.191	23:29
Breath Volume	2.292	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/27/2024 01:07:17

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.185	01:01
Breath Volume	1.324	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.181	01:05
Breath Volume	1.578	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 04/27/2024 01:22:51

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.040	01:18
Breath Volume	3.167	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:20
Subject Sample # 2	0.040	01:21
Breath Volume	3.312	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MIRANDA VARGAS

DATE/TIME OF TEST: 04/27/2024 04:01:44

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.163	03:57
Breath Volume	2.906	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.168	04:00
Breath Volume	2.929	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: TRAYLOR

DATE/TIME OF TEST: 04/27/2024 04:40:13

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.147	04:35
Breath Volume	3.003	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.146	04:38
Breath Volume	2.699	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RENGIFO CASTRO

DATE/TIME OF TEST: 04/27/2024 04:52:50

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.194	04:48
Breath Volume	1.460	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.205	04:51
Breath Volume	1.144	04:51
Air Blank	0.000	04:51
Control Test	0.079	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: EMENIKE

DATE/TIME OF TEST: 04/27/2024 18:43:07

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.081	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.237	18:38
Breath Volume	2.191	18:38
Air Blank	0.000	18:39
Air Blank	0.000	18:40
Subject Sample # 2	0.239	18:41
Breath Volume	2.472	18:41
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: DUEFFER

DATE/TIME OF TEST: 04/27/2024 23:33:59

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.136	23:29
Breath Volume	1.453	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.148	23:32
Breath Volume	1.671	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAAR,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 04/28/2024 03:47:53

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.081	03:43
Breath Volume	3.085	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.080	03:46
Breath Volume	3.339	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 04/28/2024 04:08:46

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.130	04:04
Breath Volume	1.492	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.131	04:07
Breath Volume	1.828	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ, C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: GARCIA DA SILVA

DATE/TIME OF TEST: 04/28/2024 04:54:47

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.080	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.000	04:50
Breath Volume	3.406	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.000	04:53
Breath Volume	3.039	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CLARK,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BURGOS JR

DATE/TIME OF TEST: 04/28/2024 05:11:00

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.081	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.000	05:06
Breath Volume	1.480	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.000	05:09
Breath Volume	1.476	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GUATEMALA

DATE/TIME OF TEST: 04/29/2024 04:44:05

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.122	04:39
Breath Volume	2.250	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.119	04:42
Breath Volume	1.542	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: LEHIGH

DATE/TIME OF TEST: 04/29/2024 22:13:01

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.114	22:08
Breath Volume	3.246	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.116	22:11
Breath Volume	2.949	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/08/2024 14:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LOGSDON,DEREK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: CREPACK

DATE/TIME OF TEST: 04/04/2024 00:23:43

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1 0.311	00:17	Volume Not Met
Breath Volume	0.128	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 2	00:21	Subject Test Refused
Breath Volume	0.199	00:21	
Air Blank	0.000	00:22	
Control Test	0.076	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CELESTINO-HERNANDEZ

DATE/TIME OF TEST: 04/05/2024 22:38:16

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.143	22:28	
Breath Volume	1.554	22:28	
Air Blank	0.000	22:28	
Air Blank	0.000	22:30	
Subject Sample	# 2 0.114	22:33	Volume Not Met
Breath Volume	0.546	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 3 0.124	22:36	
Breath Volume	1.164	22:36	
Air Blank	0.000	22:37	
Control Test	0.077	22:37	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FANSLER,MICHEAL
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ABT

DATE/TIME OF TEST: 04/05/2024 23:27:20

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.161	23:22
Breath Volume	1.105	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.159	23:25
Breath Volume	1.433	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PRENTISS,HAYLEIGH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: BUTEAU

DATE/TIME OF TEST: 04/07/2024 18:49:38

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.082	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:48	Subject Test Refused
Breath Volume	0.000	18:48	
Air Blank	0.000	18:48	
Control Test	0.082	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:49	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CORAM

DATE/TIME OF TEST: 04/08/2024 22:47:38

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.082	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.000	22:43
Breath Volume	3.765	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:46
Breath Volume	2.988	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 04/13/2024 01:46:41

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.082	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.082	01:46	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SNYDER,M
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/17/2024 03:45:18

Diagnostics Check	OK	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:43	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,CARL,L
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PARKER,CARL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: MIMS

DATE/TIME OF TEST: 04/17/2024 18:14:01

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.082	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.203	18:09
Breath Volume	4.160	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:11
Subject Sample # 2	0.192	18:12
Breath Volume	2.847	18:12
Air Blank	0.000	18:13
Control Test	0.081	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:14

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: LOUGH

DATE/TIME OF TEST: 04/21/2024 05:04:38

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.223	05:00
Breath Volume	4.332	05:00
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.241	05:03
Breath Volume	3.617	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: KILPACRICK,A
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CAUDILL

DATE/TIME OF TEST: 04/24/2024 00:32:06

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.082	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.212	00:27
Breath Volume	1.890	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.194	00:30
Breath Volume	1.730	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LIPTAK, LAUREN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SPADE

DATE/TIME OF TEST: 04/24/2024 16:12:09

Diagnostics Check	OK	16:06
Air Blank	0.000	16:06
Control Test	0.082	16:06
Air Blank	0.000	16:07
Subject Sample # 1	0.000	16:07
Breath Volume	1.339	16:07
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.000	16:10
Breath Volume	1.277	16:10
Air Blank	0.000	16:11
Control Test	0.082	16:11
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HENNIGER

DATE/TIME OF TEST: 04/24/2024 22:18:38

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:07	
Control Test	0.082	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1 0.097	22:11	Volume Not Met
Breath Volume	0.910	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:13	
Subject Sample	# 2 0.103	22:17	Volume Not Met
Breath Volume	1.019	22:17	
Air Blank	0.000	22:17	
Control Test	0.080	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 04/25/2024 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: BRADLEY,WILLIAM
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 03/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RICKETT

DATE/TIME OF TEST: 04/12/2024 23:16:57

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.081	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:09	No Sample Provided
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Air Blank	0.000	23:12	
Subject Sample	# 2	0.013	23:12
Breath Volume	1.703	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 3	23:15	Interferent Detect
Breath Volume	2.484	23:15	
Air Blank	0.000	23:16	
Control Test	0.079	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 04/14/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: BRADLEY,WILLIAM
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 03/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: RICKETT

DATE/TIME OF TEST: 04/12/2024 23:26:37

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.079	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.012	23:22	
Breath Volume	2.343	23:22	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2	23:25	Interferent Detect
Breath Volume	2.480	23:25	
Air Blank	0.000	23:25	
Control Test	0.078	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 04/14/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SNIEZYK

DATE/TIME OF TEST: 04/01/2024 01:14:50

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.077	01:10
Breath Volume	2.238	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.083	01:13
Breath Volume	3.476	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 04/16/2024 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KIANA,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: HERNANDEZ-GERENA

DATE/TIME OF TEST: 04/01/2024 07:19:25

Diagnostics Check	OK	07:16	
Air Blank	0.000	07:17	
Control Test	0.079	07:17	
Air Blank	0.000	07:17	
Subject Sample	# 1	07:18	Subject Test Refused
Breath Volume	0.000	07:18	
Air Blank	0.000	07:18	
Control Test	0.080	07:18	
Air Blank	0.000	07:19	
Diagnostics Check	OK	07:19	

Data Download Date/Time: 04/16/2024 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 04/03/2024 02:43:10

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.191	02:38
Breath Volume	2.582	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.176	02:41
Breath Volume	2.371	02:41
Air Blank	0.000	02:42
Control Test	0.078	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/16/2024 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADOTTA,LIONEL,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALMEIDA,TIMOTHY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VEGA AGUILAR

DATE/TIME OF TEST: 04/07/2024 04:22:50

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.078	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.250	04:16
Breath Volume	1.824	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.240	04:20
Breath Volume	2.308	04:20
Air Blank	0.000	04:21
Control Test	0.077	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 04/16/2024 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: DELGADO,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 04/17/2024 02:56:46

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 05/14/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICCIARDI,AMANDA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALMEIDA,TIMOTHY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A2
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/20/2024 01:07:20

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 05/14/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SERRANO RONCAL

DATE/TIME OF TEST: 04/23/2024 02:10:02

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Subject Sample # 1	0.223	02:08	
Breath Volume	2.417	02:08	
Air Blank		02:09	Purge Fail
Air Blank	0.000	02:10	

Data Download Date/Time: 04/23/2024 21:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SERRANO RONCAL

DATE/TIME OF TEST: 04/23/2024 02:15:11

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.076	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	RFI Detect
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.078	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 04/23/2024 21:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SERRANO RONCAL

DATE/TIME OF TEST: 04/23/2024 02:24:01

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.079	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1 0.232	02:19	
Breath Volume	1.156	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 2	02:22	RFI Detect
Breath Volume	0.097	02:22	
Air Blank		02:23	Purge Fail
Air Blank	0.000	02:24	

Data Download Date/Time: 04/23/2024 21:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1506449
DRY GAS EXPIRY DATE: 06/10/2024
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 04/27/2024 05:04:59

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.154	05:00
Breath Volume	1.914	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.147	05:03
Breath Volume	2.312	05:03
Air Blank	0.000	05:04
Control Test	0.077	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 05/31/2024 01:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PLEASANT

DATE/TIME OF TEST: 04/01/2024 01:35:46

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.078	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.078	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 04/08/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENTLAND,ALEX,N
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: KING,SANJULO
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MAYNER

DATE/TIME OF TEST: 04/06/2024 02:23:41

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.079	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1 0.115	02:17	
Breath Volume	1.484	02:17	
Air Blank	0.000	02:18	
Air Blank	0.000	02:20	
Subject Sample	# 2 0.103	02:22	
Breath Volume	1.339	02:22	
Air Blank	0.000	02:22	
Control Test	0.073	02:23	Control Outside Tolerance
Air Blank	0.000	02:23	

Data Download Date/Time: 04/08/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TINFINA,LOGAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DAVIDSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 07:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: UILI JR

DATE/TIME OF TEST: 04/07/2024 09:46:43

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:35	
Control Test	0.080	09:36	
Air Blank	0.000	09:36	
Subject Sample	# 1 0.117	09:39	Volume Not Met
Breath Volume	0.843	09:39	
Air Blank	0.000	09:40	
Air Blank	0.000	09:42	
Subject Sample	# 2	09:45	No Sample Provided
Breath Volume	0.000	09:45	
Air Blank	0.000	09:45	
Control Test	0.077	09:46	
Air Blank	0.000	09:46	
Diagnostics Check	OK	09:46	

Data Download Date/Time: 04/08/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 04/05/2024 03:16:57

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.060	03:12
Breath Volume	2.156	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.062	03:15
Breath Volume	2.277	03:15
Air Blank	0.000	03:15
Control Test	0.081	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/27/2024 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402231696
DRY GAS EXPIRY DATE: 10/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SURUJBALLI

DATE/TIME OF TEST: 04/12/2024 02:40:07

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.091	02:35
Breath Volume	2.187	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.088	02:38
Breath Volume	2.070	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/27/2024 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SURUJBALLI

DATE/TIME OF TEST: 04/12/2024 02:06:17

Diagnostics Check	Fail	02:05	Voltage/Current Test
Air Blank	0.000	02:06	

Data Download Date/Time: 04/27/2024 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SURUJBALLI

DATE/TIME OF TEST: 04/12/2024 02:09:25

Diagnostics Check	Fail	02:08	Voltage/Current Test
Air Blank	0.000	02:09	

Data Download Date/Time: 04/27/2024 21:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FYFE

DATE/TIME OF TEST: 04/20/2024 03:07:06

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 04/27/2024 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/23/2024 23:48:34

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.084	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.225	23:47	
Breath Volume	1.683	23:47	
Air Blank		23:48	Purge Fail
Air Blank	0.000	23:48	

Data Download Date/Time: 04/27/2024 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/23/2024 23:57:38

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.231	23:52
Breath Volume	1.593	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.217	23:55
Breath Volume	1.535	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 04/27/2024 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: POLTONOWICZ

DATE/TIME OF TEST: 04/24/2024 04:14:06

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.076	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.145	04:09
Breath Volume	1.714	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.135	04:12
Breath Volume	1.589	04:12
Air Blank	0.000	04:13
Control Test	0.076	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:14

Data Download Date/Time: 04/27/2024 21:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 04/29/2024 23:56:13

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.084	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.087	23:51
Breath Volume	1.222	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.084	23:54
Breath Volume	2.160	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 05/31/2024 03:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY, KENNETH, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY, KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MANUCY

DATE/TIME OF TEST: 04/04/2024 01:15:10

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.082	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.166	01:10
Breath Volume	1.292	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.179	01:13
Breath Volume	1.363	01:13
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 04/22/2024 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CRUZADO

DATE/TIME OF TEST: 04/05/2024 02:20:50

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.083	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.149	02:14
Breath Volume	1.558	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.143	02:19
Breath Volume	2.292	02:19
Air Blank	0.000	02:19
Control Test	0.083	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 04/22/2024 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: TEWARI

DATE/TIME OF TEST: 04/14/2024 01:49:20

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.083	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.129	01:44
Breath Volume	3.195	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.126	01:47
Breath Volume	2.937	01:47
Air Blank	0.000	01:48
Control Test	0.082	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 04/22/2024 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 04/19/2024 02:17:10

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.083	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.090	02:12
Breath Volume	1.207	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.094	02:15
Breath Volume	1.453	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/22/2024 19:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SAEY

DATE/TIME OF TEST: 04/23/2024 02:19:10

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.083	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.179	02:14
Breath Volume	1.308	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.162	02:17
Breath Volume	1.113	02:17
Air Blank	0.000	02:18
Control Test	0.083	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 05/08/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SIWICA

DATE/TIME OF TEST: 04/26/2024 23:21:37

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.082	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.097	23:17
Breath Volume	2.960	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.097	23:20
Breath Volume	3.855	23:20
Air Blank	0.000	23:20
Control Test	0.082	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 05/08/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP TROOP C
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24229080A1
DRY GAS EXPIRY DATE: 02/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/27/2024 02:44:37

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.082	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.070	02:40
Breath Volume	2.039	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.073	02:43
Breath Volume	2.507	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/08/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 03/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/14/2024 00:50:07

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.162	00:45
Breath Volume	2.285	00:45
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.164	00:48
Breath Volume	1.160	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 05/01/2024 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: NAFE

DATE/TIME OF TEST: 04/09/2024 20:47:59

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.072	20:43
Breath Volume	1.246	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:45
Subject Sample # 2	0.064	20:46
Breath Volume	1.226	20:46
Air Blank	0.000	20:47
Control Test	0.076	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 05/13/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SHELBY

DATE/TIME OF TEST: 04/14/2024 00:06:08

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.164	00:01
Breath Volume	1.410	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.157	00:04
Breath Volume	1.441	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 05/13/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: MERINO MARTINEZ

DATE/TIME OF TEST: 04/24/2024 18:40:51

Diagnostics Check	OK	18:34
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:35
Subject Sample # 1	0.143	18:36
Breath Volume	2.105	18:36
Air Blank	0.000	18:37
Air Blank	0.000	18:38
Subject Sample # 2	0.135	18:39
Breath Volume	1.503	18:39
Air Blank	0.000	18:39
Control Test	0.080	18:40
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 05/13/2024 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ARROYO QUIRINDONGO

DATE/TIME OF TEST: 04/30/2024 02:38:37

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.135	02:30	Volume Not Met
Breath Volume	0.210	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.141	02:33	
Breath Volume	1.375	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 3 0.146	02:36	
Breath Volume	2.507	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 05/13/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PEREZ GREGORIO

DATE/TIME OF TEST: 04/30/2024 22:56:58

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.080	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1 0.184	22:52	
Breath Volume	1.855	22:52	
Air Blank	0.000	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 2 0.181	22:55	
Breath Volume	1.660	22:55	
Air Blank	0.000	22:56	
Control Test	0.077	22:56	
Air Blank		22:56	RFI Detect
Air Blank		22:56	RFI Detect

Data Download Date/Time: 05/13/2024 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 04/14/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PEREZ GREGORIO

DATE/TIME OF TEST: 04/30/2024 23:05:34

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.184	23:00
Breath Volume	1.347	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.186	23:03
Breath Volume	1.757	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 05/13/2024 09:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BARTER

DATE/TIME OF TEST: 04/02/2024 01:49:07

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:48	
Control Test	0.081	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 04/08/2024 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 04/06/2024 03:16:23

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.079	03:15	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 04/08/2024 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/07/2024 02:41:16

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:39	Subject Test Refused
Breath Volume	0.515	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 04/08/2024 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CRUZ COLON

DATE/TIME OF TEST: 04/12/2024 02:08:52

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2	0.099	02:03
Breath Volume	2.183	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 3	0.103	02:07
Breath Volume	2.269	02:07	
Air Blank	0.000	02:07	
Control Test	0.079	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 05/13/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 04/16/2024 04:03:08

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.080	04:00	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 05/13/2024 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CLERVIUS

DATE/TIME OF TEST: 04/20/2024 00:47:39

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.055	00:43
Breath Volume	2.222	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.057	00:46
Breath Volume	2.144	00:46
Air Blank	0.000	00:46
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 05/13/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GOHEEN

DATE/TIME OF TEST: 04/20/2024 23:16:19

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 05/13/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: KEEN,CHARLES
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: NORTON JR

DATE/TIME OF TEST: 04/04/2024 00:02:52

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.083	23:57
Breath Volume	2.339	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.083	00:01
Breath Volume	1.562	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 04/19/2024 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: MCCAUSLIN

DATE/TIME OF TEST: 04/05/2024 01:00:39

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.226	00:55
Breath Volume	1.300	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.234	00:58
Breath Volume	2.226	00:58
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 04/19/2024 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 04/18/2024 03:00:12

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.202	02:54
Breath Volume	1.722	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.189	02:58
Breath Volume	1.742	02:58
Air Blank	0.000	02:59
Control Test	0.076	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/19/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: INACIO

DATE/TIME OF TEST: 04/18/2024 21:39:09

Diagnostics Check	OK	21:32
Air Blank	0.000	21:33
Control Test	0.078	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.202	21:34
Breath Volume	1.828	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.200	21:37
Breath Volume	1.867	21:37
Air Blank	0.000	21:38
Control Test	0.077	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 04/19/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 10/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 04/19/2024 03:31:19

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.077	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.106	03:26
Breath Volume	1.394	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.106	03:29
Breath Volume	1.769	03:29
Air Blank	0.000	03:30
Control Test	0.077	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/19/2024 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: YOUNKMAN

DATE/TIME OF TEST: 04/20/2024 03:30:15

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.148	03:24
Breath Volume	1.421	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.146	03:28
Breath Volume	1.171	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/23/2024 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 04/21/2024 01:17:57

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	RFI Detect
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 04/23/2024 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 04/21/2024 01:26:11

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.160	01:21
Breath Volume	1.464	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.148	01:24
Breath Volume	1.480	01:24
Air Blank	0.000	01:25
Control Test	0.078	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 04/23/2024 17:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HOLCY

DATE/TIME OF TEST: 04/30/2024 21:09:24

Diagnostics Check	OK	21:03
Air Blank	0.000	21:03
Control Test	0.079	21:04
Air Blank	0.000	21:04
Subject Sample # 1	0.053	21:05
Breath Volume	1.480	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.050	21:07
Breath Volume	1.785	21:07
Air Blank	0.000	21:08
Control Test	0.078	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 05/10/2024 06:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BROWN,JEREMIAH
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: LUCIANO

DATE/TIME OF TEST: 04/03/2024 07:02:47

Diagnostics Check	OK	06:56
Air Blank	0.000	06:56
Control Test	0.079	06:57
Air Blank	0.000	06:57
Subject Sample # 1	0.161	06:58
Breath Volume	1.363	06:58
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.149	07:01
Breath Volume	1.800	07:01
Air Blank	0.000	07:01
Control Test	0.079	07:02
Air Blank	0.000	07:02
Diagnostics Check	OK	07:02

Data Download Date/Time: 04/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REDFEARN,JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ZUCKERMAN

DATE/TIME OF TEST: 04/05/2024 03:03:32

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.151	02:58
Breath Volume	1.988	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.148	03:01
Breath Volume	1.949	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STEPHENS,BRANDEN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 04/06/2024 12:50:02

Diagnostics Check	OK	12:42
Air Blank	0.000	12:42
Control Test	0.079	12:43
Air Blank	0.000	12:43
Subject Sample # 1	0.247	12:45
Breath Volume	2.250	12:45
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.244	12:48
Breath Volume	2.367	12:48
Air Blank	0.000	12:49
Control Test	0.078	12:49
Air Blank	0.000	12:49
Diagnostics Check	OK	12:50

Data Download Date/Time: 04/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CORAS

DATE/TIME OF TEST: 04/07/2024 01:24:42

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.183	01:20
Breath Volume	3.355	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.176	01:23
Breath Volume	3.535	01:23
Air Blank	0.000	01:23
Control Test	0.078	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 04/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 04/18/2024 02:26:26

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.082	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.164	02:21
Breath Volume	2.593	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.171	02:24
Breath Volume	2.156	02:24
Air Blank	0.000	02:25
Control Test	0.082	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 04/30/2024 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 04/20/2024 00:43:53

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.148	00:39
Breath Volume	2.789	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.151	00:42
Breath Volume	2.953	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/30/2024 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: OSTMAN

DATE/TIME OF TEST: 04/26/2024 23:22:39

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.185	23:18
Breath Volume	1.296	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.175	23:20
Breath Volume	1.144	23:20
Air Blank	0.000	23:21
Control Test	0.082	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 04/30/2024 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,P
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: BILOTTI

DATE/TIME OF TEST: 04/05/2024 17:45:08

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.080	17:39	
Air Blank	0.000	17:39	
Subject Sample # 1	0.077	17:40	
Breath Volume	2.234	17:40	
Air Blank	0.000	17:40	
Air Blank	0.000	17:42	
Subject Sample # 2	0.080	17:43	
Breath Volume	1.808	17:43	
Air Blank	0.000	17:43	
Control Test	0.079	17:43	
Air Blank	0.000	17:44	
Diagnostics Check	Fail	17:44	Internal Printer Test
Air Blank	0.000	17:45	

Data Download Date/Time: 04/30/2024 22:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,P
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: BILOTTI

DATE/TIME OF TEST: 04/05/2024 17:56:04

Diagnostics Check	OK	17:49
Air Blank	0.000	17:49
Control Test	0.079	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.078	17:51
Breath Volume	2.292	17:51
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.080	17:54
Breath Volume	2.691	17:54
Air Blank	0.000	17:55
Control Test	0.079	17:55
Air Blank	0.000	17:55
Diagnostics Check	OK	17:56

Data Download Date/Time: 04/30/2024 22:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: CHECA MACHADO

DATE/TIME OF TEST: 04/05/2024 23:19:56

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.181	23:15
Breath Volume	1.996	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.183	23:18
Breath Volume	1.945	23:18
Air Blank	0.000	23:19
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/10/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FAULKNER

DATE/TIME OF TEST: 04/06/2024 01:10:35

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.090	01:06
Breath Volume	2.222	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.090	01:09
Breath Volume	1.664	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/10/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MCDOUGLE

DATE/TIME OF TEST: 04/06/2024 01:36:08

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:31
Breath Volume	1.679	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:34
Breath Volume	1.914	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 04/10/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007006

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOOLLEY,KYLE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 04/06/2024 03:04:18

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.180	02:58
Breath Volume	1.132	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.183	03:02
Breath Volume	1.363	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/10/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VAZQUEZ,JUSTIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: DIAZ MORALES

DATE/TIME OF TEST: 04/12/2024 22:33:15

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.211	22:28
Breath Volume	1.203	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.218	22:31
Breath Volume	1.621	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 05/08/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SPENCER JR.

DATE/TIME OF TEST: 04/13/2024 01:09:45

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.158	01:05
Breath Volume	1.218	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.159	01:08
Breath Volume	1.828	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/08/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DE SOUZA

DATE/TIME OF TEST: 04/14/2024 21:02:43

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.199	20:58
Breath Volume	1.996	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.196	21:01
Breath Volume	1.832	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DOMINGUEZ,ALAIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 04/15/2024 03:31:03

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.197	03:26
Breath Volume	1.351	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.187	03:29
Breath Volume	1.171	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CORTES COLON

DATE/TIME OF TEST: 04/17/2024 20:35:44

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.082	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.000	20:31
Breath Volume	1.238	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.000	20:34
Breath Volume	2.269	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:35
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMPARO,AURIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 04/26/2024 19:14:56

Diagnostics Check	OK	19:08
Air Blank	0.000	19:08
Control Test	0.082	19:09
Air Blank	0.000	19:09
Subject Sample # 1	0.237	19:10
Breath Volume	2.570	19:10
Air Blank	0.000	19:11
Air Blank	0.000	19:12
Subject Sample # 2	0.243	19:13
Breath Volume	2.500	19:13
Air Blank	0.000	19:13
Control Test	0.081	19:14
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 04/26/2024 23:58:35

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.304	23:53
Breath Volume	2.679	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.291	23:56
Breath Volume	2.195	23:56
Air Blank	0.000	23:57
Control Test	0.081	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARNECO,TAYLOR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 04/30/2024 01:17:41

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.146	01:13
Breath Volume	1.980	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.148	01:16
Breath Volume	1.550	01:16
Air Blank	0.000	01:16
Control Test	0.081	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 05/08/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/06/2024 02:27:11

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.175	02:22
Breath Volume	1.835	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.175	02:25
Breath Volume	1.894	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/30/2024 23:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007009

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 03/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/26/2024 00:58:22

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.079	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.201	00:53
Breath Volume	2.550	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.188	00:56
Breath Volume	1.500	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 04/30/2024 23:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: PIETRANGELO

DATE/TIME OF TEST: 04/02/2024 02:00:12

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.191	01:55
Breath Volume	3.660	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.189	01:58
Breath Volume	3.445	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIERRE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/03/2024 01:51:28

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.078	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BATZ YUS

DATE/TIME OF TEST: 04/06/2024 02:58:04

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.216	02:53
Breath Volume	1.843	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.223	02:56
Breath Volume	2.601	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:58

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/06/2024 03:58:59

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.089	03:54
Breath Volume	1.433	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.087	03:57
Breath Volume	1.726	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: SIRMONS

DATE/TIME OF TEST: 04/06/2024 05:55:36

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.081	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.058	05:51
Breath Volume	2.812	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.058	05:54
Breath Volume	2.789	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 04/07/2024 00:20:00

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.064	00:15
Breath Volume	1.242	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.070	00:18
Breath Volume	1.375	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RENTSCHLER

DATE/TIME OF TEST: 04/07/2024 02:21:33

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.262	02:16
Breath Volume	1.117	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.269	02:19
Breath Volume	2.210	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/07/2024 06:14:30

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:00	
Control Test	0.080	06:00	
Air Blank	0.000	06:01	
Subject Sample	# 1 0.164	06:04	Volume Not Met
Breath Volume	0.394	06:04	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample	# 2 0.202	06:09	
Breath Volume	2.066	06:09	
Air Blank	0.000	06:10	
Air Blank	0.000	06:11	
Subject Sample	# 3 0.206	06:12	
Breath Volume	1.417	06:12	
Air Blank	0.000	06:13	
Control Test	0.077	06:13	
Air Blank	0.000	06:14	
Diagnostics Check	OK	06:14	

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HERNANDEZ RAMIREZ

DATE/TIME OF TEST: 04/07/2024 21:03:12

Diagnostics Check	OK	20:55
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:56
Subject Sample # 1	0.312	20:58
Breath Volume	2.113	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.303	21:01
Breath Volume	2.101	21:01
Air Blank	0.000	21:02
Control Test	0.078	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WAKEFIELD

DATE/TIME OF TEST: 04/07/2024 21:57:21

Diagnostics Check	OK	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.192	21:52
Breath Volume	2.445	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.196	21:55
Breath Volume	2.441	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 04/07/2024 22:55:25

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.205	22:50
Breath Volume	2.191	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.203	22:53
Breath Volume	1.949	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 04/08/2024 00:05:31

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.189	00:00
Breath Volume	3.074	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.195	00:03
Breath Volume	3.535	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PORUBCAN

DATE/TIME OF TEST: 04/09/2024 00:15:48

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.234	00:10
Breath Volume	1.167	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.247	00:14
Breath Volume	1.550	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 04/09/2024 03:29:03

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.053	03:24
Breath Volume	1.695	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:26
Subject Sample # 2	0.055	03:27
Breath Volume	1.785	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 04/10/2024 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: AG22203
BREATH TEST OPERATOR: KLEIS,KEVIN,	DRY GAS EXPIRY DATE: 08/10/2024
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: HOPPE,AUSTIN	ARREST TIME: 19:18
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 04/23/2024	OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: VEIGA	DATE/TIME OF TEST: 04/28/2024	20:04:32
----------------------------	-------------------------------	----------

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.080	19:58
Air Blank	0.000	19:59
Subject Sample # 1	0.192	19:59
Breath Volume	2.332	19:59
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.194	20:02
Breath Volume	3.285	20:02
Air Blank	0.000	20:03
Control Test	0.079	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 05/24/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007165

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 1380119
BREATH TEST OPERATOR: BEGERA,JERRY,J	DRY GAS EXPIRY DATE: 08/10/2024
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)	VIOLATION CODE: BUI
ARRESTING OFFICER: BEGERA,JERRY	ARREST TIME: 20:42
ARRESTING OFFICER AGENCY: FFWCC (West Palm)	GENDER: Male
LAST AGENCY INSPECTION: 04/23/2024	OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: CUEVAS GARCIA	DATE/TIME OF TEST: 04/28/2024	21:18:22
------------------------------------	-------------------------------	----------

Diagnostics Check	OK	21:15	
Air Blank	0.000	21:15	
Control Test	0.080	21:16	
Air Blank	0.000	21:16	
Subject Sample	# 1	21:16	Subject Test Refused
Breath Volume	0.000	21:16	
Air Blank	0.000	21:17	
Control Test	0.080	21:17	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 05/24/2024 12:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPIRITO,GEORGE,W

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: WILLIS,TYLER

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402498227

DRY GAS EXPIRY DATE: 07/22/2025

VIOLATION CODE: DUI

ARREST TIME: 17:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 17:11

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/28/2024 18:14:20

Diagnostics Check	OK	18:02	
Air Blank	0.000	18:03	
Control Test	0.080	18:03	
Air Blank	0.000	18:04	
Subject Sample	# 1	18:07	No Sample Provided
Breath Volume	0.000	18:07	
Air Blank	0.000	18:07	
Air Blank	0.000	18:09	
Subject Sample	# 2	18:12	No Sample Provided
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Control Test	0.080	18:13	
Air Blank	0.000	18:14	
Diagnostics Check	OK	18:14	

Data Download Date/Time: 05/24/2024 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPIRITO,GEORGE,W

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: EUSTICE,JOSH

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402498227

DRY GAS EXPIRY DATE: 07/22/2025

VIOLATION CODE: BUI

ARREST TIME: 19:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: MALAGON

DATE/TIME OF TEST: 04/28/2024 19:56:15

Diagnostics Check	OK	19:49
Air Blank	0.000	19:50
Control Test	0.080	19:50
Air Blank	0.000	19:51
Subject Sample # 1	0.056	19:51
Breath Volume	3.199	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.059	19:54
Breath Volume	3.035	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 05/24/2024 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPIRITO,GEORGE,W

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: KLEIN,JUSTIN

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402498227

DRY GAS EXPIRY DATE: 07/22/2025

VIOLATION CODE: BUI

ARREST TIME: 19:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: BILLITIER

DATE/TIME OF TEST: 04/28/2024 20:30:13

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.080	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.149	20:25
Breath Volume	1.246	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.146	20:28
Breath Volume	1.863	20:28
Air Blank	0.000	20:29
Control Test	0.080	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 05/24/2024 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007166

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPIRITO,GEORGE,W

BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)

ARRESTING OFFICER: RIVARD,MCKENZIE

ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)

LAST AGENCY INSPECTION: 04/23/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402498227

DRY GAS EXPIRY DATE: 07/22/2025

VIOLATION CODE: BUI

ARREST TIME: 20:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SOUSA

DATE/TIME OF TEST: 04/28/2024 21:24:38

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.085	21:19
Breath Volume	3.562	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.092	21:23
Breath Volume	2.554	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:24
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 05/24/2024 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: BOLTZ

DATE/TIME OF TEST: 04/01/2024 00:56:36

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.162	00:51
Breath Volume	1.257	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.171	00:54
Breath Volume	2.425	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/03/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TUREK,TOMASZ
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: KOVAL

DATE/TIME OF TEST: 04/01/2024 20:10:02

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.080	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.148	20:05
Breath Volume	3.328	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.154	20:08
Breath Volume	2.738	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:10

Data Download Date/Time: 04/03/2024 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODDY,AARON,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDY,AARON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 04/03/2024 14:25:47

Diagnostics Check	OK	14:20	
Air Blank	0.000	14:20	
Control Test	0.080	14:21	
Air Blank	0.000	14:21	
Subject Sample	# 1	14:24	Subject Test Refused
Breath Volume	0.000	14:24	
Air Blank	0.000	14:24	
Control Test	0.081	14:25	
Air Blank	0.000	14:25	
Diagnostics Check	OK	14:25	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LADD,J
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DAYAO

DATE/TIME OF TEST: 04/04/2024 23:50:26

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:47	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.081	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DOWLING

DATE/TIME OF TEST: 04/05/2024 22:39:21

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.079	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:38	
Control Test	0.080	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHAM,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SIMPSON,JAMES
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 04/07/2024 01:09:35

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.249	01:03
Breath Volume	1.335	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.250	01:07
Breath Volume	1.707	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 04/07/2024 19:28:37

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.080	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.232	19:22
Breath Volume	1.339	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.224	19:26
Breath Volume	1.871	19:26
Air Blank	0.000	19:27
Control Test	0.077	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JOHN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALLEN,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 04/14/2024 07:31:28

Diagnostics Check	OK	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:25
Subject Sample # 1	0.208	07:26
Breath Volume	2.843	07:26
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.225	07:29
Breath Volume	2.597	07:29
Air Blank	0.000	07:30
Control Test	0.078	07:30
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: LACOMBE

DATE/TIME OF TEST: 04/14/2024 21:18:23

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.144	21:13
Breath Volume	1.488	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:16
Subject Sample # 2	0.142	21:16
Breath Volume	1.910	21:16
Air Blank	0.000	21:17
Control Test	0.079	21:17
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ALLENSWORTH

DATE/TIME OF TEST: 04/15/2024 19:51:02

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.080	19:48	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.081	19:50	
Air Blank	0.000	19:50	
Diagnostics Check	OK	19:51	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: SANTIAGO CRUZ

DATE/TIME OF TEST: 04/17/2024 04:40:33

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.081	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.195	04:30	
Breath Volume	2.027	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2 0.162	04:36	Volume Not Met
Breath Volume	0.488	04:36	
Air Blank	0.000	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 3	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.075	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: LOMBARDI

DATE/TIME OF TEST: 04/17/2024 14:22:55

Diagnostics Check	OK	14:16
Air Blank	0.000	14:16
Control Test	0.081	14:16
Air Blank	0.000	14:17
Subject Sample # 1	0.369	14:17
Breath Volume	1.421	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.353	14:21
Breath Volume	1.855	14:21
Air Blank	0.000	14:21
Control Test	0.078	14:22
Air Blank	0.000	14:22
Diagnostics Check	OK	14:22

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHLEBANOWSKI, RYAN, P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: LAURA HURTADO

DATE/TIME OF TEST: 04/18/2024 04:37:29

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.182	04:29
Breath Volume	1.410	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.153	04:32
Breath Volume	2.429	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 3	0.181	04:35
Breath Volume	1.382	04:35
Air Blank	0.000	04:36
Control Test	0.077	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BURNS,ALYSON
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 04/19/2024 03:20:43

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Control Test	0.080	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: AJANEL MAS

DATE/TIME OF TEST: 04/20/2024 05:25:45

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:20	
Control Test	0.080	05:20	
Air Blank	0.000	05:21	
Subject Sample # 1	0.227	05:22	
Breath Volume	1.250	05:22	
Air Blank	0.000	05:22	
Air Blank		05:24	RFI Detect
Air Blank	0.000	05:24	
Control Test	0.080	05:25	
Air Blank	0.000	05:25	
Diagnostics Check	OK	05:25	

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERNATOVA,ANDREA,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BERNATOVA,ANDREA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: AJANEL MAS

DATE/TIME OF TEST: 04/20/2024 05:35:02

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.203	05:30
Breath Volume	1.679	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.201	05:33
Breath Volume	1.488	05:33
Air Blank	0.000	05:34
Control Test	0.076	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:35

Data Download Date/Time: 05/23/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: ALLAN

DATE/TIME OF TEST: 04/22/2024 23:56:37

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.081	23:54	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:55	Subject Test Refused
Breath Volume	0.000	23:55	
Air Blank	0.000	23:55	
Control Test	0.080	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:07

SUBJECT'S LAST NAME: CARBAUGH

DATE/TIME OF TEST: 04/23/2024 14:38:02

Diagnostics Check	OK	14:34	
Air Blank	0.000	14:35	
Control Test	0.079	14:35	
Air Blank	0.000	14:36	
Subject Sample	# 1	14:36	Subject Test Refused
Breath Volume	0.000	14:36	
Air Blank	0.000	14:37	
Control Test	0.080	14:37	
Air Blank	0.000	14:37	
Diagnostics Check	OK	14:38	

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BRABLC,M
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 04/25/2024 06:55:24

Diagnostics Check	OK	06:49
Air Blank	0.000	06:49
Control Test	0.081	06:49
Air Blank	0.000	06:50
Subject Sample # 1	0.198	06:50
Breath Volume	1.441	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.189	06:53
Breath Volume	1.511	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,ELAM,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: THORNBRUGH,E
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/26/2024 19:44:11

Diagnostics Check	OK	19:35
Air Blank	0.000	19:35
Control Test	0.081	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.097	19:38
Breath Volume	1.679	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.096	19:42
Breath Volume	1.812	19:42
Air Blank	0.000	19:43
Control Test	0.078	19:43
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: PEREZ,RANDY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 04/27/2024 16:59:38

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.079	16:56	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:58	Subject Test Refused
Breath Volume	0.000	16:58	
Air Blank	0.000	16:58	
Control Test	0.080	16:59	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 04/28/2024 07:13:19

Diagnostics Check	OK	07:12	
Air Blank		07:12	RFI Detect
Air Blank	0.000	07:13	

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALAK,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 04/28/2024 07:22:09

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.080	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.141	07:17
Breath Volume	3.328	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:19
Subject Sample # 2	0.144	07:20
Breath Volume	2.683	07:20
Air Blank	0.000	07:21
Control Test	0.080	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSE,JOSHUA,C
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HESSE,JOSHUA
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: CLEMENTS

DATE/TIME OF TEST: 04/28/2024 23:16:35

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1	23:15	Subject Test Refused
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Control Test	0.080	23:15	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BRABLC,M
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: SUGAMELI

DATE/TIME OF TEST: 04/29/2024 06:24:48

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:18
Subject Sample # 1	0.133	06:20
Breath Volume	1.781	06:20
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.122	06:23
Breath Volume	1.132	06:23
Air Blank	0.000	06:23
Control Test	0.079	06:24
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 05/23/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 04/05/2024 03:42:53

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,MEGHAN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DULLY,STEPHEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: KONYK

DATE/TIME OF TEST: 04/07/2024 00:13:18

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.177	00:08
Breath Volume	1.617	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.166	00:11
Breath Volume	3.292	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODDA, TIFFANY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RATHBUN

DATE/TIME OF TEST: 04/09/2024 21:20:02

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.080	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:18	Subject Test Refused
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.080	21:19	
Air Blank	0.000	21:19	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: CABALLER VARGAS

DATE/TIME OF TEST: 04/14/2024 04:41:31

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CHLEBANOWSKI,RYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 04/17/2024 23:34:37

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.080	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,THOMAS,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ARDREY-GIGUERE,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: TRENT

DATE/TIME OF TEST: 04/25/2024 20:29:36

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:27	
Control Test	0.080	20:27	
Air Blank	0.000	20:28	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:28	
Control Test	0.080	20:28	
Air Blank	0.000	20:29	
Diagnostics Check	OK	20:29	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: DEASY

DATE/TIME OF TEST: 04/26/2024 21:28:11

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.081	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.128	21:23
Breath Volume	3.277	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.127	21:26
Breath Volume	3.156	21:26
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/28/2024 02:18:51

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:17	
Control Test	0.080	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 05/22/2024 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/26/2024 16:47:13

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:42	
Control Test	0.079	16:42	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:45	RFI Detect
Breath Volume	0.800	16:45	
Air Blank	0.000	16:46	
Control Test	0.080	16:46	
Air Blank	0.000	16:47	
Diagnostics Check	OK	16:47	

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/26/2024 16:49:46

Diagnostics Check	Fail	16:49	Voltage/Current Test
Air Blank	0.000	16:49	

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/26/2024 17:00:03

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.078	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.191	16:53
Breath Volume	1.554	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.194	16:58
Breath Volume	2.867	16:58
Air Blank	0.000	16:59
Control Test	0.077	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	17:00

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVILA,VICTOR,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LISS,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TULLONGE

DATE/TIME OF TEST: 04/28/2024 03:41:51

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.169	03:36
Breath Volume	1.476	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.163	03:39
Breath Volume	1.398	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECONTE,JAMES,N
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: AULTMAN,ANDREW
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: BUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: LECONTE

DATE/TIME OF TEST: 04/28/2024 20:20:54

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:13	
Control Test	0.079	20:13	
Air Blank	0.000	20:14	
Subject Sample	# 1 0.101	20:16	
Breath Volume	2.457	20:16	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2 0.100	20:19	
Breath Volume	1.718	20:19	
Air Blank	0.000	20:19	
Control Test	0.076	20:20	
Air Blank		20:20	RFI Detect
Air Blank	0.000	20:20	

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 04/30/2024 03:16:38

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Subject Sample # 1	0.000	03:13	
Breath Volume	5.722	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample # 2	0.000	03:16	
Breath Volume	6.300	03:16	
Air Blank		03:16	RFI Detect
Air Blank	0.000	03:16	

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WHITTINGTON,CAROLINE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 04/30/2024 03:24:50

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.077	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.000	03:20
Breath Volume	4.531	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.000	03:23
Breath Volume	4.382	03:23
Air Blank	0.000	03:23
Control Test	0.077	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 05/29/2024 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/07/2024 23:09:12

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.083	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.225	23:04
Breath Volume	1.601	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.213	23:07
Breath Volume	1.347	23:07
Air Blank	0.000	23:08
Control Test	0.084	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAY,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MAY,WILLIAM
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MARCIAL BENITEZ

DATE/TIME OF TEST: 04/15/2024 19:43:37

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.139	19:39
Breath Volume	3.554	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.133	19:41
Breath Volume	2.484	19:41
Air Blank	0.000	19:42
Control Test	0.085	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/15/2024 20:29:11

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.082	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:24
Breath Volume	2.058	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:27
Breath Volume	2.800	20:27
Air Blank	0.000	20:28
Control Test	0.085	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCALIA, ADAM, C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KAPPENMACHER, THOMAS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/17/2024 19:26:55

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.082	19:21
Air Blank	0.000	19:21
Subject Sample # 1	0.170	19:22
Breath Volume	1.546	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:24
Subject Sample # 2	0.176	19:25
Breath Volume	1.628	19:25
Air Blank	0.000	19:25
Control Test	0.082	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:26

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VANCE,SPENCER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/19/2024 20:25:15

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:15	
Control Test	0.081	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 1	20:19	No Sample Provided
Breath Volume	0.000	20:19	
Air Blank	0.000	20:20	
Air Blank	0.000	20:21	
Subject Sample	# 2	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:24	
Control Test	0.082	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CARNAHAN,TREVOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 04/20/2024 21:10:18

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.163	21:05
Breath Volume	2.765	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.153	21:08
Breath Volume	2.921	21:08
Air Blank	0.000	21:09
Control Test	0.085	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: KENNEL

DATE/TIME OF TEST: 04/20/2024 22:25:57

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:23	
Control Test	0.082	22:23	
Air Blank	0.000	22:24	
Subject Sample	# 1	22:24	Subject Test Refused
Breath Volume	0.000	22:24	
Air Blank	0.000	22:25	
Control Test	0.082	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: CAN

DATE/TIME OF TEST: 04/21/2024 19:35:22

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:29
Subject Sample # 1	0.248	19:30
Breath Volume	3.320	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.257	19:33
Breath Volume	2.554	19:33
Air Blank	0.000	19:34
Control Test	0.080	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: SCHROEDER

DATE/TIME OF TEST: 04/25/2024 23:18:50

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.083	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 1 0.000	23:14	
Breath Volume	3.730	23:14	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample	# 2 0.000	23:17	
Breath Volume	3.738	23:17	
Air Blank	0.000	23:17	
Control Test	0.086	23:18	Control Outside Tolerance
Air Blank	0.000	23:18	

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 04/29/2024 23:48:01

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.082	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.055	23:43
Breath Volume	2.945	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.051	23:46
Breath Volume	1.589	23:46
Air Blank	0.000	23:47
Control Test	0.083	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

Data Download Date/Time: 04/30/2024 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKEY,A,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MINOSO,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: LOYOLA

DATE/TIME OF TEST: 04/26/2024 03:51:38

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:51	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 05/13/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACKEY,A,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MINOSO,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: LOYOLA

DATE/TIME OF TEST: 04/26/2024 04:08:08

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.093	04:03
Breath Volume	1.406	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.089	04:06
Breath Volume	1.230	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/13/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: LEDBETTER,CALEB

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 222152

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 19:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: NEELY

DATE/TIME OF TEST: 04/10/2024 20:43:20

Diagnostics Check	OK	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.232	20:38
Breath Volume	4.437	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.224	20:41
Breath Volume	3.496	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FLINN,AARON,D

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: FLINN,AARON

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 07:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: HILLBURN

DATE/TIME OF TEST: 04/13/2024 08:37:02

Diagnostics Check	OK	08:29
Air Blank	0.000	08:29
Control Test	0.080	08:30
Air Blank	0.000	08:30
Subject Sample # 1	0.163	08:32
Breath Volume	1.527	08:32
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.156	08:35
Breath Volume	2.589	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:36
Diagnostics Check	OK	08:37

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: WERNER,DANIEL

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 20:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: TREES

DATE/TIME OF TEST: 04/15/2024 22:01:46

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.081	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.000	21:56
Breath Volume	2.984	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	22:00
Breath Volume	2.839	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 01:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 04/17/2024 02:26:55

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.123	02:22
Breath Volume	4.394	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.123	02:25
Breath Volume	4.203	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 04/20/2024 03:59:36

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:58	RFI Detect
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 04/20/2024 04:08:53

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.152	04:03
Breath Volume	2.636	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.148	04:07
Breath Volume	2.886	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BECKER,JONATHON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221783
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LOWREY

DATE/TIME OF TEST: 04/22/2024 04:21:15

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.142	04:16
Breath Volume	1.906	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.135	04:19
Breath Volume	1.289	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: MAYFIELD,MATTHEW

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 03:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: PALERMO

DATE/TIME OF TEST: 04/28/2024 05:01:50

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.149	04:57
Breath Volume	1.261	04:57
Air Blank	0.000	04:58
Air Blank	0.000	04:59
Subject Sample # 2	0.145	05:00
Breath Volume	1.445	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASWELL,AUSTIN,J

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: MAYFIELD,MATTHEW

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 02:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 04/29/2024 04:06:32

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.000	04:01
Breath Volume	1.558	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	1.222	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KENT,ROBERT,T

BREATH TEST OPERATOR AGENCY: Pensacola PD

ARRESTING OFFICER: KENT,ROBERT

ARRESTING OFFICER AGENCY: Pensacola PD

LAST AGENCY INSPECTION: 03/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 221783

DRY GAS EXPIRY DATE: 12/22/2025

VIOLATION CODE: DUI

ARREST TIME: 17:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/29/2024 18:12:37

Diagnostics Check	OK	18:06
Air Blank	0.000	18:06
Control Test	0.081	18:06
Air Blank	0.000	18:07
Subject Sample # 1	0.109	18:08
Breath Volume	3.121	18:08
Air Blank	0.000	18:08
Air Blank	0.000	18:10
Subject Sample # 2	0.099	18:11
Breath Volume	2.000	18:11
Air Blank	0.000	18:11
Control Test	0.081	18:12
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

Data Download Date/Time: 04/30/2024 15:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANDILE,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: KAYMORE

DATE/TIME OF TEST: 04/01/2024 01:51:44

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.147	01:47
Breath Volume	1.898	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.146	01:50
Breath Volume	3.398	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 04/02/2024 23:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: ALVES,MATHEUS,T	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: GREENE,LETOUSHA	ARREST TIME: 17:42
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: KINGDESKI DATE/TIME OF TEST: 04/05/2024 18:55:02

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.081	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.000	18:50
Breath Volume	3.250	18:50
Air Blank	0.000	18:51
Air Blank	0.000	18:53
Subject Sample # 2	0.000	18:53
Breath Volume	3.019	18:53
Air Blank	0.000	18:54
Control Test	0.079	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:55

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: UTLEY,BRETT	ARREST TIME: 01:09
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: COMER	DATE/TIME OF TEST: 04/06/2024 02:04:45
----------------------------	--

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:00
Breath Volume	1.531	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:03
Breath Volume	2.656	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MILLS,SCOTT,T	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MILLS,SCOTT	ARREST TIME: 06:08
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: HUGGINS DATE/TIME OF TEST: 04/06/2024 06:50:50

Diagnostics Check	OK	06:44
Air Blank	0.000	06:44
Control Test	0.080	06:44
Air Blank	0.000	06:45
Subject Sample # 1	0.147	06:46
Breath Volume	2.039	06:46
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.148	06:49
Breath Volume	1.503	06:49
Air Blank	0.000	06:49
Control Test	0.080	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 20:51
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STAINO	DATE/TIME OF TEST: 04/06/2024 21:46:35
-----------------------------	--

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.202	21:42
Breath Volume	2.617	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.204	21:45
Breath Volume	3.468	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GODWIN

DATE/TIME OF TEST: 04/07/2024 01:39:19

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.081	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B

BREATH TEST OPERATOR AGENCY: St. John's County SO

ARRESTING OFFICER: MONIE,JARED

ARRESTING OFFICER AGENCY: St. John's County SO

LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 01:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/07/2024 02:37:47

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.241	02:33
Breath Volume	1.625	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.233	02:36
Breath Volume	1.566	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: YANTES

DATE/TIME OF TEST: 04/09/2024 03:48:47

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.244	03:44
Breath Volume	2.785	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.245	03:47
Breath Volume	2.695	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DIAZ ROSARIO

DATE/TIME OF TEST: 04/09/2024 23:43:15

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.024	23:38
Breath Volume	1.222	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.024	23:41
Breath Volume	1.980	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MANGOLD,K,R	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: SOLES,R	ARREST TIME: 16:14
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: POOLE	DATE/TIME OF TEST: 04/17/2024	18:14:24
----------------------------	-------------------------------	----------

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.078	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.225	18:10
Breath Volume	2.531	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.228	18:12
Breath Volume	2.460	18:12
Air Blank	0.000	18:13
Control Test	0.075	18:13
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SUNDSTROM

DATE/TIME OF TEST: 04/17/2024 21:02:08

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.089	20:57
Breath Volume	1.960	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.090	21:00
Breath Volume	1.882	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRERA,ALEXANDER,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/18/2024 09:25:08

Diagnostics Check	OK	09:19
Air Blank	0.000	09:19
Control Test	0.080	09:19
Air Blank	0.000	09:20
Subject Sample # 1	0.173	09:20
Breath Volume	1.757	09:20
Air Blank	0.000	09:21
Air Blank	0.000	09:23
Subject Sample # 2	0.175	09:23
Breath Volume	1.585	09:23
Air Blank	0.000	09:24
Control Test	0.079	09:24
Air Blank	0.000	09:25
Diagnostics Check	OK	09:25

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: CANOVA

DATE/TIME OF TEST: 04/19/2024 17:15:41

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:04	
Control Test	0.080	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1 0.000	17:07	Volume Not Met
Breath Volume	0.394	17:07	
Air Blank	0.000	17:08	
Air Blank	0.000	17:10	
Subject Sample	# 2 0.000	17:11	
Breath Volume	2.652	17:11	
Air Blank	0.000	17:11	
Air Blank	0.000	17:13	
Subject Sample	# 3 0.000	17:14	
Breath Volume	2.949	17:14	
Air Blank	0.000	17:14	
Control Test	0.079	17:15	
Air Blank	0.000	17:15	
Diagnostics Check	OK	17:15	

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 01:29
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MONCRIEF	DATE/TIME OF TEST: 04/20/2024 02:24:27
-------------------------------	--

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.183	02:19
Breath Volume	1.980	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.174	02:22
Breath Volume	2.183	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 00:35
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Male
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: FULMER	DATE/TIME OF TEST: 04/21/2024 01:24:58
-----------------------------	--

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.122	01:20
Breath Volume	1.347	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.124	01:23
Breath Volume	1.605	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/24/2024 00:01:16

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.000	23:57
Breath Volume	4.164	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.000	23:59
Breath Volume	2.710	23:59
Air Blank	0.000	00:00
Control Test	0.081	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 220703
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B	DRY GAS EXPIRY DATE: 10/03/2025
BREATH TEST OPERATOR AGENCY: St. John's County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: MCCANDLESS,MATTHEW	ARREST TIME: 02:24
ARRESTING OFFICER AGENCY: St. John's County SO	GENDER: Female
LAST AGENCY INSPECTION: 04/02/2024	OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BOWEN	DATE/TIME OF TEST: 04/24/2024 03:30:35
----------------------------	--

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.082	03:19	
Air Blank	0.000	03:19	
Subject Sample # 1	0.197	03:20	
Breath Volume	1.257	03:20	
Air Blank	0.000	03:21	
Air Blank	0.000	03:22	
Subject Sample # 2	0.173	03:26	Volume Not Met
Breath Volume	0.324	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample # 3	0.190	03:29	
Breath Volume	1.203	03:29	
Air Blank	0.000	03:29	
Control Test	0.079	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SMITH,TRISTAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: SABEAN

DATE/TIME OF TEST: 04/28/2024 19:40:25

Diagnostics Check	OK	19:33
Air Blank	0.000	19:34
Control Test	0.081	19:34
Air Blank	0.000	19:35
Subject Sample # 1	0.197	19:35
Breath Volume	1.421	19:35
Air Blank	0.000	19:36
Air Blank	0.000	19:38
Subject Sample # 2	0.213	19:38
Breath Volume	1.613	19:38
Air Blank	0.000	19:39
Control Test	0.081	19:39
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SPEROS,KOBE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220703
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/29/2024 02:11:23

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.082	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.000	02:06
Breath Volume	2.777	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.000	02:10
Breath Volume	2.554	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

INSTRUMENT REGISTERED TO: ST JOHNS SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WHITE,TIMOTHY,E

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: WHITE,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 220703

DRY GAS EXPIRY DATE: 10/03/2025

VIOLATION CODE: DUI

ARREST TIME: 05:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: TRISTAN

DATE/TIME OF TEST: 04/29/2024 07:12:48

Diagnostics Check	OK	06:58	
Air Blank	0.000	06:59	
Control Test	0.081	06:59	
Air Blank	0.000	07:00	
Subject Sample	# 1 0.155	07:03	Volume Not Met
Breath Volume	0.402	07:03	
Air Blank	0.000	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 2 0.143	07:06	
Breath Volume	1.910	07:06	
Air Blank	0.000	07:07	
Air Blank	0.000	07:09	
Subject Sample	# 3 0.139	07:11	
Breath Volume	1.531	07:11	
Air Blank	0.000	07:11	
Control Test	0.076	07:12	
Air Blank	0.000	07:12	
Diagnostics Check	OK	07:12	

Data Download Date/Time: 05/02/2024 21:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: WIENCKOWSKI,CHANCE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BURFIELD

DATE/TIME OF TEST: 04/02/2024 00:03:11

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 2 0.119	23:54	Volume Not Met
Breath Volume	0.070	23:54	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.119	23:57	
Breath Volume	1.527	23:57	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.108	00:01	
Breath Volume	1.605	00:01	
Air Blank	0.000	00:02	
Control Test	0.079	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VANCE,SPENCER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DEWALT II

DATE/TIME OF TEST: 04/02/2024 23:47:18

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.000	23:42
Breath Volume	1.503	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.000	23:45
Breath Volume	2.378	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: EDDY,DANIEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/04/2024 21:05:25

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.082	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.200	21:00
Breath Volume	4.746	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.197	21:03
Breath Volume	4.582	21:03
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: JACOB

DATE/TIME OF TEST: 04/06/2024 19:18:16

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.082	19:13
Air Blank	0.000	19:13
Subject Sample # 1	0.000	19:14
Breath Volume	2.570	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.000	19:16
Breath Volume	2.476	19:16
Air Blank	0.000	19:17
Control Test	0.082	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS,A,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CARDENAS,A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 04/07/2024 16:50:29

Diagnostics Check	OK	16:44
Air Blank	0.000	16:44
Control Test	0.083	16:45
Air Blank	0.000	16:45
Subject Sample # 1	0.000	16:46
Breath Volume	2.949	16:46
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.000	16:49
Breath Volume	3.183	16:49
Air Blank	0.000	16:49
Control Test	0.083	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 04/10/2024 01:49:41

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.319	01:45
Breath Volume	2.277	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.332	01:48
Breath Volume	2.167	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOKERN,VICTORIA,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SYVERSON,PAUL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: MORRILL

DATE/TIME OF TEST: 04/13/2024 15:53:56

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:43	
Control Test	0.082	15:43	
Air Blank	0.000	15:43	
Subject Sample	# 1	15:46	No Sample Provided
Breath Volume	0.000	15:46	
Air Blank	0.000	15:47	
Air Blank	0.000	15:49	
Subject Sample	# 2	15:52	No Sample Provided
Breath Volume	0.000	15:52	
Air Blank	0.000	15:53	
Control Test	0.082	15:53	
Air Blank	0.000	15:53	
Diagnostics Check	OK	15:53	

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DE LA CRUZ,JUAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SANCHEZ DE LA CRUZ

DATE/TIME OF TEST: 04/15/2024 03:15:43

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.082	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.182	03:10
Breath Volume	2.183	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.182	03:14
Breath Volume	3.242	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 04/15/2024 17:38:39

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.082	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.000	17:34
Breath Volume	2.691	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.000	17:37
Breath Volume	2.667	17:37
Air Blank	0.000	17:37
Control Test	0.083	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VENTRY,WILLIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KAITZ

DATE/TIME OF TEST: 04/15/2024 23:47:33

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.082	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.083	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS, BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GRANDPRE

DATE/TIME OF TEST: 04/16/2024 03:39:41

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.173	03:35
Breath Volume	3.640	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.164	03:38
Breath Volume	3.929	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/16/2024 17:46:22

Diagnostics Check	OK	17:40
Air Blank	0.000	17:40
Control Test	0.082	17:41
Air Blank	0.000	17:41
Subject Sample # 1	0.000	17:41
Breath Volume	1.371	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 2	0.000	17:45
Breath Volume	1.640	17:45
Air Blank	0.000	17:45
Control Test	0.082	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOKERN,VICTORIA,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BOKERN,VICTORIA
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: STANGE

DATE/TIME OF TEST: 04/18/2024 09:21:05

Diagnostics Check	OK	09:15
Air Blank	0.000	09:15
Control Test	0.082	09:15
Air Blank	0.000	09:16
Subject Sample # 1	0.000	09:16
Breath Volume	2.316	09:16
Air Blank	0.000	09:17
Air Blank	0.000	09:19
Subject Sample # 2	0.000	09:19
Breath Volume	2.164	09:19
Air Blank	0.000	09:20
Control Test	0.082	09:20
Air Blank	0.000	09:21
Diagnostics Check	OK	09:21

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/24/2024 00:52:36

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.082	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.027	00:47
Breath Volume	1.386	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.027	00:51
Breath Volume	1.191	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJAMIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/24/2024 18:16:54

Diagnostics Check	OK	18:11
Air Blank	0.000	18:11
Control Test	0.083	18:11
Air Blank	0.000	18:12
Subject Sample # 1	0.000	18:12
Breath Volume	1.347	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.000	18:15
Breath Volume	1.503	18:15
Air Blank	0.000	18:15
Control Test	0.083	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 04/30/2024 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: JACOBS,BENJMAIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 04/25/2024 17:58:32

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.082	17:53
Air Blank	0.000	17:53
Subject Sample # 1	0.000	17:54
Breath Volume	2.566	17:54
Air Blank	0.000	17:54
Air Blank	0.000	17:56
Subject Sample # 2	0.000	17:57
Breath Volume	2.757	17:57
Air Blank	0.000	17:57
Control Test	0.082	17:57
Air Blank	0.000	17:58
Diagnostics Check	OK	17:58

Data Download Date/Time: 04/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,BENJAMIN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/25/2024 22:27:06

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.082	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.172	22:22
Breath Volume	1.128	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.161	22:25
Breath Volume	1.351	22:25
Air Blank	0.000	22:26
Control Test	0.082	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 04/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: VANCE, SPENCER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 04/26/2024 01:18:31

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:14	
Air Blank	0.000	01:14	
Subject Sample # 1	0.231	01:15	
Breath Volume	1.464	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample # 2	0.231	01:17	
Breath Volume	1.335	01:17	
Air Blank		01:18	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 04/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 03/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370891
DRY GAS EXPIRY DATE: 02/23/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: VIJACKA

DATE/TIME OF TEST: 04/28/2024 19:24:22

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:16	
Control Test	0.082	19:16	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:20	No Sample Provided
Breath Volume	0.000	19:20	
Air Blank	0.000	19:20	
Air Blank	0.000	19:22	
Subject Sample	# 2	19:23	Subject Test Refused
Breath Volume	0.000	19:23	
Air Blank	0.000	19:23	
Control Test	0.082	19:23	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 04/30/2024 22:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 04/05/2024 03:30:06

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.111	03:25
Breath Volume	2.648	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.103	03:28
Breath Volume	1.398	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: KELLICI

DATE/TIME OF TEST: 04/10/2024 19:58:09

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.078	19:49	
Air Blank	0.000	19:49	
Subject Sample	# 1 0.234	19:51	
Breath Volume	1.671	19:51	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample	# 2 0.221	19:56	Slope Not Level
Breath Volume	1.761	19:56	
Air Blank		19:57	Purge Fail
Air Blank	0.000	19:58	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: KELLICI

DATE/TIME OF TEST: 04/10/2024 20:00:36

Diagnostics Check	OK	19:59	
Air Blank	0.000	19:59	
Control Test	0.072	20:00	Control Outside Tolerance
Air Blank	0.000	20:00	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: KELLICI

DATE/TIME OF TEST: 04/10/2024 20:06:43

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:02	
Control Test	0.075	20:02	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.218	20:05	
Breath Volume	1.585	20:05	
Air Blank		20:06	Purge Fail
Air Blank	0.000	20:06	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 04/19/2024 21:40:55

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.076	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:39	Subject Test Refused
Breath Volume	0.644	21:39	
Air Blank	0.000	21:39	
Control Test	0.076	21:40	
Air Blank	0.000	21:40	
Diagnostics Check	OK	21:40	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 04/19/2024 23:26:58

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.077	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.188	23:22
Breath Volume	1.531	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.183	23:25
Breath Volume	1.847	23:25
Air Blank	0.000	23:26
Control Test	0.076	23:26
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: READ

DATE/TIME OF TEST: 04/20/2024 22:20:52

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.078	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1 0.300	22:13	
Breath Volume	2.718	22:13	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 2 0.271	22:16	
Breath Volume	1.636	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:18	
Subject Sample	# 3 0.286	22:19	
Breath Volume	2.464	22:19	
Air Blank	0.000	22:19	
Control Test	0.073	22:20	Control Outside Tolerance
Air Blank	0.000	22:20	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: READ

DATE/TIME OF TEST: 04/20/2024 22:23:33

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.074	22:22	Control Outside Tolerance
Air Blank	0.000	22:23	

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,JASON,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: READ

DATE/TIME OF TEST: 04/20/2024 22:31:10

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.075	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.298	22:26
Breath Volume	2.996	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.288	22:29
Breath Volume	2.453	22:29
Air Blank	0.000	22:30
Control Test	0.075	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LILES,R
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PAKENHAM

DATE/TIME OF TEST: 04/20/2024 23:47:13

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.077	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.155	23:42
Breath Volume	1.375	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.155	23:45
Breath Volume	1.605	23:45
Air Blank	0.000	23:46
Control Test	0.076	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: UTLEY,B
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/21/2024 01:24:40

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.077	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.165	01:19
Breath Volume	2.593	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.160	01:23
Breath Volume	2.351	01:23
Air Blank	0.000	01:23
Control Test	0.077	01:24
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,J,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ARMENTA,B
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PASSMORE

DATE/TIME OF TEST: 04/21/2024 01:52:24

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.157	01:47
Breath Volume	2.101	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.156	01:50
Breath Volume	1.949	01:50
Air Blank	0.000	01:51
Control Test	0.075	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/02/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAAB,COLTON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: AKERLIND

DATE/TIME OF TEST: 04/10/2024 17:28:39

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.080	17:22
Air Blank	0.000	17:23
Subject Sample # 1	0.144	17:23
Breath Volume	1.675	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 2	0.151	17:26
Breath Volume	2.433	17:26
Air Blank	0.000	17:27
Control Test	0.079	17:28
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTONINI,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RODRIGUEZ DIAZ

DATE/TIME OF TEST: 04/11/2024 02:00:46

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.202	01:56
Breath Volume	1.222	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.207	01:58
Breath Volume	1.589	01:58
Air Blank	0.000	01:59
Control Test	0.077	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRANOVSKY,ROBERT
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: CERON GARZON

DATE/TIME OF TEST: 04/12/2024 09:41:24

Diagnostics Check	OK	09:33
Air Blank	0.000	09:34
Control Test	0.081	09:34
Air Blank	0.000	09:35
Subject Sample # 1	0.144	09:36
Breath Volume	2.332	09:36
Air Blank	0.000	09:37
Air Blank	0.000	09:38
Subject Sample # 2	0.142	09:39
Breath Volume	2.113	09:39
Air Blank	0.000	09:40
Control Test	0.077	09:40
Air Blank	0.000	09:41
Diagnostics Check	OK	09:41

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/13/2024 06:26:56

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.082	06:21
Air Blank	0.000	06:21
Subject Sample # 1	0.000	06:22
Breath Volume	1.425	06:22
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.000	06:25
Breath Volume	2.281	06:25
Air Blank	0.000	06:25
Control Test	0.081	06:26
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RIGALI,ORLANDO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 04/14/2024 18:00:57

Diagnostics Check	OK	17:53
Air Blank	0.000	17:54
Control Test	0.082	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.264	17:56
Breath Volume	2.664	17:56
Air Blank	0.000	17:57
Air Blank	0.000	17:58
Subject Sample # 2	0.251	17:59
Breath Volume	2.066	17:59
Air Blank	0.000	17:59
Control Test	0.077	18:00
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MEJIA CLAROS

DATE/TIME OF TEST: 04/16/2024 02:21:23

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.100	02:16
Breath Volume	2.136	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.105	02:19
Breath Volume	2.339	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MASEILLE,BRYANT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 04/17/2024 15:20:59

Diagnostics Check	OK	15:14
Air Blank	0.000	15:14
Control Test	0.081	15:15
Air Blank	0.000	15:15
Subject Sample # 1	0.168	15:16
Breath Volume	1.363	15:16
Air Blank	0.000	15:16
Air Blank	0.000	15:18
Subject Sample # 2	0.162	15:19
Breath Volume	1.359	15:19
Air Blank	0.000	15:20
Control Test	0.080	15:20
Air Blank	0.000	15:20
Diagnostics Check	OK	15:20

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,JOMAR
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/19/2024 05:40:42

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.081	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.192	05:36
Breath Volume	2.074	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.201	05:38
Breath Volume	1.628	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAPOTA,JOSHUA
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GARCIA RIOS

DATE/TIME OF TEST: 04/20/2024 04:31:18

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank	0.000	04:19	
Subject Sample	# 1 0.171	04:23	Volume Not Met
Breath Volume	0.894	04:23	
Air Blank	0.000	04:24	
Air Blank	0.000	04:25	
Subject Sample	# 2 0.177	04:26	
Breath Volume	2.523	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 3 0.173	04:29	
Breath Volume	1.859	04:29	
Air Blank	0.000	04:30	
Control Test	0.077	04:30	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SGUROS,BRIAN
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: BISsoon

DATE/TIME OF TEST: 04/21/2024 18:53:42

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.082	18:47
Air Blank	0.000	18:47
Subject Sample # 1	0.175	18:48
Breath Volume	1.449	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.179	18:51
Breath Volume	1.914	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/21/2024 19:40:25

Diagnostics Check	OK	19:28	
Air Blank	0.000	19:28	
Control Test	0.080	19:29	
Air Blank	0.000	19:29	
Subject Sample	# 1 0.213	19:32	Volume Not Met
Breath Volume	0.726	19:32	
Air Blank	0.000	19:33	
Air Blank	0.000	19:35	
Subject Sample	# 2 0.235	19:35	
Breath Volume	1.902	19:35	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample	# 3 0.242	19:38	
Breath Volume	2.281	19:38	
Air Blank	0.000	19:39	
Control Test	0.077	19:39	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DOVI,FRANK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BERT

DATE/TIME OF TEST: 04/22/2024 02:22:20

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.136	02:17
Breath Volume	3.492	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.141	02:20
Breath Volume	3.675	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 04/24/2024 00:09:01

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.275	00:03
Breath Volume	2.347	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.265	00:07
Breath Volume	2.578	00:07
Air Blank	0.000	00:08
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:09

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FAUL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/25/2024 04:48:58

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.189	04:41
Breath Volume	1.316	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 2	0.195	04:47
Breath Volume	1.597	04:47
Air Blank	0.000	04:47
Control Test	0.079	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMANO,MICHAEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: CAUSEY

DATE/TIME OF TEST: 04/25/2024 18:40:53

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1 0.153	18:31	Volume Not Met
Breath Volume	0.292	18:31	
Air Blank	0.000	18:32	
Air Blank	0.000	18:34	
Subject Sample	# 2 0.178	18:34	
Breath Volume	1.554	18:34	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 3 0.168	18:39	
Breath Volume	1.417	18:39	
Air Blank	0.000	18:39	
Control Test	0.079	18:40	
Air Blank	0.000	18:40	
Diagnostics Check	OK	18:40	

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOVEE,VINCENT
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MUNOZ SEPULVEDA

DATE/TIME OF TEST: 04/25/2024 20:56:04

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.104	20:51
Breath Volume	2.167	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:53
Subject Sample # 2	0.092	20:54
Breath Volume	1.445	20:54
Air Blank	0.000	20:55
Control Test	0.079	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:56

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 04/29/2024 02:22:03

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.211	02:17
Breath Volume	1.589	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.194	02:20
Breath Volume	1.210	02:20
Air Blank	0.000	02:21
Control Test	0.076	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUERRA,ALVARO
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: GAITAN

DATE/TIME OF TEST: 04/30/2024 23:17:20

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.103	23:12
Breath Volume	1.300	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.095	23:15
Breath Volume	2.679	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/08/2024 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402816921
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,	DRY GAS EXPIRY DATE: 08/09/2026
BREATH TEST OPERATOR AGENCY: FHP Troop B	VIOLATION CODE: DUI
ARRESTING OFFICER: HERNANDEZ,ALAIN	ARREST TIME: 19:07
ARRESTING OFFICER AGENCY: FHP Troop B	GENDER: Male
LAST AGENCY INSPECTION: 04/01/2024	OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: GONZALEZ	DATE/TIME OF TEST: 04/08/2024	19:27:27
-------------------------------	-------------------------------	----------

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.081	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.279	19:22
Breath Volume	3.066	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.269	19:25
Breath Volume	2.761	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 05/01/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SABO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HAMILTON SANTIAGO

DATE/TIME OF TEST: 04/13/2024 20:03:32

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:01	
Control Test	0.083	20:01	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:02	
Control Test	0.081	20:02	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

Data Download Date/Time: 05/01/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LISS,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TULLONGE

DATE/TIME OF TEST: 04/28/2024 02:55:57

Diagnostics Check	Fail	02:55	Internal Printer Test
Air Blank	0.000	02:55	

Data Download Date/Time: 05/01/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LISS,CODY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: BUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TULLONGE

DATE/TIME OF TEST: 04/28/2024 02:58:38

Diagnostics Check	Fail	02:58	Internal Printer Test
Air Blank	0.000	02:58	

Data Download Date/Time: 05/01/2024 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,JOHN,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HAMMOND,JAXSON
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 03/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402754628
DRY GAS EXPIRY DATE: 05/30/2026
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: BLANTON

DATE/TIME OF TEST: 04/14/2024 03:43:52

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.078	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1 0.068	03:36	Volume Not Met
Breath Volume	0.046	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2 0.076	03:42	Volume Not Met
Breath Volume	0.156	03:42	
Air Blank	0.000	03:42	
Control Test	0.076	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 04/24/2024 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: TENISON

DATE/TIME OF TEST: 04/14/2024 07:19:41

Diagnostics Check	OK	07:18	
Air Blank		07:19	Ambient Fail
Air Blank	0.000	07:19	

Data Download Date/Time: 05/08/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: TENISON

DATE/TIME OF TEST: 04/14/2024 07:21:59

Diagnostics Check	OK	07:21	
Air Blank		07:21	Ambient Fail
Air Blank		07:21	Ambient Fail

Data Download Date/Time: 05/08/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: TENISON

DATE/TIME OF TEST: 04/14/2024 07:30:31

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.082	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.210	07:25
Breath Volume	2.078	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.209	07:28
Breath Volume	2.105	07:28
Air Blank	0.000	07:29
Control Test	0.078	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 05/08/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: ARONIA SIFUENTES

DATE/TIME OF TEST: 04/13/2024 19:28:30

Diagnostics Check	OK	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.196	19:23
Breath Volume	1.902	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.201	19:26
Breath Volume	1.757	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SANTIAGO CARTAGENA

DATE/TIME OF TEST: 04/17/2024 01:38:18

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:36	RFI Detect
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.079	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: SANTIAGO CARTAGENA

DATE/TIME OF TEST: 04/17/2024 01:47:44

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.274	01:43
Breath Volume	2.265	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.261	01:46
Breath Volume	1.796	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ZAMUDIO

DATE/TIME OF TEST: 04/22/2024 19:37:01

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:21	
Control Test	0.080	19:21	
Air Blank	0.000	19:22	
Subject Sample	# 1 0.084	19:25	Volume Not Met
Breath Volume	0.851	19:25	
Air Blank	0.000	19:25	
Air Blank	0.000	19:27	
Subject Sample	# 2 0.096	19:29	
Breath Volume	1.117	19:29	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample	# 3 0.086	19:35	Volume Not Met
Breath Volume	0.890	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	
Air Blank	0.000	19:36	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ZAMUDIO

DATE/TIME OF TEST: 04/22/2024 19:50:19

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.079	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.097	19:44
Breath Volume	1.367	19:44
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.094	19:48
Breath Volume	1.609	19:48
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SANTAMARIA

DATE/TIME OF TEST: 04/26/2024 01:31:39

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.145	01:27
Breath Volume	3.656	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.141	01:30
Breath Volume	3.386	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GONZALEZ TISTA

DATE/TIME OF TEST: 04/29/2024 03:09:53

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.116	03:05
Breath Volume	1.695	03:05
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.112	03:08
Breath Volume	1.753	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/08/2024 14:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: GUERRERO ABUNDIS

DATE/TIME OF TEST: 04/27/2024 05:34:51

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.189	05:30
Breath Volume	2.472	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.174	05:33
Breath Volume	1.199	05:33
Air Blank	0.000	05:33
Control Test	0.079	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 05/08/2024 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, DESMAND, M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: AMBROCIO AMBROCIO

DATE/TIME OF TEST: 04/29/2024 00:52:45

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.146	00:47
Breath Volume	1.742	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.137	00:51
Breath Volume	3.269	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 05/08/2024 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: KING,TYLER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 04/17/2024 03:33:34

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.082	03:28
Breath Volume	2.511	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.084	03:31
Breath Volume	2.328	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 04/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH,LAURA,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DIEDESCH,LAURA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 04/21/2024 23:52:22

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.247	23:47
Breath Volume	2.093	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.236	23:50
Breath Volume	1.796	23:50
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 04/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOFFMMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: CRISERA,MATTHEW
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/26/2024 09:33:11

Diagnostics Check	OK	09:25
Air Blank	0.000	09:26
Control Test	0.080	09:26
Air Blank	0.000	09:27
Subject Sample # 1	0.066	09:28
Breath Volume	1.148	09:28
Air Blank	0.000	09:29
Air Blank	0.000	09:30
Subject Sample # 2	0.063	09:31
Breath Volume	1.562	09:31
Air Blank	0.000	09:32
Control Test	0.079	09:32
Air Blank	0.000	09:33
Diagnostics Check	OK	09:33

Data Download Date/Time: 04/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRESLEY,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/27/2024 00:52:35

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.111	00:47
Breath Volume	1.593	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.106	00:50
Breath Volume	1.125	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 04/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SPENCER,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: OTHER
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/06/2024 18:47:13

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.082	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.152	18:42
Breath Volume	2.035	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.152	18:45
Breath Volume	1.121	18:45
Air Blank	0.000	18:46
Control Test	0.081	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 04/29/2024 04:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 04/01/2024 00:01:47

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.171	23:57
Breath Volume	1.398	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.167	00:00
Breath Volume	1.457	00:00
Air Blank	0.000	00:00
Control Test	0.081	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,C,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 04/05/2024 23:58:14

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.081	23:56	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.082	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,JOSE
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: EMMERT

DATE/TIME OF TEST: 04/06/2024 18:34:28

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.082	18:32	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:33	Subject Test Refused
Breath Volume	0.000	18:33	
Air Blank	0.000	18:33	
Control Test	0.081	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 04/12/2024 10:46:54

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.081	10:41
Air Blank	0.000	10:41
Subject Sample # 1	0.000	10:42
Breath Volume	1.867	10:42
Air Blank	0.000	10:42
Air Blank	0.000	10:44
Subject Sample # 2	0.000	10:45
Breath Volume	1.652	10:45
Air Blank	0.000	10:45
Control Test	0.080	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN, WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 04/15/2024 15:20:49

Diagnostics Check	OK	15:14	
Air Blank	0.000	15:15	
Control Test	0.082	15:15	
Air Blank	0.000	15:15	
Subject Sample # 1	0.000	15:16	
Breath Volume	2.437	15:16	
Air Blank	0.000	15:16	
Air Blank	0.000	15:19	
Subject Sample # 2	0.000	15:19	
Breath Volume	1.648	15:19	
Air Blank	0.000	15:19	
Control Test	0.074	15:20	Interferent Detect
Air Blank	0.000	15:20	

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 04/17/2024 17:17:15

Diagnostics Check	OK	17:11
Air Blank	0.000	17:11
Control Test	0.082	17:11
Air Blank	0.000	17:12
Subject Sample # 1	0.000	17:12
Breath Volume	2.257	17:12
Air Blank	0.000	17:13
Air Blank	0.000	17:15
Subject Sample # 2	0.000	17:15
Breath Volume	1.695	17:15
Air Blank	0.000	17:16
Control Test	0.081	17:16
Air Blank	0.000	17:17
Diagnostics Check	OK	17:17

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/22/2024 16:26:43

Diagnostics Check	OK	16:20
Air Blank	0.000	16:20
Control Test	0.081	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:22
Breath Volume	2.507	16:22
Air Blank	0.000	16:22
Air Blank	0.000	16:24
Subject Sample # 2	0.000	16:25
Breath Volume	2.453	16:25
Air Blank	0.000	16:25
Control Test	0.081	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: DIAZ ASCANI

DATE/TIME OF TEST: 04/23/2024 16:58:58

Diagnostics Check	OK	16:52
Air Blank	0.000	16:52
Control Test	0.082	16:52
Air Blank	0.000	16:53
Subject Sample # 1	0.000	16:54
Breath Volume	1.851	16:54
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:57
Breath Volume	1.925	16:57
Air Blank	0.000	16:58
Control Test	0.081	16:58
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/23/2024 20:04:40

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.082	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.000	20:00
Breath Volume	2.074	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.000	20:03
Breath Volume	2.410	20:03
Air Blank	0.000	20:03
Control Test	0.082	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 04/24/2024 12:24:43

Diagnostics Check	OK	12:18
Air Blank	0.000	12:18
Control Test	0.081	12:19
Air Blank	0.000	12:19
Subject Sample # 1	0.000	12:20
Breath Volume	3.128	12:20
Air Blank	0.000	12:20
Air Blank	0.000	12:22
Subject Sample # 2	0.000	12:23
Breath Volume	3.113	12:23
Air Blank	0.000	12:23
Control Test	0.080	12:24
Air Blank	0.000	12:24
Diagnostics Check	OK	12:24

Data Download Date/Time: 04/29/2024 03:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTMAN,BRANDON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:06

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 04/25/2024 08:29:27

Diagnostics Check	OK	08:26	
Air Blank	0.000	08:26	
Control Test	0.081	08:27	
Air Blank	0.000	08:27	
Subject Sample	# 1	08:28	Subject Test Refused
Breath Volume	0.000	08:28	
Air Blank	0.000	08:28	
Control Test	0.080	08:28	
Air Blank	0.000	08:29	
Diagnostics Check	OK	08:29	

Data Download Date/Time: 04/29/2024 03:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: GOPHER

DATE/TIME OF TEST: 04/26/2024 10:06:43

Diagnostics Check	OK	10:00
Air Blank	0.000	10:00
Control Test	0.081	10:01
Air Blank	0.000	10:01
Subject Sample # 1	0.000	10:02
Breath Volume	1.828	10:02
Air Blank	0.000	10:02
Air Blank	0.000	10:04
Subject Sample # 2	0.000	10:05
Breath Volume	1.738	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Diagnostics Check	OK	10:06

Data Download Date/Time: 04/29/2024 03:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SCHROEDER,ETHAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 03/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/28/2024 01:43:24

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.166	01:38
Breath Volume	1.710	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.178	01:41
Breath Volume	2.242	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 04/29/2024 03:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ELLINGTON,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MCNELLIS

DATE/TIME OF TEST: 04/30/2024 20:54:34

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.081	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 05/31/2024 01:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO,RAILY,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LANE,DALTON
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 03/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: BURDETTE

DATE/TIME OF TEST: 04/06/2024 17:33:23

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.081	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.156	17:27
Breath Volume	1.816	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.171	17:31
Breath Volume	2.296	17:31
Air Blank	0.000	17:32
Control Test	0.078	17:32
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 04/30/2024 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAQUETTE,AARON,Y
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 03/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CHAVEZ ARDON

DATE/TIME OF TEST: 04/12/2024 22:11:24

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.217	22:06
Breath Volume	3.078	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.219	22:09
Breath Volume	2.289	22:09
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 04/30/2024 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 03/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 04/13/2024 21:33:44

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.176	21:28
Breath Volume	2.402	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.177	21:32
Breath Volume	1.210	21:32
Air Blank	0.000	21:32
Control Test	0.076	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 04/30/2024 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,JASON,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: ROOSA,JOHN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 03/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: GUERRO-GUDINO

DATE/TIME OF TEST: 04/25/2024 15:19:32

Diagnostics Check	OK	15:12
Air Blank	0.000	15:13
Control Test	0.079	15:13
Air Blank	0.000	15:14
Subject Sample # 1	0.172	15:14
Breath Volume	1.855	15:14
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.157	15:17
Breath Volume	1.121	15:17
Air Blank	0.000	15:18
Control Test	0.079	15:18
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 04/30/2024 19:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDILL, CHERI, L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL, DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 03/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A1
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:49

SUBJECT'S LAST NAME: JIMENEZ-JIMENEZ

DATE/TIME OF TEST: 04/28/2024 14:17:57

Diagnostics Check	OK	14:11
Air Blank	0.000	14:12
Control Test	0.080	14:12
Air Blank	0.000	14:12
Subject Sample # 1	0.221	14:13
Breath Volume	1.378	14:13
Air Blank	0.000	14:14
Air Blank	0.000	14:15
Subject Sample # 2	0.227	14:16
Breath Volume	2.367	14:16
Air Blank	0.000	14:17
Control Test	0.077	14:17
Air Blank	0.000	14:17
Diagnostics Check	OK	14:17

Data Download Date/Time: 04/30/2024 19:09