

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 06/08/2024 06:17:41

Diagnostics Check	OK	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:10
Air Blank	0.000	06:10
Subject Sample # 1	0.104	06:12
Breath Volume	1.300	06:12
Air Blank	0.000	06:12
Air Blank	0.000	06:14
Subject Sample # 2	0.100	06:16
Breath Volume	1.457	06:16
Air Blank	0.000	06:16
Control Test	0.077	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 06/20/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ANDERSON,J
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: WHITE,ANDERSON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 06/12/2024 02:49:27

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.246	02:44
Breath Volume	1.554	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.256	02:47
Breath Volume	2.816	02:47
Air Blank	0.000	02:48
Control Test	0.077	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 06/20/2024 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLER,BRANDON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ACUNAMILAN

DATE/TIME OF TEST: 06/15/2024 03:46:51

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.078	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.140	03:45	
Air Blank	0.000	03:45	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 06/20/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN-
ROSQUETE,STEPHANNE,J
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE CASCARET,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024

VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GAVINO MARFIL

DATE/TIME OF TEST: 06/18/2024 03:54:25

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.244	03:49
Breath Volume	1.218	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.229	03:52
Breath Volume	2.035	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/20/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: STENBACK,AUSTIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/18/2024 23:41:51

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.073	23:41	Control Outside Tolerance
Air Blank	0.000	23:41	

Data Download Date/Time: 06/20/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: STENBACK,AUSTIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/18/2024 23:52:12

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.077	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.240	23:47	
Breath Volume	2.457	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.077	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 06/20/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE CASCARET,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: CARIAS

DATE/TIME OF TEST: 06/19/2024 00:39:10

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.078	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.078	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 06/20/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COLUMBO,DANIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: SEKOYAN

DATE/TIME OF TEST: 06/21/2024 05:39:57

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.211	05:35
Breath Volume	1.546	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.196	05:38
Breath Volume	1.378	05:38
Air Blank	0.000	05:39
Control Test	0.078	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 07/23/2024 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CARBALLO,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DESRAMEAUX

DATE/TIME OF TEST: 06/28/2024 22:04:52

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.079	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:03	
Control Test	0.079	22:04	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:04	

Data Download Date/Time: 07/23/2024 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BANNISTER,JACQUELINE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FEOLA

DATE/TIME OF TEST: 06/30/2024 01:31:50

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.166	01:27
Breath Volume	1.953	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.174	01:30
Breath Volume	1.957	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 07/23/2024 13:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKEWELL,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: BAKEWELL,ZACHARY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BROSS

DATE/TIME OF TEST: 06/27/2024 02:32:18

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 07/11/2024 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/02/2024 07:16:31

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.079	07:11
Air Blank	0.000	07:11
Subject Sample # 1	0.325	07:12
Breath Volume	2.031	07:12
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.312	07:14
Breath Volume	1.269	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 06/06/2024 02:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MAYAHERNANDEZ

DATE/TIME OF TEST: 06/14/2024 02:31:42

Diagnostics Check	OK	02:31	
Air Blank		02:31	RFI Detect
Air Blank		02:31	RFI Detect

Data Download Date/Time: 07/04/2024 05:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MAYAHERNANDEZ

DATE/TIME OF TEST: 06/14/2024 02:37:23

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:35	
Control Test	0.082	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 07/04/2024 05:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: PERALTA,L
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: JARQUINRIVAS

DATE/TIME OF TEST: 06/14/2024 03:09:48

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.235	03:05
Breath Volume	1.769	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:07
Subject Sample # 2	0.242	03:08
Breath Volume	1.367	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/04/2024 05:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,HENRIQUE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: COOPER,HENRIQUE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 05/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 06/05/2024 00:21:58

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 06/17/2024 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIVAL,PHANEL,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: LAURENT,BRENDA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 06/16/2024 13:54:01

Diagnostics Check	OK	13:51	
Air Blank	0.000	13:51	
Control Test	0.078	13:51	
Air Blank	0.000	13:52	
Subject Sample	# 1	13:52	Subject Test Refused
Breath Volume	0.000	13:52	
Air Blank	0.000	13:53	
Control Test	0.078	13:53	
Air Blank	0.000	13:53	
Diagnostics Check	OK	13:54	

Data Download Date/Time: 06/17/2024 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,CALEB,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILLIS,CALEB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: GOSCHA

DATE/TIME OF TEST: 06/09/2024 17:30:55

Diagnostics Check	OK	17:24
Air Blank	0.000	17:25
Control Test	0.080	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.226	17:26
Breath Volume	2.558	17:26
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.220	17:29
Breath Volume	2.722	17:29
Air Blank	0.000	17:29
Control Test	0.080	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 07/07/2024 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/10/2024 05:11:23

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.080	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.146	05:06
Breath Volume	1.832	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.150	05:09
Breath Volume	2.503	05:09
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/07/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/13/2024 15:27:02

Diagnostics Check	OK	15:20
Air Blank	0.000	15:20
Control Test	0.079	15:21
Air Blank	0.000	15:21
Subject Sample # 1	0.098	15:22
Breath Volume	2.941	15:22
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.097	15:25
Breath Volume	2.773	15:25
Air Blank	0.000	15:26
Control Test	0.080	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:27

Data Download Date/Time: 07/07/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: KILMORE

DATE/TIME OF TEST: 06/15/2024 02:28:02

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.130	02:23
Breath Volume	3.003	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.129	02:26
Breath Volume	3.039	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/07/2024 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/01/2024 04:10:11

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.078	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1 0.246	04:03	Volume Not Met
Breath Volume	0.035	04:03	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 2 0.202	04:08	Volume Not Met
Breath Volume	0.417	04:08	
Air Blank	0.000	04:09	
Control Test	0.078	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 06/13/2024 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/01/2024 04:21:44

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1 0.223	04:17	Volume Not Met
Breath Volume	0.050	04:17	
Air Blank	0.000	04:17	
Air Blank	0.000	04:19	
Subject Sample	# 2	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.076	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 06/13/2024 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: REY

DATE/TIME OF TEST: 06/01/2024 19:05:51

Diagnostics Check	OK	18:59
Air Blank	0.000	19:00
Control Test	0.081	19:00
Air Blank	0.000	19:01
Subject Sample # 1	0.000	19:01
Breath Volume	1.875	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.000	19:04
Breath Volume	1.820	19:04
Air Blank	0.000	19:04
Control Test	0.081	19:05
Air Blank	0.000	19:05
Diagnostics Check	OK	19:05

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: ZAMOT COLON

DATE/TIME OF TEST: 06/01/2024 22:31:25

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.120	22:26
Breath Volume	1.683	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.130	22:29
Breath Volume	1.804	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,DION
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PERKO

DATE/TIME OF TEST: 06/02/2024 22:33:35

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.082	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:32	
Control Test	0.081	22:32	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:33	

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUIRGUIS,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: TERRAZAS

DATE/TIME OF TEST: 06/03/2024 14:32:32

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.082	14:27
Air Blank	0.000	14:27
Subject Sample # 1	0.371	14:28
Breath Volume	2.390	14:28
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.362	14:30
Breath Volume	2.054	14:30
Air Blank	0.000	14:31
Control Test	0.080	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: OSSIAN,MARK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 11:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/07/2024 12:11:55

Diagnostics Check	OK	12:05
Air Blank	0.000	12:05
Control Test	0.081	12:06
Air Blank	0.000	12:06
Subject Sample # 1	0.298	12:07
Breath Volume	2.777	12:07
Air Blank	0.000	12:07
Air Blank	0.000	12:09
Subject Sample # 2	0.292	12:10
Breath Volume	2.925	12:10
Air Blank	0.000	12:10
Control Test	0.081	12:11
Air Blank	0.000	12:11
Diagnostics Check	OK	12:11

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ, ANTHONY, N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOBOSIEWICZ, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/09/2024 00:33:23

Diagnostics Check	OK	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.166	00:28
Breath Volume	1.308	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.175	00:31
Breath Volume	1.593	00:31
Air Blank	0.000	00:32
Control Test	0.081	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUGHES,CONNOR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: SHOCKEY

DATE/TIME OF TEST: 06/11/2024 23:05:30

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.309	23:00
Breath Volume	2.656	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.321	23:03
Breath Volume	2.597	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PASKALAKIS,BRIANNA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 06/12/2024 00:54:53

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 06/13/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,OFC B,
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: KLIPIN,OFC A
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 06/16/2024 01:59:19

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.081	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.163	01:54
Breath Volume	1.515	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.170	01:57
Breath Volume	2.226	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/17/2024 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE, ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: ROJAS CASTILLO

DATE/TIME OF TEST: 06/16/2024 03:31:54

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.111	03:27
Breath Volume	2.757	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.108	03:30
Breath Volume	2.269	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 07/17/2024 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: RODRIGUEZ GOMEZ

DATE/TIME OF TEST: 06/16/2024 03:54:39

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.081	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATAROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GASTON

DATE/TIME OF TEST: 06/16/2024 04:49:11

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.081	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.142	04:44
Breath Volume	2.433	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.136	04:47
Breath Volume	1.742	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: NICKCHEN

DATE/TIME OF TEST: 06/16/2024 19:41:50

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.082	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.192	19:36
Breath Volume	1.457	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:38
Subject Sample # 2	0.177	19:40
Breath Volume	1.601	19:40
Air Blank	0.000	19:40
Control Test	0.081	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ROSENWALD

DATE/TIME OF TEST: 06/20/2024 00:06:58

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.247	00:02
Breath Volume	1.816	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.249	00:05
Breath Volume	1.804	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: HERZ

DATE/TIME OF TEST: 06/20/2024 06:23:11

Diagnostics Check	OK	06:16
Air Blank	0.000	06:16
Control Test	0.081	06:17
Air Blank	0.000	06:17
Subject Sample # 1	0.127	06:18
Breath Volume	1.250	06:18
Air Blank	0.000	06:18
Air Blank	0.000	06:20
Subject Sample # 2	0.129	06:21
Breath Volume	1.527	06:21
Air Blank	0.000	06:22
Control Test	0.080	06:22
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 06/22/2024 03:08:59

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.065	03:04
Breath Volume	1.941	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.062	03:07
Breath Volume	1.714	03:07
Air Blank	0.000	03:08
Control Test	0.081	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATAROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 06/23/2024 02:28:09

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.024	02:22
Breath Volume	1.425	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.023	02:26
Breath Volume	1.390	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUNDH,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BINGHAM

DATE/TIME OF TEST: 06/23/2024 03:26:11

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.081	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.259	03:17
Breath Volume	1.128	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:19
Subject Sample # 2	0.231	03:21
Breath Volume	1.312	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 3	0.229	03:24
Breath Volume	1.773	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 06/24/2024 00:16:50

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.130	00:12
Breath Volume	3.035	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.130	00:15
Breath Volume	2.492	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG233401
DRY GAS EXPIRY DATE: 11/30/2024
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 06/29/2024 22:22:13

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:18
Breath Volume	2.261	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:20
Breath Volume	1.382	22:20
Air Blank	0.000	22:21
Control Test	0.082	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/17/2024 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VASQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/01/2024 00:53:56

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.081	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.085	00:49
Breath Volume	1.277	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.083	00:52
Breath Volume	1.246	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VASQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 06/01/2024 03:41:28

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.133	03:36
Breath Volume	1.449	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.132	03:39
Breath Volume	1.609	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: PROFETA

DATE/TIME OF TEST: 06/01/2024 05:29:13

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.163	05:24
Breath Volume	2.335	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.161	05:27
Breath Volume	2.488	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CHAPPELL

DATE/TIME OF TEST: 06/01/2024 21:14:45

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.286	21:09
Breath Volume	1.371	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.289	21:13
Breath Volume	1.492	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KING,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 06/01/2024 23:27:48

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.081	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.197	23:22
Breath Volume	2.191	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.195	23:26
Breath Volume	1.316	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: KIPNESS

DATE/TIME OF TEST: 06/02/2024 07:46:59

Diagnostics Check	OK	07:44	
Air Blank	0.000	07:44	
Control Test	0.080	07:44	
Air Blank	0.000	07:45	
Subject Sample	# 1	07:45	Subject Test Refused
Breath Volume	0.000	07:45	
Air Blank	0.000	07:46	
Control Test	0.080	07:46	
Air Blank	0.000	07:46	
Diagnostics Check	OK	07:46	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BJURHOLM,JACOB,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHULTZ,DYLAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TOOKES

DATE/TIME OF TEST: 06/03/2024 02:28:49

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 06/04/2024 22:23:39

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NICHOLS,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ADCOCK

DATE/TIME OF TEST: 06/06/2024 01:49:44

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.000	01:45
Breath Volume	2.089	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.000	01:48
Breath Volume	2.300	01:48
Air Blank	0.000	01:48
Control Test	0.081	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 06/07/2024 01:40:44

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.000	01:36
Breath Volume	1.746	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.000	01:39
Breath Volume	2.062	01:39
Air Blank	0.000	01:39
Control Test	0.082	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LIGHTFOOT,LANE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/07/2024 20:39:19

Diagnostics Check	OK	20:33
Air Blank	0.000	20:33
Control Test	0.082	20:34
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:35
Breath Volume	2.511	20:35
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:37
Breath Volume	3.003	20:37
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Diagnostics Check	OK	20:39

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EDDINGER,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ANDRUS

DATE/TIME OF TEST: 06/08/2024 01:12:30

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.262	01:07
Breath Volume	2.718	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.257	01:10
Breath Volume	2.742	01:10
Air Blank	0.000	01:11
Control Test	0.079	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, DEP GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, DEP GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/08/2024 03:07:44

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.152	03:02
Breath Volume	1.273	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.157	03:06
Breath Volume	1.332	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,M,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VERH,JORDAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 06/08/2024 07:16:35

Diagnostics Check	OK	07:13	
Air Blank	0.000	07:14	
Control Test	0.080	07:14	
Air Blank	0.000	07:14	
Subject Sample	# 1	07:15	Subject Test Refused
Breath Volume	0.000	07:15	
Air Blank	0.000	07:15	
Control Test	0.080	07:15	
Air Blank	0.000	07:16	
Diagnostics Check	OK	07:16	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,M,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,AMIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:42

SUBJECT'S LAST NAME: SILVIO-RAMON

DATE/TIME OF TEST: 06/08/2024 11:25:49

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:14	
Control Test	0.080	11:15	
Air Blank	0.000	11:15	
Subject Sample	# 1 0.000	11:18	Volume Not Met
Breath Volume	0.000	11:18	
Air Blank	0.000	11:19	
Air Blank	0.000	11:21	
Subject Sample	# 2 0.000	11:24	Volume Not Met
Breath Volume	0.019	11:24	
Air Blank	0.000	11:24	
Control Test	0.080	11:25	
Air Blank	0.000	11:25	
Diagnostics Check	OK	11:25	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,CHRISTOPHER,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CRYDERMAN,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 06/09/2024 05:54:16

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:38	
Control Test	0.080	05:38	
Air Blank	0.000	05:39	
Subject Sample	# 1 0.205	05:42	Volume Not Met
Breath Volume	0.031	05:42	
Air Blank	0.000	05:43	
Air Blank	0.000	05:45	
Subject Sample	# 2 0.179	05:48	
Breath Volume	1.792	05:48	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 3 0.175	05:52	
Breath Volume	1.550	05:52	
Air Blank	0.000	05:53	
Control Test	0.076	05:53	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:54	

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 06/10/2024 03:15:22

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.083	03:10
Breath Volume	2.820	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.084	03:13
Breath Volume	3.062	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUSBANDS, TYLER, J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HUSBANDS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: MARTINEZ RODRIGUEZ

DATE/TIME OF TEST: 06/10/2024 05:00:26

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.081	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.226	04:55
Breath Volume	2.007	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.228	04:58
Breath Volume	2.035	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: KASLANDER

DATE/TIME OF TEST: 06/11/2024 01:22:39

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.244	01:18
Breath Volume	1.609	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.244	01:21
Breath Volume	1.859	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SHORT,JACLYN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/11/2024 02:34:40

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.265	02:26
Breath Volume	1.304	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.244	02:29
Breath Volume	1.312	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 3	0.255	02:33
Breath Volume	2.121	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATAROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/13/2024 00:15:31

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.137	00:10
Breath Volume	1.195	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.137	00:13
Breath Volume	1.410	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 06/13/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATAROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BURROUGH

DATE/TIME OF TEST: 06/13/2024 20:08:05

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.081	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:06	Subject Test Refused
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	
Control Test	0.080	20:07	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PAPPAMIHIEL,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 06/13/2024 21:14:01

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.082	21:07
Air Blank	0.000	21:07
Subject Sample # 1	0.238	21:09
Breath Volume	2.078	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.234	21:12
Breath Volume	1.507	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:14

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATAROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZAPATAROLON,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ELFISHAWY

DATE/TIME OF TEST: 06/14/2024 01:12:40

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.082	01:07
Breath Volume	1.726	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.080	01:11
Breath Volume	1.449	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHGRABER,BRIAN,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PAPPAMIHIEL,ELAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PREAL

DATE/TIME OF TEST: 06/14/2024 03:45:29

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.000	03:40
Breath Volume	1.847	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:44
Breath Volume	1.984	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: ODAK

DATE/TIME OF TEST: 06/14/2024 18:49:57

Diagnostics Check	OK	18:43
Air Blank	0.000	18:43
Control Test	0.081	18:44
Air Blank	0.000	18:44
Subject Sample # 1	0.115	18:45
Breath Volume	2.539	18:45
Air Blank	0.000	18:46
Air Blank	0.000	18:47
Subject Sample # 2	0.117	18:48
Breath Volume	2.007	18:48
Air Blank	0.000	18:49
Control Test	0.080	18:49
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHGRABER,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PAULK

DATE/TIME OF TEST: 06/14/2024 22:29:53

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.173	22:24
Breath Volume	1.917	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.179	22:28
Breath Volume	1.902	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRCHGRABER,BRIAN,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANDERSON,JASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 06/15/2024 00:08:28

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.118	00:03
Breath Volume	1.441	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.118	00:06
Breath Volume	1.558	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: RHOADS

DATE/TIME OF TEST: 06/15/2024 04:25:36

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.081	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.081	04:24	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUHLMEIER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 06/15/2024 21:15:58

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.146	21:11
Breath Volume	1.535	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:13
Subject Sample # 2	0.147	21:14
Breath Volume	1.441	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MANNO

DATE/TIME OF TEST: 06/16/2024 00:17:03

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.000	00:12
Breath Volume	3.027	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:15
Breath Volume	2.796	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEISER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GEISER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: HOOK

DATE/TIME OF TEST: 06/16/2024 01:54:32

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.080	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1 0.180	01:46	Volume Not Met
Breath Volume	0.851	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:49	
Subject Sample	# 2 0.193	01:49	
Breath Volume	1.601	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 3 0.200	01:52	
Breath Volume	2.273	01:52	
Air Blank	0.000	01:53	
Control Test	0.079	01:53	
Air Blank	0.000	01:54	
Diagnostics Check	OK	01:54	

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HUGHES,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/16/2024 03:14:48

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.254	03:10
Breath Volume	1.613	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.250	03:13
Breath Volume	1.605	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MADDALENA

DATE/TIME OF TEST: 06/16/2024 23:59:41

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.000	23:55
Breath Volume	2.484	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	1.988	23:58
Air Blank	0.000	23:58
Control Test	0.082	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EDDINGER,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/19/2024 03:33:48

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 06/19/2024 14:38:11

Diagnostics Check	OK	14:35	
Air Blank	0.000	14:35	
Control Test	0.080	14:36	
Air Blank	0.000	14:36	
Subject Sample	# 1	14:36	Subject Test Refused
Breath Volume	0.000	14:36	
Air Blank	0.000	14:37	
Control Test	0.080	14:37	
Air Blank	0.000	14:38	
Diagnostics Check	OK	14:38	

Data Download Date/Time: 07/17/2024 09:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/20/2024 00:04:03

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.081	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LUTH,TIMOTHY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/20/2024 20:05:45

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.080	20:00
Air Blank	0.000	20:00
Subject Sample # 1	0.000	20:01
Breath Volume	1.628	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.000	20:04
Breath Volume	2.296	20:04
Air Blank	0.000	20:04
Control Test	0.081	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/20/2024 22:51:50

Diagnostics Check	OK	22:48	
Air Blank	0.000	22:49	
Control Test	0.081	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Control Test	0.081	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/21/2024 03:51:09

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.000	03:46
Breath Volume	2.027	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.000	03:49
Breath Volume	2.195	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,HOWARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/21/2024 15:13:59

Diagnostics Check	OK	15:11	
Air Blank	0.000	15:11	
Control Test	0.081	15:11	
Air Blank	0.000	15:12	
Subject Sample	# 1	15:12	Subject Test Refused
Breath Volume	0.000	15:12	
Air Blank	0.000	15:13	
Control Test	0.081	15:13	
Air Blank	0.000	15:13	
Diagnostics Check	OK	15:13	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KATSAREAS,NIKOLAOS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEAL,LOUIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: AUPPERLEE

DATE/TIME OF TEST: 06/22/2024 15:45:57

Diagnostics Check	OK	15:42	
Air Blank	0.000	15:43	
Control Test	0.081	15:43	
Air Blank	0.000	15:44	
Subject Sample	# 1	15:44	Subject Test Refused
Breath Volume	0.000	15:44	
Air Blank	0.000	15:45	
Control Test	0.080	15:45	
Air Blank	0.000	15:45	
Diagnostics Check	OK	15:45	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STACHNIK,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: FORLINI

DATE/TIME OF TEST: 06/22/2024 16:44:01

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.081	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.133	16:39
Breath Volume	1.632	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.138	16:42
Breath Volume	1.453	16:42
Air Blank	0.000	16:43
Control Test	0.080	16:43
Air Blank	0.000	16:43
Diagnostics Check	OK	16:44

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO,ERIK,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CUTINO,ERIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CALDERONE

DATE/TIME OF TEST: 06/23/2024 02:20:10

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.108	02:15
Breath Volume	1.664	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:17
Subject Sample # 2	0.104	02:18
Breath Volume	1.597	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLYWINE,GARRETT,V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,GARRETT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HENDRICKSON

DATE/TIME OF TEST: 06/23/2024 03:37:37

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.107	03:32
Breath Volume	2.140	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.113	03:36
Breath Volume	2.750	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROSS,JILLIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 06/23/2024 23:53:34

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASS,JAKOB,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLASS,JAKOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ZAMBARDINO

DATE/TIME OF TEST: 06/24/2024 00:24:03

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:19
Breath Volume	2.339	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.000	00:22
Breath Volume	1.527	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:24

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GIORDANO,TANYA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: SARANTOS

DATE/TIME OF TEST: 06/24/2024 19:37:54

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:35	
Control Test	0.080	19:35	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:36	
Control Test	0.080	19:37	
Air Blank	0.000	19:37	
Diagnostics Check	OK	19:37	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: LUNDBERG

DATE/TIME OF TEST: 06/25/2024 00:38:03

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.081	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.219	00:32
Breath Volume	1.628	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.226	00:36
Breath Volume	1.429	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LOPEZ-HERNANDEZ

DATE/TIME OF TEST: 06/26/2024 22:40:09

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.224	22:31
Breath Volume	2.339	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.200	22:35
Breath Volume	1.382	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 3	0.217	22:38
Breath Volume	1.121	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASS,JAKOB,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLASS,JAKOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: PEPE

DATE/TIME OF TEST: 06/27/2024 22:36:10

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:31
Breath Volume	2.187	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.000	22:34
Breath Volume	1.800	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:08

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 06/28/2024 17:51:31

Diagnostics Check	OK	17:39	
Air Blank	0.000	17:40	
Control Test	0.081	17:40	
Air Blank	0.000	17:40	
Subject Sample	# 1 0.145	17:43	Volume Not Met
Breath Volume	0.945	17:43	
Air Blank	0.000	17:44	
Air Blank	0.000	17:46	
Subject Sample	# 2 0.152	17:46	
Breath Volume	1.808	17:46	
Air Blank	0.000	17:47	
Air Blank	0.000	17:49	
Subject Sample	# 3 0.161	17:49	
Breath Volume	1.914	17:49	
Air Blank	0.000	17:50	
Control Test	0.079	17:50	
Air Blank	0.000	17:51	
Diagnostics Check	OK	17:51	

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAINES,TYLER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: BAURES

DATE/TIME OF TEST: 06/29/2024 17:06:44

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.081	17:01
Air Blank	0.000	17:01
Subject Sample # 1	0.161	17:02
Breath Volume	1.976	17:02
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.178	17:05
Breath Volume	2.382	17:05
Air Blank	0.000	17:05
Control Test	0.080	17:06
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 06/30/2024 02:22:10

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:17
Breath Volume	2.707	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	2.371	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 07/17/2024 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VALDEZ,SEAN,D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: VALDEZ,SEAN

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402783528

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 18:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: FROBENIUS

DATE/TIME OF TEST: 06/01/2024 05:03:00

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:54	
Control Test	0.081	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1	04:58	No Sample Provided
Breath Volume	0.000	04:58	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 2	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.082	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 06/26/2024 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,SEAN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: VALDEZ,SEAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: FROBENIUS

DATE/TIME OF TEST: 06/01/2024 05:17:13

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.199	05:11
Breath Volume	1.953	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.182	05:15
Breath Volume	1.625	05:15
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 06/26/2024 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402469736
BREATH TEST OPERATOR: STINSON,ERIC,S	DRY GAS EXPIRY DATE: 06/16/2025
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: STINSON,ERIC	ARREST TIME: 16:01
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/29/2024	OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BROWN	DATE/TIME OF TEST: 06/02/2024	02:06:15
----------------------------	-------------------------------	----------

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.157	01:58
Breath Volume	2.464	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.135	02:01
Breath Volume	1.351	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 3	0.134	02:04
Breath Volume	1.515	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: JUDKINS

DATE/TIME OF TEST: 06/07/2024 18:11:58

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:09	
Control Test	0.083	18:09	
Air Blank	0.000	18:09	
Subject Sample	# 1	18:10	Subject Test Refused
Breath Volume	0.000	18:10	
Air Blank	0.000	18:11	
Control Test	0.082	18:11	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402783528
BREATH TEST OPERATOR: MURSCH,THEODORE,R	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: MURSCH,THEODORE	ARREST TIME: 15:50
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/29/2024	OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: OATES	DATE/TIME OF TEST: 06/08/2024 16:30:23
----------------------------	--

Diagnostics Check	OK	16:24
Air Blank	0.000	16:24
Control Test	0.083	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.000	16:25
Breath Volume	3.175	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.000	16:29
Breath Volume	3.289	16:29
Air Blank	0.000	16:29
Control Test	0.083	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PRITCHARD,AHSLEY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: BLYTHE

DATE/TIME OF TEST: 06/09/2024 23:03:37

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:00	
Control Test	0.082	23:01	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.083	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/13/2024 19:05:56

Diagnostics Check	OK	18:54	
Air Blank	0.000	18:55	
Control Test	0.082	18:55	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:58	No Sample Provided
Breath Volume	0.000	18:58	
Air Blank	0.000	18:59	
Air Blank	0.000	19:01	
Subject Sample	# 2	19:04	No Sample Provided
Breath Volume	0.000	19:04	
Air Blank	0.000	19:05	
Control Test	0.082	19:05	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: SENST,CASEY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 15:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:14

SUBJECT'S LAST NAME: HUCKABY

DATE/TIME OF TEST: 06/21/2024 15:48:52

Diagnostics Check	OK	15:45	
Air Blank	0.000	15:46	
Control Test	0.082	15:46	
Air Blank	0.000	15:47	
Subject Sample	# 1	15:47	Subject Test Refused
Breath Volume	0.000	15:47	
Air Blank	0.000	15:47	
Control Test	0.082	15:48	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402783528
BREATH TEST OPERATOR: MURSCH,THEODORE,R	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PRITCHARD,AHSLEY	ARREST TIME: 16:50
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 05/29/2024	OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: GARRETT	DATE/TIME OF TEST: 06/23/2024	17:26:36
------------------------------	-------------------------------	----------

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.083	17:19
Air Blank	0.000	17:19
Subject Sample # 1	0.094	17:22
Breath Volume	1.960	17:22
Air Blank	0.000	17:22
Air Blank	0.000	17:24
Subject Sample # 2	0.090	17:25
Breath Volume	2.035	17:25
Air Blank	0.000	17:25
Control Test	0.082	17:26
Air Blank	0.000	17:26
Diagnostics Check	OK	17:26

Data Download Date/Time: 06/26/2024 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PITRE

DATE/TIME OF TEST: 06/28/2024 19:40:00

Diagnostics Check	OK	19:31	
Air Blank	0.000	19:31	
Control Test	0.082	19:32	
Air Blank	0.000	19:32	
Subject Sample	# 1	19:35	No Sample Provided
Breath Volume	0.000	19:35	
Air Blank	0.000	19:36	
Air Blank	0.000	19:38	
Subject Sample	# 2	19:38	Subject Test Refused
Breath Volume	0.000	19:38	
Air Blank	0.000	19:39	
Control Test	0.083	19:39	
Air Blank	0.000	19:39	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 08/02/2024 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURSCH,THEODORE,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MURSCH,THEODORE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MCPHERSON

DATE/TIME OF TEST: 06/29/2024 20:25:45

Diagnostics Check	OK	20:18
Air Blank	0.000	20:19
Control Test	0.082	20:19
Air Blank	0.000	20:20
Subject Sample # 1	0.104	20:21
Breath Volume	3.250	20:21
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.097	20:24
Breath Volume	1.394	20:24
Air Blank	0.000	20:24
Control Test	0.083	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 08/02/2024 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, LANDON, M

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: CHRISTMAS, FOLIN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 05/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: BUI

ARREST TIME: 16:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/22/2024 17:18:45

Diagnostics Check	OK	17:09
Air Blank	0.000	17:10
Control Test	0.080	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.173	17:14
Breath Volume	1.890	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.167	17:17
Breath Volume	1.535	17:17
Air Blank	0.000	17:17
Control Test	0.080	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

Data Download Date/Time: 06/28/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A3
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 06/20/2024 21:01:27

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.080	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.056	20:57
Breath Volume	2.382	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.054	21:00
Breath Volume	2.003	21:00
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 07/26/2024 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: COLE,BRANDON
ARRESTING OFFICER AGENCY: GULF COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A3
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 06/20/2024 23:06:33

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	23:02	No Sample Provided
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 2	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 07/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A3
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HEDGECOCK

DATE/TIME OF TEST: 06/29/2024 20:54:46

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.080	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.045	20:50
Breath Volume	3.019	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.041	20:53
Breath Volume	2.601	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 07/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: PORT ST. JOE PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: PORT ST. JOE PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A3
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 06/30/2024 02:27:28

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1 0.290	02:19	Volume Not Met
Breath Volume	0.750	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2 0.284	02:22	
Breath Volume	1.359	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 3 0.286	02:26	
Breath Volume	1.371	02:26	
Air Blank	0.000	02:26	
Control Test	0.079	02:26	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 07/26/2024 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: CRUZ ROMUALDO

DATE/TIME OF TEST: 06/01/2024 04:26:52

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.083	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.214	04:19
Breath Volume	1.207	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.210	04:25
Breath Volume	1.847	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: BRAVO,GEORGE
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: GALE

DATE/TIME OF TEST: 06/01/2024 05:22:19

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.083	05:20	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.083	05:21	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: BLANDON,ERICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/05/2024 23:36:21

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.084	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.147	23:30
Breath Volume	1.402	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.142	23:34
Breath Volume	1.359	23:34
Air Blank	0.000	23:35
Control Test	0.082	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/06/2024 02:51:19

Diagnostics Check	OK	02:51	
Air Blank		02:51	RFI Detect
Air Blank		02:51	RFI Detect

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/06/2024 02:55:52

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.083	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	RFI Detect
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.084	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/06/2024 03:00:52

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.083	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.083	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 06/09/2024 23:24:41

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.084	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.083	23:24	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MAS

DATE/TIME OF TEST: 06/22/2024 01:44:46

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.083	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	
Control Test	0.083	01:44	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 06/29/2024 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: SANTOS,JAVIER
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 06/30/2024 02:54:19

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.083	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1	02:50	No Sample Provided
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.083	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 07/18/2024 23:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,K,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 06/12/2024 23:20:30

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.168	23:15
Breath Volume	2.031	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.169	23:18
Breath Volume	2.703	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAEDER,STEVEN,P
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FRANCOIS,WELBY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: RAEDER

DATE/TIME OF TEST: 06/14/2024 01:09:56

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.000	01:05
Breath Volume	2.660	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.000	01:08
Breath Volume	2.757	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNODLE,KADE,I
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SHIELDS

DATE/TIME OF TEST: 06/14/2024 02:40:40

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.078	02:36
Breath Volume	1.429	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.077	02:39
Breath Volume	1.617	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,TAREQ,H
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BERMUDEZ,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/16/2024 03:26:48

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.137	03:22
Breath Volume	2.296	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.135	03:25
Breath Volume	2.164	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMANDO,TINA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STANLEY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: GAVASTO

DATE/TIME OF TEST: 06/16/2024 21:13:57

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.310	21:09
Breath Volume	2.007	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.298	21:12
Breath Volume	1.386	21:12
Air Blank	0.000	21:13
Control Test	0.080	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JAMES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AUSTIN,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: POLLOM

DATE/TIME OF TEST: 06/21/2024 16:58:31

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.080	16:49
Air Blank	0.000	16:50
Subject Sample # 1	0.177	16:53
Breath Volume	1.328	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:55
Subject Sample # 2	0.187	16:56
Breath Volume	1.496	16:56
Air Blank	0.000	16:57
Control Test	0.079	16:57
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCANDREW,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MOLINA TAMIRANO

DATE/TIME OF TEST: 06/21/2024 21:55:52

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:48
Air Blank	0.000	21:49
Subject Sample # 1	0.196	21:51
Breath Volume	1.195	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:53
Subject Sample # 2	0.213	21:54
Breath Volume	1.339	21:54
Air Blank	0.000	21:54
Control Test	0.079	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 07/04/2024 10:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 06/22/2024 02:44:33

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.101	02:40
Breath Volume	1.386	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.103	02:43
Breath Volume	2.328	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RUIZ DUENAS

DATE/TIME OF TEST: 06/22/2024 04:00:19

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.121	03:55
Breath Volume	1.402	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.122	03:58
Breath Volume	1.332	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PENASANCHEZ

DATE/TIME OF TEST: 06/22/2024 04:11:14

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.180	04:06
Breath Volume	1.625	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.179	04:09
Breath Volume	1.296	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KORPOLINSKI,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 06/22/2024 04:20:53

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.220	04:16
Breath Volume	2.800	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.213	04:19
Breath Volume	2.718	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KELLEY,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PEREZ ROSETE

DATE/TIME OF TEST: 06/24/2024 03:37:54

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.175	03:32
Breath Volume	1.367	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.175	03:36
Breath Volume	1.402	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JONES,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: BARRIENTOS TRUJILLO

DATE/TIME OF TEST: 06/24/2024 05:01:14

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.322	04:56
Breath Volume	2.218	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.327	04:59
Breath Volume	1.945	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMONA,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SANTIAGO ORTIZ

DATE/TIME OF TEST: 06/28/2024 00:04:59

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.218	00:00
Breath Volume	3.183	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.221	00:03
Breath Volume	3.148	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZTRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RAMIREZ,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 06/28/2024 20:41:00

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.168	20:36
Breath Volume	2.148	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.158	20:39
Breath Volume	1.906	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:41

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTT,RYAN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HEBEN,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: BONANNO

DATE/TIME OF TEST: 06/30/2024 04:02:56

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.108	03:58
Breath Volume	1.929	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.107	04:01
Breath Volume	2.132	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 07/04/2024 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SCHIPANSKY,CHAD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 06/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: CLINKSCALES

DATE/TIME OF TEST: 06/15/2024 18:13:52

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.080	18:07
Air Blank	0.000	18:07
Subject Sample # 1	0.271	18:08
Breath Volume	1.347	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:10
Subject Sample # 2	0.256	18:11
Breath Volume	1.921	18:11
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 07/07/2024 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACZAK, DAVID, M
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: TACZAK, DAVID
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: MAGILL

DATE/TIME OF TEST: 06/06/2024 14:21:22

Diagnostics Check	OK	14:16	
Air Blank	0.000	14:16	
Control Test	0.079	14:16	
Air Blank	0.000	14:17	
Subject Sample # 1	0.323	14:17	
Breath Volume	3.398	14:17	
Air Blank	0.000	14:18	
Air Blank		14:19	Improper Sample
Air Blank	0.000	14:20	
Control Test	0.078	14:20	
Air Blank	0.000	14:21	
Diagnostics Check	OK	14:21	

Data Download Date/Time: 06/17/2024 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TACZAK, DAVID, M
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: TACZAK, DAVID
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: MAGILL

DATE/TIME OF TEST: 06/06/2024 14:30:25

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.077	14:25
Air Blank	0.000	14:25
Subject Sample # 1	0.312	14:26
Breath Volume	2.554	14:26
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.321	14:28
Breath Volume	3.562	14:28
Air Blank	0.000	14:29
Control Test	0.075	14:29
Air Blank	0.000	14:30
Diagnostics Check	OK	14:30

Data Download Date/Time: 06/17/2024 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMCZYK,ANDREW,P
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: ADAMCZYK,ANDREW
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: CALEJAS-GARCIA

DATE/TIME OF TEST: 06/24/2024 17:20:48

Diagnostics Check	OK	17:13
Air Blank	0.000	17:14
Control Test	0.078	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.277	17:16
Breath Volume	1.445	17:16
Air Blank	0.000	17:17
Air Blank	0.000	17:18
Subject Sample # 2	0.285	17:19
Breath Volume	2.640	17:19
Air Blank	0.000	17:19
Control Test	0.075	17:20
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 08/06/2024 10:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOOLEY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: DOOLEY

DATE/TIME OF TEST: 06/27/2024 03:29:41

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.166	03:24
Breath Volume	2.425	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.167	03:28
Breath Volume	2.074	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 07/31/2024 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, MATTHEW, C
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MORALES, BOBBY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: BUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 06/17/2024 18:41:43

Diagnostics Check	OK	18:35
Air Blank	0.000	18:36
Control Test	0.083	18:36
Air Blank	0.000	18:36
Subject Sample # 1	0.000	18:37
Breath Volume	3.574	18:37
Air Blank	0.000	18:38
Air Blank	0.000	18:39
Subject Sample # 2	0.000	18:40
Breath Volume	3.980	18:40
Air Blank	0.000	18:40
Control Test	0.082	18:41
Air Blank	0.000	18:41
Diagnostics Check	OK	18:41

Data Download Date/Time: 06/26/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL, HEATHER, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL, HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 06/29/2024 02:12:02

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.082	02:09	
Air Blank	0.000	02:09	
Subject Sample # 1	0.196	02:10	
Breath Volume	4.617	02:10	
Air Blank		02:11	Purge Fail
Air Blank	0.000	02:12	

Data Download Date/Time: 07/09/2024 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL,HEATHER,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL,HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 06/29/2024 02:20:59

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.184	02:16
Breath Volume	3.914	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.181	02:19
Breath Volume	4.878	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 07/09/2024 09:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WACHTEL,BRUCE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: TAVERASGABOT

DATE/TIME OF TEST: 06/01/2024 07:49:47

Diagnostics Check	OK	07:43
Air Blank	0.000	07:44
Control Test	0.082	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.000	07:45
Breath Volume	1.757	07:45
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.000	07:48
Breath Volume	3.375	07:48
Air Blank	0.000	07:48
Control Test	0.082	07:49
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIAZ-BUENO,YENIER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: COYLE

DATE/TIME OF TEST: 06/01/2024 23:17:30

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.082	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.007	23:13
Breath Volume	4.675	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.000	23:16
Breath Volume	4.574	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ALVAREZ QUINTANA

DATE/TIME OF TEST: 06/02/2024 01:57:15

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.166	01:52
Breath Volume	1.183	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.162	01:55
Breath Volume	1.726	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DE JESUS, HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: RAMOS DA SILVA

DATE/TIME OF TEST: 06/02/2024 02:24:25

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.235	02:19
Breath Volume	1.652	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.227	02:22
Breath Volume	1.562	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOSEPH, FRANKY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ORTIZ GONZALEZ

DATE/TIME OF TEST: 06/04/2024 03:53:49

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.146	03:49
Breath Volume	1.609	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.145	03:52
Breath Volume	1.136	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:32

SUBJECT'S LAST NAME: GONZALEZMOLDONADO

DATE/TIME OF TEST: 06/04/2024 11:02:32

Diagnostics Check	OK	10:56
Air Blank	0.000	10:57
Control Test	0.082	10:57
Air Blank	0.000	10:57
Subject Sample # 1	0.099	10:58
Breath Volume	1.832	10:58
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.100	11:01
Breath Volume	2.238	11:01
Air Blank	0.000	11:01
Control Test	0.081	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO, EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RAMOS CATALA

DATE/TIME OF TEST: 06/06/2024 01:36:33

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.082	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.082	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEONARD,WAKENS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:01

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/06/2024 15:32:36

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.081	15:25
Air Blank	0.000	15:26
Subject Sample # 1	0.180	15:27
Breath Volume	1.597	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.185	15:31
Breath Volume	1.429	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIERRA, EDGAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/07/2024 01:01:28

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.185	00:57
Breath Volume	2.722	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.183	00:59
Breath Volume	3.027	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: BLAKES

DATE/TIME OF TEST: 06/07/2024 03:00:29

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.082	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.175	02:56
Breath Volume	3.527	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.167	02:58
Breath Volume	3.742	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ,ISABELLA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 06/08/2024 03:32:59

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.106	03:28
Breath Volume	1.121	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.108	03:31
Breath Volume	2.257	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/09/2024 01:56:41

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.150	01:52
Breath Volume	1.175	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.151	01:55
Breath Volume	1.394	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 06/20/2024 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONANNO,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: ROJASCOVA

DATE/TIME OF TEST: 06/09/2024 06:57:47

Diagnostics Check	OK	06:51
Air Blank	0.000	06:52
Control Test	0.082	06:52
Air Blank	0.000	06:52
Subject Sample # 1	0.163	06:53
Breath Volume	1.769	06:53
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.170	06:56
Breath Volume	2.796	06:56
Air Blank	0.000	06:56
Control Test	0.081	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: MORANGONZALEZ

DATE/TIME OF TEST: 06/09/2024 07:40:54

Diagnostics Check	OK	07:34
Air Blank	0.000	07:34
Control Test	0.081	07:35
Air Blank	0.000	07:35
Subject Sample # 1	0.195	07:36
Breath Volume	2.968	07:36
Air Blank	0.000	07:36
Air Blank	0.000	07:38
Subject Sample # 2	0.197	07:39
Breath Volume	4.234	07:39
Air Blank	0.000	07:40
Control Test	0.081	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:40

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TAVERAS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 06/09/2024 21:55:15

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.081	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.120	21:50
Breath Volume	2.699	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:52
Subject Sample # 2	0.113	21:53
Breath Volume	1.683	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVAS, TABITHA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: SERAFIN OLMOS

DATE/TIME OF TEST: 06/13/2024 21:43:46

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.083	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.192	21:39
Breath Volume	2.835	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.174	21:42
Breath Volume	1.195	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TEURO,ADRIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JUAREZ BAUTISTA

DATE/TIME OF TEST: 06/13/2024 23:24:13

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.172	23:19
Breath Volume	1.316	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.176	23:22
Breath Volume	1.703	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: CHINCHILLA

DATE/TIME OF TEST: 06/14/2024 07:25:00

Diagnostics Check	OK	07:18
Air Blank	0.000	07:19
Control Test	0.082	07:19
Air Blank	0.000	07:19
Subject Sample # 1	0.122	07:20
Breath Volume	2.925	07:20
Air Blank	0.000	07:21
Air Blank	0.000	07:22
Subject Sample # 2	0.124	07:23
Breath Volume	3.601	07:23
Air Blank	0.000	07:24
Control Test	0.081	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:25

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/15/2024 00:26:26

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.186	00:21
Breath Volume	1.265	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.185	00:24
Breath Volume	1.945	00:24
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONANNO,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/15/2024 05:13:20

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.082	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.127	05:08
Breath Volume	1.105	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.129	05:11
Breath Volume	1.148	05:11
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LARTONOIX,BRYCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: SAINTIL

DATE/TIME OF TEST: 06/15/2024 06:18:41

Diagnostics Check	OK	06:12
Air Blank	0.000	06:13
Control Test	0.082	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.171	06:14
Breath Volume	1.957	06:14
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.178	06:17
Breath Volume	1.664	06:17
Air Blank	0.000	06:17
Control Test	0.081	06:18
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PERDOMO,ARCADIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: ORTIZBUSTAMANTE

DATE/TIME OF TEST: 06/15/2024 07:22:23

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.082	07:17
Air Blank	0.000	07:17
Subject Sample # 1	0.141	07:18
Breath Volume	2.328	07:18
Air Blank	0.000	07:18
Air Blank	0.000	07:20
Subject Sample # 2	0.142	07:20
Breath Volume	3.343	07:20
Air Blank	0.000	07:21
Control Test	0.081	07:21
Air Blank	0.000	07:22
Diagnostics Check	OK	07:22

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WACHTEL,BRUCE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/15/2024 08:04:16

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.081	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.161	07:59
Breath Volume	2.152	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.172	08:02
Breath Volume	2.796	08:02
Air Blank	0.000	08:03
Control Test	0.081	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONNANO, CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: HERNANDEZ SALAZAR

DATE/TIME OF TEST: 06/16/2024 05:02:34

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.082	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.125	04:58
Breath Volume	2.785	04:58
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.116	05:01
Breath Volume	1.804	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: BALMONTE

DATE/TIME OF TEST: 06/16/2024 07:15:25

Diagnostics Check	OK	07:03	
Air Blank	0.000	07:04	
Control Test	0.082	07:04	
Air Blank	0.000	07:05	
Subject Sample	# 1 0.081	07:08	Volume Not Met
Breath Volume	0.304	07:08	
Air Blank	0.000	07:08	
Air Blank	0.000	07:10	
Subject Sample	# 2 0.078	07:13	Volume Not Met
Breath Volume	0.539	07:13	
Air Blank	0.000	07:14	
Control Test	0.082	07:14	
Air Blank	0.000	07:15	
Diagnostics Check	OK	07:15	

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: BALMONTE

DATE/TIME OF TEST: 06/16/2024 07:25:04

Diagnostics Check	OK	07:16
Air Blank	0.000	07:17
Control Test	0.081	07:17
Air Blank	0.000	07:18
Subject Sample # 1	0.083	07:19
Breath Volume	1.328	07:19
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.085	07:23
Breath Volume	1.402	07:23
Air Blank	0.000	07:24
Control Test	0.081	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:25

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: RODRIGUEZ RIVERA

DATE/TIME OF TEST: 06/17/2024 03:50:31

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.071	03:45
Breath Volume	1.210	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.068	03:49
Breath Volume	1.347	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 06/18/2024 21:58:28

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.083	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.000	21:54
Breath Volume	3.335	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:57
Breath Volume	3.609	21:57
Air Blank	0.000	21:57
Control Test	0.082	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 06/20/2024 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CASAME

DATE/TIME OF TEST: 06/19/2024 00:44:54

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.199	00:37
Breath Volume	2.457	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.222	00:40
Breath Volume	2.925	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:42
Subject Sample # 3	0.212	00:43
Breath Volume	1.925	00:43
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 06/20/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LINTON,RYAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/20/2024 04:13:40

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.082	04:07	
Air Blank	0.000	04:08	
Subject Sample # 1	0.164	04:08	
Breath Volume	1.148	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample # 2		04:12	Slope Not Met
Breath Volume	2.117	04:12	
Air Blank	0.000	04:12	
Control Test	0.081	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/20/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LINTON,RYAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/20/2024 04:44:18

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.081	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.149	04:39
Breath Volume	1.921	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.142	04:42
Breath Volume	2.453	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 06/20/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/22/2024 02:54:21

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.081	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:47	No Sample Provided
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.082	02:53	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FERNANDEZ,JOSHUA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: RESZECKIFERREIRADACO

DATE/TIME OF TEST: 06/23/2024 00:58:51

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.082	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.161	00:54
Breath Volume	1.390	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.155	00:57
Breath Volume	2.253	00:57
Air Blank	0.000	00:57
Control Test	0.083	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PATTERSON,JACAARA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CARUS

DATE/TIME OF TEST: 06/23/2024 01:45:55

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.082	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.181	01:41
Breath Volume	1.300	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.185	01:44
Breath Volume	2.300	01:44
Air Blank	0.000	01:45
Control Test	0.081	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 06/23/2024 02:32:52

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.082	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:28	
Subject Sample	# 2	02:31	No Sample Provided
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: OLIVEROS

DATE/TIME OF TEST: 06/24/2024 02:38:08

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.206	02:33
Breath Volume	3.156	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.210	02:36
Breath Volume	3.429	02:36
Air Blank	0.000	02:37
Control Test	0.082	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARMON,JEREMY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: BRUMAIRE

DATE/TIME OF TEST: 06/24/2024 04:18:31

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.190	04:12
Breath Volume	1.343	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.184	04:16
Breath Volume	1.437	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CINTOPEREZ

DATE/TIME OF TEST: 06/29/2024 01:42:49

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.214	01:38
Breath Volume	1.613	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.222	01:41
Breath Volume	1.761	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/30/2024 02:45:35

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.082	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:41
Breath Volume	3.593	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	2.757	02:44
Air Blank	0.000	02:44
Control Test	0.082	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 07/04/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCKINNON,SIR FRANCIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: VARGAS RODRIGUEZ

DATE/TIME OF TEST: 06/02/2024 15:47:16

Diagnostics Check	OK	15:40
Air Blank	0.000	15:41
Control Test	0.083	15:41
Air Blank	0.000	15:42
Subject Sample # 1	0.160	15:42
Breath Volume	1.683	15:42
Air Blank	0.000	15:43
Air Blank	0.000	15:45
Subject Sample # 2	0.158	15:45
Breath Volume	2.054	15:45
Air Blank	0.000	15:46
Control Test	0.081	15:46
Air Blank	0.000	15:47
Diagnostics Check	OK	15:47

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GARZA,JUSTIN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: PORRAS VARGAS

DATE/TIME OF TEST: 06/05/2024 04:32:33

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.083	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.079	04:28
Breath Volume	2.792	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.080	04:31
Breath Volume	3.039	04:31
Air Blank	0.000	04:31
Control Test	0.083	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BAPTISE

DATE/TIME OF TEST: 06/07/2024 23:50:25

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.083	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.082	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROYO,ALEJANDRO,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ARROYO,ALEJANDRO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: PEREZ RAMOS

DATE/TIME OF TEST: 06/08/2024 01:35:25

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.083	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.173	01:30
Breath Volume	2.117	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.174	01:33
Breath Volume	2.007	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CORINA,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ROSE,SHANE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: GONZALEZ ORTIZ

DATE/TIME OF TEST: 06/08/2024 21:13:42

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.083	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.200	21:09
Breath Volume	3.210	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.195	21:12
Breath Volume	1.753	21:12
Air Blank	0.000	21:12
Control Test	0.083	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,C
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 06/08/2024 22:37:05

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
		22:36	RFI Detect
Air Blank	0.000	22:37	

Data Download Date/Time: 06/17/2024 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PARZYGnat,DESTINY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 06/15/2024 08:24:08

Diagnostics Check	OK	08:21	
Air Blank	0.000	08:21	
Control Test	0.083	08:22	
Air Blank	0.000	08:22	
Subject Sample	# 1	08:23	Sequence Aborted
Breath Volume	0.000	08:23	
Air Blank	0.000	08:24	

Data Download Date/Time: 06/17/2024 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PARZYGnat,DESTINY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 06/15/2024 08:35:05

Diagnostics Check	OK	08:27
Air Blank	0.000	08:28
Control Test	0.083	08:28
Air Blank	0.000	08:29
Subject Sample # 1	0.179	08:30
Breath Volume	3.625	08:30
Air Blank	0.000	08:31
Air Blank	0.000	08:32
Subject Sample # 2	0.178	08:33
Breath Volume	3.585	08:33
Air Blank	0.000	08:34
Control Test	0.079	08:34
Air Blank	0.000	08:35
Diagnostics Check	OK	08:35

Data Download Date/Time: 06/17/2024 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BANKSTON

DATE/TIME OF TEST: 06/23/2024 01:26:57

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.083	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.007	01:22
Breath Volume	2.140	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.000	01:25
Breath Volume	2.144	01:25
Air Blank	0.000	01:26
Control Test	0.082	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/19/2024 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUADALUPE,GABRIEL,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GUADALUPE,GABRIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JIMENEZ ALVAREZ

DATE/TIME OF TEST: 06/23/2024 03:17:57

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.083	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.249	03:13
Breath Volume	1.492	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.249	03:16
Breath Volume	1.312	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/19/2024 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BORGES-LOMBANO

DATE/TIME OF TEST: 06/23/2024 04:59:01

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.083	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.083	04:54
Breath Volume	1.355	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.082	04:57
Breath Volume	1.933	04:57
Air Blank	0.000	04:58
Control Test	0.083	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:59

Data Download Date/Time: 07/19/2024 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: COBIAN

DATE/TIME OF TEST: 06/24/2024 05:19:42

Diagnostics Check	OK	05:13
Air Blank	0.000	05:13
Control Test	0.083	05:14
Air Blank	0.000	05:14
Subject Sample # 1	0.248	05:15
Breath Volume	2.917	05:15
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.248	05:18
Breath Volume	1.679	05:18
Air Blank	0.000	05:18
Control Test	0.083	05:19
Air Blank	0.000	05:19
Diagnostics Check	OK	05:19

Data Download Date/Time: 07/19/2024 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSCIS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: NUSCIS,NICHOLAS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402330433
DRY GAS EXPIRY DATE: 01/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MINER-MARRERO

DATE/TIME OF TEST: 06/28/2024 00:29:46

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.083	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.185	00:24
Breath Volume	1.144	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.185	00:28
Breath Volume	1.351	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 07/19/2024 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BERRIOS,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: RIDLER

DATE/TIME OF TEST: 06/02/2024 03:35:22

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.217	03:30
Breath Volume	1.808	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.211	03:33
Breath Volume	1.503	03:33
Air Blank	0.000	03:34
Control Test	0.075	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COLLINS,THOMAS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DUARTE CASTILLO

DATE/TIME OF TEST: 06/02/2024 04:02:30

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.077	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.119	03:57
Breath Volume	1.777	03:57
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.110	04:00
Breath Volume	1.734	04:00
Air Blank	0.000	04:01
Control Test	0.077	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CURBELO,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GRAYS

DATE/TIME OF TEST: 06/15/2024 00:27:29

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.165	00:23
Breath Volume	1.441	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.152	00:25
Breath Volume	1.187	00:25
Air Blank	0.000	00:26
Control Test	0.077	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,EVAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: O'NIELL

DATE/TIME OF TEST: 06/15/2024 03:04:18

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.144	02:59
Breath Volume	1.152	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.143	03:02
Breath Volume	1.273	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEAL,DANIAL,J
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BERRIOS,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/16/2024 03:20:39

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.077	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1 0.151	03:13	Volume Not Met
Breath Volume	0.101	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 2 0.153	03:19	Volume Not Met
Breath Volume	0.296	03:19	
Air Blank	0.000	03:19	
Control Test	0.075	03:20	
Air Blank	0.000	03:20	
Diagnostics Check	OK	03:20	

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEAL,DANIAL,J
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BERRIOS,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/16/2024 03:29:57

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.076	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.153	03:25
Breath Volume	1.292	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.164	03:28
Breath Volume	2.062	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKE,PALMER,
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: EATON,M.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/16/2024 06:40:58

Diagnostics Check	OK	06:34
Air Blank	0.000	06:35
Control Test	0.077	06:35
Air Blank	0.000	06:35
Subject Sample # 1	0.154	06:36
Breath Volume	1.382	06:36
Air Blank	0.000	06:37
Air Blank	0.000	06:38
Subject Sample # 2	0.152	06:39
Breath Volume	1.675	06:39
Air Blank	0.000	06:40
Control Test	0.077	06:40
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEAL,DANIAL,J
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BERRIOS,JOSE
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: QUINNEY

DATE/TIME OF TEST: 06/26/2024 00:50:47

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:46
Breath Volume	2.027	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:49
Breath Volume	1.886	00:49
Air Blank	0.000	00:49
Control Test	0.077	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 06/28/2024 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 06/02/2024 00:27:10

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.251	00:22
Breath Volume	2.296	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.256	00:25
Breath Volume	3.011	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: LARKINS

DATE/TIME OF TEST: 06/02/2024 22:06:28

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.172	22:01
Breath Volume	1.472	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.177	22:04
Breath Volume	1.410	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 06/03/2024 01:54:26

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.078	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.185	01:49
Breath Volume	1.234	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.201	01:52
Breath Volume	1.843	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RAMOS PLAZA

DATE/TIME OF TEST: 06/05/2024 22:48:35

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.080	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1 0.130	22:41	Volume Not Met
Breath Volume	0.750	22:41	
Air Blank	0.000	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 2 0.125	22:47	Volume Not Met
Breath Volume	0.972	22:47	
Air Blank	0.000	22:47	
Control Test	0.079	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GRIMALDI,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RODRIGUEZ-DIAZ

DATE/TIME OF TEST: 06/06/2024 02:42:17

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.176	02:37
Breath Volume	2.523	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.174	02:40
Breath Volume	2.574	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROMERO STORY

DATE/TIME OF TEST: 06/12/2024 02:42:29

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.140	02:37
Breath Volume	1.714	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.137	02:40
Breath Volume	1.335	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WRIGHT,LAMONT
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MEDA CORADO

DATE/TIME OF TEST: 06/17/2024 03:16:42

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.167	03:12
Breath Volume	2.179	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.162	03:15
Breath Volume	2.921	03:15
Air Blank	0.000	03:15
Control Test	0.078	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/17/2024 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABREIL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402397742
DRY GAS EXPIRY DATE: 03/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SAVINE

DATE/TIME OF TEST: 06/20/2024 00:43:09

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.107	00:38
Breath Volume	2.015	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:40
Subject Sample # 2	0.103	00:41
Breath Volume	1.492	00:41
Air Blank	0.000	00:42
Control Test	0.078	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/19/2024 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TATE,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 06/02/2024 01:29:41

Diagnostics Check	OK	01:29	
Air Blank		01:29	RFI Detect
Air Blank		01:29	RFI Detect

Data Download Date/Time: 06/04/2024 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TATE,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 06/02/2024 01:34:47

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:34	RFI Detect
Breath Volume	0.000	01:34	
Air Blank		01:34	RFI Detect
Air Blank	0.000	01:34	

Data Download Date/Time: 06/04/2024 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TATE,BRANDON
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SANDOVAL

DATE/TIME OF TEST: 06/02/2024 01:52:51

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.217	01:48
Breath Volume	1.574	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.214	01:51
Breath Volume	1.515	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 06/04/2024 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: ABELL

DATE/TIME OF TEST: 06/04/2024 02:55:52

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.166	02:51
Breath Volume	2.175	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.164	02:54
Breath Volume	1.867	02:54
Air Blank	0.000	02:54
Control Test	0.078	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 06/04/2024 05:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: CLANTON,JESSICA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: WOSAR

DATE/TIME OF TEST: 06/05/2024 00:11:03

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.078	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.223	00:06
Breath Volume	3.421	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.228	00:09
Breath Volume	3.671	00:09
Air Blank	0.000	00:10
Control Test	0.077	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:11

Data Download Date/Time: 07/05/2024 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 06/08/2024 03:48:20

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.105	03:43
Breath Volume	2.585	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.097	03:46
Breath Volume	1.800	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 07/05/2024 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: CLANTON,JESSICA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: DINOVI

DATE/TIME OF TEST: 06/09/2024 23:09:51

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.121	23:04
Breath Volume	2.089	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.130	23:08
Breath Volume	2.156	23:08
Air Blank	0.000	23:08
Control Test	0.079	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 07/05/2024 19:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: KERLEY

DATE/TIME OF TEST: 06/10/2024 02:18:56

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.075	02:14
Breath Volume	2.683	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.073	02:17
Breath Volume	3.109	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/05/2024 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMES,R,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,N
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: LAWS

DATE/TIME OF TEST: 06/11/2024 01:10:25

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.183	01:05
Breath Volume	1.769	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.177	01:08
Breath Volume	1.148	01:08
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/05/2024 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: CLANTON,JESSICA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 06/13/2024 23:23:21

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.190	23:18
Breath Volume	1.839	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.190	23:21
Breath Volume	2.718	23:21
Air Blank	0.000	23:22
Control Test	0.079	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMES,R,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ADAMES,R
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 06/16/2024 03:38:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.060	03:34
Breath Volume	1.562	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.062	03:37
Breath Volume	1.859	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES, NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/17/2024 01:49:08

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.110	01:43
Breath Volume	1.378	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.102	01:47
Breath Volume	1.710	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: SIPES

DATE/TIME OF TEST: 06/18/2024 22:37:44

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.312	22:33
Breath Volume	3.062	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.308	22:36
Breath Volume	2.468	22:36
Air Blank	0.000	22:36
Control Test	0.077	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: ALFONSO RAVELO

DATE/TIME OF TEST: 06/19/2024 01:23:09

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.064	01:18
Breath Volume	1.550	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.065	01:21
Breath Volume	1.675	01:21
Air Blank	0.000	01:22
Control Test	0.078	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLARD,SEAN,M
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: CLANTON,JESSICA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: COHN

DATE/TIME OF TEST: 06/22/2024 22:59:45

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.078	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.223	22:55
Breath Volume	2.722	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.222	22:58
Breath Volume	2.843	22:58
Air Blank	0.000	22:58
Control Test	0.076	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HILL,ROBERT
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 06/23/2024 02:28:45

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.079	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.150	02:24
Breath Volume	3.414	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.151	02:27
Breath Volume	2.738	02:27
Air Blank	0.000	02:27
Control Test	0.078	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: INGERSOLL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: KABRAN

DATE/TIME OF TEST: 06/26/2024 23:22:32

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.079	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1 0.135	23:11	
Breath Volume	1.894	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.127	23:17	Volume Not Met
Breath Volume	0.382	23:17	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 3 0.127	23:21	
Breath Volume	1.335	23:21	
Air Blank	0.000	23:21	
Control Test	0.078	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MARQUEZ MAYORA

DATE/TIME OF TEST: 06/27/2024 00:29:03

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.118	00:23
Breath Volume	1.582	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.128	00:27
Breath Volume	1.433	00:27
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LANGDON

DATE/TIME OF TEST: 06/28/2024 00:26:26

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.154	00:22
Breath Volume	1.941	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.155	00:24
Breath Volume	2.261	00:24
Air Blank	0.000	00:25
Control Test	0.078	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 07/05/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMES,RAUL,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: TATE,B
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HLAVSA

DATE/TIME OF TEST: 06/30/2024 02:14:45

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.114	02:09
Breath Volume	1.187	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.109	02:13
Breath Volume	1.531	02:13
Air Blank	0.000	02:13
Control Test	0.078	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/05/2024 19:44