

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GOMEZ GARCIA

DATE/TIME OF TEST: 06/11/2024 02:09:47

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.199	02:05
Breath Volume	1.351	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.214	02:08
Breath Volume	3.363	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/23/2024 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: AKHREMENKO

DATE/TIME OF TEST: 06/14/2024 04:52:10

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.183	04:47
Breath Volume	2.023	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.178	04:50
Breath Volume	2.023	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 06/23/2024 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HURST,RAY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 06/18/2024 05:06:36

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.000	05:02
Breath Volume	1.640	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.000	05:05
Breath Volume	1.511	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 06/23/2024 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: LANTZ

DATE/TIME OF TEST: 06/20/2024 15:52:37

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.079	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.219	15:48
Breath Volume	2.949	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.219	15:51
Breath Volume	2.812	15:51
Air Blank	0.000	15:51
Control Test	0.079	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 06/26/2024 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: FERRIS

DATE/TIME OF TEST: 06/24/2024 21:13:44

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.079	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.000	21:09
Breath Volume	1.710	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:12
Breath Volume	3.375	21:12
Air Blank	0.000	21:12
Control Test	0.079	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 06/26/2024 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 06/24/2024 23:50:39

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.071	23:45
Breath Volume	1.578	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.082	23:49
Breath Volume	2.230	23:49
Air Blank	0.000	23:49
Control Test	0.078	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 06/26/2024 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,J,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HOWARD,J
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:23

SUBJECT'S LAST NAME: REVERON

DATE/TIME OF TEST: 06/25/2024 16:49:30

Diagnostics Check	OK	16:43
Air Blank	0.000	16:43
Control Test	0.079	16:43
Air Blank	0.000	16:44
Subject Sample # 1	0.166	16:45
Breath Volume	4.316	16:45
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.169	16:47
Breath Volume	4.175	16:47
Air Blank	0.000	16:48
Control Test	0.077	16:48
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 06/26/2024 21:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MOOSE

DATE/TIME OF TEST: 06/01/2024 05:00:31

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.083	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.123	04:56
Breath Volume	3.183	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.121	04:59
Breath Volume	2.968	04:59
Air Blank	0.000	04:59
Control Test	0.082	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/30/2024 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: SKEANS

DATE/TIME OF TEST: 06/04/2024 00:53:36

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.177	00:49
Breath Volume	2.675	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.168	00:52
Breath Volume	2.457	00:52
Air Blank	0.000	00:52
Control Test	0.082	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/30/2024 15:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: GREENDALE

DATE/TIME OF TEST: 06/02/2024 03:41:55

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1 0.121	03:36	Volume Not Met
Breath Volume	0.167	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	03:41	Sequence Aborted
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	

Data Download Date/Time: 06/03/2024 15:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEP,J,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PERKINS,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: BLACKBURN

DATE/TIME OF TEST: 06/09/2024 17:11:47

Diagnostics Check	OK	17:08	
Air Blank	0.000	17:09	
Control Test	0.079	17:09	
Air Blank	0.000	17:10	
Subject Sample	# 1	17:10	Subject Test Refused
Breath Volume	0.000	17:10	
Air Blank	0.000	17:10	
Control Test	0.079	17:11	
Air Blank	0.000	17:11	
Diagnostics Check	OK	17:11	

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: WILLIS-LAWRENCE

DATE/TIME OF TEST: 06/14/2024 09:37:14

Diagnostics Check	OK	09:25	
Air Blank	0.000	09:26	
Control Test	0.079	09:26	
Air Blank	0.000	09:26	
Subject Sample	# 1 0.000	09:27	
Breath Volume	1.144	09:27	
Air Blank	0.000	09:27	
Air Blank	0.000	09:29	
Subject Sample	# 2 0.000	09:32	Volume Not Met
Breath Volume	1.003	09:32	
Air Blank	0.000	09:33	
Air Blank	0.000	09:35	
Subject Sample	# 3 0.000	09:35	
Breath Volume	1.320	09:35	
Air Blank	0.000	09:36	
Control Test	0.078	09:36	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:37	

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAUN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/15/2024 03:04:53

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.042	03:00
Breath Volume	1.726	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.043	03:03
Breath Volume	1.355	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI, PHILIP, P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COFFMAN, DYLAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/16/2024 03:37:08

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.121	03:32
Breath Volume	1.941	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.123	03:35
Breath Volume	1.707	03:35
Air Blank	0.000	03:36
Control Test	0.077	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 06/16/2024 05:20:26

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.078	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1 0.249	05:16	
Breath Volume	3.023	05:16	
Air Blank	0.000	05:17	
Air Blank		05:19	RFI Detect
Air Blank	0.000	05:19	
Control Test	0.078	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 06/16/2024 05:28:43

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.239	05:24
Breath Volume	2.269	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.235	05:27
Breath Volume	2.011	05:27
Air Blank	0.000	05:27
Control Test	0.078	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOTOWSKI,PHILIP,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BROWN,IVERSON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 06/21/2024 01:00:23

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.000	00:55
Breath Volume	2.597	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.000	00:59
Breath Volume	1.843	00:59
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/21/2024 09:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,JAXON,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DORSEY,JAXON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SAVITSKIY

DATE/TIME OF TEST: 06/22/2024 23:59:53

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1	23:59	Sequence Aborted
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	

Data Download Date/Time: 07/18/2024 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: KISNER

DATE/TIME OF TEST: 06/23/2024 06:44:42

Diagnostics Check	OK	06:37
Air Blank	0.000	06:38
Control Test	0.079	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.178	06:40
Breath Volume	2.886	06:40
Air Blank	0.000	06:40
Air Blank	0.000	06:42
Subject Sample # 2	0.177	06:43
Breath Volume	2.339	06:43
Air Blank	0.000	06:43
Control Test	0.078	06:44
Air Blank	0.000	06:44
Diagnostics Check	OK	06:44

Data Download Date/Time: 07/18/2024 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CARDENALES

DATE/TIME OF TEST: 06/29/2024 19:52:49

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.238	19:48
Breath Volume	2.773	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.240	19:51
Breath Volume	2.250	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 07/18/2024 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:38

SUBJECT'S LAST NAME: APOLLON

DATE/TIME OF TEST: 06/30/2024 13:02:31

Diagnostics Check	OK	12:58	
Air Blank	0.000	12:58	
Control Test	0.079	12:58	
Air Blank	0.000	12:59	
Subject Sample	# 1	13:00	Subject Test Refused
Breath Volume	0.000	13:00	
Air Blank	0.000	13:01	
Control Test	0.079	13:01	
Air Blank	0.000	13:02	
Diagnostics Check	OK	13:02	

Data Download Date/Time: 07/18/2024 08:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERIA,MARIA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: HOPLER

DATE/TIME OF TEST: 06/06/2024 14:04:12

Diagnostics Check	OK	13:57
Air Blank	0.000	13:57
Control Test	0.080	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.255	13:59
Breath Volume	1.742	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.252	14:02
Breath Volume	1.746	14:02
Air Blank	0.000	14:03
Control Test	0.080	14:03
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 06/07/2024 03:51:17

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.233	03:46
Breath Volume	2.406	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.220	03:49
Breath Volume	1.847	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VALDES,YANELIS
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: CORZO MEDINA

DATE/TIME OF TEST: 06/07/2024 05:30:01

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.137	05:25
Breath Volume	2.089	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.141	05:28
Breath Volume	1.242	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:30

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MENDOZA,STEVEN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MCDEVITT

DATE/TIME OF TEST: 06/08/2024 00:44:08

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.173	00:39
Breath Volume	2.640	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.174	00:42
Breath Volume	3.066	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DORFMAN

DATE/TIME OF TEST: 06/08/2024 03:05:52

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.134	03:01
Breath Volume	3.074	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.130	03:04
Breath Volume	3.019	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REDFEARN,JAMES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/08/2024 04:18:09

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.163	04:13
Breath Volume	1.503	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.164	04:16
Breath Volume	1.472	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NORENA,ALEX
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 06/15/2024 02:36:44

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:35	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: RAMAGE JR.

DATE/TIME OF TEST: 06/15/2024 02:58:13

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.172	02:53
Breath Volume	1.437	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.177	02:56
Breath Volume	1.632	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: VU

DATE/TIME OF TEST: 06/15/2024 05:02:11

Diagnostics Check	OK	04:56
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.159	04:57
Breath Volume	1.867	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.148	05:00
Breath Volume	1.882	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: JIMENEZ URENA

DATE/TIME OF TEST: 06/17/2024 03:01:00

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.101	02:56
Breath Volume	2.375	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.099	02:59
Breath Volume	1.878	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMERENCIANO,MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 06/19/2024 21:30:39

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.080	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1 0.183	21:22	Volume Not Met
Breath Volume	1.078	21:22	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample	# 2 0.191	21:26	
Breath Volume	1.890	21:26	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 3 0.194	21:29	
Breath Volume	1.843	21:29	
Air Blank	0.000	21:29	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MIGUEL LOPEZ

DATE/TIME OF TEST: 06/23/2024 02:01:04

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.244	01:56
Breath Volume	1.453	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.239	01:59
Breath Volume	1.964	01:59
Air Blank	0.000	02:00
Control Test	0.081	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 06/23/2024 02:59:20

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.118	02:54
Breath Volume	1.566	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.129	02:57
Breath Volume	1.292	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: TITUS

DATE/TIME OF TEST: 06/24/2024 23:04:09

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:02	Slope Not Met
Breath Volume	0.976	23:02	
Air Blank	0.000	23:03	
Control Test	0.079	23:03	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAMPBELL,SHELBY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: TITUS

DATE/TIME OF TEST: 06/24/2024 23:32:58

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.239	23:28
Breath Volume	1.199	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.246	23:31
Breath Volume	1.449	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIL,RAMON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 06/27/2024 01:17:07

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.082	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.211	01:12
Breath Volume	1.855	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.210	01:15
Breath Volume	1.453	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BOONE,MICHAEL
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BONILLA MARTINEZ

DATE/TIME OF TEST: 06/28/2024 23:46:34

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.131	23:41
Breath Volume	1.597	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.132	23:45
Breath Volume	2.789	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 07/03/2024 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 06/12/2024 21:53:33

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.134	21:48
Breath Volume	1.109	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.135	21:51
Breath Volume	1.144	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: CONTRERAS-LOPEZ

DATE/TIME OF TEST: 06/13/2024 01:07:18

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.142	01:02
Breath Volume	1.574	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.141	01:05
Breath Volume	1.765	01:05
Air Blank	0.000	01:06
Control Test	0.077	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MUNSEN

DATE/TIME OF TEST: 06/13/2024 04:01:41

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.198	03:56
Breath Volume	2.812	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.189	04:00
Breath Volume	2.570	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 06/14/2024 02:21:24

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CRUZ-JIMENEZ

DATE/TIME OF TEST: 06/15/2024 01:17:40

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.131	01:13
Breath Volume	1.523	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.120	01:16
Breath Volume	2.484	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SHORES

DATE/TIME OF TEST: 06/18/2024 02:54:26

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.054	02:49
Breath Volume	2.761	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.051	02:52
Breath Volume	2.523	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: BATALITZKY

DATE/TIME OF TEST: 06/19/2024 05:48:04

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:46	
Subject Sample	# 1	05:46	Subject Test Refused
Breath Volume	0.000	05:46	
Air Blank	0.000	05:47	
Control Test	0.079	05:47	
Air Blank	0.000	05:47	
Diagnostics Check	OK	05:48	

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 06/20/2024 05:48:22

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.080	05:41
Air Blank	0.000	05:42
Subject Sample # 1	0.124	05:43
Breath Volume	1.207	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:45
Subject Sample # 2	0.115	05:46
Breath Volume	1.730	05:46
Air Blank	0.000	05:47
Control Test	0.078	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: GAMMELL

DATE/TIME OF TEST: 06/20/2024 14:26:13

Diagnostics Check	OK	14:19
Air Blank	0.000	14:19
Control Test	0.080	14:20
Air Blank	0.000	14:20
Subject Sample # 1	0.000	14:21
Breath Volume	2.140	14:21
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.000	14:24
Breath Volume	1.968	14:24
Air Blank	0.000	14:25
Control Test	0.080	14:25
Air Blank	0.000	14:26
Diagnostics Check	OK	14:26

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ANGELL

DATE/TIME OF TEST: 06/22/2024 01:06:36

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.086	01:01
Breath Volume	2.886	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.087	01:04
Breath Volume	3.039	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: LOPEZ AREOS

DATE/TIME OF TEST: 06/22/2024 04:27:20

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.079	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: YANDLE

DATE/TIME OF TEST: 06/23/2024 00:30:44

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Air Blank	0.000	00:26	
Subject Sample	# 2	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Control Test	0.082	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GUNTER

DATE/TIME OF TEST: 06/23/2024 03:34:29

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.145	03:29
Breath Volume	2.320	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.152	03:32
Breath Volume	2.585	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: GOMEZ ROBLERO

DATE/TIME OF TEST: 06/23/2024 23:16:29

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.184	23:11
Breath Volume	3.574	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.178	23:14
Breath Volume	3.503	23:14
Air Blank	0.000	23:15
Control Test	0.079	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 06/25/2024 00:51:28

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.080	00:47	
Air Blank	0.000	00:48	
Subject Sample	# 1 0.170	00:49	
Breath Volume	2.195	00:49	
Air Blank		00:50	Purge Fail
Air Blank		00:51	Purge Fail

Data Download Date/Time: 07/08/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CARSTENS

DATE/TIME OF TEST: 06/25/2024 02:08:09

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.077	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.166	02:03
Breath Volume	2.210	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.172	02:06
Breath Volume	1.785	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: KLEFFNER

DATE/TIME OF TEST: 06/25/2024 02:42:52

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.078	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.000	02:38
Breath Volume	2.914	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:41
Breath Volume	2.535	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HELMI

DATE/TIME OF TEST: 06/27/2024 05:08:38

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:06	
Control Test	0.080	05:06	
Air Blank	0.000	05:07	
Subject Sample	# 1	05:07	Subject Test Refused
Breath Volume	0.000	05:07	
Air Blank	0.000	05:07	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: HERNANDEZ CAMPOS

DATE/TIME OF TEST: 06/27/2024 22:09:14

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.215	22:04
Breath Volume	2.136	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.218	22:07
Breath Volume	2.136	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SUAREZ BALDERAS

DATE/TIME OF TEST: 06/28/2024 02:18:24

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.188	02:13
Breath Volume	2.328	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.181	02:16
Breath Volume	1.523	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ANTUNES SANTANA

DATE/TIME OF TEST: 06/29/2024 03:27:53

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.159	03:23
Breath Volume	1.769	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.157	03:26
Breath Volume	1.281	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 07/08/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: RULE

DATE/TIME OF TEST: 06/01/2024 02:00:20

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.081	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: STROMAN

DATE/TIME OF TEST: 06/01/2024 03:37:35

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.079	03:33
Breath Volume	1.859	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.076	03:36
Breath Volume	1.421	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: FULWOOD

DATE/TIME OF TEST: 06/02/2024 00:39:10

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.178	00:31
Breath Volume	2.656	00:31
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.155	00:34
Breath Volume	1.281	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 3	0.173	00:37
Breath Volume	2.781	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 06/03/2024 11:03:18

Diagnostics Check	OK	10:58	
Air Blank	0.000	10:58	
Control Test	0.081	10:59	
Air Blank	0.000	10:59	
Subject Sample	# 1	11:01	Subject Test Refused
Breath Volume	0.000	11:01	
Air Blank	0.000	11:02	
Control Test	0.080	11:02	
Air Blank	0.000	11:03	
Diagnostics Check	OK	11:03	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MCANENEY

DATE/TIME OF TEST: 06/03/2024 21:28:03

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.082	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	Subject Test Refused
Breath Volume	0.000	21:26	
Air Blank	0.000	21:27	
Control Test	0.081	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FEBLES RODRIGUEZ

DATE/TIME OF TEST: 06/05/2024 19:55:23

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.081	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.193	19:50
Breath Volume	1.550	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.182	19:53
Breath Volume	1.964	19:53
Air Blank	0.000	19:54
Control Test	0.081	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE, RONALD, W
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKENZIE, RONALD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: SABELLA

DATE/TIME OF TEST: 06/05/2024 20:28:58

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.108	20:24
Breath Volume	1.695	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.107	20:27
Breath Volume	1.425	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AYACHI

DATE/TIME OF TEST: 06/06/2024 02:12:13

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.098	02:07
Breath Volume	1.753	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.095	02:10
Breath Volume	1.503	02:10
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/06/2024 04:59:47

Diagnostics Check	OK	04:48	
Air Blank	0.000	04:48	
Control Test	0.082	04:48	
Air Blank	0.000	04:49	
Subject Sample	# 1 0.100	04:52	Volume Not Met
Breath Volume	0.121	04:52	
Air Blank	0.000	04:53	
Air Blank	0.000	04:55	
Subject Sample	# 2 0.090	04:58	Volume Not Met
Breath Volume	0.164	04:58	
Air Blank	0.000	04:58	
Control Test	0.080	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 06/06/2024 23:21:57

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.000	23:17
Breath Volume	1.785	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.000	23:20
Breath Volume	1.230	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FARINA,JOSEPH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: QUARANTA

DATE/TIME OF TEST: 06/07/2024 03:38:37

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANISZEWSKI,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 06/07/2024 23:29:27

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.081	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.198	23:24
Breath Volume	2.128	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.195	23:27
Breath Volume	2.089	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FARINA,JOSEPH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/08/2024 00:17:15

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.081	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/08/2024 02:26:52

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1 0.154	02:19	Volume Not Met
Breath Volume	0.031	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:25	No Sample Provided
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/08/2024 02:57:11

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1 0.162	02:49	Volume Not Met
Breath Volume	0.066	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 2 0.169	02:55	Volume Not Met
Breath Volume	0.066	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 06/13/2024 07:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BUHRTS

DATE/TIME OF TEST: 06/08/2024 04:28:07

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.217	04:23
Breath Volume	1.234	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.199	04:26
Breath Volume	1.414	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: MCCALLISTER

DATE/TIME OF TEST: 06/08/2024 20:28:37

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.019	20:24
Breath Volume	1.796	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.019	20:27
Breath Volume	2.085	20:27
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,KRISTINA
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PIZZINI

DATE/TIME OF TEST: 06/09/2024 20:58:57

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.081	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.306	20:54
Breath Volume	2.093	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:56
Subject Sample # 2	0.291	20:57
Breath Volume	1.867	20:57
Air Blank	0.000	20:57
Control Test	0.080	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,CLAYTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: RAMIREZ OSORIO

DATE/TIME OF TEST: 06/11/2024 23:19:43

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	No Sample Provided
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2	23:18	No Sample Provided
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.081	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: PENTECOST

DATE/TIME OF TEST: 06/12/2024 10:32:44

Diagnostics Check	OK	10:26
Air Blank	0.000	10:27
Control Test	0.081	10:27
Air Blank	0.000	10:27
Subject Sample # 1	0.000	10:28
Breath Volume	3.019	10:28
Air Blank	0.000	10:28
Air Blank	0.000	10:30
Subject Sample # 2	0.000	10:31
Breath Volume	2.914	10:31
Air Blank	0.000	10:31
Control Test	0.082	10:32
Air Blank	0.000	10:32
Diagnostics Check	OK	10:32

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: LESHORE

DATE/TIME OF TEST: 06/13/2024 01:06:36

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.085	01:01
Breath Volume	1.242	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.086	01:05
Breath Volume	1.250	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/13/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: DODSON

DATE/TIME OF TEST: 06/13/2024 21:45:21

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.161	21:40
Breath Volume	2.394	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.156	21:43
Breath Volume	1.546	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SHIPP,JUSTIS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KEANE

DATE/TIME OF TEST: 06/14/2024 00:24:18

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.081	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 06/14/2024 02:28:59

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.080	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BEEM

DATE/TIME OF TEST: 06/14/2024 03:37:12

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.000	03:32
Breath Volume	1.589	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:35
Breath Volume	1.253	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ANDERSEN HOOPER

DATE/TIME OF TEST: 06/15/2024 02:35:55

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.184	02:30
Breath Volume	2.710	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:32
Subject Sample # 2	0.178	02:34
Breath Volume	1.406	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,ADAM
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: LANCASHIRE

DATE/TIME OF TEST: 06/15/2024 04:40:30

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:38	
Control Test	0.081	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: AVEY,J
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: KOCIELKO

DATE/TIME OF TEST: 06/15/2024 06:50:51

Diagnostics Check	OK	06:44
Air Blank	0.000	06:45
Control Test	0.081	06:45
Air Blank	0.000	06:45
Subject Sample # 1	0.011	06:46
Breath Volume	3.992	06:46
Air Blank	0.000	06:46
Air Blank	0.000	06:48
Subject Sample # 2	0.010	06:49
Breath Volume	3.285	06:49
Air Blank	0.000	06:49
Control Test	0.082	06:50
Air Blank	0.000	06:50
Diagnostics Check	OK	06:50

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PROVAU

DATE/TIME OF TEST: 06/15/2024 23:11:55

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SISTI,JORDAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: MAISTO

DATE/TIME OF TEST: 06/16/2024 00:14:18

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.057	00:09
Breath Volume	2.636	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.059	00:12
Breath Volume	2.714	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 06/16/2024 00:45:50

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:44	
Control Test	0.081	00:45	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON, C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/16/2024 03:29:06

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.087	03:24
Breath Volume	2.710	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.085	03:27
Breath Volume	1.890	03:27
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 06/16/2024 05:17:14

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.123	05:12
Breath Volume	2.402	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.117	05:15
Breath Volume	2.750	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HUTCHINSON, SHANE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: CARRASQUILLO-HERRERA

DATE/TIME OF TEST: 06/16/2024 19:35:58

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.081	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:29	No Sample Provided
Breath Volume	0.000	19:29	
Air Blank	0.000	19:30	
Air Blank	0.000	19:32	
Subject Sample	# 2	19:34	Subject Test Refused
Breath Volume	0.000	19:34	
Air Blank	0.000	19:35	
Control Test	0.081	19:35	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS, KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/16/2024 19:57:10

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.081	19:51
Air Blank	0.000	19:51
Subject Sample # 1	0.000	19:52
Breath Volume	2.332	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.000	19:55
Breath Volume	2.386	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 07/03/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/17/2024 10:38:41

Diagnostics Check	OK	10:32
Air Blank	0.000	10:32
Control Test	0.082	10:33
Air Blank	0.000	10:33
Subject Sample # 1	0.000	10:34
Breath Volume	1.582	10:34
Air Blank	0.000	10:34
Air Blank	0.000	10:36
Subject Sample # 2	0.000	10:37
Breath Volume	1.292	10:37
Air Blank	0.000	10:37
Control Test	0.080	10:38
Air Blank	0.000	10:38
Diagnostics Check	OK	10:38

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/18/2024 08:49:18

Diagnostics Check	OK	08:42
Air Blank	0.000	08:43
Control Test	0.081	08:43
Air Blank	0.000	08:43
Subject Sample # 1	0.000	08:44
Breath Volume	3.253	08:44
Air Blank	0.000	08:45
Air Blank	0.000	08:47
Subject Sample # 2	0.000	08:47
Breath Volume	2.757	08:47
Air Blank	0.000	08:48
Control Test	0.081	08:48
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/18/2024 17:13:15

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.081	17:07
Air Blank	0.000	17:08
Subject Sample # 1	0.000	17:08
Breath Volume	2.062	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:11
Breath Volume	1.105	17:11
Air Blank	0.000	17:12
Control Test	0.081	17:12
Air Blank	0.000	17:13
Diagnostics Check	OK	17:13

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 06/18/2024 20:55:00

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.054	20:50
Breath Volume	1.238	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.059	20:53
Breath Volume	1.289	20:53
Air Blank	0.000	20:54
Control Test	0.081	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:55

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: O'LEARY

DATE/TIME OF TEST: 06/18/2024 23:17:03

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.218	23:12
Breath Volume	2.312	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.206	23:15
Breath Volume	2.746	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:17

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: DEJOHN

DATE/TIME OF TEST: 06/20/2024 01:28:10

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.082	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:21	No Sample Provided
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:26	No Sample Provided
Breath Volume	0.000	01:26	
Air Blank	0.000	01:27	
Control Test	0.082	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 07/03/2024 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BRENTS

DATE/TIME OF TEST: 06/21/2024 02:42:22

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.145	02:37
Breath Volume	1.773	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.135	02:40
Breath Volume	1.941	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: DEBOE

DATE/TIME OF TEST: 06/21/2024 21:11:06

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
Control Test	0.081	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:10	
Control Test	0.080	21:10	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 06/21/2024 22:50:59

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.226	22:46
Breath Volume	2.078	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.236	22:49
Breath Volume	1.871	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III, KENNETH, M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: NIYAZOV, DIMASH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: VAILLANT

DATE/TIME OF TEST: 06/22/2024 01:00:12

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.088	00:55
Breath Volume	1.296	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.082	00:58
Breath Volume	1.441	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/22/2024 04:39:04

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1 0.154	04:31	Volume Not Met
Breath Volume	0.140	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2 0.140	04:37	Volume Not Met
Breath Volume	0.496	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: WESTINGHOUSE

DATE/TIME OF TEST: 06/22/2024 12:10:47

Diagnostics Check	OK	12:03
Air Blank	0.000	12:03
Control Test	0.080	12:04
Air Blank	0.000	12:04
Subject Sample # 1	0.000	12:05
Breath Volume	1.175	12:05
Air Blank	0.000	12:06
Air Blank	0.000	12:08
Subject Sample # 2	0.000	12:09
Breath Volume	1.300	12:09
Air Blank	0.000	12:09
Control Test	0.081	12:10
Air Blank	0.000	12:10
Diagnostics Check	OK	12:10

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD, MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: BRESKOW

DATE/TIME OF TEST: 06/25/2024 00:51:51

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:45	No Sample Provided
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Air Blank	0.000	00:47	
Subject Sample	# 2	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Diagnostics Check	OK	00:51	

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MCKEEVER

DATE/TIME OF TEST: 06/25/2024 20:15:38

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.081	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:08	No Sample Provided
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Air Blank	0.000	20:11	
Subject Sample	# 2	20:14	No Sample Provided
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.081	20:15	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 06/26/2024 02:38:03

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.082	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.082	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WONG

DATE/TIME OF TEST: 06/28/2024 01:07:21

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.125	01:02
Breath Volume	1.578	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:04
Subject Sample # 2	0.126	01:05
Breath Volume	1.960	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 06/28/2024 20:22:55

Diagnostics Check	OK	20:16
Air Blank	0.000	20:16
Control Test	0.081	20:17
Air Blank	0.000	20:17
Subject Sample # 1	0.238	20:18
Breath Volume	1.421	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample # 2	0.244	20:21
Breath Volume	1.445	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: OSTMAN

DATE/TIME OF TEST: 06/28/2024 23:58:50

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.204	23:54
Breath Volume	2.023	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.209	23:57
Breath Volume	2.226	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/29/2024 00:44:02

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.125	00:39
Breath Volume	1.289	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:41
Subject Sample # 2	0.120	00:42
Breath Volume	1.335	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SWAIN

DATE/TIME OF TEST: 06/29/2024 02:28:25

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.152	02:23
Breath Volume	1.722	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.148	02:26
Breath Volume	1.503	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 07/03/2024 11:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: CORNELL

DATE/TIME OF TEST: 06/29/2024 17:14:57

Diagnostics Check	OK	17:08
Air Blank	0.000	17:09
Control Test	0.081	17:09
Air Blank	0.000	17:10
Subject Sample # 1	0.000	17:10
Breath Volume	2.441	17:10
Air Blank	0.000	17:11
Air Blank	0.000	17:13
Subject Sample # 2	0.000	17:13
Breath Volume	2.187	17:13
Air Blank	0.000	17:14
Control Test	0.081	17:14
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

Data Download Date/Time: 07/03/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/29/2024 23:19:38

Diagnostics Check	OK	23:07	
Air Blank	0.000	23:07	
Control Test	0.081	23:08	
Air Blank	0.000	23:08	
Subject Sample	# 1 0.118	23:11	Volume Not Met
Breath Volume	0.921	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.147	23:14	
Breath Volume	1.531	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 3 0.147	23:18	
Breath Volume	1.882	23:18	
Air Blank	0.000	23:18	
Control Test	0.080	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 07/03/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 06/29/2024 23:39:08

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.081	23:28	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Air Blank	0.000	23:34	
Subject Sample	# 2	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:38	
Control Test	0.081	23:38	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:39	

Data Download Date/Time: 07/03/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST. PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: TINGLING

DATE/TIME OF TEST: 06/30/2024 04:00:37

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.144	03:55
Breath Volume	1.296	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.137	03:59
Breath Volume	1.738	03:59
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/03/2024 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,LANDON,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 06/01/2024 02:42:05

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.156	02:33
Breath Volume	1.445	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.178	02:36
Breath Volume	1.640	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 3	0.170	02:40
Breath Volume	1.937	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/25/2024 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HARRIS, LODRICK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/02/2024 21:12:08

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.081	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:05	No Sample Provided
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 2	21:10	No Sample Provided
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 06/25/2024 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HARRIS, LODRICK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 06/02/2024 21:21:20

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.081	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.173	21:16
Breath Volume	1.378	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:18
Subject Sample # 2	0.168	21:19
Breath Volume	1.304	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 06/25/2024 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN, D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 06/05/2024 02:02:02

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.174	01:56
Breath Volume	1.890	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.172	02:00
Breath Volume	2.367	02:00
Air Blank	0.000	02:01
Control Test	0.082	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:02

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KING,CASEY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SCRIVEN

DATE/TIME OF TEST: 06/05/2024 03:44:16

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.116	03:43	
Breath Volume	1.996	03:43	
Air Blank		03:43	RFI Detect
Air Blank		03:43	RFI Detect
Air Blank	0.000	03:44	

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KING,CASEY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SCRIVEN

DATE/TIME OF TEST: 06/05/2024 03:52:26

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.118	03:47
Breath Volume	1.988	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.118	03:50
Breath Volume	2.117	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWLER,ADDISON,W
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PRADO,ALEXIS
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/08/2024 23:21:58

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.201	23:17
Breath Volume	1.800	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.190	23:20
Breath Volume	1.816	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: SYMONS

DATE/TIME OF TEST: 06/14/2024 19:20:57

Diagnostics Check	OK	19:14
Air Blank	0.000	19:15
Control Test	0.082	19:15
Air Blank	0.000	19:16
Subject Sample # 1	0.000	19:16
Breath Volume	2.082	19:16
Air Blank	0.000	19:17
Air Blank	0.000	19:19
Subject Sample # 2	0.000	19:19
Breath Volume	1.527	19:19
Air Blank	0.000	19:20
Control Test	0.082	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ELLINGTON,ADAM
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 06/16/2024 03:48:03

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.083	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.188	03:43
Breath Volume	2.851	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.191	03:46
Breath Volume	2.046	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/21/2024 02:19:05

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.174	02:14
Breath Volume	1.117	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.171	02:17
Breath Volume	2.640	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WARREN,ZACHARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 06/21/2024 22:44:37

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.122	22:39
Breath Volume	2.312	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.117	22:43
Breath Volume	2.085	22:43
Air Blank	0.000	22:43
Control Test	0.082	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 06/25/2024 12:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPELL,E,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,E
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CROWDER

DATE/TIME OF TEST: 06/22/2024 00:37:15

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.201	00:32
Breath Volume	3.269	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.195	00:35
Breath Volume	3.328	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 06/25/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWLER, ADDISON, W
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BENVENISTE, JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 06/23/2024 20:18:49

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.083	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.218	20:13
Breath Volume	1.250	20:13
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.220	20:17
Breath Volume	2.660	20:17
Air Blank	0.000	20:17
Control Test	0.081	20:18
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 06/25/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/26/2024 04:26:07

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.169	04:20
Breath Volume	1.355	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.179	04:24
Breath Volume	2.269	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 07/01/2024 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WARREN,ZACHARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33021080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: STAHL

DATE/TIME OF TEST: 06/28/2024 02:53:02

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.094	02:48
Breath Volume	2.324	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.093	02:51
Breath Volume	1.953	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/01/2024 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: PIRONE

DATE/TIME OF TEST: 06/23/2024 01:08:58

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.199	01:04
Breath Volume	1.734	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.204	01:07
Breath Volume	2.101	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ARRESEIGOR

DATE/TIME OF TEST: 06/25/2024 02:15:42

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.135	02:11
Breath Volume	1.535	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.131	02:14
Breath Volume	1.484	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SMITH,NICKOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HUPPE

DATE/TIME OF TEST: 06/25/2024 22:33:12

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.134	22:28
Breath Volume	1.101	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.134	22:31
Breath Volume	1.433	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: EVELER,JONATHAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 06/25/2024 23:02:51

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.237	22:58
Breath Volume	2.250	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.220	23:01
Breath Volume	3.164	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,NICKOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/29/2024 22:55:46

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.080	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1 0.122	22:48	Volume Not Met
Breath Volume	0.324	22:48	
Air Blank	0.000	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 2 0.130	22:54	Volume Not Met
Breath Volume	1.035	22:54	
Air Blank	0.000	22:54	
Control Test	0.079	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,NICKOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GRAY,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/29/2024 23:10:09

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.151	23:04
Breath Volume	3.039	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.155	23:08
Breath Volume	1.792	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 07/27/2024 18:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CLIFFORD,ALAIN,R

BREATH TEST OPERATOR AGENCY: Eglin AFB PD

ARRESTING OFFICER: POOLE,TONYA

ARRESTING OFFICER AGENCY: Eglin AFB PD

LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402648278

DRY GAS EXPIRY DATE: 01/27/2026

VIOLATION CODE: ADMIN

ARREST TIME: 14:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: MCWHIRTER

DATE/TIME OF TEST: 06/20/2024 15:24:18

Diagnostics Check	OK	15:18
Air Blank	0.000	15:18
Control Test	0.081	15:19
Air Blank	0.000	15:19
Subject Sample # 1	0.019	15:19
Breath Volume	3.640	15:19
Air Blank	0.000	15:20
Air Blank	0.000	15:22
Subject Sample # 2	0.018	15:22
Breath Volume	3.140	15:22
Air Blank	0.000	15:23
Control Test	0.080	15:23
Air Blank	0.000	15:24
Diagnostics Check	OK	15:24

Data Download Date/Time: 07/26/2024 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HAYS

DATE/TIME OF TEST: 06/16/2024 03:35:58

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.133	03:31
Breath Volume	1.375	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.120	03:34
Breath Volume	1.898	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 06/30/2024 23:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LAVENTURE,KAITLYN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 06/23/2024 17:37:05

Diagnostics Check	OK	17:27	
Air Blank	0.000	17:27	
Control Test	0.081	17:28	
Air Blank	0.000	17:28	
Subject Sample	# 1 0.048	17:30	
Breath Volume	1.160	17:30	
Air Blank	0.000	17:31	
Air Blank	0.000	17:33	
Subject Sample	# 2	17:35	Subject Test Refused
Breath Volume	0.636	17:35	
Air Blank	0.000	17:36	
Control Test	0.081	17:36	
Air Blank	0.000	17:37	
Diagnostics Check	OK	17:37	

Data Download Date/Time: 06/30/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EVANS,MAURINE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: HAN

DATE/TIME OF TEST: 06/30/2024 09:13:30

Diagnostics Check	OK	09:01	
Air Blank	0.000	09:02	
Control Test	0.082	09:02	
Air Blank	0.000	09:03	
Subject Sample	# 1 0.119	09:06	Volume Not Met
Breath Volume	0.273	09:06	
Air Blank	0.000	09:06	
Air Blank	0.000	09:08	
Subject Sample	# 2 0.153	09:11	Volume Not Met
Breath Volume	0.156	09:11	
Air Blank	0.000	09:12	
Control Test	0.079	09:12	
Air Blank	0.000	09:13	
Diagnostics Check	OK	09:13	

Data Download Date/Time: 06/30/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EVANS,MAURINE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: HAN

DATE/TIME OF TEST: 06/30/2024 09:27:53

Diagnostics Check	OK	09:16	
Air Blank	0.000	09:16	
Control Test	0.081	09:17	
Air Blank	0.000	09:17	
Subject Sample	# 1 0.139	09:20	Volume Not Met
Breath Volume	0.339	09:20	
Air Blank	0.000	09:21	
Air Blank	0.000	09:23	
Subject Sample	# 2 0.166	09:26	Volume Not Met
Breath Volume	0.015	09:26	
Air Blank	0.000	09:26	
Control Test	0.079	09:27	
Air Blank	0.000	09:27	
Diagnostics Check	OK	09:27	

Data Download Date/Time: 06/30/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,RICHARD,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: EVANS,MAURINE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:43

SUBJECT'S LAST NAME: HAN

DATE/TIME OF TEST: 06/30/2024 10:20:44

Diagnostics Check	OK	10:10	
Air Blank	0.000	10:10	
Control Test	0.080	10:10	
Air Blank	0.000	10:11	
Subject Sample # 1	0.139	10:13	
Breath Volume	1.667	10:13	
Air Blank	0.000	10:14	
Air Blank	0.000	10:15	
Subject Sample # 2	0.139	10:19	Volume Not Met
Breath Volume	0.117	10:19	
Air Blank		10:19	RFI Detect
Air Blank	0.000	10:19	
Control Test	0.079	10:20	
Air Blank	0.000	10:20	
Diagnostics Check	OK	10:20	

Data Download Date/Time: 06/30/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001157

INSTRUMENT REGISTERED TO: POLK CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 06/22/2024 11:52:47

Diagnostics Check	OK	11:46
Air Blank	0.000	11:47
Control Test	0.078	11:47
Air Blank	0.000	11:47
Subject Sample # 1	0.014	11:48
Breath Volume	3.507	11:48
Air Blank	0.000	11:48
Air Blank	0.000	11:50
Subject Sample # 2	0.012	11:51
Breath Volume	3.675	11:51
Air Blank	0.000	11:51
Control Test	0.077	11:52
Air Blank	0.000	11:52
Diagnostics Check	OK	11:52

Data Download Date/Time: 08/13/2024 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 23:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HARNARAIN

DATE/TIME OF TEST: 06/01/2024 00:37:31

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.082	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.157	00:32
Breath Volume	1.964	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.147	00:35
Breath Volume	2.332	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 06/26/2024 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BUSHONG,NICOLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 04:14

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MORALES CARDENAS

DATE/TIME OF TEST: 06/01/2024 05:10:36

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.082	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.062	05:06
Breath Volume	2.277	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.056	05:09
Breath Volume	2.152	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 06/26/2024 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BERRIOS,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 06/08/2024 00:35:13

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.081	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ZOLLWEG,B
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BITTERMAN

DATE/TIME OF TEST: 06/09/2024 22:54:56

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.081	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 21:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LICATA

DATE/TIME OF TEST: 06/11/2024 22:15:58

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:13	
Control Test	0.081	22:13	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:14	Subject Test Refused
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Control Test	0.082	22:15	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BARKER,D
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DYE

DATE/TIME OF TEST: 06/13/2024 22:28:16

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:26	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.081	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: MENCARELLI,D
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEABLER

DATE/TIME OF TEST: 06/14/2024 02:53:06

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.081	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 02:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MORRISSEY

DATE/TIME OF TEST: 06/15/2024 03:06:32

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.082	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.147	03:01
Breath Volume	2.300	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.145	03:04
Breath Volume	2.082	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 01:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SCORSUNE

DATE/TIME OF TEST: 06/16/2024 02:13:11

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.137	02:08
Breath Volume	1.105	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.148	02:11
Breath Volume	3.796	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LEE,H
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 06/24/2024 05:37:22

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.065	05:32
Breath Volume	3.296	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.065	05:35
Breath Volume	2.378	05:35
Air Blank	0.000	05:36
Control Test	0.081	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 06/26/2024 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: PEREZ,F
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 05/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GUALLE CHICO

DATE/TIME OF TEST: 06/08/2024 01:43:30

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.174	01:38
Breath Volume	1.347	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.166	01:41
Breath Volume	1.773	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/10/2024 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: ZEPEDA

DATE/TIME OF TEST: 06/21/2024 05:28:17

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.244	05:23
Breath Volume	1.792	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.238	05:26
Breath Volume	2.640	05:26
Air Blank	0.000	05:27
Control Test	0.079	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 07/31/2024 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RAMIREZ,CRISTIAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 06/21/2024 05:52:40

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:50	
Control Test	0.079	05:50	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:51	
Control Test	0.079	05:52	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 07/31/2024 16:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 06/02/2024 01:39:30

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.082	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.153	01:34
Breath Volume	1.402	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.150	01:38
Breath Volume	1.175	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 07/22/2024 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GULOTTA,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33922080A1
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MALDONADOLEON

DATE/TIME OF TEST: 06/02/2024 00:11:43

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.157	00:06
Breath Volume	1.683	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.173	00:10
Breath Volume	1.199	00:10
Air Blank	0.000	00:10
Control Test	0.077	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 06/13/2024 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,WILLIAM,W
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: MILLER,WILLIAM
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 05/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BOLDREGHINI

DATE/TIME OF TEST: 06/13/2024 23:24:58

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.197	23:19
Breath Volume	2.125	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.183	23:23
Breath Volume	2.539	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 06/18/2024 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RODRIGUEZ,LARRY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KERVENS

DATE/TIME OF TEST: 06/29/2024 02:11:50

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.203	02:05
Breath Volume	1.742	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.219	02:10
Breath Volume	1.785	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 07/01/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FACCHINI

DATE/TIME OF TEST: 06/01/2024 00:55:29

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.210	00:50
Breath Volume	3.195	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.211	00:53
Breath Volume	2.996	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KERSEY

DATE/TIME OF TEST: 06/01/2024 01:13:36

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1 0.112	01:09	
Breath Volume	2.703	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 2	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GARDINER,PAUL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SANTIAGO ARZOLA

DATE/TIME OF TEST: 06/01/2024 02:46:44

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:42
Breath Volume	3.128	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.000	02:45
Breath Volume	2.515	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: TERON

DATE/TIME OF TEST: 06/01/2024 03:18:28

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: CARRERA

DATE/TIME OF TEST: 06/01/2024 20:52:21

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.080	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:48	No Sample Provided
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 2	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.080	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER, JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CASTRO PEREZ

DATE/TIME OF TEST: 06/01/2024 23:02:03

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.110	22:57
Breath Volume	2.765	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.107	23:00
Breath Volume	2.464	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KNAPP, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SUSHIL

DATE/TIME OF TEST: 06/02/2024 02:08:26

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.000	02:04
Breath Volume	2.804	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.000	02:07
Breath Volume	2.941	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 06/02/2024 02:27:15

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.239	02:22
Breath Volume	3.742	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.235	02:25
Breath Volume	4.003	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 06/02/2024 02:40:17

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.079	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.000	02:36
Breath Volume	3.351	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.000	02:38
Breath Volume	3.683	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/02/2024 19:20:40

Diagnostics Check	OK	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:15
Subject Sample # 1	0.400	19:15
Breath Volume	2.050	19:15
Air Blank	0.000	19:16
Air Blank	0.000	19:18
Subject Sample # 2	0.410	19:19
Breath Volume	1.433	19:19
Air Blank	0.000	19:19
Control Test	0.078	19:20
Air Blank	0.000	19:20
Diagnostics Check	OK	19:20

Data Download Date/Time: 06/26/2024 10:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CONSUEGRA-PEREZ,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GUEVARA HERNANDEZ

DATE/TIME OF TEST: 06/02/2024 19:29:53

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.079	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.170	19:25
Breath Volume	1.234	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.174	19:28
Breath Volume	1.761	19:28
Air Blank	0.000	19:28
Control Test	0.079	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DOMINGUEZ POLANCO

DATE/TIME OF TEST: 06/03/2024 02:34:01

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.259	02:29
Breath Volume	1.367	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.243	02:32
Breath Volume	1.503	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 06/03/2024 03:24:52

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.213	03:20
Breath Volume	2.437	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.212	03:23
Breath Volume	2.750	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,MOSES,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BOYSEN,DERRIDE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: US TOJIN

DATE/TIME OF TEST: 06/04/2024 03:44:03

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.101	03:39
Breath Volume	1.414	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.099	03:42
Breath Volume	1.828	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:44

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MYNAUGH, JOSEPH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GOTTSCHLING

DATE/TIME OF TEST: 06/06/2024 04:18:53

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1 0.220	04:13	
Breath Volume	1.464	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:17	Subject Test Refused
Breath Volume	0.777	04:17	
Air Blank	0.000	04:17	
Control Test	0.079	04:18	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER, JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MATA VALENCIA

DATE/TIME OF TEST: 06/07/2024 03:31:42

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.126	03:27
Breath Volume	1.410	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.125	03:30
Breath Volume	2.988	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CEDILLO BERNAL

DATE/TIME OF TEST: 06/08/2024 02:42:52

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.189	02:38
Breath Volume	2.898	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.177	02:41
Breath Volume	2.289	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GAAR

DATE/TIME OF TEST: 06/08/2024 03:36:04

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.086	03:31
Breath Volume	2.511	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.088	03:34
Breath Volume	3.605	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 06/09/2024 01:59:18

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.138	01:54
Breath Volume	1.957	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.137	01:57
Breath Volume	1.648	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FUGIEL,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: HERNANDEZ CAPELES

DATE/TIME OF TEST: 06/09/2024 03:11:29

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.118	03:06
Breath Volume	2.511	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.113	03:09
Breath Volume	2.472	03:09
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: OLVERA SANCHEZ

DATE/TIME OF TEST: 06/09/2024 23:33:55

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.140	23:29
Breath Volume	1.343	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.132	23:32
Breath Volume	2.308	23:32
Air Blank	0.000	23:33
Control Test	0.078	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 06/10/2024 04:11:18

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 06/26/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, DAVID, A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CLEMENTS, TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 06/10/2024 05:05:00

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.079	04:56	
Air Blank	0.000	04:57	
Subject Sample	# 1	05:00	No Sample Provided
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 2	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: FREESE

DATE/TIME OF TEST: 06/10/2024 07:00:21

Diagnostics Check	OK	06:54
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Subject Sample # 1	0.147	06:55
Breath Volume	2.691	06:55
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.141	06:58
Breath Volume	1.429	06:58
Air Blank	0.000	06:59
Control Test	0.078	06:59
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/10/2024 18:59:40

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.080	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:55
Breath Volume	1.675	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:58
Breath Volume	1.335	18:58
Air Blank	0.000	18:58
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 06/10/2024 20:15:39

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.080	20:12	
Air Blank	0.000	20:13	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.343	20:14	
Air Blank	0.000	20:14	
Control Test	0.080	20:15	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLAUSNER,JAMIE,N
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MORALES,HECTOR
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: A325462590060

DATE/TIME OF TEST: 06/10/2024 20:32:10

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.080	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.167	20:30	
Air Blank	0.000	20:31	
Control Test	0.079	20:31	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:32	

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FRAZIN,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SERVELLIERI PERREIRA

DATE/TIME OF TEST: 06/12/2024 02:15:15

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.110	02:10
Breath Volume	2.406	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.101	02:13
Breath Volume	1.941	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HARRINGTON,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: PICKENS

DATE/TIME OF TEST: 06/12/2024 20:40:57

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.118	20:36
Breath Volume	1.761	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.112	20:39
Breath Volume	1.605	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: DIETRICH

DATE/TIME OF TEST: 06/13/2024 01:13:53

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.087	01:09
Breath Volume	2.558	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.086	01:12
Breath Volume	2.574	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINEZ,FRANCISCO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: REGO

DATE/TIME OF TEST: 06/13/2024 20:20:13

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:17	
Control Test	0.080	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:19	
Control Test	0.079	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/14/2024 00:22:14

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.111	00:17
Breath Volume	1.878	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.115	00:20
Breath Volume	1.738	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,KWESI,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MAHON,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 06/15/2024 01:21:17

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.086	01:16
Breath Volume	1.640	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.089	01:19
Breath Volume	2.964	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER, JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ELIAS-YOUNG

DATE/TIME OF TEST: 06/15/2024 03:47:31

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.211	03:39
Breath Volume	1.707	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.189	03:42
Breath Volume	1.250	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 3	0.200	03:45
Breath Volume	1.863	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: BUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/15/2024 20:32:02

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:26
Air Blank	0.000	20:26
Subject Sample # 1	0.225	20:27
Breath Volume	1.664	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.223	20:30
Breath Volume	1.582	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN, TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CLAVETTE

DATE/TIME OF TEST: 06/15/2024 23:16:04

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.170	23:11
Breath Volume	1.824	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.167	23:14
Breath Volume	2.011	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:16

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CRUZ PINEDA

DATE/TIME OF TEST: 06/16/2024 03:24:03

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.123	03:19
Breath Volume	3.199	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.127	03:22
Breath Volume	1.800	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JILLIAN,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CARROLL,JILLIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: DARR

DATE/TIME OF TEST: 06/16/2024 08:39:58

Diagnostics Check	OK	08:33
Air Blank	0.000	08:34
Control Test	0.080	08:34
Air Blank	0.000	08:34
Subject Sample # 1	0.129	08:35
Breath Volume	1.988	08:35
Air Blank	0.000	08:36
Air Blank	0.000	08:37
Subject Sample # 2	0.131	08:38
Breath Volume	1.312	08:38
Air Blank	0.000	08:39
Control Test	0.079	08:39
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SERRES,DEVAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CORONADO

DATE/TIME OF TEST: 06/17/2024 00:48:40

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.080	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.134	00:43
Breath Volume	1.503	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.141	00:47
Breath Volume	1.839	00:47
Air Blank	0.000	00:47
Control Test	0.079	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HERNANDEZ DE LA CRUZ

DATE/TIME OF TEST: 06/19/2024 00:53:38

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.101	00:49
Breath Volume	1.597	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.101	00:52
Breath Volume	1.421	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KNAPP, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/20/2024 02:59:11

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.155	02:54
Breath Volume	1.718	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.149	02:57
Breath Volume	3.585	02:57
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 06/22/2024 03:48:58

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.164	03:44
Breath Volume	1.445	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.171	03:47
Breath Volume	1.199	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HARRINGTON,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LUKE-SAID

DATE/TIME OF TEST: 06/22/2024 04:14:00

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.136	04:09
Breath Volume	1.546	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.139	04:12
Breath Volume	1.136	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:14

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: COLOMINA

DATE/TIME OF TEST: 06/22/2024 04:22:39

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.079	04:18	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:21	
Control Test	0.079	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 06/26/2024 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: DAMON,RUSSELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/02/2024 03:31:33

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.075	03:26
Breath Volume	1.816	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.076	03:29
Breath Volume	1.910	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/06/2024 05:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY, PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TORVESTAD

DATE/TIME OF TEST: 06/02/2024 02:03:44

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.266	01:58
Breath Volume	2.222	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.282	02:02
Breath Volume	2.613	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 06/25/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/15/2024 03:10:59

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1 0.000	03:04	Volume Not Met
Breath Volume	0.042	03:04	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.077	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 06/25/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/16/2024 02:49:10

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.275	02:44
Breath Volume	3.113	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.257	02:47
Breath Volume	2.433	02:47
Air Blank	0.000	02:48
Control Test	0.076	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 06/25/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: MOWBARY,MARK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BELISLE

DATE/TIME OF TEST: 06/18/2024 01:28:20

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.077	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.130	01:23
Breath Volume	1.746	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.133	01:26
Breath Volume	1.359	01:26
Air Blank	0.000	01:27
Control Test	0.075	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 06/25/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: HOMEN, LEANNA
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/20/2024 21:09:44

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:57	
Control Test	0.078	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	21:01	No Sample Provided
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Air Blank	0.000	21:03	
Subject Sample	# 2	21:04	
Breath Volume	1.843	21:04	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 3	21:08	
Breath Volume	1.828	21:08	
Air Blank	0.000	21:08	
Control Test	0.077	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 06/25/2024 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY, PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 06/26/2024 00:12:23

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.078	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.175	00:07
Breath Volume	2.460	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.177	00:10
Breath Volume	2.253	00:10
Air Blank	0.000	00:11
Control Test	0.078	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/22/2024 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 06/27/2024 03:17:56

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.078	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.153	03:13
Breath Volume	2.445	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.154	03:16
Breath Volume	2.218	03:16
Air Blank	0.000	03:17
Control Test	0.076	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 07/22/2024 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 06/30/2024 00:36:26

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.077	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.151	00:25	Volume Not Met
Breath Volume	0.632	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:28	
Subject Sample	# 2 0.166	00:29	
Breath Volume	1.453	00:29	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 3 0.140	00:34	Volume Not Met
Breath Volume	0.402	00:34	
Air Blank	0.000	00:35	
Control Test	0.076	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 07/22/2024 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEGER,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: FEGER,JOSHUA
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477282
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 06/30/2024 15:27:00

Diagnostics Check	OK	15:21
Air Blank	0.000	15:21
Control Test	0.078	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.000	15:22
Breath Volume	1.933	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.000	15:25
Breath Volume	2.070	15:25
Air Blank	0.000	15:26
Control Test	0.077	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:27

Data Download Date/Time: 07/22/2024 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,JESSE,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: LANGER,MAX
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 06/04/2024 15:10:34

Diagnostics Check	OK	15:03
Air Blank	0.000	15:03
Control Test	0.081	15:04
Air Blank	0.000	15:04
Subject Sample # 1	0.253	15:05
Breath Volume	1.609	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.254	15:09
Breath Volume	1.128	15:09
Air Blank	0.000	15:09
Control Test	0.081	15:09
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

Data Download Date/Time: 06/28/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEISLAHN, ERIC
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ANGLIM

DATE/TIME OF TEST: 06/12/2024 04:15:22

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:14	Slope Not Met
Breath Volume	3.269	04:14	
Air Blank	0.000	04:14	
Control Test	0.081	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 06/28/2024 15:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOVANOVICH, THOMAS,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MEISLAHN, ERIC
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ANGLIM

DATE/TIME OF TEST: 06/12/2024 04:24:17

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.000	04:19
Breath Volume	3.441	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:22
Breath Volume	3.390	04:22
Air Blank	0.000	04:23
Control Test	0.082	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 06/28/2024 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOFE,DEREK,T
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/18/2024 16:06:40

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.081	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.316	15:57
Breath Volume	1.164	15:57
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.337	16:01
Breath Volume	1.785	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 3	0.316	16:05
Breath Volume	1.550	16:05
Air Blank	0.000	16:05
Control Test	0.077	16:06
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 06/28/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANCHESTER,KAYLIE,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: VARNER,ETHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ESPRE

DATE/TIME OF TEST: 06/21/2024 02:58:16

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.081	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.179	02:53
Breath Volume	3.472	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.177	02:56
Breath Volume	2.304	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 06/28/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHNIBBEN,BRIAN,K
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: DUNN,ROGERS
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 06/23/2024 03:29:00

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.058	03:24
Breath Volume	3.761	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.055	03:27
Breath Volume	2.996	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/28/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,LANDON,L
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: INDAHL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:29

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 06/26/2024 09:58:29

Diagnostics Check	OK	09:50
Air Blank	0.000	09:50
Control Test	0.081	09:50
Air Blank	0.000	09:51
Subject Sample # 1	0.206	09:52
Breath Volume	1.359	09:52
Air Blank	0.000	09:52
Air Blank	0.000	09:54
Subject Sample # 2	0.208	09:56
Breath Volume	1.875	09:56
Air Blank	0.000	09:57
Control Test	0.082	09:57
Air Blank	0.000	09:58
Diagnostics Check	OK	09:58

Data Download Date/Time: 06/28/2024 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAFFNEY,CASEY,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MENDOZA,JONOTHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: GRANATO

DATE/TIME OF TEST: 06/09/2024 21:17:07

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.082	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.150	21:12
Breath Volume	2.078	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.141	21:15
Breath Volume	1.949	21:15
Air Blank	0.000	21:16
Control Test	0.082	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 06/12/2024 16:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JONATHAN,F
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MENDOZA,JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 06/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: QUEEN

DATE/TIME OF TEST: 06/22/2024 22:59:33

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.083	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.220	22:54
Breath Volume	1.902	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.212	22:57
Breath Volume	1.710	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/17/2024 03:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RICARDO,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: VALDES BERNAL

DATE/TIME OF TEST: 06/01/2024 00:50:09

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.110	00:45
Breath Volume	2.195	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:47
Subject Sample # 2	0.103	00:48
Breath Volume	2.804	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RICARDO,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: VALDES BERNAL

DATE/TIME OF TEST: 06/01/2024 01:05:16

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.078	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.105	01:00
Breath Volume	2.796	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.101	01:03
Breath Volume	2.433	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLORZANO,DYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SOLORZANO,DYLAN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BEERS

DATE/TIME OF TEST: 06/02/2024 00:11:01

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.140	00:06
Breath Volume	1.441	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:08
Subject Sample # 2	0.144	00:09
Breath Volume	1.355	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:11

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBO,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ALBO,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GUZMAN COBO

DATE/TIME OF TEST: 06/02/2024 00:46:57

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.192	00:42
Breath Volume	3.136	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.187	00:45
Breath Volume	3.347	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,BILL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CASTILLO VALDES

DATE/TIME OF TEST: 06/02/2024 03:58:57

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Sequence Aborted
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,BILL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: CASTILLO VALDES

DATE/TIME OF TEST: 06/02/2024 04:31:43

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.171	04:22
Breath Volume	1.675	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.141	04:27
Breath Volume	3.507	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 3	0.166	04:30
Breath Volume	1.785	04:30
Air Blank	0.000	04:30
Control Test	0.076	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 06/19/2024 08:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,TERENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HICKS,TERRENCE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SCHEUERMANN

DATE/TIME OF TEST: 06/03/2024 00:23:44

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank		00:23	RFI Detect
Air Blank	0.000	00:23	

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/06/2024 18:51:55

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.079	18:49	
Air Blank	0.000	18:50	
Subject Sample	# 1	18:50	RFI Detect
Breath Volume	0.000	18:50	
Air Blank	0.000	18:50	
Control Test	0.079	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/06/2024 19:01:16

Diagnostics Check	OK	18:54
Air Blank	0.000	18:55
Control Test	0.080	18:55
Air Blank	0.000	18:56
Subject Sample # 1	0.301	18:56
Breath Volume	2.625	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.301	18:59
Breath Volume	2.914	18:59
Air Blank	0.000	19:00
Control Test	0.079	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: RESTREPO VAHOS

DATE/TIME OF TEST: 06/08/2024 05:43:15

Diagnostics Check	OK	05:36	
Air Blank	0.000	05:37	
Control Test	0.080	05:37	
Air Blank	0.000	05:37	
Subject Sample	# 1 0.132	05:39	
Breath Volume	1.726	05:39	
Air Blank	0.000	05:40	
Air Blank		05:41	RFI Detect
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: RESTREPO VAHOS

DATE/TIME OF TEST: 06/08/2024 05:53:01

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.079	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.127	05:47
Breath Volume	1.484	05:47
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.132	05:51
Breath Volume	1.332	05:51
Air Blank	0.000	05:52
Control Test	0.078	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:53

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FERRO,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: GUTIERREZ MARTINEZ

DATE/TIME OF TEST: 06/08/2024 21:11:15

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.191	21:04
Breath Volume	1.863	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.184	21:09
Breath Volume	2.191	21:09
Air Blank	0.000	21:10
Control Test	0.077	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CIDRON,JULIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: BATISTA MARTINEZ

DATE/TIME OF TEST: 06/16/2024 21:02:00

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Subject Sample # 1	0.155	20:57
Breath Volume	1.527	20:57
Air Blank	0.000	20:58
Air Blank	0.000	20:59
Subject Sample # 2	0.161	21:00
Breath Volume	1.222	21:00
Air Blank	0.000	21:01
Control Test	0.077	21:01
Air Blank	0.000	21:01
Diagnostics Check	OK	21:02

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: CORONADO

DATE/TIME OF TEST: 06/19/2024 02:18:52

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:14
Breath Volume	2.144	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:17
Breath Volume	3.144	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 06/19/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: JEREZ

DATE/TIME OF TEST: 06/30/2024 17:55:34

Diagnostics Check	OK	17:50	
Air Blank	0.000	17:51	
Control Test	0.080	17:51	
Air Blank	0.000	17:52	
Subject Sample	# 1	17:54	Subject Test Refused
Breath Volume	0.000	17:54	
Air Blank	0.000	17:54	
Control Test	0.081	17:54	
Air Blank	0.000	17:55	
Diagnostics Check	OK	17:55	

Data Download Date/Time: 07/03/2024 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/02/2024 22:56:31

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.207	22:52
Breath Volume	1.878	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.224	22:55
Breath Volume	2.144	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 06/25/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 06/15/2024 20:57:19

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.082	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.314	20:52
Breath Volume	4.468	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.319	20:55
Breath Volume	3.761	20:55
Air Blank	0.000	20:56
Control Test	0.082	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 06/25/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CASTILLO GIRALDO

DATE/TIME OF TEST: 06/16/2024 06:12:19

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:58	
Control Test	0.079	05:59	
Air Blank	0.000	05:59	
Subject Sample	# 1 0.095	06:02	Volume Not Met
Breath Volume	0.785	06:02	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 2 0.139	06:07	
Breath Volume	1.167	06:07	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 3 0.128	06:10	
Breath Volume	1.507	06:10	
Air Blank	0.000	06:11	
Control Test	0.078	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 06/25/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BELIARD

DATE/TIME OF TEST: 06/18/2024 00:24:24

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.080	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.000	00:20
Breath Volume	2.062	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.000	00:23
Breath Volume	2.132	00:23
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 06/25/2024 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/01/2024 05:19:22

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.083	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.107	05:09	
Breath Volume	1.554	05:09	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2 0.096	05:14	Volume Not Met
Breath Volume	1.070	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 3 0.102	05:17	
Breath Volume	1.957	05:17	
Air Blank	0.000	05:18	
Control Test	0.079	05:18	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 06/17/2024 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HAMILTON,E
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: UFLAND

DATE/TIME OF TEST: 06/17/2024 20:23:56

Diagnostics Check	OK	20:17
Air Blank	0.000	20:17
Control Test	0.082	20:17
Air Blank	0.000	20:18
Subject Sample # 1	0.000	20:19
Breath Volume	1.519	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.000	20:22
Breath Volume	1.468	20:22
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 07/17/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HITCHCOCK,LUCAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MOUALA

DATE/TIME OF TEST: 06/30/2024 23:12:06

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.176	23:07
Breath Volume	1.746	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:09
Subject Sample # 2	0.177	23:10
Breath Volume	3.289	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 07/17/2024 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MORENOGARCIA

DATE/TIME OF TEST: 06/01/2024 00:32:21

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.082	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.129	00:27
Breath Volume	1.820	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.119	00:30
Breath Volume	1.484	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERCADO, TYLER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BUITRAGO

DATE/TIME OF TEST: 06/01/2024 03:09:59

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.157	02:58	Volume Not Met
Breath Volume	1.050	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 2 0.182	03:02	
Breath Volume	1.539	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 3 0.137	03:08	Volume Not Met
Breath Volume	0.078	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERCADO, TYLER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BUITRAGO

DATE/TIME OF TEST: 06/01/2024 03:18:11

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.177	03:13
Breath Volume	2.027	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.186	03:16
Breath Volume	3.351	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRACKNELL,JASON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: OFOSUADDO

DATE/TIME OF TEST: 06/02/2024 01:56:36

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.082	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.233	01:51	Volume Not Met
Breath Volume	0.167	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 2	01:55	Subject Test Refused
Breath Volume	0.000	01:55	
Air Blank	0.000	01:55	
Control Test	0.077	01:55	
Air Blank	0.000	01:56	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARONA,GENESSIS
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: PERALTARAMIREZ

DATE/TIME OF TEST: 06/02/2024 03:41:58

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.131	03:37
Breath Volume	1.398	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.135	03:40
Breath Volume	1.226	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JEAN BAPTISTE, ALBENS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 06/02/2024 04:07:04

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.207	04:02
Breath Volume	2.820	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.198	04:05
Breath Volume	2.437	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAYES,HUNTER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/03/2024 01:59:46

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.283	01:54
Breath Volume	1.535	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.278	01:57
Breath Volume	2.496	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: WEXLER

DATE/TIME OF TEST: 06/04/2024 04:40:14

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.082	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.000	04:36
Breath Volume	2.441	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.000	04:38
Breath Volume	2.941	04:38
Air Blank	0.000	04:39
Control Test	0.082	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA,NIKOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/08/2024 03:31:29

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.090	03:27
Breath Volume	1.894	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.080	03:29
Breath Volume	1.191	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMALL,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 06/09/2024 00:30:12

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.082	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.211	00:25
Breath Volume	2.593	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.216	00:28
Breath Volume	2.335	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: CUSI

DATE/TIME OF TEST: 06/09/2024 01:43:24

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:41	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.082	01:42	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATSON,ISAIAH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 06/09/2024 02:50:07

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.168	02:45
Breath Volume	1.277	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.174	02:48
Breath Volume	2.296	02:48
Air Blank	0.000	02:49
Control Test	0.082	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA, NIKOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/09/2024 04:19:30

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.082	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.147	04:15
Breath Volume	1.808	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.139	04:17
Breath Volume	1.226	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URBANSKI,ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 06/15/2024 01:07:52

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.082	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.081	01:03
Breath Volume	1.105	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.088	01:06
Breath Volume	1.164	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: LENGELCLAWSON

DATE/TIME OF TEST: 06/16/2024 02:45:00

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.083	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.187	02:40
Breath Volume	2.929	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.194	02:43
Breath Volume	3.843	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/20/2024 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: AGUDELOHOYOS

DATE/TIME OF TEST: 06/16/2024 04:08:35

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.180	04:03
Breath Volume	1.476	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.185	04:06
Breath Volume	1.316	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 06/20/2024 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEMP,MARK
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MAYNARD JR

DATE/TIME OF TEST: 06/17/2024 04:25:25

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.176	04:20
Breath Volume	2.031	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.171	04:23
Breath Volume	1.644	04:23
Air Blank	0.000	04:24
Control Test	0.082	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 06/20/2024 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOVELL,NATHAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SOSTRE JR

DATE/TIME OF TEST: 06/18/2024 02:49:19

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.082	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.266	02:44
Breath Volume	2.156	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:46
Subject Sample # 2	0.264	02:47
Breath Volume	2.546	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 06/20/2024 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BOIVIN

DATE/TIME OF TEST: 06/23/2024 01:48:16

Diagnostics Check	OK	01:47	
Air Blank		01:47	RFI Detect
Air Blank	0.000	01:48	

Data Download Date/Time: 07/04/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLANDINGHAM, OLIVIA, B
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE, JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BOIVIN

DATE/TIME OF TEST: 06/23/2024 01:55:34

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.077	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.123	01:51
Breath Volume	1.253	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.143	01:54
Breath Volume	1.488	01:54
Air Blank	0.000	01:54
Control Test	0.078	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 07/04/2024 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BUCKLEY,D
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 06/15/2024 23:01:05

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:58	
Air Blank	0.000	22:59	
Subject Sample	# 1	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	23:00	
Control Test	0.080	23:00	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 06/21/2024 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYNY,E,C
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 06/20/2024 18:49:59

Diagnostics Check	OK	18:37	
Air Blank	0.000	18:38	
Control Test	0.080	18:38	
Air Blank	0.000	18:39	
Subject Sample	# 1 0.150	18:42	Volume Not Met
Breath Volume	0.500	18:42	
Air Blank	0.000	18:43	
Air Blank	0.000	18:44	
Subject Sample	# 2 0.178	18:45	
Breath Volume	1.667	18:45	
Air Blank	0.000	18:46	
Air Blank	0.000	18:47	
Subject Sample	# 3 0.172	18:48	
Breath Volume	1.558	18:48	
Air Blank	0.000	18:49	
Control Test	0.077	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:49	

Data Download Date/Time: 06/21/2024 20:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: JONGSMA

DATE/TIME OF TEST: 06/24/2024 21:51:23

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.150	21:46
Breath Volume	2.546	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.152	21:49
Breath Volume	2.390	21:49
Air Blank	0.000	21:50
Control Test	0.079	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 07/27/2024 19:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,C,B
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: JANKOWSKI,B
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 05/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: RIESER

DATE/TIME OF TEST: 06/09/2024 06:54:20

Diagnostics Check	OK	06:47
Air Blank	0.000	06:47
Control Test	0.081	06:47
Air Blank	0.000	06:48
Subject Sample # 1	0.150	06:49
Breath Volume	1.652	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.145	06:52
Breath Volume	1.882	06:52
Air Blank	0.000	06:53
Control Test	0.082	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 07/01/2024 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: WALKER,JAMES
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 06/10/2024 20:57:32

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.081	20:46	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:50	No Sample Provided
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Air Blank	0.000	20:52	
Subject Sample	# 2	0.042	20:53
Breath Volume	1.601	20:53	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 3	0.043	20:56
Breath Volume	1.582	20:56	
Air Blank	0.000	20:56	
Control Test	0.081	20:56	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 07/01/2024 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,ANDREW,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: KELLEY,CHAISE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: AQUINO FLORIAN

DATE/TIME OF TEST: 06/16/2024 22:41:24

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.012	22:37
Breath Volume	1.667	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	1.332	22:40
Air Blank	0.000	22:40
Control Test	0.077	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 07/01/2024 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: LAKATOS

DATE/TIME OF TEST: 06/11/2024 16:00:55

Diagnostics Check	OK	15:55
Air Blank	0.000	15:55
Control Test	0.080	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.028	15:56
Breath Volume	1.546	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.027	15:59
Breath Volume	1.871	15:59
Air Blank	0.000	15:59
Control Test	0.080	16:00
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 06/21/2024 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 19:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: COVER

DATE/TIME OF TEST: 06/11/2024 20:17:56

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.107	20:13
Breath Volume	2.468	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.113	20:16
Breath Volume	2.593	20:16
Air Blank	0.000	20:16
Control Test	0.080	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 06/21/2024 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 15:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: HOLLINGER

DATE/TIME OF TEST: 06/19/2024 15:37:29

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:25	
Control Test	0.080	15:25	
Air Blank	0.000	15:25	
Subject Sample	# 1 0.194	15:26	
Breath Volume	1.656	15:26	
Air Blank	0.000	15:27	
Air Blank	0.000	15:28	
Subject Sample	# 2 0.160	15:32	Volume Not Met
Breath Volume	0.000	15:32	
Air Blank	0.000	15:32	
Air Blank	0.000	15:34	
Subject Sample	# 3 0.174	15:35	
Breath Volume	2.062	15:35	
Air Blank	0.000	15:36	
Control Test	0.079	15:36	
Air Blank	0.000	15:37	
Diagnostics Check	OK	15:37	

Data Download Date/Time: 06/21/2024 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 18:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: WARTHEN

DATE/TIME OF TEST: 06/20/2024 19:03:46

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:59	
Control Test	0.080	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:01	Subject Test Refused
Breath Volume	0.000	19:01	
Air Blank	0.000	19:02	
Control Test	0.080	19:02	
Air Blank	0.000	19:03	
Diagnostics Check	Fail	19:03	Voltage/Current Test
Air Blank	0.000	19:03	

Data Download Date/Time: 06/21/2024 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29122080A1

DRY GAS EXPIRY DATE: 12/05/2024

VIOLATION CODE: DUI

ARREST TIME: 22:04

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HYLTON

DATE/TIME OF TEST: 06/10/2024 22:31:26

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.170	22:26
Breath Volume	1.554	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.174	22:29
Breath Volume	1.199	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 06/21/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,SEAN,
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: SALGADO,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: MCCLENNON

DATE/TIME OF TEST: 06/02/2024 18:36:34

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:31
Subject Sample # 1	0.153	18:31
Breath Volume	2.074	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.144	18:35
Breath Volume	2.378	18:35
Air Blank	0.000	18:35
Control Test	0.080	18:35
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 06/10/2024 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOEL,K
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 05/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: FLORES-SOTO

DATE/TIME OF TEST: 06/05/2024 04:13:50

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.192	04:09
Breath Volume	2.378	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.190	04:12
Breath Volume	2.492	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 06/10/2024 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANROMPA,V,
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: STANLEY,J
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: CROWELL

DATE/TIME OF TEST: 06/07/2024 20:18:04

Diagnostics Check	OK	20:06
Air Blank	0.000	20:07
Control Test	0.078	20:07
Air Blank	0.000	20:08
Subject Sample # 1	0.156	20:11
Breath Volume	1.171	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:13
Subject Sample # 2	0.146	20:16
Breath Volume	1.121	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:18

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TEEHAN,M
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: TUCCI

DATE/TIME OF TEST: 06/07/2024 23:07:18

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.166	23:02
Breath Volume	2.628	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.164	23:05
Breath Volume	2.359	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: GRIER

DATE/TIME OF TEST: 06/08/2024 04:46:05

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.073	04:40
Breath Volume	2.136	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.070	04:44
Breath Volume	1.265	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALLEN,IRA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: OTALVARO

DATE/TIME OF TEST: 06/09/2024 02:40:19

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.174	02:34
Breath Volume	2.050	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.165	02:38
Breath Volume	1.441	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: KESLAR

DATE/TIME OF TEST: 06/09/2024 03:30:47

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.174	03:26
Breath Volume	2.011	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.180	03:29
Breath Volume	1.828	03:29
Air Blank	0.000	03:29
Control Test	0.078	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GRUNAT,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ABEHZELE FUMAT

DATE/TIME OF TEST: 06/09/2024 06:31:21

Diagnostics Check	OK	06:24
Air Blank	0.000	06:24
Control Test	0.079	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.141	06:26
Breath Volume	1.312	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:29
Subject Sample # 2	0.126	06:29
Breath Volume	2.066	06:29
Air Blank	0.000	06:30
Control Test	0.078	06:30
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 06/10/2024 23:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOLAND,T
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 06/12/2024 03:42:31

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.079	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.079	03:41	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 06/14/2024 01:50:55

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.116	01:46
Breath Volume	1.386	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.115	01:49
Breath Volume	1.269	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,SEAN,
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: HILL,SEAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 06/15/2024 19:28:04

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.080	19:20
Air Blank	0.000	19:20
Subject Sample # 1	0.191	19:21
Breath Volume	1.558	19:21
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.185	19:26
Breath Volume	1.992	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:28

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLANC,E
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MAZARIEGOS BORRALLAS

DATE/TIME OF TEST: 06/15/2024 23:13:28

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.206	23:09
Breath Volume	2.921	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.202	23:11
Breath Volume	2.074	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KRUPINSKI,L
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 06/16/2024 04:42:01

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.144	04:37
Breath Volume	1.855	04:37
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.132	04:40
Breath Volume	1.953	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:42

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLAKE,RYAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JETT

DATE/TIME OF TEST: 06/19/2024 22:54:15

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.078	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.228	22:47
Breath Volume	1.296	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.235	22:52
Breath Volume	1.968	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FLAKE,RYAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: LUCAS MARTINEZ

DATE/TIME OF TEST: 06/20/2024 02:01:48

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.243	01:52
Breath Volume		1.714	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.282	01:57
Breath Volume		3.578	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 3	0.290	02:00
Breath Volume		1.871	02:00
Air Blank		0.000	02:00
Control Test		0.077	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,LEONARD
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 06/21/2024 03:52:47

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.171	03:48
Breath Volume	2.539	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.167	03:51
Breath Volume	3.667	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: TOLKOV,ISABELLA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ZUSSMAN

DATE/TIME OF TEST: 06/22/2024 05:52:11

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:49	
Control Test	0.081	05:50	
Air Blank	0.000	05:50	
Subject Sample	# 1	05:50	RFI Detect
Breath Volume	0.000	05:50	
Air Blank	0.000	05:51	
Control Test	0.080	05:51	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: TOLKOV,I
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ZUSSMAN

DATE/TIME OF TEST: 06/22/2024 05:59:31

Diagnostics Check	OK	05:53	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:55	
Subject Sample	# 1	05:58	Slope Not Met
Breath Volume	0.687	05:58	
Air Blank		05:58	Purge Fail
Air Blank	0.000	05:59	

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS,E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: TOLKOV,I
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ZUSSMAN

DATE/TIME OF TEST: 06/22/2024 06:02:53

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:01	
Control Test	0.071	06:02	Control Outside Tolerance
Air Blank	0.000	06:02	

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAND,THOMAS',E
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: TOLKOV,I
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ZUSSMAN

DATE/TIME OF TEST: 06/22/2024 06:18:36

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:04	
Control Test	0.079	06:05	
Air Blank	0.000	06:05	
Subject Sample	# 1 0.131	06:06	
Breath Volume	1.820	06:06	
Air Blank	0.000	06:07	
Air Blank	0.000	06:09	
Subject Sample	# 2 0.133	06:12	Volume Not Met
Breath Volume	0.000	06:12	
Air Blank	0.000	06:13	
Air Blank	0.000	06:15	
Subject Sample	# 3 0.117	06:16	
Breath Volume	1.511	06:16	
Air Blank	0.000	06:17	
Control Test	0.079	06:18	
Air Blank	0.000	06:18	
Diagnostics Check	OK	06:18	

Data Download Date/Time: 07/04/2024 04:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CARBONE,A
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: LEVANDER

DATE/TIME OF TEST: 06/26/2024 22:40:10

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.000	22:34
Breath Volume	1.636	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.000	22:38
Breath Volume	1.816	22:38
Air Blank	0.000	22:39
Control Test	0.081	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BYRD,RYAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/29/2024 01:14:42

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.238	01:10
Breath Volume	2.074	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.237	01:13
Breath Volume	1.503	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BRYANT,STEVEN
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HELSEN

DATE/TIME OF TEST: 06/29/2024 02:54:03

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.176	02:49
Breath Volume	4.417	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:51
Subject Sample # 2	0.166	02:52
Breath Volume	2.972	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELORBE,DOMINICK
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VASQUEZ GERONIMO

DATE/TIME OF TEST: 06/29/2024 03:24:33

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.203	03:20
Breath Volume	1.230	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.208	03:23
Breath Volume	1.507	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MATTEW,OVERALL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: NEGRON-MORALES

DATE/TIME OF TEST: 06/29/2024 03:40:36

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.191	03:35
Breath Volume	2.062	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.181	03:39
Breath Volume	2.062	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LOSO,JOHN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/29/2024 22:34:59

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.078	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.182	22:30
Breath Volume	2.218	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.187	22:33
Breath Volume	2.054	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 07/04/2024 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 06/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: WIEGMANN HINST

DATE/TIME OF TEST: 06/30/2024 03:25:03

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.128	03:19
Breath Volume	1.871	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.126	03:23
Breath Volume	3.351	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:25

Data Download Date/Time: 07/04/2024 04:14