

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 06/04/2024 23:54:35

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.079	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.096	23:49	Volume Not Met
Breath Volume	0.601	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 2	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.078	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: RAMIREZ PEREZ

DATE/TIME OF TEST: 06/07/2024 19:52:53

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.138	19:48
Breath Volume	1.679	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.136	19:51
Breath Volume	1.832	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: FLORES MARTINEZ

DATE/TIME OF TEST: 06/10/2024 00:53:19

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.000	00:49
Breath Volume	4.011	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.000	00:51
Breath Volume	3.957	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/10/2024 22:37:28

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.080	22:24	
Air Blank	0.000	22:24	
Subject Sample	# 1 0.286	22:25	
Breath Volume	1.523	22:25	
Air Blank	0.000	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.256	22:30	Volume Not Met
Breath Volume	0.136	22:30	
Air Blank	0.000	22:32	
Air Blank	0.000	22:33	
Subject Sample	# 3 0.275	22:35	
Breath Volume	1.160	22:35	
Air Blank	0.000	22:36	
Control Test	0.075	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KERNAN

DATE/TIME OF TEST: 06/13/2024 01:09:52

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.087	01:04
Breath Volume	2.093	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.085	01:08
Breath Volume	2.312	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/13/2024 02:32:08

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.111	02:27
Breath Volume	1.695	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.118	02:30
Breath Volume	2.031	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DENNARD

DATE/TIME OF TEST: 06/16/2024 01:13:27

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.144	01:08
Breath Volume	2.722	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.148	01:11
Breath Volume	2.800	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: WHITE,BRANDON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: PIROLO

DATE/TIME OF TEST: 06/16/2024 17:03:51

Diagnostics Check	OK	16:54
Air Blank	0.000	16:54
Control Test	0.080	16:54
Air Blank	0.000	16:55
Subject Sample # 1	0.225	16:57
Breath Volume	1.734	16:57
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.230	17:02
Breath Volume	1.957	17:02
Air Blank	0.000	17:02
Control Test	0.078	17:03
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: QUINLIN

DATE/TIME OF TEST: 06/19/2024 04:16:16

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.193	04:11
Breath Volume	4.156	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.182	04:14
Breath Volume	3.730	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BURCKHALTER

DATE/TIME OF TEST: 06/22/2024 05:05:17

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.204	05:00
Breath Volume	1.402	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.199	05:03
Breath Volume	1.878	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HERNANDEZ CAHUAMARI

DATE/TIME OF TEST: 06/23/2024 01:44:03

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.186	01:39
Breath Volume	1.964	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.192	01:42
Breath Volume	2.449	01:42
Air Blank	0.000	01:43
Control Test	0.081	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: HAGG

DATE/TIME OF TEST: 06/23/2024 22:05:21

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.234	22:00
Breath Volume	1.492	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.228	22:03
Breath Volume	2.832	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 06/25/2024 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DORESTAN-DOREUS, WISLINE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/24/2024 02:25:48

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.162	02:20
Breath Volume	2.515	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.162	02:24
Breath Volume	2.390	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 06/25/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BUTTS

DATE/TIME OF TEST: 06/25/2024 05:12:01

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.377	05:07
Breath Volume	3.457	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.375	05:10
Breath Volume	3.628	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:12

Data Download Date/Time: 06/25/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KISWAH, KEES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: DOMINGO JIMENEZ

DATE/TIME OF TEST: 06/25/2024 23:39:46

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.204	23:35
Breath Volume	2.828	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.192	23:38
Breath Volume	1.812	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/17/2024 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: RICHARDS BROWN

DATE/TIME OF TEST: 06/27/2024 00:21:38

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample	# 2	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 07/17/2024 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: TAMBRIZ

DATE/TIME OF TEST: 06/29/2024 00:36:37

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.233	00:32
Breath Volume	2.402	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.240	00:35
Breath Volume	3.515	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 07/17/2024 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITELLO,MICHAEL,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: VITELLO,MICHAEL
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: MACPHERSON

DATE/TIME OF TEST: 06/29/2024 09:31:51

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:22
Air Blank	0.000	09:23
Subject Sample # 1	0.208	09:24
Breath Volume	3.093	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.179	09:27
Breath Volume	1.718	09:27
Air Blank	0.000	09:27
Air Blank	0.000	09:29
Subject Sample # 3	0.192	09:30
Breath Volume	3.800	09:30
Air Blank	0.000	09:30
Control Test	0.080	09:31
Air Blank	0.000	09:31
Diagnostics Check	OK	09:31

Data Download Date/Time: 07/17/2024 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MACRAE-LAMBERT

DATE/TIME OF TEST: 06/30/2024 01:23:34

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.172	01:19
Breath Volume	2.039	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.170	01:22
Breath Volume	3.734	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/17/2024 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ETELAMAKI, RYAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15922080A5
DRY GAS EXPIRY DATE: 08/05/2024
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BEDOYA

DATE/TIME OF TEST: 06/30/2024 21:21:28

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.080	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1 0.043	21:12	Volume Not Met
Breath Volume	0.226	21:12	
Air Blank	0.000	21:12	
Air Blank	0.000	21:14	
Subject Sample	# 2 0.059	21:16	
Breath Volume	1.402	21:16	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 3 0.054	21:19	
Breath Volume	1.132	21:19	
Air Blank	0.000	21:20	
Control Test	0.079	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 07/17/2024 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 06/20/2024 04:21:11

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.079	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.265	04:16
Breath Volume	2.636	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.255	04:19
Breath Volume	3.160	04:19
Air Blank	0.000	04:20
Control Test	0.078	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 06/30/2024 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PACHECO LAYNEZ

DATE/TIME OF TEST: 06/22/2024 03:48:14

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.244	03:43
Breath Volume	2.847	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.253	03:46
Breath Volume	2.988	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/30/2024 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: VELASQUEZ AGUILAR

DATE/TIME OF TEST: 06/29/2024 02:40:59

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.196	02:35
Breath Volume	1.843	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.184	02:39
Breath Volume	2.179	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 06/30/2024 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/30/2024 08:44:34

Diagnostics Check	OK	08:38	
Air Blank	0.000	08:39	
Control Test	0.079	08:39	
Air Blank	0.000	08:40	
Subject Sample	# 1	08:43	Subject Test Refused
Breath Volume	0.000	08:43	
Air Blank	0.000	08:43	
Control Test	0.081	08:43	
Air Blank	0.000	08:44	
Diagnostics Check	OK	08:44	

Data Download Date/Time: 06/30/2024 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,RORY,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BATCH,JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CULLEY

DATE/TIME OF TEST: 06/06/2024 02:56:10

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.149	02:51
Breath Volume	1.550	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.147	02:54
Breath Volume	1.300	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/17/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE, JESSE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 06/09/2024 01:22:50

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.229	01:17
Breath Volume	1.699	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.230	01:21
Breath Volume	1.519	01:21
Air Blank	0.000	01:21
Control Test	0.076	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/17/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FEDRO, DARIEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 06/09/2024 03:07:48

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.161	03:03
Breath Volume	3.656	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.156	03:06
Breath Volume	3.132	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 06/17/2024 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIALOBRZESKI,M
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: GOLD

DATE/TIME OF TEST: 06/20/2024 13:35:28

Diagnostics Check	OK	13:23	
Air Blank	0.000	13:23	
Control Test	0.080	13:24	
Air Blank	0.000	13:24	
Subject Sample	# 1 0.032	13:27	Volume Not Met
Breath Volume	0.000	13:27	
Air Blank	0.000	13:28	
Air Blank	0.000	13:30	
Subject Sample	# 2 0.037	13:30	
Breath Volume	1.156	13:30	
Air Blank	0.000	13:31	
Air Blank	0.000	13:33	
Subject Sample	# 3 0.035	13:33	
Breath Volume	1.652	13:33	
Air Blank	0.000	13:34	
Control Test	0.079	13:34	
Air Blank	0.000	13:35	
Diagnostics Check	OK	13:35	

Data Download Date/Time: 07/17/2024 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING,BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 06/21/2024 05:31:39

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.079	05:26	
Air Blank	0.000	05:26	
Subject Sample	# 1 0.145	05:27	
Breath Volume	4.000	05:27	
Air Blank	0.000	05:27	
Air Blank	0.000	05:29	
Subject Sample	# 2 0.137	05:30	
Breath Volume	4.558	05:30	
Air Blank	0.000	05:30	
Control Test	0.073	05:31	Control Outside Tolerance
Air Blank	0.000	05:31	

Data Download Date/Time: 07/17/2024 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: KING, BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 06/21/2024 05:41:05

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.079	05:33	
Air Blank	0.000	05:34	
Subject Sample # 1	0.141	05:36	
Breath Volume	4.777	05:36	
Air Blank	0.000	05:37	
Air Blank	0.000	05:39	
Subject Sample # 2	0.135	05:39	
Breath Volume	4.656	05:39	
Air Blank	0.000	05:40	
Control Test	0.067	05:40	Control Outside Tolerance
Air Blank	0.000	05:41	

Data Download Date/Time: 07/17/2024 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY,MATTHEW,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCARLEY,MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MOREAU

DATE/TIME OF TEST: 06/22/2024 23:35:55

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.190	23:31
Breath Volume	1.960	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:33
Subject Sample # 2	0.188	23:34
Breath Volume	2.343	23:34
Air Blank	0.000	23:34
Control Test	0.077	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/17/2024 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DAVIS,C
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: ALEGRIA

DATE/TIME OF TEST: 06/24/2024 06:29:07

Diagnostics Check	OK	06:22
Air Blank	0.000	06:23
Control Test	0.079	06:23
Air Blank	0.000	06:23
Subject Sample # 1	0.077	06:24
Breath Volume	1.945	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.078	06:27
Breath Volume	2.355	06:27
Air Blank	0.000	06:28
Control Test	0.079	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 07/17/2024 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRAYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BUTENSCHOEN

DATE/TIME OF TEST: 06/08/2024 03:39:10

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.083	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.174	03:34
Breath Volume	1.363	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:36
Subject Sample # 2	0.174	03:37
Breath Volume	1.812	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RODRIGUEZ,LARRY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MEDINA BERNAZA

DATE/TIME OF TEST: 06/11/2024 04:50:02

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.083	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.094	04:44
Breath Volume	1.160	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.095	04:48
Breath Volume	2.574	04:48
Air Blank	0.000	04:49
Control Test	0.083	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:50

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: LUGO TREJO

DATE/TIME OF TEST: 06/11/2024 14:01:09

Diagnostics Check	OK	13:56	
Air Blank	0.000	13:57	
Control Test	0.082	13:57	
Air Blank	0.000	13:57	
Subject Sample	# 1	13:59	Subject Test Refused
Breath Volume	0.000	13:59	
Air Blank	0.000	14:00	
Control Test	0.084	14:00	
Air Blank	0.000	14:01	
Diagnostics Check	OK	14:01	

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ISABEL,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: REAVES,DONOVAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ALVARADO VIZCARRA

DATE/TIME OF TEST: 06/15/2024 03:01:11

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.082	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.126	02:59	
Breath Volume	1.828	02:59	
Air Blank		03:00	Purge Fail
Air Blank	0.000	03:01	

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ISABEL,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: REAVES,DONOVAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ALVARADO VIZCARRA

DATE/TIME OF TEST: 06/15/2024 03:10:37

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.081	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.118	03:05
Breath Volume	1.300	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.122	03:08
Breath Volume	1.343	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCLEOD,BUFFIE
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: ALDANA

DATE/TIME OF TEST: 06/16/2024 16:35:06

Diagnostics Check	OK	16:27
Air Blank	0.000	16:27
Control Test	0.084	16:27
Air Blank	0.000	16:28
Subject Sample # 1	0.210	16:29
Breath Volume	1.917	16:29
Air Blank	0.000	16:30
Air Blank	0.000	16:31
Subject Sample # 2	0.202	16:33
Breath Volume	1.480	16:33
Air Blank	0.000	16:34
Control Test	0.082	16:34
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ISABEL,M
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRAYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SANCHEZ GAYOSSO

DATE/TIME OF TEST: 06/17/2024 20:58:23

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.084	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.172	20:53
Breath Volume	1.894	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.157	20:56
Breath Volume	1.730	20:56
Air Blank	0.000	20:57
Control Test	0.082	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AFONSO,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRAYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 05/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: AGUILAR OLAN

DATE/TIME OF TEST: 06/23/2024 23:03:09

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.083	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.165	22:58
Breath Volume	2.687	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:00
Subject Sample # 2	0.168	23:01
Breath Volume	3.187	23:01
Air Blank	0.000	23:02
Control Test	0.082	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 07/01/2024 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 06/02/2024 01:06:45

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.170	00:59
Breath Volume	1.906	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.180	01:04
Breath Volume	1.355	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 06/10/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 06/05/2024 18:48:44

Diagnostics Check	OK	18:42
Air Blank	0.000	18:42
Control Test	0.080	18:42
Air Blank	0.000	18:43
Subject Sample # 1	0.000	18:44
Breath Volume	2.859	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.000	18:47
Breath Volume	4.601	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:48

Data Download Date/Time: 06/10/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: CORSO

DATE/TIME OF TEST: 06/13/2024 23:43:33

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.283	23:38
Breath Volume	1.410	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.285	23:41
Breath Volume	1.121	23:41
Air Blank	0.000	23:42
Control Test	0.077	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 06/15/2024 22:51:44

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:46
Air Blank	0.000	22:46
Subject Sample # 1	0.207	22:47
Breath Volume	1.851	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.218	22:49
Breath Volume	1.871	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: LOPEZ-CHULIN

DATE/TIME OF TEST: 06/16/2024 03:14:40

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.193	03:09
Breath Volume	2.875	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.188	03:12
Breath Volume	2.726	03:12
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ARREAGA PEREZ

DATE/TIME OF TEST: 06/17/2024 03:29:05

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:24	
Subject Sample # 1	0.156	03:25	
Breath Volume	2.191	03:25	
Air Blank	0.000	03:26	
Air Blank		03:27	RFI Detect
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ARREAGAPEREZ

DATE/TIME OF TEST: 06/17/2024 03:37:52

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.174	03:32
Breath Volume	1.285	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.171	03:36
Breath Volume	2.695	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HAUSER

DATE/TIME OF TEST: 06/21/2024 02:54:08

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.136	02:49
Breath Volume	2.343	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:51
Subject Sample # 2	0.137	02:52
Breath Volume	1.464	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: LANDSBERG

DATE/TIME OF TEST: 06/24/2024 21:28:28

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:23
Subject Sample # 1	0.173	21:23
Breath Volume	1.187	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.178	21:26
Breath Volume	1.781	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ATKISSON

DATE/TIME OF TEST: 06/26/2024 20:50:14

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.215	20:45
Breath Volume	1.234	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.215	20:48
Breath Volume	1.707	20:48
Air Blank	0.000	20:49
Control Test	0.079	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CHICO MANZANARES

DATE/TIME OF TEST: 06/28/2024 23:23:06

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.114	23:18
Breath Volume	2.453	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.111	23:21
Breath Volume	2.968	23:21
Air Blank	0.000	23:22
Control Test	0.076	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/30/2024 00:34:56

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADEN,JASON,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BRADEN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/30/2024 00:51:22

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1 0.241	00:39	
Breath Volume	1.261	00:39	
Air Blank	0.000	00:40	
Air Blank	0.000	00:42	
Subject Sample	# 2 0.206	00:45	Volume Not Met
Breath Volume	0.082	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 3	00:49	Subject Test Refused
Breath Volume	0.421	00:49	
Air Blank	0.000	00:50	
Control Test	0.074	00:50	Control Outside Tolerance
Air Blank	0.000	00:51	

Data Download Date/Time: 07/08/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/01/2024 03:26:26

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.263	03:18
Breath Volume	1.613	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.242	03:21
Breath Volume	2.007	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 3	0.243	03:24
Breath Volume	2.585	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 06/14/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CUTZ TORRES

DATE/TIME OF TEST: 06/02/2024 01:32:10

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
		01:31	RFI Detect
Air Blank	0.000	01:32	

Data Download Date/Time: 06/14/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: CUTZ TORRES

DATE/TIME OF TEST: 06/02/2024 01:41:32

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.245	01:36
Breath Volume	3.156	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.239	01:39
Breath Volume	2.699	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 06/14/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ASHLEY,AUSTIN
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: EXUMAT

DATE/TIME OF TEST: 06/02/2024 04:36:19

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:22	
Control Test	0.080	04:22	
Air Blank	0.000	04:23	
Subject Sample	# 1 0.209	04:26	Volume Not Met
Breath Volume	0.351	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2 0.268	04:31	
Breath Volume	1.835	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 3 0.278	04:34	
Breath Volume	1.472	04:34	
Air Blank	0.000	04:35	
Control Test	0.076	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 06/14/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/02/2024 22:57:45

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.207	22:53
Breath Volume	2.445	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.215	22:56
Breath Volume	2.343	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 06/14/2024 16:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: FLORESTAL

DATE/TIME OF TEST: 06/03/2024 01:19:54

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.161	01:15
Breath Volume	1.816	01:15
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.168	01:18
Breath Volume	1.851	01:18
Air Blank	0.000	01:18
Control Test	0.080	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DILWORTH

DATE/TIME OF TEST: 06/03/2024 22:13:02

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.080	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:08
Breath Volume	1.878	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:11
Breath Volume	2.375	22:11
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:13

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: GOMEZ REYES

DATE/TIME OF TEST: 06/04/2024 22:39:39

Diagnostics Check	OK	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.187	22:35
Breath Volume	1.746	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.185	22:38
Breath Volume	1.492	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 06/06/2024 01:24:31

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.130	01:19
Breath Volume	1.671	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.123	01:22
Breath Volume	1.847	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SAUCEDO CANO

DATE/TIME OF TEST: 06/06/2024 01:47:16

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.159	01:42
Breath Volume	1.328	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.156	01:45
Breath Volume	1.332	01:45
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MANZANO SOLORIO

DATE/TIME OF TEST: 06/06/2024 23:02:56

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.201	22:58
Breath Volume	2.785	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.204	23:01
Breath Volume	2.656	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD, ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CORNETTA

DATE/TIME OF TEST: 06/07/2024 22:30:33

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.000	22:26
Breath Volume	1.886	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:29
Breath Volume	3.460	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FURNESS, SCOTT
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BRAVO SOTO

DATE/TIME OF TEST: 06/08/2024 00:08:12

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.244	00:03
Breath Volume	2.558	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.238	00:06
Breath Volume	1.769	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EGINTON, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: PEREZ AGUILAR

DATE/TIME OF TEST: 06/08/2024 09:55:04

Diagnostics Check	OK	09:48
Air Blank	0.000	09:48
Control Test	0.080	09:49
Air Blank	0.000	09:49
Subject Sample # 1	0.235	09:50
Breath Volume	1.277	09:50
Air Blank	0.000	09:50
Air Blank	0.000	09:52
Subject Sample # 2	0.236	09:53
Breath Volume	1.667	09:53
Air Blank	0.000	09:54
Control Test	0.079	09:54
Air Blank	0.000	09:54
Diagnostics Check	OK	09:55

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MONEUS

DATE/TIME OF TEST: 06/09/2024 01:11:43

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.243	01:07
Breath Volume	2.496	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.249	01:10
Breath Volume	2.644	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IVANOVIC,MATTHEW
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MORALES PEREZ

DATE/TIME OF TEST: 06/09/2024 04:07:16

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.227	04:02
Breath Volume	2.683	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.231	04:05
Breath Volume	2.417	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: JAUQUET

DATE/TIME OF TEST: 06/10/2024 03:47:01

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	0.137	03:39
Breath Volume	1.531	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3	0.143	03:45
Breath Volume	2.484	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MENDEZ, ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 06/10/2024 23:41:08

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.162	23:36
Breath Volume	2.332	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.152	23:39
Breath Volume	1.593	23:39
Air Blank	0.000	23:40
Control Test	0.080	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: CITALANSAQUIC

DATE/TIME OF TEST: 06/11/2024 22:26:41

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1 0.175	22:18	Volume Not Met
Breath Volume	0.781	22:18	
Air Blank	0.000	22:19	
Air Blank	0.000	22:21	
Subject Sample	# 2 0.181	22:21	
Breath Volume	1.378	22:21	
Air Blank	0.000	22:22	
Air Blank	0.000	22:24	
Subject Sample	# 3 0.187	22:25	
Breath Volume	1.347	22:25	
Air Blank	0.000	22:25	
Control Test	0.079	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: REVELO MAFLA

DATE/TIME OF TEST: 06/12/2024 02:27:31

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.203	02:22
Breath Volume	1.226	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.216	02:25
Breath Volume	1.796	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRIMBLE, DELANEY
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 06/12/2024 22:56:40

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.193	22:52
Breath Volume	2.515	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.206	22:55
Breath Volume	2.214	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VILLARIN,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: RESIGNAC

DATE/TIME OF TEST: 06/13/2024 06:27:11

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.081	06:21
Air Blank	0.000	06:21
Subject Sample # 1	0.117	06:22
Breath Volume	1.941	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:24
Subject Sample # 2	0.115	06:25
Breath Volume	1.976	06:25
Air Blank	0.000	06:26
Control Test	0.081	06:26
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SASOV, ALEKSEY
ARRESTING OFFICER AGENCY: Ocean Ridge PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SCALO

DATE/TIME OF TEST: 06/13/2024 20:17:53

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.080	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.172	20:13
Breath Volume	1.796	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.167	20:16
Breath Volume	1.613	20:16
Air Blank	0.000	20:16
Control Test	0.081	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/14/2024 01:48:51

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.304	01:44
Breath Volume	1.343	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.305	01:47
Breath Volume	2.542	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/14/2024 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: PEREZ VELASQUEZ

DATE/TIME OF TEST: 06/01/2024 23:09:31

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.215	23:04
Breath Volume	1.445	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.206	23:07
Breath Volume	1.105	23:07
Air Blank	0.000	23:08
Control Test	0.079	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERA,FRANKLYN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 06/01/2024 23:59:22

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:46	
Control Test	0.079	23:46	
Air Blank	0.000	23:47	
Subject Sample	# 1 0.102	23:50	Volume Not Met
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:52	
Subject Sample	# 2 0.117	23:53	
Breath Volume	2.714	23:53	
Air Blank	0.000	23:54	
Air Blank	0.000	23:56	
Subject Sample	# 3 0.118	23:57	
Breath Volume	3.464	23:57	
Air Blank	0.000	23:58	
Control Test	0.078	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: STYGAR

DATE/TIME OF TEST: 06/02/2024 05:34:39

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.146	05:30
Breath Volume	1.472	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.162	05:33
Breath Volume	3.179	05:33
Air Blank	0.000	05:33
Control Test	0.079	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: OCHOA LOPEZ

DATE/TIME OF TEST: 06/02/2024 23:58:12

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:51
Subject Sample # 1	0.098	23:52
Breath Volume	1.351	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.100	23:56
Breath Volume	1.378	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JENNINGS,DAVON
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LEPSCH THORNE

DATE/TIME OF TEST: 06/06/2024 23:01:12

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.249	22:54
Breath Volume	1.191	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.253	22:59
Breath Volume	1.523	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LARRABEE,JONATHAN
ARRESTING OFFICER AGENCY: FL Atlantic UPD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SCHOTTKE

DATE/TIME OF TEST: 06/07/2024 02:28:43

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.212	02:23
Breath Volume	1.144	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.226	02:27
Breath Volume	3.261	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 06/07/2024 22:48:46

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.095	22:43
Breath Volume	1.445	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.095	22:47
Breath Volume	1.640	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MANNING,WESLEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: LAINEZ RAMIREZ

DATE/TIME OF TEST: 06/07/2024 23:40:56

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.208	23:36
Breath Volume	2.613	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.211	23:39
Breath Volume	1.917	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 06/09/2024 03:06:02

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.169	03:01
Breath Volume	2.144	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.183	03:04
Breath Volume	2.730	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:06

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAPPIN,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 06/09/2024 05:45:48

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.080	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.194	05:40
Breath Volume	1.671	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.205	05:44
Breath Volume	2.628	05:44
Air Blank	0.000	05:44
Control Test	0.079	05:45
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DIAS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/11/2024 14:15:22

Diagnostics Check	OK	14:09
Air Blank	0.000	14:09
Control Test	0.079	14:09
Air Blank	0.000	14:10
Subject Sample # 1	0.000	14:11
Breath Volume	1.214	14:11
Air Blank	0.000	14:11
Air Blank	0.000	14:13
Subject Sample # 2	0.000	14:14
Breath Volume	1.503	14:14
Air Blank	0.000	14:14
Control Test	0.079	14:14
Air Blank	0.000	14:15
Diagnostics Check	OK	14:15

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: FOSCHINI

DATE/TIME OF TEST: 06/11/2024 22:35:07

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.151	22:30
Breath Volume	2.117	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.156	22:33
Breath Volume	1.238	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: KUBENEZ

DATE/TIME OF TEST: 06/12/2024 23:50:36

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.079	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2	23:49	No Sample Provided
Breath Volume	0.000	23:49	
Air Blank	0.000	23:49	
Control Test	0.079	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: LOPEZ MENDEZ

DATE/TIME OF TEST: 06/14/2024 01:15:31

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:14	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXANDER,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 06/14/2024 03:54:04

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.174	03:49
Breath Volume	1.386	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.165	03:52
Breath Volume	2.515	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:54

Data Download Date/Time: 06/14/2024 13:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: CORDOVA

DATE/TIME OF TEST: 06/14/2024 22:34:59

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.171	22:28
Breath Volume	1.984	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.159	22:33
Breath Volume	1.222	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/15/2024 01:54:17

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.142	01:49
Breath Volume	1.417	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.142	01:52
Breath Volume	1.550	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD, ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LOPEZ RAMIREZ

DATE/TIME OF TEST: 06/15/2024 03:07:59

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.231	03:02
Breath Volume	2.519	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.237	03:06
Breath Volume	1.312	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BIRENBAUM

DATE/TIME OF TEST: 06/15/2024 22:01:38

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.000	21:57
Breath Volume	2.144	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	22:00
Breath Volume	2.109	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLAKE, XAVIAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SCHLOEMER

DATE/TIME OF TEST: 06/16/2024 00:46:16

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.188	00:41
Breath Volume	2.058	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.195	00:44
Breath Volume	1.437	00:44
Air Blank	0.000	00:45
Control Test	0.078	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALONSO, ROXANA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: JANVIER

DATE/TIME OF TEST: 06/16/2024 07:33:04

Diagnostics Check	OK	07:25
Air Blank	0.000	07:26
Control Test	0.080	07:26
Air Blank	0.000	07:27
Subject Sample # 1	0.140	07:28
Breath Volume	1.691	07:28
Air Blank	0.000	07:28
Air Blank	0.000	07:30
Subject Sample # 2	0.141	07:31
Breath Volume	1.363	07:31
Air Blank	0.000	07:32
Control Test	0.079	07:32
Air Blank	0.000	07:32
Diagnostics Check	OK	07:33

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHEHAB,KHODR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: CASTELLON BLANCO

DATE/TIME OF TEST: 06/16/2024 08:31:10

Diagnostics Check	OK	08:21
Air Blank	0.000	08:21
Control Test	0.080	08:21
Air Blank	0.000	08:22
Subject Sample # 1	0.143	08:24
Breath Volume	1.144	08:24
Air Blank	0.000	08:25
Air Blank	0.000	08:27
Subject Sample # 2	0.136	08:29
Breath Volume	1.148	08:29
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BASHAR

DATE/TIME OF TEST: 06/16/2024 22:08:26

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.000	22:03
Breath Volume	1.425	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.000	22:07
Breath Volume	1.585	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 06/18/2024 01:37:13

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.229	01:32
Breath Volume	1.933	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.243	01:35
Breath Volume	2.058	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: HIDALGO LOPEZ

DATE/TIME OF TEST: 06/19/2024 02:26:44

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.080	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.170	02:22
Breath Volume	1.937	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.170	02:25
Breath Volume	1.429	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 06/20/2024 00:33:44

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.079	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.177	00:29
Breath Volume	2.351	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.173	00:32
Breath Volume	2.265	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: DI ROBERTO JR

DATE/TIME OF TEST: 06/20/2024 01:12:49

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.114	01:08
Breath Volume	2.300	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.116	01:11
Breath Volume	2.445	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PINZON,JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 06/20/2024 18:17:31

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.080	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.181	18:13
Breath Volume	3.429	18:13
Air Blank	0.000	18:13
Air Blank	0.000	18:15
Subject Sample # 2	0.164	18:15
Breath Volume	1.660	18:15
Air Blank	0.000	18:16
Control Test	0.079	18:16
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/21/2024 03:43:12

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.173	03:38
Breath Volume	2.019	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.165	03:41
Breath Volume	2.449	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJADA, ALBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ANDRES RAYMUNDO

DATE/TIME OF TEST: 06/21/2024 23:18:14

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.226	23:13
Breath Volume	2.199	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.223	23:16
Breath Volume	1.632	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/22/2024 00:21:22

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.163	00:16
Breath Volume	2.710	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.166	00:19
Breath Volume	3.449	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: POZO GUERRERO

DATE/TIME OF TEST: 06/22/2024 01:11:24

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.176	01:06
Breath Volume	2.062	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.171	01:09
Breath Volume	2.675	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HLUCH,MARIO
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BARTHELEMY

DATE/TIME OF TEST: 06/22/2024 04:39:47

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.139	04:35
Breath Volume	1.250	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.145	04:38
Breath Volume	1.136	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: MATEO MORALES

DATE/TIME OF TEST: 06/22/2024 06:55:14

Diagnostics Check	OK	06:48
Air Blank	0.000	06:49
Control Test	0.079	06:49
Air Blank	0.000	06:49
Subject Sample # 1	0.177	06:50
Breath Volume	1.527	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:53
Subject Sample # 2	0.173	06:53
Breath Volume	1.968	06:53
Air Blank	0.000	06:54
Control Test	0.079	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: IXCOY LOPEZ

DATE/TIME OF TEST: 06/22/2024 23:27:33

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.174	23:22
Breath Volume	1.312	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.172	23:25
Breath Volume	1.335	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILORD,JHAMIL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/23/2024 01:42:14

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.205	01:37
Breath Volume	1.480	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.205	01:40
Breath Volume	1.925	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ, JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MANGINI

DATE/TIME OF TEST: 06/23/2024 06:15:45

Diagnostics Check	OK	06:12	
Air Blank	0.000	06:12	
Control Test	0.080	06:13	
Air Blank	0.000	06:13	
Subject Sample	# 1	06:14	Slope Not Met
Breath Volume	0.914	06:14	
Air Blank	0.000	06:14	
Control Test	0.079	06:15	
Air Blank	0.000	06:15	
Diagnostics Check	OK	06:15	

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JUAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: MANGINI

DATE/TIME OF TEST: 06/23/2024 06:47:28

Diagnostics Check	OK	06:40
Air Blank	0.000	06:41
Control Test	0.079	06:41
Air Blank	0.000	06:42
Subject Sample # 1	0.256	06:42
Breath Volume	1.285	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.251	06:45
Breath Volume	2.503	06:45
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR, MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: SELENIUS

DATE/TIME OF TEST: 06/23/2024 21:45:08

Diagnostics Check	OK	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.000	21:40
Breath Volume	3.394	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:43
Breath Volume	1.285	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KYSOR, MITCHELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MILLAN HERNANDEZ

DATE/TIME OF TEST: 06/24/2024 02:40:29

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.263	02:35
Breath Volume	1.210	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.270	02:38
Breath Volume	1.796	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH, GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: MAKI

DATE/TIME OF TEST: 06/24/2024 21:27:12

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.079	21:19
Air Blank	0.000	21:19
Subject Sample # 1	0.146	21:20
Breath Volume	1.496	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.145	21:25
Breath Volume	1.652	21:25
Air Blank	0.000	21:26
Control Test	0.078	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: VASQUEZ GARCIA

DATE/TIME OF TEST: 06/25/2024 02:09:36

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.234	02:04
Breath Volume	2.730	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.230	02:08
Breath Volume	2.328	02:08
Air Blank	0.000	02:08
Control Test	0.078	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KENNEDY,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: KACZMARSKI

DATE/TIME OF TEST: 06/25/2024 20:32:03

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.079	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.230	20:27
Breath Volume	1.394	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.217	20:30
Breath Volume	1.449	20:30
Air Blank	0.000	20:31
Control Test	0.079	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:32

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: MORADEL

DATE/TIME OF TEST: 06/25/2024 21:58:03

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.079	21:50
Air Blank	0.000	21:50
Subject Sample # 1	0.144	21:52
Breath Volume	1.628	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.154	21:56
Breath Volume	1.988	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:58

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/25/2024 23:36:00

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.149	23:31
Breath Volume	1.500	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.151	23:34
Breath Volume	1.906	23:34
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:36

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MIGUEL

DATE/TIME OF TEST: 06/27/2024 06:04:27

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.080	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.161	05:59
Breath Volume	1.601	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.159	06:02
Breath Volume	2.054	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BODENSCHATZ

DATE/TIME OF TEST: 06/28/2024 21:46:27

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.229	21:41
Breath Volume	2.382	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.226	21:44
Breath Volume	2.394	21:44
Air Blank	0.000	21:45
Control Test	0.078	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 06/29/2024 00:40:39

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.224	00:36
Breath Volume	2.843	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.209	00:39
Breath Volume	2.652	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 07/02/2024 06:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ALFARO ESCOBAR

DATE/TIME OF TEST: 06/29/2024 01:08:49

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.175	01:04
Breath Volume	1.734	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.177	01:07
Breath Volume	1.886	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROTHENBURG,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: MEDINA FERNANDEZ

DATE/TIME OF TEST: 06/29/2024 08:47:18

Diagnostics Check	OK	08:40
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:41
Subject Sample # 1	0.244	08:42
Breath Volume	1.976	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.246	08:45
Breath Volume	1.949	08:45
Air Blank	0.000	08:46
Control Test	0.078	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 06/29/2024 23:19:43

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank	0.000	23:09	
Subject Sample	# 1 0.182	23:12	Volume Not Met
Breath Volume	0.144	23:12	
Air Blank	0.000	23:13	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.204	23:18	Volume Not Met
Breath Volume	0.019	23:18	
Air Blank	0.000	23:18	
Control Test	0.078	23:19	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/30/2024 02:31:08

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.068	02:25
Breath Volume	1.757	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.070	02:29
Breath Volume	1.765	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 06/30/2024 04:19:08

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Subject Sample # 1	0.191	04:09	
Breath Volume	1.382	04:09	
Air Blank	0.000	04:09	
Air Blank	0.000	04:11	
Subject Sample # 2	0.172	04:14	Volume Not Met
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample # 3	0.178	04:17	
Breath Volume	1.250	04:17	
Air Blank	0.000	04:18	
Control Test	0.078	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON, DAVIS
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: PALKO

DATE/TIME OF TEST: 06/30/2024 17:41:38

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.079	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.124	17:37
Breath Volume	3.332	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.135	17:40
Breath Volume	3.593	17:40
Air Blank	0.000	17:40
Control Test	0.080	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 07/02/2024 06:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SCHAEFFER,WILLIAM
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: STRANSKY

DATE/TIME OF TEST: 06/03/2024 18:34:16

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:28	
Control Test	0.078	18:28	
Air Blank	0.000	18:29	
Subject Sample	# 1 0.303	18:29	
Breath Volume	3.277	18:29	
Air Blank	0.000	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 2 0.306	18:32	
Breath Volume	2.925	18:32	
Air Blank	0.000	18:33	
Control Test	0.074	18:33	Control Outside Tolerance
Air Blank	0.000	18:34	

Data Download Date/Time: 06/27/2024 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GONZALEZ,JOSE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: NOBLES

DATE/TIME OF TEST: 06/05/2024 00:26:40

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.132	00:21
Breath Volume	2.695	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.124	00:24
Breath Volume	2.621	00:24
Air Blank	0.000	00:25
Control Test	0.076	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 06/27/2024 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: VALDES,LYNN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:32

SUBJECT'S LAST NAME: KOORS

DATE/TIME OF TEST: 06/07/2024 17:05:08

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.079	16:58
Air Blank	0.000	16:59
Subject Sample # 1	0.293	17:00
Breath Volume	3.355	17:00
Air Blank	0.000	17:01
Air Blank	0.000	17:02
Subject Sample # 2	0.289	17:03
Breath Volume	3.394	17:03
Air Blank	0.000	17:04
Control Test	0.077	17:04
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

Data Download Date/Time: 06/27/2024 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GONZALEZ,GILBERTO
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 05/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: BISOGNO

DATE/TIME OF TEST: 06/23/2024 15:48:04

Diagnostics Check	OK	15:36	
Air Blank	0.000	15:36	
Control Test	0.079	15:37	
Air Blank	0.000	15:37	
Subject Sample	# 1	15:41	No Sample Provided
Breath Volume	0.000	15:41	
Air Blank	0.000	15:41	
Air Blank	0.000	15:43	
Subject Sample	# 2	15:46	No Sample Provided
Breath Volume	0.000	15:46	
Air Blank	0.000	15:47	
Control Test	0.079	15:47	
Air Blank	0.000	15:47	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 06/27/2024 21:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: CARLISLE,KEITH
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:07

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/28/2024 14:34:01

Diagnostics Check	OK	14:27
Air Blank	0.000	14:27
Control Test	0.079	14:28
Air Blank	0.000	14:28
Subject Sample # 1	0.000	14:29
Breath Volume	2.363	14:29
Air Blank	0.000	14:29
Air Blank	0.000	14:31
Subject Sample # 2	0.000	14:32
Breath Volume	2.308	14:32
Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:33
Diagnostics Check	OK	14:34

Data Download Date/Time: 07/28/2024 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: HOTHEM,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/01/2024 02:58:06

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.034	02:53
Breath Volume	1.578	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.033	02:56
Breath Volume	1.800	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 06/05/2024 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARRASQUILLO

DATE/TIME OF TEST: 06/01/2024 04:28:26

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.176	04:23
Breath Volume	2.761	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.168	04:26
Breath Volume	1.746	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 06/05/2024 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: TORO

DATE/TIME OF TEST: 06/01/2024 05:57:16

Diagnostics Check	OK	05:49
Air Blank	0.000	05:50
Control Test	0.079	05:50
Air Blank	0.000	05:51
Subject Sample # 1	0.071	05:52
Breath Volume	1.269	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.073	05:55
Breath Volume	2.093	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 06/05/2024 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LOPEZ ACROS

DATE/TIME OF TEST: 06/02/2024 00:22:31

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.180	00:17
Breath Volume	2.273	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.172	00:20
Breath Volume	1.808	00:20
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 06/05/2024 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: TABARES SANCHEZ

DATE/TIME OF TEST: 06/02/2024 05:07:37

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.181	05:03
Breath Volume	1.300	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.175	05:06
Breath Volume	2.449	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 06/05/2024 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/03/2024 02:56:47

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.119	02:52
Breath Volume	2.824	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.113	02:55
Breath Volume	2.835	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/05/2024 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/07/2024 02:25:56

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.102	02:20
Breath Volume	1.238	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.108	02:24
Breath Volume	2.785	02:24
Air Blank	0.000	02:24
Control Test	0.078	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORREZ JARQUIN

DATE/TIME OF TEST: 06/08/2024 02:56:27

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.134	02:52
Breath Volume	1.742	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.138	02:54
Breath Volume	2.527	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 06/09/2024 01:43:21

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.007	01:38
Breath Volume	2.375	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.007	01:41
Breath Volume	2.128	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HABBISON

DATE/TIME OF TEST: 06/09/2024 03:09:44

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.175	03:03
Breath Volume	1.160	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.160	03:08
Breath Volume	1.496	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 06/09/2024 04:16:32

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.074	04:11
Breath Volume	2.109	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.073	04:14
Breath Volume	1.667	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: HINES II

DATE/TIME OF TEST: 06/09/2024 05:11:52

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.119	05:07
Breath Volume	2.093	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.119	05:10
Breath Volume	1.832	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 06/10/2024 03:24:36

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1 0.230	03:13	
Breath Volume	2.453	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 2 0.271	03:17	
Breath Volume	1.402	03:17	
Air Blank	0.000	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 3	03:22	Subject Test Refused
Breath Volume	0.562	03:22	
Air Blank	0.000	03:23	
Control Test	0.078	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: MONROY RAMIREZ

DATE/TIME OF TEST: 06/11/2024 05:13:46

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.152	05:08
Breath Volume	2.816	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.151	05:12
Breath Volume	1.964	05:12
Air Blank	0.000	05:12
Control Test	0.079	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HOYTE

DATE/TIME OF TEST: 06/13/2024 03:51:58

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.192	03:47
Breath Volume	3.410	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.187	03:50
Breath Volume	3.054	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: SELVEY

DATE/TIME OF TEST: 06/13/2024 23:45:52

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.237	23:37
Breath Volume	1.585	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.264	23:41
Breath Volume	1.855	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 3	0.229	23:44
Breath Volume	1.730	23:44
Air Blank	0.000	23:44
Control Test	0.077	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 06/14/2024 03:06:59

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.129	03:02
Breath Volume	1.187	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.125	03:05
Breath Volume	1.269	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COOPER, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SUAREZ VALDES

DATE/TIME OF TEST: 06/14/2024 03:44:26

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 07/03/2024 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 06/15/2024 01:17:24

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:10	No Sample Provided
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2	01:15	Volume Not Met
Breath Volume	0.070	01:15	
Air Blank	0.000	01:16	
Control Test	0.077	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 06/15/2024 03:19:23

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.000	03:14
Breath Volume	1.667	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.000	03:18
Breath Volume	1.828	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 06/15/2024 03:47:55

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.079	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.078	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/16/2024 02:37:19

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.059	02:32
Breath Volume	2.300	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.058	02:35
Breath Volume	2.281	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VILLANUEVA III

DATE/TIME OF TEST: 06/16/2024 03:52:59

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.066	03:47
Breath Volume	1.117	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.065	03:51
Breath Volume	1.238	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORSEN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 06/17/2024 05:40:43

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.138	05:36
Breath Volume	2.644	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.145	05:39
Breath Volume	3.125	05:39
Air Blank	0.000	05:39
Control Test	0.078	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 06/19/2024 04:22:47

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.137	04:18
Breath Volume	2.621	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.140	04:21
Breath Volume	2.187	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 06/19/2024 23:50:46

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.182	23:45
Breath Volume	2.070	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.192	23:49
Breath Volume	1.671	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BANCRAFF ACOSTA

DATE/TIME OF TEST: 06/21/2024 03:23:44

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.125	03:19
Breath Volume	2.000	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.109	03:22
Breath Volume	1.242	03:22
Air Blank	0.000	03:22
Control Test	0.079	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ANGERMILLER

DATE/TIME OF TEST: 06/22/2024 00:32:34

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.213	00:27
Breath Volume	2.312	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.218	00:30
Breath Volume	2.074	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MASOCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 06/22/2024 04:47:43

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.079	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.185	04:42
Breath Volume	2.003	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.166	04:46
Breath Volume	1.769	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: RIVERA CASTRO

DATE/TIME OF TEST: 06/22/2024 05:03:31

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.188	04:59
Breath Volume	1.820	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.192	05:01
Breath Volume	2.578	05:01
Air Blank	0.000	05:02
Control Test	0.079	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: RAMIREZ RAMIREZ

DATE/TIME OF TEST: 06/23/2024 01:08:59

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.186	01:04
Breath Volume	2.011	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.183	01:07
Breath Volume	2.488	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: BUONANNO JR

DATE/TIME OF TEST: 06/23/2024 01:51:13

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DEJOHN

DATE/TIME OF TEST: 06/23/2024 04:35:30

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.109	04:30
Breath Volume	2.421	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.101	04:33
Breath Volume	1.597	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 06/23/2024 05:35:06

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.079	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.144	05:30
Breath Volume	2.976	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.141	05:33
Breath Volume	1.933	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAIRSTON,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/23/2024 15:37:44

Diagnostics Check	OK	15:30
Air Blank	0.000	15:31
Control Test	0.079	15:31
Air Blank	0.000	15:31
Subject Sample # 1	0.272	15:33
Breath Volume	1.484	15:33
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.265	15:36
Breath Volume	1.238	15:36
Air Blank	0.000	15:36
Control Test	0.077	15:37
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 06/23/2024 20:36:06

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.079	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.089	20:31
Breath Volume	2.656	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.089	20:34
Breath Volume	2.625	20:34
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: BAISDEN

DATE/TIME OF TEST: 06/23/2024 21:49:45

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.079	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.162	21:45
Breath Volume	2.527	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.158	21:48
Breath Volume	2.234	21:48
Air Blank	0.000	21:48
Control Test	0.078	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MEJIA ESCALANTE

DATE/TIME OF TEST: 06/24/2024 23:36:50

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.192	23:32
Breath Volume	1.507	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.201	23:35
Breath Volume	2.078	23:35
Air Blank	0.000	23:35
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FULKER

DATE/TIME OF TEST: 06/25/2024 03:10:05

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.000	03:05
Breath Volume	2.796	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.000	03:08
Breath Volume	2.570	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/03/2024 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: CEDIEL GONZALES

DATE/TIME OF TEST: 06/25/2024 21:46:14

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.145	21:41
Breath Volume	1.527	21:41
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.143	21:44
Breath Volume	1.636	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GUZMAN JR

DATE/TIME OF TEST: 06/26/2024 03:00:53

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.135	02:55
Breath Volume	2.035	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.129	02:59
Breath Volume	1.187	02:59
Air Blank	0.000	02:59
Control Test	0.076	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEISCH,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 06/29/2024 01:45:00

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.208	01:39
Breath Volume	1.640	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:41
Subject Sample # 2	0.196	01:43
Breath Volume	1.136	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN, T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ALSHROUF

DATE/TIME OF TEST: 06/29/2024 02:27:02

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.166	02:18
Breath Volume	1.738	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.145	02:22
Breath Volume	1.582	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 3	0.153	02:25
Breath Volume	1.824	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WILKINS

DATE/TIME OF TEST: 06/29/2024 04:05:38

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.084	04:01
Breath Volume	2.230	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.079	04:04
Breath Volume	1.894	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SUTPHIN-GARCIA

DATE/TIME OF TEST: 06/30/2024 01:08:06

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.041	01:03
Breath Volume	2.699	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.038	01:06
Breath Volume	3.281	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PEREZ RAMOS

DATE/TIME OF TEST: 06/30/2024 04:33:25

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.152	04:28
Breath Volume	1.343	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:30
Subject Sample # 2	0.155	04:31
Breath Volume	2.257	04:31
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORSEN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: SANTIAGO MARTINEZ

DATE/TIME OF TEST: 06/30/2024 23:00:54

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.241	22:56
Breath Volume	1.480	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.247	22:59
Breath Volume	1.597	22:59
Air Blank	0.000	22:59
Control Test	0.077	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 07/03/2024 13:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/01/2024 01:03:03

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Sequence Aborted
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	

Data Download Date/Time: 06/18/2024 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/01/2024 01:22:26

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.078	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.201	01:17
Breath Volume	1.914	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.196	01:20
Breath Volume	1.625	01:20
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORBIN,BRITTON,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: FRETWELL

DATE/TIME OF TEST: 06/01/2024 22:52:55

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.023	22:48
Breath Volume	2.632	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.021	22:51
Breath Volume	1.554	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 06/02/2024 00:28:31

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.194	00:23
Breath Volume	1.398	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.206	00:26
Breath Volume	2.515	00:26
Air Blank	0.000	00:27
Control Test	0.079	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENAFIELD,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BENAFIELD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 06/09/2024 20:35:14

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.108	20:30
Breath Volume	1.828	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.096	20:33
Breath Volume	2.554	20:33
Air Blank	0.000	20:34
Control Test	0.077	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEGGET,GARRETT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 06/15/2024 16:23:21

Diagnostics Check	OK	16:17
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.247	16:18
Breath Volume	1.843	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.238	16:21
Breath Volume	2.472	16:21
Air Blank	0.000	16:22
Control Test	0.077	16:22
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATERS,DEVON,J
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: WATERS,DEVON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 06/15/2024 19:20:24

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.080	19:15	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:19	Subject Test Refused
Breath Volume	0.000	19:19	
Air Blank	0.000	19:19	
Control Test	0.080	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	OK	19:20	

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MCINTYRE

DATE/TIME OF TEST: 06/16/2024 01:41:38

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.000	01:37
Breath Volume	1.195	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.000	01:40
Breath Volume	1.410	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROEBER,ANTHONY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROEBER,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: MCCLUSKEY

DATE/TIME OF TEST: 06/17/2024 18:25:52

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:13	
Control Test	0.079	18:13	
Air Blank	0.000	18:13	
Subject Sample	# 1 0.327	18:14	
Breath Volume	1.113	18:14	
Air Blank	0.000	18:15	
Air Blank	0.000	18:17	
Subject Sample	# 2 0.285	18:20	Volume Not Met
Breath Volume	0.554	18:20	
Air Blank	0.000	18:21	
Air Blank	0.000	18:23	
Subject Sample	# 3 0.311	18:24	
Breath Volume	1.214	18:24	
Air Blank	0.000	18:24	
Control Test	0.074	18:25	Control Outside Tolerance
Air Blank	0.000	18:25	

Data Download Date/Time: 06/18/2024 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: BUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:51

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 06/29/2024 19:28:30

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 07/08/2024 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: BUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: NESMITH

DATE/TIME OF TEST: 06/09/2024 18:23:56

Diagnostics Check	OK	18:20	
Air Blank	0.000	18:21	
Control Test	0.080	18:21	
Air Blank	0.000	18:22	
Subject Sample	# 1	18:22	Subject Test Refused
Breath Volume	0.000	18:22	
Air Blank	0.000	18:23	
Control Test	0.080	18:23	
Air Blank	0.000	18:23	
Diagnostics Check	OK	18:23	

Data Download Date/Time: 06/18/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESTEK,JOHN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MESTEK,JOHN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 06/13/2024 12:16:43

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.080	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.000	12:12
Breath Volume	4.390	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.000	12:15
Breath Volume	3.550	12:15
Air Blank	0.000	12:15
Control Test	0.078	12:16
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 06/18/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL, N
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/20/2024 05:22:53

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.081	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.000	05:17
Breath Volume	2.191	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.000	05:21
Breath Volume	1.171	05:21
Air Blank	0.000	05:21
Control Test	0.078	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 07/08/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,TIMYRON,N
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: YOUNG,TIMYRON
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HONAKER

DATE/TIME OF TEST: 06/21/2024 21:09:32

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.105	21:04
Breath Volume	1.394	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.103	21:07
Breath Volume	1.621	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 07/08/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/22/2024 00:06:36

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.080	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:02
Breath Volume	2.273	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:05
Breath Volume	2.210	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 07/08/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT,COLTEN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,N
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 06/23/2024 03:35:08

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.079	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	RFI Detect
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 07/08/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN PELT, COLTEN, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL, N
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 06/23/2024 03:43:16

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.125	03:38
Breath Volume	1.113	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.129	03:41
Breath Volume	2.281	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/08/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/01/2024 01:12:30

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.253	01:07
Breath Volume	2.621	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.249	01:10
Breath Volume	2.085	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTRO,SHANE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 06/01/2024 22:45:25

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.211	22:40
Breath Volume	1.554	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.205	22:43
Breath Volume	2.023	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,DANIEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VALDES,DANIEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 06/02/2024 07:29:44

Diagnostics Check	OK	07:23
Air Blank	0.000	07:23
Control Test	0.079	07:23
Air Blank	0.000	07:24
Subject Sample # 1	0.124	07:25
Breath Volume	1.894	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:27
Subject Sample # 2	0.112	07:28
Breath Volume	1.906	07:28
Air Blank	0.000	07:28
Control Test	0.078	07:29
Air Blank	0.000	07:29
Diagnostics Check	OK	07:29

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 06/05/2024 03:54:05

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.081	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	Subject Test Refused
Breath Volume	0.000	03:52	
Air Blank	0.000	03:53	
Control Test	0.081	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,SHANE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORVEC,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: PASCUAL

DATE/TIME OF TEST: 06/06/2024 20:37:25

Diagnostics Check	OK	20:34	
Air Blank	0.000	20:34	
Control Test	0.080	20:35	
Air Blank	0.000	20:35	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:36	
Control Test	0.079	20:36	
Air Blank	0.000	20:37	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUFIN,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARDNER,LANDON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: CANALS VALDEZ

DATE/TIME OF TEST: 06/07/2024 17:37:42

Diagnostics Check	OK	17:34	
Air Blank	0.000	17:34	
Control Test	0.080	17:34	
Air Blank	0.000	17:35	
Subject Sample	# 1	17:36	Subject Test Refused
Breath Volume	0.000	17:36	
Air Blank	0.000	17:36	
Control Test	0.080	17:37	
Air Blank	0.000	17:37	
Diagnostics Check	OK	17:37	

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SYLVESTRE

DATE/TIME OF TEST: 06/10/2024 14:56:21

Diagnostics Check	Fail	14:55	Temperature Regulation Test
Air Blank	0.000	14:56	

Data Download Date/Time: 06/14/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: SLVESTRE

DATE/TIME OF TEST: 06/10/2024 15:04:10

Diagnostics Check	OK	14:58
Air Blank	0.000	14:58
Control Test	0.081	14:58
Air Blank	0.000	14:59
Subject Sample # 1	0.161	14:59
Breath Volume	1.394	14:59
Air Blank	0.000	15:00
Air Blank	0.000	15:02
Subject Sample # 2	0.152	15:02
Breath Volume	1.625	15:02
Air Blank	0.000	15:03
Control Test	0.079	15:03
Air Blank	0.000	15:04
Diagnostics Check	OK	15:04

Data Download Date/Time: 06/14/2024 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 06/11/2024 21:43:47

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.208	21:38
Breath Volume	1.250	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.208	21:42
Breath Volume	2.117	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 06/14/2024 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GARDNER,LANDON
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 05/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 06/12/2024 23:29:29

Diagnostics Check	OK	23:13	
Air Blank	0.000	23:14	
Control Test	0.080	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.167	23:18	Volume Not Met
Breath Volume	0.023	23:18	
Air Blank	0.000	23:18	
Air Blank	0.000	23:20	
Subject Sample	# 2 0.171	23:23	
Breath Volume	1.476	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 3 0.182	23:27	
Breath Volume	1.464	23:27	
Air Blank	0.000	23:28	
Control Test	0.076	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 06/14/2024 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 06/14/2024 23:57:46

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:53	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Control Test	0.079	23:57	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JAMIE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MILLER,JAMIE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 06/15/2024 20:23:44

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:19
Breath Volume	1.640	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.000	20:22
Breath Volume	1.707	20:22
Air Blank	0.000	20:22
Control Test	0.080	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ABSTEN

DATE/TIME OF TEST: 06/17/2024 21:46:14

Diagnostics Check	Fail	21:45	Temperature Regulation Test
Air Blank	0.000	21:46	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ABSTEN

DATE/TIME OF TEST: 06/17/2024 22:00:39

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.080	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1 0.077	21:53	Volume Not Met
Breath Volume	0.445	21:53	
Air Blank	0.000	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 2 0.086	21:59	Volume Not Met
Breath Volume	0.746	21:59	
Air Blank	0.000	21:59	
Control Test	0.079	22:00	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: ABSTEN

DATE/TIME OF TEST: 06/17/2024 22:09:52

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.078	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.113	22:05
Breath Volume	1.187	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.123	22:08
Breath Volume	1.214	22:08
Air Blank	0.000	22:08
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 06/21/2024 22:17:14

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:15	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,DAVID,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUZMAN,DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GARAY JIMENEZ

DATE/TIME OF TEST: 06/23/2024 02:12:35

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:05	No Sample Provided
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Control Test	0.080	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BURGAN

DATE/TIME OF TEST: 06/23/2024 04:30:35

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.100	04:26
Breath Volume	1.281	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.097	04:29
Breath Volume	1.410	04:29
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: OLIVO

DATE/TIME OF TEST: 06/29/2024 01:57:35

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:50	No Sample Provided
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 2	0.159	01:52
Breath Volume	2.277	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 3	0.156	01:56
Breath Volume	2.234	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILROY,ZACHARY,I
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SMITS,JESSICA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/29/2024 21:26:14

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.081	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.240	21:18
Breath Volume	3.261	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.214	21:21
Breath Volume	1.421	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 3	0.237	21:24
Breath Volume	1.355	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 06/30/2024 00:32:57

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEJIA,ELVIS,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: THOMAS,CALEB
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 06/30/2024 19:29:58

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.079	19:29	
Air Blank	0.000	19:29	
Diagnostics Check	OK	19:29	

Data Download Date/Time: 07/02/2024 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 06/04/2024 19:45:13

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:42	
Control Test	0.078	19:42	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Control Test	0.078	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 07/12/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 06/14/2024 01:41:17

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.078	01:39	
Air Blank	0.000	01:39	
Subject Sample # 1	0.191	01:40	
Breath Volume	1.734	01:40	
Air Blank		01:40	RFI Detect
Air Blank		01:40	RFI Detect
Air Blank	0.000	01:41	

Data Download Date/Time: 07/12/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 06/14/2024 01:49:29

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.077	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.207	01:44
Breath Volume	3.843	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.190	01:47
Breath Volume	2.792	01:47
Air Blank	0.000	01:48
Control Test	0.077	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/12/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 06/15/2024 03:43:26

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.221	03:37
Breath Volume	1.738	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.217	03:41
Breath Volume	2.832	03:41
Air Blank	0.000	03:42
Control Test	0.078	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/12/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MOFFAT

DATE/TIME OF TEST: 06/16/2024 23:18:05

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.223	23:13
Breath Volume	2.960	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.224	23:16
Breath Volume	3.105	23:16
Air Blank	0.000	23:17
Control Test	0.077	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:18

Data Download Date/Time: 07/12/2024 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SCHERMERHORN

DATE/TIME OF TEST: 06/18/2024 03:09:35

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.235	03:04
Breath Volume	3.933	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.224	03:07
Breath Volume	3.007	03:07
Air Blank	0.000	03:08
Control Test	0.078	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/12/2024 14:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MENDOZA ROMULO

DATE/TIME OF TEST: 06/15/2024 01:26:13

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:20
Subject Sample # 1	0.245	01:21
Breath Volume	1.660	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.249	01:24
Breath Volume	2.304	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: SANTOS RAMIREZ

DATE/TIME OF TEST: 06/15/2024 02:14:05

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.167	02:09
Breath Volume	2.343	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.171	02:12
Breath Volume	2.980	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LUIS MARTINEZ

DATE/TIME OF TEST: 06/16/2024 01:56:22

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.242	01:51
Breath Volume	2.101	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.241	01:54
Breath Volume	2.527	01:54
Air Blank	0.000	01:55
Control Test	0.077	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOSEPH,JOEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 06/16/2024 08:02:15

Diagnostics Check	OK	07:55
Air Blank	0.000	07:55
Control Test	0.079	07:56
Air Blank	0.000	07:56
Subject Sample # 1	0.169	07:57
Breath Volume	1.261	07:57
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.168	08:00
Breath Volume	2.042	08:00
Air Blank	0.000	08:01
Control Test	0.078	08:01
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: CORDERO BARRIOS

DATE/TIME OF TEST: 06/17/2024 09:28:10

Diagnostics Check	OK	09:20
Air Blank	0.000	09:20
Control Test	0.079	09:21
Air Blank	0.000	09:21
Subject Sample # 1	0.162	09:22
Breath Volume	1.207	09:22
Air Blank	0.000	09:23
Air Blank	0.000	09:25
Subject Sample # 2	0.162	09:26
Breath Volume	1.890	09:26
Air Blank	0.000	09:27
Control Test	0.079	09:27
Air Blank	0.000	09:28
Diagnostics Check	OK	09:28

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FENDER, NICHOLAS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: RIOS LA LUZ

DATE/TIME OF TEST: 06/19/2024 23:05:52

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.080	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.000	23:01
Breath Volume	1.886	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.000	23:04
Breath Volume	1.222	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR, ROBINSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 06/20/2024 01:36:07

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.150	01:31
Breath Volume	1.976	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.144	01:34
Breath Volume	2.781	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU, YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 06/21/2024 03:59:55

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.122	03:55
Breath Volume	2.269	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.117	03:58
Breath Volume	1.195	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MARTINEZ GARCIA

DATE/TIME OF TEST: 06/22/2024 00:40:53

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.189	00:35
Breath Volume	1.390	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.171	00:39
Breath Volume	2.007	00:39
Air Blank	0.000	00:39
Control Test	0.077	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HERRERA CIFUENTES

DATE/TIME OF TEST: 06/22/2024 01:56:08

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.185	01:51
Breath Volume	1.789	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.179	01:54
Breath Volume	1.980	01:54
Air Blank	0.000	01:55
Control Test	0.077	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROSARIO, MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 06/22/2024 02:44:13

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DOMINGO FELIPE

DATE/TIME OF TEST: 06/22/2024 04:44:42

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.191	04:40
Breath Volume	1.804	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.193	04:43
Breath Volume	2.656	04:43
Air Blank	0.000	04:43
Control Test	0.078	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RIDOUT

DATE/TIME OF TEST: 06/23/2024 04:03:06

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.252	03:58
Breath Volume	2.242	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.254	04:01
Breath Volume	2.394	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PICARD,RONNIE
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MENDOZA CHOLOTIO

DATE/TIME OF TEST: 06/26/2024 03:11:53

Diagnostics Check	OK	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.130	03:06
Breath Volume	1.621	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.132	03:10
Breath Volume	2.218	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 07/26/2024 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DAGHER

DATE/TIME OF TEST: 06/27/2024 02:37:49

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.042	02:33
Breath Volume	2.417	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.041	02:36
Breath Volume	2.332	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/26/2024 13:39