

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: PIDGEON,ANDREW
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:41

SUBJECT'S LAST NAME: MELCHER

DATE/TIME OF TEST: 07/04/2024 09:10:45

Diagnostics Check	OK	09:05	
Air Blank	0.000	09:05	
Control Test	0.079	09:06	
Air Blank	0.000	09:06	
Subject Sample	# 1 0.110	09:07	
Breath Volume	1.855	09:07	
Air Blank	0.000	09:07	
Air Blank	0.000	09:09	
Subject Sample	# 2 0.115	09:10	
Breath Volume	1.882	09:10	
Air Blank		09:10	RFI Detect
Air Blank	0.000	09:10	

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: TODD,N

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 20:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: BIENES

DATE/TIME OF TEST: 07/04/2024 21:22:10

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.141	21:16
Breath Volume	2.031	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.138	21:20
Breath Volume	1.738	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO, WESLEY, G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO, WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 04:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: DOMINGO VELASQUEZ

DATE/TIME OF TEST: 07/05/2024 05:13:26

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.222	05:08
Breath Volume	2.972	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.206	05:11
Breath Volume	3.152	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELBIG,DEREK,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: APONTE,DANIEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 07/13/2024 05:54:48

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:49
Subject Sample # 1	0.157	05:50
Breath Volume	1.457	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:52
Subject Sample # 2	0.163	05:53
Breath Volume	2.324	05:53
Air Blank	0.000	05:53
Control Test	0.077	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:54

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: OGBORN

DATE/TIME OF TEST: 07/19/2024 21:49:09

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.144	21:44
Breath Volume	1.601	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.150	21:47
Breath Volume	2.011	21:47
Air Blank	0.000	21:48
Control Test	0.079	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DUMONT,TRAVIS,H

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: SHUCK,MONICA

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 04:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: EDMONDSON

DATE/TIME OF TEST: 07/20/2024 05:10:09

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.079	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.133	05:05
Breath Volume	2.542	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.125	05:08
Breath Volume	1.105	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCCLANAHAN

DATE/TIME OF TEST: 07/21/2024 02:41:05

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.081	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:34	No Sample Provided
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:39	Subject Test Refused
Breath Volume	0.117	02:39	
Air Blank	0.000	02:40	
Control Test	0.081	02:40	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARNETT,JAMES,C
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: OOLEY,JAMES
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: M245112857090

DATE/TIME OF TEST: 07/21/2024 02:54:08

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.110	02:47
Breath Volume	1.695	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.110	02:52
Breath Volume	1.367	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO,WESLEY,G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/21/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 01:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/25/2024 02:32:26

Diagnostics Check	OK	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.144	02:27
Breath Volume	2.996	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.143	02:30
Breath Volume	2.718	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 07/28/2024 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLLY,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 07/01/2024 22:08:23

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.000	22:04
Breath Volume	2.734	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.000	22:06
Breath Volume	1.781	22:06
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 07/16/2024 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,MOSES,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LOLLIO

DATE/TIME OF TEST: 07/03/2024 00:25:21

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.267	00:17
Breath Volume	1.496	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.230	00:20
Breath Volume	1.195	00:20
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 3	0.258	00:23
Breath Volume	2.531	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS, MOSES, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY, DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: OROZCO SERVIN

DATE/TIME OF TEST: 07/03/2024 04:12:07

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.231	04:06
Breath Volume	1.257	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.247	04:10
Breath Volume	2.339	04:10
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS, MOSES, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROOS, JOSHUA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 07/07/2024 00:58:17

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.251	00:53
Breath Volume	2.761	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.253	00:56
Breath Volume	2.781	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,MOSES,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,DEREK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: LUX RODRIGUEZ

DATE/TIME OF TEST: 07/07/2024 01:13:28

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.208	01:08
Breath Volume	3.371	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.218	01:11
Breath Volume	3.863	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS, MOSES, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BERMUDEZ, LAWRENCE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GUTIERREZ MENDOZA

DATE/TIME OF TEST: 07/07/2024 02:53:44

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.196	02:48
Breath Volume	2.078	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.188	02:51
Breath Volume	3.101	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINEBERGER,RYAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: GARCIA CASTILLO

DATE/TIME OF TEST: 07/07/2024 05:37:53

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.051	05:33
Breath Volume	1.773	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:35
Subject Sample # 2	0.052	05:36
Breath Volume	1.449	05:36
Air Blank	0.000	05:36
Control Test	0.080	05:37
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,MOSES,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: GUZMAN PAREDES

DATE/TIME OF TEST: 07/08/2024 02:47:07

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.179	02:42
Breath Volume	1.386	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.176	02:45
Breath Volume	1.628	02:45
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,MOSES,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PABLO

DATE/TIME OF TEST: 07/11/2024 03:39:00

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.153	03:34
Breath Volume	2.683	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.148	03:37
Breath Volume	2.554	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLFF,NATHANIEL,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINEBERGER,RYAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/11/2024 05:57:32

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.170	05:52
Breath Volume	3.675	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.180	05:55
Breath Volume	3.218	05:55
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ULLOA

DATE/TIME OF TEST: 07/12/2024 03:21:24

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.101	03:16
Breath Volume	2.273	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.099	03:19
Breath Volume	2.652	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GEHLE,STEPHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SHERLOCK-MASON

DATE/TIME OF TEST: 07/12/2024 21:42:59

Diagnostics Check	OK	21:36
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.124	21:38
Breath Volume	1.140	21:38
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.121	21:41
Breath Volume	1.179	21:41
Air Blank	0.000	21:42
Control Test	0.080	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HARDENFELDER, GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/13/2024 02:23:33

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.080	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 07/14/2024 00:07:05

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.109	00:02
Breath Volume	1.351	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.114	00:05
Breath Volume	2.539	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KMAPP,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KOPEC

DATE/TIME OF TEST: 07/14/2024 02:48:12

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.146	02:43
Breath Volume	1.882	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.143	02:46
Breath Volume	1.828	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: TOTA

DATE/TIME OF TEST: 07/14/2024 05:56:56

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.080	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.198	05:52
Breath Volume	3.070	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.198	05:55
Breath Volume	3.214	05:55
Air Blank	0.000	05:56
Control Test	0.079	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COMAS ZALDIVAR

DATE/TIME OF TEST: 07/15/2024 02:23:13

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.011	02:18
Breath Volume	4.343	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.009	02:21
Breath Volume	4.078	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WEISS,MASON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MARIN-ALDANA

DATE/TIME OF TEST: 07/15/2024 02:50:26

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.175	02:45
Breath Volume	4.902	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.195	02:48
Breath Volume	5.269	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 07/16/2024 12:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SHERRILL

DATE/TIME OF TEST: 07/18/2024 01:58:56

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.120	01:54
Breath Volume	2.683	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.122	01:57
Breath Volume	3.253	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK, KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PAVON ORTIZ

DATE/TIME OF TEST: 07/18/2024 03:25:18

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.126	03:20
Breath Volume	3.085	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.129	03:23
Breath Volume	3.140	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MULET SUAREZ

DATE/TIME OF TEST: 07/19/2024 05:29:03

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.112	05:24
Breath Volume	1.183	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.103	05:27
Breath Volume	1.195	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: BLIER

DATE/TIME OF TEST: 07/20/2024 02:45:49

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.101	02:41
Breath Volume	1.718	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.102	02:44
Breath Volume	1.750	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HILL,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/20/2024 04:06:54

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.219	04:02
Breath Volume	2.476	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:04
Subject Sample # 2	0.218	04:05
Breath Volume	2.339	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MONCLERC

DATE/TIME OF TEST: 07/20/2024 04:43:02

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.056	04:38
Breath Volume	2.691	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.055	04:41
Breath Volume	2.386	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:43

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CLEMENTS,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: GREENAWAY

DATE/TIME OF TEST: 07/20/2024 23:33:18

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.120	23:25
Breath Volume	1.261	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.144	23:28
Breath Volume	2.140	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 3	0.127	23:31
Breath Volume	1.296	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/21/2024 03:31:46

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.149	03:26
Breath Volume	1.656	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.145	03:30
Breath Volume	1.292	03:30
Air Blank	0.000	03:30
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MATERA,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CASTRO MEREGILDO

DATE/TIME OF TEST: 07/22/2024 00:56:09

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.161	00:51
Breath Volume	1.695	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.158	00:54
Breath Volume	2.199	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SELVA

DATE/TIME OF TEST: 07/24/2024 03:48:48

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.147	03:44
Breath Volume	4.281	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.142	03:47
Breath Volume	3.472	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: TELLIN

DATE/TIME OF TEST: 07/26/2024 03:11:07

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.141	03:06
Breath Volume	2.273	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.148	03:09
Breath Volume	2.082	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 08/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KITTLESON,SEAN
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: GALL

DATE/TIME OF TEST: 07/26/2024 14:57:08

Diagnostics Check	OK	14:54	
Air Blank	0.000	14:54	
Control Test	0.081	14:55	
Air Blank	0.000	14:55	
Subject Sample	# 1	14:55	Subject Test Refused
Breath Volume	0.000	14:55	
Air Blank	0.000	14:56	
Control Test	0.081	14:56	
Air Blank	0.000	14:57	
Diagnostics Check	OK	14:57	

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PANOZO LAREDO

DATE/TIME OF TEST: 07/27/2024 03:48:54

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.165	03:44
Breath Volume	2.277	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.165	03:47
Breath Volume	2.093	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SHOEMAKER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/27/2024 05:29:25

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.151	05:24
Breath Volume	2.078	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.145	05:27
Breath Volume	1.566	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: CHIAPPY,CHRISTOFER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 07/27/2024 20:02:05

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:56
Subject Sample # 1	0.146	19:57
Breath Volume	4.835	19:57
Air Blank	0.000	19:58
Air Blank	0.000	19:59
Subject Sample # 2	0.146	20:00
Breath Volume	5.042	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER, JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: AGUSTIN MENDOZA

DATE/TIME OF TEST: 07/28/2024 00:20:33

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:15	
Control Test	0.080	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.273	00:18	
Air Blank	0.000	00:19	
Control Test	0.080	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 07/28/2024 04:03:26

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.148	03:58
Breath Volume	4.003	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.144	04:01
Breath Volume	3.300	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KNAPP, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: DIAZ LOPEZ

DATE/TIME OF TEST: 07/28/2024 18:53:14

Diagnostics Check	OK	18:46
Air Blank	0.000	18:47
Control Test	0.080	18:47
Air Blank	0.000	18:48
Subject Sample # 1	0.150	18:48
Breath Volume	3.503	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.155	18:51
Breath Volume	3.187	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHOEMAKER, JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: LANTZ

DATE/TIME OF TEST: 07/28/2024 19:46:12

Diagnostics Check	OK	19:39
Air Blank	0.000	19:39
Control Test	0.080	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.223	19:41
Breath Volume	3.394	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.220	19:44
Breath Volume	3.488	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BOMMARITO-MALTA

DATE/TIME OF TEST: 07/29/2024 02:55:22

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.189	02:50
Breath Volume	2.503	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.191	02:53
Breath Volume	2.441	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF, LANCE, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CROTHERS, AUSTIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: YURGELAITIS

DATE/TIME OF TEST: 07/29/2024 04:44:52

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.080	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.220	04:40
Breath Volume	2.187	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.232	04:43
Breath Volume	2.972	04:43
Air Blank	0.000	04:43
Control Test	0.079	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 08/26/2024 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CLIFFORD,UNK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: ROCHE

DATE/TIME OF TEST: 07/04/2024 21:02:58

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.240	20:58
Breath Volume	2.265	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.250	21:01
Breath Volume	2.019	21:01
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PARRISH,UNK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/05/2024 00:59:54

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.177	00:55
Breath Volume	1.660	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.182	00:58
Breath Volume	1.429	00:58
Air Blank	0.000	00:59
Control Test	0.081	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK',
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BARONE,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: AMODEI

DATE/TIME OF TEST: 07/05/2024 22:21:49

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.081	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.172	22:17
Breath Volume	3.761	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.159	22:20
Breath Volume	2.882	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FERNANDEZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/05/2024 22:59:30

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.000	22:55
Breath Volume	2.062	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.000	22:58
Breath Volume	2.164	22:58
Air Blank	0.000	22:58
Control Test	0.082	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FERNANDEZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: OLSEN JOHNSON

DATE/TIME OF TEST: 07/07/2024 00:05:54

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.082	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.178	00:01
Breath Volume	2.046	00:01
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.169	00:04
Breath Volume	2.250	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRIEND,S
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 07/07/2024 03:05:41

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.053	03:01
Breath Volume	1.722	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.052	03:04
Breath Volume	2.582	03:04
Air Blank	0.000	03:04
Control Test	0.081	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTOFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: YOHO

DATE/TIME OF TEST: 07/13/2024 16:15:03

Diagnostics Check	OK	16:08
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:10
Subject Sample # 1	0.161	16:10
Breath Volume	1.894	16:10
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.154	16:13
Breath Volume	1.691	16:13
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:14
Diagnostics Check	OK	16:15

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIS,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: OGRODOWSKI

DATE/TIME OF TEST: 07/20/2024 16:19:55

Diagnostics Check	OK	16:13
Air Blank	0.000	16:13
Control Test	0.081	16:13
Air Blank	0.000	16:14
Subject Sample # 1	0.208	16:15
Breath Volume	2.453	16:15
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.206	16:18
Breath Volume	2.949	16:18
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARARO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 07/21/2024 03:39:29

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.081	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.201	03:34
Breath Volume	2.062	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.190	03:37
Breath Volume	1.582	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 07/21/2024 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DIPIAZZA,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BEHPOUR

DATE/TIME OF TEST: 07/21/2024 22:08:54

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.086	22:03
Breath Volume	1.265	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.085	22:07
Breath Volume	1.718	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JOHN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORRITZ,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/22/2024 00:15:35

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.207	00:11	Volume Not Met
Breath Volume	0.050	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEE,E
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SANTIAGO MALTES

DATE/TIME OF TEST: 07/24/2024 22:22:32

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.200	22:17
Breath Volume	2.367	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.210	22:21
Breath Volume	2.218	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,TAMARA,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WYMER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 07/25/2024 03:30:47

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.192	03:26
Breath Volume	2.558	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.190	03:29
Breath Volume	2.003	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CCSO,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BELLO,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 07/27/2024 04:25:46

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.127	04:21
Breath Volume	1.554	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.130	04:24
Breath Volume	2.636	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CCSO K,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DODSON,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: SNOWDON

DATE/TIME OF TEST: 07/27/2024 21:11:45

Diagnostics Check	OK	21:03
Air Blank	0.000	21:04
Control Test	0.082	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.195	21:05
Breath Volume	1.859	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.187	21:10
Breath Volume	1.683	21:10
Air Blank	0.000	21:10
Control Test	0.081	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SITKO,DEP.
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREZ FLORES

DATE/TIME OF TEST: 07/28/2024 02:12:33

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.156	02:07
Breath Volume	1.687	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.165	02:11
Breath Volume	1.742	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BELLO-RUIZ,Y
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: LUX-TUM

DATE/TIME OF TEST: 07/29/2024 00:51:56

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.081	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.213	00:47
Breath Volume	2.843	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.215	00:50
Breath Volume	2.539	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 08/15/2024 22:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,OFFICER,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/11/2024 19:01:47

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.077	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:56
Breath Volume	1.105	18:56
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.000	19:00
Breath Volume	1.121	19:00
Air Blank	0.000	19:00
Control Test	0.079	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HITE

DATE/TIME OF TEST: 07/13/2024 01:11:48

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.101	01:07
Breath Volume	2.121	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.093	01:10
Breath Volume	1.195	01:10
Air Blank	0.000	01:10
Control Test	0.077	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELATORRE-CASTELLANO,OFC,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: MOSQUERA JIMENEZ

DATE/TIME OF TEST: 07/13/2024 05:55:57

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.115	05:51
Breath Volume	1.894	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.115	05:54
Breath Volume	2.191	05:54
Air Blank	0.000	05:55
Control Test	0.078	05:55
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REICH,OWEN,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,OWEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/14/2024 00:05:34

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 2	23:58	No Sample Provided
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:03	Volume Not Met
Breath Volume	0.777	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REICH,OWEN,S
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REICH,OWEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/14/2024 00:15:29

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.078	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.103	00:10
Breath Volume	1.953	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.111	00:13
Breath Volume	3.015	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/20/2024 03:24:40

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.122	03:20
Breath Volume	2.613	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.124	03:23
Breath Volume	2.253	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 08/20/2024 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: RESENDIZ RESENDIZ

DATE/TIME OF TEST: 07/20/2024 23:49:08

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.177	23:44
Breath Volume	1.550	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.175	23:47
Breath Volume	1.875	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SAMBHI

DATE/TIME OF TEST: 07/21/2024 02:36:36

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.223	02:31
Breath Volume	1.449	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.223	02:34
Breath Volume	1.945	02:34
Air Blank	0.000	02:35
Control Test	0.078	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,JOEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DE LA ROSA

DATE/TIME OF TEST: 07/21/2024 03:01:22

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.077	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.136	02:56
Breath Volume	1.371	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.138	02:59
Breath Volume	2.082	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/23/2024 23:00:08

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:54	
Control Test	0.079	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1 0.158	22:55	
Breath Volume	1.593	22:55	
Air Blank	0.000	22:56	
Air Blank	0.000	22:58	
Subject Sample	# 2	22:58	RFI Detect
Breath Volume	2.660	22:58	
Air Blank	0.000	22:59	
Control Test	0.078	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/23/2024 23:09:14

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Subject Sample # 1	0.154	23:04
Breath Volume	2.292	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.150	23:07
Breath Volume	2.562	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELCHNER,M,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KELCHNER,M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ROSSY

DATE/TIME OF TEST: 07/25/2024 04:32:40

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.153	04:27
Breath Volume	2.894	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.155	04:31
Breath Volume	3.242	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELCHNER,M,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KELCHNER,M
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 07/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CORTES JERONIMO

DATE/TIME OF TEST: 07/28/2024 01:48:53

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.120	01:44
Breath Volume	2.484	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.122	01:47
Breath Volume	1.652	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 08/20/2024 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 07/03/2024 03:46:18

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.153	03:40
Breath Volume	1.414	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.148	03:44
Breath Volume	1.347	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 07/03/2024 19:51:54

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.079	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.262	19:47
Breath Volume	1.656	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.259	19:50
Breath Volume	1.558	19:50
Air Blank	0.000	19:50
Control Test	0.078	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD, EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: UTUY

DATE/TIME OF TEST: 07/03/2024 22:02:14

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.264	21:57
Breath Volume	2.593	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.282	22:00
Breath Volume	2.933	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/04/2024 01:49:00

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.122	01:44
Breath Volume	1.433	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.114	01:47
Breath Volume	1.355	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:49

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GREENBERG

DATE/TIME OF TEST: 07/04/2024 03:32:21

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:28
Breath Volume	1.390	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:30
Breath Volume	2.117	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAMILTON,CORY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ANDRES MARTIN

DATE/TIME OF TEST: 07/04/2024 05:41:15

Diagnostics Check	OK	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Subject Sample # 1	0.277	05:35
Breath Volume	1.292	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.277	05:39
Breath Volume	1.621	05:39
Air Blank	0.000	05:40
Control Test	0.078	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVIS, MILJEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: DEROSIER

DATE/TIME OF TEST: 07/05/2024 01:06:48

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.208	01:01
Breath Volume	1.312	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.201	01:05
Breath Volume	1.226	01:05
Air Blank	0.000	01:05
Control Test	0.079	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HEY

DATE/TIME OF TEST: 07/05/2024 01:35:30

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.204	01:31
Breath Volume	3.429	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.199	01:33
Breath Volume	1.921	01:33
Air Blank	0.000	01:34
Control Test	0.078	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SANITZO GONZALES

DATE/TIME OF TEST: 07/05/2024 04:32:15

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.140	04:27
Breath Volume	2.660	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.135	04:30
Breath Volume	2.343	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RENTERIA, ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: MIRSKY

DATE/TIME OF TEST: 07/05/2024 05:30:23

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.079	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:29	
Control Test	0.080	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 07/06/2024 01:41:14

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.162	01:33	Volume Not Met
Breath Volume	1.097	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 2 0.171	01:36	
Breath Volume	1.792	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 3 0.178	01:39	
Breath Volume	2.734	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: FELIX JUAN

DATE/TIME OF TEST: 07/06/2024 05:27:51

Diagnostics Check	OK	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.197	05:22
Breath Volume	2.320	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.206	05:26
Breath Volume	2.976	05:26
Air Blank	0.000	05:26
Control Test	0.078	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 07/06/2024 23:31:32

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.327	23:23
Breath Volume	1.484	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.302	23:26
Breath Volume	1.597	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 3	0.313	23:29
Breath Volume	1.328	23:29
Air Blank	0.000	23:30
Control Test	0.077	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 07/07/2024 02:05:39

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.158	02:01
Breath Volume	1.156	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.158	02:04
Breath Volume	1.414	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MARCOS CARDONA

DATE/TIME OF TEST: 07/07/2024 04:00:23

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.079	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.197	03:55
Breath Volume	2.460	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.201	03:58
Breath Volume	2.863	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: REYES MONTES

DATE/TIME OF TEST: 07/08/2024 00:15:33

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.079	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.093	00:11
Breath Volume	1.250	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.094	00:13
Breath Volume	1.785	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 07/08/2024 06:56:43

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.080	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.087	06:52
Breath Volume	1.667	06:52
Air Blank	0.000	06:52
Air Blank	0.000	06:54
Subject Sample # 2	0.086	06:55
Breath Volume	2.242	06:55
Air Blank	0.000	06:55
Control Test	0.080	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BOLAND, CONNOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/08/2024 16:36:33

Diagnostics Check	OK	16:30
Air Blank	0.000	16:30
Control Test	0.080	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.098	16:31
Breath Volume	1.996	16:31
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.091	16:34
Breath Volume	1.457	16:34
Air Blank	0.000	16:35
Control Test	0.080	16:35
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO, AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: JEAN NOEL

DATE/TIME OF TEST: 07/08/2024 22:09:01

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.216	22:04
Breath Volume	1.199	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:06
Subject Sample # 2	0.225	22:07
Breath Volume	1.296	22:07
Air Blank	0.000	22:08
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:09

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MORALES RAMIREZ

DATE/TIME OF TEST: 07/08/2024 23:35:30

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.221	23:30
Breath Volume	2.531	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.223	23:33
Breath Volume	2.339	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/09/2024 02:15:07

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:10
Breath Volume	2.656	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	2.101	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CANTONE

DATE/TIME OF TEST: 07/09/2024 21:54:02

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.254	21:48
Breath Volume	2.109	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.257	21:52
Breath Volume	1.339	21:52
Air Blank	0.000	21:53
Control Test	0.079	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:54

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MOLL

DATE/TIME OF TEST: 07/11/2024 02:58:29

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.135	02:53
Breath Volume	1.730	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.132	02:56
Breath Volume	1.449	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PASCUAL FRANCISCO

DATE/TIME OF TEST: 07/11/2024 23:42:22

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.128	23:37
Breath Volume	1.855	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.124	23:40
Breath Volume	2.015	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BOLAND,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 07/12/2024 23:21:47

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.189	23:17
Breath Volume	2.214	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.171	23:20
Breath Volume	1.484	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: THESEE,JACQUES
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/13/2024 04:28:25

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.045	04:23
Breath Volume	1.507	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.046	04:26
Breath Volume	1.750	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CATHERMAN

DATE/TIME OF TEST: 07/14/2024 02:34:39

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.141	02:29
Breath Volume	3.535	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.136	02:33
Breath Volume	3.324	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CARTAGENA JACKSON

DATE/TIME OF TEST: 07/14/2024 04:35:18

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.212	04:30
Breath Volume	1.847	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.201	04:33
Breath Volume	1.812	04:33
Air Blank	0.000	04:34
Control Test	0.079	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: JONATHAN

DATE/TIME OF TEST: 07/14/2024 05:32:20

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.080	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.163	05:27
Breath Volume	1.570	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.164	05:30
Breath Volume	1.699	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO, ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: FELIZ GOMEZ

DATE/TIME OF TEST: 07/14/2024 16:45:40

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.181	16:41
Breath Volume	1.167	16:41
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.170	16:44
Breath Volume	1.148	16:44
Air Blank	0.000	16:44
Control Test	0.079	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CONROY

DATE/TIME OF TEST: 07/14/2024 22:45:29

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.146	22:40
Breath Volume	1.437	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.144	22:43
Breath Volume	1.390	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MURDOCK,LOGAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CHAMPAGNE

DATE/TIME OF TEST: 07/15/2024 01:11:04

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.092	01:06
Breath Volume	1.734	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.095	01:09
Breath Volume	1.273	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:11

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: HENRILUS

DATE/TIME OF TEST: 07/15/2024 02:16:07

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.137	02:11
Breath Volume	2.175	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.137	02:14
Breath Volume	1.882	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORRISON,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GAMARRA

DATE/TIME OF TEST: 07/15/2024 11:21:43

Diagnostics Check	OK	11:15
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:17
Breath Volume	2.300	11:17
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:20
Breath Volume	2.644	11:20
Air Blank	0.000	11:20
Control Test	0.080	11:21
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COLON,VICTOR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 07/16/2024 00:56:30

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.208	00:52
Breath Volume	1.742	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.205	00:54
Breath Volume	1.585	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: WHITTINGHAM

DATE/TIME OF TEST: 07/16/2024 01:52:16

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:47
Breath Volume	2.109	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:50
Breath Volume	1.761	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD, ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: BARRESI

DATE/TIME OF TEST: 07/16/2024 23:05:59

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.080	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ,DINEEN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VELASQUEZ PEREZ

DATE/TIME OF TEST: 07/17/2024 01:23:26

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.257	01:18
Breath Volume	2.625	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.250	01:21
Breath Volume	2.382	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO, AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: GRAEVE

DATE/TIME OF TEST: 07/17/2024 18:46:03

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.080	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.098	18:41
Breath Volume	2.699	18:41
Air Blank	0.000	18:41
Air Blank	0.000	18:43
Subject Sample # 2	0.094	18:44
Breath Volume	1.570	18:44
Air Blank	0.000	18:45
Control Test	0.079	18:45
Air Blank	0.000	18:45
Diagnostics Check	OK	18:46

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NEWTON,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MERIDA ORTIZ

DATE/TIME OF TEST: 07/20/2024 00:33:44

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.250	00:28
Breath Volume	2.019	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.240	00:32
Breath Volume	1.578	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BOLAND,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 07/20/2024 01:59:11

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.123	01:54
Breath Volume	2.722	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.119	01:57
Breath Volume	2.468	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LOPEZ RAMOS

DATE/TIME OF TEST: 07/20/2024 02:39:58

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.141	02:35
Breath Volume	1.410	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.152	02:38
Breath Volume	2.070	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STRIPLING,CASEY
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: TIRANO PAREDES

DATE/TIME OF TEST: 07/20/2024 08:47:06

Diagnostics Check	OK	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Subject Sample # 1	0.192	08:42
Breath Volume	1.433	08:42
Air Blank	0.000	08:43
Air Blank	0.000	08:45
Subject Sample # 2	0.199	08:45
Breath Volume	1.835	08:45
Air Blank	0.000	08:46
Control Test	0.080	08:46
Air Blank	0.000	08:47
Diagnostics Check	OK	08:47

Data Download Date/Time: 07/26/2024 16:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DENAUD

DATE/TIME OF TEST: 07/21/2024 00:46:34

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.256	00:41
Breath Volume	2.441	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.258	00:44
Breath Volume	1.414	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,EDWARD
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 07/21/2024 01:51:12

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.165	01:46
Breath Volume	2.519	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.171	01:49
Breath Volume	2.625	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 07/21/2024 02:53:22

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.153	02:48
Breath Volume	1.546	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.153	02:51
Breath Volume	1.667	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 07/21/2024 03:29:36

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.106	03:25
Breath Volume	1.363	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.107	03:28
Breath Volume	1.148	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CARRILLO CARRILLO

DATE/TIME OF TEST: 07/21/2024 03:58:10

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.252	03:53
Breath Volume	2.507	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.259	03:56
Breath Volume	1.167	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHIRE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: GABRIEL VELASQUEZ

DATE/TIME OF TEST: 07/21/2024 23:48:48

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.210	23:44
Breath Volume	1.417	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.210	23:47
Breath Volume	2.332	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 07/22/2024 00:12:01

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.234	00:07
Breath Volume	2.484	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.218	00:10
Breath Volume	1.175	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BELLO, LUCIA
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GOLDROD

DATE/TIME OF TEST: 07/22/2024 01:56:05

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.100	01:51
Breath Volume	1.363	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.104	01:54
Breath Volume	1.976	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MORALES DIAZ

DATE/TIME OF TEST: 07/22/2024 02:46:30

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.172	02:41
Breath Volume	1.523	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.175	02:44
Breath Volume	2.085	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: REYES VASQUEZ

DATE/TIME OF TEST: 07/23/2024 22:24:31

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.079	22:20	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.273	22:22	
Air Blank	0.000	22:23	
Control Test	0.080	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALLEN, BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/25/2024 07:48:23

Diagnostics Check	OK	07:41
Air Blank	0.000	07:42
Control Test	0.080	07:42
Air Blank	0.000	07:43
Subject Sample # 1	0.148	07:43
Breath Volume	1.695	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:46
Subject Sample # 2	0.151	07:46
Breath Volume	1.410	07:46
Air Blank	0.000	07:47
Control Test	0.080	07:47
Air Blank	0.000	07:48
Diagnostics Check	OK	07:48

Data Download Date/Time: 07/26/2024 16:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

INSTRUMENT REGISTERED TO: NAS KEY WEST
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,LEE,B
BREATH TEST OPERATOR AGENCY: NAS KEY WEST
ARRESTING OFFICER: DYESS,CONNOR
ARRESTING OFFICER AGENCY: NAS KEY WEST
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1514466
DRY GAS EXPIRY DATE: 06/18/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: GOVERO

DATE/TIME OF TEST: 07/14/2024 04:30:05

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.175	04:25
Breath Volume	3.234	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.184	04:28
Breath Volume	3.390	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:30

Data Download Date/Time: 09/06/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES',LANDON,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: WALKER,DANIEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/06/2024 02:38:07

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.081	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.148	02:33
Breath Volume	1.273	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.143	02:36
Breath Volume	1.398	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 07/23/2024 17:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PIGNATARO, THOMAS
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 07/10/2024 01:38:08

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.208	01:33
Breath Volume	1.843	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.210	01:36
Breath Volume	1.765	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 07/23/2024 17:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNNEFELD,RALPH,A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JOHNSON,REBECCA
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 07/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:51

SUBJECT'S LAST NAME: OFFUTT

DATE/TIME OF TEST: 07/26/2024 09:19:06

Diagnostics Check	OK	09:16	
Air Blank	0.000	09:16	
Control Test	0.080	09:16	
Air Blank	0.000	09:17	
Subject Sample	# 1	09:17	Subject Test Refused
Breath Volume	0.000	09:17	
Air Blank	0.000	09:18	
Control Test	0.081	09:18	
Air Blank	0.000	09:19	
Diagnostics Check	OK	09:19	

Data Download Date/Time: 08/30/2024 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIALLO MORELL,RAUL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: PAVON ROBAINA,JUAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MUNOZ PRESTAMO

DATE/TIME OF TEST: 07/01/2024 05:25:27

Diagnostics Check	OK	05:18
Air Blank	0.000	05:18
Control Test	0.080	05:18
Air Blank	0.000	05:19
Subject Sample # 1	0.259	05:20
Breath Volume	1.195	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:23
Subject Sample # 2	0.260	05:23
Breath Volume	1.148	05:23
Air Blank	0.000	05:24
Control Test	0.077	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 07/11/2024 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,SEAN,A
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ALFARO,DANIEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 07/01/2024 05:55:36

Diagnostics Check	OK	05:52	
Air Blank	0.000	05:53	
Control Test	0.080	05:53	
Air Blank	0.000	05:53	
Subject Sample	# 1	05:54	Subject Test Refused
Breath Volume	0.000	05:54	
Air Blank	0.000	05:54	
Control Test	0.080	05:54	
Air Blank	0.000	05:55	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 07/11/2024 10:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,ADRIAN,H
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: PRATT,MEWZAM
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: KELLAM

DATE/TIME OF TEST: 07/21/2024 10:22:22

Diagnostics Check	OK	10:06	
Air Blank	0.000	10:07	
Control Test	0.081	10:07	
Air Blank	0.000	10:08	
Subject Sample	# 1 0.163	10:11	Volume Not Met
Breath Volume	0.093	10:11	
Air Blank	0.000	10:12	
Air Blank	0.000	10:13	
Subject Sample	# 2 0.194	10:15	
Breath Volume	2.402	10:15	
Air Blank	0.000	10:16	
Air Blank	0.000	10:17	
Subject Sample	# 3 0.190	10:20	
Breath Volume	1.203	10:20	
Air Blank	0.000	10:21	
Control Test	0.075	10:21	
Air Blank	0.000	10:22	
Diagnostics Check	OK	10:22	

Data Download Date/Time: 07/31/2024 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OLSON, D
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: KADUKOV

DATE/TIME OF TEST: 07/01/2024 00:47:35

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.079	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLESTER,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/02/2024 22:57:59

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.152	22:53
Breath Volume	3.203	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.164	22:56
Breath Volume	3.730	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURNS,R,E

BREATH TEST OPERATOR AGENCY: Indian River Co SO

ARRESTING OFFICER: STEWART,S

ARRESTING OFFICER AGENCY: Indian River Shores

LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: DUI

ARREST TIME: 17:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ROMO

DATE/TIME OF TEST: 07/03/2024 18:47:32

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.079	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.000	18:42
Breath Volume	3.207	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.000	18:46
Breath Volume	3.000	18:46
Air Blank	0.000	18:46
Control Test	0.078	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DIBIASIO

DATE/TIME OF TEST: 07/03/2024 20:13:24

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.079	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	
Control Test	0.080	20:12	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,SEBASTIEN,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ,SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 07/05/2024 02:37:52

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.134	02:33
Breath Volume	1.378	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.141	02:36
Breath Volume	3.011	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 07/05/2024 23:02:42

Diagnostics Check	OK	22:59	
Air Blank	0.000	23:00	
Control Test	0.078	23:00	
Air Blank	0.000	23:01	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	0.000	23:01	
Air Blank	0.000	23:01	
Control Test	0.078	23:02	
Air Blank	0.000	23:02	
Diagnostics Check	OK	23:02	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILAR,RODRIGO,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AGUILAR,RODRIGO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/06/2024 00:49:49

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.078	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.214	00:45	Volume Not Met
Breath Volume	0.554	00:45	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.077	00:49	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAYES,SEAN,K

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: RIVARD,MCKENZIE

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: BUI

ARREST TIME: 16:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/06/2024 17:54:10

Diagnostics Check	OK	17:48	
Air Blank	0.000	17:48	
Control Test	0.078	17:49	
Air Blank	0.000	17:49	
Subject Sample	# 1	17:52	RFI Detect
Breath Volume	0.128	17:52	
Air Blank	0.000	17:53	
Control Test	0.080	17:53	
Air Blank	0.000	17:54	
Diagnostics Check	OK	17:54	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAYES,SEAN,K

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: RIVARD,MCKENZIE

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: BUI

ARREST TIME: 16:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/06/2024 18:13:29

Diagnostics Check	OK	17:59	
Air Blank	0.000	18:00	
Control Test	0.079	18:00	
Air Blank	0.000	18:01	
Subject Sample	# 1 0.076	18:04	Volume Not Met
Breath Volume	0.613	18:04	
Air Blank	0.000	18:04	
Air Blank	0.000	18:06	
Subject Sample	# 2 0.075	18:07	
Breath Volume	1.707	18:07	
Air Blank	0.000	18:08	
Air Blank	0.000	18:10	
Subject Sample	# 3 0.072	18:11	
Breath Volume	1.218	18:11	
Air Blank	0.000	18:12	
Control Test	0.078	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLESTER,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: TORRES ESCOBAR

DATE/TIME OF TEST: 07/06/2024 23:35:44

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.192	23:31
Breath Volume	1.375	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.185	23:34
Breath Volume	1.328	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: TAURCHINI

DATE/TIME OF TEST: 07/08/2024 02:28:06

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Air Blank	0.000	02:19	
Subject Sample	# 2	0.214	02:21
Breath Volume	2.867	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 3	0.206	02:26
Breath Volume	1.472	02:26	
Air Blank	0.000	02:27	
Control Test	0.078	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PORTOCARRERO,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: PRIETO

DATE/TIME OF TEST: 07/09/2024 19:27:40

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.000	19:23
Breath Volume	1.921	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.000	19:26
Breath Volume	1.816	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GUTIERREZ HERNANDEZ

DATE/TIME OF TEST: 07/10/2024 04:23:18

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.177	04:18
Breath Volume	1.835	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.179	04:21
Breath Volume	1.554	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENTAS,N,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OOLEY,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG310302
DRY GAS EXPIRY DATE: 04/13/2025
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DEPALMENARY

DATE/TIME OF TEST: 07/11/2024 20:11:04

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.153	20:03
Breath Volume	1.371	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.176	20:06
Breath Volume	3.636	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:09
Subject Sample # 3	0.155	20:09
Breath Volume	1.972	20:09
Air Blank	0.000	20:10
Control Test	0.078	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:11

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: AGUILAR,RODRIGO,

BREATH TEST OPERATOR AGENCY: Vero Beach PD

ARRESTING OFFICER: AGUILAR,RODRIGO

ARRESTING OFFICER AGENCY: Vero Beach PD

LAST AGENCY INSPECTION: 06/10/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: DUI

ARREST TIME: 23:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GOTSCHALL

DATE/TIME OF TEST: 07/12/2024 03:19:51

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.195	03:15
Breath Volume	2.285	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.209	03:18
Breath Volume	2.273	03:18
Air Blank	0.000	03:18
Control Test	0.078	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/06/2024 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: THOMAS,MEGAM,A

BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)

ARRESTING OFFICER: THOMAS,MEGAN

ARRESTING OFFICER AGENCY: FFWCC (Ocala)

LAST AGENCY INSPECTION: 06/28/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: AG310302

DRY GAS EXPIRY DATE: 04/13/2025

VIOLATION CODE: BUI

ARREST TIME: 20:35

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/05/2024 21:28:44

Diagnostics Check	OK	21:22
Air Blank	0.000	21:22
Control Test	0.077	21:23
Air Blank	0.000	21:23
Subject Sample # 1	0.111	21:24
Breath Volume	3.191	21:24
Air Blank	0.000	21:24
Air Blank	0.000	21:26
Subject Sample # 2	0.111	21:27
Breath Volume	3.671	21:27
Air Blank	0.000	21:27
Control Test	0.078	21:28
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 07/22/2024 18:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAHR,DANIEL,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HAHR,DANIEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: BUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CAUDILLO

DATE/TIME OF TEST: 07/13/2024 17:44:27

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:42	
Control Test	0.077	17:42	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:43	Subject Test Refused
Breath Volume	0.000	17:43	
Air Blank	0.000	17:43	
Control Test	0.077	17:43	
Air Blank	0.000	17:44	
Diagnostics Check	OK	17:44	

Data Download Date/Time: 08/02/2024 09:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: BUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 07/13/2024 18:28:10

Diagnostics Check	OK	18:19	
Air Blank	0.000	18:19	
Control Test	0.077	18:20	
Air Blank	0.000	18:20	
Subject Sample	# 1	18:23	No Sample Provided
Breath Volume	0.000	18:23	
Air Blank	0.000	18:24	
Air Blank	0.000	18:26	
Subject Sample	# 2	18:26	Subject Test Refused
Breath Volume	0.000	18:26	
Air Blank	0.000	18:27	
Control Test	0.078	18:27	
Air Blank	0.000	18:28	
Diagnostics Check	OK	18:28	

Data Download Date/Time: 08/02/2024 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 07/01/2024 00:54:46

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.083	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.126	00:50
Breath Volume	2.750	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.125	00:53
Breath Volume	2.199	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SANTOS RIVERA

DATE/TIME OF TEST: 07/01/2024 01:10:06

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.198	01:05
Breath Volume	2.738	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.187	01:08
Breath Volume	1.437	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LAMPP

DATE/TIME OF TEST: 07/01/2024 01:19:04

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.174	01:14
Breath Volume	1.738	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.182	01:17
Breath Volume	1.632	01:17
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:19

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 07/01/2024 02:24:09

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:19
Breath Volume	2.113	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:22
Breath Volume	1.359	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: GUBBALA

DATE/TIME OF TEST: 07/01/2024 22:12:50

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.082	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.032	22:08
Breath Volume	1.917	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.033	22:11
Breath Volume	1.941	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 07/01/2024 23:15:03

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.082	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.201	23:10
Breath Volume	2.222	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.196	23:13
Breath Volume	1.769	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:15

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/02/2024 02:31:52

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.055	02:27
Breath Volume	1.898	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:29
Subject Sample # 2	0.054	02:30
Breath Volume	1.589	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PACHON PINEROS

DATE/TIME OF TEST: 07/03/2024 00:53:42

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.044	00:49
Breath Volume	2.878	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.044	00:52
Breath Volume	2.808	00:52
Air Blank	0.000	00:52
Control Test	0.081	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: O'BRIEN JR

DATE/TIME OF TEST: 07/03/2024 04:31:06

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.113	04:26
Breath Volume	1.953	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.116	04:29
Breath Volume	1.484	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: DUPONT

DATE/TIME OF TEST: 07/03/2024 06:10:29

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:01	
Control Test	0.082	06:02	
Air Blank	0.000	06:02	
Subject Sample	# 1 0.269	06:05	Volume Not Met
Breath Volume	0.152	06:05	
Air Blank	0.000	06:06	
Air Blank	0.000	06:08	
Subject Sample	# 2	06:09	Subject Test Refused
Breath Volume	0.000	06:09	
Air Blank	0.000	06:09	
Control Test	0.078	06:09	
Air Blank	0.000	06:10	
Diagnostics Check	OK	06:10	

Data Download Date/Time: 07/03/2024 14:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EARLYWINE,G
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/07/2024 05:18:18

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.082	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.063	05:13
Breath Volume	2.050	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.064	05:16
Breath Volume	1.812	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 08/06/2024 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 07/08/2024 01:29:22

Diagnostics Check	OK	01:18
Air Blank	0.000	01:19
Control Test	0.083	01:19
Air Blank	0.000	01:19
Subject Sample # 1	0.207	01:21
Breath Volume	1.660	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:23
Subject Sample # 2	0.241	01:24
Breath Volume	1.968	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 3	0.223	01:27
Breath Volume	2.195	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 08/06/2024 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DE LA CRUZ SANTIZ

DATE/TIME OF TEST: 07/10/2024 01:32:22

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.148	01:27
Breath Volume	1.585	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.149	01:30
Breath Volume	2.480	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/11/2024 01:47:25

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1 0.142	01:42	Volume Not Met
Breath Volume	0.128	01:42	
Air Blank	0.000	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 2	01:45	Interferent Detect
Breath Volume	1.566	01:45	
Air Blank	0.000	01:46	
Control Test	0.080	01:46	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: GALEANO MEJIA

DATE/TIME OF TEST: 07/12/2024 22:12:00

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.082	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.132	22:07
Breath Volume	3.226	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.134	22:10
Breath Volume	3.269	22:10
Air Blank	0.000	22:11
Control Test	0.082	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:12

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/13/2024 03:14:57

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.082	03:08	
Air Blank	0.000	03:09	
Subject Sample # 1	0.132	03:10	
Breath Volume	3.722	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:12	
Subject Sample # 2	0.128	03:13	
Breath Volume	2.742	03:13	
Air Blank	0.000	03:14	
Control Test	0.082	03:14	Interferent Detect
Air Blank	0.000	03:14	

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 07/13/2024 03:27:00

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.121	03:22
Breath Volume	3.050	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.129	03:25
Breath Volume	3.691	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:27

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS, P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ACHILLEA

DATE/TIME OF TEST: 07/13/2024 05:06:53

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.082	05:02
Breath Volume	2.328	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.083	05:05
Breath Volume	1.796	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/14/2024 00:48:54

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.082	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.095	00:44
Breath Volume	1.738	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.091	00:47
Breath Volume	1.523	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: OSTROWSKI JR

DATE/TIME OF TEST: 07/14/2024 02:56:20

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.082	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.117	02:51
Breath Volume	2.316	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:53
Subject Sample # 2	0.114	02:54
Breath Volume	2.285	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: RAHE

DATE/TIME OF TEST: 07/14/2024 06:05:06

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.082	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.167	06:00
Breath Volume	2.925	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:02
Subject Sample # 2	0.172	06:03
Breath Volume	2.265	06:03
Air Blank	0.000	06:04
Control Test	0.081	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: CORTES FLORES

DATE/TIME OF TEST: 07/16/2024 01:54:01

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:49
Breath Volume	2.765	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:52
Breath Volume	2.562	01:52
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 07/18/2024 03:37:25

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.083	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.083	03:32
Breath Volume	1.355	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.074	03:35
Breath Volume	1.179	03:35
Air Blank	0.000	03:36
Control Test	0.082	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: OBREGON CRUZ

DATE/TIME OF TEST: 07/18/2024 23:09:30

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.083	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.082	23:04
Breath Volume	1.730	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.085	23:07
Breath Volume	1.769	23:07
Air Blank	0.000	23:08
Control Test	0.082	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 07/20/2024 05:18:38

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.082	05:11
Air Blank	0.000	05:11
Subject Sample # 1	0.119	05:14
Breath Volume	1.445	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.117	05:17
Breath Volume	1.523	05:17
Air Blank	0.000	05:17
Control Test	0.081	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MACLIN

DATE/TIME OF TEST: 07/21/2024 03:41:01

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.146	03:36
Breath Volume	2.605	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.145	03:39
Breath Volume	1.960	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: AGUIAR AGUIAR

DATE/TIME OF TEST: 07/21/2024 05:07:16

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.082	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.221	05:01
Breath Volume	1.726	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.230	05:05
Breath Volume	2.066	05:05
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 07/26/2024 03:45:50

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: CHACAJVELAZQUEZ

DATE/TIME OF TEST: 07/26/2024 21:45:24

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.083	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.197	21:40
Breath Volume	1.410	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.197	21:43
Breath Volume	1.132	21:43
Air Blank	0.000	21:44
Control Test	0.080	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 07/31/2024 23:09:46

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.083	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.082	23:09	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 08/06/2024 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/06/2024 23:35:31

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.082	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.042	23:31
Breath Volume	2.535	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.043	23:33
Breath Volume	1.933	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 07/18/2024 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FHP TROOP K
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MERCEDES CACERES

DATE/TIME OF TEST: 07/21/2024 01:52:33

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank		01:51	Purge Fail
Air Blank		01:52	Purge Fail

Data Download Date/Time: 08/23/2024 01:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402909409
DRY GAS EXPIRY DATE: 11/29/2026
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/31/2024 08:50:10

Diagnostics Check	OK	08:44
Air Blank	0.000	08:44
Control Test	0.079	08:44
Air Blank	0.000	08:45
Subject Sample # 1	0.000	08:45
Breath Volume	2.742	08:45
Air Blank	0.000	08:46
Air Blank	0.000	08:48
Subject Sample # 2	0.000	08:48
Breath Volume	2.527	08:48
Air Blank	0.000	08:49
Control Test	0.079	08:49
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 08/28/2024 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/31/2024 14:47:41

Diagnostics Check	OK	14:44	
Air Blank	0.000	14:45	
Control Test	0.080	14:45	
Air Blank	0.000	14:46	
Subject Sample	# 1	14:46	Subject Test Refused
Breath Volume	0.000	14:46	
Air Blank	0.000	14:46	
Control Test	0.081	14:47	
Air Blank	0.000	14:47	
Diagnostics Check	OK	14:47	

Data Download Date/Time: 08/28/2024 13:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MULDERRIG,SEAN,F

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: VON PREYSING,R

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A3

DRY GAS EXPIRY DATE: 09/05/2025

VIOLATION CODE: DUI

ARREST TIME: 14:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/18/2024 16:12:54

Diagnostics Check	OK	16:06
Air Blank	0.000	16:06
Control Test	0.080	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.151	16:08
Breath Volume	2.257	16:08
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.155	16:11
Breath Volume	2.066	16:11
Air Blank	0.000	16:11
Control Test	0.080	16:12
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEAS,DALE,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: HERMANN,MATTHEW

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A3

DRY GAS EXPIRY DATE: 09/05/2025

VIOLATION CODE: DUI

ARREST TIME: 08:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: GOMEZ ANDRES

DATE/TIME OF TEST: 07/19/2024 09:04:59

Diagnostics Check	OK	08:58
Air Blank	0.000	08:59
Control Test	0.081	08:59
Air Blank	0.000	08:59
Subject Sample # 1	0.144	09:00
Breath Volume	3.281	09:00
Air Blank	0.000	09:01
Air Blank	0.000	09:02
Subject Sample # 2	0.143	09:03
Breath Volume	3.046	09:03
Air Blank	0.000	09:04
Control Test	0.078	09:04
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: COTHREN

DATE/TIME OF TEST: 07/28/2024 22:55:19

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.221	22:50
Breath Volume	2.843	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.230	22:53
Breath Volume	2.808	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 07/31/2024 00:16:18

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.126	00:11
Breath Volume	2.472	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.125	00:14
Breath Volume	1.496	00:14
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KINZER, ZACHERY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 07/31/2024 05:24:12

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.081	05:19	
Air Blank	0.000	05:19	
Subject Sample # 1	0.234	05:20	
Breath Volume	1.117	05:20	
Air Blank	0.000	05:21	
Air Blank		05:22	RFI Detect
Air Blank	0.000	05:23	
Control Test	0.081	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAFONTANT, SEDONA, S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KINZER, ZACHERY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 07/31/2024 05:33:04

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.234	05:28
Breath Volume	1.359	05:28
Air Blank	0.000	05:29
Air Blank	0.000	05:30
Subject Sample # 2	0.230	05:31
Breath Volume	1.351	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:33

Data Download Date/Time: 08/26/2024 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BURGER, WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: AJAO

DATE/TIME OF TEST: 07/10/2024 03:28:03

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.080	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: HASTY

DATE/TIME OF TEST: 07/12/2024 21:24:44

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.081	21:14	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:17	No Sample Provided
Breath Volume	0.000	21:17	
Air Blank	0.000	21:18	
Air Blank	0.000	21:20	
Subject Sample	# 2	21:23	No Sample Provided
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.081	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 07/14/2024 03:05:07

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1 0.148	02:55	
Breath Volume	1.355	02:55	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 3	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Control Test	0.079	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, EDGARDO,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: OROURKE, ROCKY
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: LOPEZ FRANCO

DATE/TIME OF TEST: 07/14/2024 04:51:13

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.156	04:46
Breath Volume	1.515	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.154	04:49
Breath Volume	1.507	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BURGER, WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 07/14/2024 14:44:35

Diagnostics Check	OK	14:41	
Air Blank	0.000	14:42	
Control Test	0.080	14:42	
Air Blank	0.000	14:42	
Subject Sample	# 1	14:43	Subject Test Refused
Breath Volume	0.000	14:43	
Air Blank	0.000	14:43	
Control Test	0.080	14:44	
Air Blank	0.000	14:44	
Diagnostics Check	OK	14:44	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGER, WILLIAM, T
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BURGER, WILLIAM
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 07/14/2024 15:07:13

Diagnostics Check	OK	15:00
Air Blank	0.000	15:01
Control Test	0.080	15:01
Air Blank	0.000	15:02
Subject Sample # 1	0.000	15:02
Breath Volume	4.566	15:02
Air Blank	0.000	15:03
Air Blank	0.000	15:05
Subject Sample # 2	0.000	15:05
Breath Volume	4.472	15:05
Air Blank	0.000	15:06
Control Test	0.079	15:06
Air Blank	0.000	15:07
Diagnostics Check	OK	15:07

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIECK, DAVID, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: BIECK, DAVID
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SCHLAU

DATE/TIME OF TEST: 07/29/2024 19:11:17

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.079	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.120	19:06
Breath Volume	1.179	19:06
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.127	19:09
Breath Volume	2.339	19:09
Air Blank	0.000	19:10
Control Test	0.081	19:10
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS, TYLER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: STARKS

DATE/TIME OF TEST: 07/20/2024 04:10:28

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.000	04:06
Breath Volume	2.597	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.000	04:09
Breath Volume	2.785	04:09
Air Blank	0.000	04:09
Control Test	0.077	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOWARD, AUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 07/22/2024 00:16:34

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.111	00:12
Breath Volume	3.046	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.114	00:15
Breath Volume	2.835	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, SETH, D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ORDOQUI, DAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 07/24/2024 03:31:23

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 1 0.200	03:24	Volume Not Met
Breath Volume	0.371	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.197	03:29	Volume Not Met
Breath Volume	0.128	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, SETH, D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ORDOQUI, DAN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 07/24/2024 03:38:51

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.079	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.359	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRICK,JASMIN,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOWARD,AUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BERTUBIN

DATE/TIME OF TEST: 07/25/2024 03:03:08

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.178	02:59	
Breath Volume	1.582	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.079	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRICK,JASMIN,H
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOWARD,AUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BERTUBIN

DATE/TIME OF TEST: 07/25/2024 03:35:54

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.172	03:31
Breath Volume	1.808	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.164	03:34
Breath Volume	1.839	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TYLER, J
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOWARD, AUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BOGAN

DATE/TIME OF TEST: 07/27/2024 20:56:13

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.029	20:51
Breath Volume	2.281	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.029	20:54
Breath Volume	2.332	20:54
Air Blank	0.000	20:55
Control Test	0.079	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,BENJAMIN,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HOWARD,AUSTIN
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 06/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 07/30/2024 22:16:31

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Subject Sample	# 1 0.112	22:05	Volume Not Met
Breath Volume	0.445	22:05	
Air Blank	0.000	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 2 0.139	22:09	
Breath Volume	1.539	22:09	
Air Blank	0.000	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 3 0.117	22:15	Volume Not Met
Breath Volume	0.203	22:15	
Air Blank	0.000	22:15	
Control Test	0.079	22:15	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 08/03/2024 03:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOONEY,KEVIN,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL,SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: ROMERO RAMIREZ

DATE/TIME OF TEST: 07/01/2024 00:38:24

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.256	00:33
Breath Volume	2.171	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.253	00:36
Breath Volume	2.441	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARROLL, SHELBY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BALLACK

DATE/TIME OF TEST: 07/02/2024 21:38:28

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEHRENS,JACOB,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BEHRENS,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: YANEZ ULLOA

DATE/TIME OF TEST: 07/05/2024 04:01:57

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.143	03:57
Breath Volume	1.937	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.146	04:00
Breath Volume	1.937	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNA,JAVIER,O
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REYNA,JAVIER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BUSTILLO MORALES

DATE/TIME OF TEST: 07/05/2024 04:58:02

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:46	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1 0.126	04:50	Volume Not Met
Breath Volume	0.171	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2 0.122	04:56	Volume Not Met
Breath Volume	0.117	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNA,JAVIER,O
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REYNA,JAVIER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BUSTILLO MORALES

DATE/TIME OF TEST: 07/05/2024 05:02:56

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.078	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.079	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SMALL, BRANNON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LA HAYE

DATE/TIME OF TEST: 07/06/2024 22:01:46

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.238	21:57
Breath Volume	2.894	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.235	22:00
Breath Volume	2.777	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 07/25/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JASON,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REGISTER,JASON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 07/07/2024 05:34:17

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.080	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Control Test	0.080	05:33	
Air Blank	0.000	05:34	
Diagnostics Check	OK	05:34	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGANK,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: THOMPSON,CLYDE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/07/2024 06:31:28

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:17	
Control Test	0.079	06:18	
Air Blank	0.000	06:18	
Subject Sample	# 1 0.095	06:21	Volume Not Met
Breath Volume	0.570	06:21	
Air Blank	0.000	06:22	
Air Blank	0.000	06:24	
Subject Sample	# 2 0.114	06:25	
Breath Volume	2.558	06:25	
Air Blank	0.000	06:26	
Air Blank	0.000	06:28	
Subject Sample	# 3 0.118	06:29	
Breath Volume	1.363	06:29	
Air Blank	0.000	06:30	
Control Test	0.079	06:30	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCRUGGS,JADEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCRUGGS,JADEN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WOODHOUSE

DATE/TIME OF TEST: 07/08/2024 01:37:10

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:32
Breath Volume	1.406	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:35
Breath Volume	1.792	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: OTHER
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 07/10/2024 05:27:50

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.088	05:23
Breath Volume	1.476	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:25
Subject Sample # 2	0.087	05:26
Breath Volume	1.628	05:26
Air Blank	0.000	05:26
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ESNEUL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/11/2024 21:08:34

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.080	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT, ADAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GIBSON, KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/12/2024 02:46:53

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.208	02:42
Breath Volume	3.855	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.205	02:45
Breath Volume	3.636	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ALNABULSI,MOHAMMAD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: GARCIA SANTOS

DATE/TIME OF TEST: 07/13/2024 22:45:52

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.190	22:41
Breath Volume	2.953	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.186	22:44
Breath Volume	2.214	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARTER,LARRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/14/2024 05:55:02

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.080	05:44	
Air Blank	0.000	05:44	
Subject Sample	# 1 0.166	05:47	Volume Not Met
Breath Volume	0.191	05:47	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 2 0.174	05:50	
Breath Volume	1.367	05:50	
Air Blank	0.000	05:51	
Air Blank	0.000	05:53	
Subject Sample	# 3	05:53	RFI Detect
Breath Volume	0.000	05:53	
Air Blank	0.000	05:54	
Control Test	0.078	05:54	
Air Blank	0.000	05:54	
Diagnostics Check	OK	05:55	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,LARRY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CARTER,LARRY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/14/2024 06:04:52

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.079	05:58
Air Blank	0.000	05:58
Subject Sample # 1	0.174	05:59
Breath Volume	1.562	05:59
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.182	06:03
Breath Volume	1.289	06:03
Air Blank	0.000	06:03
Control Test	0.080	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENNIS,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HENNIS,TIMOTHY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MUNOZ SANCHEZ

DATE/TIME OF TEST: 07/15/2024 02:15:40

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.197	02:11
Breath Volume	1.742	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.196	02:14
Breath Volume	1.296	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 07/15/2024 21:46:42

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:44	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:45	Subject Test Refused
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.079	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: BUI
ARREST TIME: 16:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 07/17/2024 17:52:36

Diagnostics Check	OK	17:45
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.121	17:47
Breath Volume	2.242	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.124	17:51
Breath Volume	2.742	17:51
Air Blank	0.000	17:51
Control Test	0.078	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 07/19/2024 00:05:17

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.025	00:00
Breath Volume	2.246	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.026	00:03
Breath Volume	2.187	00:03
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBO,GREGORY,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUBO,GREGORY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BOWLING

DATE/TIME OF TEST: 07/19/2024 04:16:20

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.220	04:11
Breath Volume	1.906	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.218	04:14
Breath Volume	1.539	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEWART, WILLIAM, R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SEWART, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: EWASHKO

DATE/TIME OF TEST: 07/20/2024 04:49:44

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CALLEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 07/22/2024 05:21:46

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.000	05:17
Breath Volume	3.128	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.000	05:20
Breath Volume	3.179	05:20
Air Blank	0.000	05:20
Control Test	0.080	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WAGNER,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: RUIZ MARTINEZ

DATE/TIME OF TEST: 07/25/2024 05:44:58

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:42	
Control Test	0.080	05:42	
Air Blank	0.000	05:43	
Subject Sample	# 1	05:43	Subject Test Refused
Breath Volume	0.000	05:43	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 07/25/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,ROBERT,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RASHEED,AHMAD
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: LEEBOY

DATE/TIME OF TEST: 07/25/2024 22:56:26

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.302	22:48
Breath Volume	3.457	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.270	22:51
Breath Volume	1.929	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 3	0.300	22:54
Breath Volume	1.199	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 09/04/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VANNIEUWENHOVEN,SCOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 07/27/2024 01:17:30

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.194	01:13
Breath Volume	2.425	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.189	01:15
Breath Volume	2.066	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 09/04/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: CUPPLES,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: ARRUE ESPARZA

DATE/TIME OF TEST: 07/27/2024 01:49:45

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.000	01:45
Breath Volume	2.769	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.000	01:48
Breath Volume	2.609	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 09/04/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ABEL,ANTHONY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/29/2024 19:32:19

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:23	
Control Test	0.080	19:24	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:27	No Sample Provided
Breath Volume	0.000	19:27	
Air Blank	0.000	19:28	
Air Blank	0.000	19:30	
Subject Sample	# 2	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:31	
Control Test	0.080	19:31	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 09/04/2024 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: YORK,MICHAEL
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 07/04/2024 00:02:56

Diagnostics Check	OK	23:54	
Air Blank	0.000	23:55	
Control Test	0.084	23:55	
Air Blank	0.000	23:55	
Subject Sample	# 2	23:58	No Sample Provided
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.084	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MINNIFIELD,SAM,M

BREATH TEST OPERATOR AGENCY: Crestview PD

ARRESTING OFFICER: MINNIFIELD,SAM

ARRESTING OFFICER AGENCY: Crestview PD

LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402590345

DRY GAS EXPIRY DATE: 11/04/2025

VIOLATION CODE: DUI

ARREST TIME: 02:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: W416110051010

DATE/TIME OF TEST: 07/19/2024 03:03:56

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.084	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.104	02:59
Breath Volume	4.093	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.113	03:02
Breath Volume	4.640	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRY,NICHOLAS
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: RAMIREZ SILVA

DATE/TIME OF TEST: 07/23/2024 04:11:23

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:03	
Control Test	0.083	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1 0.103	04:07	Volume Not Met
Breath Volume	0.097	04:07	
Air Blank	0.000	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 2	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Control Test	0.080	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINNIFIELD,SAM,M
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: MINNIFIELD,SAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: HENDRIX

DATE/TIME OF TEST: 07/24/2024 00:46:53

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.083	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 07/27/2024 05:22:20

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.159	05:17
Breath Volume	1.898	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.159	05:20
Breath Volume	2.101	05:20
Air Blank	0.000	05:21
Control Test	0.082	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/28/2024 05:38:19

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.083	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.191	05:33
Breath Volume	1.371	05:33
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.175	05:36
Breath Volume	1.683	05:36
Air Blank	0.000	05:37
Control Test	0.082	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 07/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/30/2024 01:26:23

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.083	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1 0.175	01:18	Volume Not Met
Breath Volume	0.566	01:18	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2 0.186	01:24	Volume Not Met
Breath Volume	0.015	01:24	
Air Blank	0.000	01:25	
Control Test	0.081	01:25	Interferent Detect
Air Blank		01:26	Purge Fail

Data Download Date/Time: 08/03/2024 20:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DIBIASE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LAMPP

DATE/TIME OF TEST: 07/01/2024 01:08:09

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	No .020 agreement
Breath Volume	2.539	01:03	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3	01:06	No .020 agreement
Breath Volume	1.210	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 07/03/2024 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOAGLAN,BRANDON,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 07/26/2024 16:12:38

Diagnostics Check	OK	16:10	
Air Blank	0.000	16:11	
Control Test	0.077	16:11	
Air Blank		16:12	RFI Detect
Air Blank	0.000	16:12	

Data Download Date/Time: 08/15/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOAGLAN,BRANDON,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 07/26/2024 16:21:13

Diagnostics Check	OK	16:14
Air Blank	0.000	16:15
Control Test	0.077	16:15
Air Blank	0.000	16:16
Subject Sample # 1	0.227	16:16
Breath Volume	3.843	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:18
Subject Sample # 2	0.230	16:19
Breath Volume	3.945	16:19
Air Blank	0.000	16:20
Control Test	0.075	16:20
Air Blank	0.000	16:21
Diagnostics Check	OK	16:21

Data Download Date/Time: 08/15/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: MUNION

DATE/TIME OF TEST: 07/10/2024 11:36:48

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:31
Subject Sample # 1	0.000	11:32
Breath Volume	2.640	11:32
Air Blank	0.000	11:33
Air Blank	0.000	11:35
Subject Sample # 2	0.000	11:35
Breath Volume	2.574	11:35
Air Blank	0.000	11:35
Control Test	0.080	11:36
Air Blank	0.000	11:36
Diagnostics Check	OK	11:36

Data Download Date/Time: 08/21/2024 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CHESTNUT

DATE/TIME OF TEST: 07/20/2024 03:39:02

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 08/19/2024 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 07/22/2024 03:19:33

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.080	03:17	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:18	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 08/19/2024 16:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: DULIN

DATE/TIME OF TEST: 07/29/2024 22:29:45

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.131	22:24
Breath Volume	1.812	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.130	22:27
Breath Volume	1.820	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 08/19/2024 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: REYES,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/04/2024 03:56:42

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.104	03:52
Breath Volume	1.726	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.107	03:55
Breath Volume	2.050	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 07/11/2024 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,R
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: CASTRO PORTILLO

DATE/TIME OF TEST: 07/04/2024 06:26:13

Diagnostics Check	OK	06:18
Air Blank	0.000	06:19
Control Test	0.079	06:19
Air Blank	0.000	06:19
Subject Sample # 1	0.154	06:20
Breath Volume	1.929	06:20
Air Blank	0.000	06:21
Air Blank	0.000	06:23
Subject Sample # 2	0.150	06:24
Breath Volume	1.929	06:24
Air Blank	0.000	06:25
Control Test	0.078	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 07/11/2024 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BATIST,J
ARRESTING OFFICER AGENCY: NORTH MIAMI PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SENAT

DATE/TIME OF TEST: 07/05/2024 01:13:05

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.079	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1 0.000	01:06	Volume Not Met
Breath Volume	0.015	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 2	01:11	Subject Test Refused
Breath Volume	0.089	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 07/11/2024 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BARBOZA MORALES

DATE/TIME OF TEST: 07/18/2024 04:42:11

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:41	
Control Test	0.081	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SALVANT,E
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: OROZCO BALAGUERA

DATE/TIME OF TEST: 07/20/2024 06:59:47

Diagnostics Check	OK	06:52
Air Blank	0.000	06:53
Control Test	0.079	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.090	06:55
Breath Volume	2.507	06:55
Air Blank	0.000	06:55
Air Blank	0.000	06:57
Subject Sample # 2	0.095	06:58
Breath Volume	1.269	06:58
Air Blank	0.000	06:58
Control Test	0.078	06:59
Air Blank	0.000	06:59
Diagnostics Check	OK	06:59

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: PENA BURGOS

DATE/TIME OF TEST: 07/20/2024 07:28:06

Diagnostics Check	OK	07:21
Air Blank	0.000	07:22
Control Test	0.078	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.142	07:23
Breath Volume	1.679	07:23
Air Blank	0.000	07:24
Air Blank	0.000	07:25
Subject Sample # 2	0.149	07:26
Breath Volume	2.035	07:26
Air Blank	0.000	07:27
Control Test	0.078	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:28

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,J
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BIGNOTTE PORTELA

DATE/TIME OF TEST: 07/21/2024 01:52:31

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.077	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.122	01:47
Breath Volume	1.800	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.116	01:50
Breath Volume	1.593	01:50
Air Blank	0.000	01:51
Control Test	0.077	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PEREZ DOMINGUEZ

DATE/TIME OF TEST: 07/21/2024 05:39:11

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.165	05:30
Breath Volume	1.554	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.188	05:34
Breath Volume	1.714	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 3	0.177	05:37
Breath Volume	1.738	05:37
Air Blank	0.000	05:38
Control Test	0.077	05:38
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 07/25/2024 05:32:14

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.190	05:27
Breath Volume	2.214	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:29
Subject Sample # 2	0.187	05:30
Breath Volume	1.886	05:30
Air Blank	0.000	05:31
Control Test	0.080	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: DUKES

DATE/TIME OF TEST: 07/27/2024 02:42:19

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1 0.137	02:33	Volume Not Met
Breath Volume	0.253	02:33	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2 0.136	02:36	
Breath Volume	1.441	02:36	
Air Blank	0.000	02:37	
Air Blank	0.000	02:39	
Subject Sample	# 3 0.128	02:40	
Breath Volume	1.453	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SALVANT,E
ARRESTING OFFICER AGENCY: NORTH BAY VILLAGE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: EDOUARD

DATE/TIME OF TEST: 07/28/2024 02:33:18

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 08/07/2024 04:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/05/2024 03:13:00

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.077	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.077	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 07/11/2024 04:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: REYES,JOEL
ARRESTING OFFICER AGENCY: SOUTH MIAMI PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LIZARDO

DATE/TIME OF TEST: 07/13/2024 00:35:58

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.256	00:29
Breath Volume	2.105	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.250	00:34
Breath Volume	2.089	00:34
Air Blank	0.000	00:35
Control Test	0.077	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 08/15/2024 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RODRIGUEZ BERITAN,LUIS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/18/2024 04:10:46

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.196	04:06
Breath Volume	3.222	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.184	04:09
Breath Volume	1.847	04:09
Air Blank	0.000	04:09
Control Test	0.076	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 08/15/2024 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,JORGE,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: DARDEN

DATE/TIME OF TEST: 07/25/2024 06:20:58

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.077	06:15
Air Blank	0.000	06:15
Subject Sample # 1	0.252	06:15
Breath Volume	1.460	06:15
Air Blank	0.000	06:16
Air Blank	0.000	06:18
Subject Sample # 2	0.241	06:19
Breath Volume	2.332	06:19
Air Blank	0.000	06:20
Control Test	0.077	06:20
Air Blank	0.000	06:20
Diagnostics Check	OK	06:20

Data Download Date/Time: 08/15/2024 00:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 07/04/2024 08:00:55

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.078	07:53
Air Blank	0.000	07:53
Subject Sample # 1	0.244	07:56
Breath Volume	1.921	07:56
Air Blank	0.000	07:57
Air Blank	0.000	07:58
Subject Sample # 2	0.235	07:59
Breath Volume	1.421	07:59
Air Blank	0.000	07:59
Control Test	0.078	08:00
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

Data Download Date/Time: 07/11/2024 03:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABIB,EMAD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LABIB,EMAD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: AUGUSTIN

DATE/TIME OF TEST: 07/18/2024 05:00:45

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.151	04:56
Breath Volume	1.906	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.146	04:59
Breath Volume	1.539	04:59
Air Blank	0.000	04:59
Control Test	0.078	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 08/07/2024 03:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERNANDEZ,ORLANDO
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TELLEZ MOLINA

DATE/TIME OF TEST: 07/26/2024 23:40:14

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.094	23:35
Breath Volume	1.960	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.096	23:38
Breath Volume	2.339	23:38
Air Blank	0.000	23:39
Control Test	0.078	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 08/07/2024 03:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CACARI

DATE/TIME OF TEST: 07/05/2024 02:01:20

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.206	01:56
Breath Volume	2.804	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:58
Subject Sample # 2	0.209	01:59
Breath Volume	3.695	01:59
Air Blank	0.000	02:00
Control Test	0.077	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/11/2024 03:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/13/2024 04:13:44

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.129	04:09
Breath Volume	2.156	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.123	04:12
Breath Volume	2.203	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 08/07/2024 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PUGA-CORRALES

DATE/TIME OF TEST: 07/18/2024 03:08:36

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.077	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.208	03:04
Breath Volume	1.867	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.217	03:06
Breath Volume	2.257	03:06
Air Blank	0.000	03:07
Control Test	0.078	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 08/07/2024 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/21/2024 02:32:44

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 08/07/2024 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HOYTE,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 07/26/2024 22:33:04

Diagnostics Check	OK	22:26
Air Blank	0.000	22:27
Control Test	0.078	22:27
Air Blank	0.000	22:28
Subject Sample # 1	0.101	22:28
Breath Volume	1.144	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.103	22:31
Breath Volume	1.546	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Diagnostics Check	OK	22:33

Data Download Date/Time: 08/07/2024 03:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADFORD, HARLEE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TIRADO, B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 07/03/2024 10:30:02

Diagnostics Check	OK	10:28	
Air Blank	0.000	10:29	
Control Test	0.081	10:29	
Air Blank		10:29	RFI Detect
Air Blank		10:30	RFI Detect

Data Download Date/Time: 07/30/2024 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADFORD, HARLEE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TIRADO, B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: CAULEY

DATE/TIME OF TEST: 07/03/2024 10:41:49

Diagnostics Check	OK	10:32
Air Blank	0.000	10:33
Control Test	0.080	10:33
Air Blank	0.000	10:34
Subject Sample # 1	0.000	10:36
Breath Volume	1.632	10:36
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample # 2	0.000	10:40
Breath Volume	1.621	10:40
Air Blank	0.000	10:40
Control Test	0.079	10:41
Air Blank	0.000	10:41
Diagnostics Check	OK	10:41

Data Download Date/Time: 07/30/2024 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: TERCERO-SILES

DATE/TIME OF TEST: 07/05/2024 04:55:26

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.163	04:50
Breath Volume	1.878	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.153	04:53
Breath Volume	1.160	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 07/30/2024 07:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA,S,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ZUNIGA,S
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: CARMENATE

DATE/TIME OF TEST: 07/06/2024 18:13:33

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.081	18:11	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.080	18:12	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 07/30/2024 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,C,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SIMONETTI,R
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: FARINELLA

DATE/TIME OF TEST: 07/06/2024 21:03:51

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.337	20:58
Breath Volume	1.273	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.349	21:01
Breath Volume	2.781	21:01
Air Blank	0.000	21:02
Control Test	0.078	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 07/30/2024 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: SCHLEY

DATE/TIME OF TEST: 07/10/2024 01:40:54

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:33	
Air Blank	0.000	01:33	
Subject Sample # 1	0.367	01:34	
Breath Volume	1.355	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:36	
Subject Sample # 2		01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.077	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JACOB,C
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MARTIN,JACOB
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 07/14/2024 05:07:44

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.131	05:02
Breath Volume	3.187	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.133	05:05
Breath Volume	3.136	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BROUGHTON

DATE/TIME OF TEST: 07/15/2024 04:35:45

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.080	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1 0.150	04:30	Volume Not Met
Breath Volume	0.070	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 2	04:33	Subject Test Refused
Breath Volume	0.265	04:33	
Air Blank	0.000	04:34	
Control Test	0.077	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: FOLEY, DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: ASUNCION MARTINEZ

DATE/TIME OF TEST: 07/15/2024 23:16:59

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.122	23:12
Breath Volume	1.972	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:14
Subject Sample # 2	0.120	23:15
Breath Volume	2.109	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDEMAR,KYLE,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VANDEMAR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MARTIN BALTAZAR

DATE/TIME OF TEST: 07/26/2024 03:10:19

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:04
Subject Sample # 1	0.212	03:05
Breath Volume	1.414	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.228	03:08
Breath Volume	3.660	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADFORD, HARLEE,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: PASSERI, MARLON
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: RAMOS SICA

DATE/TIME OF TEST: 07/28/2024 07:24:52

Diagnostics Check	OK	07:17
Air Blank	0.000	07:18
Control Test	0.081	07:18
Air Blank	0.000	07:18
Subject Sample # 1	0.092	07:20
Breath Volume	2.613	07:20
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.092	07:23
Breath Volume	2.812	07:23
Air Blank	0.000	07:23
Control Test	0.080	07:24
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 07/30/2024 07:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: LABOY

DATE/TIME OF TEST: 07/07/2024 00:13:18

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.077	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.199	00:08
Breath Volume	1.660	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.205	00:11
Breath Volume	1.773	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TEJADA,ALBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 07/08/2024 22:55:54

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.207	22:51
Breath Volume	3.261	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.201	22:54
Breath Volume	3.109	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: DOMINGO SALES

DATE/TIME OF TEST: 07/12/2024 21:38:20

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.110	21:33
Breath Volume	2.527	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.110	21:36
Breath Volume	2.726	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTIZ,JOSE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BARRIOS GONZALEZ

DATE/TIME OF TEST: 07/12/2024 22:55:31

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.158	22:51
Breath Volume	2.890	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.153	22:53
Breath Volume	2.347	22:53
Air Blank	0.000	22:54
Control Test	0.080	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 07/12/2024 23:43:49

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.205	23:39
Breath Volume	1.894	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.200	23:42
Breath Volume	1.886	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GLENNON

DATE/TIME OF TEST: 07/13/2024 01:00:17

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.137	00:55
Breath Volume	2.078	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.138	00:58
Breath Volume	2.058	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 07/26/2024 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,IVO,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: ESPINOSA,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: BUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 07/06/2024 16:24:33

Diagnostics Check	OK	16:19	
Air Blank	0.000	16:19	
Control Test	0.079	16:19	
Air Blank	0.000	16:20	
Subject Sample	# 1 0.161	16:22	
Breath Volume	1.164	16:22	
Air Blank		16:23	Purge Fail
Air Blank		16:24	Purge Fail

Data Download Date/Time: 07/23/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,NICHOLAS,
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: ESPINOSA,MIGUEL
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: BUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 07/06/2024 16:45:48

Diagnostics Check	OK	16:38
Air Blank	0.000	16:39
Control Test	0.077	16:39
Air Blank	0.000	16:39
Subject Sample # 1	0.159	16:40
Breath Volume	2.023	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.155	16:44
Breath Volume	1.945	16:44
Air Blank	0.000	16:44
Control Test	0.078	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 07/23/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,CORIN,L
BREATH TEST OPERATOR AGENCY: FFWCC (SOUTH A)
ARRESTING OFFICER: MARTINEZ,HENRY
ARRESTING OFFICER AGENCY: FFWCC (SOUTH A)
LAST AGENCY INSPECTION: 06/20/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: VENERIO

DATE/TIME OF TEST: 07/13/2024 13:33:05

Diagnostics Check	OK	13:28	
Air Blank	0.000	13:28	
Control Test	0.080	13:28	
Air Blank	0.000	13:29	
Subject Sample	# 1	13:31	Subject Test Refused
Breath Volume	0.000	13:31	
Air Blank	0.000	13:32	
Control Test	0.080	13:32	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:33	

Data Download Date/Time: 07/23/2024 13:56