

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES III,ISRAEL,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: REYES III,ISRAEL
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DE CAMPS

DATE/TIME OF TEST: 10/05/2024 03:44:20

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	RFI Detect
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 11/04/2024 21:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES III,ISRAEL,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: REYES III,ISRAEL
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: DE CAMPS

DATE/TIME OF TEST: 10/05/2024 03:49:19

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 11/04/2024 21:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ, DOUGLAS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BARQUERO, JOSE
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: REINA GOMEZ

DATE/TIME OF TEST: 10/01/2024 17:25:16

Diagnostics Check	OK	17:17
Air Blank	0.000	17:17
Control Test	0.078	17:18
Air Blank	0.000	17:18
Subject Sample # 1	0.314	17:19
Breath Volume	1.214	17:19
Air Blank	0.000	17:20
Air Blank	0.000	17:21
Subject Sample # 2	0.323	17:23
Breath Volume	1.398	17:23
Air Blank	0.000	17:24
Control Test	0.077	17:24
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 10/11/2024 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: SOLORZANO,WILLIAM
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 09/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SARMIENTO PRADO

DATE/TIME OF TEST: 10/02/2024 02:52:10

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.169	02:47
Breath Volume	1.367	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.175	02:50
Breath Volume	2.242	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 10/11/2024 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBAN,OMAR,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: MCCLURE,TIM
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 09/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MENDOZA CELIS

DATE/TIME OF TEST: 10/05/2024 05:30:45

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.135	05:24
Breath Volume	1.511	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.140	05:29
Breath Volume	1.507	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/11/2024 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTONIO,
BREATH TEST OPERATOR AGENCY: OTHER LOCAL
ARRESTING OFFICER: MENOCAL,MANUEL
ARRESTING OFFICER AGENCY: OTHER LOCAL
LAST AGENCY INSPECTION: 09/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PULVIRENTI

DATE/TIME OF TEST: 10/08/2024 23:34:30

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.078	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.139	23:29
Breath Volume	3.574	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.142	23:32
Breath Volume	4.292	23:32
Air Blank	0.000	23:33
Control Test	0.077	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/11/2024 02:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,O,
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: FERNANDEZ,O
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROMEROAVENDANO

DATE/TIME OF TEST: 10/14/2024 21:52:41

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1 0.144	21:45	Volume Not Met
Breath Volume	0.777	21:45	
Air Blank	0.000	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 2 0.158	21:48	
Breath Volume	2.070	21:48	
Air Blank	0.000	21:48	
Air Blank	0.000	21:50	
Subject Sample	# 3 0.158	21:51	
Breath Volume	2.140	21:51	
Air Blank	0.000	21:51	
Control Test	0.076	21:52	
Air Blank	0.000	21:52	
Diagnostics Check	OK	21:52	

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBAN,OMAR,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GALBAN,OMAR
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RODRIGUEZ BARRUETO

DATE/TIME OF TEST: 10/19/2024 01:18:07

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.195	01:13
Breath Volume	3.589	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.190	01:16
Breath Volume	3.281	01:16
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALBAN,OMAR,
BREATH TEST OPERATOR AGENCY: MEDLEY PD
ARRESTING OFFICER: GODINEZ,RANDY
ARRESTING OFFICER AGENCY: MEDLEY PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 10/19/2024 03:58:55

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.069	03:54
Breath Volume	1.363	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.066	03:57
Breath Volume	1.632	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOON,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAFAEL,ROBERT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MCDOUGALL

DATE/TIME OF TEST: 10/29/2024 00:35:36

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.164	00:34	Control Outside Tolerance
Air Blank	0.000	00:35	

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOON,SAMANTHA,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: RAFAEL,ROBERT
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MCDOUGALL

DATE/TIME OF TEST: 10/29/2024 00:55:49

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.162	00:50
Breath Volume	1.328	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.163	00:54
Breath Volume	2.222	00:54
Air Blank	0.000	00:54
Control Test	0.078	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,MOISES,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: BLANCO,MOISES
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/31/2024 23:37:02

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.078	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	Subject Test Refused
Breath Volume	0.000	23:35	
Air Blank	0.000	23:36	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:37	

Data Download Date/Time: 11/07/2024 02:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: SAXTON

DATE/TIME OF TEST: 10/04/2024 02:29:28

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.079	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.078	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LOPEZ RODRIGUEZ

DATE/TIME OF TEST: 10/04/2024 22:56:49

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.078	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.088	22:52
Breath Volume	3.257	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.078	22:55
Breath Volume	2.050	22:55
Air Blank	0.000	22:55
Control Test	0.078	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CARRACTER

DATE/TIME OF TEST: 10/05/2024 01:20:26

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.087	01:15
Breath Volume	1.503	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.082	01:18
Breath Volume	1.265	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: FORD-JONES

DATE/TIME OF TEST: 10/05/2024 01:58:36

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.078	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.078	01:58	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MUNOZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: KORNEGAY

DATE/TIME OF TEST: 10/05/2024 04:47:29

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.078	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.227	04:42
Breath Volume	2.046	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.230	04:45
Breath Volume	1.765	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARRAR,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:38

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 10/05/2024 14:14:26

Diagnostics Check	OK	13:59	
Air Blank	0.000	13:59	
Control Test	0.078	14:00	
Air Blank	0.000	14:00	
Subject Sample	# 1 0.157	14:03	
Breath Volume	1.152	14:03	
Air Blank	0.000	14:04	
Air Blank	0.000	14:06	
Subject Sample	# 2 0.146	14:09	Volume Not Met
Breath Volume	0.792	14:09	
Air Blank	0.000	14:09	
Air Blank	0.000	14:11	
Subject Sample	# 3 0.155	14:12	
Breath Volume	2.324	14:12	
Air Blank	0.000	14:13	
Control Test	0.078	14:13	
Air Blank	0.000	14:14	
Diagnostics Check	OK	14:14	

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/07/2024 02:56:38

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 10/24/2024 12:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: HOCKENBERY

DATE/TIME OF TEST: 10/09/2024 21:49:32

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.078	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.079	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CUBAS FLORES

DATE/TIME OF TEST: 10/10/2024 23:19:38

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.011	23:14
Breath Volume	1.500	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.009	23:18
Breath Volume	1.156	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE',M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/11/2024 01:41:15

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.078	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 10/11/2024 03:17:10

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.142	03:12
Breath Volume	1.996	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.135	03:15
Breath Volume	1.351	03:15
Air Blank	0.000	03:16
Control Test	0.078	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/11/2024 04:23:29

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	
Air Blank	0.000	04:21	
Subject Sample	# 1	04:22	Subject Test Refused
Breath Volume	0.000	04:22	
Air Blank	0.000	04:22	
Control Test	0.078	04:22	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: CORLEY

DATE/TIME OF TEST: 10/11/2024 22:29:20

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:25
Breath Volume	2.261	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:27
Breath Volume	2.187	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MEDINA CARDONA

DATE/TIME OF TEST: 10/12/2024 00:37:26

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.382	00:35	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 10/12/2024 01:47:48

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Control Test	0.078	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 10/12/2024 02:59:42

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.022	02:55
Breath Volume	2.304	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.020	02:58
Breath Volume	1.519	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ALBERTINI

DATE/TIME OF TEST: 10/12/2024 04:36:17

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.139	04:31
Breath Volume	2.445	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.136	04:34
Breath Volume	2.023	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MARADIAGA MORENO

DATE/TIME OF TEST: 10/13/2024 00:02:36

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.079	00:00	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WINGATE

DATE/TIME OF TEST: 10/13/2024 00:44:52

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.033	00:40
Breath Volume	1.996	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.033	00:43
Breath Volume	3.035	00:43
Air Blank	0.000	00:43
Control Test	0.079	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TANNER,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SHKURTAJ

DATE/TIME OF TEST: 10/13/2024 02:06:05

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.079	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.078	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 10/13/2024 03:12:10

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.079	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/13/2024 23:09:07

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.047	23:04
Breath Volume	1.113	23:04
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.045	23:07
Breath Volume	1.207	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 10/16/2024 04:25:57

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/16/2024 23:45:37

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:44	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BIBLE,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRISBANE

DATE/TIME OF TEST: 10/19/2024 03:42:02

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.130	03:37
Breath Volume	1.648	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.129	03:40
Breath Volume	1.339	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MEDINA CARDONA

DATE/TIME OF TEST: 10/19/2024 21:10:36

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.078	21:08	
Air Blank	0.000	21:09	
Subject Sample	# 1	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Control Test	0.079	21:10	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 10/24/2024 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MONTOYA HERNANDEZ

DATE/TIME OF TEST: 10/20/2024 02:50:00

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.079	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 10/24/2024 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CURRY,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: SILVA GOTAY

DATE/TIME OF TEST: 10/20/2024 07:53:56

Diagnostics Check	OK	07:42	
Air Blank	0.000	07:42	
Control Test	0.078	07:42	
Air Blank	0.000	07:43	
Subject Sample	# 1 0.112	07:46	Volume Not Met
Breath Volume	0.109	07:46	
Air Blank	0.000	07:47	
Air Blank	0.000	07:48	
Subject Sample	# 2 0.159	07:49	
Breath Volume	4.574	07:49	
Air Blank	0.000	07:50	
Air Blank	0.000	07:52	
Subject Sample	# 3 0.159	07:52	
Breath Volume	4.781	07:52	
Air Blank	0.000	07:53	
Control Test	0.077	07:53	
Air Blank	0.000	07:53	
Diagnostics Check	OK	07:53	

Data Download Date/Time: 10/24/2024 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 09/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: RAMIREZ PORTILLO

DATE/TIME OF TEST: 10/22/2024 23:07:45

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.145	23:03
Breath Volume	1.281	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.142	23:06
Breath Volume	2.097	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 10/24/2024 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GURR

DATE/TIME OF TEST: 10/25/2024 03:07:42

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.211	03:03
Breath Volume	3.789	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.201	03:06
Breath Volume	3.039	03:06
Air Blank	0.000	03:06
Control Test	0.078	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 10/25/2024 04:08:05

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.175	04:03
Breath Volume	1.527	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.163	04:06
Breath Volume	1.296	04:06
Air Blank	0.000	04:07
Control Test	0.077	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: BAEZ TORRES

DATE/TIME OF TEST: 10/26/2024 21:01:10

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.079	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	21:00	
Control Test	0.079	21:00	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:01	

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: QUINTERO GARCIA

DATE/TIME OF TEST: 10/27/2024 00:47:26

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.183	00:42
Breath Volume	1.968	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.188	00:45
Breath Volume	1.757	00:45
Air Blank	0.000	00:46
Control Test	0.078	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/27/2024 04:15:21

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.125	04:10
Breath Volume	1.832	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.121	04:13
Breath Volume	1.531	04:13
Air Blank	0.000	04:14
Control Test	0.079	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/29/2024 03:56:48

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.093	03:52
Breath Volume	2.277	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.097	03:55
Breath Volume	1.902	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: LAZO

DATE/TIME OF TEST: 10/30/2024 02:56:57

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.131	02:52
Breath Volume	1.390	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.129	02:55
Breath Volume	1.562	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CASTELLANOS MORALES

DATE/TIME OF TEST: 10/30/2024 04:23:31

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.194	04:15
Breath Volume	1.597	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.215	04:18
Breath Volume	2.707	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 3	0.205	04:21
Breath Volume	1.941	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MANINI,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 10/31/2024 19:38:36

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.079	19:36	
Air Blank	0.000	19:37	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.079	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 11/26/2024 08:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DEAL,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: COURT ORD
ARREST TIME: 07:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 10/28/2024 08:02:29

Diagnostics Check	OK	07:55
Air Blank	0.000	07:56
Control Test	0.079	07:56
Air Blank	0.000	07:57
Subject Sample # 1	0.000	07:57
Breath Volume	3.054	07:57
Air Blank	0.000	07:58
Air Blank	0.000	08:00
Subject Sample # 2	0.000	08:01
Breath Volume	2.902	08:01
Air Blank	0.000	08:01
Control Test	0.077	08:01
Air Blank	0.000	08:02
Diagnostics Check	OK	08:02

Data Download Date/Time: 12/02/2024 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, DAKOTA, J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: EDWARDS, DAKOTA
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: COURT ORD
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 10/29/2024 07:26:36

Diagnostics Check	OK	07:20
Air Blank	0.000	07:20
Control Test	0.079	07:21
Air Blank	0.000	07:21
Subject Sample # 1	0.000	07:22
Breath Volume	2.402	07:22
Air Blank	0.000	07:22
Air Blank	0.000	07:24
Subject Sample # 2	0.000	07:25
Breath Volume	2.746	07:25
Air Blank	0.000	07:25
Control Test	0.078	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 12/02/2024 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNGO, WILLIAM, F
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES, JASON
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: COURT ORD
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 10/30/2024 07:26:51

Diagnostics Check	OK	07:20
Air Blank	0.000	07:21
Control Test	0.079	07:21
Air Blank	0.000	07:21
Subject Sample # 1	0.000	07:22
Breath Volume	1.101	07:22
Air Blank	0.000	07:22
Air Blank	0.000	07:24
Subject Sample # 2	0.000	07:25
Breath Volume	1.812	07:25
Air Blank	0.000	07:25
Control Test	0.078	07:26
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 12/02/2024 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNGO, WILLIAM, F
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES, JASON
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 10/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: COURT ORD
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: LEDFORD

DATE/TIME OF TEST: 10/31/2024 07:16:59

Diagnostics Check	OK	07:10
Air Blank	0.000	07:11
Control Test	0.079	07:11
Air Blank	0.000	07:12
Subject Sample # 1	0.000	07:12
Breath Volume	2.390	07:12
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.000	07:15
Breath Volume	2.050	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 12/02/2024 19:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: HERNDON,DAVID
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 10/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/12/2024 20:05:43

Diagnostics Check	OK	19:59
Air Blank	0.000	20:00
Control Test	0.081	20:00
Air Blank	0.000	20:00
Subject Sample # 1	0.244	20:01
Breath Volume	3.312	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:03
Subject Sample # 2	0.244	20:04
Breath Volume	2.281	20:04
Air Blank	0.000	20:04
Control Test	0.080	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 12/04/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: JACKSON,DWIGHT
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 10/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 10/21/2024 20:39:10

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.081	20:37	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:38	
Control Test	0.081	20:38	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 12/04/2024 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: ELKINS,JUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/01/2024 00:48:10

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.081	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HOGAN,JERENY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: RENNIE

DATE/TIME OF TEST: 10/02/2024 02:23:45

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.160	02:19
Breath Volume	2.421	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.160	02:22
Breath Volume	2.710	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: DELANEY,TYLER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/05/2024 20:33:29

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.082	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:32	Subject Test Refused
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.082	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 10/11/2024 22:31:34

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.082	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.182	22:26
Breath Volume	1.910	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.163	22:30
Breath Volume	3.550	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HARRISON,MARAH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GORDY PUGLIESSI

DATE/TIME OF TEST: 10/12/2024 00:31:44

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.082	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:24	No Sample Provided
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Air Blank	0.000	00:27	
Subject Sample	# 2	00:30	Volume Not Met
Breath Volume	1.035	00:30	
Air Blank	0.000	00:30	
Control Test	0.084	00:31	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SCOTT,DYLAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/18/2024 20:52:10

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.083	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.057	20:47
Breath Volume	3.363	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.055	20:50
Breath Volume	2.167	20:50
Air Blank	0.000	20:51
Control Test	0.082	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JAXSON,W
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: HAMMOND,JAXSON
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: VOSSELLER

DATE/TIME OF TEST: 10/22/2024 22:14:41

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.082	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:10
Breath Volume	1.953	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:13
Breath Volume	1.804	22:13
Air Blank	0.000	22:13
Control Test	0.082	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 10/22/2024 22:46:58

Diagnostics Check	OK	22:39
Air Blank	0.000	22:39
Control Test	0.081	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.140	22:42
Breath Volume	1.472	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.139	22:45
Breath Volume	2.476	22:45
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 10/28/2024 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISON,ALLISON,K
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NEWELL,JESSICA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/24/2024 00:45:38

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.081	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.172	00:41
Breath Volume	1.195	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.161	00:44
Breath Volume	1.121	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/28/2024 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/25/2024 16:07:41

Diagnostics Check	OK	16:00
Air Blank	0.000	16:00
Control Test	0.081	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.239	16:03
Breath Volume	1.531	16:03
Air Blank	0.000	16:03
Air Blank	0.000	16:05
Subject Sample # 2	0.226	16:06
Breath Volume	2.382	16:06
Air Blank	0.000	16:06
Control Test	0.079	16:07
Air Blank	0.000	16:07
Diagnostics Check	OK	16:07

Data Download Date/Time: 10/28/2024 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:33

SUBJECT'S LAST NAME: RUTTER

DATE/TIME OF TEST: 10/26/2024 17:59:56

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.081	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.100	17:55
Breath Volume	1.535	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.102	17:58
Breath Volume	2.078	17:58
Air Blank	0.000	17:59
Control Test	0.082	17:59
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 10/28/2024 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: ALVEY

DATE/TIME OF TEST: 10/27/2024 20:23:03

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:12	
Control Test	0.081	20:12	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:16	No Sample Provided
Breath Volume	0.000	20:16	
Air Blank	0.000	20:16	
Air Blank	0.000	20:18	
Subject Sample	# 2	20:21	No Sample Provided
Breath Volume	0.000	20:21	
Air Blank	0.000	20:22	
Control Test	0.082	20:22	
Air Blank	0.000	20:22	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 10/28/2024 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: ALVEY

DATE/TIME OF TEST: 10/27/2024 20:52:44

Diagnostics Check	OK	20:45
Air Blank	0.000	20:45
Control Test	0.082	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.161	20:48
Breath Volume	1.175	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.169	20:51
Breath Volume	1.425	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 10/28/2024 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSME,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: COSME,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: DRALLE

DATE/TIME OF TEST: 10/31/2024 23:17:15

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.198	23:12
Breath Volume	1.750	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.212	23:15
Breath Volume	1.863	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 11/22/2024 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,COTY,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RUTHERFORD,SYDNEY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/06/2024 21:24:22

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.080	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:19
Breath Volume	1.285	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.000	21:23
Breath Volume	1.242	21:23
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 10/07/2024 12:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,COTY,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RUTHERFORD,SYDNEY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/09/2024 20:53:53

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.080	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1	20:52	Subject Test Refused
Breath Volume	0.000	20:52	
Air Blank	0.000	20:52	
Control Test	0.081	20:53	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 11/25/2024 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON, THOMAS, T
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: RUTHERFORD, SYDNEY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MOELLER

DATE/TIME OF TEST: 10/28/2024 22:28:33

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.161	22:24
Breath Volume	1.949	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.160	22:26
Breath Volume	5.070	22:26
Air Blank	0.000	22:27
Control Test	0.079	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 11/25/2024 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DUGAS

DATE/TIME OF TEST: 10/04/2024 22:24:14

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.080	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUCHNIR,EUGENE,V
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KOUCHNIR,EUGENE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: COOLEY

DATE/TIME OF TEST: 10/05/2024 13:35:03

Diagnostics Check	OK	13:28
Air Blank	0.000	13:29
Control Test	0.079	13:29
Air Blank	0.000	13:30
Subject Sample # 1	0.017	13:30
Breath Volume	3.816	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:33
Subject Sample # 2	0.016	13:33
Breath Volume	3.066	13:33
Air Blank	0.000	13:34
Control Test	0.080	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:35

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 10/06/2024 01:12:41

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.079	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.080	01:12	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITONIAK,ROBERT,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PITONIAK,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: WANN

DATE/TIME OF TEST: 10/13/2024 20:33:22

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.163	20:28
Breath Volume	1.257	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.163	20:31
Breath Volume	1.343	20:31
Air Blank	0.000	20:32
Control Test	0.078	20:32
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COKONOUGH,ROBERT,F
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 10/16/2024 00:51:09

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Subject Sample # 1	0.000	00:46
Breath Volume	3.148	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:49
Breath Volume	3.132	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTRY, WILLIS, E
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF, HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402697419
DRY GAS EXPIRY DATE: 03/20/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: LOBMEYER

DATE/TIME OF TEST: 10/19/2024 23:24:29

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.227	23:19
Breath Volume	1.351	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.239	23:22
Breath Volume	1.585	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 10/25/2024 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: LANGE

DATE/TIME OF TEST: 10/04/2024 01:29:08

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.088	01:24
Breath Volume	2.867	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.091	01:27
Breath Volume	4.167	01:27
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 10/22/2024 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 06:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 10/04/2024 07:13:00

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.082	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.000	07:08
Breath Volume	2.398	07:08
Air Blank	0.000	07:08
Air Blank	0.000	07:10
Subject Sample # 2	0.000	07:11
Breath Volume	2.109	07:11
Air Blank	0.000	07:12
Control Test	0.081	07:12
Air Blank	0.000	07:12
Diagnostics Check	OK	07:13

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE, ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: WRIGHT, MARK
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/04/2024 23:25:59

Diagnostics Check	OK	23:23	
Air Blank	0.000	23:24	
Control Test	0.082	23:24	
Air Blank	0.000	23:24	
Subject Sample	# 1 0.156	23:25	
Breath Volume	1.707	23:25	
Air Blank		23:25	RFI Detect
Air Blank		23:25	RFI Detect
Air Blank		23:25	RFI Detect

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE, ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: WRIGHT, MARK
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 10/04/2024 23:34:04

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.149	23:29
Breath Volume	2.140	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.134	23:32
Breath Volume	1.484	23:32
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 10/06/2024 11:43:53

Diagnostics Check	OK	11:40	
Air Blank	0.000	11:40	
Control Test	0.081	11:41	
Air Blank	0.000	11:41	
Subject Sample # 1	0.000	11:42	
Breath Volume	2.875	11:42	
Air Blank		11:42	RFI Detect
Air Blank	0.000	11:42	
Control Test	0.081	11:43	
Air Blank	0.000	11:43	
Diagnostics Check	OK	11:43	

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 10/06/2024 11:53:13

Diagnostics Check	OK	11:47
Air Blank	0.000	11:47
Control Test	0.081	11:47
Air Blank	0.000	11:48
Subject Sample # 1	0.000	11:48
Breath Volume	2.929	11:48
Air Blank	0.000	11:49
Air Blank	0.000	11:51
Subject Sample # 2	0.000	11:51
Breath Volume	2.820	11:51
Air Blank	0.000	11:52
Control Test	0.082	11:52
Air Blank	0.000	11:53
Diagnostics Check	OK	11:53

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 10/07/2024 22:28:56

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.082	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.119	22:24
Breath Volume	2.000	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.118	22:27
Breath Volume	1.433	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/09/2024 04:18:02

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.233	04:13
Breath Volume	3.625	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:15
Subject Sample # 2	0.227	04:16
Breath Volume	3.222	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: HAMNER

DATE/TIME OF TEST: 10/10/2024 22:09:13

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.081	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.169	22:04
Breath Volume	1.660	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:06
Subject Sample # 2	0.178	22:07
Breath Volume	1.750	22:07
Air Blank	0.000	22:08
Control Test	0.081	22:08
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 10/11/2024 22:58:02

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.010	22:53
Breath Volume	4.121	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:56
Subject Sample # 2	0.009	22:56
Breath Volume	4.265	22:56
Air Blank	0.000	22:57
Control Test	0.082	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:58

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: GARCIA TREJO

DATE/TIME OF TEST: 10/12/2024 19:31:12

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.310	19:26
Breath Volume	3.281	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.307	19:29
Breath Volume	3.238	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/13/2024 02:33:32

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.082	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:33	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 10/14/2024 01:20:26

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/22/2024 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURTAGH,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MURTAGH,ROBERT
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: MORALES ESCALANTE

DATE/TIME OF TEST: 10/14/2024 06:00:07

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.081	05:53
Air Blank	0.000	05:54
Subject Sample # 1	0.239	05:55
Breath Volume	2.445	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:57
Subject Sample # 2	0.222	05:58
Breath Volume	1.429	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 10/22/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 09/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: FERRAZ CARMONA

DATE/TIME OF TEST: 10/16/2024 21:53:43

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.082	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.081	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 10/22/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 10/23/2024 02:25:16

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:21
Breath Volume	1.496	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:23
Breath Volume	1.515	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: JEUNE

DATE/TIME OF TEST: 10/26/2024 04:10:48

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:08	
Control Test	0.082	04:08	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:09	
Control Test	0.082	04:10	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RAIMONDA

DATE/TIME OF TEST: 10/26/2024 22:14:43

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.082	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.177	22:10
Breath Volume	2.093	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.167	22:13
Breath Volume	2.804	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: RENO,KARSON
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: BUSSIERE

DATE/TIME OF TEST: 10/26/2024 23:15:44

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.082	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.000	23:11
Breath Volume	2.843	23:11
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.000	23:14
Breath Volume	2.699	23:14
Air Blank	0.000	23:14
Control Test	0.082	23:15
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PONCE DE LEON

DATE/TIME OF TEST: 10/27/2024 03:13:50

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.138	03:09
Breath Volume	2.718	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.142	03:12
Breath Volume	3.035	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCHAMPS,KASEN,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISSI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: DALMATA

DATE/TIME OF TEST: 10/27/2024 20:23:32

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.082	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.285	20:18
Breath Volume	3.152	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 2	0.285	20:21
Breath Volume	3.304	20:21
Air Blank	0.000	20:22
Control Test	0.082	20:22
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/27/2024 20:56:34

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.255	20:51
Breath Volume	3.035	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.255	20:54
Breath Volume	2.914	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: RODRIGUEZ,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 10/31/2024 20:31:51

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:29	
Control Test	0.082	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.082	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 11/21/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: APPELBAUM,KEITH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BASTIEN

DATE/TIME OF TEST: 10/04/2024 02:32:46

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.142	02:28
Breath Volume	2.808	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.149	02:31
Breath Volume	3.074	02:31
Air Blank	0.000	02:31
Control Test	0.076	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THERIAULT,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: BRUTON

DATE/TIME OF TEST: 10/06/2024 05:21:19

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.077	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.204	05:16
Breath Volume	2.238	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.201	05:19
Breath Volume	2.824	05:19
Air Blank	0.000	05:20
Control Test	0.076	05:20
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITES,DEJAS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WAITES,DEJAS
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HANCZARUK

DATE/TIME OF TEST: 10/10/2024 23:54:59

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.077	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.138	23:50
Breath Volume	2.035	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.138	23:53
Breath Volume	2.421	23:53
Air Blank	0.000	23:54
Control Test	0.076	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILBERT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GILBERT,MATTHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: TAIT

DATE/TIME OF TEST: 10/10/2024 00:55:27

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.127	00:51
Breath Volume	2.539	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.128	00:53
Breath Volume	2.460	00:53
Air Blank	0.000	00:54
Control Test	0.077	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: 5L29,TOMCZYK
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GRACI

DATE/TIME OF TEST: 10/11/2024 18:27:19

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.077	18:17	
Air Blank	0.000	18:17	
Subject Sample	# 1 0.297	18:21	Volume Not Met
Breath Volume	0.015	18:21	
Air Blank	0.000	18:21	
Air Blank	0.000	18:23	
Subject Sample	# 2 0.320	18:24	
Breath Volume	1.734	18:24	
Air Blank	0.000	18:24	
Air Blank	0.000	18:26	
Subject Sample	# 3	18:26	RFI Detect
Breath Volume	0.000	18:26	
Air Blank		18:26	RFI Detect
Air Blank	0.000	18:27	

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GRACI

DATE/TIME OF TEST: 10/11/2024 18:41:29

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:29	
Control Test	0.076	18:29	
Air Blank	0.000	18:30	
Subject Sample # 1	0.311	18:31	
Breath Volume	1.421	18:31	
Air Blank	0.000	18:31	
Air Blank	0.000	18:33	
Subject Sample # 2	0.272	18:36	Volume Not Met
Breath Volume	0.269	18:36	
Air Blank	0.000	18:37	
Air Blank	0.000	18:39	
Subject Sample # 3	0.265	18:39	
Breath Volume	1.562	18:39	
Air Blank	0.000	18:40	
Control Test	0.074	18:40	Control Outside Tolerance
Air Blank	0.000	18:41	

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMCZYK,STENLY,
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: LAWTON,BRENNEN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: RAYMONVIL

DATE/TIME OF TEST: 10/11/2024 21:22:46

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.076	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.267	21:10	Volume Not Met
Breath Volume	0.675	21:10	
Air Blank	0.000	21:11	
Air Blank	0.000	21:13	
Subject Sample	# 2 0.285	21:15	
Breath Volume	1.109	21:15	
Air Blank	0.000	21:16	
Air Blank	0.000	21:18	
Subject Sample	# 3	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Control Test	0.074	21:22	Control Outside Tolerance
Air Blank	0.000	21:22	

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TEMINO,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CALIXTE

DATE/TIME OF TEST: 10/13/2024 03:22:50

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.077	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.144	03:18
Breath Volume	1.621	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.161	03:21
Breath Volume	2.785	03:21
Air Blank	0.000	03:21
Control Test	0.076	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILBERT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TOMCZYK,STENLY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LAVENDER

DATE/TIME OF TEST: 10/15/2024 23:13:02

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.078	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.055	23:08
Breath Volume	3.023	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.054	23:11
Breath Volume	3.230	23:11
Air Blank	0.000	23:12
Control Test	0.077	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:13

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILBERT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: GILBERT,MATTHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 09/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: PLASENCIA LAMAR

DATE/TIME OF TEST: 10/18/2024 23:45:36

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.078	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.054	23:41
Breath Volume	3.234	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.052	23:44
Breath Volume	3.066	23:44
Air Blank	0.000	23:44
Control Test	0.078	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 10/23/2024 13:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ, SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SPRADLING

DATE/TIME OF TEST: 10/01/2024 21:53:27

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.083	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.082	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 10/02/2024 00:50:15

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:47	
Control Test	0.083	00:48	
Air Blank	0.000	00:48	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.081	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ, SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HAYS

DATE/TIME OF TEST: 10/02/2024 01:30:10

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.082	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.194	01:25
Breath Volume	1.632	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.203	01:28
Breath Volume	2.296	01:28
Air Blank	0.000	01:29
Control Test	0.082	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STAFF,P,S
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: JENTER

DATE/TIME OF TEST: 10/02/2024 15:55:16

Diagnostics Check	OK	15:48
Air Blank	0.000	15:49
Control Test	0.083	15:49
Air Blank	0.000	15:49
Subject Sample # 1	0.152	15:50
Breath Volume	1.566	15:50
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.148	15:53
Breath Volume	1.152	15:53
Air Blank	0.000	15:54
Control Test	0.082	15:54
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUREN,T,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCOTT,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 10/03/2024 00:27:11

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.058	00:22
Breath Volume	2.042	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.062	00:25
Breath Volume	1.921	00:25
Air Blank	0.000	00:26
Control Test	0.082	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MULCAHY

DATE/TIME OF TEST: 10/03/2024 02:43:17

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:40	
Control Test	0.082	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUREN,T,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLISTER,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GEROME

DATE/TIME OF TEST: 10/03/2024 03:39:10

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.000	03:34
Breath Volume	2.882	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.000	03:37
Breath Volume	2.621	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,BRITTANY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RIVERA,A
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 10/04/2024 17:17:13

Diagnostics Check	OK	17:03	
Air Blank	0.000	17:03	
Control Test	0.083	17:04	
Air Blank	0.000	17:04	
Subject Sample	# 1 0.161	17:06	
Breath Volume	3.492	17:06	
Air Blank	0.000	17:07	
Air Blank	0.000	17:09	
Subject Sample	# 2 0.110	17:12	Volume Not Met
Breath Volume	0.269	17:12	
Air Blank	0.000	17:13	
Air Blank	0.000	17:15	
Subject Sample	# 3	17:15	Subject Test Refused
Breath Volume	0.000	17:15	
Air Blank	0.000	17:16	
Control Test	0.080	17:16	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARRIENTOS, S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MCCULLERS

DATE/TIME OF TEST: 10/04/2024 21:55:54

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.083	21:53	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.082	21:55	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MONTGOMERY,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WYNN

DATE/TIME OF TEST: 10/04/2024 23:33:30

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.083	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:32	
Control Test	0.082	23:32	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:33	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: QUINT, K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BARROWS

DATE/TIME OF TEST: 10/04/2024 23:53:56

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.083	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.168	23:49
Breath Volume	2.222	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:51
Subject Sample # 2	0.165	23:52
Breath Volume	1.890	23:52
Air Blank	0.000	23:53
Control Test	0.082	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,OSMAY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: VAZSQUEZ

DATE/TIME OF TEST: 10/05/2024 00:18:40

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.082	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:17	No Sample Provided
Breath Volume	0.000	00:17	
Air Blank	0.000	00:17	
Control Test	0.083	00:18	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,OSMAY,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PEREZ,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/05/2024 00:27:43

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.083	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.153	00:23
Breath Volume	2.398	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.152	00:26
Breath Volume	2.648	00:26
Air Blank	0.000	00:26
Control Test	0.082	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 10/19/2024 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,SEBASTIEN,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GONZALEZ,SEBASTIEN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 10/05/2024 03:00:33

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.176	02:55
Breath Volume	3.882	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.177	02:58
Breath Volume	3.753	02:58
Air Blank	0.000	02:59
Control Test	0.079	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Sebastian PD
ARRESTING OFFICER: DIDIER, M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ODRODO

DATE/TIME OF TEST: 10/05/2024 04:27:39

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.083	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.000	04:23
Breath Volume	1.789	04:23
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.000	04:26
Breath Volume	1.320	04:26
Air Blank	0.000	04:26
Control Test	0.083	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,SEBASTIEN,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: GONZALEZ,SEBASTIEN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: TUSSING

DATE/TIME OF TEST: 10/05/2024 23:34:03

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.083	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.082	23:33	
Air Blank	0.000	23:33	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ, SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: GAGLIARDI

DATE/TIME OF TEST: 10/06/2024 02:27:46

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.083	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.102	02:22
Breath Volume	1.746	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.120	02:26
Breath Volume	2.445	02:26
Air Blank	0.000	02:26
Control Test	0.081	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCOTT,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/06/2024 19:53:44

Diagnostics Check	OK	19:50	
Air Blank	0.000	19:51	
Control Test	0.082	19:51	
Air Blank	0.000	19:51	
Subject Sample	# 1	19:52	Subject Test Refused
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Control Test	0.082	19:53	
Air Blank	0.000	19:53	
Diagnostics Check	OK	19:53	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ, M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ROSANO

DATE/TIME OF TEST: 10/07/2024 02:57:31

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.209	02:51
Breath Volume	1.566	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.211	02:55
Breath Volume	1.394	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCROGGIN,JOHN,C
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: SCROGGIN,JOHN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/11/2024 16:01:30

Diagnostics Check	OK	15:54
Air Blank	0.000	15:54
Control Test	0.083	15:55
Air Blank	0.000	15:55
Subject Sample # 1	0.167	15:56
Breath Volume	1.738	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.165	15:59
Breath Volume	1.687	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BOLES,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:02

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/11/2024 20:09:14

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.082	20:07	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:08	
Control Test	0.081	20:08	
Air Blank	0.000	20:09	
Diagnostics Check	OK	20:09	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: QUINT,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/12/2024 22:46:45

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.083	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Control Test	0.000	22:46	RFI Detect
Air Blank	0.000	22:46	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTINEZ,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 10/12/2024 23:57:50

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.082	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	4.175	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:56
Breath Volume	4.265	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCOTT, C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MESSER

DATE/TIME OF TEST: 10/14/2024 21:27:11

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.082	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.215	21:19
Breath Volume	1.238	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:21
Subject Sample # 2	0.244	21:22
Breath Volume	2.253	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 3	0.220	21:25
Breath Volume	1.800	21:25
Air Blank	0.000	21:26
Control Test	0.080	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARRIENTOS,S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LUCCHINI

DATE/TIME OF TEST: 10/14/2024 23:11:43

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.082	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.081	23:11	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 10/15/2024 22:40:57

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.083	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:40	
Control Test	0.083	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER, K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: POSTELL

DATE/TIME OF TEST: 10/18/2024 20:04:50

Diagnostics Check	OK	19:58
Air Blank	0.000	19:58
Control Test	0.082	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.000	20:00
Breath Volume	1.902	20:00
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.000	20:03
Breath Volume	2.171	20:03
Air Blank	0.000	20:03
Control Test	0.083	20:04
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,K,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BANSIL,H
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 09/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/19/2024 09:51:43

Diagnostics Check	OK	09:45
Air Blank	0.000	09:46
Control Test	0.083	09:46
Air Blank	0.000	09:47
Subject Sample # 1	0.000	09:47
Breath Volume	1.847	09:47
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.000	09:50
Breath Volume	1.863	09:50
Air Blank	0.000	09:50
Control Test	0.083	09:51
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 10/19/2024 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDY,KEVIN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTINEZ,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/20/2024 06:29:03

Diagnostics Check	OK	06:21
Air Blank	0.000	06:22
Control Test	0.082	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.160	06:23
Breath Volume	2.281	06:23
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.152	06:27
Breath Volume	1.660	06:27
Air Blank	0.000	06:28
Control Test	0.082	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:29

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUEVAS,J,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 10/23/2024 22:34:02

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:31	
Control Test	0.083	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:32	Subject Test Refused
Breath Volume	0.000	22:32	
Air Blank	0.000	22:33	
Control Test	0.082	22:33	
Air Blank	0.000	22:33	
Diagnostics Check	OK	22:34	

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ, SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 10/24/2024 03:11:00

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.082	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank		03:10	RFI Detect
Air Blank	0.000	03:11	

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DEGETTE,J
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 10/25/2024 20:50:31

Diagnostics Check	OK	20:47	
Air Blank	0.000	20:48	
Control Test	0.083	20:48	
Air Blank	0.000	20:48	
Subject Sample	# 1	20:49	Subject Test Refused
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Control Test	0.082	20:49	
Air Blank	0.000	20:50	
Diagnostics Check	OK	20:50	

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MARTINEZ,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/26/2024 00:21:57

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Control Test	0.082	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SCOTT,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: EIX

DATE/TIME OF TEST: 10/26/2024 01:01:04

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.082	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1	00:59	Subject Test Refused
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Control Test	0.082	01:00	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORDAZ, SERGIO,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: AGUILAR, RODRIGO
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 10/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 10/28/2024 00:28:05

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.082	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.122	00:23
Breath Volume	2.093	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.128	00:26
Breath Volume	1.273	00:26
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:28

Data Download Date/Time: 11/17/2024 15:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZILINSKIY

DATE/TIME OF TEST: 10/04/2024 01:40:04

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.115	01:32	Volume Not Met
Breath Volume	0.750	01:32	
Air Blank	0.000	01:33	
Air Blank	0.000	01:35	
Subject Sample	# 2 0.106	01:38	Volume Not Met
Breath Volume	0.484	01:38	
Air Blank	0.000	01:39	
Control Test	0.077	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZILINSKIY

DATE/TIME OF TEST: 10/04/2024 01:57:08

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1 0.105	01:45	Slope Not Level
Breath Volume	1.378	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.106	01:51	
Breath Volume	1.152	01:51	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 3 0.096	01:55	
Breath Volume	1.164	01:55	
Air Blank	0.000	01:56	
Control Test	0.075	01:56	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 10/06/2024 01:14:40

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.082	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.196	01:10
Breath Volume	2.121	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.193	01:13
Breath Volume	2.203	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 10/06/2024 04:13:48

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.206	04:08
Breath Volume	1.820	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.193	04:12
Breath Volume	1.328	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FRAME

DATE/TIME OF TEST: 10/11/2024 02:23:07

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:19	
Control Test	0.081	02:19	
Air Blank	0.000	02:20	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.082	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/12/2024 01:37:22

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.200	01:29
Breath Volume	1.605	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.168	01:32
Breath Volume	1.527	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 3	0.165	01:35
Breath Volume	2.050	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 10/14/2024 23:41:50

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:30	
Control Test	0.082	23:31	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:34	No Sample Provided
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Control Test	0.083	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 10/14/2024 23:50:56

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.141	23:45
Breath Volume	1.656	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.143	23:49
Breath Volume	1.695	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JACOBSON

DATE/TIME OF TEST: 10/18/2024 01:10:21

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.084	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.140	01:05
Breath Volume	1.562	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.150	01:08
Breath Volume	1.835	01:08
Air Blank	0.000	01:09
Control Test	0.083	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: PEREIRA-ROCHA

DATE/TIME OF TEST: 10/18/2024 23:55:13

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.195	23:50
Breath Volume	1.152	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.200	23:53
Breath Volume	1.468	23:53
Air Blank	0.000	23:54
Control Test	0.082	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 10/19/2024 06:01:03

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.083	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.251	05:56
Breath Volume	1.613	05:56
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.253	05:59
Breath Volume	1.847	05:59
Air Blank	0.000	06:00
Control Test	0.082	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:01

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE',TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 10/20/2024 00:31:50

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.082	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.207	00:27
Breath Volume	2.566	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.208	00:30
Breath Volume	2.550	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,RAYCHEL,E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: DOUGAN,KASEY
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ ALMANZA

DATE/TIME OF TEST: 10/20/2024 04:19:43

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.082	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.208	04:15
Breath Volume	2.363	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.224	04:18
Breath Volume	2.363	04:18
Air Blank	0.000	04:18
Control Test	0.083	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 11/01/2024 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,RAYCHEL,E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: SIPE,RAYCHEL
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LORIO GLASGOW

DATE/TIME OF TEST: 10/20/2024 04:33:28

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.082	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.143	04:29
Breath Volume	1.843	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.148	04:31
Breath Volume	2.039	04:31
Air Blank	0.000	04:32
Control Test	0.080	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MALAK,TIM
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MLADENOVA

DATE/TIME OF TEST: 10/21/2024 02:46:52

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.122	02:42
Breath Volume	2.593	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.110	02:45
Breath Volume	1.644	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,RAYCHEL,E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: PEREZ,ORLANDO
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/21/2024 05:58:46

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:47	
Control Test	0.082	05:48	
Air Blank	0.000	05:48	
Subject Sample	# 1 0.107	05:51	Volume Not Met
Breath Volume	0.089	05:51	
Air Blank	0.000	05:52	
Air Blank	0.000	05:54	
Subject Sample	# 2 0.106	05:57	Volume Not Met
Breath Volume	0.125	05:57	
Air Blank	0.000	05:57	
Control Test	0.079	05:58	
Air Blank	0.000	05:58	
Diagnostics Check	OK	05:58	

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ALLAN

DATE/TIME OF TEST: 10/22/2024 02:56:49

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.083	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.163	02:52
Breath Volume	1.382	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.171	02:55
Breath Volume	1.539	02:55
Air Blank	0.000	02:55
Control Test	0.082	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VAN DYKE,ANDREW
ARRESTING OFFICER AGENCY: Longboat Key PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: KUHN

DATE/TIME OF TEST: 10/22/2024 03:38:33

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.148	03:33
Breath Volume	1.468	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.151	03:37
Breath Volume	1.554	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: VARGAS WISLOW

DATE/TIME OF TEST: 10/23/2024 01:40:48

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.083	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.045	01:36
Breath Volume	1.250	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.046	01:39
Breath Volume	1.222	01:39
Air Blank	0.000	01:39
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERSTEN,DANIEL,N
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NERSTEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/23/2024 02:28:52

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.081	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.161	02:24
Breath Volume	1.449	02:24
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.159	02:27
Breath Volume	1.648	02:27
Air Blank	0.000	02:27
Control Test	0.082	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 10/24/2024 02:23:09

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.083	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.083	02:22	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,ANDREW,G
BREATH TEST OPERATOR AGENCY: Venice PD
ARRESTING OFFICER: NELSON,ANDREW
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: CRAIN

DATE/TIME OF TEST: 10/25/2024 22:15:17

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.083	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:11
Breath Volume	2.187	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	2.503	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,R,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ,O
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GUERRA CRUZ

DATE/TIME OF TEST: 10/26/2024 01:07:36

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.165	01:03
Breath Volume	2.199	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.179	01:06
Breath Volume	2.062	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: RUPERT

DATE/TIME OF TEST: 10/26/2024 01:51:24

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.202	01:46
Breath Volume	1.808	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.207	01:49
Breath Volume	1.898	01:49
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/26/2024 05:58:23

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.083	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.195	05:50
Breath Volume	1.453	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.173	05:53
Breath Volume	1.261	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:56
Subject Sample # 3	0.167	05:56
Breath Volume	1.457	05:56
Air Blank	0.000	05:57
Control Test	0.082	05:57
Air Blank	0.000	05:58
Diagnostics Check	OK	05:58

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CARRILLO PEREZ

DATE/TIME OF TEST: 10/27/2024 03:08:42

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.160	03:04
Breath Volume	1.753	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.152	03:07
Breath Volume	2.480	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/29/2024 22:24:24

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.131	22:19
Breath Volume	1.652	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.128	22:22
Breath Volume	2.070	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACEY,BRANDON,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STACEY,BRANDON
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 10/30/2024 01:20:55

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.083	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.084	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MCINTIRE

DATE/TIME OF TEST: 10/30/2024 21:04:36

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.082	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.153	21:00
Breath Volume	1.511	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.146	21:03
Breath Volume	1.515	21:03
Air Blank	0.000	21:03
Control Test	0.081	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 10/31/2024 18:36:47

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.082	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.051	18:31
Breath Volume	1.328	18:31
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.047	18:35
Breath Volume	1.285	18:35
Air Blank	0.000	18:35
Control Test	0.081	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:36

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: ROSEN

DATE/TIME OF TEST: 10/31/2024 20:55:57

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.082	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.000	20:51
Breath Volume	2.167	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.000	20:54
Breath Volume	2.390	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 11/01/2024 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SUNG

DATE/TIME OF TEST: 10/25/2024 23:58:24

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.150	23:53
Breath Volume	2.500	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.149	23:56
Breath Volume	2.367	23:56
Air Blank	0.000	23:57
Control Test	0.081	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/01/2024 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORTON,GREGORY,W
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NORTON,GREGORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CRUZ LOPEZ

DATE/TIME OF TEST: 10/28/2024 00:36:54

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.082	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.211	00:32
Breath Volume	1.351	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.194	00:35
Breath Volume	1.687	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 11/01/2024 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: BARTOW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Bartow PD
ARRESTING OFFICER: ROQUE,ALEXANDER
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 10/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/07/2024 06:52:59

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.080	06:47
Air Blank	0.000	06:47
Subject Sample # 1	0.062	06:48
Breath Volume	1.812	06:48
Air Blank	0.000	06:48
Air Blank	0.000	06:50
Subject Sample # 2	0.059	06:51
Breath Volume	1.468	06:51
Air Blank	0.000	06:52
Control Test	0.078	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:52

Data Download Date/Time: 11/16/2024 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTZER,P,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LEIARZAR,CHRIS
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 09/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/20/2024 17:36:59

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.079	17:30
Air Blank	0.000	17:30
Subject Sample # 1	0.198	17:32
Breath Volume	3.089	17:32
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.200	17:35
Breath Volume	2.621	17:35
Air Blank	0.000	17:36
Control Test	0.078	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 10/22/2024 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ELLIS,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 10/22/2024 19:31:36

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Subject Sample # 1	0.133	19:26
Breath Volume	1.703	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.128	19:30
Breath Volume	1.722	19:30
Air Blank	0.000	19:30
Control Test	0.079	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARARO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SCHUTTER

DATE/TIME OF TEST: 10/23/2024 00:15:17

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:10
Breath Volume	1.203	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:13
Breath Volume	1.234	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,D,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GERHART,P
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/23/2024 03:28:07

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.054	03:24	
Breath Volume	1.429	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.052	03:27	
Breath Volume	1.265	03:27	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank		03:28	RFI Detect
Air Blank		03:28	RFI Detect

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEGRASSE,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 10/23/2024 17:52:03

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.079	17:46
Air Blank	0.000	17:47
Subject Sample # 1	0.000	17:47
Breath Volume	1.617	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.000	17:50
Breath Volume	1.667	17:50
Air Blank	0.000	17:51
Control Test	0.078	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:52

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STAUB,CPL.
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: PILLINGER

DATE/TIME OF TEST: 10/24/2024 14:15:14

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.079	14:09
Air Blank	0.000	14:09
Subject Sample # 1	0.000	14:10
Breath Volume	1.257	14:10
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.000	14:13
Breath Volume	1.464	14:13
Air Blank	0.000	14:14
Control Test	0.079	14:14
Air Blank	0.000	14:15
Diagnostics Check	OK	14:15

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUOMO,J,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CURTIS,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: LOPSHIRE

DATE/TIME OF TEST: 10/27/2024 23:45:42

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.128	23:41
Breath Volume	1.621	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.127	23:44
Breath Volume	1.781	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LAVANCE,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: RUTLAND

DATE/TIME OF TEST: 10/28/2024 00:14:14

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.155	00:09
Breath Volume	1.648	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.150	00:12
Breath Volume	1.972	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 12/02/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EDWARDS,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BERNSTEIN

DATE/TIME OF TEST: 10/29/2024 01:46:30

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.109	01:41
Breath Volume	1.707	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.104	01:45
Breath Volume	2.000	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 12/02/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CURTIS,R
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GREENUP

DATE/TIME OF TEST: 10/31/2024 01:20:07

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:08	
Control Test	0.079	01:09	
Air Blank	0.000	01:09	
Subject Sample # 1	0.264	01:10	
Breath Volume	1.722	01:10	
Air Blank	0.000	01:11	
Air Blank	0.000	01:12	
Subject Sample # 2		01:16	No Sample Provided
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample # 3		01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 12/02/2024 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODKIN,JASON,M
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: GODKIN,JASON
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 09/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: KNUTSON

DATE/TIME OF TEST: 10/03/2024 01:13:54

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.079	01:09
Breath Volume	2.488	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.079	01:12
Breath Volume	2.328	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 10/18/2024 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELBOGIS,DAVID,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DOWLING,DAKOTA
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 09/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ZELLER

DATE/TIME OF TEST: 10/11/2024 17:06:33

Diagnostics Check	OK	17:00
Air Blank	0.000	17:00
Control Test	0.081	17:00
Air Blank	0.000	17:01
Subject Sample # 1	0.120	17:01
Breath Volume	1.585	17:01
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.118	17:04
Breath Volume	1.539	17:04
Air Blank	0.000	17:05
Control Test	0.080	17:05
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 10/18/2024 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELLERT, DAVID, C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SWIGETT, TIM
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 09/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 10/05/2024
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 10/14/2024 18:42:05

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.080	18:31	
Air Blank	0.000	18:31	
Subject Sample	# 1 0.124	18:34	Volume Not Met
Breath Volume	0.289	18:34	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 2 0.124	18:40	Volume Not Met
Breath Volume	0.855	18:40	
Air Blank	0.000	18:41	
Control Test	0.079	18:41	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:42	

Data Download Date/Time: 10/18/2024 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22722080A1
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/20/2024 00:22:50

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.102	00:18
Breath Volume	1.781	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.100	00:21
Breath Volume	1.691	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 11/08/2024 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN, JOSHUA, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALLEN, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 10/26/2024 22:52:38

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.148	22:47
Breath Volume	2.328	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.141	22:51
Breath Volume	2.234	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 11/08/2024 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: ALLENDESUAREZ

DATE/TIME OF TEST: 10/04/2024 22:40:20

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.078	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.097	22:35
Breath Volume	1.125	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.099	22:38
Breath Volume	3.164	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBINSON, MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 10/07/2024 04:19:16

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.125	04:14
Breath Volume	1.503	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.128	04:17
Breath Volume	1.453	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BUENAFEBA

DATE/TIME OF TEST: 10/09/2024 02:28:18

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.083	02:23
Breath Volume	1.398	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.081	02:26
Breath Volume	1.535	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BONANNO,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 10/12/2024 03:39:09

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.175	03:34
Breath Volume	2.152	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.170	03:37
Breath Volume	2.285	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/13/2024 04:53:32

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.078	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.167	04:49
Breath Volume	2.042	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.168	04:52
Breath Volume	2.523	04:52
Air Blank	0.000	04:52
Control Test	0.077	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: PEREZLOPEZ

DATE/TIME OF TEST: 10/15/2024 00:35:21

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.183	00:30
Breath Volume	2.222	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.187	00:33
Breath Volume	1.757	00:33
Air Blank	0.000	00:34
Control Test	0.077	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SILVESTREDELGADO

DATE/TIME OF TEST: 10/16/2024 01:14:00

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.078	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.170	01:09
Breath Volume	1.203	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.165	01:12
Breath Volume	1.511	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:14

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: OCANDOAGUADO

DATE/TIME OF TEST: 10/16/2024 03:48:42

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.186	03:44
Breath Volume	1.292	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:46
Subject Sample # 2	0.176	03:47
Breath Volume	1.406	03:47
Air Blank	0.000	03:47
Control Test	0.076	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 10/16/2024 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOSEPH,FRANKY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 10/27/2024 02:46:01

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.212	02:41
Breath Volume	1.523	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:43
Subject Sample # 2	0.209	02:44
Breath Volume	1.437	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:46

Data Download Date/Time: 11/19/2024 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERICKSON, JASMINE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MOMPLAISIR

DATE/TIME OF TEST: 10/29/2024 22:55:33

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.161	22:51
Breath Volume	2.285	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.161	22:54
Breath Volume	2.175	22:54
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 11/19/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOSEPH,FRANKY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ALMONTEVASQUEZ

DATE/TIME OF TEST: 10/31/2024 02:41:18

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.151	02:36
Breath Volume	1.886	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.140	02:39
Breath Volume	1.144	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/19/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANCHEZ, ISABELLA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: HULEN

DATE/TIME OF TEST: 10/31/2024 07:14:19

Diagnostics Check	OK	07:07
Air Blank	0.000	07:08
Control Test	0.081	07:08
Air Blank	0.000	07:09
Subject Sample # 1	0.164	07:09
Breath Volume	3.523	07:09
Air Blank	0.000	07:10
Air Blank	0.000	07:12
Subject Sample # 2	0.164	07:12
Breath Volume	3.207	07:12
Air Blank	0.000	07:13
Control Test	0.081	07:13
Air Blank	0.000	07:14
Diagnostics Check	OK	07:14

Data Download Date/Time: 11/19/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA, NIKOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MUNAR

DATE/TIME OF TEST: 10/31/2024 23:10:43

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.092	23:06
Breath Volume	2.609	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.090	23:09
Breath Volume	2.656	23:09
Air Blank	0.000	23:09
Control Test	0.080	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 11/19/2024 12:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT, WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: OAKOWSKY

DATE/TIME OF TEST: 10/08/2024 01:38:08

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.208	01:33
Breath Volume	2.570	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.217	01:36
Breath Volume	2.578	01:36
Air Blank	0.000	01:37
Control Test	0.077	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 10/27/2024 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: COTRAMANO

DATE/TIME OF TEST: 10/12/2024 21:37:01

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.250	21:32
Breath Volume	1.832	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.249	21:35
Breath Volume	1.593	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:37

Data Download Date/Time: 10/27/2024 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MCELLIGATT

DATE/TIME OF TEST: 10/15/2024 01:14:50

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.079	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.077	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 10/27/2024 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PETERS,DARIN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RIVAS-LANCERIO

DATE/TIME OF TEST: 10/20/2024 01:40:44

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.078	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.195	01:36
Breath Volume	1.453	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.186	01:39
Breath Volume	1.621	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 10/27/2024 18:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP, MATTHEW, J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS, CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 10/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/02/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: NOSKOWICZ

DATE/TIME OF TEST: 10/31/2024 01:04:25

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.077	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 11/20/2024 17:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: CRONIN III

DATE/TIME OF TEST: 10/02/2024 15:18:07

Diagnostics Check	OK	15:09
Air Blank	0.000	15:10
Control Test	0.082	15:10
Air Blank	0.000	15:11
Subject Sample # 1	0.236	15:12
Breath Volume	3.765	15:12
Air Blank	0.000	15:13
Air Blank	0.000	15:14
Subject Sample # 2	0.221	15:16
Breath Volume	1.453	15:16
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:18
Diagnostics Check	OK	15:18

Data Download Date/Time: 10/29/2024 21:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAEUTER,DARIAN,R
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DASZUTA,DYLAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:31

SUBJECT'S LAST NAME: GUTBROD

DATE/TIME OF TEST: 10/03/2024 10:59:30

Diagnostics Check	OK	10:52
Air Blank	0.000	10:53
Control Test	0.082	10:53
Air Blank	0.000	10:53
Subject Sample # 1	0.117	10:54
Breath Volume	4.246	10:54
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.105	10:57
Breath Volume	3.265	10:57
Air Blank	0.000	10:58
Control Test	0.080	10:58
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 10/29/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: BAKER,DALTON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BRITT

DATE/TIME OF TEST: 10/21/2024 00:47:04

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.254	00:42
Breath Volume	1.636	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.246	00:45
Breath Volume	1.328	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:47

Data Download Date/Time: 10/29/2024 21:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221264
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/01/2024 22:57:42

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.082	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.168	22:52
Breath Volume	1.277	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.164	22:56
Breath Volume	2.039	22:56
Air Blank	0.000	22:56
Control Test	0.083	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 10/03/2024 21:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVES,J,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: CHAVES,J
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ALMONTE SORIANO

DATE/TIME OF TEST: 10/05/2024 23:49:23

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:47	Subject Test Refused
Breath Volume	0.000	23:47	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 10/20/2024 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEANITE,DUHAMEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JEANITE,DUHAMEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/06/2024 10:16:49

Diagnostics Check	OK	10:10
Air Blank	0.000	10:10
Control Test	0.079	10:10
Air Blank	0.000	10:11
Subject Sample # 1	0.125	10:12
Breath Volume	1.132	10:12
Air Blank	0.000	10:12
Air Blank	0.000	10:14
Subject Sample # 2	0.121	10:15
Breath Volume	1.378	10:15
Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16
Diagnostics Check	OK	10:16

Data Download Date/Time: 10/20/2024 17:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: DAVIS,R
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29122080A1
DRY GAS EXPIRY DATE: 12/05/2024
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: CENOBLE

DATE/TIME OF TEST: 10/11/2024 05:36:31

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.080	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.214	05:31
Breath Volume	1.757	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.223	05:34
Breath Volume	2.285	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 10/20/2024 17:51