

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: GUYTON

DATE/TIME OF TEST: 10/16/2024 21:16:02

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Subject Sample	# 1	21:14	Subject Test Refused
Breath Volume	0.000	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 10/31/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/26/2024 22:40:41

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.079	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.278	22:35
Breath Volume	1.363	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.282	22:38
Breath Volume	1.535	22:38
Air Blank	0.000	22:39
Control Test	0.078	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/31/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SALDIVAR

DATE/TIME OF TEST: 10/30/2024 22:54:38

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:46	
Control Test	0.078	22:46	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:50	No Sample Provided
Breath Volume	0.000	22:50	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:53	
Control Test	0.080	22:54	
Air Blank	0.000	22:54	
Diagnostics Check	OK	22:54	

Data Download Date/Time: 10/31/2024 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/31/2024 19:08:23

Diagnostics Check	OK	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.222	19:03
Breath Volume	2.085	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.209	19:06
Breath Volume	1.554	19:06
Air Blank	0.000	19:07
Control Test	0.078	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 11/28/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806340
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PEREZ CARRILLO

DATE/TIME OF TEST: 10/31/2024 23:39:18

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.234	23:34
Breath Volume	3.085	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.238	23:37
Breath Volume	2.890	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 11/28/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ADAMS, J
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CHIQUIN

DATE/TIME OF TEST: 10/26/2024 23:00:08

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.222	22:55
Breath Volume	3.761	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.214	22:58
Breath Volume	3.996	22:58
Air Blank	0.000	22:59
Control Test	0.076	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 11/23/2024 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: CRUMPTON

DATE/TIME OF TEST: 10/02/2024 23:16:34

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.112	23:12
Breath Volume	1.792	23:12
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.114	23:14
Breath Volume	4.089	23:14
Air Blank	0.000	23:15
Control Test	0.078	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 10/18/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: ENRIGUEZ

DATE/TIME OF TEST: 10/08/2024 23:14:41

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.197	23:09
Breath Volume	1.179	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.190	23:12
Breath Volume	1.574	23:12
Air Blank	0.000	23:13
Control Test	0.077	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 10/18/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 10/11/2024 02:38:14

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.078	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 10/18/2024 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 10/21/2024 23:45:50

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.078	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.122	23:41
Breath Volume	2.605	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.125	23:44
Breath Volume	2.542	23:44
Air Blank	0.000	23:44
Control Test	0.076	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 11/22/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 10/22/2024 23:57:52

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.054	23:53
Breath Volume	2.531	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.051	23:56
Breath Volume	3.136	23:56
Air Blank	0.000	23:56
Control Test	0.075	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 11/22/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: BLENNIS

DATE/TIME OF TEST: 10/25/2024 22:02:12

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.079	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.178	21:57
Breath Volume	1.824	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.176	22:00
Breath Volume	1.859	22:00
Air Blank	0.000	22:01
Control Test	0.076	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: RAMIREZ MORALES

DATE/TIME OF TEST: 10/27/2024 02:23:47

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.194	02:19
Breath Volume	3.085	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.190	02:22
Breath Volume	2.644	02:22
Air Blank	0.000	02:22
Control Test	0.076	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: GORGONIO JIMENEZ

DATE/TIME OF TEST: 10/28/2024 00:14:59

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.197	00:10
Breath Volume	1.968	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.210	00:13
Breath Volume	2.437	00:13
Air Blank	0.000	00:14
Control Test	0.077	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 10/31/2024 21:22:32

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.079	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.197	21:18
Breath Volume	2.804	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.197	21:20
Breath Volume	2.359	21:20
Air Blank	0.000	21:21
Control Test	0.078	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 11/22/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,BRYCE,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMPBELL,AMY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 09/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: JARVIS

DATE/TIME OF TEST: 10/11/2024 14:52:49

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.080	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.355	14:48
Breath Volume	3.035	14:48
Air Blank	0.000	14:49
Air Blank	0.000	14:50
Subject Sample # 2	0.362	14:51
Breath Volume	3.960	14:51
Air Blank	0.000	14:51
Control Test	0.079	14:52
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 10/26/2024 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: MONTANO LASERNA

DATE/TIME OF TEST: 10/25/2024 23:03:19

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.162	22:58
Breath Volume	1.867	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.162	23:01
Breath Volume	1.906	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GIBBS, ANDRE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: PERALTA ZELAYA

DATE/TIME OF TEST: 10/26/2024 03:40:38

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.201	03:34
Breath Volume	1.101	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.201	03:38
Breath Volume	1.218	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRATHER, TRENTON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: PEREZ MORALES

DATE/TIME OF TEST: 10/26/2024 05:47:12

Diagnostics Check	OK	05:40
Air Blank	0.000	05:40
Control Test	0.080	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.212	05:42
Breath Volume	1.910	05:42
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.206	05:45
Breath Volume	2.230	05:45
Air Blank	0.000	05:46
Control Test	0.078	05:46
Air Blank	0.000	05:47
Diagnostics Check	OK	05:47

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERGUSON,RYAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: FRECHETTE

DATE/TIME OF TEST: 10/26/2024 18:57:03

Diagnostics Check	OK	18:45	
Air Blank	0.000	18:45	
Control Test	0.079	18:45	
Air Blank	0.000	18:46	
Subject Sample	# 1 0.211	18:49	Volume Not Met
Breath Volume	1.082	18:49	
Air Blank	0.000	18:50	
Air Blank	0.000	18:52	
Subject Sample	# 2 0.231	18:52	
Breath Volume	1.171	18:52	
Air Blank	0.000	18:53	
Air Blank	0.000	18:55	
Subject Sample	# 3 0.238	18:55	
Breath Volume	1.855	18:55	
Air Blank	0.000	18:56	
Control Test	0.077	18:56	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:57	

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/27/2024 03:35:40

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.199	03:29
Breath Volume	1.246	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.211	03:34
Breath Volume	1.652	03:34
Air Blank	0.000	03:34
Control Test	0.078	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ALCANTARA ARQUINIGO

DATE/TIME OF TEST: 10/27/2024 04:26:49

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.181	04:22
Breath Volume	1.824	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.169	04:25
Breath Volume	2.816	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HABER,ZACHARY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: ROBLERO PEREZ

DATE/TIME OF TEST: 10/27/2024 08:03:26

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:57
Subject Sample # 1	0.174	07:58
Breath Volume	1.996	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:01
Subject Sample # 2	0.165	08:01
Breath Volume	2.335	08:01
Air Blank	0.000	08:02
Control Test	0.080	08:02
Air Blank	0.000	08:03
Diagnostics Check	OK	08:03

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MORERA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: DEY

DATE/TIME OF TEST: 10/31/2024 10:18:47

Diagnostics Check	OK	10:11
Air Blank	0.000	10:11
Control Test	0.081	10:11
Air Blank	0.000	10:12
Subject Sample # 1	0.163	10:14
Breath Volume	1.945	10:14
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.166	10:17
Breath Volume	2.035	10:17
Air Blank	0.000	10:17
Control Test	0.079	10:18
Air Blank	0.000	10:18
Diagnostics Check	OK	10:18

Data Download Date/Time: 11/15/2024 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS, THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 10/01/2024 18:32:52

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.099	18:27
Breath Volume	1.265	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.102	18:31
Breath Volume	1.484	18:31
Air Blank	0.000	18:31
Control Test	0.081	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HLUCH,MARIO
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: FENDER

DATE/TIME OF TEST: 10/04/2024 01:03:17

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:52	
Subject Sample	# 1 0.193	00:55	Volume Not Met
Breath Volume	0.226	00:55	
Air Blank	0.000	00:56	
Air Blank	0.000	00:58	
Subject Sample	# 2 0.211	01:01	Volume Not Met
Breath Volume	0.953	01:01	
Air Blank	0.000	01:02	
Control Test	0.077	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO, AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ARIAS RANGEL

DATE/TIME OF TEST: 10/04/2024 22:05:55

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.000	22:01
Breath Volume	1.621	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.000	22:04
Breath Volume	1.238	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BIBB

DATE/TIME OF TEST: 10/05/2024 01:42:38

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:38
Breath Volume	1.460	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:41
Breath Volume	2.691	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: BARILLAS CORTES

DATE/TIME OF TEST: 10/05/2024 20:52:13

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.225	20:45
Breath Volume	1.316	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.235	20:50
Breath Volume	1.207	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHEHAB,KHODR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: NAGLER

DATE/TIME OF TEST: 10/06/2024 09:59:55

Diagnostics Check	OK	09:53
Air Blank	0.000	09:54
Control Test	0.081	09:54
Air Blank	0.000	09:54
Subject Sample # 1	0.060	09:55
Breath Volume	1.328	09:55
Air Blank	0.000	09:56
Air Blank	0.000	09:57
Subject Sample # 2	0.061	09:58
Breath Volume	1.667	09:58
Air Blank	0.000	09:59
Control Test	0.080	09:59
Air Blank	0.000	09:59
Diagnostics Check	OK	09:59

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ,DINEEN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HERNANDEZ MORENO

DATE/TIME OF TEST: 10/07/2024 00:47:25

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.042	00:42
Breath Volume	1.859	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.044	00:45
Breath Volume	3.113	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: CEDILLO CETO

DATE/TIME OF TEST: 10/08/2024 23:38:57

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.216	23:34
Breath Volume	1.882	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.212	23:37
Breath Volume	1.121	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HILL,NATALIE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 10/09/2024 05:09:16

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1 0.255	05:01	Volume Not Met
Breath Volume	0.125	05:01	
Air Blank	0.000	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 2 0.260	05:04	
Breath Volume	1.246	05:04	
Air Blank	0.000	05:05	
Air Blank	0.000	05:07	
Subject Sample	# 3 0.272	05:07	
Breath Volume	2.820	05:07	
Air Blank	0.000	05:08	
Control Test	0.078	05:08	
Air Blank	0.000	05:09	
Diagnostics Check	OK	05:09	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COLON, VICTOR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 10/09/2024 06:05:47

Diagnostics Check	OK	05:59
Air Blank	0.000	05:59
Control Test	0.079	06:00
Air Blank	0.000	06:00
Subject Sample # 1	0.106	06:01
Breath Volume	1.855	06:01
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.104	06:04
Breath Volume	1.695	06:04
Air Blank	0.000	06:04
Control Test	0.079	06:05
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOSCIALKOWSKI,SEBASTIAN
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: PEPE

DATE/TIME OF TEST: 10/09/2024 19:53:03

Diagnostics Check	OK	19:46
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Subject Sample # 1	0.000	19:48
Breath Volume	1.351	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.000	19:51
Breath Volume	3.429	19:51
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:53

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SACCO

DATE/TIME OF TEST: 10/11/2024 01:47:28

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.000	01:43
Breath Volume	3.476	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.000	01:46
Breath Volume	2.882	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 10/11/2024 23:01:25

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.194	22:56
Breath Volume	4.222	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.207	22:59
Breath Volume	3.722	22:59
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MORALES DOMINGO

DATE/TIME OF TEST: 10/12/2024 01:56:23

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.080	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.176	01:51
Breath Volume	1.867	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.186	01:54
Breath Volume	1.285	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOBLE,RILEY
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: TATE JR

DATE/TIME OF TEST: 10/12/2024 02:13:26

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.153	02:08
Breath Volume	1.316	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.155	02:11
Breath Volume	1.960	02:11
Air Blank	0.000	02:12
Control Test	0.079	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 10/12/2024 03:11:33

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1 0.200	02:59	
Breath Volume	1.304	02:59	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2 0.190	03:04	Volume Not Met
Breath Volume	0.050	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 3 0.185	03:09	
Breath Volume	1.289	03:09	
Air Blank	0.000	03:10	
Control Test	0.078	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: WALTER

DATE/TIME OF TEST: 10/12/2024 04:06:34

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:02
Breath Volume	1.726	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	1.144	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 10/12/2024 18:35:06

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.081	18:29
Air Blank	0.000	18:30
Subject Sample # 1	0.000	18:30
Breath Volume	1.988	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.000	18:33
Breath Volume	2.363	18:33
Air Blank	0.000	18:34
Control Test	0.081	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RIVAS, LUIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: KIELB

DATE/TIME OF TEST: 10/13/2024 00:03:04

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.195	23:58
Breath Volume	1.113	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.200	00:01
Breath Volume	1.523	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:03

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLAKE, XAVIAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: KLEE

DATE/TIME OF TEST: 10/13/2024 01:38:26

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.131	01:33
Breath Volume	1.121	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:35
Subject Sample # 2	0.133	01:36
Breath Volume	1.316	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CANO GALICIA

DATE/TIME OF TEST: 10/13/2024 02:21:33

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.196	02:16
Breath Volume	2.273	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.201	02:19
Breath Volume	2.093	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRUDHOMME, DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: COLAS

DATE/TIME OF TEST: 10/13/2024 03:50:01

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:49	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARABALLO,CRISTIAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BARRIOS GARCIA

DATE/TIME OF TEST: 10/14/2024 00:20:54

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1 0.159	00:11	Slope Not Level
Breath Volume	1.195	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2 0.169	00:16	
Breath Volume	2.101	00:16	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample	# 3 0.166	00:19	
Breath Volume	1.671	00:19	
Air Blank	0.000	00:19	
Control Test	0.078	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LOPEZ MENDOZA

DATE/TIME OF TEST: 10/15/2024 02:55:57

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1 0.176	02:46	Volume Not Met
Breath Volume	0.769	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:49	
Subject Sample	# 2 0.171	02:50	
Breath Volume	2.156	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 3 0.165	02:54	
Breath Volume	1.746	02:54	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: JIMENEZ ORTIZ

DATE/TIME OF TEST: 10/18/2024 01:27:33

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.213	01:23
Breath Volume	2.238	01:23
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.211	01:25
Breath Volume	2.855	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERA, FRANKLYN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: CALIX CRUZ

DATE/TIME OF TEST: 10/18/2024 21:07:56

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.081	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.131	21:03
Breath Volume	2.625	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.125	21:06
Breath Volume	2.382	21:06
Air Blank	0.000	21:07
Control Test	0.081	21:07
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: OXLAJ CHIC

DATE/TIME OF TEST: 10/18/2024 23:26:22

Diagnostics Check	OK	23:19
Air Blank	0.000	23:20
Control Test	0.081	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.155	23:21
Breath Volume	1.757	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.154	23:24
Breath Volume	1.496	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LOPEZ ROMERO

DATE/TIME OF TEST: 10/19/2024 04:49:24

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.165	04:44
Breath Volume	1.261	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.161	04:47
Breath Volume	3.132	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE, DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: LUZURIAGA

DATE/TIME OF TEST: 10/19/2024 21:09:09

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.081	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.206	21:04
Breath Volume	1.839	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.189	21:07
Breath Volume	1.589	21:07
Air Blank	0.000	21:08
Control Test	0.080	21:08
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BARCO GARCIA

DATE/TIME OF TEST: 10/19/2024 21:30:38

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.187	21:26
Breath Volume	2.628	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.178	21:29
Breath Volume	1.101	21:29
Air Blank	0.000	21:29
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: FUENTES RODRIGUEZ

DATE/TIME OF TEST: 10/20/2024 01:35:02

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.163	01:30
Breath Volume	2.132	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.159	01:33
Breath Volume	1.953	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SANTOS VALDES

DATE/TIME OF TEST: 10/20/2024 04:31:50

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.131	04:27
Breath Volume	1.347	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.128	04:30
Breath Volume	1.335	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 10/25/2024 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAPPIN, KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 10/20/2024 23:41:30

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Subject Sample	# 1	23:34	No Sample Provided
Breath Volume	0.000	23:34	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample	# 2	0.057	23:39
Breath Volume	0.718	23:39	Volume Not Met
Air Blank	0.000	23:40	
Control Test	0.081	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAPPIN, KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 10/20/2024 23:53:08

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.056	23:48
Breath Volume	1.335	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.057	23:51
Breath Volume	1.257	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: DIEGO

DATE/TIME OF TEST: 10/21/2024 02:36:02

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.142	02:31
Breath Volume	1.117	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.145	02:34
Breath Volume	1.707	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:36

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BALES,DOUGLAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 10/21/2024 16:02:31

Diagnostics Check	OK	15:56
Air Blank	0.000	15:56
Control Test	0.080	15:56
Air Blank	0.000	15:57
Subject Sample # 1	0.137	15:58
Breath Volume	1.496	15:58
Air Blank	0.000	15:58
Air Blank	0.000	16:00
Subject Sample # 2	0.141	16:00
Breath Volume	1.406	16:00
Air Blank	0.000	16:01
Control Test	0.080	16:01
Air Blank	0.000	16:02
Diagnostics Check	OK	16:02

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HECTOR,BRADY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/23/2024 01:29:23

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1 0.000	01:20	Volume Not Met
Breath Volume	1.121	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2 0.000	01:24	
Breath Volume	1.476	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 3 0.000	01:28	
Breath Volume	1.148	01:28	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOPEZ, BRANDON
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: LOUIS JEAN

DATE/TIME OF TEST: 10/24/2024 16:58:05

Diagnostics Check	OK	16:51
Air Blank	0.000	16:52
Control Test	0.081	16:52
Air Blank	0.000	16:53
Subject Sample # 1	0.344	16:53
Breath Volume	1.265	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.332	16:56
Breath Volume	1.183	16:56
Air Blank	0.000	16:57
Control Test	0.080	16:57
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MOREL

DATE/TIME OF TEST: 10/25/2024 03:07:41

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1 0.226	02:56	
Breath Volume	1.640	02:56	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 3	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.080	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 10/25/2024 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: VILAHOMAT

DATE/TIME OF TEST: 10/26/2024 03:30:37

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.173	03:25
Breath Volume	2.125	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.164	03:29
Breath Volume	2.503	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRAFF,HEATHER
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/26/2024 07:37:45

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.082	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.156	07:32
Breath Volume	1.351	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.153	07:36
Breath Volume	1.765	07:36
Air Blank	0.000	07:36
Control Test	0.079	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,DYLAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GUZMAN VASQUEZ

DATE/TIME OF TEST: 10/27/2024 03:08:25

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.144	03:04
Breath Volume	2.394	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.145	03:06
Breath Volume	2.085	03:06
Air Blank	0.000	03:07
Control Test	0.081	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: AHMED

DATE/TIME OF TEST: 10/27/2024 03:33:41

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.180	03:25
Breath Volume	2.968	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.158	03:29
Breath Volume	3.300	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 3	0.169	03:32
Breath Volume	3.000	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PRATHER,TRENTON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ANDERT

DATE/TIME OF TEST: 10/27/2024 04:38:23

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.156	04:33
Breath Volume	2.175	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.161	04:36
Breath Volume	2.792	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/27/2024 22:19:27

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:15
Breath Volume	1.226	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.000	22:18
Breath Volume	1.753	22:18
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: POINTU, PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: NOELSAINT

DATE/TIME OF TEST: 10/28/2024 05:12:37

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.082	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.153	05:08
Breath Volume	1.210	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.158	05:11
Breath Volume	2.078	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEXIS, VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/30/2024 22:38:10

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.081	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.152	22:33
Breath Volume	2.980	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.154	22:36
Breath Volume	2.765	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DESUYO,AISON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MIELE

DATE/TIME OF TEST: 10/30/2024 23:36:41

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.081	23:30
Air Blank	0.000	23:30
Subject Sample # 1	0.209	23:31
Breath Volume	1.886	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.223	23:35
Breath Volume	2.140	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: LOPEZ ORTIZ

DATE/TIME OF TEST: 10/31/2024 00:14:47

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.193	00:07
Breath Volume	3.039	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.171	00:10
Breath Volume	1.195	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 3	0.178	00:13
Breath Volume	1.480	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 11/15/2024 16:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEDESCH, LAURA, R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BUETTNER, CADEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/01/2024 20:55:22

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.084	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.162	20:50
Breath Volume	2.156	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:52
Subject Sample # 2	0.171	20:53
Breath Volume	2.152	20:53
Air Blank	0.000	20:54
Control Test	0.082	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 10/11/2024 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BABA,TANNER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: FANN JR

DATE/TIME OF TEST: 10/02/2024 00:05:53

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.082	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	1.863	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:04
Breath Volume	1.914	00:04
Air Blank	0.000	00:04
Control Test	0.080	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/11/2024 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,TYLER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HARDRICK

DATE/TIME OF TEST: 10/04/2024 22:47:14

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.082	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:41	No Sample Provided
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2	22:46	Sequence Aborted
Breath Volume	0.000	22:46	
Air Blank	0.000	22:47	

Data Download Date/Time: 10/11/2024 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,TYLER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: HARDRICK

DATE/TIME OF TEST: 10/04/2024 23:14:38

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.082	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.213	23:09
Breath Volume	1.468	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.196	23:12
Breath Volume	1.242	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 10/11/2024 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402498227
DRY GAS EXPIRY DATE: 07/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 10/10/2024 22:36:24

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.083	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.235	22:30
Breath Volume	1.546	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.221	22:34
Breath Volume	1.156	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 10/11/2024 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIDGEN,STEPHANIE,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STOKES,BRAYDEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PAGEL

DATE/TIME OF TEST: 10/01/2024 19:33:09

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.304	19:28
Breath Volume	1.496	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:30
Subject Sample # 2	0.301	19:31
Breath Volume	1.207	19:31
Air Blank	0.000	19:32
Control Test	0.077	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 10/02/2024 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STOKES,BRAYDEN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/04/2024 23:40:56

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:38	
Control Test	0.079	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.078	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: LAWSON,SKYLER
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 10/07/2024 05:22:30

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.078	05:14	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:18	No Sample Provided
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Air Blank	0.000	05:20	
Subject Sample	# 2	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:21	
Control Test	0.078	05:21	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MESHES,GREG
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/11/2024 02:41:30

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.125	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GARRIDO,ANTHONY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/12/2024 02:54:39

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.174	02:49
Breath Volume	2.003	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.172	02:52
Breath Volume	1.683	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENICK,MARK,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GARRIDO,ANTHONY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: PILSON

DATE/TIME OF TEST: 10/13/2024 23:01:37

Diagnostics Check	OK	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.254	22:56
Breath Volume	2.296	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.254	22:59
Breath Volume	2.746	22:59
Air Blank	0.000	23:00
Control Test	0.078	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,KYLE,R
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MESHES,GREGORY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 09/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/15/2024 22:52:31

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.078	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.078	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 10/22/2024 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: DREW,LIAM
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: RICO ZARATE

DATE/TIME OF TEST: 10/11/2024 00:10:38

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.121	00:06
Breath Volume	3.289	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.126	00:08
Breath Volume	3.363	00:08
Air Blank	0.000	00:09
Control Test	0.077	00:10
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,GARY,B
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HULL,GARY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: RUBIO HERNANDEZ

DATE/TIME OF TEST: 10/17/2024 15:40:27

Diagnostics Check	OK	15:33
Air Blank	0.000	15:34
Control Test	0.079	15:34
Air Blank	0.000	15:35
Subject Sample # 1	0.246	15:35
Breath Volume	1.847	15:35
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.247	15:38
Breath Volume	1.617	15:38
Air Blank	0.000	15:39
Control Test	0.077	15:39
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: IZMA,KENNY
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: ALONZO HERRERA

DATE/TIME OF TEST: 10/20/2024 04:23:56

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.200	04:16
Breath Volume	1.804	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.200	04:21
Breath Volume	2.457	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: RAMIREZ GARCIA,HERNAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DIAZ-CARRENO

DATE/TIME OF TEST: 10/26/2024 03:08:58

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BARRIRA,BRAYAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: VILLAGOMEZ

DATE/TIME OF TEST: 10/27/2024 02:28:29

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.244	02:23
Breath Volume	2.644	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.236	02:26
Breath Volume	1.773	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LARRY,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TOMBLIN,JASON
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MARTINEZ SANCHEZ

DATE/TIME OF TEST: 10/27/2024 23:43:21

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:40	
Control Test	0.080	23:41	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:41	Subject Test Refused
Breath Volume	0.000	23:41	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 11/04/2024 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIAGET,PHILIP,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VANPELT,COLTEN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 10/04/2024 23:53:18

Diagnostics Check	OK	23:46	
Air Blank	0.000	23:47	
Control Test	0.081	23:47	
Air Blank	0.000	23:48	
Subject Sample	# 1 0.192	23:48	
Breath Volume	1.652	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2 0.189	23:51	
Breath Volume	1.652	23:51	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank		23:52	RFI Detect
Air Blank	0.000	23:53	

Data Download Date/Time: 10/16/2024 10:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/13/2024 21:46:59

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:45	RFI Detect
Breath Volume	0.308	21:45	
Air Blank	0.000	21:46	
Control Test	0.079	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/13/2024 21:55:45

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.076	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1 0.134	21:54	Volume Not Met
Breath Volume	0.019	21:54	
Air Blank		21:55	Purge Fail
Air Blank	0.000	21:55	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/13/2024 22:05:08

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:04	
Control Test	0.074	22:04	Control Outside Tolerance
Air Blank	0.000	22:05	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: BUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: L120280946270

DATE/TIME OF TEST: 10/13/2024 22:19:47

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.077	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1 0.146	22:12	Volume Not Met
Breath Volume	0.062	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 2 0.136	22:18	Volume Not Met
Breath Volume	0.328	22:18	
Air Blank		22:19	Purge Fail
Air Blank	0.000	22:19	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUBBS,MILTON,B
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 10/20/2024 23:11:34

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:06	
Control Test	0.078	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:10	RFI Detect
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.078	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUBBS,MILTON,B
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 10/20/2024 23:21:21

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:15	
Control Test	0.079	23:15	
Air Blank	0.000	23:16	
Subject Sample # 1	0.220	23:18	
Breath Volume	2.015	23:18	
Air Blank	0.000	23:19	
Air Blank		23:20	Ambient Fail
Air Blank	0.000	23:21	

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUBBS,MILTON,B
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WATTLEY,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SEQUIERA

DATE/TIME OF TEST: 10/20/2024 23:38:06

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.216	23:33
Breath Volume	2.082	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.208	23:36
Breath Volume	1.933	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARANA,MARLON,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROQUE,EUGENE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 10/27/2024 00:58:23

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Subject Sample # 1	0.156	00:53
Breath Volume	2.839	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.161	00:56
Breath Volume	1.707	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 10/30/2024 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS 2, LEONARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DOMINGUEZ, OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BURGOS 2

DATE/TIME OF TEST: 10/01/2024 04:22:00

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.000	04:17
Breath Volume	1.457	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:20
Breath Volume	1.300	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:21
Diagnostics Check	OK	04:22

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/01/2024 22:45:40

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:44	
Control Test	0.079	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 10/05/2024 02:16:02

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.080	02:13	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: SMITS,JESSICA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: OSTEEN

DATE/TIME OF TEST: 10/05/2024 22:36:42

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.079	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE, NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GILBERT HERNANDEZ

DATE/TIME OF TEST: 10/05/2024 23:23:48

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.108	23:18
Breath Volume	1.488	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.106	23:22
Breath Volume	2.339	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/07/2024 02:15:52

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN, DAVID,
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: GUZMAN, DAVID
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 09/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/08/2024 16:50:14

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:40
Subject Sample # 1	0.304	16:42
Breath Volume	1.191	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.320	16:48
Breath Volume	1.789	16:48
Air Blank	0.000	16:49
Control Test	0.079	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 10/10/2024 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN JACQUES, WEDLET,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: JEAN JACQUES, WEDLET
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:11

SUBJECT'S LAST NAME: MCDOUGALL

DATE/TIME OF TEST: 10/12/2024 06:37:34

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.079	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.108	06:32
Breath Volume	2.199	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.112	06:36
Breath Volume	2.179	06:36
Air Blank	0.000	06:36
Control Test	0.078	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA, OSCAR, A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MARIN, MARIO
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: FALCON BEOVIDES

DATE/TIME OF TEST: 10/12/2024 22:52:36

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.000	22:51	
Air Blank	0.000	22:51	
Control Test	0.079	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BERKOWITZ

DATE/TIME OF TEST: 10/15/2024 00:07:57

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:04	
Control Test	0.079	00:04	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.078	00:07	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN, JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: PUNSHON

DATE/TIME OF TEST: 10/18/2024 20:10:43

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.086	20:05
Breath Volume	1.789	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.080	20:09
Breath Volume	1.542	20:09
Air Blank	0.000	20:09
Control Test	0.077	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: BRADY,JOSHUA
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/18/2024 23:04:42

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.219	22:59
Breath Volume	1.375	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.211	23:03
Breath Volume	1.812	23:03
Air Blank	0.000	23:03
Control Test	0.076	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 10/19/2024 02:19:24

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.053	02:14
Breath Volume	1.800	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.052	02:17
Breath Volume	2.054	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/25/2024 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 10/19/2024 03:10:15

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.079	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 10/19/2024 21:02:05

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.079	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:01	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 10/20/2024 01:00:16

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.080	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:59	
Control Test	0.080	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN,JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: GROARK

DATE/TIME OF TEST: 10/20/2024 02:26:27

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.123	02:21
Breath Volume	1.207	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:23
Subject Sample # 2	0.125	02:24
Breath Volume	1.550	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,SHANE,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MULLIS

DATE/TIME OF TEST: 10/20/2024 05:19:50

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:17	
Control Test	0.079	05:17	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.079	05:19	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DIEZ-ARGUELLES

DATE/TIME OF TEST: 10/21/2024 00:36:27

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.078	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: DOMINGUEZ,OSCAR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 10/21/2024 01:16:03

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.079	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA, WILLIAM, J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: GUERRA, WILLIAM
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: FALCUCCI

DATE/TIME OF TEST: 10/23/2024 19:05:06

Diagnostics Check	OK	19:02	
Air Blank	0.000	19:02	
Control Test	0.079	19:03	
Air Blank	0.000	19:03	
Subject Sample	# 1	19:03	Subject Test Refused
Breath Volume	0.000	19:03	
Air Blank	0.000	19:04	
Control Test	0.078	19:04	
Air Blank	0.000	19:05	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/24/2024 22:20:04

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.079	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.000	22:15
Breath Volume	2.003	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:18
Breath Volume	3.449	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:20

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GONZALEZ GUTIERREZ

DATE/TIME OF TEST: 10/24/2024 23:30:25

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.079	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: LACY

DATE/TIME OF TEST: 10/27/2024 22:39:56

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.079	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.079	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR, D
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: ORINSTEIN, JOSEPH
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PALLAVICINI

DATE/TIME OF TEST: 10/30/2024 01:32:45

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.138	01:28
Breath Volume	1.609	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.142	01:31
Breath Volume	1.695	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 11/25/2024 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MILLLS

DATE/TIME OF TEST: 10/12/2024 03:45:40

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.227	03:41
Breath Volume	2.250	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.216	03:44
Breath Volume	1.468	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/28/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ROMERO MUNGUIA

DATE/TIME OF TEST: 10/13/2024 03:18:56

Diagnostics Check	OK	03:18	
Air Blank		03:18	RFI Detect
Air Blank	0.000	03:18	

Data Download Date/Time: 10/28/2024 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ROMERO MUNGUIA

DATE/TIME OF TEST: 10/13/2024 03:40:21

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.086	03:35
Breath Volume	1.667	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.089	03:38
Breath Volume	1.546	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/28/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 10/17/2024 22:57:55

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.138	22:53	
Breath Volume	2.640	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2 0.146	22:56	
Breath Volume	2.460	22:56	
Air Blank		22:57	RFI Detect
Air Blank	0.000	22:57	

Data Download Date/Time: 10/28/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 10/17/2024 23:07:31

Diagnostics Check	OK	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.145	23:02
Breath Volume	3.093	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.142	23:05
Breath Volume	2.097	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 10/28/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 09/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HILTON

DATE/TIME OF TEST: 10/19/2024 03:02:05

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.079	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 10/28/2024 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WADDELL,NATHANIEL,J
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 10/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402310313
DRY GAS EXPIRY DATE: 01/15/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/18/2024 04:30:57

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:28	
Control Test	0.081	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Control Test	0.082	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 11/22/2024 15:38