

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MATIAS

DATE/TIME OF TEST: 02/01/2025 02:31:59

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.068	02:26
Breath Volume	1.308	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.068	02:30
Breath Volume	1.605	02:30
Air Blank	0.000	02:31
Control Test	0.075	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/21/2025 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/06/2025 02:04:59

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.024	02:00
Breath Volume	1.367	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.020	02:03
Breath Volume	1.242	02:03
Air Blank	0.000	02:04
Control Test	0.076	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 02/21/2025 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 02/08/2025 00:17:30

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.078	00:12	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.133	00:15	
Breath Volume	1.414	00:15	
Air Blank		00:16	RFI Detect
Air Blank	0.000	00:16	
Control Test	0.078	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 02/21/2025 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 02/08/2025 00:27:02

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.076	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.135	00:22
Breath Volume	1.394	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.132	00:25
Breath Volume	1.207	00:25
Air Blank	0.000	00:26
Control Test	0.075	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

Data Download Date/Time: 02/21/2025 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: POPOVICH

DATE/TIME OF TEST: 02/14/2025 04:53:14

Diagnostics Check	OK	04:47
Air Blank	0.000	04:47
Control Test	0.078	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.000	04:48
Breath Volume	3.957	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.000	04:51
Breath Volume	2.820	04:51
Air Blank	0.000	04:52
Control Test	0.077	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 02/21/2025 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 02/14/2025 21:02:59

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.077	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.018	20:58
Breath Volume	2.027	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.015	21:01
Breath Volume	1.679	21:01
Air Blank	0.000	21:02
Control Test	0.077	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 02/21/2025 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/15/2025 21:00:04

Diagnostics Check	OK	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.161	20:55
Breath Volume	1.730	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.158	20:58
Breath Volume	1.601	20:58
Air Blank	0.000	20:59
Control Test	0.076	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	21:00

Data Download Date/Time: 02/21/2025 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,JOHN,A
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: RIVERS,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: RIVERA AVILES

DATE/TIME OF TEST: 02/21/2025 03:50:10

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.078	03:44
Air Blank	0.000	03:45
Subject Sample # 1	0.160	03:45
Breath Volume	1.710	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.150	03:48
Breath Volume	2.007	03:48
Air Blank	0.000	03:49
Control Test	0.078	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 02/21/2025 12:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: PALMISANO

DATE/TIME OF TEST: 02/13/2025 21:51:01

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.096	21:44
Breath Volume	1.578	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.090	21:49
Breath Volume	1.929	21:49
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:51

Data Download Date/Time: 02/19/2025 10:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: ALITZ

DATE/TIME OF TEST: 02/04/2025 19:45:17

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:42	
Control Test	0.079	19:43	
Air Blank	0.000	19:43	
Subject Sample	# 1	19:43	Subject Test Refused
Breath Volume	0.000	19:43	
Air Blank	0.000	19:44	
Control Test	0.079	19:44	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: KIEKBUSH

DATE/TIME OF TEST: 02/05/2025 05:38:30

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:36	
Control Test	0.079	05:36	
Air Blank	0.000	05:37	
Subject Sample	# 1	05:37	Subject Test Refused
Breath Volume	0.000	05:37	
Air Blank	0.000	05:37	
Control Test	0.080	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ, JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/08/2025 02:13:55

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.147	02:09
Breath Volume	1.199	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.151	02:12
Breath Volume	1.320	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: STACY, CONNOR
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: PRITTY

DATE/TIME OF TEST: 02/08/2025 02:44:01

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.208	02:39
Breath Volume	1.371	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.198	02:42
Breath Volume	1.226	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AREOSTATICO,NICHOLAS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PAZ HERNANDEZ

DATE/TIME OF TEST: 02/08/2025 03:06:31

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.079	03:04	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:05	Subject Test Refused
Breath Volume	0.000	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:05	
Air Blank	0.000	03:06	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GERMOND

DATE/TIME OF TEST: 02/09/2025 22:53:34

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.085	22:48
Breath Volume	1.238	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.081	22:52
Breath Volume	1.800	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RUCKER

DATE/TIME OF TEST: 02/13/2025 01:31:43

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.078	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.079	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 02/13/2025 20:10:08

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.080	20:08	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.079	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SCHUPP

DATE/TIME OF TEST: 02/18/2025 03:10:16

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:07	
Control Test	0.079	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 02/18/2025 20:27:17

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:24	
Control Test	0.079	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1	20:25	Subject Test Refused
Breath Volume	0.000	20:25	
Air Blank	0.000	20:26	
Control Test	0.080	20:26	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 02/19/2025 10:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANKOWSKI,BENJAMIN,M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BROWN,BRITTANY
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 01/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MASSA REYES

DATE/TIME OF TEST: 02/14/2025 22:00:16

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.019	21:55
Breath Volume	2.304	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.016	21:58
Breath Volume	2.453	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/24/2025 22:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMERI, DOMINIC, J
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: THOMPSON, ERIC
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 01/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/22/2025 04:28:21

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.224	04:23
Breath Volume	2.375	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.222	04:26
Breath Volume	2.367	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 02/24/2025 22:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWAB,JOSEPH,S
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: SHUMATE,TODD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:21

SUBJECT'S LAST NAME: WUEST

DATE/TIME OF TEST: 02/01/2025 12:51:02

Diagnostics Check	OK	12:44
Air Blank	0.000	12:45
Control Test	0.081	12:45
Air Blank	0.000	12:46
Subject Sample # 1	0.209	12:46
Breath Volume	1.707	12:46
Air Blank	0.000	12:47
Air Blank	0.000	12:49
Subject Sample # 2	0.197	12:49
Breath Volume	1.937	12:49
Air Blank	0.000	12:50
Control Test	0.080	12:50
Air Blank	0.000	12:50
Diagnostics Check	OK	12:51

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KAPLAN,MICHAEL
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: FLANDERS

DATE/TIME OF TEST: 02/01/2025 18:31:35

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.081	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.000	18:27
Breath Volume	1.710	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.000	18:30
Breath Volume	1.855	18:30
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: ROCK,JAYQUAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: DEPACE

DATE/TIME OF TEST: 02/01/2025 23:32:30

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.219	23:27
Breath Volume	1.847	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.209	23:30
Breath Volume	1.890	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: GOLDSTEIN,SHELLI
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/09/2025 23:15:18

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.082	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.193	23:10
Breath Volume	2.464	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.199	23:13
Breath Volume	2.425	23:13
Air Blank	0.000	23:14
Control Test	0.080	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: O'CONNELL, WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GUANDOLO

DATE/TIME OF TEST: 02/15/2025 03:13:18

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.274	03:08
Breath Volume	4.441	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.264	03:11
Breath Volume	3.812	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: O'CONNELL, WILLIAM
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: COLLIS

DATE/TIME OF TEST: 02/16/2025 03:00:29

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.181	02:55
Breath Volume	1.222	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.181	02:58
Breath Volume	1.542	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,JASON,B
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CARROLL,STEVEN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 01/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: OVALLE

DATE/TIME OF TEST: 02/21/2025 08:42:12

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.081	08:37
Air Blank	0.000	08:37
Subject Sample # 1	0.000	08:38
Breath Volume	4.394	08:38
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.000	08:40
Breath Volume	4.082	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 02/25/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: SANCHEZ TERAQ

DATE/TIME OF TEST: 02/09/2025 21:30:44

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.222	21:26
Breath Volume	2.496	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.218	21:29
Breath Volume	2.695	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 02/26/2025 12:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUSSEF, ANDREW,
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: ADAMS, J
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: PEREZ NAVARIJO

DATE/TIME OF TEST: 02/15/2025 06:56:05

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:47
Subject Sample # 1	0.154	06:49
Breath Volume	2.140	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:51
Subject Sample # 2	0.154	06:54
Breath Volume	1.996	06:54
Air Blank	0.000	06:55
Control Test	0.077	06:55
Air Blank	0.000	06:55
Diagnostics Check	OK	06:56

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RUNKEY

DATE/TIME OF TEST: 02/19/2025 05:04:11

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.081	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.076	04:59
Breath Volume	1.515	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.083	05:02
Breath Volume	2.261	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MORSE,K
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/21/2025 21:10:41

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:09	
Control Test	0.082	21:09	
Air Blank	0.000	21:09	
Subject Sample	# 1	21:10	Sequence Aborted
Breath Volume	0.000	21:10	
Air Blank	0.000	21:10	

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: MORSE,K
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/21/2025 21:20:00

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.168	21:15
Breath Volume	2.226	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:17
Subject Sample # 2	0.174	21:18
Breath Volume	1.593	21:18
Air Blank	0.000	21:19
Control Test	0.081	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HAMM,MARSHALL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CORNETT

DATE/TIME OF TEST: 02/23/2025 00:03:11

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.082	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: COCOA BEACH PD
ARRESTING OFFICER: HAMM,MARSHALL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: CORNETT

DATE/TIME OF TEST: 02/23/2025 00:22:35

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.204	00:17
Breath Volume	1.546	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.218	00:20
Breath Volume	2.035	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 02/24/2025 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEY,KYLE,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: CALDERON,ANDREA
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: SANTIZ

DATE/TIME OF TEST: 02/02/2025 20:55:12

Diagnostics Check	OK	20:48
Air Blank	0.000	20:49
Control Test	0.081	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.137	20:50
Breath Volume	3.375	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.130	20:53
Breath Volume	1.878	20:53
Air Blank	0.000	20:54
Control Test	0.079	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 02/26/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEY,KYLE,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXI
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: PULESSOTO

DATE/TIME OF TEST: 02/03/2025 00:34:24

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.133	00:29
Breath Volume	4.144	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.134	00:32
Breath Volume	3.843	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 02/26/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: WALKER,BRIAN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/03/2025 01:23:26

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.000	01:19
Breath Volume	2.769	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.000	01:22
Breath Volume	2.613	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 02/26/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEY,KYLE,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: SAINTVICTOR

DATE/TIME OF TEST: 02/06/2025 19:03:59

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.080	18:58
Air Blank	0.000	18:59
Subject Sample # 1	0.285	18:59
Breath Volume	3.445	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:02
Subject Sample # 2	0.275	19:02
Breath Volume	1.171	19:02
Air Blank	0.000	19:03
Control Test	0.080	19:03
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 02/26/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEX
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: PONSI

DATE/TIME OF TEST: 02/06/2025 23:30:45

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.000	23:26
Breath Volume	2.015	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.000	23:29
Breath Volume	2.156	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 02/26/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 15:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/07/2025 00:19:38

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.188	00:14
Breath Volume	1.433	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.182	00:18
Breath Volume	1.628	00:18
Air Blank	0.000	00:18
Control Test	0.081	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 02/26/2025 10:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MOREMAN

DATE/TIME OF TEST: 02/01/2025 03:46:13

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.015	03:41
Breath Volume	1.980	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.013	03:44
Breath Volume	1.898	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ALAMO

DATE/TIME OF TEST: 02/02/2025 00:13:06

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.090	00:07
Breath Volume	1.359	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.089	00:11
Breath Volume	1.953	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: STUKEL

DATE/TIME OF TEST: 02/02/2025 03:02:32

Diagnostics Check	OK	02:59	
Air Blank	0.000	03:00	
Control Test	0.082	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: HOYOS

DATE/TIME OF TEST: 02/02/2025 06:42:28

Diagnostics Check	OK	06:35
Air Blank	0.000	06:36
Control Test	0.082	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.105	06:37
Breath Volume	1.371	06:37
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.100	06:40
Breath Volume	1.367	06:40
Air Blank	0.000	06:41
Control Test	0.081	06:41
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 02/03/2025 01:13:35

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.155	01:09
Breath Volume	2.167	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.143	01:11
Breath Volume	2.312	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 02/03/2025 23:28:46

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.082	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.082	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 02/05/2025 00:59:48

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.000	00:55
Breath Volume	1.257	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.000	00:58
Breath Volume	1.589	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: GRUBBS

DATE/TIME OF TEST: 02/05/2025 22:41:25

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.082	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 02/06/2025 02:25:40

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.082	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.081	02:25	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: SOUTHARD

DATE/TIME OF TEST: 02/06/2025 06:32:27

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.082	06:30	
Air Blank	0.000	06:30	
Subject Sample	# 1	06:31	Subject Test Refused
Breath Volume	0.000	06:31	
Air Blank	0.000	06:31	
Control Test	0.081	06:31	
Air Blank	0.000	06:32	
Diagnostics Check	OK	06:32	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 02/07/2025 05:42:15

Diagnostics Check	OK	05:35
Air Blank	0.000	05:36
Control Test	0.082	05:36
Air Blank	0.000	05:37
Subject Sample # 1	0.236	05:37
Breath Volume	1.578	05:37
Air Blank	0.000	05:38
Air Blank	0.000	05:40
Subject Sample # 2	0.217	05:40
Breath Volume	1.367	05:40
Air Blank	0.000	05:41
Control Test	0.081	05:41
Air Blank	0.000	05:42
Diagnostics Check	OK	05:42

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 02/07/2025 23:42:03

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:39	
Control Test	0.083	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.082	23:41	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 02/08/2025 04:40:18

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.082	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.179	04:34
Breath Volume	1.238	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.181	04:38
Breath Volume	2.312	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 02/08/2025 05:57:10

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.082	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.106	05:52
Breath Volume	1.289	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:54
Subject Sample # 2	0.114	05:55
Breath Volume	2.230	05:55
Air Blank	0.000	05:56
Control Test	0.081	05:56
Air Blank	0.000	05:57
Diagnostics Check	OK	05:57

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GROTH

DATE/TIME OF TEST: 02/09/2025 04:11:02

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.082	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.176	04:06
Breath Volume	1.687	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.171	04:09
Breath Volume	1.847	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARSLON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 02/10/2025 00:27:59

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.082	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 02/24/2025 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: CABASSA MISKIMEN

DATE/TIME OF TEST: 02/10/2025 01:47:33

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.082	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.135	01:42
Breath Volume	3.167	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.131	01:45
Breath Volume	2.535	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SHAW,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 02/10/2025 03:59:34

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:57	
Control Test	0.083	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:58	
Control Test	0.082	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/13/2025 21:25:40

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.083	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.099	21:20
Breath Volume	2.796	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.088	21:23
Breath Volume	1.796	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/14/2025 04:51:34

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.082	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.094	04:46
Breath Volume	3.425	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.084	04:49
Breath Volume	2.640	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: PINCOMB

DATE/TIME OF TEST: 02/14/2025 22:46:20

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:41
Breath Volume	2.195	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.000	22:44
Breath Volume	2.921	22:44
Air Blank	0.000	22:45
Control Test	0.082	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KRAMER,S
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 02/14/2025 23:57:17

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.082	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.174	23:52
Breath Volume	2.156	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.173	23:55
Breath Volume	1.652	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: WANG

DATE/TIME OF TEST: 02/15/2025 03:30:54

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.082	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:29	
Control Test	0.081	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/15/2025 04:41:47

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.082	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.082	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 02/15/2025 23:36:19

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.082	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JEAN ANTOINE

DATE/TIME OF TEST: 02/16/2025 03:57:15

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.082	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/16/2025 19:54:28

Diagnostics Check	OK	19:47
Air Blank	0.000	19:48
Control Test	0.082	19:48
Air Blank	0.000	19:49
Subject Sample # 1	0.187	19:49
Breath Volume	3.347	19:49
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.186	19:52
Breath Volume	3.398	19:52
Air Blank	0.000	19:53
Control Test	0.079	19:53
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RODRIGUEZ IZZA

DATE/TIME OF TEST: 02/17/2025 03:03:25

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.000	02:58
Breath Volume	2.003	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:02
Breath Volume	2.226	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ZENNI

DATE/TIME OF TEST: 02/18/2025 04:41:06

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.175	04:36
Breath Volume	4.308	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.174	04:39
Breath Volume	2.687	04:39
Air Blank	0.000	04:40
Control Test	0.077	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: IUEN

DATE/TIME OF TEST: 02/22/2025 02:28:37

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.083	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: LONDON

DATE/TIME OF TEST: 02/22/2025 18:15:12

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.083	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.197	18:10
Breath Volume	1.625	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:12
Subject Sample # 2	0.200	18:13
Breath Volume	2.226	18:13
Air Blank	0.000	18:14
Control Test	0.079	18:14
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BAWI

DATE/TIME OF TEST: 02/22/2025 22:55:31

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.082	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.090	22:50
Breath Volume	1.675	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.089	22:53
Breath Volume	1.792	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRUBOR,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BARRINGTON

DATE/TIME OF TEST: 02/23/2025 03:06:17

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.082	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.181	03:01
Breath Volume	1.910	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.165	03:04
Breath Volume	1.476	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: TOOKES

DATE/TIME OF TEST: 02/23/2025 04:33:18

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.083	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:31	Subject Test Refused
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Control Test	0.083	04:32	
Air Blank	0.000	04:33	
Diagnostics Check	OK	04:33	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 02/24/2025 05:00:05

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.082	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:59	
Control Test	0.082	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	05:00	

Data Download Date/Time: 02/24/2025 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAUMGARTNER,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/01/2025 20:53:23

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.080	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.116	20:48
Breath Volume	3.218	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.112	20:51
Breath Volume	3.218	20:51
Air Blank	0.000	20:52
Control Test	0.079	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 02/04/2025 20:39:52

Diagnostics Check	OK	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:25	
Subject Sample	# 1 0.157	20:29	Volume Not Met
Breath Volume	1.089	20:29	
Air Blank	0.000	20:29	
Air Blank	0.000	20:31	
Subject Sample	# 2 0.154	20:33	
Breath Volume	1.511	20:33	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 3 0.158	20:38	
Breath Volume	2.023	20:38	
Air Blank	0.000	20:38	
Control Test	0.079	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 02/08/2025 03:55:33

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:51	No Sample Provided
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Air Blank	0.000	03:53	
Subject Sample	# 2	03:54	Subject Test Refused
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 02/09/2025 01:59:53

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:48	
Control Test	0.081	01:49	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:53	
Air Blank	0.000	01:55	
Subject Sample	# 2	01:58	No Sample Provided
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: SANDIFORD

DATE/TIME OF TEST: 02/10/2025 00:38:31

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.061	00:34
Breath Volume	3.394	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.060	00:37
Breath Volume	3.304	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 02/12/2025 21:41:25

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.079	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.193	21:36
Breath Volume	1.617	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:39
Subject Sample # 2	0.207	21:39
Breath Volume	1.222	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 02/14/2025 04:17:10

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.049	04:12
Breath Volume	1.285	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.047	04:15
Breath Volume	1.386	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HUMES,U
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: TREMBLAY

DATE/TIME OF TEST: 02/14/2025 22:18:39

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.080	22:08	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:11	No Sample Provided
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	
Air Blank	0.000	22:14	
Subject Sample	# 2	22:17	No Sample Provided
Breath Volume	0.000	22:17	
Air Blank	0.000	22:17	
Control Test	0.080	22:18	
Air Blank	0.000	22:18	
Diagnostics Check	OK	22:18	

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN, MATTHEW,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GUILLEN, A
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: PRITCHETT

DATE/TIME OF TEST: 02/15/2025 05:37:54

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:29	
Control Test	0.080	05:30	
Air Blank	0.000	05:30	
Subject Sample	# 1	05:33	No Sample Provided
Breath Volume	0.000	05:33	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample	# 2	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.080	05:37	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: FLOWE

DATE/TIME OF TEST: 02/16/2025 04:20:57

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.216	04:16
Breath Volume	2.871	04:16
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.206	04:19
Breath Volume	1.746	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BABCOCK, M
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: CLICK

DATE/TIME OF TEST: 02/18/2025 20:49:40

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.080	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.274	20:44
Breath Volume	1.453	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.264	20:48
Breath Volume	1.601	20:48
Air Blank	0.000	20:48
Control Test	0.079	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HEARD,T
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LUDLOFF

DATE/TIME OF TEST: 02/19/2025 00:14:50

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.187	00:10
Breath Volume	3.117	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.187	00:13
Breath Volume	2.289	00:13
Air Blank	0.000	00:13
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BLAKE, T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: WULFF

DATE/TIME OF TEST: 02/21/2025 20:39:03

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.256	20:34
Breath Volume	3.281	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.243	20:37
Breath Volume	2.371	20:37
Air Blank	0.000	20:38
Control Test	0.080	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:39

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOLENA, B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 02/23/2025 04:35:49

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.079	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.161	04:31
Breath Volume	1.964	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.165	04:34
Breath Volume	1.921	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 02/28/2025 06:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: BRAGG,RANDOLPH
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 01/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FORGATSCH

DATE/TIME OF TEST: 02/27/2025 02:31:04

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.166	02:26
Breath Volume	1.648	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.166	02:29
Breath Volume	1.621	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 03/10/2025 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: HAMMACK

DATE/TIME OF TEST: 02/01/2025 00:22:50

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.078	00:20	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 02/17/2025 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASSANTE,AUGUSTINE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PASSANTE,AUGUSTINE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/02/2025 03:22:05

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.136	03:17
Breath Volume	2.367	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.142	03:20
Breath Volume	2.589	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 02/17/2025 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: MONTIE,DAVID
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HUNT-MAYFIELD

DATE/TIME OF TEST: 02/22/2025 03:48:31

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.082	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1 0.221	03:35	
Breath Volume	1.101	03:35	
Air Blank	0.000	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.214	03:41	Volume Not Met
Breath Volume	0.960	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 3 0.200	03:46	Volume Not Met
Breath Volume	0.378	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 02/28/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: MONTIE,DAVID
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: HUNT-MAYFIELD

DATE/TIME OF TEST: 02/22/2025 04:20:29

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.203	04:15
Breath Volume	1.890	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample # 2	0.192	04:18
Breath Volume	1.433	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/28/2025 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ZACHARY,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: GISCHEL,MATTHEW
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 02/07/2025 23:28:03

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.083	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.251	23:20
Breath Volume	2.125	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.229	23:23
Breath Volume	1.855	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:25
Subject Sample # 3	0.243	23:26
Breath Volume	1.585	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:28

Data Download Date/Time: 02/25/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,EVA,
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: PARRISH,T
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 02/12/2025 16:06:12

Diagnostics Check	OK	16:00
Air Blank	0.000	16:00
Control Test	0.081	16:01
Air Blank	0.000	16:01
Subject Sample # 1	0.000	16:01
Breath Volume	2.027	16:01
Air Blank	0.000	16:02
Air Blank	0.000	16:04
Subject Sample # 2	0.000	16:04
Breath Volume	1.667	16:04
Air Blank	0.000	16:05
Control Test	0.080	16:05
Air Blank	0.000	16:06
Diagnostics Check	OK	16:06

Data Download Date/Time: 02/25/2025 03:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINCIN,WESLEY,P
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: MINCIN,WESLEY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 02/23/2025 19:25:56

Diagnostics Check	OK	19:22	
Air Blank	0.000	19:23	
Control Test	0.082	19:23	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:25	
Control Test	0.082	19:25	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 02/25/2025 03:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNEMANN,M,W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PACHECO-MARTINEZ,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: JR LLOYD

DATE/TIME OF TEST: 02/02/2025 03:34:18

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.052	03:29
Breath Volume	1.644	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.047	03:32
Breath Volume	1.738	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARBOLEDA,JUAN,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:32

SUBJECT'S LAST NAME: LIRIANO PINO

DATE/TIME OF TEST: 02/09/2025 14:58:40

Diagnostics Check	OK	14:52
Air Blank	0.000	14:53
Control Test	0.080	14:53
Air Blank	0.000	14:54
Subject Sample # 1	0.000	14:54
Breath Volume	3.195	14:54
Air Blank	0.000	14:54
Air Blank	0.000	14:56
Subject Sample # 2	0.000	14:57
Breath Volume	2.917	14:57
Air Blank	0.000	14:57
Control Test	0.080	14:58
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HALL,JORDAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/11/2025 21:27:45

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:25	
Control Test	0.080	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 1	21:26	RFI Detect
Breath Volume	0.000	21:26	
Air Blank	0.000	21:26	
Control Test	0.080	21:27	
Air Blank	0.000	21:27	
Diagnostics Check	OK	21:27	

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: HALL,JORDAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/11/2025 21:35:50

Diagnostics Check	OK	21:29
Air Blank	0.000	21:30
Control Test	0.081	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.000	21:31
Breath Volume	1.425	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:34
Breath Volume	1.726	21:34
Air Blank	0.000	21:34
Control Test	0.080	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: TRIGLIANOS-FROQUIHNO,TIM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ADKISON

DATE/TIME OF TEST: 02/12/2025 04:10:21

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.000	04:06
Breath Volume	2.410	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.000	04:08
Breath Volume	1.878	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: SPINELLI,A
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 02/12/2025 19:57:46

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.121	19:53
Breath Volume	1.824	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:55
Subject Sample # 2	0.118	19:56
Breath Volume	1.656	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LISS,K,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CHLEBOWSKI,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/13/2025 02:20:32

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:16
Breath Volume	5.183	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:19
Breath Volume	4.203	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 02/16/2025 19:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 02/09/2025 00:40:14

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 02/19/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NIKOLOV,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/09/2025 02:47:55

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.080	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 02/19/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/09/2025 22:48:56

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.296	22:44
Breath Volume	1.808	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.284	22:47
Breath Volume	1.164	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 02/19/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 02/15/2025 20:08:53

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Subject Sample	# 1	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:07	
Control Test	0.078	20:08	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 02/19/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RAY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: POWERS-BLANCO

DATE/TIME OF TEST: 02/02/2025 03:30:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.105	03:25
Breath Volume	2.371	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.102	03:29
Breath Volume	1.695	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 02/21/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BLAZER

DATE/TIME OF TEST: 02/11/2025 03:17:21

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.140	03:12
Breath Volume	3.390	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.136	03:15
Breath Volume	1.531	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 02/21/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/01/2025 00:03:15

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	2.980	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:01
Breath Volume	3.199	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 02/12/2025 02:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: FENTON

DATE/TIME OF TEST: 02/01/2025 00:18:09

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.080	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 02/12/2025 02:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUNARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 02/01/2025 03:02:38

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.080	02:58	
Air Blank	0.000	02:58	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 02/12/2025 02:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/01/2025 03:13:56

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.244	03:09
Breath Volume	1.617	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.243	03:12
Breath Volume	1.457	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 02/12/2025 02:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,C,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 02/01/2025 21:38:28

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.174	21:33
Breath Volume	3.269	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.174	21:36
Breath Volume	3.218	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 02/12/2025 02:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,C,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: CLIMES

DATE/TIME OF TEST: 02/01/2025 22:44:26

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.000	22:40
Breath Volume	2.703	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.000	22:43
Breath Volume	2.859	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GALLARDO RODRIGUEZ

DATE/TIME OF TEST: 02/02/2025 02:28:53

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.000	02:24
Breath Volume	1.320	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:27
Breath Volume	1.503	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: IOMELLI,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/03/2025 02:23:53

Diagnostics Check	OK	02:19	
Air Blank	0.000	02:20	
Control Test	0.080	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:22	RFI Detect
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: IOMELLI,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/03/2025 02:36:58

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:26	
Control Test	0.080	02:26	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:30	No Sample Provided
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Air Blank	0.000	02:32	
Subject Sample	# 2	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MILLER,MORGAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: AL RAWAHNEH

DATE/TIME OF TEST: 02/05/2025 02:43:53

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.282	02:38
Breath Volume	2.250	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.295	02:42
Breath Volume	1.250	02:42
Air Blank	0.000	02:42
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: SANDLER

DATE/TIME OF TEST: 02/05/2025 08:52:24

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.080	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.000	08:48
Breath Volume	2.078	08:48
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.000	08:51
Breath Volume	1.777	08:51
Air Blank	0.000	08:51
Control Test	0.080	08:51
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DARLING

DATE/TIME OF TEST: 02/06/2025 02:25:27

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.272	02:20
Breath Volume	1.808	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.260	02:23
Breath Volume	1.527	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: RUBIANO

DATE/TIME OF TEST: 02/06/2025 03:21:14

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:18	
Control Test	0.080	03:18	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.080	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: RANDLE

DATE/TIME OF TEST: 02/06/2025 05:11:47

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:58	
Control Test	0.080	04:59	
Air Blank	0.000	04:59	
Subject Sample	# 1	05:02	No Sample Provided
Breath Volume	0.000	05:02	
Air Blank	0.000	05:03	
Air Blank	0.000	05:05	
Subject Sample	# 2	0.161	05:06
Breath Volume	2.160	05:06	
Air Blank	0.000	05:06	
Air Blank	0.000	05:08	
Subject Sample	# 3	0.158	05:10
Breath Volume	1.347	05:10	
Air Blank	0.000	05:10	
Control Test	0.080	05:11	
Air Blank	0.000	05:11	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALI,B
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BONEL

DATE/TIME OF TEST: 02/08/2025 01:43:52

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:40	
Control Test	0.081	01:40	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RIMAS

DATE/TIME OF TEST: 02/08/2025 03:16:48

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: VASHALIFSKI

DATE/TIME OF TEST: 02/11/2025 20:54:30

Diagnostics Check	OK	20:47
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.319	20:49
Breath Volume	2.371	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.329	20:52
Breath Volume	1.148	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 02/12/2025 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 01/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VIRGEN

DATE/TIME OF TEST: 02/08/2025 02:50:49

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:49	Subject Test Refused
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 02/13/2025 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 01/20/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 02:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MARROQUIN

DATE/TIME OF TEST: 02/08/2025 03:02:45

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.081	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Control Test	0.081	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 02/13/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 16:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: MOTT

DATE/TIME OF TEST: 02/14/2025 18:32:56

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:30	
Control Test	0.081	18:30	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:31	Subject Test Refused
Breath Volume	0.000	18:31	
Air Blank	0.000	18:32	
Control Test	0.081	18:32	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: TESSIER

DATE/TIME OF TEST: 02/15/2025 20:00:45

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.081	19:55
Air Blank	0.000	19:55
Subject Sample # 1	0.041	19:56
Breath Volume	1.972	19:56
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.041	19:59
Breath Volume	2.242	19:59
Air Blank	0.000	19:59
Control Test	0.080	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HANNON

DATE/TIME OF TEST: 02/16/2025 00:22:34

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: HAGGANS

DATE/TIME OF TEST: 02/19/2025 23:51:36

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:50	Subject Test Refused
Breath Volume	0.000	23:50	
Air Blank	0.000	23:50	
Control Test	0.079	23:51	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SILVERSTEIN,CHRISTINA

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 22:07

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: KINNEY

DATE/TIME OF TEST: 02/20/2025 22:57:15

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.081	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:55	Subject Test Refused
Breath Volume	0.000	22:55	
Air Blank	0.000	22:56	
Control Test	0.081	22:56	
Air Blank	0.000	22:57	
Diagnostics Check	OK	22:57	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 02/22/2025 00:43:55

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 23:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: RIDLER

DATE/TIME OF TEST: 02/23/2025 00:07:21

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MCNULTY

DATE/TIME OF TEST: 02/28/2025 01:21:04

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.080	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.064	01:16
Breath Volume	2.273	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.061	01:19
Breath Volume	2.191	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:21

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/01/2025 00:40:05

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.021	00:35
Breath Volume	1.785	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.020	00:38
Breath Volume	1.582	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/11/2025 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/01/2025 01:42:18

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2	01:40	No Sample Provided
Breath Volume	0.000	01:40	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 02/11/2025 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,EDSON,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHAVEZ,EDSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 02/01/2025 21:31:26

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:30	Sequence Aborted
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	

Data Download Date/Time: 02/11/2025 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,EDSON,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHAVEZ,EDSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 02/01/2025 21:49:32

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.079	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:46	No Sample Provided
Breath Volume	0.000	21:46	
Air Blank	0.000	21:47	
Air Blank		21:49	Sequence Aborted
Air Blank	0.000	21:49	

Data Download Date/Time: 02/11/2025 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 02/01/2025 22:52:40

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.144	22:48
Breath Volume	1.695	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.136	22:51
Breath Volume	1.734	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 02/11/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,HUNTER,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 02/01/2025 23:46:51

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.080	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:39	No Sample Provided
Breath Volume	0.000	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2	23:45	No Sample Provided
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.079	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 02/11/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,KEVIN,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLE,KEVIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 02/03/2025 15:48:20

Diagnostics Check	OK	15:42
Air Blank	0.000	15:43
Control Test	0.081	15:43
Air Blank	0.000	15:43
Subject Sample # 1	0.000	15:44
Breath Volume	1.312	15:44
Air Blank	0.000	15:44
Air Blank	0.000	15:46
Subject Sample # 2	0.000	15:46
Breath Volume	1.828	15:46
Air Blank	0.000	15:47
Control Test	0.079	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 02/11/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE, BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 02/04/2025 03:49:07

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:37	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.155	03:40	Volume Not Met
Breath Volume	0.523	03:40	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.191	03:43	
Breath Volume	1.589	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 3 0.191	03:47	
Breath Volume	1.507	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 02/11/2025 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SERGIO

DATE/TIME OF TEST: 02/01/2025 01:37:22

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.000	01:33
Breath Volume	2.839	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.000	01:36
Breath Volume	1.953	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 02/21/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/05/2025 01:36:24

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.081	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.000	01:31
Breath Volume	1.753	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:35
Breath Volume	1.289	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 02/21/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 02/05/2025 02:47:57

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.166	02:43
Breath Volume	2.035	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.173	02:46
Breath Volume	1.492	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 02/21/2025 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER, ROBERT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: P417780640000

DATE/TIME OF TEST: 02/06/2025 05:53:27

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.080	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.000	05:48
Breath Volume	3.199	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.000	05:52
Breath Volume	2.515	05:52
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 02/21/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: TURKOVICS

DATE/TIME OF TEST: 02/08/2025 06:53:02

Diagnostics Check	OK	06:46
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.117	06:48
Breath Volume	1.554	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:50
Subject Sample # 2	0.104	06:51
Breath Volume	1.144	06:51
Air Blank	0.000	06:52
Control Test	0.079	06:52
Air Blank	0.000	06:52
Diagnostics Check	OK	06:53

Data Download Date/Time: 02/21/2025 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOMEZ,DESTINEE,R

BREATH TEST OPERATOR AGENCY: Plantation PD

ARRESTING OFFICER: GOMEZ,DESTINEE

ARRESTING OFFICER AGENCY: Plantation PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231603

DRY GAS EXPIRY DATE: 03/05/2026

VIOLATION CODE: DUI

ARREST TIME: 02:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HOMS

DATE/TIME OF TEST: 02/14/2025 03:43:05

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:31	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.138	03:34	Volume Not Met
Breath Volume	0.710	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 2 0.136	03:38	
Breath Volume	1.285	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3 0.140	03:41	
Breath Volume	1.832	03:41	
Air Blank	0.000	03:42	
Control Test	0.077	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 02/28/2025 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 02/04/2025 23:16:37

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.082	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1 0.126	23:08	Volume Not Met
Breath Volume	0.378	23:08	
Air Blank	0.000	23:09	
Air Blank	0.000	23:11	
Subject Sample	# 2 0.148	23:11	
Breath Volume	1.414	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 3 0.151	23:15	
Breath Volume	2.187	23:15	
Air Blank	0.000	23:15	
Control Test	0.082	23:16	
Air Blank	0.000	23:16	
Diagnostics Check	OK	23:16	

Data Download Date/Time: 02/05/2025 02:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BOER

DATE/TIME OF TEST: 02/19/2025 16:35:52

Diagnostics Check	OK	16:30	
Air Blank	0.000	16:30	
Control Test	0.081	16:31	
Air Blank	0.000	16:31	
Subject Sample # 1	0.127	16:32	
Breath Volume	2.554	16:32	
Air Blank	0.000	16:32	
Air Blank	0.000	16:34	
Subject Sample # 2	0.121	16:35	
Breath Volume	2.855	16:35	
Air Blank		16:35	RFI Detect
Air Blank	0.000	16:35	

Data Download Date/Time: 03/02/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BOER

DATE/TIME OF TEST: 02/19/2025 16:45:16

Diagnostics Check	OK	16:39
Air Blank	0.000	16:39
Control Test	0.082	16:39
Air Blank	0.000	16:40
Subject Sample # 1	0.121	16:40
Breath Volume	2.718	16:40
Air Blank	0.000	16:41
Air Blank	0.000	16:43
Subject Sample # 2	0.119	16:43
Breath Volume	2.804	16:43
Air Blank	0.000	16:44
Control Test	0.083	16:44
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 03/02/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LEMONS

DATE/TIME OF TEST: 02/05/2025 03:29:19

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.197	03:19	Slope Not Level
Breath Volume	1.476	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2 0.210	03:22	
Breath Volume	1.828	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 3 0.191	03:27	
Breath Volume	1.925	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 02/28/2025 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHADL,CAMERON,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SCHADL,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 02/14/2025 23:22:41

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Subject Sample	# 1 0.121	23:11	Volume Not Met
Breath Volume	0.316	23:11	
Air Blank	0.000	23:11	
Air Blank	0.000	23:13	
Subject Sample	# 2 0.153	23:15	
Breath Volume	1.195	23:15	
Air Blank	0.000	23:16	
Air Blank	0.000	23:18	
Subject Sample	# 3	23:21	Subject Test Refused
Breath Volume	0.015	23:21	
Air Blank	0.000	23:21	
Control Test	0.080	23:22	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 02/28/2025 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 02/16/2025 02:14:56

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.081	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.243	02:10
Breath Volume	1.777	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:12
Subject Sample # 2	0.233	02:13
Breath Volume	2.742	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 02/28/2025 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: ICHICH CUCUL

DATE/TIME OF TEST: 02/23/2025 17:00:20

Diagnostics Check	OK	16:53
Air Blank	0.000	16:54
Control Test	0.081	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.153	16:55
Breath Volume	2.003	16:55
Air Blank	0.000	16:56
Air Blank	0.000	16:58
Subject Sample # 2	0.146	16:58
Breath Volume	1.671	16:58
Air Blank	0.000	16:59
Control Test	0.080	16:59
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 02/28/2025 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUMONT,TRAVIS,H
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: GOMEZ, RAMON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/01/2025 02:54:24

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.176	02:49
Breath Volume	1.136	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.182	02:52
Breath Volume	1.523	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SIAS,MARK,D

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: ABERLE,AJ

ARRESTING OFFICER AGENCY: Volusia County SO

LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/12/2026

VIOLATION CODE: DUI

ARREST TIME: 06:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: SOMERVILLE

DATE/TIME OF TEST: 02/01/2025 07:34:21

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.081	07:28
Air Blank	0.000	07:29
Subject Sample # 1	0.165	07:29
Breath Volume	2.730	07:29
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.158	07:32
Breath Volume	1.949	07:32
Air Blank	0.000	07:33
Control Test	0.081	07:33
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEDINGER,RYAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 02/10/2025 00:30:26

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.081	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.193	00:25
Breath Volume	2.011	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.185	00:28
Breath Volume	2.273	00:28
Air Blank	0.000	00:29
Control Test	0.082	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEDINGER,RYAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: LOVELESS,P
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 02/10/2025 01:53:15

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.157	01:48
Breath Volume	4.046	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.149	01:51
Breath Volume	3.359	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWERS,ABIGAYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: POWERS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/17/2025 00:55:18

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.082	00:52	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEDINGER,R,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: MERHIB,M
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: DUNTON JR

DATE/TIME OF TEST: 02/23/2025 00:52:18

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.111	00:47
Breath Volume	3.296	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.093	00:50
Breath Volume	1.851	00:50
Air Blank	0.000	00:51
Control Test	0.081	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 02/28/2025 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,ANDER,P
BREATH TEST OPERATOR AGENCY: Bonifay PD
ARRESTING OFFICER: SULLIVAN,ANDER
ARRESTING OFFICER AGENCY: Bonifay PD
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: CRISWELL

DATE/TIME OF TEST: 02/09/2025 02:08:13

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.079	01:57	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.129	02:00	Volume Not Met
Breath Volume	0.796	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:03	
Subject Sample	# 2 0.160	02:03	
Breath Volume	1.632	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 3 0.150	02:06	
Breath Volume	2.371	02:06	
Air Blank	0.000	02:07	
Control Test	0.078	02:07	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 03/09/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMPSON,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: VAUGHN,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: BERSCHE

DATE/TIME OF TEST: 02/07/2025 21:26:15

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.082	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.234	21:21
Breath Volume	3.007	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.230	21:24
Breath Volume	1.882	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 02/26/2025 12:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRIZARRY GONZALEZ,MIRIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BIFANO,JASMINE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: IRIZARRY GONZALEZ

DATE/TIME OF TEST: 02/18/2025 17:04:18

Diagnostics Check	OK	16:57
Air Blank	0.000	16:57
Control Test	0.081	16:58
Air Blank	0.000	16:58
Subject Sample # 1	0.339	16:59
Breath Volume	1.582	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.355	17:02
Breath Volume	2.250	17:02
Air Blank	0.000	17:03
Control Test	0.079	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 02/26/2025 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUDAY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RICKERSON,KENTON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 13:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/20/2025 14:12:28

Diagnostics Check	OK	14:05
Air Blank	0.000	14:06
Control Test	0.081	14:06
Air Blank	0.000	14:07
Subject Sample # 1	0.088	14:08
Breath Volume	3.359	14:08
Air Blank	0.000	14:08
Air Blank	0.000	14:10
Subject Sample # 2	0.081	14:10
Breath Volume	2.917	14:10
Air Blank	0.000	14:11
Control Test	0.081	14:11
Air Blank	0.000	14:12
Diagnostics Check	OK	14:12

Data Download Date/Time: 02/26/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MATA-ORTIZ

DATE/TIME OF TEST: 02/01/2025 04:35:56

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.173	04:31
Breath Volume	1.535	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.171	04:34
Breath Volume	1.261	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS, WILLIAM
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/01/2025 05:29:16

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.038	05:24
Breath Volume	1.132	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.037	05:27
Breath Volume	1.222	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,EDSON,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHAVEZ,EDSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 02/01/2025 22:00:31

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:56
Breath Volume	4.000	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:59
Breath Volume	2.875	21:59
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MORELLO

DATE/TIME OF TEST: 02/01/2025 22:39:45

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:33
Air Blank	0.000	22:34
Subject Sample # 1	0.157	22:35
Breath Volume	1.718	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.156	22:38
Breath Volume	2.316	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,HUNTER,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOTO,MAVERICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 02/01/2025 23:59:59

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.078	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.178	23:54
Breath Volume	2.253	23:54
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.174	23:58
Breath Volume	1.218	23:58
Air Blank	0.000	23:59
Control Test	0.078	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FARRER

DATE/TIME OF TEST: 02/02/2025 03:39:01

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.202	03:34
Breath Volume	2.683	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.194	03:37
Breath Volume	2.593	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:39

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/02/2025 04:18:26

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.096	04:14
Breath Volume	1.492	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.102	04:16
Breath Volume	1.277	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELBER,RYAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: LARISON

DATE/TIME OF TEST: 02/02/2025 04:47:04

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.090	04:42
Breath Volume	1.738	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.091	04:45
Breath Volume	1.496	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 02/02/2025 19:47:33

Diagnostics Check	OK	19:40
Air Blank	0.000	19:41
Control Test	0.080	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.098	19:42
Breath Volume	1.382	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.099	19:46
Breath Volume	2.878	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELABRA,HUGO,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MOSTAFA,RYAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: DELABRA

DATE/TIME OF TEST: 02/05/2025 01:08:08

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:01	No Sample Provided
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 2	01:06	No Sample Provided
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.081	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KALEN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MOSTAFA,RYAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: DELABRA

DATE/TIME OF TEST: 02/05/2025 01:31:13

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:20	
Control Test	0.081	01:20	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 02/05/2025 20:08:13

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.055	20:03
Breath Volume	1.167	20:03
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.056	20:06
Breath Volume	1.226	20:06
Air Blank	0.000	20:07
Control Test	0.080	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 02/05/2025 22:46:41

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:35	
Control Test	0.081	22:35	
Air Blank	0.000	22:36	
Subject Sample	# 1	22:39	No Sample Provided
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Air Blank	0.000	22:41	
Subject Sample	# 2	0.183	22:42
Breath Volume	2.000	22:42	
Air Blank	0.000	22:42	
Air Blank	0.000	22:44	
Subject Sample	# 3	0.173	22:45
Breath Volume	1.792	22:45	
Air Blank	0.000	22:45	
Control Test	0.078	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ROONEY JR

DATE/TIME OF TEST: 02/06/2025 23:42:41

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.081	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.166	23:35	Volume Not Met
Breath Volume	0.031	23:35	
Air Blank	0.000	23:36	
Air Blank	0.000	23:38	
Subject Sample	# 2 0.150	23:41	Volume Not Met
Breath Volume	0.210	23:41	
Air Blank	0.000	23:41	
Control Test	0.079	23:42	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,HUNTER,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SAUNDERS,HUNTER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 02/08/2025 00:50:30

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.118	00:46
Breath Volume	3.648	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.119	00:49
Breath Volume	3.417	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/09/2025 02:45:31

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:41
Breath Volume	1.414	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	2.218	02:44
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/11/2025 08:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KALEN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FARLESS,MITCHELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/09/2025 04:20:28

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.126	04:15
Breath Volume	1.898	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.126	04:19
Breath Volume	2.304	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 02/11/2025 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KALEN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PERALTA,ALLISON
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 02/09/2025 04:40:00

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	0.124	04:34
Breath Volume	2.117	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 3	0.120	04:38
Breath Volume	1.558	04:38	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 02/11/2025 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 02/09/2025 19:22:54

Diagnostics Check	OK	19:16
Air Blank	0.000	19:17
Control Test	0.081	19:17
Air Blank	0.000	19:18
Subject Sample # 1	0.129	19:18
Breath Volume	1.398	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.128	19:21
Breath Volume	1.664	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 02/11/2025 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,PAUL,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BYRAMS,JADA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 02/10/2025 22:51:57

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.128	22:46
Breath Volume	1.792	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.124	22:50
Breath Volume	1.289	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 02/11/2025 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LYNDON

DATE/TIME OF TEST: 02/06/2025 00:37:45

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.159	00:33
Breath Volume	3.308	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.159	00:36
Breath Volume	3.406	00:36
Air Blank	0.000	00:36
Control Test	0.079	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 02/27/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGLE,SHAWN,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SALZL,GUENTHER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: FEARON

DATE/TIME OF TEST: 02/09/2025 22:02:11

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.000	21:57
Breath Volume	1.785	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.000	22:00
Breath Volume	2.105	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 02/27/2025 18:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PONDER,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GALIVER BENOIT

DATE/TIME OF TEST: 02/04/2025 04:55:48

Diagnostics Check	OK	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.202	04:50
Breath Volume	1.750	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.201	04:54
Breath Volume	1.558	04:54
Air Blank	0.000	04:54
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 02/27/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/06/2025 23:56:02

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.000	23:51
Breath Volume	1.671	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:54
Breath Volume	1.941	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:56

Data Download Date/Time: 02/27/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MADRAK

DATE/TIME OF TEST: 02/09/2025 00:27:11

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.080	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.130	00:22
Breath Volume	2.675	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.136	00:25
Breath Volume	2.437	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 02/27/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: GOMEZ GARCIA

DATE/TIME OF TEST: 02/09/2025 01:09:44

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.090	01:04
Breath Volume	2.496	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.095	01:08
Breath Volume	1.355	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 02/27/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: MELNIK

DATE/TIME OF TEST: 02/20/2025 23:08:25

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 02/27/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RICAFORT,KEANA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: FABIANSKI

DATE/TIME OF TEST: 02/26/2025 04:16:10

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.045	04:11
Breath Volume	2.523	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.046	04:14
Breath Volume	2.894	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 02/27/2025 18:50