

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CUMMING,WILLIAM
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/01/2025 02:16:16

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.214	02:11
Breath Volume	2.734	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.211	02:14
Breath Volume	2.625	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FORDE

DATE/TIME OF TEST: 02/01/2025 02:31:33

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.143	02:27
Breath Volume	2.312	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.144	02:30
Breath Volume	1.976	02:30
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CRICI

DATE/TIME OF TEST: 02/01/2025 20:36:50

Diagnostics Check	OK	20:30
Air Blank	0.000	20:30
Control Test	0.079	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.072	20:32
Breath Volume	2.355	20:32
Air Blank	0.000	20:32
Air Blank	0.000	20:34
Subject Sample # 2	0.071	20:35
Breath Volume	1.367	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: KUPLIN

DATE/TIME OF TEST: 02/01/2025 23:57:20

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.157	23:52
Breath Volume	3.914	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.150	23:55
Breath Volume	2.746	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/02/2025 00:30:41

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.083	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.085	00:26
Breath Volume	1.355	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.084	00:29
Breath Volume	1.710	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 02/02/2025 01:53:19

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.153	01:48
Breath Volume	3.035	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.142	01:51
Breath Volume	1.933	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HEYWARD

DATE/TIME OF TEST: 02/02/2025 02:39:49

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.100	02:34
Breath Volume	1.398	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.099	02:38
Breath Volume	1.375	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GRUBBS

DATE/TIME OF TEST: 02/06/2025 22:53:12

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.032	22:48
Breath Volume	2.933	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.031	22:51
Breath Volume	2.332	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GOWIN

DATE/TIME OF TEST: 02/07/2025 00:06:15

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:02
Breath Volume	1.296	00:02
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:04
Breath Volume	1.660	00:04
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SPARROW

DATE/TIME OF TEST: 02/07/2025 01:21:14

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.246	01:16
Breath Volume	1.468	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.240	01:19
Breath Volume	2.164	01:19
Air Blank	0.000	01:20
Control Test	0.079	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CLARY

DATE/TIME OF TEST: 02/08/2025 02:45:42

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.046	02:41
Breath Volume	1.816	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.045	02:44
Breath Volume	2.414	02:44
Air Blank	0.000	02:44
Control Test	0.079	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 02/14/2025 01:51:46

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.237	01:47
Breath Volume	1.910	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.244	01:50
Breath Volume	2.476	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: NICODEMUS

DATE/TIME OF TEST: 02/21/2025 02:17:28

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.147	02:12
Breath Volume	1.476	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.150	02:15
Breath Volume	1.324	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GOSLIN,TYLER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: ZEILMAN

DATE/TIME OF TEST: 02/26/2025 01:00:43

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.162	00:55
Breath Volume	2.695	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.143	00:59
Breath Volume	1.203	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 02/26/2025 19:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SCHERR, EVAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 02/01/2025 02:54:18

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.077	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.223	02:49
Breath Volume	1.238	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.213	02:52
Breath Volume	1.570	02:52
Air Blank	0.000	02:53
Control Test	0.077	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,J,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHINN,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: OBERLE

DATE/TIME OF TEST: 02/01/2025 03:22:03

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPIASI,JAKE,L
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SCHMIDT,JONATHAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: AHONKHAI

DATE/TIME OF TEST: 02/02/2025 00:31:53

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.153	00:27
Breath Volume	3.035	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.149	00:30
Breath Volume	3.058	00:30
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERICKSON,NATHAN,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MEDINA,JACQULYN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/02/2025 14:29:47

Diagnostics Check	OK	14:18	
Air Blank	0.000	14:18	
Control Test	0.080	14:18	
Air Blank	0.000	14:19	
Subject Sample	# 1 0.163	14:22	Volume Not Met
Breath Volume	0.164	14:22	
Air Blank	0.000	14:23	
Air Blank	0.000	14:25	
Subject Sample	# 2 0.138	14:28	Volume Not Met
Breath Volume	0.050	14:28	
Air Blank	0.000	14:28	
Control Test	0.079	14:29	
Air Blank	0.000	14:29	
Diagnostics Check	OK	14:29	

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FANSLER,M
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 02/04/2025 23:01:13

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.111	22:56
Breath Volume	1.589	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.109	22:59
Breath Volume	1.980	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 02/06/2025 21:09:12

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:06	
Control Test	0.079	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LULU,ABD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: DAWSON-ANDRZEJCZYK

DATE/TIME OF TEST: 02/06/2025 21:47:25

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.270	21:42
Breath Volume	2.152	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.268	21:45
Breath Volume	2.484	21:45
Air Blank	0.000	21:46
Control Test	0.077	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MARSAN,NECOLE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: G625432793490

DATE/TIME OF TEST: 02/07/2025 21:33:17

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:25	
Control Test	0.079	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.106	21:27	
Breath Volume	1.480	21:27	
Air Blank	0.000	21:28	
Air Blank	0.000	21:30	
Subject Sample	# 2	21:31	Subject Test Refused
Breath Volume	0.113	21:31	
Air Blank	0.000	21:32	
Control Test	0.079	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SAUNDERS,JOHNNY
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BOYLE

DATE/TIME OF TEST: 02/07/2025 22:01:34

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.194	21:57
Breath Volume	2.101	21:57
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.188	22:00
Breath Volume	1.406	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:00
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 02/12/2025 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARSAN,NECOLE,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FANSLER,MICHEAL
ARRESTING OFFICER AGENCY: Bunnell PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: REEVES-JOYNES

DATE/TIME OF TEST: 02/07/2025 22:51:08

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.090	22:46
Breath Volume	1.808	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.093	22:49
Breath Volume	1.875	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 02/12/2025 08:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFMAN,DYLAN,R
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: COFFMAN,DYLAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: BROWNFIELD

DATE/TIME OF TEST: 02/09/2025 18:44:47

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.079	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.187	18:40
Breath Volume	2.164	18:40
Air Blank	0.000	18:41
Air Blank	0.000	18:42
Subject Sample # 2	0.184	18:43
Breath Volume	1.816	18:43
Air Blank	0.000	18:43
Control Test	0.080	18:44
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 02/12/2025 08:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 02/01/2025 02:55:55

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.080	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.173	02:45	Volume Not Met
Breath Volume	0.140	02:45	
Air Blank	0.000	02:46	
Air Blank	0.000	02:48	
Subject Sample	# 2 0.150	02:49	
Breath Volume	1.566	02:49	
Air Blank	0.000	02:50	
Air Blank	0.000	02:52	
Subject Sample	# 3 0.157	02:54	
Breath Volume	1.527	02:54	
Air Blank	0.000	02:54	
Control Test	0.078	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 02/07/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GRECCO

DATE/TIME OF TEST: 02/02/2025 01:51:14

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.164	01:46
Breath Volume	1.261	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:48
Subject Sample # 2	0.153	01:49
Breath Volume	1.277	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 02/07/2025 09:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: LORENTZEN

DATE/TIME OF TEST: 02/07/2025 01:33:35

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.145	01:29
Breath Volume	1.914	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.146	01:31
Breath Volume	1.617	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 02/07/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: HEPPE

DATE/TIME OF TEST: 02/07/2025 21:14:59

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.080	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.213	21:10
Breath Volume	2.167	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample # 2	0.213	21:13
Breath Volume	1.687	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,R,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: CASTRO,J
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: PABON VELANDIA

DATE/TIME OF TEST: 02/08/2025 13:38:19

Diagnostics Check	OK	13:31
Air Blank	0.000	13:31
Control Test	0.082	13:31
Air Blank	0.000	13:32
Subject Sample # 1	0.228	13:33
Breath Volume	2.664	13:33
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.237	13:36
Breath Volume	2.632	13:36
Air Blank	0.000	13:37
Control Test	0.078	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKEN,CHELSEA,N
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: BIRMINGHAM

DATE/TIME OF TEST: 02/08/2025 16:57:29

Diagnostics Check	OK	16:49
Air Blank	0.000	16:49
Control Test	0.080	16:50
Air Blank	0.000	16:50
Subject Sample # 1	0.233	16:51
Breath Volume	1.433	16:51
Air Blank	0.000	16:52
Air Blank	0.000	16:53
Subject Sample # 2	0.223	16:55
Breath Volume	2.035	16:55
Air Blank	0.000	16:56
Control Test	0.080	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: L330676026700

DATE/TIME OF TEST: 02/09/2025 02:47:58

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.161	02:42
Breath Volume	1.753	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.152	02:46
Breath Volume	1.507	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VERYCKEN,MATTHEW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/10/2025 00:41:15

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.154	00:36
Breath Volume	2.050	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.148	00:39
Breath Volume	1.562	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 02/12/2025 02:33:03

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LOPEZ MENJIVAR

DATE/TIME OF TEST: 02/17/2025 04:16:45

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.137	04:11
Breath Volume	3.234	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.132	04:15
Breath Volume	1.902	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MERRELL

DATE/TIME OF TEST: 02/17/2025 04:41:56

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.000	04:37
Breath Volume	1.640	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.000	04:40
Breath Volume	1.515	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/07/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ROGALSKI

DATE/TIME OF TEST: 02/21/2025 00:08:40

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:06	
Control Test	0.081	00:06	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:07	Subject Test Refused
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 03/07/2025 14:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,BRANDEN,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: THOMAS,BRANDEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: SARGENT

DATE/TIME OF TEST: 02/21/2025 05:46:12

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.081	05:44	
Air Blank	0.000	05:44	
Subject Sample	# 1	05:44	Subject Test Refused
Breath Volume	0.000	05:44	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:46	
Diagnostics Check	OK	05:46	

Data Download Date/Time: 03/07/2025 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 02/28/2025 02:00:32

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
		02:00	RFI Detect
Air Blank	0.000	02:00	

Data Download Date/Time: 03/07/2025 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 02/28/2025 02:03:38

Diagnostics Check	OK	02:02	
Air Blank		02:03	RFI Detect
Air Blank	0.000	02:03	

Data Download Date/Time: 03/07/2025 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES GONZALEZ,PEDRO,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES GONZALEZ,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 02/28/2025 02:09:27

Diagnostics Check	Fail	02:08	Internal Printer Test
Air Blank	0.000	02:09	

Data Download Date/Time: 03/07/2025 14:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: FLANAGAN, ROY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 02/23/2025 01:54:34

Diagnostics Check	Fail	01:54	Voltage/Current Test
Air Blank	0.000	01:54	

Data Download Date/Time: 02/28/2025 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES, LANDON, A
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: FLANAGAN, ROY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 02/23/2025 02:04:57

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.075	01:59
Breath Volume	1.515	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.078	02:03
Breath Volume	1.242	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 02/28/2025 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BIDDLE,MICAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: STROMAN

DATE/TIME OF TEST: 02/01/2025 06:32:10

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.160	06:27
Breath Volume	1.988	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.161	06:30
Breath Volume	1.750	06:30
Air Blank	0.000	06:31
Control Test	0.081	06:31
Air Blank	0.000	06:32
Diagnostics Check	OK	06:32

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HARLEY

DATE/TIME OF TEST: 02/02/2025 00:43:07

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.082	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.176	00:35	Volume Not Met
Breath Volume	1.050	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 2 0.172	00:41	Volume Not Met
Breath Volume	0.746	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 02/07/2025 04:40:51

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.272	04:36
Breath Volume	1.242	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.286	04:39
Breath Volume	1.242	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: WEINTRAUB

DATE/TIME OF TEST: 02/08/2025 02:01:09

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1 0.160	01:54	Volume Not Met
Breath Volume	0.488	01:54	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2 0.170	01:59	Volume Not Met
Breath Volume	0.828	01:59	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WEINTRAUB

DATE/TIME OF TEST: 02/08/2025 02:37:54

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.200	02:33
Breath Volume	1.175	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.186	02:36
Breath Volume	1.449	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LOPEZ,J
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 02/08/2025 19:56:08

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.081	19:45	
Air Blank	0.000	19:45	
Subject Sample	# 1 0.158	19:48	Slope Not Level
Breath Volume	1.730	19:48	
Air Blank	0.000	19:49	
Air Blank	0.000	19:51	
Subject Sample	# 2 0.150	19:54	Volume Not Met
Breath Volume	0.015	19:54	
Air Blank	0.000	19:55	
Control Test	0.081	19:55	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LOPEZ,J
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 02/08/2025 20:11:00

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:02	
Control Test	0.079	20:02	
Air Blank	0.000	20:03	
Subject Sample	# 1 0.157	20:06	Volume Not Met
Breath Volume	0.015	20:06	
Air Blank	0.000	20:06	
Air Blank	0.000	20:08	
Subject Sample	# 2	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.080	20:09	
Air Blank		20:10	Purge Fail
Air Blank	0.000	20:11	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 02/09/2025 04:13:05

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.162	04:00	
Breath Volume	1.234	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2 0.152	04:05	Volume Not Met
Breath Volume	0.914	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 3 0.139	04:11	Volume Not Met
Breath Volume	0.265	04:11	
Air Blank	0.000	04:12	
Control Test	0.080	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GRIMES

DATE/TIME OF TEST: 02/09/2025 04:46:51

Diagnostics Check	OK	04:32	
Air Blank	0.000	04:33	
Control Test	0.080	04:33	
Air Blank	0.000	04:33	
Subject Sample	# 1 0.145	04:35	
Breath Volume	1.214	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2 0.137	04:41	Volume Not Met
Breath Volume	1.093	04:41	
Air Blank	0.000	04:41	
Air Blank	0.000	04:43	
Subject Sample	# 3 0.139	04:45	
Breath Volume	1.316	04:45	
Air Blank	0.000	04:45	
Control Test	0.077	04:46	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 02/10/2025 01:42:30

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.000	01:38
Breath Volume	2.878	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.000	01:41
Breath Volume	3.031	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 02/11/2025 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 02/27/2025 01:32:09

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.037	01:27
Breath Volume	2.636	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.038	01:30
Breath Volume	1.601	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 02/27/2025 19:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,CHAD,T
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: JIMENEZ,B
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: RAMPERSAD

DATE/TIME OF TEST: 02/06/2025 15:28:54

Diagnostics Check	OK	15:22
Air Blank	0.000	15:22
Control Test	0.081	15:22
Air Blank	0.000	15:23
Subject Sample # 1	0.000	15:24
Breath Volume	1.343	15:24
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.000	15:27
Breath Volume	1.890	15:27
Air Blank	0.000	15:27
Control Test	0.080	15:28
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: LATIMER,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 19:17

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 02/12/2025 20:07:05

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.080	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:05	Subject Test Refused
Breath Volume	0.000	20:05	
Air Blank	0.000	20:06	
Control Test	0.080	20:06	
Air Blank	0.000	20:07	
Diagnostics Check	OK	20:07	

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REBANDO,BRIAN,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: MACAVOY,JASON
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 02/15/2025 00:52:30

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.254	00:47
Breath Volume	2.320	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.254	00:50
Breath Volume	2.074	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REASONER,D,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: MENCARELLI,DAVE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/16/2025 11:52:56

Diagnostics Check	OK	11:40	
Air Blank	0.000	11:40	
Control Test	0.080	11:40	
Air Blank	0.000	11:41	
Subject Sample	# 1 0.262	11:41	
Breath Volume	1.605	11:41	
Air Blank	0.000	11:42	
Air Blank	0.000	11:44	
Subject Sample	# 2 0.200	11:47	Volume Not Met
Breath Volume	0.335	11:47	
Air Blank	0.000	11:48	
Air Blank	0.000	11:49	
Subject Sample	# 3 0.254	11:51	
Breath Volume	3.066	11:51	
Air Blank	0.000	11:52	
Control Test	0.077	11:52	
Air Blank	0.000	11:52	
Diagnostics Check	OK	11:52	

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KHAN,S

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 00:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VIAMONTES

DATE/TIME OF TEST: 02/18/2025 01:33:11

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: REBANDO,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 21:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: APONTE

DATE/TIME OF TEST: 02/18/2025 22:31:35

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.081	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:30	
Control Test	0.080	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 02/25/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: KHAN,K

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 03:02

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SILVIA DORADO

DATE/TIME OF TEST: 02/19/2025 03:52:32

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.069	03:48
Breath Volume	2.218	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.068	03:50
Breath Volume	2.417	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/25/2025 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MACAVOY,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ZIMMERMAN

DATE/TIME OF TEST: 02/19/2025 04:05:31

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.129	04:00
Breath Volume	2.285	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.127	04:03
Breath Volume	2.242	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 02/25/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MACAVOY,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 01:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BOYKIN JR

DATE/TIME OF TEST: 02/22/2025 03:26:58

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.082	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 02/25/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: MACAVOY,J

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 01:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: NAFIELD

DATE/TIME OF TEST: 02/23/2025 02:34:29

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.164	02:29
Breath Volume	1.308	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.163	02:32
Breath Volume	2.093	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 02/25/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: LOPEZ DE LEON

DATE/TIME OF TEST: 02/12/2025 23:10:02

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.008	23:05
Breath Volume	2.113	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.000	23:08
Breath Volume	2.558	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:10

Data Download Date/Time: 03/03/2025 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: LOPEZ DE LEON

DATE/TIME OF TEST: 02/12/2025 23:20:02

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.007	23:15
Breath Volume	2.507	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.000	23:18
Breath Volume	2.238	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:20

Data Download Date/Time: 03/03/2025 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ANGEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA-RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ALEMAN-CRUZ

DATE/TIME OF TEST: 02/15/2025 01:13:11

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.137	01:07
Breath Volume	1.984	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.150	01:11
Breath Volume	2.726	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 02/22/2025 00:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: NUNEZ TEJEDA

DATE/TIME OF TEST: 02/22/2025 01:03:36

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.082	01:03	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 02/22/2025 03:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEITFELD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: MCKEE RUIZ

DATE/TIME OF TEST: 02/15/2025 01:22:32

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.107	01:17
Breath Volume	4.960	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.106	01:21
Breath Volume	4.894	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 03/10/2025 16:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,AUSTIN,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HAMPTON,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: PORRETTA

DATE/TIME OF TEST: 02/01/2025 20:22:21

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.081	20:16
Air Blank	0.000	20:17
Subject Sample # 1	0.120	20:17
Breath Volume	1.539	20:17
Air Blank	0.000	20:18
Air Blank	0.000	20:20
Subject Sample # 2	0.118	20:20
Breath Volume	1.222	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Diagnostics Check	OK	20:22

Data Download Date/Time: 02/05/2025 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEATON,J,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WENNLUND,F
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 02/02/2025 02:18:10

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.098	02:13
Breath Volume	2.890	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.099	02:16
Breath Volume	2.910	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 02/05/2025 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,SAMUEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROHRAFF,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BARON

DATE/TIME OF TEST: 02/03/2025 02:49:35

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.238	02:44
Breath Volume	1.648	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.231	02:48
Breath Volume	1.730	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 02/05/2025 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIMON,KARLHEINZ,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BELTRAN,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: ORDONEZCARDONA

DATE/TIME OF TEST: 02/02/2025 14:44:30

Diagnostics Check	OK	14:37
Air Blank	0.000	14:38
Control Test	0.082	14:38
Air Blank	0.000	14:39
Subject Sample # 1	0.274	14:39
Breath Volume	1.785	14:39
Air Blank	0.000	14:40
Air Blank	0.000	14:42
Subject Sample # 2	0.281	14:42
Breath Volume	1.519	14:42
Air Blank	0.000	14:43
Control Test	0.081	14:43
Air Blank	0.000	14:44
Diagnostics Check	OK	14:44

Data Download Date/Time: 02/06/2025 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 02/03/2025 21:53:54

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.078	21:50	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:53	Sequence Aborted
Breath Volume	0.000	21:53	
Air Blank	0.000	21:53	

Data Download Date/Time: 03/02/2025 00:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANSON,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MANSON,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: HOLAND

DATE/TIME OF TEST: 02/08/2025 17:07:41

Diagnostics Check	OK	17:01
Air Blank	0.000	17:02
Control Test	0.079	17:02
Air Blank	0.000	17:02
Subject Sample # 1	0.044	17:03
Breath Volume	2.351	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.040	17:06
Breath Volume	1.859	17:06
Air Blank	0.000	17:06
Control Test	0.079	17:07
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 03/02/2025 00:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KYLE,F
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LEE,NICHOLAS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: LABAN

DATE/TIME OF TEST: 02/08/2025 22:00:30

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.292	21:55
Breath Volume	2.882	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.277	21:59
Breath Volume	1.214	21:59
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 03/02/2025 00:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANSON,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MANSON,TIMOTHY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: HIRE

DATE/TIME OF TEST: 02/10/2025 00:27:40

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.079	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 03/02/2025 00:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HALE,CALEB
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SALES ORDONEZ

DATE/TIME OF TEST: 02/16/2025 02:42:47

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.188	02:38
Breath Volume	2.304	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.187	02:41
Breath Volume	2.035	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOBEL,EARL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KNOBEL,EARL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:03

SUBJECT'S LAST NAME: HAUN

DATE/TIME OF TEST: 02/19/2025 12:29:59

Diagnostics Check	OK	12:23
Air Blank	0.000	12:24
Control Test	0.081	12:24
Air Blank	0.000	12:24
Subject Sample # 1	0.000	12:25
Breath Volume	1.972	12:25
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample # 2	0.000	12:28
Breath Volume	1.125	12:28
Air Blank	0.000	12:29
Control Test	0.080	12:29
Air Blank	0.000	12:29
Diagnostics Check	OK	12:29

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SHAVER, COLTON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BJORKMAN

DATE/TIME OF TEST: 02/22/2025 21:44:15

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.205	21:39
Breath Volume	1.746	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.190	21:42
Breath Volume	1.800	21:42
Air Blank	0.000	21:43
Control Test	0.078	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KYLE,F
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOLMES,KYLE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HALCOVICH

DATE/TIME OF TEST: 02/25/2025 01:35:43

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.079	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY, TYLER, J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ALVARADO ALVARADO

DATE/TIME OF TEST: 02/28/2025 01:15:38

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1 0.020	01:07	Volume Not Met
Breath Volume	0.167	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2 0.020	01:11	
Breath Volume	1.550	01:11	
Air Blank	0.000	01:11	
Air Blank	0.000	01:13	
Subject Sample	# 3 0.019	01:14	
Breath Volume	2.210	01:14	
Air Blank	0.000	01:14	
Control Test	0.080	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,R
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: HALLMARK

DATE/TIME OF TEST: 02/28/2025 19:48:40

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.080	19:43
Air Blank	0.000	19:43
Subject Sample # 1	0.000	19:44
Breath Volume	1.511	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.000	19:47
Breath Volume	1.980	19:47
Air Blank	0.000	19:47
Control Test	0.080	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 03/02/2025 00:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINK,CHRIS,D
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: LINK,CHRIS
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 01/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 02/19/2025 23:24:57

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.174	23:20
Breath Volume	2.226	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.164	23:23
Breath Volume	2.195	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 03/03/2025 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOW,DAIMON,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: DOW,DAIMON
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: OGORMAN

DATE/TIME OF TEST: 02/03/2025 00:28:35

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.226	00:23
Breath Volume	1.687	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.217	00:27
Breath Volume	1.703	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 02/04/2025 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILTON,GREGORY,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: HILTON,GREGORY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: ADMIN
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:28

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 02/05/2025 14:03:03

Diagnostics Check	OK	13:56
Air Blank	0.000	13:57
Control Test	0.081	13:57
Air Blank	0.000	13:57
Subject Sample # 1	0.042	13:58
Breath Volume	3.632	13:58
Air Blank	0.000	13:59
Air Blank	0.000	14:01
Subject Sample # 2	0.043	14:01
Breath Volume	4.378	14:01
Air Blank	0.000	14:02
Control Test	0.079	14:02
Air Blank	0.000	14:02
Diagnostics Check	OK	14:03

Data Download Date/Time: 03/04/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/03/2025 22:40:10

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 02/25/2025 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MARTINEZ,JACOB
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LALOR-CUEVAS

DATE/TIME OF TEST: 02/09/2025 04:44:14

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.195	04:39
Breath Volume	1.410	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.208	04:42
Breath Volume	1.277	04:42
Air Blank	0.000	04:43
Control Test	0.079	04:43
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 02/25/2025 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ALLEN,BROOKE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 02/10/2025 00:07:32

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:06	
Control Test	0.080	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 02/25/2025 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: MENDEZ,ALINA
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: VILLALBA OJEDA

DATE/TIME OF TEST: 02/15/2025 19:59:05

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.080	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.249	19:54
Breath Volume	1.230	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.248	19:57
Breath Volume	1.949	19:57
Air Blank	0.000	19:58
Control Test	0.079	19:58
Air Blank	0.000	19:59
Diagnostics Check	OK	19:59

Data Download Date/Time: 02/25/2025 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,ALINA,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: GITTENS,THOMAS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: VASQUEZ GONZALEZ

DATE/TIME OF TEST: 02/22/2025 20:54:49

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.079	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.242	20:49
Breath Volume	1.433	20:49
Air Blank	0.000	20:50
Air Blank	0.000	20:52
Subject Sample # 2	0.243	20:53
Breath Volume	1.105	20:53
Air Blank	0.000	20:53
Control Test	0.078	20:54
Air Blank	0.000	20:54
Diagnostics Check	OK	20:54

Data Download Date/Time: 02/25/2025 13:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMIN, PEYTON, C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: FRANK, BILL
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 01/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 02/02/2025 00:03:32

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.147	23:58
Breath Volume	2.847	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.151	00:01
Breath Volume	3.132	00:01
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 02/13/2025 05:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMIN, PEYTON, C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CROFUTT, BRITTANY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: PANDELOS

DATE/TIME OF TEST: 02/12/2025 02:20:55

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.223	02:16
Breath Volume	2.835	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.224	02:19
Breath Volume	1.250	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 02/13/2025 05:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOKER, RONALD, J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: ASHWORTH

DATE/TIME OF TEST: 02/09/2025 18:49:18

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.077	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.193	18:44
Breath Volume	1.871	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:47
Subject Sample # 2	0.189	18:47
Breath Volume	2.007	18:47
Air Blank	0.000	18:48
Control Test	0.076	18:48
Air Blank	0.000	18:49
Diagnostics Check	OK	18:49

Data Download Date/Time: 02/20/2025 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUMMOND,ALEX,V
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: DRUMMOND,ALEX
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 02/04/2025 16:19:28

Diagnostics Check	OK	16:12
Air Blank	0.000	16:13
Control Test	0.080	16:13
Air Blank	0.000	16:13
Subject Sample # 1	0.215	16:14
Breath Volume	2.140	16:14
Air Blank	0.000	16:15
Air Blank	0.000	16:17
Subject Sample # 2	0.221	16:17
Breath Volume	2.363	16:17
Air Blank	0.000	16:18
Control Test	0.078	16:18
Air Blank	0.000	16:19
Diagnostics Check	OK	16:19

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: HIGGINBOTHAM,JACOB
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 02/08/2025 03:24:10

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.000	03:19
Breath Volume	3.429	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.000	03:22
Breath Volume	3.464	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: YARBROUGH,JONATHAN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: MESSMAN

DATE/TIME OF TEST: 02/08/2025 15:02:02

Diagnostics Check	OK	14:55
Air Blank	0.000	14:56
Control Test	0.080	14:56
Air Blank	0.000	14:57
Subject Sample # 1	0.000	14:57
Breath Volume	3.867	14:57
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.000	15:00
Breath Volume	3.746	15:00
Air Blank	0.000	15:01
Control Test	0.080	15:01
Air Blank	0.000	15:01
Diagnostics Check	OK	15:02

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TISDALE,RICHARD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/08/2025 21:07:04

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.081	21:01
Air Blank	0.000	21:01
Subject Sample # 1	0.180	21:02
Breath Volume	1.875	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.174	21:05
Breath Volume	1.867	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:07

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: YARBROUGH,JONATHAN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 02/09/2025 23:55:08

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.067	23:50
Breath Volume	1.617	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.069	23:53
Breath Volume	1.246	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: YARBROUGH,JONATHAN
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 02/14/2025 19:17:42

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.174	19:13
Breath Volume	3.660	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.161	19:16
Breath Volume	2.679	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: LINGOLD,BOBBY
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 02/14/2025 20:54:21

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Subject Sample # 1	0.196	20:42	
Breath Volume	2.035	20:42	
Air Blank	0.000	20:43	
Air Blank	0.000	20:44	
Subject Sample # 2		20:48	No Sample Provided
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample # 3		20:53	Subject Test Refused
Breath Volume	0.000	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:54	
Diagnostics Check	OK	20:54	

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: WILLIAMS,ALEXIS
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 02/17/2025 23:50:26

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1 0.240	23:43	Volume Not Met
Breath Volume	0.718	23:43	
Air Blank	0.000	23:43	
Air Blank	0.000	23:45	
Subject Sample	# 2 0.281	23:46	
Breath Volume	1.984	23:46	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 3 0.275	23:48	
Breath Volume	2.039	23:48	
Air Blank	0.000	23:49	
Control Test	0.081	23:49	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:50	

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TISDALE,RICHARD,S
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TISDALE,RICHARD
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TERRERO VASQUEZ

DATE/TIME OF TEST: 02/22/2025 03:38:18

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.190	03:33
Breath Volume	1.363	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.203	03:36
Breath Volume	2.042	03:36
Air Blank	0.000	03:37
Control Test	0.082	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 02/28/2025 23:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JONATHAN,F
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: SCALISE,ANTHONY
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/05/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096062
DRY GAS EXPIRY DATE: 07/15/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 02/03/2025 01:12:57

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.162	01:08
Breath Volume	2.132	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.167	01:11
Breath Volume	2.105	01:11
Air Blank	0.000	01:12
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/03/2025 03:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 02/20/2025 01:15:47

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:14	RFI Detect
Breath Volume	0.312	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 03/01/2025 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 02/20/2025 01:20:31

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:18	RFI Detect
Breath Volume	1.425	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 03/01/2025 14:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS, LOURDES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MCNALLY

DATE/TIME OF TEST: 02/08/2025 02:53:37

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.161	02:47	Volume Not Met
Breath Volume	0.050	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2	02:52	Subject Test Refused
Breath Volume	0.273	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 02/24/2025 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS, LOURDES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 02/09/2025 03:08:05

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.167	03:03
Breath Volume	1.742	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.166	03:06
Breath Volume	2.031	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 02/24/2025 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LOPEZ ROBLERO

DATE/TIME OF TEST: 02/01/2025 01:11:32

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.082	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.171	01:07
Breath Volume	1.910	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.161	01:10
Breath Volume	1.484	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HERNANDEZ NIEVES

DATE/TIME OF TEST: 02/01/2025 02:29:49

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.083	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.097	02:25
Breath Volume	1.414	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.099	02:28
Breath Volume	1.808	02:28
Air Blank	0.000	02:28
Control Test	0.082	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAPTISTE,JEAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: AROCHO

DATE/TIME OF TEST: 02/01/2025 03:37:20

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.224	03:30	Volume Not Met
Breath Volume	0.707	03:30	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2 0.217	03:35	Volume Not Met
Breath Volume	0.039	03:35	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JACKSON, DONAVAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 02/01/2025 05:26:05

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.082	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.204	05:21
Breath Volume	1.425	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.205	05:24
Breath Volume	2.300	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:25
Diagnostics Check	OK	05:26

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATIAS,ROLANDO
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 02/01/2025 21:05:57

Diagnostics Check	OK	20:59
Air Blank	0.000	21:00
Control Test	0.083	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.240	21:01
Breath Volume	1.867	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.245	21:04
Breath Volume	2.035	21:04
Air Blank	0.000	21:05
Control Test	0.082	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA
SANTILLAN,ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male

OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: VELASQUEZ GABRIEL

DATE/TIME OF TEST: 02/02/2025 00:22:26

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.083	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Slope Not Met
Breath Volume	0.625	00:20	
Air Blank	0.000	00:21	
Control Test	0.083	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA
SANTILLAN,ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: VELASQUEZ GABRIEL

DATE/TIME OF TEST: 02/02/2025 00:52:58

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.083	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.245	00:48
Breath Volume	1.367	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.240	00:51
Breath Volume	1.316	00:51
Air Blank	0.000	00:52
Control Test	0.082	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: PEREZ AMBROCIO

DATE/TIME OF TEST: 02/02/2025 01:17:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.182	01:12
Breath Volume	1.949	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.182	01:15
Breath Volume	1.933	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARARD,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 02/02/2025 05:41:09

Diagnostics Check	OK	05:35
Air Blank	0.000	05:35
Control Test	0.082	05:35
Air Blank	0.000	05:36
Subject Sample # 1	0.195	05:36
Breath Volume	2.851	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.192	05:39
Breath Volume	2.625	05:39
Air Blank	0.000	05:40
Control Test	0.082	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIDEAUX,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: ALIENDRES HIDALGO

DATE/TIME OF TEST: 02/02/2025 07:02:58

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:51	
Control Test	0.083	06:51	
Air Blank	0.000	06:52	
Subject Sample	# 1 0.209	06:55	Volume Not Met
Breath Volume	0.972	06:55	
Air Blank	0.000	06:56	
Air Blank	0.000	06:57	
Subject Sample	# 2 0.229	06:58	
Breath Volume	1.601	06:58	
Air Blank	0.000	06:59	
Air Blank	0.000	07:00	
Subject Sample	# 3 0.226	07:01	
Breath Volume	1.703	07:01	
Air Blank	0.000	07:02	
Control Test	0.082	07:02	
Air Blank	0.000	07:02	
Diagnostics Check	OK	07:02	

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BROWN,KAMAR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:13

SUBJECT'S LAST NAME: AZBAEZA

DATE/TIME OF TEST: 02/02/2025 07:45:15

Diagnostics Check	OK	07:39
Air Blank	0.000	07:39
Control Test	0.082	07:39
Air Blank	0.000	07:40
Subject Sample # 1	0.174	07:40
Breath Volume	1.343	07:40
Air Blank	0.000	07:41
Air Blank	0.000	07:43
Subject Sample # 2	0.177	07:43
Breath Volume	1.367	07:43
Air Blank	0.000	07:44
Control Test	0.082	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHANNESSEN,ETHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:07

SUBJECT'S LAST NAME: ALAWI

DATE/TIME OF TEST: 02/02/2025 08:39:40

Diagnostics Check	OK	08:33
Air Blank	0.000	08:34
Control Test	0.083	08:34
Air Blank	0.000	08:34
Subject Sample # 1	0.000	08:35
Breath Volume	1.750	08:35
Air Blank	0.000	08:35
Air Blank	0.000	08:37
Subject Sample # 2	0.000	08:38
Breath Volume	1.753	08:38
Air Blank	0.000	08:38
Control Test	0.083	08:39
Air Blank	0.000	08:39
Diagnostics Check	OK	08:39

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA, ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: MATEO THOMAS

DATE/TIME OF TEST: 02/03/2025 01:23:44

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.082	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.318	01:13
Breath Volume	1.238	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.293	01:19
Breath Volume	1.984	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 3	0.302	01:22
Breath Volume	1.839	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 02/03/2025 08:52:40

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.083	08:47
Air Blank	0.000	08:47
Subject Sample # 1	0.197	08:48
Breath Volume	1.457	08:48
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.205	08:51
Breath Volume	1.515	08:51
Air Blank	0.000	08:51
Control Test	0.083	08:52
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,KELVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:29

SUBJECT'S LAST NAME: CHESSMORE

DATE/TIME OF TEST: 02/03/2025 10:00:33

Diagnostics Check	OK	09:54
Air Blank	0.000	09:55
Control Test	0.083	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.000	09:56
Breath Volume	2.355	09:56
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.000	09:59
Breath Volume	2.617	09:59
Air Blank	0.000	09:59
Control Test	0.082	09:59
Air Blank	0.000	10:00
Diagnostics Check	OK	10:00

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 02/03/2025 19:40:48

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.082	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.000	19:36
Breath Volume	1.468	19:36
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.000	19:39
Breath Volume	1.675	19:39
Air Blank	0.000	19:39
Control Test	0.083	19:40
Air Blank	0.000	19:40
Diagnostics Check	OK	19:40

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ, MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: GABRIEL ZACARIAS

DATE/TIME OF TEST: 02/03/2025 21:35:32

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.082	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.288	21:31
Breath Volume	1.890	21:31
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.282	21:34
Breath Volume	2.484	21:34
Air Blank	0.000	21:34
Control Test	0.082	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL, PHILLIP
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/04/2025 03:14:30

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.082	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.180	03:10
Breath Volume	1.312	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.189	03:12
Breath Volume	2.441	03:12
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: HERNANDEZ GUTIERREZ

DATE/TIME OF TEST: 02/04/2025 03:50:29

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.244	03:45
Breath Volume	1.714	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:47
Subject Sample # 2	0.246	03:48
Breath Volume	1.734	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,ALFRED
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ALQAHTANI

DATE/TIME OF TEST: 02/04/2025 04:50:37

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:41	
Control Test	0.082	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:45	No Sample Provided
Breath Volume	0.000	04:45	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample	# 2	04:49	Subject Test Refused
Breath Volume	0.000	04:49	
Air Blank	0.000	04:49	
Control Test	0.082	04:50	
Air Blank	0.000	04:50	
Diagnostics Check	OK	04:50	

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: VALDEZ ESTRADA

DATE/TIME OF TEST: 02/04/2025 23:08:45

Diagnostics Check	OK	23:02
Air Blank	0.000	23:03
Control Test	0.083	23:03
Air Blank	0.000	23:03
Subject Sample # 1	0.000	23:04
Breath Volume	3.574	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.000	23:07
Breath Volume	3.996	23:07
Air Blank	0.000	23:07
Control Test	0.083	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 02/06/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESPO, BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 02/04/2025 23:42:16

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.082	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:35	No Sample Provided
Breath Volume	0.000	23:35	
Air Blank	0.000	23:35	
Air Blank	0.000	23:37	
Subject Sample	# 2	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.082	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 02/05/2025 04:15:05

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.083	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.170	04:07
Breath Volume	2.269	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.170	04:13
Breath Volume	1.328	04:13
Air Blank	0.000	04:14
Control Test	0.082	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:15

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PUTMAN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 02/05/2025 17:53:09

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.083	17:47
Air Blank	0.000	17:48
Subject Sample # 1	0.000	17:48
Breath Volume	1.335	17:48
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.000	17:51
Breath Volume	1.449	17:51
Air Blank	0.000	17:52
Control Test	0.082	17:52
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: MORALES JR

DATE/TIME OF TEST: 02/06/2025 00:08:55

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.083	00:03
Air Blank	0.000	00:04
Subject Sample # 1	0.156	00:04
Breath Volume	1.132	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.153	00:07
Breath Volume	3.160	00:07
Air Blank	0.000	00:08
Control Test	0.082	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MCCORMICK JR

DATE/TIME OF TEST: 02/06/2025 01:32:21

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.083	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.100	01:27
Breath Volume	2.582	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.100	01:30
Breath Volume	2.562	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OCASIO,JONATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MANZANO

DATE/TIME OF TEST: 02/06/2025 04:01:55

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.083	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.082	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 02/06/2025 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: VENTURA CORONADO

DATE/TIME OF TEST: 02/06/2025 23:03:11

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.083	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.272	22:55
Breath Volume	3.218	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.250	22:58
Breath Volume	1.933	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 3	0.258	23:01
Breath Volume	1.511	23:01
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 02/06/2025 23:44:56

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.082	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.249	23:40
Breath Volume	3.128	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.247	23:43
Breath Volume	2.496	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SUAREZ GONZALES

DATE/TIME OF TEST: 02/07/2025 22:47:47

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.117	22:42
Breath Volume	1.835	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.114	22:46
Breath Volume	1.492	22:46
Air Blank	0.000	22:46
Control Test	0.083	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NDOUMBE, FREDERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SUICH

DATE/TIME OF TEST: 02/08/2025 00:11:07

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.083	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.213	00:06
Breath Volume	2.656	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.213	00:09
Breath Volume	2.894	00:09
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO,GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CASTLEBERRY III

DATE/TIME OF TEST: 02/08/2025 00:53:43

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.083	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:49
Breath Volume	1.292	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.000	00:52
Breath Volume	2.144	00:52
Air Blank	0.000	00:52
Control Test	0.083	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MELENDEZ OPIO

DATE/TIME OF TEST: 02/08/2025 01:43:07

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.083	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:38
Breath Volume	1.832	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.000	01:41
Breath Volume	2.910	01:41
Air Blank	0.000	01:42
Control Test	0.083	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: CASTANEDA SORIANO

DATE/TIME OF TEST: 02/08/2025 21:50:10

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.083	21:43
Air Blank	0.000	21:44
Subject Sample # 1	0.167	21:45
Breath Volume	1.156	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.171	21:48
Breath Volume	1.746	21:48
Air Blank	0.000	21:49
Control Test	0.083	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRAW,LOGAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 02/09/2025 02:16:27

Diagnostics Check	OK	02:09
Air Blank	0.000	02:10
Control Test	0.083	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.168	02:11
Breath Volume	1.917	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.170	02:14
Breath Volume	1.859	02:14
Air Blank	0.000	02:15
Control Test	0.083	02:15
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING, PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VISHWAKARMA

DATE/TIME OF TEST: 02/09/2025 04:08:53

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.082	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1	04:01	No Sample Provided
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2	04:07	No Sample Provided
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.083	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZURBANSKI, ZACHARY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: ANDRAE

DATE/TIME OF TEST: 02/10/2025 01:04:35

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.084	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2	01:03	No Sample Provided
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.083	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CASAS RAMIREZ

DATE/TIME OF TEST: 02/10/2025 03:34:59

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.083	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1 0.228	03:24	
Breath Volume	1.695	03:24	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.196	03:29	Volume Not Met
Breath Volume	0.441	03:29	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 3 0.214	03:33	
Breath Volume	1.128	03:33	
Air Blank	0.000	03:34	
Control Test	0.081	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO, EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: CUNEO

DATE/TIME OF TEST: 02/10/2025 21:16:51

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.083	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.000	21:12
Breath Volume	1.707	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.000	21:15
Breath Volume	1.164	21:15
Air Blank	0.000	21:15
Control Test	0.082	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA, SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 02/11/2025 05:01:51

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.084	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Control Test	0.083	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY, BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 02/12/2025 00:35:10

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.083	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:30
Breath Volume	3.636	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:33
Breath Volume	3.828	00:33
Air Blank	0.000	00:34
Control Test	0.083	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 02/13/2025 01:01:18

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.083	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:57
Breath Volume	2.132	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	00:59
Breath Volume	2.121	00:59
Air Blank	0.000	01:00
Control Test	0.082	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SAMBA RESEMOND

DATE/TIME OF TEST: 02/13/2025 01:35:30

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.282	01:31
Breath Volume	1.496	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.280	01:33
Breath Volume	1.859	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ, DANIELLI
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CASSEUS

DATE/TIME OF TEST: 02/13/2025 03:16:05

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.083	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.132	03:09
Breath Volume	1.351	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:11
Subject Sample # 2	0.129	03:14
Breath Volume	1.507	03:14
Air Blank	0.000	03:15
Control Test	0.083	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING, PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 02/13/2025 03:39:25

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.082	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.201	03:34
Breath Volume	5.089	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.202	03:37
Breath Volume	4.808	03:37
Air Blank	0.000	03:38
Control Test	0.082	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA, AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 02/13/2025 04:09:27

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.082	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1 0.131	03:58	
Breath Volume	1.339	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 2 0.140	04:04	Volume Not Met
Breath Volume	0.058	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 3 0.124	04:07	
Breath Volume	1.281	04:07	
Air Blank	0.000	04:08	
Control Test	0.081	04:08	
Air Blank	0.000	04:09	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING, PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: MARTIN JR

DATE/TIME OF TEST: 02/13/2025 19:32:10

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.083	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.297	19:27
Breath Volume	1.378	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.278	19:30
Breath Volume	2.882	19:30
Air Blank	0.000	19:31
Control Test	0.082	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 02/18/2025 12:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ARCHAMBAULT

DATE/TIME OF TEST: 02/14/2025 00:35:52

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.082	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.192	00:30
Breath Volume	1.835	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.186	00:34
Breath Volume	1.683	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 02/14/2025 04:04:24

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.083	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.117	03:59
Breath Volume	2.558	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.124	04:02
Breath Volume	1.601	04:02
Air Blank	0.000	04:03
Control Test	0.082	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALTINERLIELMAS,ADEM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 02/15/2025 02:44:29

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.083	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.144	02:39
Breath Volume	2.027	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.141	02:42
Breath Volume	1.476	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA, SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: LO BIONDO JR

DATE/TIME OF TEST: 02/15/2025 23:49:30

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.083	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.065	23:45
Breath Volume	2.269	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.067	23:48
Breath Volume	1.574	23:48
Air Blank	0.000	23:48
Control Test	0.083	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: DOMINGO DOMINGO

DATE/TIME OF TEST: 02/16/2025 07:48:39

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.083	07:40
Air Blank	0.000	07:41
Subject Sample # 1	0.182	07:43
Breath Volume	1.753	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:46
Subject Sample # 2	0.181	07:47
Breath Volume	1.425	07:47
Air Blank	0.000	07:47
Control Test	0.080	07:48
Air Blank	0.000	07:48
Diagnostics Check	OK	07:48

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALTINERLIELMAS,ADEM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 02/16/2025 19:47:23

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.082	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.269	19:43
Breath Volume	1.371	19:43
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.273	19:45
Breath Volume	1.914	19:45
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO, EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 02/16/2025 20:53:30

Diagnostics Check	OK	20:47
Air Blank	0.000	20:47
Control Test	0.082	20:48
Air Blank	0.000	20:48
Subject Sample # 1	0.284	20:49
Breath Volume	1.855	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.290	20:51
Breath Volume	2.125	20:51
Air Blank	0.000	20:52
Control Test	0.082	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO, SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: PROTHRO JR

DATE/TIME OF TEST: 02/18/2025 01:47:40

Diagnostics Check	OK	01:44	
Air Blank	0.000	01:45	
Control Test	0.083	01:45	
Air Blank	0.000	01:46	
Subject Sample	# 1	01:46	Subject Test Refused
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Control Test	0.083	01:47	
Air Blank	0.000	01:47	
Diagnostics Check	OK	01:47	

Data Download Date/Time: 02/18/2025 12:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITRIGHT, MARK, D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SYMEON, SAMANTHA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 10:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: RENNINGER

DATE/TIME OF TEST: 02/03/2025 12:46:36

Diagnostics Check	OK	12:40
Air Blank	0.000	12:40
Control Test	0.082	12:41
Air Blank	0.000	12:41
Subject Sample # 1	0.000	12:42
Breath Volume	3.562	12:42
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.000	12:45
Breath Volume	1.355	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 02/19/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: RENKEN,E
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GUNIER

DATE/TIME OF TEST: 02/07/2025 03:09:47

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.146	03:03
Breath Volume	1.187	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.146	03:07
Breath Volume	1.464	03:07
Air Blank	0.000	03:08
Control Test	0.083	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 02/19/2025 14:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 02/16/2025 03:41:44

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.167	03:40	Volume Not Met
Breath Volume	0.683	03:40	
Air Blank		03:40	Purge Fail
Air Blank	0.000	03:41	

Data Download Date/Time: 02/19/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 02/16/2025 03:53:06

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.161	03:46
Breath Volume	1.246	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.157	03:51
Breath Volume	1.617	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 02/19/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GUTIERREZ RODRIGUEZ

DATE/TIME OF TEST: 02/17/2025 00:56:46

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.145	00:52
Breath Volume	2.234	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.144	00:55
Breath Volume	1.457	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 02/19/2025 14:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 02/01/2025 00:23:57

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.080	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.197	00:19
Breath Volume	2.187	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:21
Subject Sample # 2	0.200	00:22
Breath Volume	2.152	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 02/26/2025 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 03:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 02/01/2025 04:08:07

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.172	04:03
Breath Volume	2.246	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.165	04:06
Breath Volume	1.496	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 02/26/2025 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 02/08/2025 00:08:34

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.204	00:04
Breath Volume	2.761	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.196	00:07
Breath Volume	2.496	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 02/26/2025 12:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,RICARDO
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: SANTIAGO FIGUEROA

DATE/TIME OF TEST: 02/01/2025 05:17:50

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.108	05:13
Breath Volume	3.410	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:15
Subject Sample # 2	0.110	05:16
Breath Volume	3.488	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 02/26/2025 23:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANTIAGO,L
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: TARNOVSKA

DATE/TIME OF TEST: 02/02/2025 02:03:26

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.082	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.134	01:58
Breath Volume	1.417	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.133	02:01
Breath Volume	1.929	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MANNS,HUNTER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: BOLORIN VELILLA

DATE/TIME OF TEST: 02/03/2025 00:41:11

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.082	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.234	00:36
Breath Volume	1.500	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.233	00:39
Breath Volume	1.839	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRISSEY,ZACHARY,M
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: MORRISSEY,ZACHARY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: OLVERA

DATE/TIME OF TEST: 02/04/2025 07:31:05

Diagnostics Check	OK	07:24
Air Blank	0.000	07:25
Control Test	0.082	07:25
Air Blank	0.000	07:25
Subject Sample # 1	0.166	07:26
Breath Volume	1.468	07:26
Air Blank	0.000	07:27
Air Blank	0.000	07:28
Subject Sample # 2	0.174	07:29
Breath Volume	1.492	07:29
Air Blank	0.000	07:30
Control Test	0.081	07:30
Air Blank	0.000	07:30
Diagnostics Check	OK	07:31

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RITTGER,NICHOLAS
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MINICUCCI

DATE/TIME OF TEST: 02/06/2025 00:10:24

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.085	00:09	Interferent Detect
Air Blank	0.000	00:10	

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GARCIA ABREGO

DATE/TIME OF TEST: 02/08/2025 03:34:21

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.187	03:28
Breath Volume	1.679	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.179	03:32
Breath Volume	1.210	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,L
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: OLIVEIRA

DATE/TIME OF TEST: 02/08/2025 04:01:22

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.084	03:56
Breath Volume	1.550	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.082	03:59
Breath Volume	1.628	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NAKAMURA,L
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 02/08/2025 04:52:37

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.094	04:48
Breath Volume	1.933	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.095	04:51
Breath Volume	2.425	04:51
Air Blank	0.000	04:51
Control Test	0.081	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PLAZAS CRUZ

DATE/TIME OF TEST: 02/08/2025 05:31:27

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.082	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.111	05:25
Breath Volume	1.421	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.106	05:29
Breath Volume	1.285	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STEVENS,JAMES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: RODRIGUEZ MAURA

DATE/TIME OF TEST: 02/08/2025 23:00:37

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.082	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.208	22:53
Breath Volume	1.578	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.211	22:59
Breath Volume	1.769	22:59
Air Blank	0.000	22:59
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RUIZ,FREDDY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: POLANCO

DATE/TIME OF TEST: 02/09/2025 02:49:23

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.083	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.223	02:40
Breath Volume	2.812	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.191	02:43
Breath Volume	2.156	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 3	0.205	02:47
Breath Volume	2.691	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,MICHAEL
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MUENZNER

DATE/TIME OF TEST: 02/10/2025 03:52:25

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.082	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.199	03:47
Breath Volume	1.566	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.206	03:50
Breath Volume	2.089	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHMIEDER,NICOLAS,G
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CORDOVA,ODELIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 02/10/2025 17:08:08

Diagnostics Check	OK	17:02
Air Blank	0.000	17:02
Control Test	0.081	17:02
Air Blank	0.000	17:03
Subject Sample # 1	0.258	17:03
Breath Volume	1.839	17:03
Air Blank	0.000	17:04
Air Blank	0.000	17:06
Subject Sample # 2	0.251	17:06
Breath Volume	1.847	17:06
Air Blank	0.000	17:07
Control Test	0.080	17:07
Air Blank	0.000	17:08
Diagnostics Check	OK	17:08

Data Download Date/Time: 02/26/2025 23:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KELLER,K
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 02/10/2025 23:11:27

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.081	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.198	23:06
Breath Volume	2.621	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.190	23:09
Breath Volume	2.359	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLUMSTEIN,R
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MCLACKLAN

DATE/TIME OF TEST: 02/11/2025 23:56:45

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.082	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:51
Breath Volume	2.546	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:55
Breath Volume	2.492	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PETERSON,JO
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DILS

DATE/TIME OF TEST: 02/14/2025 03:36:18

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.083	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.129	03:31
Breath Volume	1.488	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.125	03:34
Breath Volume	1.726	03:34
Air Blank	0.000	03:35
Control Test	0.082	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: THOMAS,A
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 02/14/2025 04:04:50

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.083	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.295	04:00
Breath Volume	1.507	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.297	04:03
Breath Volume	1.574	04:03
Air Blank	0.000	04:03
Control Test	0.082	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MANNS,HUNTER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 02/15/2025 03:37:13

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.083	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.170	03:32
Breath Volume	2.605	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.160	03:35
Breath Volume	1.621	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COOK,ANTHONY
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 02/15/2025 04:04:11

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.168	03:58
Breath Volume	1.207	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.149	04:02
Breath Volume	1.824	04:02
Air Blank	0.000	04:03
Control Test	0.081	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KELLER,KOBE
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 02/15/2025 04:32:04

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.132	04:27
Breath Volume	2.953	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.129	04:30
Breath Volume	2.718	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPINSKI,LUKE,W
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: QUIROGA HUERTAS

DATE/TIME OF TEST: 02/15/2025 19:56:21

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.082	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.277	19:51
Breath Volume	1.570	19:51
Air Blank	0.000	19:52
Air Blank	0.000	19:54
Subject Sample # 2	0.269	19:54
Breath Volume	1.945	19:54
Air Blank	0.000	19:55
Control Test	0.076	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GILLIAN

DATE/TIME OF TEST: 02/16/2025 03:41:37

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.099	03:41	Interferent Detect
Air Blank	0.000	03:41	

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GILLIAN

DATE/TIME OF TEST: 02/16/2025 03:47:35

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Interferent Detect
Breath Volume	3.179	03:46	
Air Blank	0.000	03:46	
Control Test	0.065	03:47	Interferent Detect
Air Blank	0.000	03:47	

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RAMIREZ,ANDRES
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GILLIAN

DATE/TIME OF TEST: 02/16/2025 03:59:21

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.082	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.157	03:54
Breath Volume	3.937	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.163	03:57
Breath Volume	3.656	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURKE,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: GANUS

DATE/TIME OF TEST: 02/16/2025 22:20:15

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.081	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1 0.235	22:08	Volume Not Met
Breath Volume	0.574	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:11	
Subject Sample	# 2 0.281	22:13	
Breath Volume	1.617	22:13	
Air Blank	0.000	22:14	
Air Blank	0.000	22:16	
Subject Sample	# 3 0.275	22:18	
Breath Volume	1.781	22:18	
Air Blank	0.000	22:19	
Control Test	0.076	22:19	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 02/26/2025 23:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,L
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: KRAUS

DATE/TIME OF TEST: 02/18/2025 03:08:42

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.082	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.025	03:04
Breath Volume	2.042	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.025	03:07
Breath Volume	1.578	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:08
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SANDOVAL,R
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 02/20/2025 04:01:23

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.218	03:56
Breath Volume	2.734	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.210	03:59
Breath Volume	2.359	03:59
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STEVENS,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WORLEY

DATE/TIME OF TEST: 02/24/2025 01:59:11

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.082	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.162	01:54
Breath Volume	3.343	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.157	01:57
Breath Volume	3.019	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: KELLER,KOBE
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 02/25/2025 00:58:30

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.082	00:45	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.211	00:48	Volume Not Met
Breath Volume	0.527	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 2 0.197	00:51	
Breath Volume	1.238	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 3 0.194	00:56	
Breath Volume	1.906	00:56	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VELAZQUEZ,R
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 02/25/2025 22:26:12

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.083	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.014	22:21
Breath Volume	1.769	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.012	22:24
Breath Volume	1.777	22:24
Air Blank	0.000	22:25
Control Test	0.082	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MANNS,H
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 02/26/2025 00:40:57

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.082	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.164	00:36
Breath Volume	1.277	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.157	00:39
Breath Volume	2.203	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/26/2025 23:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GONZALEZ,S
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: MACHADO FERNANDEZ

DATE/TIME OF TEST: 02/26/2025 22:20:42

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.083	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.186	22:15
Breath Volume	1.460	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.183	22:19
Breath Volume	1.824	22:19
Air Blank	0.000	22:19
Control Test	0.082	22:20
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 02/26/2025 23:27