

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO,WESLEY,G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: FERINO,WESLEY

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 14:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: BRAGDON

DATE/TIME OF TEST: 02/01/2025 15:32:33

Diagnostics Check	OK	15:25
Air Blank	0.000	15:26
Control Test	0.079	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.241	15:27
Breath Volume	1.921	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.230	15:30
Breath Volume	1.765	15:30
Air Blank	0.000	15:31
Control Test	0.079	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERINO,WESLEY,G

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: NEGRON,JORGE

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 18:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: ZAVODNY

DATE/TIME OF TEST: 02/01/2025 19:32:34

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.000	19:28
Breath Volume	2.339	19:28
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.000	19:31
Breath Volume	1.449	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNETT,JAMES,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BRANDOFINO,JOHN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/03/2025 22:05:01

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Subject Sample	# 1 0.181	21:59
Breath Volume	4.457	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample	# 2 0.179	22:03
Breath Volume	4.578	22:03
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:05

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 14:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: RAIFORD

DATE/TIME OF TEST: 02/05/2025 14:51:00

Diagnostics Check	OK	14:44
Air Blank	0.000	14:45
Control Test	0.080	14:45
Air Blank	0.000	14:46
Subject Sample # 1	0.000	14:46
Breath Volume	1.824	14:46
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.000	14:49
Breath Volume	1.832	14:49
Air Blank	0.000	14:50
Control Test	0.080	14:50
Air Blank	0.000	14:50
Diagnostics Check	OK	14:51

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STEPP,V,

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: STEPP,V

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 20:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: SHIRER

DATE/TIME OF TEST: 02/08/2025 21:48:38

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.158	21:44
Breath Volume	2.585	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.156	21:47
Breath Volume	3.648	21:47
Air Blank	0.000	21:47
Control Test	0.079	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: MIXON,GERMAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: JAYCOX

DATE/TIME OF TEST: 02/09/2025 08:19:31

Diagnostics Check	OK	08:13
Air Blank	0.000	08:13
Control Test	0.079	08:14
Air Blank	0.000	08:14
Subject Sample # 1	0.000	08:15
Breath Volume	2.347	08:15
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.000	08:18
Breath Volume	2.199	08:18
Air Blank	0.000	08:18
Control Test	0.080	08:18
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIARA,KENDALL,R
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HODGE,KAMOI
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 02/09/2025 08:42:07

Diagnostics Check	OK	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Subject Sample # 1	0.000	08:37
Breath Volume	1.785	08:37
Air Blank	0.000	08:38
Air Blank	0.000	08:40
Subject Sample # 2	0.000	08:40
Breath Volume	1.835	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Diagnostics Check	OK	08:42

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARNETT,JAMES,C

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: BARNETT,JAMES

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 04:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/16/2025 05:08:49

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.148	05:03
Breath Volume	1.222	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.145	05:07
Breath Volume	1.335	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

INSTRUMENT REGISTERED TO: PALM BAY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: MATCHETT,I

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 01/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BLUM

DATE/TIME OF TEST: 02/16/2025 22:24:33

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.081	22:20	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.080	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 02/26/2025 11:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MYNAUGH,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DISSOTTLE

DATE/TIME OF TEST: 02/01/2025 22:52:25

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.169	22:47
Breath Volume	2.703	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.163	22:50
Breath Volume	1.886	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 02/19/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,SUZETTE,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: PARKS,M
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 02/02/2025 06:34:34

Diagnostics Check	OK	06:27
Air Blank	0.000	06:27
Control Test	0.081	06:28
Air Blank	0.000	06:28
Subject Sample # 1	0.162	06:29
Breath Volume	2.125	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:31
Subject Sample # 2	0.165	06:32
Breath Volume	2.132	06:32
Air Blank	0.000	06:33
Control Test	0.079	06:33
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 02/19/2025 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUM,JACOB,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LORENZO,GABRIEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 02/03/2025 09:49:02

Diagnostics Check	OK	09:42
Air Blank	0.000	09:43
Control Test	0.081	09:43
Air Blank	0.000	09:43
Subject Sample # 1	0.215	09:44
Breath Volume	1.257	09:44
Air Blank	0.000	09:45
Air Blank	0.000	09:46
Subject Sample # 2	0.215	09:47
Breath Volume	1.476	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:48
Diagnostics Check	OK	09:49

Data Download Date/Time: 02/19/2025 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: GRICAR

DATE/TIME OF TEST: 02/04/2025 20:17:37

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.096	20:12
Breath Volume	1.863	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.103	20:15
Breath Volume	1.718	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: VEGA-CRUZ

DATE/TIME OF TEST: 02/07/2025 23:16:31

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.130	23:11
Breath Volume	1.765	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.135	23:14
Breath Volume	2.101	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: LUTZ,DARIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: POTGIETER

DATE/TIME OF TEST: 02/07/2025 23:49:27

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.080	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1 0.112	23:36	
Breath Volume	2.757	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 2 0.101	23:42	Volume Not Met
Breath Volume	0.285	23:42	
Air Blank	0.000	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 3 0.084	23:47	Volume Not Met
Breath Volume	0.128	23:47	
Air Blank	0.000	23:48	
Control Test	0.075	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GABRIEL YOC

DATE/TIME OF TEST: 02/08/2025 01:39:14

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.105	01:33
Breath Volume	1.769	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.102	01:37
Breath Volume	1.851	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: PALLES

DATE/TIME OF TEST: 02/08/2025 03:40:25

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.156	03:35
Breath Volume	1.437	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.148	03:38
Breath Volume	1.898	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KARSNICK,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CASTRO LEON

DATE/TIME OF TEST: 02/08/2025 21:21:23

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:16
Air Blank	0.000	21:16
Subject Sample # 1	0.083	21:16
Breath Volume	3.410	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.080	21:19
Breath Volume	3.292	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOYSEN,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SCHMITZ

DATE/TIME OF TEST: 02/08/2025 23:23:19

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.038	23:18
Breath Volume	2.207	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.036	23:21
Breath Volume	2.156	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KARSNICK,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BOLAN

DATE/TIME OF TEST: 02/10/2025 03:07:24

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.084	03:02
Breath Volume	2.875	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.081	03:05
Breath Volume	2.578	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGNER, BRIAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SEARCY

DATE/TIME OF TEST: 02/12/2025 23:19:44

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.119	23:15
Breath Volume	2.824	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.118	23:18
Breath Volume	2.468	23:18
Air Blank	0.000	23:18
Control Test	0.079	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINAMORE, ABIGAIL, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FISHMAN, ALEXANDER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/13/2025 01:54:11

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.169	01:49
Breath Volume	1.292	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.166	01:52
Breath Volume	2.140	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: EXTEROVICH

DATE/TIME OF TEST: 02/13/2025 23:56:14

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.081	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.384	23:48
Breath Volume	3.785	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.330	23:51
Breath Volume	1.414	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 3	0.341	23:54
Breath Volume	2.015	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METCALF,LANCE,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MANUEL MANUEL

DATE/TIME OF TEST: 02/14/2025 01:04:58

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.233	01:00
Breath Volume	2.550	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:02
Subject Sample # 2	0.234	01:03
Breath Volume	2.894	01:03
Air Blank	0.000	01:04
Control Test	0.080	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/14/2025 03:36:11

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.028	03:31
Breath Volume	2.253	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.027	03:34
Breath Volume	1.953	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRINKWATER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: STEARNS

DATE/TIME OF TEST: 02/15/2025 01:41:40

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.207	01:36
Breath Volume	1.390	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.212	01:40
Breath Volume	1.847	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KETTGEN,G
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 02/15/2025 20:59:59

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:53	
Control Test	0.080	20:53	
Air Blank	0.000	20:54	
Subject Sample # 1	0.293	20:55	
Breath Volume	2.328	20:55	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample # 2	0.290	20:58	
Breath Volume	2.421	20:58	
Air Blank	0.000	20:58	
Control Test	0.080	20:59	Interferent Detect
Air Blank		20:59	Purge Fail

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KETTGEN,G
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 02/15/2025 21:06:56

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.079	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:05	Interferent Detect
Breath Volume	2.066	21:05	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	Interferent Detect
Air Blank	0.000	21:06	

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,EARTHEN,A
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: FERNANDEZ TORRES

DATE/TIME OF TEST: 02/16/2025 01:57:32

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.178	01:52
Breath Volume	1.996	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.177	01:55
Breath Volume	2.542	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WHITUS,O
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 02/16/2025 04:38:53

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.227	04:33
Breath Volume	2.156	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.212	04:37
Breath Volume	1.910	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,LARRY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WINCELOWICZ,LARRY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: MARSTON

DATE/TIME OF TEST: 02/16/2025 06:29:17

Diagnostics Check	OK	06:22
Air Blank	0.000	06:23
Control Test	0.081	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.135	06:24
Breath Volume	2.722	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.129	06:27
Breath Volume	2.406	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 02/19/2025 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: RIVIECCIO

DATE/TIME OF TEST: 02/01/2025 04:35:23

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.037	04:30
Breath Volume	1.316	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.038	04:33
Breath Volume	1.476	04:33
Air Blank	0.000	04:34
Control Test	0.078	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 02/01/2025 19:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONGA,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ELLIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: UVEGES

DATE/TIME OF TEST: 02/01/2025 07:28:00

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.082	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.094	07:23
Breath Volume	2.613	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.096	07:26
Breath Volume	1.949	07:26
Air Blank	0.000	07:27
Control Test	0.081	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:28

Data Download Date/Time: 02/04/2025 21:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,JAMES,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DEP,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARCHMAN

DATE/TIME OF TEST: 02/05/2025 04:41:20

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.065	04:36
Breath Volume	2.167	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.067	04:39
Breath Volume	2.464	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HOUSEMAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GALPER

DATE/TIME OF TEST: 02/05/2025 22:50:10

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.081	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.116	22:45
Breath Volume	2.175	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.113	22:48
Breath Volume	1.890	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JOHN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WYMER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: KENNING

DATE/TIME OF TEST: 02/07/2025 02:26:21

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.153	02:21
Breath Volume	4.710	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.160	02:24
Breath Volume	2.660	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANO,MARDEE,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STAUB,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 02/08/2025 17:04:20

Diagnostics Check	OK	16:58
Air Blank	0.000	16:58
Control Test	0.081	16:58
Air Blank	0.000	16:59
Subject Sample # 1	0.264	16:59
Breath Volume	1.687	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.274	17:02
Breath Volume	1.527	17:02
Air Blank	0.000	17:03
Control Test	0.080	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN, CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: POLEND

DATE/TIME OF TEST: 02/09/2025 18:48:02

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.312	18:43
Breath Volume	1.539	18:43
Air Blank	0.000	18:44
Air Blank	0.000	18:45
Subject Sample # 2	0.295	18:46
Breath Volume	1.476	18:46
Air Blank	0.000	18:47
Control Test	0.080	18:47
Air Blank	0.000	18:47
Diagnostics Check	OK	18:48

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DAVIDSON,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: HANEY

DATE/TIME OF TEST: 02/12/2025 20:38:47

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:27	
Control Test	0.081	20:28	
Air Blank	0.000	20:28	
Subject Sample	# 1 0.122	20:31	Volume Not Met
Breath Volume	0.070	20:31	
Air Blank	0.000	20:32	
Air Blank	0.000	20:34	
Subject Sample	# 2 0.121	20:37	Volume Not Met
Breath Volume	0.320	20:37	
Air Blank	0.000	20:37	
Control Test	0.081	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COURTNEY,PATRICK,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARRIS,DEP
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 02/15/2025 01:07:14

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.117	01:02
Breath Volume	1.664	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.109	01:05
Breath Volume	1.320	01:05
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSO,CHARLES,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MOROZYUK,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 02/16/2025 11:27:43

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.081	11:21
Air Blank	0.000	11:22
Subject Sample # 1	0.000	11:23
Breath Volume	1.910	11:23
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.000	11:26
Breath Volume	1.324	11:26
Air Blank	0.000	11:26
Control Test	0.081	11:27
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JOHN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARARO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BINGAMAN

DATE/TIME OF TEST: 02/16/2025 23:58:07

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.205	23:53
Breath Volume	4.285	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.207	23:56
Breath Volume	4.421	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEINMETZ,CCSO J,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WILSON,CCSO D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CULLEN

DATE/TIME OF TEST: 02/18/2025 23:03:21

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.081	22:52	
Air Blank	0.000	22:53	
Subject Sample	# 1 0.129	22:56	Volume Not Met
Breath Volume	0.941	22:56	
Air Blank	0.000	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 2	23:02	No Sample Provided
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.080	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YENNA,BRITTANY,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEIHS,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 02/19/2025 18:34:18

Diagnostics Check	OK	18:21	
Air Blank	0.000	18:21	
Control Test	0.081	18:22	
Air Blank	0.000	18:22	
Subject Sample # 1	0.144	18:24	
Breath Volume	1.152	18:24	
Air Blank	0.000	18:25	
Air Blank	0.000	18:27	
Subject Sample # 2	0.125	18:30	Volume Not Met
Breath Volume	0.031	18:30	
Air Blank	0.000	18:30	
Air Blank	0.000	18:32	
Subject Sample # 3		18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:33	
Control Test	0.080	18:33	
Air Blank	0.000	18:34	
Diagnostics Check	OK	18:34	

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARAO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 02/20/2025 03:39:01

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:38	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:39	

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MARRO,M
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 02/21/2025 01:19:25

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.132	01:14
Breath Volume	2.199	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.121	01:17
Breath Volume	1.546	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,JOHN,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARARO,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: YUNKER

DATE/TIME OF TEST: 02/21/2025 01:43:58

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.163	01:38
Breath Volume	1.105	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.164	01:42
Breath Volume	1.277	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KAMPHAUS,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ALTERS

DATE/TIME OF TEST: 02/22/2025 01:03:25

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.132	00:58
Breath Volume	1.312	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.122	01:01
Breath Volume	1.175	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,JOSHUA,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NORRIS,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/23/2025 17:53:52

Diagnostics Check	OK	17:47
Air Blank	0.000	17:47
Control Test	0.081	17:48
Air Blank	0.000	17:48
Subject Sample # 1	0.039	17:49
Breath Volume	2.207	17:49
Air Blank	0.000	17:49
Air Blank	0.000	17:51
Subject Sample # 2	0.037	17:52
Breath Volume	2.613	17:52
Air Blank	0.000	17:52
Control Test	0.081	17:53
Air Blank	0.000	17:53
Diagnostics Check	OK	17:53

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CCSO K,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COLON,CCSO
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PILECKI

DATE/TIME OF TEST: 02/23/2025 23:08:38

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.185	23:04
Breath Volume	3.863	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.170	23:07
Breath Volume	2.570	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:08
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN, CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KAMPHAUS, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 02/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: MCCLURE

DATE/TIME OF TEST: 02/27/2025 18:38:44

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.081	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.000	18:34
Breath Volume	1.597	18:34
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.000	18:37
Breath Volume	1.484	18:37
Air Blank	0.000	18:37
Control Test	0.081	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 03/09/2025 21:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ARKANGELO

DATE/TIME OF TEST: 02/01/2025 03:45:54

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.231	03:41
Breath Volume	1.566	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.233	03:44
Breath Volume	1.562	03:44
Air Blank	0.000	03:44
Control Test	0.076	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 02/03/2025 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: OLIVENCIA

DATE/TIME OF TEST: 02/02/2025 04:59:18

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.134	04:54
Breath Volume	1.769	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.127	04:57
Breath Volume	1.710	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 02/03/2025 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIERS,SETH,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: STIERS,SETH
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:33

SUBJECT'S LAST NAME: SAVEL

DATE/TIME OF TEST: 02/02/2025 09:01:58

Diagnostics Check	OK	08:53
Air Blank	0.000	08:54
Control Test	0.080	08:54
Air Blank	0.000	08:55
Subject Sample # 1	0.023	08:56
Breath Volume	1.621	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.021	09:00
Breath Volume	1.835	09:00
Air Blank	0.000	09:01
Control Test	0.079	09:01
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 02/03/2025 23:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 02/07/2025 23:25:10

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.093	23:20
Breath Volume	2.316	23:20
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.090	23:23
Breath Volume	2.214	23:23
Air Blank	0.000	23:24
Control Test	0.080	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: RODRIGUEZ,X
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: JARDINES LOBAINA

DATE/TIME OF TEST: 02/08/2025 20:43:35

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.058	20:38
Breath Volume	1.332	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.063	20:41
Breath Volume	1.507	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BENNIS

DATE/TIME OF TEST: 02/09/2025 01:10:31

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.081	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.171	01:05
Breath Volume	3.035	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.166	01:08
Breath Volume	3.179	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: FAILS

DATE/TIME OF TEST: 02/09/2025 02:59:07

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.082	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.090	02:54
Breath Volume	1.242	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.085	02:57
Breath Volume	1.203	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIE,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: ABOAGYE

DATE/TIME OF TEST: 02/09/2025 21:36:58

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:30	
Control Test	0.081	21:30	
Air Blank	0.000	21:31	
Subject Sample # 1	0.260	21:32	
Breath Volume	1.886	21:32	
Air Blank	0.000	21:32	
Air Blank	0.000	21:34	
Subject Sample # 2	0.265	21:35	
Breath Volume	2.398	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank		21:36	RFI Detect
Air Blank	0.000	21:36	

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/10/2025 23:50:21

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.164	23:45
Breath Volume	2.511	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.165	23:48
Breath Volume	2.601	23:48
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,JOEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CABALLERO ALFONSO

DATE/TIME OF TEST: 02/12/2025 03:09:55

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1 0.220	03:03	
Breath Volume	2.261	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 2 0.244	03:06	
Breath Volume	2.574	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
		03:09	RFI Detect
Air Blank		03:09	RFI Detect
Air Blank	0.000	03:09	

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,JOEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CABALLERO ALFONSO

DATE/TIME OF TEST: 02/12/2025 03:25:18

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:19
Subject Sample # 1	0.226	03:20
Breath Volume	1.156	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:22
Subject Sample # 2	0.234	03:23
Breath Volume	1.523	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/02/2025 02:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 02/16/2025 04:12:53

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.068	04:08
Breath Volume	1.863	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.063	04:11
Breath Volume	1.585	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MITCHELL,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: HARDING III

DATE/TIME OF TEST: 02/16/2025 21:08:13

Diagnostics Check	OK	21:01
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:03
Breath Volume	2.875	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.000	21:06
Breath Volume	1.660	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 02/16/2025 23:52:44

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.129	23:47
Breath Volume	4.164	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:49
Subject Sample # 2	0.128	23:51
Breath Volume	1.789	23:51
Air Blank	0.000	23:51
Control Test	0.081	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/17/2025 02:07:58

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.182	02:03
Breath Volume	2.210	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.183	02:06
Breath Volume	1.652	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: YEATES

DATE/TIME OF TEST: 02/22/2025 00:25:31

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.000	00:21
Breath Volume	2.671	00:21
Air Blank	0.000	00:21
Air Blank	0.000	00:23
Subject Sample # 2	0.000	00:24
Breath Volume	1.707	00:24
Air Blank	0.000	00:24
Control Test	0.083	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 02/23/2025 00:02:04

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.065	23:57
Breath Volume	2.140	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.059	00:00
Breath Volume	1.707	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BOIANO

DATE/TIME OF TEST: 02/23/2025 02:19:03

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:12	No Sample Provided
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Air Blank	0.000	02:14	
Subject Sample	# 2	02:17	No Sample Provided
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.082	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LANKFORD

DATE/TIME OF TEST: 02/28/2025 01:21:10

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.080	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.000	01:16
Breath Volume	1.960	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:19
Breath Volume	1.664	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: CASSELL

DATE/TIME OF TEST: 02/28/2025 20:03:08

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.082	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.000	19:58
Breath Volume	1.425	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.000	20:01
Breath Volume	1.562	20:01
Air Blank	0.000	20:02
Control Test	0.079	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: DOW JR

DATE/TIME OF TEST: 02/28/2025 20:33:00

Diagnostics Check	OK	20:25
Air Blank	0.000	20:26
Control Test	0.081	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.111	20:28
Breath Volume	2.078	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.103	20:31
Breath Volume	1.992	20:31
Air Blank	0.000	20:32
Control Test	0.080	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:33

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NINA,JOEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 01/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231243
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BROCATO

DATE/TIME OF TEST: 02/28/2025 22:57:33

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.083	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.036	22:53
Breath Volume	2.992	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.037	22:56
Breath Volume	2.769	22:56
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 03/02/2025 02:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BERTOLON

DATE/TIME OF TEST: 02/03/2025 01:46:46

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.082	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.021	01:42
Breath Volume	2.730	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.019	01:45
Breath Volume	2.960	01:45
Air Blank	0.000	01:45
Control Test	0.082	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SALES MALDONADO

DATE/TIME OF TEST: 02/06/2025 00:06:33

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.137	00:01
Breath Volume	2.042	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:03
Subject Sample # 2	0.142	00:05
Breath Volume	2.664	00:05
Air Blank	0.000	00:05
Control Test	0.082	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 02/07/2025 02:04:38

Diagnostics Check	OK	01:57
Air Blank	0.000	01:57
Control Test	0.082	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.000	02:00
Breath Volume	2.492	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.000	02:03
Breath Volume	1.488	02:03
Air Blank	0.000	02:03
Control Test	0.083	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 02/09/2025 17:49:19

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.080	17:44
Air Blank	0.000	17:44
Subject Sample # 1	0.258	17:44
Breath Volume	1.828	17:44
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.248	17:47
Breath Volume	2.207	17:47
Air Blank	0.000	17:48
Control Test	0.080	17:48
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TANNER,M
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 02/09/2025 18:46:18

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:43	
Control Test	0.081	18:44	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:44	Subject Test Refused
Breath Volume	0.000	18:44	
Air Blank	0.000	18:45	
Control Test	0.081	18:45	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:46	

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DESVALES

DATE/TIME OF TEST: 02/11/2025 23:48:56

Diagnostics Check	OK	23:41
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.139	23:44
Breath Volume	1.335	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.156	23:47
Breath Volume	1.578	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 02/13/2025 00:20:59

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:18	
Control Test	0.081	00:18	
Air Blank	0.000	00:19	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 02/13/2025 01:31:51

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.099	01:27
Breath Volume	2.777	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.096	01:30
Breath Volume	2.859	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 02/23/2025 03:12:44

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.204	03:07
Breath Volume	1.242	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.201	03:11
Breath Volume	1.777	03:11
Air Blank	0.000	03:11
Control Test	0.082	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 02/24/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSE,O
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,JORGE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 02/01/2025 04:07:28

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.162	04:02
Breath Volume	1.261	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.179	04:05
Breath Volume	2.171	04:05
Air Blank	0.000	04:06
Control Test	0.076	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,ERIC,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PALACIOS,VICTOR
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: QUIROLA AGUINAGA

DATE/TIME OF TEST: 02/03/2025 08:14:42

Diagnostics Check	OK	08:08
Air Blank	0.000	08:09
Control Test	0.078	08:09
Air Blank	0.000	08:09
Subject Sample # 1	0.108	08:10
Breath Volume	1.753	08:10
Air Blank	0.000	08:11
Air Blank	0.000	08:12
Subject Sample # 2	0.093	08:13
Breath Volume	1.964	08:13
Air Blank	0.000	08:13
Control Test	0.077	08:14
Air Blank	0.000	08:14
Diagnostics Check	OK	08:14

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ARIAS FERNANDEZ

DATE/TIME OF TEST: 02/10/2025 02:04:44

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.207	02:00
Breath Volume	2.253	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.203	02:03
Breath Volume	2.644	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,MARCOS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,JORGE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: MIRANDA EDGAR

DATE/TIME OF TEST: 02/14/2025 06:28:49

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:26	
Control Test	0.078	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 1	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:27	
Control Test	0.078	06:28	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCCI,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: TUCCI,JOSE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: DIAZ-PORTAL

DATE/TIME OF TEST: 02/14/2025 15:49:14

Diagnostics Check	OK	15:33	
Air Blank	0.000	15:33	
Control Test	0.078	15:34	
Air Blank	0.000	15:34	
Subject Sample	# 1 0.245	15:37	Volume Not Met
Breath Volume	0.234	15:37	
Air Blank	0.000	15:38	
Air Blank	0.000	15:40	
Subject Sample	# 2 0.272	15:43	
Breath Volume	1.218	15:43	
Air Blank	0.000	15:43	
Air Blank	0.000	15:45	
Subject Sample	# 3 0.268	15:47	
Breath Volume	1.343	15:47	
Air Blank	0.000	15:48	
Control Test	0.077	15:48	
Air Blank	0.000	15:49	
Diagnostics Check	OK	15:49	

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BROWN
ROSQUETE,STEPHANNE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: CORRE

DATE/TIME OF TEST: 02/18/2025 03:42:24

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.078	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.144	03:37
Breath Volume	1.128	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.131	03:40
Breath Volume	1.687	03:40
Air Blank	0.000	03:41
Control Test	0.077	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 02/20/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROJAS,GARIEL,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: BORROTO,J
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 01/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402370890
DRY GAS EXPIRY DATE: 05/09/2025
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/01/2025 22:17:21

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.078	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.228	22:12
Breath Volume	3.640	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.224	22:15
Breath Volume	1.695	22:15
Air Blank	0.000	22:16
Control Test	0.077	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 02/06/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,IVETTE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FABIEN,JEFFREY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FURLOW

DATE/TIME OF TEST: 02/05/2025 17:07:24

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.081	17:02
Air Blank	0.000	17:02
Subject Sample # 1	0.164	17:02
Breath Volume	1.472	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.159	17:05
Breath Volume	1.343	17:05
Air Blank	0.000	17:06
Control Test	0.081	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 02/26/2025 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMALLEY,BRIAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TRAPAGA,STEVE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/04/2025 15:51:53

Diagnostics Check	OK	15:48	
Air Blank	0.000	15:49	
Control Test	0.081	15:49	
Air Blank	0.000	15:49	
Subject Sample	# 1	15:50	Subject Test Refused
Breath Volume	0.000	15:50	
Air Blank	0.000	15:50	
Control Test	0.077	15:51	
Air Blank	0.000	15:51	
Diagnostics Check	OK	15:51	

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,IVETTE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FABIEN,JEFFREY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: FURLOW

DATE/TIME OF TEST: 02/05/2025 16:24:56

Diagnostics Check	OK	16:22	
Air Blank	0.000	16:22	
Control Test	0.080	16:22	
Air Blank	0.000	16:23	
Subject Sample	# 1	16:23	RFI Detect
Breath Volume	0.000	16:23	
Air Blank	0.000	16:24	
Control Test	0.079	16:24	
Air Blank	0.000	16:24	
Diagnostics Check	OK	16:24	

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,IVETTE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FABIEN,JEFFREY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: FURKOW

DATE/TIME OF TEST: 02/05/2025 16:36:50

Diagnostics Check	OK	16:36	
Air Blank		16:36	RFI Detect
Air Blank	0.000	16:36	

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARDSON,IVETTE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: FABIEN,JEFFREY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: FURLOW

DATE/TIME OF TEST: 02/05/2025 16:53:46

Diagnostics Check	OK	16:50	
Air Blank	0.000	16:50	
Control Test	0.080	16:50	
Air Blank	0.000	16:51	
Subject Sample	# 1	16:51	RFI Detect
Breath Volume	0.082	16:51	
Air Blank	0.000	16:52	
Control Test	0.079	16:53	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:53	

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RASHAUN,K
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HALL,RASHAUN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/16/2025 02:52:32

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RASHAUN,K
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CAMPBELL,KEANDRA
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 02/19/2025 23:45:39

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.102	23:40
Breath Volume	1.328	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.097	23:43
Breath Volume	1.328	23:43
Air Blank	0.000	23:44
Control Test	0.075	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

INSTRUMENT REGISTERED TO: MIAMI GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KENY,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GONZALEZ,KENY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 02/23/2025 15:06:08

Diagnostics Check	OK	14:59
Air Blank	0.000	15:00
Control Test	0.079	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.288	15:01
Breath Volume	3.148	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:03
Subject Sample # 2	0.276	15:04
Breath Volume	2.660	15:04
Air Blank	0.000	15:05
Control Test	0.076	15:05
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 02/26/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROVIROSA, ARMANDO, R
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: ROVIROSA, ARMANDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/03/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: MOODIE

DATE/TIME OF TEST: 02/23/2025 17:55:00

Diagnostics Check	OK	17:48
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Subject Sample # 1	0.207	17:50
Breath Volume	1.921	17:50
Air Blank	0.000	17:50
Air Blank	0.000	17:52
Subject Sample # 2	0.200	17:53
Breath Volume	1.437	17:53
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:55

Data Download Date/Time: 03/10/2025 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JAIRUS,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MCCONNELL,JAIRUS
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402678400
DRY GAS EXPIRY DATE: 02/24/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: VAN NESS

DATE/TIME OF TEST: 02/15/2025 22:10:22

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.140	22:05
Breath Volume	2.167	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.130	22:08
Breath Volume	1.929	22:08
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 02/23/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 02/12/2025 21:47:40

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.124	21:43
Breath Volume	2.371	21:43
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.124	21:46
Breath Volume	1.687	21:46
Air Blank	0.000	21:46
Control Test	0.079	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 02/17/2025 20:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 02/16/2025 02:42:31

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.217	02:37
Breath Volume	1.785	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.231	02:40
Breath Volume	2.765	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 02/17/2025 20:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: GONZALEZ MATA

DATE/TIME OF TEST: 02/01/2025 23:00:04

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank		22:59	RFI Detect
Air Blank		23:00	RFI Detect

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: GONZALEZ MATA

DATE/TIME OF TEST: 02/01/2025 23:10:50

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.144	23:05
Breath Volume	1.605	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.140	23:09
Breath Volume	3.835	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: HOLLY

DATE/TIME OF TEST: 02/03/2025 23:49:44

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.184	23:44
Breath Volume	2.777	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.184	23:47
Breath Volume	3.226	23:47
Air Blank	0.000	23:48
Control Test	0.078	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: KUGEL

DATE/TIME OF TEST: 02/07/2025 20:08:21

Diagnostics Check	OK	20:04	
Air Blank	0.000	20:04	
Control Test	0.079	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:06	RFI Detect
Breath Volume	0.269	20:06	
Air Blank	0.000	20:07	
Control Test	0.080	20:07	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GRATTAN

DATE/TIME OF TEST: 02/08/2025 00:21:01

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.203	00:16
Breath Volume	2.726	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:18
Subject Sample # 2	0.205	00:19
Breath Volume	3.089	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: BERRELEZA RIVAS

DATE/TIME OF TEST: 02/13/2025 21:45:52

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.211	21:40
Breath Volume	1.371	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.227	21:44
Breath Volume	1.785	21:44
Air Blank	0.000	21:44
Control Test	0.078	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RODRIGUEZ
GONZALEZ,MARIO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male

OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GUZMAN REYES

DATE/TIME OF TEST: 02/23/2025 22:44:03

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.081	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:44	

Data Download Date/Time: 02/27/2025 13:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 02/05/2025 19:01:36

Diagnostics Check	OK	18:55
Air Blank	0.000	18:55
Control Test	0.078	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:57
Breath Volume	2.050	18:57
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.000	19:00
Breath Volume	1.871	19:00
Air Blank	0.000	19:00
Control Test	0.077	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 02/07/2025 22:17:41

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.218	22:13
Breath Volume	1.917	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.215	22:16
Breath Volume	1.546	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 02/08/2025 00:12:21

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:10	RFI Detect
Breath Volume	0.113	00:10	
Air Blank	0.000	00:11	
Control Test	0.082	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: RIGGALL,NICHOLAS

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14323080A3

DRY GAS EXPIRY DATE: 09/05/2025

VIOLATION CODE: DUI

ARREST TIME: 23:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 02/08/2025 00:20:14

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.140	00:15
Breath Volume	1.656	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.154	00:18
Breath Volume	2.468	00:18
Air Blank	0.000	00:19
Control Test	0.082	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEREZ MOLINA

DATE/TIME OF TEST: 02/08/2025 04:13:45

Diagnostics Check	Fail	04:13	Voltage/Current Test
Air Blank	0.000	04:13	

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEREZ MOLINA

DATE/TIME OF TEST: 02/08/2025 04:24:06

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.086	04:18
Breath Volume	1.425	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.087	04:22
Breath Volume	2.031	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 02/09/2025 01:08:26

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.000	01:04
Breath Volume	2.437	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.000	01:07
Breath Volume	2.800	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/09/2025 02:56:53

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.029	02:52
Breath Volume	3.132	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.028	02:55
Breath Volume	2.664	02:55
Air Blank	0.000	02:55
Control Test	0.082	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MAR OROZCO

DATE/TIME OF TEST: 02/09/2025 22:21:48

Diagnostics Check	Fail	22:21	Voltage/Current Test
Air Blank	0.000	22:21	

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MAR OROZCO

DATE/TIME OF TEST: 02/09/2025 22:31:46

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.182	22:25
Breath Volume	1.312	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.173	22:30
Breath Volume	1.500	22:30
Air Blank	0.000	22:30
Control Test	0.081	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GREBINGER

DATE/TIME OF TEST: 02/12/2025 22:47:29

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.165	22:42
Breath Volume	3.421	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.165	22:45
Breath Volume	3.398	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGALL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: RIGGALL,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PASCUAL RODRIGUEZ

DATE/TIME OF TEST: 02/24/2025 02:32:16

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.178	02:26
Breath Volume	1.312	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.185	02:30
Breath Volume	1.222	02:30
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 02/24/2025 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 02/02/2025 03:51:33

Diagnostics Check	OK	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.081	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: WOODLEY

DATE/TIME OF TEST: 02/09/2025 23:35:16

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.078	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	0.328 23:30	No .020 agreement
Breath Volume	1.343	23:30	
Air Blank	0.000	23:31	
Air Blank	0.000	23:33	
Subject Sample	# 3	0.357 23:33	No .020 agreement
Breath Volume	2.222	23:33	
Air Blank	0.000	23:34	
Control Test	0.077	23:34	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:35	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, SETH, D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HANKINS, M
ARRESTING OFFICER AGENCY: Shalimar PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 02/11/2025 19:10:54

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.079	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.416	19:06
Breath Volume	1.914	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.414	19:09
Breath Volume	2.234	19:09
Air Blank	0.000	19:09
Control Test	0.077	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING MOONEY,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: HERTEL,JOHN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: ULVEN

DATE/TIME OF TEST: 02/15/2025 21:21:25

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.080	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:17	No Sample Provided
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Air Blank	0.000	21:19	
Subject Sample	# 2	21:20	Subject Test Refused
Breath Volume	0.000	21:20	
Air Blank	0.000	21:20	
Control Test	0.077	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,KYLE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HOGAN,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: GIPSON

DATE/TIME OF TEST: 02/19/2025 21:37:37

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.249	21:32
Breath Volume	1.804	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.233	21:36
Breath Volume	1.292	21:36
Air Blank	0.000	21:36
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,GEORGE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VASQUEZ,GEORGE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 02/22/2025 05:15:58

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.081	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:15	
Control Test	0.081	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIRLING MOONEY,MORGAN,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: JONES,STUART
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/23/2025 05:20:45

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.166	05:11	Volume Not Met
Breath Volume	0.117	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 2 0.171	05:14	
Breath Volume	1.910	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 3	05:19	Subject Test Refused
Breath Volume	0.031	05:19	
Air Blank	0.000	05:19	
Control Test	0.075	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

INSTRUMENT REGISTERED TO: FORT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,RITCHIE,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: RIVERA,RITCHIE
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 01/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: NARONGIN

DATE/TIME OF TEST: 02/27/2025 03:27:41

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.163	03:19	Volume Not Met
Breath Volume	0.093	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:21	
Subject Sample	# 2 0.166	03:22	
Breath Volume	1.976	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 3 0.165	03:26	
Breath Volume	1.863	03:26	
Air Blank	0.000	03:26	
Control Test	0.078	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 02/28/2025 21:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDDY,STERLING,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: EDDY,STERLING
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 02/06/2025 02:33:02

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.083	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.017	02:28
Breath Volume	4.078	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.015	02:31
Breath Volume	4.066	02:31
Air Blank	0.000	02:32
Control Test	0.082	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/20/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SPRY,NICHOLAS
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/06/2025 19:31:53

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.083	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:30	Subject Test Refused
Breath Volume	0.000	19:30	
Air Blank	0.000	19:30	
Control Test	0.084	19:31	
Air Blank	0.000	19:31	
Diagnostics Check	OK	19:31	

Data Download Date/Time: 02/20/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 02/11/2025 00:58:58

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.083	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:58	
Control Test	0.083	00:58	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 02/20/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRY,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: GREENE,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MAZANEK

DATE/TIME OF TEST: 02/15/2025 02:45:01

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.083	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.290	02:40
Breath Volume	2.242	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:42
Subject Sample # 2	0.289	02:43
Breath Volume	2.398	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

Data Download Date/Time: 02/20/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: FANOR,J
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 01/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 02/02/2025 13:17:34

Diagnostics Check	OK	13:14	
Air Blank	0.000	13:14	
Control Test	0.081	13:15	
Air Blank	0.000	13:15	
Subject Sample	# 1	13:16	Subject Test Refused
Breath Volume	0.000	13:16	
Air Blank	0.000	13:16	
Control Test	0.081	13:16	
Air Blank	0.000	13:17	
Diagnostics Check	OK	13:17	

Data Download Date/Time: 02/23/2025 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SASSO,SCOTT,A
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: REAHL,TIM
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CEFALONE

DATE/TIME OF TEST: 02/28/2025 03:56:35

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.162	03:51
Breath Volume	2.375	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.161	03:54
Breath Volume	2.093	03:54
Air Blank	0.000	03:55
Control Test	0.076	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 03/05/2025 07:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLEASE,PHILLIP,S
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAMIREZ,K
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 13:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BOHAN

DATE/TIME OF TEST: 02/28/2025 14:52:42

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.078	14:47
Air Blank	0.000	14:47
Subject Sample # 1	0.284	14:48
Breath Volume	2.894	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.285	14:51
Breath Volume	2.718	14:51
Air Blank	0.000	14:51
Control Test	0.077	14:52
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 03/05/2025 07:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEREZ ZULBARAN

DATE/TIME OF TEST: 02/02/2025 04:53:44

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:47
Air Blank	0.000	04:47
Subject Sample # 1	0.159	04:48
Breath Volume	2.144	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.152	04:52
Breath Volume	1.136	04:52
Air Blank	0.000	04:52
Control Test	0.082	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: STRACHAN,GREGORY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: BLODGETT

DATE/TIME OF TEST: 02/02/2025 20:49:24

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.082	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.214	20:44
Breath Volume	1.542	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.211	20:47
Breath Volume	2.132	20:47
Air Blank	0.000	20:48
Control Test	0.082	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SWIDER,MATHEW
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 02/03/2025 01:36:44

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.082	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1 0.199	01:29	Volume Not Met
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 02/03/2025 22:29:49

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.083	22:24
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:25
Breath Volume	3.015	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.000	22:28
Breath Volume	2.804	22:28
Air Blank	0.000	22:28
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MCKINNEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: EASON

DATE/TIME OF TEST: 02/06/2025 01:31:51

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.081	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1	01:24	No Sample Provided
Breath Volume	0.000	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.082	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEARY,MAGEN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ANDEREGG,A
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BINDIG

DATE/TIME OF TEST: 02/09/2025 03:49:57

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.158	03:45
Breath Volume	1.488	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.178	03:48
Breath Volume	1.972	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,DARRELL,E
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ROLLE,DARRELL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: SEPULVEDA

DATE/TIME OF TEST: 02/09/2025 07:59:21

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.083	07:53
Air Blank	0.000	07:54
Subject Sample # 1	0.000	07:55
Breath Volume	3.062	07:55
Air Blank	0.000	07:55
Air Blank	0.000	07:57
Subject Sample # 2	0.000	07:57
Breath Volume	2.246	07:57
Air Blank	0.000	07:58
Control Test	0.081	07:58
Air Blank	0.000	07:59
Diagnostics Check	OK	07:59

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEARY,MAGEN,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,K
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MOGHARNESI

DATE/TIME OF TEST: 02/09/2025 22:00:40

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.082	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.043	21:56
Breath Volume	2.996	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.044	21:59
Breath Volume	2.382	21:59
Air Blank	0.000	21:59
Control Test	0.080	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GRIMAN

DATE/TIME OF TEST: 02/10/2025 04:10:19

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.083	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.162	04:05
Breath Volume	2.203	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.162	04:08
Breath Volume	3.644	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DAWIDOWICZ,SARAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SLOSS

DATE/TIME OF TEST: 02/11/2025 05:03:41

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.082	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.000	04:59
Breath Volume	2.066	04:59
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.000	05:02
Breath Volume	2.386	05:02
Air Blank	0.000	05:02
Control Test	0.083	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 02/25/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DEESE

DATE/TIME OF TEST: 02/15/2025 01:14:17

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.083	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.210	01:09
Breath Volume	1.941	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.214	01:12
Breath Volume	2.257	01:12
Air Blank	0.000	01:13
Control Test	0.083	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DINGUS,DAVID
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GUILLORY

DATE/TIME OF TEST: 02/15/2025 06:27:41

Diagnostics Check	OK	06:20
Air Blank	0.000	06:21
Control Test	0.083	06:21
Air Blank	0.000	06:22
Subject Sample # 1	0.110	06:22
Breath Volume	3.468	06:22
Air Blank	0.000	06:23
Air Blank	0.000	06:25
Subject Sample # 2	0.110	06:26
Breath Volume	3.046	06:26
Air Blank	0.000	06:26
Control Test	0.082	06:27
Air Blank	0.000	06:27
Diagnostics Check	OK	06:27

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: INDAHL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: COKER

DATE/TIME OF TEST: 02/15/2025 17:27:38

Diagnostics Check	OK	17:21
Air Blank	0.000	17:22
Control Test	0.081	17:22
Air Blank	0.000	17:22
Subject Sample # 1	0.000	17:23
Breath Volume	2.597	17:23
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.000	17:26
Breath Volume	3.597	17:26
Air Blank	0.000	17:26
Control Test	0.080	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: QUELIZ,CARLOS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FACCIOLO

DATE/TIME OF TEST: 02/16/2025 02:57:27

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.082	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.158	02:52
Breath Volume	2.023	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.159	02:55
Breath Volume	1.937	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DAWIDOWICZ,SARAH
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CAVARRETTA

DATE/TIME OF TEST: 02/17/2025 02:14:02

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.084	02:08
Air Blank	0.000	02:08
Subject Sample # 1	0.151	02:09
Breath Volume	3.121	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 2	0.143	02:12
Breath Volume	2.792	02:12
Air Blank	0.000	02:13
Control Test	0.083	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:14

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,ASHLEY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: NYENHUIS,ETHAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 02/22/2025 15:13:12

Diagnostics Check	OK	15:07
Air Blank	0.000	15:07
Control Test	0.082	15:08
Air Blank	0.000	15:08
Subject Sample # 1	0.000	15:09
Breath Volume	2.957	15:09
Air Blank	0.000	15:09
Air Blank	0.000	15:11
Subject Sample # 2	0.000	15:11
Breath Volume	2.953	15:11
Air Blank	0.000	15:12
Control Test	0.081	15:12
Air Blank	0.000	15:13
Diagnostics Check	OK	15:13

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUELIZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MCKINNEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GILMAN

DATE/TIME OF TEST: 02/25/2025 02:20:47

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.186	02:14
Breath Volume	1.906	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.170	02:19
Breath Volume	1.609	02:19
Air Blank	0.000	02:19
Control Test	0.082	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 02/25/2025 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/01/2025 23:06:15

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:02	
Control Test	0.077	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:04	Subject Test Refused
Breath Volume	0.000	23:04	
Air Blank	0.000	23:05	
Control Test	0.077	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 02/06/2025 03:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: RODRIGUEZ GARCIA

DATE/TIME OF TEST: 02/02/2025 06:52:08

Diagnostics Check	OK	06:42	
Air Blank	0.000	06:43	
Control Test	0.078	06:43	
Air Blank	0.000	06:43	
Subject Sample	# 1 0.107	06:46	Volume Not Met
Breath Volume	0.128	06:46	
Air Blank	0.000	06:47	
Air Blank	0.000	06:49	
Subject Sample	# 2	06:50	Subject Test Refused
Breath Volume	0.253	06:50	
Air Blank	0.000	06:51	
Control Test	0.077	06:51	
Air Blank	0.000	06:52	
Diagnostics Check	OK	06:52	

Data Download Date/Time: 02/06/2025 03:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/16/2025 01:33:15

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.077	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.162	01:27
Breath Volume	1.417	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.156	01:31
Breath Volume	1.531	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/06/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: FONSECA GOMEZ

DATE/TIME OF TEST: 02/02/2025 02:28:25

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.079	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1 0.000	02:24	
Breath Volume	1.128	02:24	
Air Blank	0.000	02:24	
Air Blank	0.000	02:26	
Subject Sample	# 2 0.000	02:27	
Breath Volume	1.292	02:27	
Air Blank	0.000	02:27	
Control Test	0.076	02:27	
Air Blank		02:28	RFI Detect
Air Blank		02:28	RFI Detect

Data Download Date/Time: 02/06/2025 01:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/07/2025 01:53:46

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.024	01:49
Breath Volume	1.562	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.021	01:52
Breath Volume	1.449	01:52
Air Blank	0.000	01:52
Control Test	0.076	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: WADDELL,NATHANIEL
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MORTO DIAZ

DATE/TIME OF TEST: 02/08/2025 03:41:06

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.156	03:35
Breath Volume	1.296	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.155	03:39
Breath Volume	1.484	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: CERRO RAMIREZ

DATE/TIME OF TEST: 02/08/2025 05:21:55

Diagnostics Check	OK	05:19	
Air Blank	0.000	05:19	
Control Test	0.079	05:19	
Air Blank	0.000	05:20	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:21	
Control Test	0.078	05:21	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PAUSA LINARES

DATE/TIME OF TEST: 02/09/2025 04:20:16

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.077	04:14
Air Blank	0.000	04:15
Subject Sample # 1	0.119	04:15
Breath Volume	1.707	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.115	04:18
Breath Volume	2.656	04:18
Air Blank	0.000	04:19
Control Test	0.077	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 02/15/2025 01:19:05

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.078	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.078	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 02/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ORTEGA ROJAS

DATE/TIME OF TEST: 02/20/2025 04:31:52

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.079	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.080	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 03/06/2025 08:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 02/01/2025 06:02:51

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:51	
Control Test	0.079	05:51	
Air Blank	0.000	05:52	
Subject Sample	# 1 0.136	05:55	Volume Not Met
Breath Volume	0.070	05:55	
Air Blank	0.000	05:56	
Air Blank	0.000	05:57	
Subject Sample	# 2 0.143	06:01	Volume Not Met
Breath Volume	0.750	06:01	
Air Blank	0.000	06:01	
Control Test	0.076	06:02	
Air Blank	0.000	06:02	
Diagnostics Check	OK	06:02	

Data Download Date/Time: 02/06/2025 02:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 02/01/2025 06:15:56

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:12	
Control Test	0.077	06:12	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:14	Subject Test Refused
Breath Volume	0.000	06:14	
Air Blank	0.000	06:14	
Control Test	0.078	06:15	
Air Blank	0.000	06:15	
Diagnostics Check	OK	06:15	

Data Download Date/Time: 02/06/2025 02:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 01/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/07/2025 05:00:15

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.225	04:55
Breath Volume	1.445	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.226	04:58
Breath Volume	2.308	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 02/24/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN, ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: KUNARD

DATE/TIME OF TEST: 02/02/2025 22:21:57

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.173	22:17
Breath Volume	3.875	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.173	22:20
Breath Volume	4.246	22:20
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 02/07/2025 13:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: POZ HUIX

DATE/TIME OF TEST: 02/01/2025 21:52:39

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Subject Sample # 1	0.114	21:47
Breath Volume	2.449	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.116	21:51
Breath Volume	2.671	21:51
Air Blank	0.000	21:51
Control Test	0.080	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BERGMANN,KYLA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 02/01/2025 22:48:08

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank	0.000	22:36	
Subject Sample	# 1 0.196	22:39	Volume Not Met
Breath Volume	0.785	22:39	
Air Blank	0.000	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 2 0.209	22:43	
Breath Volume	1.210	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 3 0.203	22:46	
Breath Volume	1.210	22:46	
Air Blank	0.000	22:47	
Control Test	0.078	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: SEGURA CARDOSO

DATE/TIME OF TEST: 02/02/2025 05:49:35

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.079	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.115	05:44
Breath Volume	1.742	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.113	05:47
Breath Volume	2.789	05:47
Air Blank	0.000	05:48
Control Test	0.079	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANCHEZ,JONATHAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 02/02/2025 13:34:41

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.080	13:28
Air Blank	0.000	13:29
Subject Sample # 1	0.000	13:30
Breath Volume	1.796	13:30
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.000	13:33
Breath Volume	2.195	13:33
Air Blank	0.000	13:33
Control Test	0.080	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COOPER,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: DIVITTORIO

DATE/TIME OF TEST: 02/02/2025 18:59:59

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.165	18:55
Breath Volume	2.144	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.169	18:58
Breath Volume	2.585	18:58
Air Blank	0.000	18:59
Control Test	0.079	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: FELIPE VELASQUEZ

DATE/TIME OF TEST: 02/03/2025 04:41:11

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.080	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.273	04:36
Breath Volume	3.312	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.268	04:39
Breath Volume	3.289	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: JEANTY

DATE/TIME OF TEST: 02/04/2025 05:56:26

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.080	05:44	
Air Blank	0.000	05:45	
Subject Sample	# 1 0.195	05:48	Volume Not Met
Breath Volume	0.101	05:48	
Air Blank	0.000	05:49	
Air Blank	0.000	05:50	
Subject Sample	# 2 0.235	05:51	
Breath Volume	2.113	05:51	
Air Blank	0.000	05:52	
Air Blank	0.000	05:54	
Subject Sample	# 3 0.226	05:54	
Breath Volume	1.800	05:54	
Air Blank	0.000	05:55	
Control Test	0.079	05:55	
Air Blank	0.000	05:56	
Diagnostics Check	OK	05:56	

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MATHESON, BRYAN
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: BEAR

DATE/TIME OF TEST: 02/05/2025 00:31:14

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.096	00:24
Breath Volume	1.152	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.093	00:29
Breath Volume	2.078	00:29
Air Blank	0.000	00:30
Control Test	0.080	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: TOMEY

DATE/TIME OF TEST: 02/06/2025 00:44:12

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.149	00:39
Breath Volume	2.492	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.145	00:42
Breath Volume	2.839	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TABARES, JESUS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 02/06/2025 18:38:53

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.080	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.140	18:34
Breath Volume	1.500	18:34
Air Blank	0.000	18:35
Air Blank	0.000	18:36
Subject Sample # 2	0.152	18:37
Breath Volume	1.464	18:37
Air Blank	0.000	18:37
Control Test	0.079	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVIS, MILJEN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: AMIN

DATE/TIME OF TEST: 02/06/2025 19:43:35

Diagnostics Check	OK	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.205	19:39
Breath Volume	1.441	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.206	19:41
Breath Volume	2.128	19:41
Air Blank	0.000	19:42
Control Test	0.078	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGLER,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PEREZ DOMINGO

DATE/TIME OF TEST: 02/07/2025 01:48:26

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.248	01:43
Breath Volume	2.027	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.245	01:46
Breath Volume	2.367	01:46
Air Blank	0.000	01:47
Control Test	0.078	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 02/07/2025 04:29:56

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.000	04:25
Breath Volume	3.378	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.000	04:28
Breath Volume	3.078	04:28
Air Blank	0.000	04:29
Control Test	0.080	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 02/07/2025 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: HEATER

DATE/TIME OF TEST: 02/01/2025 14:11:04

Diagnostics Check	OK	14:04
Air Blank	0.000	14:04
Control Test	0.081	14:05
Air Blank	0.000	14:05
Subject Sample # 1	0.000	14:06
Breath Volume	2.230	14:06
Air Blank	0.000	14:06
Air Blank	0.000	14:08
Subject Sample # 2	0.000	14:09
Breath Volume	2.597	14:09
Air Blank	0.000	14:10
Control Test	0.081	14:10
Air Blank	0.000	14:10
Diagnostics Check	OK	14:11

Data Download Date/Time: 02/02/2025 22:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MONTOYA,ASHLEY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/02/2025 20:42:21

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.210	20:37
Breath Volume	3.097	20:37
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.198	20:40
Breath Volume	2.179	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 02/02/2025 22:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 02/04/2025 21:33:32

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.080	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.174	21:29
Breath Volume	3.164	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.165	21:31
Breath Volume	2.839	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 03/10/2025 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: CHAPIN

DATE/TIME OF TEST: 02/06/2025 10:08:49

Diagnostics Check	OK	10:07	
Air Blank	0.000	10:07	
Control Test	0.080	10:07	
Air Blank		10:08	RFI Detect
Air Blank	0.000	10:08	

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:28

SUBJECT'S LAST NAME: CHAPIN

DATE/TIME OF TEST: 02/06/2025 10:17:56

Diagnostics Check	OK	10:11
Air Blank	0.000	10:11
Control Test	0.081	10:12
Air Blank	0.000	10:12
Subject Sample # 1	0.000	10:13
Breath Volume	1.531	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.000	10:16
Breath Volume	1.621	10:16
Air Blank	0.000	10:16
Control Test	0.080	10:17
Air Blank	0.000	10:17
Diagnostics Check	OK	10:17

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 02/07/2025 00:46:50

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.110	00:41
Breath Volume	1.648	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.109	00:45
Breath Volume	2.078	00:45
Air Blank	0.000	00:45
Control Test	0.081	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/08/2025 00:32:07

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.179	00:27
Breath Volume	3.222	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.178	00:30
Breath Volume	2.914	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HARBINSON

DATE/TIME OF TEST: 02/09/2025 02:19:10

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:13	No Sample Provided
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITLIN,CORY,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GITLIN,CORY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: MONROY AJPACAJA

DATE/TIME OF TEST: 02/10/2025 22:38:55

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.230	22:34
Breath Volume	2.796	22:34
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.231	22:37
Breath Volume	3.757	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CREACH

DATE/TIME OF TEST: 02/14/2025 02:51:18

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.080	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.323	02:46
Breath Volume	1.371	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.312	02:49
Breath Volume	1.722	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:09

SUBJECT'S LAST NAME: HENOCK

DATE/TIME OF TEST: 02/14/2025 19:51:18

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Subject Sample # 1	0.181	19:46
Breath Volume	2.062	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.184	19:49
Breath Volume	1.707	19:49
Air Blank	0.000	19:50
Control Test	0.079	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ALONZO-FELIX

DATE/TIME OF TEST: 02/14/2025 21:14:49

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.246	21:10
Breath Volume	1.472	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.248	21:13
Breath Volume	1.960	21:13
Air Blank	0.000	21:13
Control Test	0.080	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HAYWOOD

DATE/TIME OF TEST: 02/15/2025 01:01:48

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.080	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1 0.178	00:54	Volume Not Met
Breath Volume	0.257	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.163	01:00	Volume Not Met
Breath Volume	0.371	01:00	
Air Blank	0.000	01:00	
Control Test	0.078	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HAYWOOD

DATE/TIME OF TEST: 02/15/2025 01:32:21

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.079	01:21	
Air Blank	0.000	01:21	
Subject Sample	# 1 0.165	01:25	Volume Not Met
Breath Volume	0.558	01:25	
Air Blank	0.000	01:25	
Air Blank	0.000	01:27	
Subject Sample	# 2 0.156	01:30	Volume Not Met
Breath Volume	0.019	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 02/20/2025 04:31:57

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.095	04:27
Breath Volume	1.707	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.098	04:30
Breath Volume	1.828	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SCALES

DATE/TIME OF TEST: 02/20/2025 21:05:49

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.082	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.000	21:00
Breath Volume	1.171	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.000	21:04
Breath Volume	1.363	21:04
Air Blank	0.000	21:04
Control Test	0.081	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: JACINTO-PASCUAL

DATE/TIME OF TEST: 02/23/2025 02:59:25

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.218	02:55
Breath Volume	2.949	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.211	02:57
Breath Volume	1.558	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 03/10/2025 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEE,COREY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/23/2025 07:52:24

Diagnostics Check	OK	07:43
Air Blank	0.000	07:44
Control Test	0.082	07:44
Air Blank	0.000	07:45
Subject Sample # 1	0.169	07:47
Breath Volume	3.367	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.162	07:50
Breath Volume	2.800	07:50
Air Blank	0.000	07:51
Control Test	0.080	07:51
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 03/10/2025 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 02/23/2025 19:52:47

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.222	19:48
Breath Volume	3.011	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.213	19:51
Breath Volume	2.781	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 03/10/2025 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: 2253,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: RICHIE

DATE/TIME OF TEST: 02/24/2025 23:42:04

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Subject Sample # 1	0.149	23:37
Breath Volume	2.644	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.154	23:40
Breath Volume	3.554	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 03/10/2025 16:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT, WESNES,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TOUSSAINT, WESNES
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:47

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 02/26/2025 12:27:46

Diagnostics Check	OK	12:20
Air Blank	0.000	12:21
Control Test	0.081	12:21
Air Blank	0.000	12:21
Subject Sample # 1	0.183	12:23
Breath Volume	2.128	12:23
Air Blank	0.000	12:24
Air Blank	0.000	12:25
Subject Sample # 2	0.164	12:26
Breath Volume	2.242	12:26
Air Blank	0.000	12:26
Control Test	0.077	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 03/10/2025 16:38