

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: Yes
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 01524080A4
BREATH TEST OPERATOR: WELLS,CODY,T	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Bradford County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: WELLS,CODY	ARREST TIME: 22:02
ARRESTING OFFICER AGENCY: Bradford County SO	GENDER: Male
LAST AGENCY INSPECTION: 03/01/2025	OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: HYUN	DATE/TIME OF TEST: 03/30/2025 23:29:22
---------------------------	--

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:18	
Control Test	0.079	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:22	No Sample Provided
Breath Volume	0.000	23:22	
Air Blank	0.000	23:22	
Air Blank	0.000	23:24	
Subject Sample	# 2	23:28	No Sample Provided
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 06/13/2025 17:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ABREU CAMACHO

DATE/TIME OF TEST: 03/07/2025 22:07:02

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.117	22:02
Breath Volume	2.218	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.116	22:05
Breath Volume	2.191	22:05
Air Blank	0.000	22:06
Control Test	0.081	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:07

Data Download Date/Time: 03/07/2025 22:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231603

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 15:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LEAL HERNANDEZ

DATE/TIME OF TEST: 03/11/2025 16:28:47

Diagnostics Check	OK	16:27	
Air Blank	0.000	16:27	
Control Test	0.065	16:28	Control Outside Tolerance
Air Blank	0.000	16:28	

Data Download Date/Time: 04/02/2025 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231603

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 15:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LEAL HERNANDEZ

DATE/TIME OF TEST: 03/11/2025 16:36:58

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.081	16:31
Air Blank	0.000	16:32
Subject Sample # 1	0.000	16:32
Breath Volume	1.507	16:32
Air Blank	0.000	16:33
Air Blank	0.000	16:35
Subject Sample # 2	0.000	16:35
Breath Volume	1.238	16:35
Air Blank	0.000	16:36
Control Test	0.082	16:36
Air Blank	0.000	16:36
Diagnostics Check	OK	16:36

Data Download Date/Time: 04/02/2025 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HICKS,DANIEL,N

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: N/A,N/A

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231603

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: COURT ORD

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 03/13/2025 16:38:40

Diagnostics Check	OK	16:32
Air Blank	0.000	16:33
Control Test	0.082	16:33
Air Blank	0.000	16:33
Subject Sample # 1	0.000	16:34
Breath Volume	2.542	16:34
Air Blank	0.000	16:34
Air Blank	0.000	16:36
Subject Sample # 2	0.000	16:37
Breath Volume	2.402	16:37
Air Blank	0.000	16:37
Control Test	0.081	16:38
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 04/02/2025 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/17/2025 20:02:23

Diagnostics Check	OK	19:55
Air Blank	0.000	19:56
Control Test	0.082	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.141	19:57
Breath Volume	1.707	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.140	20:00
Breath Volume	2.605	20:00
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/22/2025 03:16:19

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.132	03:12
Breath Volume	1.910	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.138	03:14
Breath Volume	1.203	03:14
Air Blank	0.000	03:15
Control Test	0.081	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: TOMPKINS

DATE/TIME OF TEST: 03/23/2025 19:50:42

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.232	19:45
Breath Volume	1.917	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.236	19:49
Breath Volume	1.328	19:49
Air Blank	0.000	19:49
Control Test	0.080	19:50
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403056709

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 08:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/27/2025 09:40:10

Diagnostics Check	OK	09:33
Air Blank	0.000	09:33
Control Test	0.081	09:34
Air Blank	0.000	09:34
Subject Sample # 1	0.000	09:35
Breath Volume	1.125	09:35
Air Blank	0.000	09:35
Air Blank	0.000	09:37
Subject Sample # 2	0.000	09:38
Breath Volume	2.121	09:38
Air Blank	0.000	09:39
Control Test	0.081	09:39
Air Blank	0.000	09:40
Diagnostics Check	OK	09:40

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: TURTON

DATE/TIME OF TEST: 03/27/2025 09:53:33

Diagnostics Check	OK	09:47
Air Blank	0.000	09:48
Control Test	0.081	09:48
Air Blank	0.000	09:48
Subject Sample # 1	0.000	09:49
Breath Volume	2.191	09:49
Air Blank	0.000	09:49
Air Blank	0.000	09:51
Subject Sample # 2	0.000	09:52
Breath Volume	2.226	09:52
Air Blank	0.000	09:52
Control Test	0.081	09:52
Air Blank	0.000	09:53
Diagnostics Check	OK	09:53

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: LOZANO QUINTERO

DATE/TIME OF TEST: 03/27/2025 10:29:07

Diagnostics Check	OK	10:22
Air Blank	0.000	10:23
Control Test	0.082	10:23
Air Blank	0.000	10:23
Subject Sample # 1	0.000	10:24
Breath Volume	1.187	10:24
Air Blank	0.000	10:25
Air Blank	0.000	10:27
Subject Sample # 2	0.000	10:27
Breath Volume	1.152	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403056709

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 09:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 03/27/2025 10:45:04

Diagnostics Check	OK	10:39
Air Blank	0.000	10:39
Control Test	0.082	10:39
Air Blank	0.000	10:40
Subject Sample # 1	0.000	10:40
Breath Volume	2.273	10:40
Air Blank	0.000	10:41
Air Blank	0.000	10:43
Subject Sample # 2	0.000	10:43
Breath Volume	1.769	10:43
Air Blank	0.000	10:44
Control Test	0.080	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:45

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:08

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/27/2025 11:35:11

Diagnostics Check	OK	11:28
Air Blank	0.000	11:29
Control Test	0.082	11:29
Air Blank	0.000	11:29
Subject Sample # 1	0.000	11:30
Breath Volume	1.625	11:30
Air Blank	0.000	11:31
Air Blank	0.000	11:33
Subject Sample # 2	0.000	11:33
Breath Volume	2.562	11:33
Air Blank	0.000	11:34
Control Test	0.081	11:34
Air Blank	0.000	11:35
Diagnostics Check	OK	11:35

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HICKS,DANIEL,N

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: N/A,N/A

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403056709

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: COURT ORD

ARREST TIME: 00:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:50

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/27/2025 12:16:15

Diagnostics Check	OK	12:10
Air Blank	0.000	12:10
Control Test	0.082	12:11
Air Blank	0.000	12:11
Subject Sample # 1	0.000	12:12
Breath Volume	2.128	12:12
Air Blank	0.000	12:12
Air Blank	0.000	12:14
Subject Sample # 2	0.000	12:14
Breath Volume	2.226	12:14
Air Blank	0.000	12:15
Control Test	0.081	12:15
Air Blank	0.000	12:16
Diagnostics Check	OK	12:16

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403056709

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 12:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: BELANGE

DATE/TIME OF TEST: 03/28/2025 13:43:18

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.082	13:37
Air Blank	0.000	13:38
Subject Sample # 1	0.000	13:38
Breath Volume	1.242	13:38
Air Blank	0.000	13:39
Air Blank	0.000	13:41
Subject Sample # 2	0.000	13:41
Breath Volume	1.394	13:41
Air Blank	0.000	13:42
Control Test	0.081	13:42
Air Blank	0.000	13:43
Diagnostics Check	OK	13:43

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: TRIMARCHI

DATE/TIME OF TEST: 03/28/2025 13:56:09

Diagnostics Check	OK	13:50
Air Blank	0.000	13:50
Control Test	0.082	13:51
Air Blank	0.000	13:51
Subject Sample # 1	0.000	13:51
Breath Volume	2.621	13:51
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.000	13:54
Breath Volume	2.824	13:54
Air Blank	0.000	13:55
Control Test	0.079	13:55
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 03/28/2025 14:08:36

Diagnostics Check	OK	14:02
Air Blank	0.000	14:03
Control Test	0.081	14:03
Air Blank	0.000	14:03
Subject Sample # 1	0.000	14:04
Breath Volume	2.003	14:04
Air Blank	0.000	14:04
Air Blank	0.000	14:06
Subject Sample # 2	0.000	14:07
Breath Volume	1.582	14:07
Air Blank	0.000	14:07
Control Test	0.080	14:08
Air Blank	0.000	14:08
Diagnostics Check	OK	14:08

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 03/28/2025 14:23:01

Diagnostics Check	OK	14:16
Air Blank	0.000	14:16
Control Test	0.080	14:17
Air Blank	0.000	14:17
Subject Sample # 1	0.000	14:18
Breath Volume	1.226	14:18
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.000	14:21
Breath Volume	1.328	14:21
Air Blank	0.000	14:22
Control Test	0.081	14:22
Air Blank	0.000	14:22
Diagnostics Check	OK	14:23

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403056709

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 13:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/28/2025 14:35:48

Diagnostics Check	OK	14:30
Air Blank	0.000	14:30
Control Test	0.081	14:30
Air Blank	0.000	14:31
Subject Sample # 1	0.000	14:31
Breath Volume	2.410	14:31
Air Blank	0.000	14:32
Air Blank	0.000	14:34
Subject Sample # 2	0.000	14:34
Breath Volume	1.589	14:34
Air Blank	0.000	14:34
Control Test	0.081	14:35
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ELBER,RYAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 03/30/2025 05:38:20

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.082	05:32
Air Blank	0.000	05:32
Subject Sample # 1	0.110	05:33
Breath Volume	1.496	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.113	05:36
Breath Volume	1.734	05:36
Air Blank	0.000	05:37
Control Test	0.082	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: SANTOYA VILLARROEL

DATE/TIME OF TEST: 03/30/2025 08:32:04

Diagnostics Check	OK	08:24	
Air Blank	0.000	08:25	
Control Test	0.082	08:25	
Air Blank	0.000	08:25	
Subject Sample	# 1 0.152	08:27	
Breath Volume	1.695	08:27	
Air Blank	0.000	08:27	
Air Blank	0.000	08:29	
Subject Sample	# 2 0.147	08:31	
Breath Volume	1.812	08:31	
Air Blank		08:31	RFI Detect
Air Blank	0.000	08:32	

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 03/07/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: SANTOYA VILLARROEL

DATE/TIME OF TEST: 03/30/2025 08:40:56

Diagnostics Check	OK	08:34
Air Blank	0.000	08:34
Control Test	0.081	08:35
Air Blank	0.000	08:35
Subject Sample # 1	0.145	08:36
Breath Volume	2.226	08:36
Air Blank	0.000	08:36
Air Blank	0.000	08:38
Subject Sample # 2	0.148	08:39
Breath Volume	1.496	08:39
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:40
Diagnostics Check	OK	08:40

Data Download Date/Time: 04/02/2025 05:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22583
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: GRIZZELL

DATE/TIME OF TEST: 03/01/2025 04:59:59

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.078	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.197	04:55
Breath Volume	3.875	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.204	04:58
Breath Volume	3.734	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 03/10/2025 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: ALMARAZ

DATE/TIME OF TEST: 03/02/2025 05:38:22

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.169	05:33
Breath Volume	2.488	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.172	05:36
Breath Volume	1.246	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/04/2025 01:45:35

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.087	01:40
Breath Volume	1.933	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.081	01:44
Breath Volume	1.597	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,RUSSELL,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: COOK,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: FUGON CUBAS

DATE/TIME OF TEST: 03/04/2025 03:02:02

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.239	02:57
Breath Volume	2.281	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.242	03:00
Breath Volume	2.300	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JERRY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GATES,AMY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 03/04/2025 04:49:36

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.079	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:48	Subject Test Refused
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.079	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAICHE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TAPORCO,DAVID
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MERINO

DATE/TIME OF TEST: 03/04/2025 20:58:30

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:49
Air Blank	0.000	20:49
Subject Sample # 1	0.116	20:51
Breath Volume	1.894	20:51
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.116	20:56
Breath Volume	1.929	20:56
Air Blank	0.000	20:57
Control Test	0.078	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOVELL, WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: NEELY

DATE/TIME OF TEST: 03/08/2025 00:00:01

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	RFI Detect
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.078	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LOVELL,WILLIAM
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: NEELY

DATE/TIME OF TEST: 03/08/2025 00:09:25

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.078	00:02	
Air Blank	0.000	00:03	
Subject Sample	# 1 0.171	00:04	
Breath Volume	2.183	00:04	
Air Blank	0.000	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 2	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.076	00:08	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POWELL,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:54

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 03/08/2025 19:13:09

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:03	
Control Test	0.077	19:04	
Air Blank	0.000	19:04	
Subject Sample	# 1 0.153	19:05	
Breath Volume	3.160	19:05	
Air Blank	0.000	19:06	
Air Blank	0.000	19:08	
Subject Sample	# 2 0.132	19:08	
Breath Volume	1.316	19:08	
Air Blank	0.000	19:09	
Air Blank	0.000	19:11	
Subject Sample	# 3 0.143	19:11	
Breath Volume	2.601	19:11	
Air Blank	0.000	19:12	
Control Test	0.074	19:12	Control Outside Tolerance
Air Blank	0.000	19:13	

Data Download Date/Time: 03/10/2025 12:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 03/14/2025 22:15:45

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.077	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.097	22:11
Breath Volume	1.433	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.090	22:14
Breath Volume	1.667	22:14
Air Blank	0.000	22:14
Control Test	0.076	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARF,KYLE,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARF,KYLE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: RUTTER

DATE/TIME OF TEST: 03/17/2025 06:11:39

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.078	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.045	06:07
Breath Volume	1.496	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.045	06:10
Breath Volume	1.792	06:10
Air Blank	0.000	06:10
Control Test	0.078	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,KAYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KAYLAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: HAYSLIP

DATE/TIME OF TEST: 03/20/2025 21:00:17

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.079	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.101	20:55
Breath Volume	1.847	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:58
Subject Sample # 2	0.099	20:58
Breath Volume	2.562	20:58
Air Blank	0.000	20:59
Control Test	0.077	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK, BIANCA, V
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WARNICK, BIANCA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HARROD

DATE/TIME OF TEST: 03/24/2025 22:52:10

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.077	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:50	Subject Test Refused
Breath Volume	0.000	22:50	
Air Blank	0.000	22:51	
Control Test	0.078	22:51	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREDERICK,TAVISH,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDERICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: MORIN

DATE/TIME OF TEST: 03/27/2025 21:38:18

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.078	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:34
Breath Volume	2.753	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:36
Breath Volume	1.789	21:36
Air Blank	0.000	21:37
Control Test	0.078	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/29/2025 03:16:51

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.224	03:12
Breath Volume	1.320	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.220	03:15
Breath Volume	1.847	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/24/2025 15:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAGHEAD,JEREMY,S
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: JOHNSTON,SHANE
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VALLES

DATE/TIME OF TEST: 03/03/2025 03:29:42

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.210	03:19	Volume Not Met
Breath Volume	0.949	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2 0.247	03:22	No .020 agreement
Breath Volume	1.867	03:22	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 3 0.214	03:27	No .020 agreement
Breath Volume	1.691	03:27	
Air Blank	0.000	03:28	
Control Test	0.078	03:29	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/19/2025 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUMBAUGH,CHARLES,C
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: NGUYEN,MICHAEL
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 03/17/2025 02:37:37

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.176	02:33
Breath Volume	1.914	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.165	02:36
Breath Volume	1.289	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/19/2025 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORZO,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Valparaiso PD
ARRESTING OFFICER: STAPLES,RYAN
ARRESTING OFFICER AGENCY: Valparaiso PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SHULTS

DATE/TIME OF TEST: 03/26/2025 23:25:48

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.123	23:21
Breath Volume	1.750	23:21
Air Blank	0.000	23:21
Air Blank	0.000	23:23
Subject Sample # 2	0.127	23:24
Breath Volume	2.191	23:24
Air Blank	0.000	23:24
Control Test	0.080	23:25
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 04/07/2025 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DAVIDSON,TAYLOR
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: VINSON

DATE/TIME OF TEST: 03/03/2025 03:05:31

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.081	03:04	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 03/13/2025 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERS,STEPHEN,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: PETERS,STEPHEN
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SERANO

DATE/TIME OF TEST: 03/03/2025 23:37:26

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.261	23:31
Breath Volume	1.687	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.246	23:35
Breath Volume	1.628	23:35
Air Blank	0.000	23:36
Control Test	0.077	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 03/13/2025 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 03/06/2025 19:10:06

Diagnostics Check	OK	19:03
Air Blank	0.000	19:03
Control Test	0.080	19:04
Air Blank	0.000	19:04
Subject Sample # 1	0.000	19:05
Breath Volume	1.210	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.000	19:08
Breath Volume	1.113	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 03/13/2025 10:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUGHES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BLUEFORD

DATE/TIME OF TEST: 03/07/2025 04:45:28

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.078	04:43	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:44	
Control Test	0.077	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 03/13/2025 10:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PEREZ,MICHAEL]
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: PANKUCH

DATE/TIME OF TEST: 03/07/2025 09:14:39

Diagnostics Check	OK	09:09	
Air Blank	0.000	09:09	
Control Test	0.081	09:10	
Air Blank	0.000	09:10	
Subject Sample	# 1	09:13	Subject Test Refused
Breath Volume	0.000	09:13	
Air Blank	0.000	09:13	
Control Test	0.081	09:13	
Air Blank	0.000	09:14	
Diagnostics Check	OK	09:14	

Data Download Date/Time: 03/13/2025 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: GARRIDO,ANTHONY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 03/10/2025 00:21:11

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.245	00:16
Breath Volume	3.921	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.242	00:19
Breath Volume	1.777	00:19
Air Blank	0.000	00:20
Control Test	0.079	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 03/13/2025 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WEGESEND

DATE/TIME OF TEST: 03/10/2025 22:36:52

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.144	22:32
Breath Volume	4.820	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.138	22:35
Breath Volume	3.671	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 03/13/2025 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 03/18/2025 03:34:29

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 04/01/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPPOCK,JOEL,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TAPANES,CODY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PEREZ-HERNANDEZ

DATE/TIME OF TEST: 03/21/2025 15:43:44

Diagnostics Check	OK	15:37
Air Blank	0.000	15:38
Control Test	0.081	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.235	15:39
Breath Volume	2.402	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:41
Subject Sample # 2	0.227	15:42
Breath Volume	3.464	15:42
Air Blank	0.000	15:42
Control Test	0.081	15:43
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 04/01/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JOHNSON,KEVIN
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/05/2025 19:56:33

Diagnostics Check	OK	19:48
Air Blank	0.000	19:49
Control Test	0.078	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.188	19:50
Breath Volume	3.164	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.200	19:54
Breath Volume	4.132	19:54
Air Blank	0.000	19:55
Control Test	0.079	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 03/28/2025 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 03/09/2025 23:40:11

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.078	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.257	23:35
Breath Volume	3.031	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.243	23:38
Breath Volume	1.835	23:38
Air Blank	0.000	23:39
Control Test	0.076	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 03/28/2025 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISHER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/10/2025 21:30:34

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.079	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.139	21:26
Breath Volume	1.832	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.146	21:29
Breath Volume	1.808	21:29
Air Blank	0.000	21:29
Control Test	0.076	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 03/28/2025 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MCQUEEN

DATE/TIME OF TEST: 03/11/2025 04:37:39

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.094	04:32
Breath Volume	3.687	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.091	04:36
Breath Volume	2.886	04:36
Air Blank	0.000	04:36
Control Test	0.078	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 03/28/2025 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: LEE,MICHAEL
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ARGANBRIGHT

DATE/TIME OF TEST: 03/15/2025 20:42:34

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.110	20:38
Breath Volume	4.062	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.111	20:41
Breath Volume	3.335	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 03/28/2025 18:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,ERIC,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: LEE,MICHAEL
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MCRA Y

DATE/TIME OF TEST: 03/16/2025 22:22:41

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.225	22:18
Breath Volume	1.265	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.214	22:21
Breath Volume	2.042	22:21
Air Blank	0.000	22:21
Control Test	0.077	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SHUMAY

DATE/TIME OF TEST: 03/19/2025 04:21:56

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.078	04:19	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.079	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 03/20/2025 02:07:40

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.145	02:03
Breath Volume	2.589	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.145	02:06
Breath Volume	2.781	02:06
Air Blank	0.000	02:06
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON, ADAM, L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: WATTS, DANIEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 03/24/2025 00:11:27

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.000	00:10	RFI Detect
Air Blank	0.000	00:11	

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON, ADAM, L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: WATTS, DANIEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 03/24/2025 00:19:49

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.079	00:15	
Air Blank	0.000	00:16	
Subject Sample # 1	0.000	00:16	
Breath Volume	4.359	00:16	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample # 2	0.000	00:19	
Breath Volume	4.550	00:19	
Air Blank		00:19	RFI Detect
Air Blank	0.000	00:19	

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS, LAQUINTA, M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: JOHNS, KEVIN
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: DRAPER

DATE/TIME OF TEST: 03/24/2025 20:26:47

Diagnostics Check	OK	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:21
Subject Sample # 1	0.172	20:22
Breath Volume	3.308	20:22
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.174	20:25
Breath Volume	3.210	20:25
Air Blank	0.000	20:25
Control Test	0.077	20:26
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 03/28/2025 18:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOHDE,JAYME,L
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: GOHDE,JAYME
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WORKMAN

DATE/TIME OF TEST: 03/30/2025 01:56:33

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.210	01:51
Breath Volume	1.757	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.207	01:54
Breath Volume	3.023	01:54
Air Blank	0.000	01:55
Control Test	0.075	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 05/05/2025 07:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KHALIEL,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: KONKINE,PAULINA
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 03/29/2025 03:12:01

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.082	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 04/24/2025 14:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KHALIEL,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: OLIVER,JAYDAN
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/29/2025 03:38:37

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
		03:38	RFI Detect
Air Blank		03:38	RFI Detect

Data Download Date/Time: 04/24/2025 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KHALIEL,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: OLIVER,JAYDAN
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 03/29/2025 03:43:55

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.081	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 04/24/2025 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KHALIEL,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: FIDLER,RYAN
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: NOLAN

DATE/TIME OF TEST: 03/29/2025 03:55:08

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.026	03:50
Breath Volume	1.769	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.028	03:53
Breath Volume	1.769	03:53
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/24/2025 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,KHALIEL,L
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD
ARRESTING OFFICER: KONKINE,PAULINA
ARRESTING OFFICER AGENCY: NAS Jacksonville PD
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A3
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: OTHER
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 03/29/2025 04:18:02

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 04/24/2025 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: RECCA,MICHAEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: RAMSAY

DATE/TIME OF TEST: 03/01/2025 20:07:07

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:04	
Control Test	0.081	20:04	
Air Blank	0.000	20:05	
Subject Sample	# 1	20:06	Sequence Aborted
Breath Volume	0.000	20:06	
Air Blank	0.000	20:07	

Data Download Date/Time: 03/26/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/09/2025 04:33:52

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.109	04:29
Breath Volume	3.710	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.105	04:32
Breath Volume	2.570	04:32
Air Blank	0.000	04:32
Control Test	0.080	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 03/26/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, BRANDIE, L
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: PIERCE, MARK
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 03/09/2025 23:19:12

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.081	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:15
Breath Volume	2.589	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	2.585	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 03/26/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/05/2025 17:46:59

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.080	17:41
Air Blank	0.000	17:42
Subject Sample # 1	0.000	17:42
Breath Volume	1.941	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:45
Breath Volume	1.738	17:45
Air Blank	0.000	17:46
Control Test	0.080	17:46
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 04/01/2025 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Holmes County SO
ARRESTING OFFICER: STRICKLAND,WADE
ARRESTING OFFICER AGENCY: Holmes County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 03/10/2025 16:34:24

Diagnostics Check	OK	16:31	
Air Blank	0.000	16:32	
Control Test	0.079	16:32	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:33	Subject Test Refused
Breath Volume	0.000	16:33	
Air Blank	0.000	16:33	
Control Test	0.079	16:33	
Air Blank	0.000	16:34	
Diagnostics Check	OK	16:34	

Data Download Date/Time: 04/01/2025 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 03/15/2025 07:46:13

Diagnostics Check	OK	07:38
Air Blank	0.000	07:39
Control Test	0.082	07:39
Air Blank	0.000	07:40
Subject Sample # 1	0.176	07:41
Breath Volume	1.570	07:41
Air Blank	0.000	07:42
Air Blank	0.000	07:44
Subject Sample # 2	0.167	07:44
Breath Volume	1.203	07:44
Air Blank	0.000	07:45
Control Test	0.080	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 04/01/2025 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: STALL

DATE/TIME OF TEST: 03/09/2025 06:25:08

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:22	
Control Test	0.080	06:22	
Air Blank	0.000	06:23	
Subject Sample	# 1	06:23	Subject Test Refused
Breath Volume	0.000	06:23	
Air Blank	0.000	06:24	
Control Test	0.080	06:24	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SUBLETT

DATE/TIME OF TEST: 03/15/2025 03:09:16

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.142	03:04
Breath Volume	1.238	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.148	03:07
Breath Volume	1.812	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RUIZ JIMENEZ

DATE/TIME OF TEST: 03/16/2025 23:04:01

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.082	22:50	
Air Blank	0.000	22:51	
Subject Sample	# 1 0.190	22:54	Volume Not Met
Breath Volume	0.714	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 2 0.194	22:58	
Breath Volume	1.523	22:58	
Air Blank	0.000	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 3 0.200	23:02	
Breath Volume	1.566	23:02	
Air Blank	0.000	23:03	
Control Test	0.072	23:03	Control Outside Tolerance
Air Blank	0.000	23:04	

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 03/17/2025 00:46:49

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.122	00:41
Breath Volume	1.968	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.111	00:45
Breath Volume	1.421	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/18/2025 06:00:26

Diagnostics Check	OK	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.140	05:55
Breath Volume	2.714	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.124	05:58
Breath Volume	1.867	05:58
Air Blank	0.000	05:59
Control Test	0.078	05:59
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MUTLUTURK,FATIH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: RIGGS

DATE/TIME OF TEST: 03/21/2025 00:35:34

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.082	00:33	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 03/21/2025 05:18:22

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.082	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.074	05:13
Breath Volume	1.703	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.081	05:16
Breath Volume	1.109	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 03:29:58

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1 0.108	03:28	
Breath Volume	1.343	03:28	
Air Blank		03:29	Purge Fail
Air Blank		03:29	Purge Fail

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 03:33:44

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.047	03:33	Control Outside Tolerance
Air Blank	0.000	03:33	

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/23/2025 03:36:57

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.056	03:36	Control Outside Tolerance
Air Blank	0.000	03:36	

Data Download Date/Time: 04/01/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTER,DYLAN,F
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: HARTER,DYLAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: NESHKOFF

DATE/TIME OF TEST: 03/28/2025 03:11:29

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1 0.147	03:06	
Breath Volume	1.785	03:06	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2 0.140	03:09	
Breath Volume	2.937	03:09	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank		03:11	RFI Detect
Air Blank	0.000	03:11	

Data Download Date/Time: 04/27/2025 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOFF,BETH,B
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 03/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DEES

DATE/TIME OF TEST: 03/29/2025 02:33:04

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:32	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:33	

Data Download Date/Time: 04/27/2025 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOFF,SARA,B
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SABATINE,MATTHEW
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 03/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ERVINE

DATE/TIME OF TEST: 03/29/2025 22:59:49

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 04/27/2025 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUALE,BRANDON,P
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BONNIN,BRICESON
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 03/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 03/30/2025 01:38:33

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.197	01:33
Breath Volume	2.207	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.192	01:37
Breath Volume	2.191	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 04/27/2025 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: AMADOR COREA

DATE/TIME OF TEST: 03/01/2025 20:48:20

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:44	
Control Test	0.081	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:47	
Control Test	0.079	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 03/10/2025 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GILES,OFFICER,G

BREATH TEST OPERATOR AGENCY: Clearwater PD

ARRESTING OFFICER: GILES,OFFICER

ARRESTING OFFICER AGENCY: Clearwater PD

LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 00:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: KAZBOUR

DATE/TIME OF TEST: 03/02/2025 01:48:46

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.078	01:44
Breath Volume	2.003	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.077	01:47
Breath Volume	1.410	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 03/10/2025 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: UPCHURCH,JERRY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:04

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/02/2025 21:56:16

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.081	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 03/10/2025 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 03:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BOWIE

DATE/TIME OF TEST: 03/16/2025 04:34:37

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.143	04:29
Breath Volume	1.902	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.137	04:33
Breath Volume	1.324	04:33
Air Blank	0.000	04:33
Control Test	0.081	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 04/17/2025 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSIER,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GOSIER,MATTHEW
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 03/17/2025 23:47:47

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1	23:46	Subject Test Refused
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:47	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 04/17/2025 13:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SWIERKOWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 03/05/2025 01:59:00

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.082	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.000	01:54
Breath Volume	3.503	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.000	01:57
Breath Volume	3.742	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:59

Data Download Date/Time: 03/12/2025 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PARRIOTT

DATE/TIME OF TEST: 03/07/2025 02:10:38

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.092	02:06
Breath Volume	4.433	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.083	02:09
Breath Volume	4.890	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 03/12/2025 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: MYERS,KIRK
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 02/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 03/08/2025 18:26:12

Diagnostics Check	OK	18:19
Air Blank	0.000	18:19
Control Test	0.081	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.134	18:21
Breath Volume	1.871	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.136	18:24
Breath Volume	2.230	18:24
Air Blank	0.000	18:25
Control Test	0.079	18:25
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

Data Download Date/Time: 03/12/2025 14:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEMON,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MEMON,MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GENEST

DATE/TIME OF TEST: 03/13/2025 22:36:19

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.080	22:29	
Air Blank	0.000	22:30	
Subject Sample	# 1 0.315	22:30	
Breath Volume	2.105	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2 0.306	22:34	
Breath Volume	2.468	22:34	
Air Blank	0.000	22:35	
Control Test	0.080	22:35	
Air Blank		22:35	RFI Detect
Air Blank	0.000	22:36	

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEMON,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MEMON,MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GENEST

DATE/TIME OF TEST: 03/13/2025 22:51:09

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Subject Sample # 1	0.320	22:46
Breath Volume	1.203	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.304	22:49
Breath Volume	1.359	22:49
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MAGANA GAMEZ

DATE/TIME OF TEST: 03/16/2025 00:05:11

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.179	00:00
Breath Volume	3.347	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.178	00:03
Breath Volume	3.507	00:03
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA, ADAEL,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ACOSTA, ADAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: LOYOLA

DATE/TIME OF TEST: 03/16/2025 23:20:30

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.149	23:15
Breath Volume	2.207	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.148	23:18
Breath Volume	1.492	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNING,CHEYENNE,H
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WELLS,CODY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 03/22/2025 00:17:10

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.207	00:09
Breath Volume	3.765	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.184	00:12
Breath Volume	2.628	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 3	0.191	00:15
Breath Volume	2.796	00:15
Air Blank	0.000	00:16
Control Test	0.080	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIERKOWSKI,CHRISTIAN,T
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SWIERKOWSKI,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/22/2025 01:50:38

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.128	01:45
Breath Volume	1.925	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.118	01:49
Breath Volume	1.718	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERTIN,SKYLAR,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BRADY,Z
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/29/2025 00:49:12

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.208	00:44
Breath Volume	2.601	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.195	00:47
Breath Volume	2.445	00:47
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 04/16/2025 13:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: BRADSHAW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: KUHNER

DATE/TIME OF TEST: 03/07/2025 23:42:08

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.000	23:37
Breath Volume	2.546	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.000	23:40
Breath Volume	2.953	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 03/22/2025 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 03/12/2025 22:58:39

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.064	22:53
Breath Volume	1.441	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.068	22:57
Breath Volume	2.464	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 03/22/2025 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LITTLE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 03/16/2025 07:35:35

Diagnostics Check	OK	07:28
Air Blank	0.000	07:29
Control Test	0.080	07:29
Air Blank	0.000	07:30
Subject Sample # 1	0.146	07:31
Breath Volume	2.492	07:31
Air Blank	0.000	07:31
Air Blank	0.000	07:33
Subject Sample # 2	0.143	07:34
Breath Volume	2.281	07:34
Air Blank	0.000	07:34
Control Test	0.080	07:35
Air Blank	0.000	07:35
Diagnostics Check	OK	07:35

Data Download Date/Time: 03/22/2025 00:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 03/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: N260071821720

DATE/TIME OF TEST: 03/22/2025 17:02:50

Diagnostics Check	OK	16:59	
Air Blank	0.000	17:00	
Control Test	0.081	17:00	
Air Blank	0.000	17:01	
Subject Sample	# 1	17:01	Subject Test Refused
Breath Volume	0.000	17:01	
Air Blank	0.000	17:01	
Control Test	0.081	17:02	
Air Blank	0.000	17:02	
Diagnostics Check	OK	17:02	

Data Download Date/Time: 05/03/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: NICHOLS,RICHARD
ARRESTING OFFICER AGENCY: Port St. Joe PD
LAST AGENCY INSPECTION: 03/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: FITCH JR

DATE/TIME OF TEST: 03/22/2025 19:56:09

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.081	19:50
Air Blank	0.000	19:50
Subject Sample # 1	0.187	19:51
Breath Volume	2.308	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.172	19:54
Breath Volume	2.964	19:54
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 05/03/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: GADACZ,MATTHEW
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: OUTSEY

DATE/TIME OF TEST: 03/01/2025 00:51:13

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.083	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.110	00:46
Breath Volume	2.750	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.103	00:49
Breath Volume	2.980	00:49
Air Blank	0.000	00:50
Control Test	0.083	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/04/2025 00:56:48

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.082	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.161	00:52
Breath Volume	2.460	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.150	00:55
Breath Volume	2.375	00:55
Air Blank	0.000	00:55
Control Test	0.083	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: MANZANO MELENDEZ

DATE/TIME OF TEST: 03/06/2025 14:01:09

Diagnostics Check	OK	13:55
Air Blank	0.000	13:55
Control Test	0.083	13:56
Air Blank	0.000	13:56
Subject Sample # 1	0.000	13:57
Breath Volume	2.417	13:57
Air Blank	0.000	13:57
Air Blank	0.000	13:59
Subject Sample # 2	0.000	13:59
Breath Volume	2.066	13:59
Air Blank	0.000	14:00
Control Test	0.081	14:00
Air Blank	0.000	14:01
Diagnostics Check	OK	14:01

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: COIL

DATE/TIME OF TEST: 03/06/2025 15:10:25

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.083	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.000	15:06
Breath Volume	1.316	15:06
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.000	15:09
Breath Volume	1.746	15:09
Air Blank	0.000	15:09
Control Test	0.082	15:09
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: BARTEL

DATE/TIME OF TEST: 03/06/2025 16:57:31

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.084	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:53
Breath Volume	2.035	16:53
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.000	16:56
Breath Volume	1.812	16:56
Air Blank	0.000	16:56
Control Test	0.083	16:56
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 03/06/2025 17:10:22

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.083	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.000	17:06
Breath Volume	1.320	17:06
Air Blank	0.000	17:06
Air Blank	0.000	17:08
Subject Sample # 2	0.000	17:09
Breath Volume	1.160	17:09
Air Blank	0.000	17:09
Control Test	0.083	17:09
Air Blank	0.000	17:10
Diagnostics Check	OK	17:10

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 03/07/2025 00:37:47

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.105	00:33
Breath Volume	3.062	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:35
Subject Sample # 2	0.109	00:36
Breath Volume	2.871	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: MCGRIFF

DATE/TIME OF TEST: 03/07/2025 12:26:16

Diagnostics Check	OK	12:20
Air Blank	0.000	12:20
Control Test	0.082	12:21
Air Blank	0.000	12:21
Subject Sample # 1	0.000	12:22
Breath Volume	1.992	12:22
Air Blank	0.000	12:22
Air Blank	0.000	12:24
Subject Sample # 2	0.000	12:24
Breath Volume	1.851	12:24
Air Blank	0.000	12:25
Control Test	0.082	12:25
Air Blank	0.000	12:26
Diagnostics Check	OK	12:26

Data Download Date/Time: 03/07/2025 16:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: PUMAREJO

DATE/TIME OF TEST: 03/07/2025 12:51:27

Diagnostics Check	OK	12:45
Air Blank	0.000	12:46
Control Test	0.082	12:46
Air Blank	0.000	12:46
Subject Sample # 1	0.000	12:47
Breath Volume	1.960	12:47
Air Blank	0.000	12:47
Air Blank	0.000	12:49
Subject Sample # 2	0.000	12:50
Breath Volume	2.289	12:50
Air Blank	0.000	12:50
Control Test	0.082	12:50
Air Blank	0.000	12:51
Diagnostics Check	OK	12:51

Data Download Date/Time: 03/07/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 03/07/2025 13:38:17

Diagnostics Check	OK	13:32
Air Blank	0.000	13:32
Control Test	0.082	13:33
Air Blank	0.000	13:33
Subject Sample # 1	0.000	13:34
Breath Volume	1.875	13:34
Air Blank	0.000	13:34
Air Blank	0.000	13:36
Subject Sample # 2	0.000	13:36
Breath Volume	1.734	13:36
Air Blank	0.000	13:37
Control Test	0.082	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 03/07/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 02/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 03/07/2025 14:34:14

Diagnostics Check	OK	14:28
Air Blank	0.000	14:28
Control Test	0.083	14:29
Air Blank	0.000	14:29
Subject Sample # 1	0.000	14:30
Breath Volume	1.136	14:30
Air Blank	0.000	14:30
Air Blank	0.000	14:32
Subject Sample # 2	0.000	14:32
Breath Volume	1.144	14:32
Air Blank	0.000	14:33
Control Test	0.082	14:33
Air Blank	0.000	14:34
Diagnostics Check	OK	14:34

Data Download Date/Time: 03/07/2025 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: REYES RODRIGUEZ

DATE/TIME OF TEST: 03/13/2025 01:24:07

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.176	01:19
Breath Volume	2.484	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.181	01:22
Breath Volume	2.550	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 03/20/2025 18:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELANSON,BRADLEY,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MELANSON,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 03/01/2025 00:10:23

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.078	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:09	
Control Test	0.078	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 03/31/2025 21:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TWITTY,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 03/01/2025 03:03:45

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:01	
Control Test	0.078	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.078	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 03/31/2025 21:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/01/2025 18:06:39

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.078	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.284	18:02
Breath Volume	2.070	18:02
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.271	18:05
Breath Volume	2.062	18:05
Air Blank	0.000	18:05
Control Test	0.077	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LOCKE,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: SCARLETT

DATE/TIME OF TEST: 03/02/2025 21:11:55

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:09	
Control Test	0.079	21:09	
Air Blank	0.000	21:10	
Subject Sample	# 1	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:11	
Diagnostics Check	OK	21:11	

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/02/2025 21:51:36

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:47
Breath Volume	1.765	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:50
Breath Volume	1.457	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CROSS,DONOVAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:22

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 03/03/2025 15:52:55

Diagnostics Check	OK	15:42
Air Blank	0.000	15:43
Control Test	0.080	15:43
Air Blank	0.000	15:44
Subject Sample # 1	0.211	15:44
Breath Volume	3.011	15:44
Air Blank	0.000	15:45
Air Blank	0.000	15:47
Subject Sample # 2	0.178	15:47
Breath Volume	1.363	15:47
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 3	0.196	15:51
Breath Volume	2.578	15:51
Air Blank	0.000	15:52
Control Test	0.080	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KEITH

DATE/TIME OF TEST: 03/05/2025 00:57:06

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.118	00:51
Breath Volume	1.281	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.130	00:55
Breath Volume	1.570	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MCNEELEY,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/06/2025 04:12:24

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.140	04:07
Breath Volume	1.281	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.135	04:10
Breath Volume	2.304	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACCHIONE,VINCENT,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MORALES,OSIEL
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: BASILE

DATE/TIME OF TEST: 03/08/2025 00:15:01

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.075	00:10
Breath Volume	1.882	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.070	00:13
Breath Volume	1.742	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/08/2025 00:59:36

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.167	00:55
Breath Volume	1.242	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.170	00:58
Breath Volume	1.566	00:58
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 03/08/2025 06:33:39

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:31	
Control Test	0.080	06:31	
Air Blank	0.000	06:32	
Subject Sample	# 1	06:32	Subject Test Refused
Breath Volume	0.000	06:32	
Air Blank	0.000	06:32	
Control Test	0.079	06:33	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:33	

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/09/2025 20:29:44

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.138	20:24
Breath Volume	1.417	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.147	20:28
Breath Volume	1.785	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 03/15/2025 04:48:30

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.159	04:43
Breath Volume	2.894	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.150	04:46
Breath Volume	2.191	04:46
Air Blank	0.000	04:47
Control Test	0.079	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/16/2025 20:04:11

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.079	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 03/18/2025 03:21:51

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.183	03:16
Breath Volume	1.796	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.172	03:20
Breath Volume	1.593	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DESANTIS

DATE/TIME OF TEST: 03/18/2025 04:06:21

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.120	04:01
Breath Volume	1.636	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.124	04:04
Breath Volume	2.281	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KIBBY,BRADLY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/21/2025 12:02:20

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:57
Breath Volume	2.144	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:00
Breath Volume	2.093	12:00
Air Blank	0.000	12:01
Control Test	0.080	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 03/31/2025 21:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CORCORAN

DATE/TIME OF TEST: 03/22/2025 23:08:07

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.153	23:03
Breath Volume	1.730	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.152	23:06
Breath Volume	2.070	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACCHIONE,VINCENT,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FERRO,T
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: BACAJOL QUINAC

DATE/TIME OF TEST: 03/23/2025 03:30:32

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.201	03:25
Breath Volume	1.949	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.196	03:28
Breath Volume	1.156	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VALENTINE

DATE/TIME OF TEST: 03/23/2025 04:07:20

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.116	04:02
Breath Volume	1.648	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.107	04:05
Breath Volume	2.171	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: BRYDEN

DATE/TIME OF TEST: 03/23/2025 22:21:56

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.117	22:17
Breath Volume	1.949	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.113	22:20
Breath Volume	1.878	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 03/24/2025 01:04:07

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:01	
Control Test	0.078	01:01	
Air Blank	0.000	01:02	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.078	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRAKE,DEREK,D
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COURINGTON,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: BERNHARDT

DATE/TIME OF TEST: 03/25/2025 15:06:53

Diagnostics Check	OK	14:57
Air Blank	0.000	14:58
Control Test	0.080	14:58
Air Blank	0.000	14:58
Subject Sample # 1	0.080	15:01
Breath Volume	1.148	15:01
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample # 2	0.078	15:05
Breath Volume	1.167	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:06
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 03/25/2025 20:29:39

Diagnostics Check	OK	20:20
Air Blank	0.000	20:20
Control Test	0.080	20:21
Air Blank	0.000	20:21
Subject Sample # 1	0.155	20:22
Breath Volume	3.871	20:22
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.124	20:24
Breath Volume	3.488	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 3	0.148	20:28
Breath Volume	3.597	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYNNE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 03/26/2025 01:46:32

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.080	01:41	
Air Blank	0.000	01:42	
Subject Sample # 1	0.284	01:43	
Breath Volume	2.984	01:43	
Air Blank	0.000	01:44	
Air Blank		01:46	Ambient Fail
Air Blank	0.000	01:46	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILLA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CICHON

DATE/TIME OF TEST: 03/30/2025 22:29:21

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:26	
Control Test	0.079	22:26	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:28	
Control Test	0.079	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHOJNOWSKY,IWONA,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: DAVILA,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: POLANCO BARBERIS

DATE/TIME OF TEST: 03/31/2025 03:56:28

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:45	
Control Test	0.080	03:45	
Air Blank	0.000	03:46	
Subject Sample	# 1 0.208	03:49	Volume Not Met
Breath Volume	0.066	03:49	
Air Blank	0.000	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 2	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.075	03:55	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 03/31/2025 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOTZ,GEOFFREY,D
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: PURINTON

DATE/TIME OF TEST: 03/01/2025 00:26:16

Diagnostics Check	OK	00:10	
Air Blank	0.000	00:10	
Control Test	0.080	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1 0.140	00:13	
Breath Volume	1.679	00:13	
Air Blank	0.000	00:14	
Air Blank	0.000	00:15	
Subject Sample	# 2 0.105	00:18	Volume Not Met
Breath Volume	0.410	00:18	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 3 0.117	00:24	Volume Not Met
Breath Volume	0.093	00:24	
Air Blank	0.000	00:25	
Control Test	0.078	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: JARNAGAN

DATE/TIME OF TEST: 03/01/2025 02:13:04

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.091	02:08
Breath Volume	2.871	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.095	02:11
Breath Volume	2.660	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:13

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SHANK

DATE/TIME OF TEST: 03/03/2025 18:35:02

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.080	18:28
Air Blank	0.000	18:29
Subject Sample # 1	0.215	18:30
Breath Volume	1.574	18:30
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.225	18:33
Breath Volume	1.636	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:35

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BANNON, C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: EGGERS

DATE/TIME OF TEST: 03/04/2025 03:31:45

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.080	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.119	03:26
Breath Volume	1.550	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.120	03:30
Breath Volume	3.363	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VANWINKLE

DATE/TIME OF TEST: 03/07/2025 01:47:12

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.041	01:42
Breath Volume	3.636	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.040	01:45
Breath Volume	3.773	01:45
Air Blank	0.000	01:46
Control Test	0.082	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOVANOVA, MICHAEL, N
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOVANOVA, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:37

SUBJECT'S LAST NAME: MCATEE

DATE/TIME OF TEST: 03/08/2025 13:04:33

Diagnostics Check	OK	12:58
Air Blank	0.000	12:58
Control Test	0.080	12:58
Air Blank	0.000	12:59
Subject Sample # 1	0.230	13:00
Breath Volume	1.539	13:00
Air Blank	0.000	13:00
Air Blank	0.000	13:02
Subject Sample # 2	0.225	13:02
Breath Volume	1.625	13:02
Air Blank	0.000	13:03
Control Test	0.080	13:03
Air Blank	0.000	13:04
Diagnostics Check	OK	13:04

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIBBEY,T,J
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: LIBBEY

DATE/TIME OF TEST: 03/09/2025 15:06:40

Diagnostics Check	OK	14:59
Air Blank	0.000	14:59
Control Test	0.080	15:00
Air Blank	0.000	15:00
Subject Sample # 1	0.139	15:01
Breath Volume	2.277	15:01
Air Blank	0.000	15:02
Air Blank	0.000	15:04
Subject Sample # 2	0.146	15:05
Breath Volume	3.167	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:06
Air Blank	0.000	15:06
Diagnostics Check	OK	15:06

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ, G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: GUMBLE

DATE/TIME OF TEST: 03/10/2025 02:16:33

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.081	02:14	
Air Blank	0.000	02:14	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSMANOVIC,TAMMERA,J
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: RIOS,NICHOLAS
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: SCHRADE

DATE/TIME OF TEST: 03/12/2025 14:58:53

Diagnostics Check	OK	14:53
Air Blank	0.000	14:53
Control Test	0.082	14:53
Air Blank	0.000	14:54
Subject Sample # 1	0.000	14:54
Breath Volume	2.761	14:54
Air Blank	0.000	14:55
Air Blank	0.000	14:57
Subject Sample # 2	0.000	14:57
Breath Volume	2.597	14:57
Air Blank	0.000	14:57
Control Test	0.081	14:58
Air Blank	0.000	14:58
Diagnostics Check	OK	14:58

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSMANOVIC,TAMMERA,J
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: BELTCHER,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: SCHRADE

DATE/TIME OF TEST: 03/12/2025 15:30:52

Diagnostics Check	OK	15:24
Air Blank	0.000	15:25
Control Test	0.082	15:25
Air Blank	0.000	15:26
Subject Sample # 1	0.000	15:26
Breath Volume	1.699	15:26
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.000	15:29
Breath Volume	2.847	15:29
Air Blank	0.000	15:29
Control Test	0.081	15:30
Air Blank	0.000	15:30
Diagnostics Check	OK	15:30

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ, G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:07

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/13/2025 18:32:06

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:28	
Control Test	0.081	18:28	
Air Blank	0.000	18:29	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.484	18:30	
Air Blank	0.000	18:31	
Control Test	0.082	18:31	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 03/31/2025 10:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIELDS-GIVHAN,ANDREW,
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SHIELDS-GIVHAN

DATE/TIME OF TEST: 03/15/2025 03:16:53

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.082	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,FRANCISCO,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GETZ,KEVIN
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: GRIER

DATE/TIME OF TEST: 03/15/2025 23:00:11

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:57	
Control Test	0.081	22:57	
Air Blank	0.000	22:58	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.082	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VANSKOY, E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/18/2025 23:43:57

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:43	
Control Test	0.082	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: IKELER

DATE/TIME OF TEST: 03/21/2025 23:38:45

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.082	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.154	23:34
Breath Volume	1.464	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.147	23:37
Breath Volume	1.890	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: REDWAN, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 03/24/2025 01:07:55

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.226	01:03
Breath Volume	2.449	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.207	01:06
Breath Volume	1.164	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 03/28/2025 20:50:04

Diagnostics Check	OK	20:49	
Air Blank		20:49	RFI Detect
Air Blank	0.000	20:50	

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 03/28/2025 20:56:49

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.081	20:54	
Air Blank	0.000	20:55	
Subject Sample	# 1	20:56	RFI Detect
Breath Volume	0.000	20:56	
Air Blank		20:56	RFI Detect
Air Blank		20:56	RFI Detect

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 03/28/2025 21:06:55

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.082	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.013	21:02
Breath Volume	2.519	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.011	21:05
Breath Volume	2.457	21:05
Air Blank	0.000	21:05
Control Test	0.081	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MATTHEW

DATE/TIME OF TEST: 03/29/2025 02:03:25

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,BRITTANY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: COURTER

DATE/TIME OF TEST: 03/29/2025 13:56:51

Diagnostics Check	OK	13:50
Air Blank	0.000	13:50
Control Test	0.081	13:51
Air Blank	0.000	13:51
Subject Sample # 1	0.152	13:52
Breath Volume	3.582	13:52
Air Blank	0.000	13:53
Air Blank	0.000	13:54
Subject Sample # 2	0.151	13:55
Breath Volume	3.277	13:55
Air Blank	0.000	13:55
Control Test	0.080	13:56
Air Blank	0.000	13:56
Diagnostics Check	OK	13:56

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRISON,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: GARRISON,TIMOTHY
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232492
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: A216949909000

DATE/TIME OF TEST: 03/30/2025 01:12:44

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.081	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.097	01:08
Breath Volume	1.753	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.091	01:11
Breath Volume	2.500	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 03/31/2025 10:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN, DAVID, R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCMULLEN, Z
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 03/10/2025 20:31:40

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.081	20:29	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 03/20/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FLOURNOY,WESLEY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/17/2025 09:24:53

Diagnostics Check	OK	09:24	
Air Blank		09:24	Improper Sample
Air Blank	0.000	09:24	

Data Download Date/Time: 03/20/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: FLOURNOY,WESLEY
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/17/2025 09:33:03

Diagnostics Check	OK	09:26
Air Blank	0.000	09:27
Control Test	0.081	09:27
Air Blank	0.000	09:28
Subject Sample # 1	0.000	09:28
Breath Volume	1.441	09:28
Air Blank	0.000	09:29
Air Blank	0.000	09:31
Subject Sample # 2	0.000	09:31
Breath Volume	2.425	09:31
Air Blank	0.000	09:32
Control Test	0.078	09:32
Air Blank	0.000	09:32
Diagnostics Check	OK	09:33

Data Download Date/Time: 03/20/2025 15:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMONE,JOHN,F
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: LEONARD,PRESTON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: TUSING

DATE/TIME OF TEST: 03/01/2025 07:21:53

Diagnostics Check	OK	07:15
Air Blank	0.000	07:15
Control Test	0.081	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.171	07:17
Breath Volume	1.972	07:17
Air Blank	0.000	07:18
Air Blank	0.000	07:19
Subject Sample # 2	0.173	07:20
Breath Volume	1.281	07:20
Air Blank	0.000	07:20
Control Test	0.080	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 03/05/2025 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,A,
BREATH TEST OPERATOR AGENCY: Mascotte PD
ARRESTING OFFICER: SOTO,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 03/01/2025 23:03:31

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.133	22:58
Breath Volume	1.832	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.129	23:01
Breath Volume	1.273	23:01
Air Blank	0.000	23:02
Control Test	0.079	23:02
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 03/05/2025 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURTZLAFF,SYDNEE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 03/04/2025 02:03:40

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 03/05/2025 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,PHILIP,J
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: MEYER,PHILIP
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: RAINFORD

DATE/TIME OF TEST: 03/04/2025 18:53:04

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:48	
Control Test	0.080	18:48	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:51	Subject Test Refused
Breath Volume	0.050	18:51	
Air Blank	0.000	18:52	
Control Test	0.082	18:52	
Air Blank	0.000	18:52	
Diagnostics Check	OK	18:53	

Data Download Date/Time: 03/05/2025 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONDAHL,CARL,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: KRONDAHL,CARL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/04/2025 22:02:23

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.080	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1 0.133	21:55	Volume Not Met
Breath Volume	0.191	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.145	22:00	Volume Not Met
Breath Volume	0.156	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 03/05/2025 09:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/07/2025 00:56:11

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.189	00:50
Breath Volume	2.121	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.169	00:54
Breath Volume	1.593	00:54
Air Blank	0.000	00:55
Control Test	0.083	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/16/2025 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: SNOW,STEPHEN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 03/09/2025 03:00:57

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.080	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 04/16/2025 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARGE,MIRANDA,H
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LARGE,MIRANDA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 03/14/2025 20:13:28

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.080	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:12	RFI Detect
Breath Volume	1.203	20:12	
Air Blank		20:13	RFI Detect
Air Blank	0.000	20:13	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARGE,MIRANDA,H
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: LARGE,MIRANDA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROLLINS

DATE/TIME OF TEST: 03/14/2025 20:21:32

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.068	20:17
Breath Volume	2.496	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.065	20:20
Breath Volume	2.542	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: BRANCH

DATE/TIME OF TEST: 03/15/2025 02:00:02

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.080	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.080	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/16/2025 01:32:33

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.080	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:31	Subject Test Refused
Breath Volume	0.000	01:31	
Air Blank	0.000	01:31	
Control Test	0.080	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALLEN-SWEET,BREANNA
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/16/2025 03:01:12

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.079	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	03:00	
Control Test	0.080	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BARNES,MONIQUE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 03/16/2025 03:21:11

Diagnostics Check	OK	03:18	
Air Blank	0.000	03:18	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Subject Sample	# 1	03:19	Subject Test Refused
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOVER,DAVID,W
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LEONARD,PRESTON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: FAKESS

DATE/TIME OF TEST: 03/20/2025 06:56:25

Diagnostics Check	OK	06:49
Air Blank	0.000	06:50
Control Test	0.080	06:50
Air Blank	0.000	06:50
Subject Sample # 1	0.056	06:51
Breath Volume	1.824	06:51
Air Blank	0.000	06:52
Air Blank	0.000	06:53
Subject Sample # 2	0.050	06:54
Breath Volume	1.835	06:54
Air Blank	0.000	06:55
Control Test	0.083	06:55
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRANSUE,MICHAEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: TRANSUE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: KANDIBOVICH

DATE/TIME OF TEST: 03/23/2025 05:23:03

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.081	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.164	05:18
Breath Volume	2.300	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.154	05:21
Breath Volume	2.058	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:23

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: RAPP,ERIC
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HOLTZMAN

DATE/TIME OF TEST: 03/26/2025 00:31:16

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Diagnostics Check	OK	00:31	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAPP,ERIC,G
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: REGO,BRYAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HOLTZMAN

DATE/TIME OF TEST: 03/26/2025 01:04:50

Diagnostics Check	OK	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank	0.000	00:57	
Subject Sample	# 1	01:00	No Sample Provided
Breath Volume	0.000	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 2	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONEY,RYAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONEY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: BLESSIE

DATE/TIME OF TEST: 03/26/2025 03:25:36

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.014	03:21
Breath Volume	2.867	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.013	03:24
Breath Volume	3.332	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHETTI,KALEIGH,W
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: RICHETTI,KALEIGH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/26/2025 21:44:13

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.073	21:39
Breath Volume	2.578	21:39
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.063	21:42
Breath Volume	1.621	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GARCIA,KEVIN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MONDESIR

DATE/TIME OF TEST: 03/31/2025 01:58:28

Diagnostics Check	OK	01:57	
Air Blank		01:57	RFI Detect
Air Blank	0.000	01:58	

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURTZLAFF,SYDNEE,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: GARCIA,KEVIN
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MONDESIR

DATE/TIME OF TEST: 03/31/2025 02:07:57

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.160	02:02
Breath Volume	1.800	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.154	02:06
Breath Volume	2.593	02:06
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 04/16/2025 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HICKS,TYRECE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: AMACHER

DATE/TIME OF TEST: 03/01/2025 03:52:01

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.390	03:47
Breath Volume	1.550	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.377	03:50
Breath Volume	2.062	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GAUNTLETT

DATE/TIME OF TEST: 03/03/2025 01:16:23

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.000	01:12
Breath Volume	3.148	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.007	01:15
Breath Volume	2.988	01:15
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 03/04/2025 19:52:12

Diagnostics Check	OK	19:45
Air Blank	0.000	19:45
Control Test	0.080	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.209	19:47
Breath Volume	1.660	19:47
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.220	19:50
Breath Volume	2.050	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BERGSTROM JR.

DATE/TIME OF TEST: 03/14/2025 23:51:21

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.212	23:46
Breath Volume	1.902	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.214	23:49
Breath Volume	1.558	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 03/16/2025 19:13:01

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.000	19:08
Breath Volume	2.042	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.000	19:11
Breath Volume	2.390	19:11
Air Blank	0.000	19:12
Control Test	0.081	19:12
Air Blank	0.000	19:12
Diagnostics Check	OK	19:13

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 03/19/2025 21:06:58

Diagnostics Check	OK	21:01
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.000	21:02
Breath Volume	1.222	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.000	21:05
Breath Volume	2.167	21:05
Air Blank	0.000	21:06
Control Test	0.081	21:06
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CRUZ ARCOS

DATE/TIME OF TEST: 03/19/2025 22:35:19

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.198	22:31
Breath Volume	1.531	22:31
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.193	22:33
Breath Volume	2.437	22:33
Air Blank	0.000	22:34
Control Test	0.080	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 03/21/2025 05:18:21

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.179	05:13
Breath Volume	2.699	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.176	05:16
Breath Volume	2.863	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: NEESE

DATE/TIME OF TEST: 03/21/2025 12:22:26

Diagnostics Check	OK	12:15
Air Blank	0.000	12:16
Control Test	0.082	12:16
Air Blank	0.000	12:17
Subject Sample # 1	0.000	12:18
Breath Volume	1.242	12:18
Air Blank	0.000	12:18
Air Blank	0.000	12:20
Subject Sample # 2	0.000	12:21
Breath Volume	1.375	12:21
Air Blank	0.000	12:21
Control Test	0.082	12:21
Air Blank	0.000	12:22
Diagnostics Check	OK	12:22

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 03/21/2025 16:22:43

Diagnostics Check	OK	16:14
Air Blank	0.000	16:15
Control Test	0.081	16:15
Air Blank	0.000	16:15
Subject Sample # 1	0.129	16:16
Breath Volume	1.355	16:16
Air Blank	0.000	16:17
Air Blank	0.000	16:18
Subject Sample # 2	0.126	16:21
Breath Volume	1.781	16:21
Air Blank	0.000	16:21
Control Test	0.081	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:22

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: ALBOGMI

DATE/TIME OF TEST: 03/22/2025 05:15:52

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.081	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.054	05:11
Breath Volume	1.734	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.051	05:14
Breath Volume	1.382	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: MCCracken

DATE/TIME OF TEST: 03/22/2025 18:54:14

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.080	18:48
Air Blank	0.000	18:49
Subject Sample # 1	0.163	18:49
Breath Volume	1.910	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.163	18:52
Breath Volume	1.464	18:52
Air Blank	0.000	18:53
Control Test	0.081	18:53
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEACH,KODY,E
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/22/2025 23:57:40

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	1.652	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.000	23:56
Breath Volume	1.398	23:56
Air Blank	0.000	23:56
Control Test	0.081	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/23/2025 02:25:58

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.183	02:20
Breath Volume	1.144	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.181	02:24
Breath Volume	1.757	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/29/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: BULLARD

DATE/TIME OF TEST: 03/23/2025 06:22:25

Diagnostics Check	OK	06:14
Air Blank	0.000	06:14
Control Test	0.081	06:14
Air Blank	0.000	06:15
Subject Sample # 1	0.187	06:16
Breath Volume	1.835	06:16
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.187	06:20
Breath Volume	1.386	06:20
Air Blank	0.000	06:21
Control Test	0.081	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 03/29/2025 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,MATTHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HALL,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ARELLANO

DATE/TIME OF TEST: 03/24/2025 00:17:36

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.161	00:13
Breath Volume	2.980	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.158	00:16
Breath Volume	1.742	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 03/29/2025 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 03/29/2025 21:04:29

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.000	21:00
Breath Volume	2.074	21:00
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.000	21:03
Breath Volume	2.250	21:03
Air Blank	0.000	21:03
Control Test	0.080	21:03
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 04/05/2025 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 03/30/2025 02:51:40

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:46	
Control Test	0.081	02:46	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.118	02:47	
Breath Volume	2.246	02:47	
Air Blank	0.000	02:48	
Air Blank	0.000	02:50	
Subject Sample	# 2 0.122	02:50	
Breath Volume	2.539	02:50	
Air Blank		02:51	RFI Detect
Air Blank	0.000	02:51	

Data Download Date/Time: 04/05/2025 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 03/30/2025 03:00:42

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.121	02:56
Breath Volume	2.226	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.121	02:59
Breath Volume	2.074	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/05/2025 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SWIFT,MICHAEL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/30/2025 04:17:24

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.155	04:12
Breath Volume	1.574	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.161	04:15
Breath Volume	2.351	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/05/2025 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: HUMPHRYES

DATE/TIME OF TEST: 03/31/2025 21:44:13

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.009	21:39
Breath Volume	1.292	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.008	21:42
Breath Volume	1.296	21:42
Air Blank	0.000	21:43
Control Test	0.081	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 04/05/2025 16:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MILLER,ZACHARY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: STANLEY

DATE/TIME OF TEST: 03/01/2025 23:29:24

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.045	23:24
Breath Volume	1.371	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.044	23:27
Breath Volume	1.320	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY, TYLER, J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CASEY, TYLER
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/03/2025 03:11:00

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.170	03:05
Breath Volume	1.710	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.170	03:09
Breath Volume	2.148	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER,DILLON,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 03/06/2025 00:29:45

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.180	00:24
Breath Volume	2.054	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.171	00:28
Breath Volume	1.761	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GILKERSON

DATE/TIME OF TEST: 03/06/2025 01:19:57

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.000	01:15
Breath Volume	4.113	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.000	01:18
Breath Volume	5.078	01:18
Air Blank	0.000	01:19
Control Test	0.078	01:19
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY, ELISE, N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/12/2025 03:04:03

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.169	02:59
Breath Volume	2.183	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.171	03:02
Breath Volume	1.558	03:02
Air Blank	0.000	03:03
Control Test	0.077	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/15/2025 03:37:24

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.078	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.098	03:30	Volume Not Met
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/15/2025 03:43:20

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.078	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 03/15/2025 19:30:37

Diagnostics Check	OK	19:19	
Air Blank	0.000	19:19	
Control Test	0.078	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1 0.119	19:23	Volume Not Met
Breath Volume	0.355	19:23	
Air Blank	0.000	19:24	
Air Blank	0.000	19:26	
Subject Sample	# 2 0.114	19:29	Volume Not Met
Breath Volume	0.710	19:29	
Air Blank	0.000	19:29	
Control Test	0.079	19:30	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 04/05/2025 22:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CORBIN

DATE/TIME OF TEST: 03/15/2025 19:42:10

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:33	
Control Test	0.077	19:34	
Air Blank	0.000	19:34	
Subject Sample	# 1 0.132	19:35	
Breath Volume	1.328	19:35	
Air Blank	0.000	19:35	
Air Blank	0.000	19:37	
Subject Sample	# 2	19:40	Subject Test Refused
Breath Volume	0.148	19:40	
Air Blank	0.000	19:41	
Control Test	0.078	19:41	
Air Blank	0.000	19:42	
Diagnostics Check	OK	19:42	

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ARGUELLO MOSHAN

DATE/TIME OF TEST: 03/16/2025 01:37:55

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.186	01:33
Breath Volume	1.300	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.192	01:36
Breath Volume	3.011	01:36
Air Blank	0.000	01:37
Control Test	0.078	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/16/2025 03:48:42

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.078	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.000	03:44
Breath Volume	2.984	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.000	03:47
Breath Volume	3.210	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FORTUNE,CALVIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: HIGHFIELD

DATE/TIME OF TEST: 03/16/2025 14:21:06

Diagnostics Check	OK	14:14
Air Blank	0.000	14:15
Control Test	0.079	14:15
Air Blank	0.000	14:15
Subject Sample # 1	0.139	14:16
Breath Volume	1.695	14:16
Air Blank	0.000	14:17
Air Blank	0.000	14:19
Subject Sample # 2	0.154	14:19
Breath Volume	1.542	14:19
Air Blank	0.000	14:20
Control Test	0.078	14:20
Air Blank	0.000	14:21
Diagnostics Check	OK	14:21

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY, ELISE, N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: CADY, ELISE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/18/2025 01:18:13

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.160	01:13
Breath Volume	1.347	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.169	01:16
Breath Volume	1.382	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 03/19/2025 23:23:45

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:19
Breath Volume	4.203	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:22
Breath Volume	3.847	23:22
Air Blank	0.000	23:22
Control Test	0.078	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINGRY,HOBIE,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KINGRY,HOBIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: GOBER

DATE/TIME OF TEST: 03/21/2025 18:14:37

Diagnostics Check	OK	18:08
Air Blank	0.000	18:08
Control Test	0.080	18:09
Air Blank	0.000	18:09
Subject Sample # 1	0.000	18:10
Breath Volume	1.972	18:10
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.000	18:13
Breath Volume	3.000	18:13
Air Blank	0.000	18:13
Control Test	0.079	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:14

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY, ELISE, N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 03/22/2025 04:29:03

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.078	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.159	04:23
Breath Volume	1.351	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.159	04:27
Breath Volume	2.464	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:29

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: ASH

DATE/TIME OF TEST: 03/24/2025 23:19:28

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.201	23:15
Breath Volume	1.277	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.193	23:18
Breath Volume	1.101	23:18
Air Blank	0.000	23:18
Control Test	0.076	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KYLE,F
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BUNCH,BRADFORD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/26/2025 02:48:23

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.086	02:43
Breath Volume	1.425	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.083	02:46
Breath Volume	1.152	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER, TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KATO-REIS

DATE/TIME OF TEST: 03/27/2025 05:05:27

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:54	
Control Test	0.079	04:54	
Air Blank	0.000	04:55	
Subject Sample	# 1 0.166	04:58	Volume Not Met
Breath Volume	0.015	04:58	
Air Blank	0.000	04:58	
Air Blank	0.000	05:00	
Subject Sample	# 2 0.186	05:03	Volume Not Met
Breath Volume	0.199	05:03	
Air Blank	0.000	05:04	
Control Test	0.078	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER,TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KATO REIS

DATE/TIME OF TEST: 03/27/2025 05:17:32

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.183	05:12
Breath Volume	1.503	05:12
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.196	05:16
Breath Volume	2.128	05:16
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALLYN,WILLIAM
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: PROCTOR

DATE/TIME OF TEST: 03/28/2025 01:44:18

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.173	01:39
Breath Volume	1.597	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.166	01:42
Breath Volume	2.445	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 03/28/2025 23:22:10

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.162	23:17
Breath Volume	2.980	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.159	23:20
Breath Volume	3.187	23:20
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER, TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 03/29/2025 23:12:54

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	1.460	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.000	23:11
Breath Volume	2.031	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/30/2025 01:45:04

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.164	01:40
Breath Volume	2.390	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.154	01:43
Breath Volume	1.554	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:45

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 03/30/2025 03:41:45

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.085	03:37
Breath Volume	1.617	03:37
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.089	03:40
Breath Volume	1.656	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAY,JON,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: GAY,JON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 03/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/31/2025 00:53:42

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.078	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.215	00:47
Breath Volume	1.242	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.211	00:52
Breath Volume	1.531	00:52
Air Blank	0.000	00:52
Control Test	0.075	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 04/05/2025 22:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GOMEZ,FRANK
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GAITAN

DATE/TIME OF TEST: 03/01/2025 22:02:29

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.152	21:58
Breath Volume	1.628	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.170	22:00
Breath Volume	1.921	22:00
Air Blank	0.000	22:01
Control Test	0.078	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 03/13/2025 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,R,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ,O
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 03/03/2025 03:07:03

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.187	03:02
Breath Volume	1.675	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.172	03:05
Breath Volume	2.425	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:07

Data Download Date/Time: 03/13/2025 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: HOLCOMBE

DATE/TIME OF TEST: 03/03/2025 22:51:08

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.130	22:46
Breath Volume	1.660	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.137	22:49
Breath Volume	1.390	22:49
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 03/13/2025 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SHUGARS

DATE/TIME OF TEST: 03/04/2025 04:18:05

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.079	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:16	RFI Detect
Breath Volume	1.050	04:16	
Air Blank	0.000	04:17	
Control Test	0.077	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 03/13/2025 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SHUGARS

DATE/TIME OF TEST: 03/04/2025 04:27:06

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.191	04:22
Breath Volume	1.859	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.193	04:25
Breath Volume	1.894	04:25
Air Blank	0.000	04:26
Control Test	0.078	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/13/2025 08:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNISTON,GABRIELLE
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 03/05/2025 03:38:29

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.174	03:33
Breath Volume	1.554	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.187	03:36
Breath Volume	1.210	03:36
Air Blank	0.000	03:37
Control Test	0.077	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANDERS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 03/05/2025 15:37:28

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.080	15:32
Air Blank	0.000	15:32
Subject Sample # 1	0.060	15:33
Breath Volume	2.742	15:33
Air Blank	0.000	15:33
Air Blank	0.000	15:35
Subject Sample # 2	0.057	15:35
Breath Volume	3.234	15:35
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:37
Diagnostics Check	OK	15:37

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,ANDREW,G
BREATH TEST OPERATOR AGENCY: Venice PD
ARRESTING OFFICER: NELSON,ANDREW
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BOYKIN

DATE/TIME OF TEST: 03/06/2025 23:41:18

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.187	23:36
Breath Volume	1.562	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.196	23:39
Breath Volume	2.265	23:39
Air Blank	0.000	23:40
Control Test	0.077	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CARROTO ORDONEZ

DATE/TIME OF TEST: 03/08/2025 03:51:21

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.144	03:46
Breath Volume	2.707	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.149	03:49
Breath Volume	3.156	03:49
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: STURDY

DATE/TIME OF TEST: 03/09/2025 00:40:31

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.155	00:36
Breath Volume	2.285	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.165	00:38
Breath Volume	1.492	00:38
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGAN,KASEY,L
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: QUINONES,MARIA
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BECKHAM

DATE/TIME OF TEST: 03/09/2025 05:28:35

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.079	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.120	05:24
Breath Volume	1.421	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.119	05:27
Breath Volume	1.558	05:27
Air Blank	0.000	05:27
Control Test	0.077	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: GRANTLEY

DATE/TIME OF TEST: 03/10/2025 03:35:25

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.117	03:31
Breath Volume	1.554	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.116	03:33
Breath Volume	1.183	03:33
Air Blank	0.000	03:34
Control Test	0.077	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TILAN,ALEXANDER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: DUQUE CAMACHO

DATE/TIME OF TEST: 03/11/2025 00:34:43

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.289	00:30
Breath Volume	2.792	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.272	00:33
Breath Volume	2.449	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PALACIOS SALAZAR

DATE/TIME OF TEST: 03/13/2025 00:01:48

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.198	23:53
Breath Volume	2.105	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 3	0.222	23:57
Breath Volume	2.144	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.225	00:00
Breath Volume	2.078	00:00
Air Blank	0.000	00:00
Control Test	0.078	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 03/13/2025 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/13/2025 19:42:07

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.196	19:37
Breath Volume	1.433	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.197	19:40
Breath Volume	1.761	19:40
Air Blank	0.000	19:41
Control Test	0.077	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACEY,BRANDON,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STACEY,BRANDON
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/14/2025 01:30:26

Diagnostics Check	OK	01:19	
Air Blank	0.000	01:19	
Control Test	0.079	01:19	
Air Blank	0.000	01:20	
Subject Sample	# 1 0.219	01:23	Volume Not Met
Breath Volume	0.015	01:23	
Air Blank	0.000	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:29	No Sample Provided
Breath Volume	0.000	01:29	
Air Blank	0.000	01:29	
Control Test	0.078	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACEY,BRANDON,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: STACEY,BRANDON
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/14/2025 01:42:29

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.225	01:37
Breath Volume	1.671	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.224	01:40
Breath Volume	2.132	01:40
Air Blank	0.000	01:41
Control Test	0.077	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RIFKIN,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/14/2025 05:12:42

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.079	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,ANDREW,G
BREATH TEST OPERATOR AGENCY: Venice PD
ARRESTING OFFICER: NELSON,ANDREW
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/14/2025 23:28:09

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.058	23:22
Breath Volume	1.519	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.053	23:26
Breath Volume	1.132	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: PORTO

DATE/TIME OF TEST: 03/15/2025 03:37:43

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.189	03:33
Breath Volume	1.371	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.189	03:36
Breath Volume	1.808	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BAKER,D
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/16/2025 01:23:41

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.127	01:19
Breath Volume	1.304	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.136	01:22
Breath Volume	1.664	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ALBA

DATE/TIME OF TEST: 03/18/2025 01:43:24

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.177	01:38
Breath Volume	2.628	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.169	01:41
Breath Volume	2.773	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FUENTES PAIZ

DATE/TIME OF TEST: 03/18/2025 02:29:00

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.155	02:24
Breath Volume	2.675	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.154	02:27
Breath Volume	3.597	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:29

Data Download Date/Time: 04/04/2025 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BUGIS WITCZAK

DATE/TIME OF TEST: 03/20/2025 02:04:48

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.143	02:00
Breath Volume	2.375	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.138	02:03
Breath Volume	2.343	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/04/2025 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTINEZ,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CHRISTMAN

DATE/TIME OF TEST: 03/21/2025 00:20:48

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.222	00:16
Breath Volume	2.480	00:16
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.225	00:19
Breath Volume	2.468	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/04/2025 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TROYER,OFFICER
ARRESTING OFFICER AGENCY: Longboat Key PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: POLLARD

DATE/TIME OF TEST: 03/21/2025 03:03:17

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.041	02:58
Breath Volume	3.324	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.041	03:01
Breath Volume	2.316	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 04/04/2025 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIMITROV,ASEN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: JONES,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BROEKERE

DATE/TIME OF TEST: 03/21/2025 04:24:23

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.116	04:19
Breath Volume	1.773	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.117	04:22
Breath Volume	2.191	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 04/04/2025 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 02/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/01/2025 10:20:39

Diagnostics Check	OK	10:14
Air Blank	0.000	10:15
Control Test	0.083	10:15
Air Blank	0.000	10:15
Subject Sample # 1	0.000	10:16
Breath Volume	2.074	10:16
Air Blank	0.000	10:16
Air Blank	0.000	10:18
Subject Sample # 2	0.000	10:19
Breath Volume	2.378	10:19
Air Blank	0.000	10:19
Control Test	0.083	10:20
Air Blank	0.000	10:20
Diagnostics Check	OK	10:20

Data Download Date/Time: 03/03/2025 16:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/11/2025 16:27:10

Diagnostics Check	OK	16:21
Air Blank	0.000	16:21
Control Test	0.083	16:21
Air Blank	0.000	16:22
Subject Sample # 1	0.000	16:22
Breath Volume	2.210	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:25
Subject Sample # 2	0.000	16:25
Breath Volume	2.199	16:25
Air Blank	0.000	16:26
Control Test	0.083	16:26
Air Blank	0.000	16:27
Diagnostics Check	OK	16:27

Data Download Date/Time: 04/04/2025 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/14/2025 15:40:22

Diagnostics Check	OK	15:34
Air Blank	0.000	15:34
Control Test	0.082	15:35
Air Blank	0.000	15:35
Subject Sample # 1	0.000	15:36
Breath Volume	2.621	15:36
Air Blank	0.000	15:36
Air Blank	0.000	15:38
Subject Sample # 2	0.000	15:39
Breath Volume	2.433	15:39
Air Blank	0.000	15:39
Control Test	0.081	15:39
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 04/04/2025 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/20/2025 17:41:38

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.083	17:36
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:37
Breath Volume	2.351	17:37
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.000	17:40
Breath Volume	2.167	17:40
Air Blank	0.000	17:40
Control Test	0.083	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 04/04/2025 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/25/2025 17:31:54

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.083	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:27
Breath Volume	2.355	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:30
Breath Volume	2.144	17:30
Air Blank	0.000	17:30
Control Test	0.083	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 04/04/2025 15:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

INSTRUMENT REGISTERED TO: BLOUNTSTOWN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 03/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 03/28/2025 14:29:54

Diagnostics Check	OK	14:24
Air Blank	0.000	14:24
Control Test	0.083	14:24
Air Blank	0.000	14:25
Subject Sample # 1	0.000	14:25
Breath Volume	2.863	14:25
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.000	14:28
Breath Volume	2.449	14:28
Air Blank	0.000	14:28
Control Test	0.083	14:29
Air Blank	0.000	14:29
Diagnostics Check	OK	14:29

Data Download Date/Time: 04/04/2025 15:33