

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ROWELL,J
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ZNOJ

DATE/TIME OF TEST: 03/04/2025 22:28:28

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.082	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1 0.223	22:18	Volume Not Met
Breath Volume	0.289	22:18	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2 0.214	22:22	
Breath Volume	2.621	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 3 0.204	22:26	
Breath Volume	2.265	22:26	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 03/14/2025 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,SHEHARA,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SCHALLON,C
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 02/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 03/08/2025 22:02:19

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.082	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1 0.112	21:54	Volume Not Met
Breath Volume	0.433	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2 0.113	21:57	
Breath Volume	1.863	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 3 0.112	22:00	
Breath Volume	1.753	22:00	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 03/14/2025 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ORTEGA ID57,IRENE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 02/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/13/2025 16:25:15

Diagnostics Check	OK	16:13	
Air Blank	0.000	16:14	
Control Test	0.082	16:14	
Air Blank	0.000	16:14	
Subject Sample	# 1 0.253	16:17	Volume Not Met
Breath Volume	0.414	16:17	
Air Blank	0.000	16:18	
Air Blank	0.000	16:20	
Subject Sample	# 2 0.256	16:23	Volume Not Met
Breath Volume	0.023	16:23	
Air Blank	0.000	16:24	
Control Test	0.081	16:24	
Air Blank	0.000	16:25	
Diagnostics Check	OK	16:25	

Data Download Date/Time: 03/14/2025 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: WEDLOWE,SHAWN
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 02/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WINSTANLEY

DATE/TIME OF TEST: 03/14/2025 01:11:39

Diagnostics Check	OK	01:08	
Air Blank	0.000	01:09	
Control Test	0.082	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:10	
Control Test	0.082	01:11	
Air Blank	0.000	01:11	
Diagnostics Check	OK	01:11	

Data Download Date/Time: 03/14/2025 21:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NIMMO,O
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 03/22/2025 01:44:40

Diagnostics Check	OK	01:44	
Air Blank		01:44	RFI Detect
Air Blank	0.000	01:44	

Data Download Date/Time: 04/11/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MODERSON,C,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NIMMO,O
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 03/22/2025 01:50:28

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.082	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.083	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 04/11/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,SHEHARA,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,KYLE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: LAVOXY

DATE/TIME OF TEST: 03/28/2025 00:39:34

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.155	00:34
Breath Volume	1.457	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.156	00:37
Breath Volume	1.765	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 04/11/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CORBLEY,LIAM
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 03/30/2025 03:21:15

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.000	03:17
Breath Volume	1.460	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:19
Breath Volume	2.054	03:19
Air Blank	0.000	03:20
Control Test	0.081	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/11/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 03/02/2025 03:16:01

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 2	03:14	No Sample Provided
Breath Volume	0.000	03:14	
Air Blank	0.000	03:15	
Control Test	0.077	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 03/18/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MOFFETT,JACKSON
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 03/02/2025 04:56:15

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.077	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.078	04:51
Breath Volume	1.429	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.080	04:54
Breath Volume	1.324	04:54
Air Blank	0.000	04:55
Control Test	0.076	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 03/18/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KANE,NOAH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 09:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/09/2025 09:51:18

Diagnostics Check	OK	09:44
Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:45
Subject Sample # 1	0.196	09:46
Breath Volume	1.492	09:46
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.204	09:49
Breath Volume	1.917	09:49
Air Blank	0.000	09:50
Control Test	0.076	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 03/18/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DIERDORFF

DATE/TIME OF TEST: 03/16/2025 03:10:59

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.077	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.176	03:06
Breath Volume	2.300	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.174	03:09
Breath Volume	2.574	03:09
Air Blank	0.000	03:10
Control Test	0.076	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 03/18/2025 16:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,EMMA
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 03/19/2025 02:55:46

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.159	02:50
Breath Volume	1.460	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.161	02:54
Breath Volume	1.113	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 04/20/2025 19:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 03/20/2025 02:40:04

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.234	02:35
Breath Volume	1.621	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.231	02:38
Breath Volume	1.542	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/20/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CLENDENIN

DATE/TIME OF TEST: 03/22/2025 02:57:07

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.106	02:52
Breath Volume	2.417	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.099	02:55
Breath Volume	2.214	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 04/20/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LAYNE,DALLAS
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 13:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:11

SUBJECT'S LAST NAME: THIES

DATE/TIME OF TEST: 03/26/2025 13:57:53

Diagnostics Check	OK	13:51
Air Blank	0.000	13:51
Control Test	0.081	13:52
Air Blank	0.000	13:52
Subject Sample # 1	0.000	13:53
Breath Volume	2.652	13:53
Air Blank	0.000	13:53
Air Blank	0.000	13:55
Subject Sample # 2	0.000	13:56
Breath Volume	2.949	13:56
Air Blank	0.000	13:56
Control Test	0.080	13:57
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 04/20/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DEMETRIUS,L
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BURELL,HNYLA
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 03/28/2025 21:48:46

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.148	21:42
Breath Volume	1.464	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.144	21:47
Breath Volume	1.683	21:47
Air Blank	0.000	21:47
Control Test	0.078	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 04/20/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HANNIS

DATE/TIME OF TEST: 03/28/2025 22:59:17

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.058	22:54
Breath Volume	2.984	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.056	22:57
Breath Volume	1.699	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 04/20/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HART,LYLE
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 03/02/2025 21:51:39

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.077	21:47	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:50	Subject Test Refused
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Control Test	0.079	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 03/18/2025 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SORRELL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FURTON

DATE/TIME OF TEST: 03/06/2025 00:44:49

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.078	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.000	00:40
Breath Volume	1.945	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.000	00:43
Breath Volume	1.605	00:43
Air Blank	0.000	00:43
Control Test	0.077	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/18/2025 03:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,REX,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: HARNED

DATE/TIME OF TEST: 03/12/2025 17:30:39

Diagnostics Check	OK	17:24
Air Blank	0.000	17:24
Control Test	0.078	17:24
Air Blank	0.000	17:25
Subject Sample # 1	0.158	17:26
Breath Volume	3.242	17:26
Air Blank	0.000	17:26
Air Blank	0.000	17:28
Subject Sample # 2	0.151	17:29
Breath Volume	2.875	17:29
Air Blank	0.000	17:29
Control Test	0.077	17:30
Air Blank	0.000	17:30
Diagnostics Check	OK	17:30

Data Download Date/Time: 03/18/2025 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MELVIN,F
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WIGGINS,ROBERT
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 02/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 14:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: CIMINO

DATE/TIME OF TEST: 03/16/2025 16:01:17

Diagnostics Check	OK	15:54
Air Blank	0.000	15:55
Control Test	0.078	15:55
Air Blank	0.000	15:56
Subject Sample # 1	0.000	15:56
Breath Volume	1.437	15:56
Air Blank	0.000	15:57
Air Blank	0.000	15:59
Subject Sample # 2	0.000	15:59
Breath Volume	1.589	15:59
Air Blank	0.000	16:00
Control Test	0.078	16:00
Air Blank	0.000	16:01
Diagnostics Check	OK	16:01

Data Download Date/Time: 03/18/2025 03:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 03/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VELASCO

DATE/TIME OF TEST: 03/18/2025 23:53:07

Diagnostics Check	OK	23:50	
Air Blank	0.000	23:50	
Control Test	0.078	23:51	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:52	
Control Test	0.078	23:52	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 04/29/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,DANIEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DORR

DATE/TIME OF TEST: 03/19/2025 22:55:54

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.082	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.000	22:51
Breath Volume	1.429	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.000	22:54
Breath Volume	1.480	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 03/20/2025 00:30:24

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.082	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.000	00:26
Breath Volume	2.425	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.000	00:29
Breath Volume	1.464	00:29
Air Blank	0.000	00:29
Control Test	0.082	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUDAS,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: D320173656000

DATE/TIME OF TEST: 03/20/2025 04:31:50

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:16	
Control Test	0.081	04:17	
Air Blank	0.000	04:17	
Subject Sample	# 1 0.161	04:19	
Breath Volume	1.386	04:19	
Air Blank	0.000	04:19	
Air Blank	0.000	04:21	
Subject Sample	# 2 0.146	04:24	Volume Not Met
Breath Volume	0.082	04:24	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 3 0.121	04:30	Volume Not Met
Breath Volume	0.062	04:30	
Air Blank	0.000	04:30	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,K,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DUDAS

DATE/TIME OF TEST: 03/20/2025 04:38:57

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.081	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.081	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/22/2025 02:17:48

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.019	02:13
Breath Volume	1.527	02:13
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.019	02:16
Breath Volume	1.726	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EULER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RUSKIN

DATE/TIME OF TEST: 03/22/2025 03:09:12

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.082	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:08	
Control Test	0.082	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEARGIN, GENESIS,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: YEARGIN, GENESIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: TORTORA

DATE/TIME OF TEST: 03/22/2025 04:27:15

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.153	04:22
Breath Volume	1.777	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.152	04:25
Breath Volume	1.605	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 03/22/2025 05:20:10

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.082	05:18	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.082	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,BRYAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HOPPE,AUSTIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: BUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BANGERTER

DATE/TIME OF TEST: 03/22/2025 17:19:23

Diagnostics Check	OK	17:16	
Air Blank	0.000	17:17	
Control Test	0.082	17:17	
Air Blank	0.000	17:17	
Subject Sample	# 1	17:18	Subject Test Refused
Breath Volume	0.000	17:18	
Air Blank	0.000	17:18	
Control Test	0.082	17:18	
Air Blank	0.000	17:19	
Diagnostics Check	OK	17:19	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: KINNEY WILDEMANN

DATE/TIME OF TEST: 03/23/2025 01:14:29

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.083	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.000	01:09
Breath Volume	1.281	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.000	01:13
Breath Volume	1.332	01:13
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: KLICH

DATE/TIME OF TEST: 03/23/2025 03:22:37

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.067	03:18
Breath Volume	1.750	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.063	03:21
Breath Volume	1.660	03:21
Air Blank	0.000	03:21
Control Test	0.082	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGANARO,JESSICA,H
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MANGANARO,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 03/23/2025 14:27:05

Diagnostics Check	OK	14:20
Air Blank	0.000	14:21
Control Test	0.082	14:21
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:22
Breath Volume	2.667	14:22
Air Blank	0.000	14:23
Air Blank	0.000	14:25
Subject Sample # 2	0.000	14:25
Breath Volume	2.953	14:25
Air Blank	0.000	14:26
Control Test	0.082	14:26
Air Blank	0.000	14:26
Diagnostics Check	OK	14:27

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANNISTER,STEPHEN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BANNISTER,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ORDUNA REINAH

DATE/TIME OF TEST: 03/23/2025 21:33:40

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.082	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.126	21:29
Breath Volume	2.601	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.119	21:32
Breath Volume	1.875	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIE,M,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUIRGUIS,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/24/2025 17:12:34

Diagnostics Check	OK	17:01	
Air Blank	0.000	17:01	
Control Test	0.082	17:01	
Air Blank	0.000	17:02	
Subject Sample	# 1	17:05	No Sample Provided
Breath Volume	0.000	17:05	
Air Blank	0.000	17:06	
Air Blank	0.000	17:08	
Subject Sample	# 2	17:11	No Sample Provided
Breath Volume	0.000	17:11	
Air Blank	0.000	17:11	
Control Test	0.081	17:11	
Air Blank	0.000	17:12	
Diagnostics Check	OK	17:12	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MATTALIANO

DATE/TIME OF TEST: 03/25/2025 05:20:50

Diagnostics Check	OK	05:06	
Air Blank	0.000	05:07	
Control Test	0.081	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1 0.389	05:11	No .020 agreement
Breath Volume	3.218	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 2 0.338	05:15	No .020 agreement
Breath Volume	1.125	05:15	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 3 0.365	05:19	No .020 agreement
Breath Volume	1.546	05:19	
Air Blank	0.000	05:19	
Control Test	0.079	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAPATA ROLON,F,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MATTALIANO

DATE/TIME OF TEST: 03/25/2025 05:39:47

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.327	05:32
Breath Volume	1.394	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.357	05:35
Breath Volume	1.402	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 3	0.343	05:38
Breath Volume	1.378	05:38
Air Blank	0.000	05:38
Control Test	0.077	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DRUMM,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: AGOLLI

DATE/TIME OF TEST: 03/25/2025 22:57:50

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.082	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.212	22:53
Breath Volume	2.074	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.213	22:56
Breath Volume	1.671	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRCHGRABER,DANIELLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/27/2025 03:30:15

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.181	03:25
Breath Volume	1.386	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.180	03:28
Breath Volume	1.191	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE, ELIZABETH, R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON, MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: VULPIS

DATE/TIME OF TEST: 03/27/2025 23:10:21

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.082	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.288	23:05
Breath Volume	1.167	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.281	23:08
Breath Volume	1.425	23:08
Air Blank	0.000	23:09
Control Test	0.082	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACHADO, LAURA, V
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MACHADO, LAURA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PARNELL

DATE/TIME OF TEST: 03/28/2025 00:37:34

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.082	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.230	00:33
Breath Volume	1.355	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.227	00:36
Breath Volume	2.621	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: EXTINE,ALYSON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 03/28/2025 01:22:24

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:13	
Control Test	0.082	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1 0.182	01:17	Volume Not Met
Breath Volume	0.894	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:20	
Subject Sample	# 2	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:21	
Control Test	0.081	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 04/23/2025 09:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BARRERAS_FLEITES

DATE/TIME OF TEST: 03/28/2025 03:05:26

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.231	03:00
Breath Volume	3.011	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.243	03:03
Breath Volume	2.761	03:03
Air Blank	0.000	03:04
Control Test	0.082	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAMMOCK,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 03/29/2025 00:59:19

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.083	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.181	00:54
Breath Volume	1.105	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.193	00:57
Breath Volume	1.648	00:57
Air Blank	0.000	00:58
Control Test	0.082	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: RODRIGUEZ MORALES

DATE/TIME OF TEST: 03/29/2025 04:32:32

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.082	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.189	04:27
Breath Volume	1.933	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.177	04:31
Breath Volume	1.257	04:31
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGHERTY,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/29/2025 05:50:09

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.082	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.163	05:45
Breath Volume	1.605	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.155	05:48
Breath Volume	1.433	05:48
Air Blank	0.000	05:49
Control Test	0.082	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLLMAN,IAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: GILLETTE

DATE/TIME OF TEST: 03/29/2025 18:25:54

Diagnostics Check	OK	18:19
Air Blank	0.000	18:20
Control Test	0.081	18:20
Air Blank	0.000	18:20
Subject Sample # 1	0.148	18:21
Breath Volume	1.175	18:21
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.146	18:24
Breath Volume	1.445	18:24
Air Blank	0.000	18:25
Control Test	0.081	18:25
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KIRKENDALL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 03/30/2025 04:42:34

Diagnostics Check	OK	04:36
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.135	04:38
Breath Volume	2.074	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.138	04:41
Breath Volume	2.476	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONZON GARCIA,ADEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONZON GARCIA,ADEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: ESCALANTE ARROYO

DATE/TIME OF TEST: 03/30/2025 06:57:58

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.081	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.203	06:52
Breath Volume	1.488	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:55
Subject Sample # 2	0.201	06:56
Breath Volume	1.812	06:56
Air Blank	0.000	06:57
Control Test	0.082	06:57
Air Blank	0.000	06:57
Diagnostics Check	OK	06:57

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,MARK,O
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCCALL,JOSH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: VANDERBILT

DATE/TIME OF TEST: 03/30/2025 14:12:41

Diagnostics Check	OK	14:09	
Air Blank	0.000	14:10	
Control Test	0.082	14:10	
Air Blank	0.000	14:11	
Subject Sample	# 1	14:11	Subject Test Refused
Breath Volume	0.000	14:11	
Air Blank	0.000	14:11	
Control Test	0.082	14:12	
Air Blank	0.000	14:12	
Diagnostics Check	OK	14:12	

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TRADOR,ZACHARY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: HROBAT

DATE/TIME OF TEST: 03/30/2025 18:31:40

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.081	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.115	18:27
Breath Volume	1.375	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.109	18:30
Breath Volume	1.250	18:30
Air Blank	0.000	18:30
Control Test	0.082	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROSS,JILLIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/31/2025 03:56:13

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.000	03:52
Breath Volume	2.250	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:54
Breath Volume	3.109	03:54
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAINT_FLEUR,MILOUSE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: KARCHESKI

DATE/TIME OF TEST: 03/31/2025 20:25:06

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.081	20:22	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:23	Subject Test Refused
Breath Volume	0.000	20:23	
Air Blank	0.000	20:24	
Control Test	0.082	20:24	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 04/23/2025 09:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,DEP,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRENNAN,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: KUMATSKY

DATE/TIME OF TEST: 03/20/2025 00:50:08

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:47	
Control Test	0.079	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: DIEKEMA

DATE/TIME OF TEST: 03/20/2025 05:31:43

Diagnostics Check	OK	05:25
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Subject Sample # 1	0.148	05:27
Breath Volume	1.265	05:27
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.148	05:30
Breath Volume	2.750	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEISER,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GEISER,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FUGAH

DATE/TIME OF TEST: 03/22/2025 01:37:31

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.221	01:33
Breath Volume	1.769	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.229	01:36
Breath Volume	1.460	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: BRAND

DATE/TIME OF TEST: 03/23/2025 19:15:30

Diagnostics Check	OK	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Subject Sample # 1	0.178	19:11
Breath Volume	2.027	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.172	19:14
Breath Volume	2.125	19:14
Air Blank	0.000	19:14
Control Test	0.079	19:14
Air Blank	0.000	19:15
Diagnostics Check	OK	19:15

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLEY,DAVID,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FOLEY,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ORTA BARINES

DATE/TIME OF TEST: 03/30/2025 04:03:31

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.123	03:59
Breath Volume	1.925	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.122	04:02
Breath Volume	1.199	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALAZZOLO,DEP,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PAVLOVIC,S
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: KRISTMUNDSOTTIR

DATE/TIME OF TEST: 03/31/2025 17:07:33

Diagnostics Check	OK	17:01
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:02
Subject Sample # 1	0.267	17:03
Breath Volume	1.718	17:03
Air Blank	0.000	17:03
Air Blank	0.000	17:05
Subject Sample # 2	0.276	17:06
Breath Volume	1.640	17:06
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:07
Diagnostics Check	OK	17:07

Data Download Date/Time: 04/23/2025 10:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMERI, DOMINIC, J
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: MORROW, ADAM
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/22/2025 19:18:56

Diagnostics Check	OK	19:16	
Air Blank	0.000	19:16	
Control Test	0.080	19:16	
Air Blank	0.000	19:17	
Subject Sample	# 1	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:18	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 04/23/2025 02:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CAMPAGNONE, ALEXANDER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: LEMAJEUR

DATE/TIME OF TEST: 03/01/2025 23:12:58

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.000	23:08
Breath Volume	3.265	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.000	23:11
Breath Volume	2.382	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNEL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RICHARDS, JAIDON
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: SALYER

DATE/TIME OF TEST: 03/02/2025 02:51:18

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Subject Sample	# 1 0.157	02:40	No .020 agreement
Breath Volume	1.425	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2 0.127	02:43	No .020 agreement
Breath Volume	1.117	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample	# 3 0.103	02:49	Volume Not Met
Breath Volume	0.519	02:49	
Air Blank	0.000	02:50	
Control Test	0.077	02:50	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RICHARDS, JAIDON
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FANCHER

DATE/TIME OF TEST: 03/06/2025 03:57:17

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.158	03:51
Breath Volume	1.589	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.154	03:55
Breath Volume	1.472	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: LUCAS

DATE/TIME OF TEST: 03/12/2025 02:15:20

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.081	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.081	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: GRIFFIN, CARY
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 03/16/2025 23:29:23

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.082	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.187	23:22
Breath Volume	2.683	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.170	23:27
Breath Volume	2.453	23:27
Air Blank	0.000	23:28
Control Test	0.080	23:28
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER, THOMAS, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CANTEES, J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 03/19/2025 00:19:31

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.081	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.113	00:15
Breath Volume	1.957	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.116	00:17
Breath Volume	1.511	00:17
Air Blank	0.000	00:18
Control Test	0.080	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: GOLDSTEIN,SHELLI
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: RODRIGUEZ URENA

DATE/TIME OF TEST: 03/19/2025 00:43:51

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:42	Subject Test Refused
Breath Volume	0.000	00:42	
Air Blank	0.000	00:42	
Control Test	0.081	00:43	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPLAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: KAPLAN,MICHAEL
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 03/19/2025 20:29:21

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.081	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.116	20:24
Breath Volume	2.011	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.116	20:27
Breath Volume	1.882	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 03/27/2025 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: MIXON,GERAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/31/2025 12:27:36

Diagnostics Check	OK	12:21
Air Blank	0.000	12:21
Control Test	0.081	12:22
Air Blank	0.000	12:22
Subject Sample # 1	0.010	12:23
Breath Volume	3.964	12:23
Air Blank	0.000	12:23
Air Blank	0.000	12:25
Subject Sample # 2	0.009	12:26
Breath Volume	4.332	12:26
Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 04/19/2025 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMMER,MIKE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 03/07/2025 17:00:46

Diagnostics Check	OK	16:52
Air Blank	0.000	16:53
Control Test	0.080	16:53
Air Blank	0.000	16:54
Subject Sample # 1	0.266	16:55
Breath Volume	2.152	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.249	16:58
Breath Volume	1.218	16:58
Air Blank	0.000	16:59
Control Test	0.080	17:00
Air Blank	0.000	17:00
Diagnostics Check	OK	17:00

Data Download Date/Time: 03/27/2025 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SANCHEZ LUCAS,HIBER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: NAVARIJO RAMIREZ

DATE/TIME OF TEST: 03/01/2025 20:15:49

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.081	20:09
Air Blank	0.000	20:09
Subject Sample # 1	0.239	20:11
Breath Volume	2.050	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.232	20:14
Breath Volume	2.031	20:14
Air Blank	0.000	20:14
Control Test	0.080	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: LEDOUX,JACOB
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BODGER

DATE/TIME OF TEST: 03/02/2025 01:15:25

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.082	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.172	01:10
Breath Volume	2.117	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.173	01:13
Breath Volume	2.050	01:13
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HULL, PAUL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:28

SUBJECT'S LAST NAME: GORGEN

DATE/TIME OF TEST: 03/02/2025 14:48:01

Diagnostics Check	OK	14:41
Air Blank	0.000	14:42
Control Test	0.082	14:42
Air Blank	0.000	14:43
Subject Sample # 1	0.101	14:43
Breath Volume	1.835	14:43
Air Blank	0.000	14:44
Air Blank	0.000	14:45
Subject Sample # 2	0.094	14:46
Breath Volume	2.039	14:46
Air Blank	0.000	14:47
Control Test	0.081	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:48

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: DUMMER,L
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SENG

DATE/TIME OF TEST: 03/03/2025 23:54:38

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.082	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.228	23:50
Breath Volume	1.660	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.231	23:53
Breath Volume	1.312	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HENRY,A
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/10/2025 02:32:21

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.082	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1 0.239	02:24	Volume Not Met
Breath Volume	0.527	02:24	
Air Blank	0.000	02:25	
Air Blank	0.000	02:27	
Subject Sample	# 2 0.256	02:30	Volume Not Met
Breath Volume	0.507	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 03/14/2025 04:42:52

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.081	04:31	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.100	04:35	Volume Not Met
Breath Volume	0.039	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2 0.115	04:41	Volume Not Met
Breath Volume	0.054	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: NAUSS

DATE/TIME OF TEST: 03/17/2025 22:06:01

Diagnostics Check	OK	21:59
Air Blank	0.000	22:00
Control Test	0.082	22:00
Air Blank	0.000	22:01
Subject Sample # 1	0.000	22:01
Breath Volume	2.250	22:01
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.000	22:04
Breath Volume	2.125	22:04
Air Blank	0.000	22:05
Control Test	0.082	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:06

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: JENKINSON,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 03/22/2025 02:37:29

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.083	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.000	02:32
Breath Volume	1.523	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.000	02:36
Breath Volume	1.652	02:36
Air Blank	0.000	02:36
Control Test	0.081	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: DOMINICIS,LOURDES
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LUBIN

DATE/TIME OF TEST: 03/22/2025 04:39:58

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.082	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1 0.224	04:35	Volume Not Met
Breath Volume	0.082	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:38	Subject Test Refused
Breath Volume	0.000	04:38	
Air Blank	0.000	04:39	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: NIGH,ALEXANDER
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BLUMFIELD

DATE/TIME OF TEST: 03/23/2025 02:09:56

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.083	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:05
Breath Volume	1.417	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	1.324	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 03/27/2025 14:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HACKETT,BRADLEY
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/23/2025 02:24:05

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.082	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.191	02:19
Breath Volume	1.488	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:21
Subject Sample # 2	0.191	02:22
Breath Volume	1.582	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/27/2025 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: CHRISTMAN, XAVIER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 03/25/2025 09:16:25

Diagnostics Check	OK	09:09
Air Blank	0.000	09:09
Control Test	0.081	09:10
Air Blank	0.000	09:10
Subject Sample # 1	0.190	09:11
Breath Volume	3.621	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:13
Subject Sample # 2	0.193	09:14
Breath Volume	3.746	09:14
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 03/27/2025 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FEDRO,DARIEN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 03/29/2025 21:25:17

Diagnostics Check	OK	21:18
Air Blank	0.000	21:18
Control Test	0.083	21:18
Air Blank	0.000	21:19
Subject Sample # 1	0.159	21:20
Breath Volume	1.632	21:20
Air Blank	0.000	21:21
Air Blank	0.000	21:22
Subject Sample # 2	0.163	21:23
Breath Volume	1.210	21:23
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 04/16/2025 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PEREZ USECHE

DATE/TIME OF TEST: 03/05/2025 23:56:59

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.149	23:52
Breath Volume	2.757	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.143	23:55
Breath Volume	2.445	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 03/26/2025 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,DONALD,P
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: LUKIANOV,ALEXANDER
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: CHIABRERA

DATE/TIME OF TEST: 03/12/2025 01:21:09

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.110	01:14
Breath Volume	1.304	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.103	01:19
Breath Volume	1.511	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 03/26/2025 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: FAGUNDESCAMPOS

DATE/TIME OF TEST: 03/26/2025 21:24:40

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:22	
Control Test	0.080	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.081	21:24	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 04/28/2025 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GETZ,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CHISMAR

DATE/TIME OF TEST: 03/01/2025 00:02:44

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.083	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.176	23:57
Breath Volume	1.253	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.191	00:00
Breath Volume	2.511	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 03/01/2025 03:11:57

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.082	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.238	03:07
Breath Volume	2.613	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.237	03:10
Breath Volume	2.894	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: VIRGO

DATE/TIME OF TEST: 03/01/2025 21:39:15

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.082	21:33
Air Blank	0.000	21:33
Subject Sample # 1	0.185	21:34
Breath Volume	3.093	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:36
Subject Sample # 2	0.183	21:37
Breath Volume	3.351	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SLOAN

DATE/TIME OF TEST: 03/02/2025 01:50:21

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.082	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.161	01:45
Breath Volume	3.250	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.154	01:48
Breath Volume	2.574	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LAKKIREDDY

DATE/TIME OF TEST: 03/02/2025 03:13:14

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.094	03:08
Breath Volume	2.191	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.089	03:11
Breath Volume	1.789	03:11
Air Blank	0.000	03:12
Control Test	0.078	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/03/2025 00:05:05

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.075	00:00
Breath Volume	1.703	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:02
Subject Sample # 2	0.074	00:03
Breath Volume	1.851	00:03
Air Blank	0.000	00:04
Control Test	0.081	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:05

Data Download Date/Time: 03/14/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: LOPEZ CARABAJO

DATE/TIME OF TEST: 03/03/2025 02:52:04

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.082	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.167	02:47
Breath Volume	1.542	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.173	02:50
Breath Volume	1.968	02:50
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 03/03/2025 04:01:34

Diagnostics Check	OK	03:55
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.043	03:57
Breath Volume	2.242	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.041	03:59
Breath Volume	2.300	03:59
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: ATTEO

DATE/TIME OF TEST: 03/04/2025 02:46:41

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:44	
Control Test	0.082	02:44	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.000	02:45	
Air Blank	0.000	02:45	
Control Test	0.081	02:46	
Air Blank	0.000	02:46	
Diagnostics Check	OK	02:46	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARIN,O
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: RILES

DATE/TIME OF TEST: 03/04/2025 03:44:21

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 03/06/2025 00:44:45

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.082	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.118	00:40
Breath Volume	2.625	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.111	00:43
Breath Volume	2.093	00:43
Air Blank	0.000	00:43
Control Test	0.080	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: TOMLIN

DATE/TIME OF TEST: 03/06/2025 21:12:31

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.082	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.254	21:07
Breath Volume	1.917	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.254	21:10
Breath Volume	1.613	21:10
Air Blank	0.000	21:11
Control Test	0.079	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SHERIFF

DATE/TIME OF TEST: 03/08/2025 02:08:45

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.082	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.101	02:03
Breath Volume	1.328	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.098	02:07
Breath Volume	1.199	02:07
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUGLAS,JAVONTAE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 03/08/2025 03:38:32

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.082	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,SANTUAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRUBOR,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/08/2025 05:31:21

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.081	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.145	05:26
Breath Volume	1.746	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.147	05:29
Breath Volume	1.875	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPENCER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:29

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/08/2025 08:56:12

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.082	08:49
Air Blank	0.000	08:50
Subject Sample # 1	0.126	08:51
Breath Volume	2.496	08:51
Air Blank	0.000	08:52
Air Blank	0.000	08:53
Subject Sample # 2	0.121	08:54
Breath Volume	2.003	08:54
Air Blank	0.000	08:55
Control Test	0.079	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: KAZMAR

DATE/TIME OF TEST: 03/08/2025 17:14:37

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:12	
Control Test	0.082	17:12	
Air Blank	0.000	17:13	
Subject Sample	# 1	17:13	Subject Test Refused
Breath Volume	0.000	17:13	
Air Blank	0.000	17:13	
Control Test	0.081	17:13	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLADDEN,A,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/08/2025 17:46:56

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:44	
Control Test	0.082	17:44	
Air Blank	0.000	17:45	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:46	
Control Test	0.081	17:46	
Air Blank	0.000	17:46	
Diagnostics Check	OK	17:46	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RAMIREZ SALES

DATE/TIME OF TEST: 03/12/2025 02:13:56

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.082	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:13	
Control Test	0.081	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: PIKE

DATE/TIME OF TEST: 03/12/2025 05:38:26

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.082	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.046	05:33
Breath Volume	1.546	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.039	05:36
Breath Volume	1.171	05:36
Air Blank	0.000	05:37
Control Test	0.081	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MATTIA

DATE/TIME OF TEST: 03/12/2025 21:49:24

Diagnostics Check	OK	21:42
Air Blank	0.000	21:43
Control Test	0.083	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.103	21:44
Breath Volume	1.214	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.103	21:47
Breath Volume	1.253	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: LEDDY

DATE/TIME OF TEST: 03/13/2025 02:51:57

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.082	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.161	02:46
Breath Volume	2.738	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.154	02:50
Breath Volume	3.093	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: IBRAHIM

DATE/TIME OF TEST: 03/14/2025 03:00:38

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.082	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.092	02:50	Volume Not Met
Breath Volume	0.210	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 2 0.088	02:55	
Breath Volume	1.464	02:55	
Air Blank	0.000	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 3 0.083	02:58	
Breath Volume	1.128	02:58	
Air Blank	0.000	02:59	
Control Test	0.077	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 03/14/2025 09:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: NESBITT

DATE/TIME OF TEST: 03/15/2025 00:54:12

Diagnostics Check	OK	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:43
Air Blank	0.000	00:44
Subject Sample # 1	0.297	00:45
Breath Volume	1.230	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.318	00:49
Breath Volume	2.351	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 3	0.307	00:52
Breath Volume	2.324	00:52
Air Blank	0.000	00:53
Control Test	0.077	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 04/15/2025 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/15/2025 03:11:24

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.082	03:09	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 03/15/2025 03:59:09

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.112	03:54
Breath Volume	1.613	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.117	03:57
Breath Volume	1.777	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMSON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/15/2025 05:03:47

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.082	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:02	Subject Test Refused
Breath Volume	0.000	05:02	
Air Blank	0.000	05:02	
Control Test	0.081	05:03	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYA,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 03/16/2025 03:17:09

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.082	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.187	03:12
Breath Volume	3.066	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.181	03:15
Breath Volume	3.910	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,D
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 03/16/2025 23:25:14

Diagnostics Check	OK	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.197	23:20
Breath Volume	1.558	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.190	23:23
Breath Volume	1.242	23:23
Air Blank	0.000	23:24
Control Test	0.078	23:24
Air Blank	0.000	23:25
Diagnostics Check	OK	23:25

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MARROQUIN

DATE/TIME OF TEST: 03/17/2025 00:27:55

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.082	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:26	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: TREW

DATE/TIME OF TEST: 03/17/2025 01:02:44

Diagnostics Check	OK	00:59	
Air Blank	0.000	01:00	
Control Test	0.082	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:01	
Control Test	0.081	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HETHERINGTON

DATE/TIME OF TEST: 03/17/2025 05:17:52

Diagnostics Check	OK	05:10
Air Blank	0.000	05:11
Control Test	0.082	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.125	05:13
Breath Volume	1.812	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.114	05:16
Breath Volume	1.242	05:16
Air Blank	0.000	05:16
Control Test	0.078	05:17
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 03/18/2025 00:15:58

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.082	00:13	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.082	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:15	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 03/18/2025 01:38:32

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.082	01:36	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.081	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOD,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/18/2025 03:30:14

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.082	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.181	03:25
Breath Volume	1.617	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.184	03:28
Breath Volume	2.289	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSSI,S
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 03/18/2025 04:35:21

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.161	04:30
Breath Volume	1.449	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.166	04:33
Breath Volume	1.605	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 03/18/2025 21:20:04

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.082	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:15
Breath Volume	1.496	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.000	21:18
Breath Volume	1.683	21:18
Air Blank	0.000	21:19
Control Test	0.082	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/18/2025 23:06:57

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.082	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Control Test	0.081	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/19/2025 03:23:35

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.081	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.076	03:18
Breath Volume	1.218	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.079	03:21
Breath Volume	1.492	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALSH,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/19/2025 05:03:54

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.062	04:59
Breath Volume	1.824	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.061	05:02
Breath Volume	1.539	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIRRI,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: CAZAREZ

DATE/TIME OF TEST: 03/19/2025 11:17:01

Diagnostics Check	OK	11:14	
Air Blank	0.000	11:14	
Control Test	0.082	11:14	
Air Blank	0.000	11:15	
Subject Sample	# 1	11:15	Subject Test Refused
Breath Volume	0.000	11:15	
Air Blank	0.000	11:16	
Control Test	0.081	11:16	
Air Blank	0.000	11:16	
Diagnostics Check	OK	11:17	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 03/19/2025 23:28:09

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.082	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.022	23:23
Breath Volume	3.324	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.019	23:26
Breath Volume	3.023	23:26
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/20/2025 02:38:53

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.140	02:33
Breath Volume	1.117	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.130	02:37
Breath Volume	1.855	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: CRUMBLEY

DATE/TIME OF TEST: 03/20/2025 22:23:33

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.095	22:18
Breath Volume	2.082	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.098	22:21
Breath Volume	2.210	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: GILCREASE

DATE/TIME OF TEST: 03/21/2025 03:11:57

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.082	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.098	03:07
Breath Volume	1.613	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:09
Subject Sample # 2	0.094	03:10
Breath Volume	1.472	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HICE,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/21/2025 05:10:25

Diagnostics Check	OK	05:03
Air Blank	0.000	05:04
Control Test	0.082	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.127	05:05
Breath Volume	1.707	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.119	05:08
Breath Volume	1.375	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STOINIOFF,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RUEN

DATE/TIME OF TEST: 03/22/2025 03:00:44

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.082	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.054	02:56
Breath Volume	1.898	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.055	02:59
Breath Volume	1.898	02:59
Air Blank	0.000	02:59
Control Test	0.082	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROSADO,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SAVIORY

DATE/TIME OF TEST: 03/22/2025 05:15:49

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.082	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:14	Subject Test Refused
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.081	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 03/23/2025 02:10:45

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.082	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.082	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CHAVEZ QUINTERO

DATE/TIME OF TEST: 03/23/2025 03:32:34

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.108	03:27
Breath Volume	1.566	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.108	03:30
Breath Volume	2.996	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DHALIWAL

DATE/TIME OF TEST: 03/23/2025 04:01:08

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.081	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.081	04:00	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JAMES,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 03/23/2025 07:53:03

Diagnostics Check	OK	07:45
Air Blank	0.000	07:46
Control Test	0.082	07:46
Air Blank	0.000	07:47
Subject Sample # 1	0.212	07:47
Breath Volume	2.296	07:47
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.201	07:51
Breath Volume	2.460	07:51
Air Blank	0.000	07:52
Control Test	0.077	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:53

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: KAMARAT

DATE/TIME OF TEST: 03/24/2025 00:01:50

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.082	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.161	23:57
Breath Volume	2.953	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.154	00:00
Breath Volume	1.589	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:01
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 03/24/2025 01:28:58

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.081	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1 0.197	01:27	
Breath Volume	1.578	01:27	
Air Blank		01:28	Purge Fail
Air Blank	0.000	01:28	

Data Download Date/Time: 04/15/2025 18:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 03/24/2025 01:39:41

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.075	01:31	
Air Blank	0.000	01:31	
Subject Sample # 1	0.189	01:33	
Breath Volume	1.117	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample # 2	0.176	01:37	
Breath Volume	1.277	01:37	
Air Blank	0.000	01:38	
Control Test	0.073	01:39	Control Outside Tolerance
Air Blank	0.000	01:39	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 03/24/2025 01:49:17

Diagnostics Check	OK	01:42
Air Blank	0.000	01:43
Control Test	0.076	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.191	01:44
Breath Volume	1.113	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.190	01:47
Breath Volume	1.199	01:47
Air Blank	0.000	01:48
Control Test	0.078	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: THORNHILL

DATE/TIME OF TEST: 03/26/2025 00:28:58

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.082	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MCDUGAL

DATE/TIME OF TEST: 03/26/2025 04:22:57

Diagnostics Check	OK	04:19	
Air Blank	0.000	04:19	
Control Test	0.081	04:20	
Air Blank	0.000	04:20	
Subject Sample	# 1	04:21	Subject Test Refused
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Control Test	0.082	04:22	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 03/26/2025 21:08:30

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:05	
Control Test	0.082	21:06	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:07	Subject Test Refused
Breath Volume	0.000	21:07	
Air Blank	0.000	21:07	
Control Test	0.082	21:07	
Air Blank	0.000	21:08	
Diagnostics Check	OK	21:08	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 03/26/2025 23:06:14

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.082	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.121	23:01
Breath Volume	1.812	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:03
Subject Sample # 2	0.117	23:04
Breath Volume	1.476	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 03/27/2025 01:33:05

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.082	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.063	01:28
Breath Volume	1.667	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.062	01:31
Breath Volume	1.878	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: CYGAN

DATE/TIME OF TEST: 03/27/2025 02:14:56

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.082	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:13	
Control Test	0.082	02:14	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/28/2025 03:58:05

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.061	03:53
Breath Volume	1.121	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.060	03:56
Breath Volume	2.121	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 03/29/2025 02:24:56

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.155	02:20
Breath Volume	1.250	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:22
Subject Sample # 2	0.152	02:23
Breath Volume	1.835	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: KOHN III

DATE/TIME OF TEST: 03/30/2025 01:30:02

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.082	01:27	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: CORTEZ BAUTISTA

DATE/TIME OF TEST: 03/30/2025 05:52:06

Diagnostics Check	OK	05:49	
Air Blank	0.000	05:49	
Control Test	0.082	05:50	
Air Blank	0.000	05:50	
Subject Sample	# 1	05:50	Subject Test Refused
Breath Volume	0.000	05:50	
Air Blank	0.000	05:51	
Control Test	0.082	05:51	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MELTON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: ORTIZ DAYA

DATE/TIME OF TEST: 03/30/2025 07:38:05

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.081	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.147	07:33
Breath Volume	2.472	07:33
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.140	07:36
Breath Volume	2.566	07:36
Air Blank	0.000	07:37
Control Test	0.079	07:37
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ROTHFELDT

DATE/TIME OF TEST: 03/30/2025 18:38:09

Diagnostics Check	OK	18:35	
Air Blank	0.000	18:35	
Control Test	0.081	18:36	
Air Blank	0.000	18:36	
Subject Sample	# 1	18:36	Subject Test Refused
Breath Volume	0.000	18:36	
Air Blank	0.000	18:37	
Control Test	0.081	18:37	
Air Blank	0.000	18:38	
Diagnostics Check	OK	18:38	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 03/30/2025 21:35:33

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:33	
Control Test	0.081	21:33	
Air Blank	0.000	21:34	
Subject Sample	# 1	21:34	Subject Test Refused
Breath Volume	0.000	21:34	
Air Blank	0.000	21:34	
Control Test	0.081	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	OK	21:35	

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: GOODSON

DATE/TIME OF TEST: 03/30/2025 22:56:57

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.082	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.136	22:52
Breath Volume	2.859	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:54
Subject Sample # 2	0.131	22:55
Breath Volume	2.550	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 04/15/2025 18:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 03/01/2025 20:10:59

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.080	20:05
Air Blank	0.000	20:05
Subject Sample # 1	0.230	20:06
Breath Volume	2.242	20:06
Air Blank	0.000	20:07
Air Blank	0.000	20:08
Subject Sample # 2	0.233	20:09
Breath Volume	2.570	20:09
Air Blank	0.000	20:10
Control Test	0.079	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: KESSINGER

DATE/TIME OF TEST: 03/02/2025 06:49:45

Diagnostics Check	OK	06:38	
Air Blank	0.000	06:38	
Control Test	0.079	06:39	
Air Blank	0.000	06:39	
Subject Sample	# 1	06:42	No Sample Provided
Breath Volume	0.000	06:42	
Air Blank	0.000	06:43	
Air Blank	0.000	06:45	
Subject Sample	# 2	06:48	No Sample Provided
Breath Volume	0.000	06:48	
Air Blank	0.000	06:48	
Control Test	0.080	06:49	
Air Blank	0.000	06:49	
Diagnostics Check	OK	06:49	

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BASHFORD, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MANZI

DATE/TIME OF TEST: 03/05/2025 00:26:52

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.079	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1	00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2	0.166	00:22
Breath Volume	1.136	00:22	
Air Blank	0.000	00:22	
Air Blank	0.000	00:24	
Subject Sample	# 3	0.173	00:25
Breath Volume	1.718	00:25	
Air Blank	0.000	00:25	
Control Test	0.078	00:26	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BENDEN

DATE/TIME OF TEST: 03/08/2025 00:24:47

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:20	No Sample Provided
Breath Volume	0.000	00:20	
Air Blank	0.000	00:21	
Air Blank	0.000	00:23	
Subject Sample	# 2	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.080	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/08/2025 01:22:11

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.080	01:21	
Air Blank		01:21	RFI Detect
Air Blank	0.000	01:22	

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BYRNE, J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/08/2025 01:30:55

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.209	01:25
Breath Volume	1.214	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.191	01:29
Breath Volume	1.550	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 03/08/2025 03:56:44

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.195	03:51
Breath Volume	2.429	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.191	03:55
Breath Volume	2.242	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: BEAUREGARD

DATE/TIME OF TEST: 03/09/2025 05:39:43

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.124	05:34
Breath Volume	1.210	05:34
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.135	05:38
Breath Volume	1.175	05:38
Air Blank	0.000	05:38
Control Test	0.079	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NEWTON, N
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: SALYER

DATE/TIME OF TEST: 03/09/2025 20:53:16

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.079	20:45	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:49	No Sample Provided
Breath Volume	0.000	20:49	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Diagnostics Check	OK	20:53	

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRISTUPA,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 03/12/2025 22:29:04

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.079	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.221	22:24
Breath Volume	2.257	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:26
Subject Sample # 2	0.210	22:27
Breath Volume	1.406	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:29

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DELEON GOMEZ

DATE/TIME OF TEST: 03/13/2025 22:41:13

Diagnostics Check	OK	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.187	22:35
Breath Volume	1.175	22:35
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.198	22:39
Breath Volume	1.542	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 03/26/2025 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 03/13/2025 22:58:10

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:54	No Sample Provided
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.079	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/14/2025 03:37:44

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:33	No Sample Provided
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.080	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/14/2025 07:09:59

Diagnostics Check	OK	07:02
Air Blank	0.000	07:03
Control Test	0.080	07:03
Air Blank	0.000	07:03
Subject Sample # 1	0.060	07:05
Breath Volume	1.972	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:07
Subject Sample # 2	0.054	07:08
Breath Volume	1.316	07:08
Air Blank	0.000	07:09
Control Test	0.079	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/15/2025 01:24:02

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.080	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:19	No Sample Provided
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Air Blank	0.000	01:21	
Subject Sample	# 2	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Control Test	0.079	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOIVIN,N
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/19/2025 03:18:37

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.011	03:14
Breath Volume	2.195	03:14
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.011	03:17
Breath Volume	2.835	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FUENTES,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: LARUE

DATE/TIME OF TEST: 03/19/2025 05:14:08

Diagnostics Check	OK	05:07
Air Blank	0.000	05:07
Control Test	0.080	05:08
Air Blank	0.000	05:08
Subject Sample # 1	0.142	05:09
Breath Volume	1.968	05:09
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.142	05:12
Breath Volume	2.082	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHLKE,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: MARCUS

DATE/TIME OF TEST: 03/19/2025 19:45:42

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:41	
Control Test	0.080	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:45	RFI Detect
Breath Volume	0.000	19:45	
Air Blank		19:45	RFI Detect
Air Blank	0.000	19:45	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHLKE,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: MARCUS

DATE/TIME OF TEST: 03/19/2025 19:55:38

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:47	
Control Test	0.080	19:47	
Air Blank	0.000	19:48	
Subject Sample	# 1	19:51	No Sample Provided
Breath Volume	0.000	19:51	
Air Blank	0.000	19:52	
Air Blank	0.000	19:54	
Subject Sample	# 2	19:54	Subject Test Refused
Breath Volume	0.000	19:54	
Air Blank	0.000	19:54	
Control Test	0.080	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ADARRAGA

DATE/TIME OF TEST: 03/20/2025 03:50:59

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 2	03:49	Subject Test Refused
Breath Volume	0.000	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:50	
Diagnostics Check	OK	03:50	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 03/21/2025 22:05:45

Diagnostics Check	OK	21:56	
Air Blank	0.000	21:57	
Control Test	0.080	21:57	
Air Blank	0.000	21:57	
Subject Sample	# 1	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Air Blank	0.000	22:03	
Subject Sample	# 2	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:04	
Control Test	0.081	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 03/22/2025 03:01:50

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 03/22/2025 21:47:31

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.200	21:42
Breath Volume	1.929	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.196	21:46
Breath Volume	2.035	21:46
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: EVANS, D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/23/2025 05:53:24

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:45	
Control Test	0.080	05:45	
Air Blank	0.000	05:46	
Subject Sample	# 1	05:49	No Sample Provided
Breath Volume	0.000	05:49	
Air Blank	0.000	05:49	
Air Blank	0.000	05:51	
Subject Sample	# 2	05:52	Subject Test Refused
Breath Volume	0.000	05:52	
Air Blank	0.000	05:52	
Control Test	0.081	05:52	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 03/24/2025 02:24:10

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.000	02:19
Breath Volume	2.050	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:22
Breath Volume	1.593	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: LAPOINTE

DATE/TIME OF TEST: 03/25/2025 00:23:09

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	1.324	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	1.300	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/26/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SEARS, C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CRANFORD

DATE/TIME OF TEST: 03/28/2025 01:40:22

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample	# 2	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 04/24/2025 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WIMER,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MASAVES

DATE/TIME OF TEST: 03/28/2025 04:17:01

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Subject Sample # 1	0.235	04:12
Breath Volume	2.675	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.225	04:15
Breath Volume	1.996	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/24/2025 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: NECESSARY

DATE/TIME OF TEST: 03/30/2025 03:42:04

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank		03:41	RFI Detect
Air Blank	0.000	03:42	

Data Download Date/Time: 04/24/2025 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: NECESARY

DATE/TIME OF TEST: 03/30/2025 03:52:12

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 04/24/2025 09:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MOSS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: BUTALEWICZ

DATE/TIME OF TEST: 03/07/2025 20:36:54

Diagnostics Check	OK	20:29
Air Blank	0.000	20:30
Control Test	0.078	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.174	20:31
Breath Volume	1.347	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.167	20:35
Breath Volume	1.328	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 03/28/2025 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: DESIR,KASSY
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SEPTEMBRE

DATE/TIME OF TEST: 03/08/2025 01:04:10

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.254	00:58
Breath Volume	2.011	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.255	01:02
Breath Volume	1.820	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 03/28/2025 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 03/13/2025 17:12:59

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.078	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.000	17:08
Breath Volume	1.269	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:11
Breath Volume	1.421	17:11
Air Blank	0.000	17:12
Control Test	0.078	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:12

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 03/15/2025 02:48:39

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.049	02:42
Breath Volume	3.449	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.051	02:47
Breath Volume	3.003	02:47
Air Blank	0.000	02:47
Control Test	0.080	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/15/2025 04:06:43

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.215	04:01
Breath Volume	1.820	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.220	04:05
Breath Volume	1.304	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: ANDERSON,TYLER
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: GALEANO

DATE/TIME OF TEST: 03/15/2025 05:21:56

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.208	05:17
Breath Volume	2.644	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.207	05:20
Breath Volume	1.589	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/16/2025 04:04:42

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1 0.186	03:53	
Breath Volume	1.574	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2 0.149	03:58	Volume Not Met
Breath Volume	0.144	03:58	
Air Blank	0.000	03:59	
Air Blank	0.000	04:01	
Subject Sample	# 3 0.168	04:03	
Breath Volume	1.433	04:03	
Air Blank	0.000	04:03	
Control Test	0.077	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPOTO,MATTHEW,R

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: ANDERSON,TYLER

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: DUI

ARREST TIME: 03:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 03/16/2025 04:19:49

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.077	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.223	04:13
Breath Volume	1.691	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.207	04:18
Breath Volume	1.523	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: GRANAS,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: TAYLOR JR

DATE/TIME OF TEST: 03/18/2025 16:33:55

Diagnostics Check	OK	16:23	
Air Blank	0.000	16:24	
Control Test	0.080	16:24	
Air Blank	0.000	16:24	
Subject Sample	# 1 0.201	16:28	Volume Not Met
Breath Volume	0.234	16:28	
Air Blank	0.000	16:29	
Air Blank	0.000	16:30	
Subject Sample	# 2	16:32	Subject Test Refused
Breath Volume	0.539	16:32	
Air Blank	0.000	16:32	
Control Test	0.076	16:33	
Air Blank	0.000	16:33	
Diagnostics Check	OK	16:33	

Data Download Date/Time: 03/27/2025 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEST,ALISHA,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WEST,ALISHA

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06723080A5

DRY GAS EXPIRY DATE: 04/05/2025

VIOLATION CODE: DUI

ARREST TIME: 00:08

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: TORRES SAEZ

DATE/TIME OF TEST: 03/22/2025 01:09:42

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.000	01:05
Breath Volume	2.449	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.000	01:08
Breath Volume	2.187	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,JRONALD
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: LABOY

DATE/TIME OF TEST: 03/23/2025 15:00:05

Diagnostics Check	OK	14:54
Air Blank	0.000	14:54
Control Test	0.079	14:54
Air Blank	0.000	14:55
Subject Sample # 1	0.043	14:55
Breath Volume	2.910	14:55
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.041	14:58
Breath Volume	1.781	14:58
Air Blank	0.000	14:59
Control Test	0.079	14:59
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 03/24/2025 03:22:35

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.000	03:18
Breath Volume	2.164	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.000	03:21
Breath Volume	1.218	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO, MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/24/2025 18:36:49

Diagnostics Check	OK	18:31	
Air Blank	0.000	18:32	
Control Test	0.076	18:32	
Air Blank	0.000	18:33	
Subject Sample	# 1	18:35	Subject Test Refused
Breath Volume	0.000	18:35	
Air Blank	0.000	18:35	
Control Test	0.077	18:36	
Air Blank	0.000	18:36	
Diagnostics Check	OK	18:36	

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/24/2025 18:43:51

Diagnostics Check	OK	18:39	
Air Blank	0.000	18:40	
Control Test	0.078	18:40	
Air Blank	0.000	18:40	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.079	18:43	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPOTO,MATTHEW,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SPOTO,MATTHEW
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 02/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: PHELAN

DATE/TIME OF TEST: 03/24/2025 21:16:30

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.079	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.187	21:12
Breath Volume	1.296	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.181	21:15
Breath Volume	1.277	21:15
Air Blank	0.000	21:15
Control Test	0.078	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 03/27/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODMAN,THOMAS,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: NIEVES,A
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 02/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 03/12/2025 14:31:02

Diagnostics Check	OK	14:23
Air Blank	0.000	14:23
Control Test	0.081	14:24
Air Blank	0.000	14:24
Subject Sample # 1	0.000	14:26
Breath Volume	5.015	14:26
Air Blank	0.000	14:26
Air Blank	0.000	14:28
Subject Sample # 2	0.000	14:29
Breath Volume	2.308	14:29
Air Blank	0.000	14:30
Control Test	0.081	14:30
Air Blank	0.000	14:30
Diagnostics Check	OK	14:31

Data Download Date/Time: 03/28/2025 02:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 02/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/02/2025 21:14:16

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.081	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.218	21:09
Breath Volume	1.707	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.211	21:12
Breath Volume	1.914	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 03/16/2025 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PAKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 02/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/06/2025 22:22:34

Diagnostics Check	OK	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.051	22:17
Breath Volume	3.433	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.045	22:21
Breath Volume	1.218	22:21
Air Blank	0.000	22:21
Control Test	0.078	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 03/16/2025 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: RUTIGLIANO,D
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: ANNIS

DATE/TIME OF TEST: 03/10/2025 23:09:31

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.046	23:05
Breath Volume	2.589	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.046	23:08
Breath Volume	2.515	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 03/16/2025 22:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CAMPANA,NICHOLAS
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 02/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: TIRADO-HERNANDEZ

DATE/TIME OF TEST: 03/15/2025 18:51:28

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:48	
Control Test	0.080	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:50	
Control Test	0.080	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 03/16/2025 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: FRAGA,JACKSON
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BINDA

DATE/TIME OF TEST: 03/16/2025 03:03:55

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.148	02:59
Breath Volume	1.734	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.155	03:02
Breath Volume	2.433	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 03/16/2025 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: PARKER,J
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: VANVLIET

DATE/TIME OF TEST: 03/29/2025 21:22:58

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.080	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.112	21:17
Breath Volume	2.683	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:19
Subject Sample # 2	0.107	21:21
Breath Volume	2.843	21:21
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 04/17/2025 21:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: COLLINS,TOM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 03/30/2025 01:08:51

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.074	01:08	Control Outside Tolerance
Air Blank	0.000	01:08	

Data Download Date/Time: 04/17/2025 21:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: COLLINS,TOM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 03/30/2025 01:22:30

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.079	01:11	
Air Blank	0.000	01:11	
Subject Sample	# 1 0.171	01:14	Volume Not Met
Breath Volume	0.171	01:14	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2 0.191	01:17	
Breath Volume	2.734	01:17	
Air Blank	0.000	01:18	
Air Blank	0.000	01:20	
Subject Sample	# 3 0.191	01:21	
Breath Volume	1.804	01:21	
Air Blank	0.000	01:21	
Control Test	0.077	01:21	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 04/17/2025 21:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,J,A
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: CURBELO,A
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 03/30/2025 01:49:16

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.162	01:44
Breath Volume	1.183	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.166	01:47
Breath Volume	1.343	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 04/17/2025 21:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORENSTEIN,RANDY,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HORENSTEIN,RANDY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MOHRBACHER

DATE/TIME OF TEST: 03/31/2025 23:39:55

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.059	23:35
Breath Volume	2.136	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.058	23:38
Breath Volume	4.410	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 04/17/2025 21:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VAZQUEZ,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GRINNELL

DATE/TIME OF TEST: 03/27/2025 03:13:21

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.190	03:08
Breath Volume	2.312	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.186	03:11
Breath Volume	1.921	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 04/23/2025 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSEN,DUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANSEN,DUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: BUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: EMBERTON

DATE/TIME OF TEST: 03/27/2025 18:30:30

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.080	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.059	18:25
Breath Volume	2.300	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.061	18:29
Breath Volume	3.222	18:29
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 04/23/2025 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 03/27/2025 20:28:04

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.079	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:26	Subject Test Refused
Breath Volume	0.000	20:26	
Air Blank	0.000	20:27	
Control Test	0.080	20:27	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:28	

Data Download Date/Time: 04/23/2025 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAZMURI,KEYSER,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GREGORY,JUSTIN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/14/2025 22:18:32

Diagnostics Check	OK	22:17	
Air Blank		22:18	RFI Detect
Air Blank	0.000	22:18	

Data Download Date/Time: 04/02/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAZMURI,KEYSER,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GREGORY,JUSTIN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/14/2025 22:29:15

Diagnostics Check	OK	22:22	
Air Blank	0.000	22:22	
Control Test	0.077	22:23	
Air Blank	0.000	22:23	
Subject Sample	# 1 0.208	22:24	
Breath Volume	1.925	22:24	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2 0.200	22:27	
Breath Volume	1.800	22:27	
Air Blank	0.000	22:28	
Control Test	0.074	22:28	Control Outside Tolerance
Air Blank	0.000	22:29	

Data Download Date/Time: 04/02/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAZMURI,KEYSER,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GREGORY,JUSTIN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/14/2025 22:50:44

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.076	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1 0.034	22:45	Volume Not Met
Breath Volume	0.257	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2	22:49	Subject Test Refused
Breath Volume	0.457	22:49	
Air Blank	0.000	22:49	
Control Test	0.076	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 04/02/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMORUYI,T,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: OMORUYI,T
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 03/24/2025 00:20:14

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.077	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1 0.113	00:13	Volume Not Met
Breath Volume	0.375	00:13	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample	# 2 0.142	00:18	Volume Not Met
Breath Volume	0.367	00:18	
Air Blank	0.000	00:19	
Control Test	0.077	00:19	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:20	

Data Download Date/Time: 04/02/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMORUYI,T,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: OMORUYI,T
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 03/24/2025 00:29:32

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.076	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.162	00:25
Breath Volume	1.429	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.153	00:28
Breath Volume	1.285	00:28
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/02/2025 12:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: LAURENTS

DATE/TIME OF TEST: 03/14/2025 21:32:45

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:22	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:26	
Air Blank	0.000	21:28	
Subject Sample	# 2	21:31	No Sample Provided
Breath Volume	0.000	21:31	
Air Blank	0.000	21:31	
Control Test	0.080	21:32	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 05/03/2025 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GIURBINO

DATE/TIME OF TEST: 03/15/2025 01:29:03

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1 0.109	01:24	
Breath Volume	2.546	01:24	
Air Blank	0.000	01:24	
Air Blank	0.000	01:26	
Subject Sample	# 2 0.108	01:27	
Breath Volume	2.703	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:28	
Air Blank		01:28	RFI Detect
Air Blank	0.000	01:29	

Data Download Date/Time: 05/03/2025 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: GIURBINO

DATE/TIME OF TEST: 03/15/2025 01:38:55

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.102	01:34
Breath Volume	2.414	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.102	01:37
Breath Volume	2.222	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 05/03/2025 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 03/16/2025 23:14:47

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.112	23:09
Breath Volume	1.164	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.114	23:13
Breath Volume	1.882	23:13
Air Blank	0.000	23:13
Control Test	0.081	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/03/2025 00:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: DEFELICE

DATE/TIME OF TEST: 03/22/2025 22:32:26

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.114	22:27
Breath Volume	1.464	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.109	22:30
Breath Volume	1.363	22:30
Air Blank	0.000	22:31
Control Test	0.081	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 03/23/2025 00:17:32

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.182	00:12
Breath Volume	1.835	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.184	00:15
Breath Volume	2.023	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: AHERN

DATE/TIME OF TEST: 03/23/2025 20:33:33

Diagnostics Check	OK	20:22	
Air Blank	0.000	20:22	
Control Test	0.081	20:22	
Air Blank	0.000	20:23	
Subject Sample	# 1	20:26	No Sample Provided
Breath Volume	0.000	20:26	
Air Blank	0.000	20:27	
Air Blank	0.000	20:29	
Subject Sample	# 2	20:32	No Sample Provided
Breath Volume	0.000	20:32	
Air Blank	0.000	20:32	
Control Test	0.081	20:32	
Air Blank	0.000	20:33	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HURT,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KIDD,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/28/2025 01:21:43

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.081	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.192	01:16
Breath Volume	1.683	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.199	01:20
Breath Volume	1.621	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,JONATHAN,T
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MITCHELL,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: DESHESKY

DATE/TIME OF TEST: 03/29/2025 04:00:00

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.081	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.129	03:55
Breath Volume	4.308	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.129	03:58
Breath Volume	4.246	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 03/29/2025 22:57:07

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.020	22:52
Breath Volume	1.574	22:52
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.020	22:55
Breath Volume	2.765	22:55
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 05/03/2025 00:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SOTO LOPEZ

DATE/TIME OF TEST: 03/01/2025 00:15:52

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.082	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.000	00:11
Breath Volume	1.242	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.000	00:14
Breath Volume	1.445	00:14
Air Blank	0.000	00:14
Control Test	0.080	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/14/2025 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/01/2025 07:46:02

Diagnostics Check	OK	07:43	
Air Blank	0.000	07:43	
Control Test	0.082	07:43	
Air Blank	0.000	07:44	
Subject Sample	# 1	07:44	Subject Test Refused
Breath Volume	0.000	07:44	
Air Blank	0.000	07:45	
Control Test	0.081	07:45	
Air Blank	0.000	07:45	
Diagnostics Check	OK	07:46	

Data Download Date/Time: 03/14/2025 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 03/01/2025 19:50:31

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.081	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.249	19:42
Breath Volume	1.832	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:44
Subject Sample # 2	0.221	19:45
Breath Volume	1.406	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 3	0.241	19:48
Breath Volume	2.398	19:48
Air Blank	0.000	19:49
Control Test	0.077	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 03/14/2025 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 03/01/2025 20:41:58

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.078	20:36
Breath Volume	1.519	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.072	20:40
Breath Volume	1.152	20:40
Air Blank	0.000	20:41
Control Test	0.078	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 03/14/2025 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: ISMAEL NAVARRO

DATE/TIME OF TEST: 03/01/2025 22:47:05

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.082	22:45	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Control Test	0.081	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:47	

Data Download Date/Time: 03/14/2025 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 03/02/2025 00:34:26

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.121	00:29
Breath Volume	2.054	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.119	00:32
Breath Volume	2.046	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SHAH

DATE/TIME OF TEST: 03/02/2025 01:56:05

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.220	01:51
Breath Volume	1.714	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.228	01:54
Breath Volume	1.386	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/02/2025 03:40:06

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.081	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.109	03:35
Breath Volume	1.773	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.111	03:38
Breath Volume	1.742	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: FERMAN

DATE/TIME OF TEST: 03/02/2025 05:53:20

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.040	05:48
Breath Volume	2.941	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.042	05:51
Breath Volume	2.890	05:51
Air Blank	0.000	05:52
Control Test	0.081	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/02/2025 09:06:39

Diagnostics Check	OK	08:59
Air Blank	0.000	09:00
Control Test	0.082	09:00
Air Blank	0.000	09:00
Subject Sample # 1	0.112	09:01
Breath Volume	3.207	09:01
Air Blank	0.000	09:02
Air Blank	0.000	09:04
Subject Sample # 2	0.111	09:05
Breath Volume	2.746	09:05
Air Blank	0.000	09:05
Control Test	0.078	09:06
Air Blank	0.000	09:06
Diagnostics Check	OK	09:06

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRESHAM,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: SANTOSPEREZ

DATE/TIME OF TEST: 03/02/2025 18:08:15

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.082	18:02
Air Blank	0.000	18:02
Subject Sample # 1	0.189	18:03
Breath Volume	1.558	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:05
Subject Sample # 2	0.196	18:06
Breath Volume	1.484	18:06
Air Blank	0.000	18:07
Control Test	0.081	18:07
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 03/02/2025 19:40:51

Diagnostics Check	OK	19:37	
Air Blank	0.000	19:38	
Control Test	0.081	19:38	
Air Blank	0.000	19:39	
Subject Sample	# 1	19:39	Subject Test Refused
Breath Volume	0.000	19:39	
Air Blank	0.000	19:39	
Control Test	0.081	19:40	
Air Blank	0.000	19:40	
Diagnostics Check	OK	19:40	

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 03/03/2025 01:39:07

Diagnostics Check	OK	01:36	
Air Blank	0.000	01:36	
Control Test	0.082	01:37	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 03/03/2025 02:23:02

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.081	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.106	02:18
Breath Volume	1.785	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.107	02:21
Breath Volume	2.867	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:23

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 03/03/2025 05:20:40

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:18	
Control Test	0.082	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.082	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 03/04/2025 00:12:05

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.090	00:07
Breath Volume	3.687	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.094	00:10
Breath Volume	3.367	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: PURVIS

DATE/TIME OF TEST: 03/04/2025 04:29:18

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.121	04:24
Breath Volume	1.812	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.124	04:27
Breath Volume	1.476	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 03/05/2025 00:53:30

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:52	Subject Test Refused
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.081	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KODOBOCZ

DATE/TIME OF TEST: 03/06/2025 01:12:36

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.081	01:10	
Air Blank	0.000	01:11	
Subject Sample	# 1	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.081	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/08/2025 03:04:14

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.082	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.139	02:59
Breath Volume	3.117	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.138	03:02
Breath Volume	2.972	03:02
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/08/2025 04:46:33

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.144	04:41
Breath Volume	1.750	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.151	04:44
Breath Volume	2.207	04:44
Air Blank	0.000	04:45
Control Test	0.077	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 03/14/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: DEVAUGHN

DATE/TIME OF TEST: 03/08/2025 20:10:05

Diagnostics Check	OK	20:03
Air Blank	0.000	20:04
Control Test	0.081	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.133	20:05
Breath Volume	1.988	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:07
Subject Sample # 2	0.129	20:08
Breath Volume	1.574	20:08
Air Blank	0.000	20:09
Control Test	0.080	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: OWENS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMJAS

DATE/TIME OF TEST: 03/08/2025 23:47:51

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.081	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.197	23:42
Breath Volume	1.402	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.188	23:46
Breath Volume	1.625	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/09/2025 00:52:36

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1 0.059	00:43	Volume Not Met
Breath Volume	1.062	00:43	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample	# 2 0.054	00:46	
Breath Volume	1.253	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 3 0.056	00:50	
Breath Volume	2.058	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 03/09/2025 03:51:52

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.034	03:47
Breath Volume	1.550	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.033	03:50
Breath Volume	1.742	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: CARCAMO MARTINEZ

DATE/TIME OF TEST: 03/09/2025 06:01:26

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.185	05:56
Breath Volume	2.464	05:56
Air Blank	0.000	05:57
Air Blank	0.000	05:58
Subject Sample # 2	0.186	05:59
Breath Volume	2.464	05:59
Air Blank	0.000	06:00
Control Test	0.079	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: CARDENAS PORRAS

DATE/TIME OF TEST: 03/09/2025 13:21:35

Diagnostics Check	OK	13:15
Air Blank	0.000	13:15
Control Test	0.079	13:15
Air Blank	0.000	13:16
Subject Sample # 1	0.081	13:17
Breath Volume	2.023	13:17
Air Blank	0.000	13:17
Air Blank	0.000	13:19
Subject Sample # 2	0.075	13:19
Breath Volume	2.171	13:19
Air Blank	0.000	13:20
Control Test	0.080	13:20
Air Blank	0.000	13:21
Diagnostics Check	OK	13:21

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHEY,A,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/09/2025 17:47:00

Diagnostics Check	OK	17:44	
Air Blank	0.000	17:44	
Control Test	0.080	17:44	
Air Blank	0.000	17:45	
Subject Sample	# 1	17:45	Subject Test Refused
Breath Volume	0.000	17:45	
Air Blank	0.000	17:46	
Control Test	0.080	17:46	
Air Blank	0.000	17:46	
Diagnostics Check	OK	17:47	

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,S,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TIRRI,5061
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: BEELER

DATE/TIME OF TEST: 03/10/2025 19:03:16

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.081	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.253	18:58
Breath Volume	1.539	18:58
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.258	19:01
Breath Volume	2.683	19:01
Air Blank	0.000	19:02
Control Test	0.081	19:02
Air Blank	0.000	19:03
Diagnostics Check	OK	19:03

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: BUNDY

DATE/TIME OF TEST: 03/10/2025 22:10:22

Diagnostics Check	OK	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.179	22:04
Breath Volume	1.359	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.191	22:08
Breath Volume	1.253	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: CASARRUBIAS GARCIA

DATE/TIME OF TEST: 03/11/2025 00:55:38

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.243	00:50
Breath Volume	2.234	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.237	00:53
Breath Volume	1.589	00:53
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 03/12/2025 01:20:26

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.081	01:18	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:19	
Control Test	0.081	01:19	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: UPTON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 03/12/2025 06:16:56

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.082	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.000	06:12
Breath Volume	2.078	06:12
Air Blank	0.000	06:13
Air Blank	0.000	06:15
Subject Sample # 2	0.000	06:15
Breath Volume	2.320	06:15
Air Blank	0.000	06:16
Control Test	0.082	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:16

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COREY,H
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 03/13/2025 00:12:13

Diagnostics Check	OK	00:05
Air Blank	0.000	00:06
Control Test	0.082	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.154	00:07
Breath Volume	1.828	00:07
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.157	00:10
Breath Volume	1.855	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 03/13/2025 01:06:49

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/13/2025 01:42:55

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.094	01:38
Breath Volume	1.515	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.100	01:41
Breath Volume	1.625	01:41
Air Blank	0.000	01:41
Control Test	0.081	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: OKAMBA

DATE/TIME OF TEST: 03/13/2025 03:46:19

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:44	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 03/14/2025 11:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHIDDEN,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: GARCIA MELENDEZ

DATE/TIME OF TEST: 03/13/2025 06:54:59

Diagnostics Check	OK	06:52	
Air Blank	0.000	06:52	
Control Test	0.081	06:52	
Air Blank	0.000	06:53	
Subject Sample	# 1	06:53	Subject Test Refused
Breath Volume	0.000	06:53	
Air Blank	0.000	06:54	
Control Test	0.081	06:54	
Air Blank	0.000	06:54	
Diagnostics Check	OK	06:54	

Data Download Date/Time: 03/14/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HORDIS

DATE/TIME OF TEST: 03/14/2025 03:08:24

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.083	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.027	03:04
Breath Volume	2.109	03:04
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.028	03:06
Breath Volume	1.804	03:06
Air Blank	0.000	03:07
Control Test	0.082	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/14/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 03/15/2025 01:23:44

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.084	01:19
Breath Volume	1.824	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.076	01:22
Breath Volume	1.250	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/15/2025 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 03/15/2025 02:06:14

Diagnostics Check	OK	02:03	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 04/15/2025 20:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GEREZ

DATE/TIME OF TEST: 03/15/2025 03:27:58

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.082	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.172	03:22
Breath Volume	1.683	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.183	03:26
Breath Volume	1.210	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 03/15/2025 04:19:31

Diagnostics Check	OK	04:16	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:17	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PEREZ-ALVAREZ

DATE/TIME OF TEST: 03/15/2025 20:24:09

Diagnostics Check	OK	20:21	
Air Blank	0.000	20:21	
Control Test	0.082	20:22	
Air Blank	0.000	20:22	
Subject Sample	# 1	20:22	Subject Test Refused
Breath Volume	0.000	20:22	
Air Blank	0.000	20:23	
Control Test	0.082	20:23	
Air Blank	0.000	20:24	
Diagnostics Check	OK	20:24	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DSOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DE LA LUZ

DATE/TIME OF TEST: 03/16/2025 00:12:33

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:10	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Control Test	0.081	00:11	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:12	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: KIRK

DATE/TIME OF TEST: 03/16/2025 00:56:08

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.100	00:51
Breath Volume	1.820	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.109	00:54
Breath Volume	1.859	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: AGUILAR HERRERA

DATE/TIME OF TEST: 03/16/2025 03:17:10

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:14	
Control Test	0.081	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/16/2025 05:05:22

Diagnostics Check	OK	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.162	05:00
Breath Volume	1.113	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.163	05:03
Breath Volume	1.121	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOHN,ANTHONY,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO-MARRERO,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 03/16/2025 05:38:06

Diagnostics Check	OK	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Subject Sample # 1	0.097	05:33
Breath Volume	1.121	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.100	05:36
Breath Volume	2.195	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LITTLE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/16/2025 20:08:44

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.082	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.221	20:02
Breath Volume	1.476	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.219	20:06
Breath Volume	1.390	20:06
Air Blank	0.000	20:07
Control Test	0.079	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BROWER

DATE/TIME OF TEST: 03/17/2025 04:14:25

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.082	04:12	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Control Test	0.082	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 03/18/2025 02:02:42

Diagnostics Check	OK	01:59	
Air Blank	0.000	02:00	
Control Test	0.082	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.082	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/18/2025 03:12:49

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.082	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.082	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VONESCHE,R
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/18/2025 03:47:45

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.082	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.124	03:43
Breath Volume	3.535	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.114	03:46
Breath Volume	1.484	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/22/2025 05:04:46

Diagnostics Check	OK	05:01	
Air Blank	0.000	05:02	
Control Test	0.082	05:02	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:03	Subject Test Refused
Breath Volume	0.000	05:03	
Air Blank	0.000	05:03	
Control Test	0.081	05:04	
Air Blank	0.000	05:04	
Diagnostics Check	OK	05:04	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SIPRES

DATE/TIME OF TEST: 03/23/2025 00:40:47

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:38	
Control Test	0.082	00:38	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.081	00:40	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ALEXANDRE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 03/23/2025 03:38:42

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.083	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: MALUEG

DATE/TIME OF TEST: 03/24/2025 00:00:50

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.082	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.181	23:55
Breath Volume	1.382	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.177	23:59
Breath Volume	1.343	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLION,S,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 03/24/2025 01:00:34

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.189	00:55
Breath Volume	2.070	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.196	00:58
Breath Volume	1.792	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HAGAN

DATE/TIME OF TEST: 03/25/2025 01:52:07

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.082	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:51	
Control Test	0.079	01:51	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/25/2025 13:31:45

Diagnostics Check	OK	13:25
Air Blank	0.000	13:25
Control Test	0.082	13:26
Air Blank	0.000	13:26
Subject Sample # 1	0.192	13:27
Breath Volume	1.660	13:27
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.193	13:30
Breath Volume	1.835	13:30
Air Blank	0.000	13:30
Control Test	0.080	13:31
Air Blank	0.000	13:31
Diagnostics Check	OK	13:31

Data Download Date/Time: 04/15/2025 20:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,J,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 03/25/2025 17:22:58

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.080	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.220	17:15
Breath Volume	1.816	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.197	17:17
Breath Volume	1.785	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 3	0.221	17:21
Breath Volume	3.691	17:21
Air Blank	0.000	17:22
Control Test	0.080	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 03/25/2025 20:51:14

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.081	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.000	20:46
Breath Volume	2.367	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.000	20:49
Breath Volume	2.253	20:49
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CROOMS,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 03/26/2025 01:18:48

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:17	
Control Test	0.080	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 03/27/2025 04:02:03

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.081	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 03/27/2025 05:35:43

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.082	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.127	05:30
Breath Volume	2.109	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.120	05:33
Breath Volume	1.382	05:33
Air Blank	0.000	05:34
Control Test	0.082	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: PASCHALL

DATE/TIME OF TEST: 03/28/2025 04:25:37

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.081	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 03/29/2025 01:26:38

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:25	Subject Test Refused
Breath Volume	0.000	01:25	
Air Blank	0.000	01:25	
Control Test	0.080	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 03/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 03/31/2025 19:52:30

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:47
Subject Sample # 1	0.096	19:48
Breath Volume	1.597	19:48
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.098	19:50
Breath Volume	1.972	19:50
Air Blank	0.000	19:51
Control Test	0.079	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 04/15/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CLOUTIER,AUSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 03/29/2025 16:24:02

Diagnostics Check	OK	16:16
Air Blank	0.000	16:17
Control Test	0.081	16:17
Air Blank	0.000	16:18
Subject Sample # 1	0.201	16:19
Breath Volume	1.464	16:19
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.216	16:22
Breath Volume	2.011	16:22
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:24

Data Download Date/Time: 04/30/2025 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,S,
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: AROCHO,J
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/30/2025 05:23:41

Diagnostics Check	OK	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Subject Sample # 1	0.214	05:18
Breath Volume	1.394	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.202	05:22
Breath Volume	2.429	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:23
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 04/23/2025 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/01/2025 01:21:18

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:18	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.082	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 03/12/2025 00:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 03/01/2025 15:52:22

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.082	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	2.691	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:50
Breath Volume	2.449	15:50
Air Blank	0.000	15:51
Control Test	0.082	15:51
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HUTCHISON,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 03/01/2025 18:54:56

Diagnostics Check	OK	18:48
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:49
Subject Sample # 1	0.000	18:50
Breath Volume	3.039	18:50
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.000	18:53
Breath Volume	3.136	18:53
Air Blank	0.000	18:54
Control Test	0.080	18:54
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HOBART,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: BUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: TALBOT

DATE/TIME OF TEST: 03/02/2025 16:59:55

Diagnostics Check	OK	16:53
Air Blank	0.000	16:53
Control Test	0.082	16:54
Air Blank	0.000	16:54
Subject Sample # 1	0.130	16:55
Breath Volume	3.359	16:55
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 2	0.118	16:58
Breath Volume	1.171	16:58
Air Blank	0.000	16:59
Control Test	0.082	16:59
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ELBERT

DATE/TIME OF TEST: 03/05/2025 02:22:06

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:15	No Sample Provided
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2	02:20	No Sample Provided
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MILON

DATE/TIME OF TEST: 03/05/2025 03:53:10

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:48	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.081	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ADDERLY

DATE/TIME OF TEST: 03/06/2025 03:18:57

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.152	03:13
Breath Volume	1.316	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.167	03:17
Breath Volume	1.265	03:17
Air Blank	0.000	03:18
Control Test	0.081	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 03/06/2025 23:26:20

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.080	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:24	Subject Test Refused
Breath Volume	0.000	23:24	
Air Blank	0.000	23:25	
Control Test	0.081	23:25	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MANKOWSKI,N
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 03/07/2025 00:54:44

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:53	Subject Test Refused
Breath Volume	0.000	00:53	
Air Blank	0.000	00:53	
Control Test	0.081	00:54	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,OFFICER,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: HATHORN

DATE/TIME OF TEST: 03/07/2025 01:22:57

Diagnostics Check	OK	01:17	
Air Blank	0.000	01:18	
Control Test	0.081	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.000	01:21	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Diagnostics Check	OK	01:22	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: TRAIL

DATE/TIME OF TEST: 03/07/2025 20:23:33

Diagnostics Check	OK	20:19	
Air Blank	0.000	20:19	
Control Test	0.081	20:20	
Air Blank	0.000	20:20	
Subject Sample	# 1	20:22	RFI Detect
Breath Volume	0.000	20:22	
Air Blank	0.000	20:22	
Control Test	0.081	20:22	
Air Blank	0.000	20:23	
Diagnostics Check	OK	20:23	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: TRAIL

DATE/TIME OF TEST: 03/07/2025 20:37:00

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:26	
Control Test	0.080	20:26	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:30	No Sample Provided
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Air Blank	0.000	20:32	
Subject Sample	# 2	20:35	No Sample Provided
Breath Volume	0.000	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:36	
Diagnostics Check	OK	20:37	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ABATTE

DATE/TIME OF TEST: 03/08/2025 02:48:38

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.081	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:48	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUNARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: QUINONES GONZALEZ

DATE/TIME OF TEST: 03/08/2025 04:04:04

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.080	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: FERNANDEZ,JULIO
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 03/08/2025 05:29:25

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.174	05:24
Breath Volume	1.683	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.168	05:27
Breath Volume	1.996	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: NARINE

DATE/TIME OF TEST: 03/09/2025 05:20:49

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.081	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.265	05:15
Breath Volume	2.171	05:15
Air Blank	0.000	05:16
Air Blank	0.000	05:18
Subject Sample # 2	0.258	05:19
Breath Volume	1.710	05:19
Air Blank	0.000	05:19
Control Test	0.080	05:20
Air Blank	0.000	05:20
Diagnostics Check	OK	05:20

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUSAM,MATTHEW,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: NAPOLI

DATE/TIME OF TEST: 03/10/2025 00:35:21

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.141	00:30
Breath Volume	2.066	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.149	00:33
Breath Volume	1.808	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,PAUL,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HINES,PAUL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: SEEDERS

DATE/TIME OF TEST: 03/10/2025 20:07:10

Diagnostics Check	OK	20:01
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:02
Subject Sample # 1	0.000	20:02
Breath Volume	3.089	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.000	20:05
Breath Volume	2.996	20:05
Air Blank	0.000	20:06
Control Test	0.081	20:06
Air Blank	0.000	20:07
Diagnostics Check	OK	20:07

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 03/11/2025 19:38:38

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.081	19:34	
Air Blank	0.000	19:35	
Subject Sample	# 1	19:37	Subject Test Refused
Breath Volume	0.000	19:37	
Air Blank	0.000	19:37	
Control Test	0.081	19:38	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: ROWLEY

DATE/TIME OF TEST: 03/11/2025 22:09:46

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.081	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.149	22:05
Breath Volume	1.984	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.141	22:08
Breath Volume	1.929	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 03/12/2025 00:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BRATTEBO

DATE/TIME OF TEST: 03/12/2025 03:15:35

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.155	03:10
Breath Volume	2.480	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.156	03:14
Breath Volume	2.093	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/12/2025 20:30:42

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.142	20:25
Breath Volume	1.835	20:25
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.136	20:29
Breath Volume	2.007	20:29
Air Blank	0.000	20:29
Control Test	0.081	20:30
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: DODENHOFF

DATE/TIME OF TEST: 03/13/2025 02:00:17

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.214	01:55
Breath Volume	1.777	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.227	01:58
Breath Volume	1.554	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PIERSMA

DATE/TIME OF TEST: 03/13/2025 02:35:22

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:31	
Control Test	0.080	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: HETRICK

DATE/TIME OF TEST: 03/13/2025 23:29:10

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 1	23:24	No Sample Provided
Breath Volume	0.000	23:24	
Air Blank	0.000	23:24	
Air Blank	0.000	23:26	
Subject Sample	# 2	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:28	
Control Test	0.080	23:28	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANDERSON,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LUKOSEVICIUS

DATE/TIME OF TEST: 03/14/2025 00:29:07

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: KORB

DATE/TIME OF TEST: 03/15/2025 01:24:18

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.081	01:12	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Air Blank	0.000	01:18	
Subject Sample	# 2	0.327	01:19
Breath Volume	2.519	01:19	
Air Blank	0.000	01:20	
Air Blank	0.000	01:22	
Subject Sample	# 3	0.334	01:22
Breath Volume	2.750	01:22	
Air Blank	0.000	01:23	
Control Test	0.080	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KAPCIO

DATE/TIME OF TEST: 03/15/2025 02:19:47

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:15	
Control Test	0.081	02:16	
Air Blank	0.000	02:16	
Subject Sample	# 1	02:18	Subject Test Refused
Breath Volume	0.000	02:18	
Air Blank	0.000	02:18	
Control Test	0.080	02:19	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 03/15/2025 04:42:25

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.081	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.080	04:41	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,CLAYTON,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HUTCHINSON,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: DRUMM

DATE/TIME OF TEST: 03/16/2025 00:26:17

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:15	
Control Test	0.081	00:15	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	No Sample Provided
Breath Volume	0.000	00:19	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 2	00:24	No Sample Provided
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:26	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 03/17/2025 23:23:16

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.168	23:18
Breath Volume	2.691	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.169	23:21
Breath Volume	1.726	23:21
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: WEILER

DATE/TIME OF TEST: 03/18/2025 01:17:42

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:09	
Control Test	0.081	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:13	No Sample Provided
Breath Volume	0.000	01:13	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 2	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 03/18/2025 03:18:19

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.249	03:13
Breath Volume	2.210	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.246	03:16
Breath Volume	2.070	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: CHAMICH

DATE/TIME OF TEST: 03/18/2025 03:25:33

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/20/2025 21:06:12

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.075	21:00	
Air Blank	0.000	21:00	
Subject Sample # 1	0.077	21:01	
Breath Volume	1.855	21:01	
Air Blank	0.000	21:02	
Air Blank	0.000	21:04	
Subject Sample # 2	0.077	21:04	
Breath Volume	1.734	21:04	
Air Blank	0.000	21:05	
Control Test	0.074	21:05	Control Outside Tolerance
Air Blank	0.000	21:06	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 03/25/2025 22:32:34

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.080	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:31	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: THURMAN

DATE/TIME OF TEST: 03/28/2025 00:26:40

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:22
Breath Volume	2.398	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:25
Breath Volume	2.234	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 04/10/2025 09:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 03/29/2025 00:11:52

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.080	00:06
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:07
Breath Volume	3.019	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:10
Breath Volume	1.429	00:10
Air Blank	0.000	00:10
Control Test	0.080	00:11
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/10/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CROSLEY

DATE/TIME OF TEST: 03/29/2025 04:37:17

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.197	04:32
Breath Volume	4.277	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.190	04:35
Breath Volume	4.003	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 04/10/2025 09:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,NICHOLAS,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MARTINEZ RODRIGUEZ

DATE/TIME OF TEST: 03/29/2025 09:17:10

Diagnostics Check	OK	09:10
Air Blank	0.000	09:11
Control Test	0.081	09:11
Air Blank	0.000	09:12
Subject Sample # 1	0.000	09:12
Breath Volume	2.003	09:12
Air Blank	0.000	09:13
Air Blank	0.000	09:15
Subject Sample # 2	0.000	09:15
Breath Volume	1.382	09:15
Air Blank	0.000	09:16
Control Test	0.081	09:16
Air Blank	0.000	09:17
Diagnostics Check	OK	09:17

Data Download Date/Time: 04/10/2025 09:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: PEREZ,ARLYNE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BERENDT

DATE/TIME OF TEST: 03/07/2025 02:48:24

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:46	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:47	Subject Test Refused
Breath Volume	0.000	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Diagnostics Check	OK	02:48	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MAXIE

DATE/TIME OF TEST: 03/08/2025 04:41:50

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	Subject Test Refused
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GRIGGS,JACK
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 02/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MENKE

DATE/TIME OF TEST: 03/09/2025 00:57:46

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.080	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 03/10/2025 12:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GAULT

DATE/TIME OF TEST: 03/20/2025 03:12:45

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.080	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 03/21/2025 02:27:54

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.117	02:23
Breath Volume	3.019	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.115	02:26
Breath Volume	2.601	02:26
Air Blank	0.000	02:26
Control Test	0.080	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ROCKWELL

DATE/TIME OF TEST: 03/22/2025 05:23:28

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.081	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.181	05:18
Breath Volume	2.371	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:21
Subject Sample # 2	0.184	05:22
Breath Volume	1.769	05:22
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 16:01

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:29

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 03/27/2025 16:58:30

Diagnostics Check	OK	16:49
Air Blank	0.000	16:50
Control Test	0.080	16:50
Air Blank	0.000	16:51
Subject Sample # 1	0.114	16:53
Breath Volume	2.324	16:53
Air Blank	0.000	16:53
Air Blank	0.000	16:55
Subject Sample # 2	0.100	16:57
Breath Volume	1.171	16:57
Air Blank	0.000	16:57
Control Test	0.080	16:57
Air Blank	0.000	16:58
Diagnostics Check	OK	16:58

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 18:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: YOUMAN

DATE/TIME OF TEST: 03/28/2025 19:00:00

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.081	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.000	18:55
Breath Volume	2.644	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.000	18:58
Breath Volume	2.179	18:58
Air Blank	0.000	18:59
Control Test	0.081	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	19:00

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HASTINGS,KENNETH,

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 15:56

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 03/29/2025 16:45:56

Diagnostics Check	OK	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:40
Air Blank	0.000	16:41
Subject Sample # 1	0.037	16:41
Breath Volume	1.125	16:41
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.033	16:44
Breath Volume	1.210	16:44
Air Blank	0.000	16:45
Control Test	0.081	16:45
Air Blank	0.000	16:45
Diagnostics Check	OK	16:45

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 03/29/2025 20:10:54

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.080	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.228	20:05
Breath Volume	1.324	20:05
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.228	20:09
Breath Volume	1.726	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 04/17/2025 13:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 03/25/2025 12:21:11

Diagnostics Check	OK	12:09	
Air Blank	0.000	12:10	
Control Test	0.079	12:10	
Air Blank	0.000	12:11	
Subject Sample	# 1	12:14	No Sample Provided
Breath Volume	0.000	12:14	
Air Blank	0.000	12:14	
Air Blank	0.000	12:16	
Subject Sample	# 2	12:19	No Sample Provided
Breath Volume	0.000	12:19	
Air Blank	0.000	12:20	
Control Test	0.080	12:20	
Air Blank	0.000	12:21	
Diagnostics Check	OK	12:21	

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ORTIZ,JESUS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: THOMPSON JR.

DATE/TIME OF TEST: 03/25/2025 19:39:38

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.080	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.128	19:35
Breath Volume	1.921	19:35
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.136	19:38
Breath Volume	1.945	19:38
Air Blank	0.000	19:38
Control Test	0.079	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/25/2025 22:03:04

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.118	21:58
Breath Volume	1.246	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.119	22:01
Breath Volume	1.398	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:03

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: NAUMANN

DATE/TIME OF TEST: 03/26/2025 20:18:22

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.078	20:07	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:11	No Sample Provided
Breath Volume	0.000	20:11	
Air Blank	0.000	20:11	
Air Blank	0.000	20:13	
Subject Sample	# 2	20:16	No Sample Provided
Breath Volume	0.000	20:16	
Air Blank	0.000	20:17	
Control Test	0.079	20:17	
Air Blank	0.000	20:18	
Diagnostics Check	OK	20:18	

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: NORTHCUTT JR

DATE/TIME OF TEST: 03/28/2025 17:35:45

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.079	17:29
Air Blank	0.000	17:29
Subject Sample # 1	0.235	17:30
Breath Volume	1.582	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.237	17:34
Breath Volume	2.351	17:34
Air Blank	0.000	17:34
Control Test	0.077	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DAWSON,MICHAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: SAMPLE TEST

DATE/TIME OF TEST: 03/31/2025 06:28:36

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:27	
Control Test	0.080	06:27	
Air Blank	0.000	06:27	
Subject Sample	# 1	06:28	Sequence Aborted
Breath Volume	0.000	06:28	
Air Blank	0.000	06:28	

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LOTT

DATE/TIME OF TEST: 03/31/2025 02:45:24

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:38	No Sample Provided
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:44	No Sample Provided
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.081	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 04/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: ALDANA

DATE/TIME OF TEST: 03/02/2025 19:31:28

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.118	19:26
Breath Volume	4.074	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.118	19:29
Breath Volume	4.054	19:29
Air Blank	0.000	19:30
Control Test	0.080	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 03/28/2025 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 03/12/2025 02:19:18

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.077	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.080	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 03/28/2025 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: KAMINSKI

DATE/TIME OF TEST: 03/18/2025 23:36:03

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.079	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:35	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: KAMINSKI

DATE/TIME OF TEST: 03/18/2025 23:55:06

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.164	23:48
Breath Volume	1.417	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.154	23:53
Breath Volume	1.332	23:53
Air Blank	0.000	23:54
Control Test	0.077	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SEIN-CABAN,G
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KOCK

DATE/TIME OF TEST: 03/19/2025 03:12:46

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.079	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:05	No Sample Provided
Breath Volume	0.000	03:05	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: HEID

DATE/TIME OF TEST: 03/23/2025 00:37:04

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.000	00:32
Breath Volume	1.199	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.000	00:35
Breath Volume	1.335	00:35
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:37

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CALCAGNO

DATE/TIME OF TEST: 03/23/2025 02:10:40

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.205	02:06
Breath Volume	2.792	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.206	02:09
Breath Volume	2.996	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIZZO,HAYLEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RIZZO,HAYLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: LANPHAR J

DATE/TIME OF TEST: 03/23/2025 16:35:33

Diagnostics Check	OK	16:29
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Subject Sample # 1	0.000	16:31
Breath Volume	1.207	16:31
Air Blank	0.000	16:31
Air Blank	0.000	16:33
Subject Sample # 2	0.000	16:34
Breath Volume	1.136	16:34
Air Blank	0.000	16:34
Control Test	0.080	16:34
Air Blank	0.000	16:35
Diagnostics Check	OK	16:35

Data Download Date/Time: 03/28/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING, KELEY, L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING, KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: PITT

DATE/TIME OF TEST: 03/29/2025 16:47:31

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.079	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.210	16:42
Breath Volume	1.941	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:45
Subject Sample # 2	0.207	16:45
Breath Volume	2.937	16:45
Air Blank	0.000	16:46
Control Test	0.079	16:46
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 04/30/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: WELL

DATE/TIME OF TEST: 03/31/2025 01:08:01

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.172	01:03
Breath Volume	2.628	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.168	01:06
Breath Volume	2.550	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:08

Data Download Date/Time: 04/30/2025 11:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/31/2025 23:37:04

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.084	23:32
Breath Volume	3.492	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.085	23:35
Breath Volume	3.207	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/30/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLEY,KEVIN,G
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BERGER,ERIC
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 02/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:06

SUBJECT'S LAST NAME: PHILIPSON

DATE/TIME OF TEST: 03/04/2025 09:37:33

Diagnostics Check	OK	09:29	
Air Blank	0.000	09:29	
Control Test	0.078	09:29	
Air Blank	0.000	09:30	
Subject Sample	# 1	09:33	No Sample Provided
Breath Volume	0.000	09:33	
Air Blank	0.000	09:33	
Air Blank	0.000	09:35	
Subject Sample	# 2	09:36	Subject Test Refused
Breath Volume	0.000	09:36	
Air Blank	0.000	09:36	
Control Test	0.078	09:36	
Air Blank	0.000	09:37	
Diagnostics Check	OK	09:37	

Data Download Date/Time: 03/31/2025 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 231603
BREATH TEST OPERATOR: GOMEZ, DESTINEE, R	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BEAUREGARD, JEFFREY	ARREST TIME: 19:53
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Female
LAST AGENCY INSPECTION: 02/28/2025	OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: TAYLOR	DATE/TIME OF TEST: 03/28/2025	20:59:49
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	20:48
Air Blank	0.000	20:48
Control Test	0.078	20:48
Air Blank	0.000	20:49
Subject Sample # 1	0.313	20:50
Breath Volume	1.609	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.287	20:54
Breath Volume	1.480	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 3	0.327	20:58
Breath Volume	2.292	20:58
Air Blank	0.000	20:58
Control Test	0.078	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 03/31/2025 19:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 03/01/2025 07:09:03

Diagnostics Check	OK	07:06	
Air Blank	0.000	07:06	
Control Test	0.082	07:07	
Air Blank	0.000	07:07	
Subject Sample	# 1	07:07	Subject Test Refused
Breath Volume	0.000	07:07	
Air Blank	0.000	07:08	
Control Test	0.084	07:08	
Air Blank	0.000	07:08	
Diagnostics Check	OK	07:09	

Data Download Date/Time: 03/02/2025 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,DARRELL,C
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NORRIS,DARRELL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 03/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: LINHARES

DATE/TIME OF TEST: 03/20/2025 18:09:05

Diagnostics Check	OK	18:02
Air Blank	0.000	18:03
Control Test	0.081	18:03
Air Blank	0.000	18:04
Subject Sample # 1	0.181	18:04
Breath Volume	2.562	18:04
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.192	18:07
Breath Volume	3.960	18:07
Air Blank	0.000	18:08
Control Test	0.082	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 05/12/2025 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: LOPEZ GUILIBALDO

DATE/TIME OF TEST: 03/24/2025 18:53:56

Diagnostics Check	OK	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.182	18:49
Breath Volume	3.113	18:49
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.182	18:52
Breath Volume	1.968	18:52
Air Blank	0.000	18:53
Control Test	0.082	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 05/12/2025 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: WILLING

DATE/TIME OF TEST: 03/29/2025 18:24:37

Diagnostics Check	OK	18:16	
Air Blank	0.000	18:17	
Control Test	0.081	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 1 0.299	18:18	
Breath Volume	1.320	18:18	
Air Blank	0.000	18:19	
Air Blank	0.000	18:21	
Subject Sample	# 2	18:23	Subject Test Refused
Breath Volume	0.000	18:23	
Air Blank	0.000	18:23	
Control Test	0.082	18:24	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 05/12/2025 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: WALSTON

DATE/TIME OF TEST: 03/29/2025 22:37:27

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:31	
Air Blank	0.000	22:32	
Subject Sample	# 1 0.137	22:32	
Breath Volume	2.316	22:32	
Air Blank	0.000	22:33	
Air Blank	0.000	22:35	
Subject Sample	# 2 0.127	22:35	
Breath Volume	2.250	22:35	
Air Blank	0.000	22:36	
Control Test	0.080	22:36	
Air Blank		22:36	RFI Detect
Air Blank	0.000	22:37	

Data Download Date/Time: 05/12/2025 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MARAIS

DATE/TIME OF TEST: 03/01/2025 01:39:26

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:32	No Sample Provided
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Air Blank	0.000	01:34	
Subject Sample	# 2	01:38	No Sample Provided
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GASIEWICS

DATE/TIME OF TEST: 03/01/2025 02:57:25

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.119	02:50
Breath Volume	2.359	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.158	02:52
Breath Volume	2.394	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 3	0.143	02:55
Breath Volume	1.597	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: NIEVES ROBLES

DATE/TIME OF TEST: 03/01/2025 03:29:43

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.153	03:25
Breath Volume	1.570	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.153	03:28
Breath Volume	1.562	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/02/2025 01:29:01

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.087	01:24
Breath Volume	1.835	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.088	01:27
Breath Volume	1.253	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: KING,PHILIP
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ARES NUNEZ

DATE/TIME OF TEST: 03/02/2025 03:03:01

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.164	02:57
Breath Volume	1.644	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.156	03:01
Breath Volume	1.359	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/02/2025 05:50:09

Diagnostics Check	OK	05:49	
Air Blank		05:50	RFI Detect
Air Blank		05:50	RFI Detect

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/02/2025 06:07:38

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.081	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.146	06:03
Breath Volume	1.148	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.158	06:06
Breath Volume	1.734	06:06
Air Blank	0.000	06:06
Control Test	0.081	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:07

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: FLORIAN TRIGUEROS

DATE/TIME OF TEST: 03/02/2025 17:31:17

Diagnostics Check	OK	17:24
Air Blank	0.000	17:25
Control Test	0.082	17:25
Air Blank	0.000	17:25
Subject Sample # 1	0.209	17:26
Breath Volume	2.441	17:26
Air Blank	0.000	17:27
Air Blank	0.000	17:29
Subject Sample # 2	0.198	17:29
Breath Volume	1.589	17:29
Air Blank	0.000	17:30
Control Test	0.081	17:30
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MARTINEZ SANTIAGO

DATE/TIME OF TEST: 03/02/2025 20:38:28

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.081	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.196	20:34
Breath Volume	1.238	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.187	20:37
Breath Volume	1.343	20:37
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX, ANDREW, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOT, JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/02/2025 22:41:56

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.167	22:37
Breath Volume	1.382	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.169	22:40
Breath Volume	1.468	22:40
Air Blank	0.000	22:41
Control Test	0.081	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: PEREZ SEQUEIRA

DATE/TIME OF TEST: 03/02/2025 23:20:46

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.081	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:13	No Sample Provided
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample	# 2	23:19	No Sample Provided
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.081	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 03/25/2025 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/04/2025 21:42:44

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.198	21:34
Breath Volume	1.300	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.221	21:37
Breath Volume	2.496	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 3	0.216	21:41
Breath Volume	2.199	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEREDIA,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 03/06/2025 04:34:07

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.000	04:29
Breath Volume	2.222	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.000	04:32
Breath Volume	1.882	04:32
Air Blank	0.000	04:33
Control Test	0.082	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 03/06/2025 17:11:31

Diagnostics Check	OK	17:05
Air Blank	0.000	17:05
Control Test	0.080	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.280	17:07
Breath Volume	1.691	17:07
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.274	17:10
Breath Volume	1.863	17:10
Air Blank	0.000	17:10
Control Test	0.080	17:10
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, KIARA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS, WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: CROSDALE

DATE/TIME OF TEST: 03/06/2025 21:29:58

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.081	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.042	21:25
Breath Volume	1.785	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.040	21:28
Breath Volume	2.035	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: IACOBUCCI,JERE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/09/2025 04:22:19

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:11	
Control Test	0.080	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample	# 2	04:20	No Sample Provided
Breath Volume	0.000	04:20	
Air Blank	0.000	04:21	
Control Test	0.081	04:21	
Air Blank	0.000	04:22	
Diagnostics Check	OK	04:22	

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 03/09/2025 06:00:44

Diagnostics Check	OK	05:57	
Air Blank	0.000	05:57	
Control Test	0.081	05:58	
Air Blank	0.000	05:58	
Subject Sample	# 1 0.077	05:58	
Breath Volume	2.250	05:58	
Air Blank		05:59	RFI Detect
Air Blank	0.000	05:59	
Control Test	0.080	06:00	
Air Blank	0.000	06:00	
Diagnostics Check	OK	06:00	

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 03/09/2025 06:09:23

Diagnostics Check	OK	06:03
Air Blank	0.000	06:03
Control Test	0.081	06:04
Air Blank	0.000	06:04
Subject Sample # 1	0.083	06:04
Breath Volume	2.425	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.078	06:07
Breath Volume	2.316	06:07
Air Blank	0.000	06:08
Control Test	0.081	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,KEVIN,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ARBONEL,JOEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/10/2025 06:38:15

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:27	
Control Test	0.080	06:27	
Air Blank	0.000	06:28	
Subject Sample	# 1	06:31	No Sample Provided
Breath Volume	0.000	06:31	
Air Blank	0.000	06:31	
Air Blank	0.000	06:33	
Subject Sample	# 2	06:36	No Sample Provided
Breath Volume	0.000	06:36	
Air Blank	0.000	06:37	
Control Test	0.081	06:37	
Air Blank	0.000	06:38	
Diagnostics Check	OK	06:38	

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HERNANDEZ,EFRAIN
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 03/11/2025 21:55:58

Diagnostics Check	OK	21:48
Air Blank	0.000	21:49
Control Test	0.081	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.106	21:50
Breath Volume	1.265	21:50
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.121	21:54
Breath Volume	1.851	21:54
Air Blank	0.000	21:55
Control Test	0.082	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CRUZ-RAMIREZ

DATE/TIME OF TEST: 03/14/2025 22:13:09

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.081	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.000	22:08
Breath Volume	2.246	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.000	22:11
Breath Volume	2.589	22:11
Air Blank	0.000	22:12
Control Test	0.080	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/14/2025 23:09:56

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.000	23:05
Breath Volume	1.460	23:05
Air Blank	0.000	23:05
Air Blank	0.000	23:07
Subject Sample # 2	0.000	23:08
Breath Volume	1.500	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: KAYWORTH

DATE/TIME OF TEST: 03/15/2025 02:50:47

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.080	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Air Blank	0.000	02:46	
Subject Sample	# 2	02:49	No Sample Provided
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: MARTINEZ REYNOSA

DATE/TIME OF TEST: 03/16/2025 00:17:14

Diagnostics Check	OK	00:05	
Air Blank	0.000	00:05	
Control Test	0.080	00:06	
Air Blank	0.000	00:06	
Subject Sample	# 1 0.151	00:07	
Breath Volume	1.359	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample	# 2 0.144	00:12	Volume Not Met
Breath Volume	0.968	00:12	
Air Blank	0.000	00:13	
Air Blank	0.000	00:15	
Subject Sample	# 3 0.146	00:15	
Breath Volume	1.371	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 03/25/2025 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,KIARA,F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CARLOS

DATE/TIME OF TEST: 03/17/2025 01:30:27

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.081	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.212	01:26
Breath Volume	1.406	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.215	01:28
Breath Volume	1.406	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 03/17/2025 14:27:08

Diagnostics Check	OK	14:20
Air Blank	0.000	14:20
Control Test	0.080	14:21
Air Blank	0.000	14:21
Subject Sample # 1	0.000	14:22
Breath Volume	1.136	14:22
Air Blank	0.000	14:22
Air Blank	0.000	14:24
Subject Sample # 2	0.000	14:25
Breath Volume	1.476	14:25
Air Blank	0.000	14:26
Control Test	0.080	14:26
Air Blank	0.000	14:27
Diagnostics Check	OK	14:27

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: IACOBUCCI,JERE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 03/17/2025 21:20:13

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:09	
Control Test	0.081	21:09	
Air Blank	0.000	21:10	
Subject Sample	# 1	21:13	No Sample Provided
Breath Volume	0.000	21:13	
Air Blank	0.000	21:13	
Air Blank	0.000	21:15	
Subject Sample	# 2	21:18	No Sample Provided
Breath Volume	0.000	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Diagnostics Check	OK	21:20	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FARLESS,MITCHELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 03/17/2025 21:48:15

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:37	
Control Test	0.081	21:37	
Air Blank	0.000	21:38	
Subject Sample	# 1 0.214	21:41	Volume Not Met
Breath Volume	0.410	21:41	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 2 0.204	21:46	Volume Not Met
Breath Volume	0.207	21:46	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX, ANDREW, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS, NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 03/17/2025 23:45:02

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:34	
Control Test	0.081	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1 0.092	23:37	Volume Not Met
Breath Volume	0.179	23:37	
Air Blank	0.000	23:38	
Air Blank	0.000	23:40	
Subject Sample	# 2 0.105	23:43	Volume Not Met
Breath Volume	0.023	23:43	
Air Blank	0.000	23:44	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:45	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: DOSS JR.

DATE/TIME OF TEST: 03/18/2025 13:58:16

Diagnostics Check	OK	13:41	
Air Blank	0.000	13:41	
Control Test	0.081	13:41	
Air Blank	0.000	13:42	
Subject Sample	# 1 0.159	13:45	Volume Not Met
Breath Volume	0.636	13:45	
Air Blank	0.000	13:46	
Air Blank	0.000	13:47	
Subject Sample	# 2 0.181	13:51	
Breath Volume	1.101	13:51	
Air Blank	0.000	13:51	
Air Blank	0.000	13:53	
Subject Sample	# 3 0.160	13:56	Volume Not Met
Breath Volume	0.960	13:56	
Air Blank	0.000	13:57	
Control Test	0.079	13:57	
Air Blank	0.000	13:58	
Diagnostics Check	OK	13:58	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DELMORAL, NATANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PEREZ-MAQUEDA

DATE/TIME OF TEST: 03/19/2025 01:53:34

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.140	01:49
Breath Volume	1.914	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.134	01:52
Breath Volume	1.464	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:07

SUBJECT'S LAST NAME: DE SANTIAGO

DATE/TIME OF TEST: 03/19/2025 13:34:39

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.080	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.233	13:30
Breath Volume	1.921	13:30
Air Blank	0.000	13:30
Air Blank	0.000	13:32
Subject Sample # 2	0.233	13:33
Breath Volume	2.308	13:33
Air Blank	0.000	13:33
Control Test	0.080	13:34
Air Blank	0.000	13:34
Diagnostics Check	OK	13:34

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: TORRES FIGUEROA

DATE/TIME OF TEST: 03/19/2025 14:18:59

Diagnostics Check	OK	14:12
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:13
Subject Sample # 1	0.000	14:14
Breath Volume	2.484	14:14
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.000	14:17
Breath Volume	2.652	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 03/20/2025 02:02:09

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.081	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.167	01:57
Breath Volume	1.527	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.170	02:00
Breath Volume	1.726	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 03/20/2025 05:20:35

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:09	
Control Test	0.081	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:13	No Sample Provided
Breath Volume	0.000	05:13	
Air Blank	0.000	05:14	
Air Blank	0.000	05:16	
Subject Sample	# 2	05:19	No Sample Provided
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.081	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BENNETT JR

DATE/TIME OF TEST: 03/21/2025 04:17:26

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.081	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.177	04:12
Breath Volume	1.671	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.171	04:15
Breath Volume	1.527	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL, NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 03/22/2025 01:41:13

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.128	01:36
Breath Volume	2.359	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.135	01:39
Breath Volume	2.757	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL, NATANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 03/22/2025 05:10:56

Diagnostics Check	OK	04:59	
Air Blank	0.000	05:00	
Control Test	0.081	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:03	No Sample Provided
Breath Volume	0.000	05:03	
Air Blank	0.000	05:04	
Air Blank	0.000	05:06	
Subject Sample	# 2	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 03/25/2025 09:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VAZQUEZSARMIENTO

DATE/TIME OF TEST: 03/23/2025 04:18:43

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.117	04:14
Breath Volume	3.027	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.119	04:17
Breath Volume	2.890	04:17
Air Blank	0.000	04:17
Control Test	0.081	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 03/25/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEEHAN,KENNETH,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MCGEEHAN,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: AKINYI

DATE/TIME OF TEST: 03/23/2025 14:32:16

Diagnostics Check	OK	14:25
Air Blank	0.000	14:26
Control Test	0.080	14:26
Air Blank	0.000	14:26
Subject Sample # 1	0.183	14:27
Breath Volume	1.308	14:27
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.187	14:30
Breath Volume	1.457	14:30
Air Blank	0.000	14:31
Control Test	0.079	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 03/25/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PERALTA,ALLISON
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HAWKINS III

DATE/TIME OF TEST: 03/23/2025 21:45:43

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.081	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.008	21:40
Breath Volume	1.210	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:44
Breath Volume	1.160	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:45
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 03/25/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CASTRO TORRES

DATE/TIME OF TEST: 03/24/2025 01:43:36

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:36	No Sample Provided
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 2	01:42	No Sample Provided
Breath Volume	0.000	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:43	
Air Blank	0.000	01:43	
Diagnostics Check	OK	01:43	

Data Download Date/Time: 03/25/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 02/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: OVALLE

DATE/TIME OF TEST: 03/25/2025 02:15:10

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.082	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:10
Breath Volume	2.546	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:13
Breath Volume	2.363	02:13
Air Blank	0.000	02:14
Control Test	0.081	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 03/25/2025 09:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECATES,C,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HUNTER,J
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 03/02/2025 22:12:18

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.126	22:07
Breath Volume	1.906	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.125	22:10
Breath Volume	1.855	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 03/26/2025 18:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEELEY,JAMES,E
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: LIRIANO,JULIO
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: BLISS

DATE/TIME OF TEST: 03/03/2025 08:16:33

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.079	08:11
Air Blank	0.000	08:11
Subject Sample # 1	0.155	08:12
Breath Volume	1.714	08:12
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.156	08:15
Breath Volume	1.253	08:15
Air Blank	0.000	08:15
Control Test	0.080	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 03/26/2025 18:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,HALEY,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: OGLE,TRISTEN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WALTZ

DATE/TIME OF TEST: 03/29/2025 23:44:22

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.199	23:40
Breath Volume	3.800	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.195	23:42
Breath Volume	2.503	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 04/28/2025 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SCHRAMM,JUSTON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 14:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: PREVATT

DATE/TIME OF TEST: 03/02/2025 17:19:19

Diagnostics Check	OK	17:12
Air Blank	0.000	17:12
Control Test	0.080	17:12
Air Blank	0.000	17:13
Subject Sample # 1	0.266	17:14
Breath Volume	1.179	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:16
Subject Sample # 2	0.272	17:17
Breath Volume	1.894	17:17
Air Blank	0.000	17:18
Control Test	0.080	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 03/26/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILL,CAMRYN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WATLINGTON,NOAH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 03/08/2025 05:24:19

Diagnostics Check	OK	05:17
Air Blank	0.000	05:18
Control Test	0.081	05:18
Air Blank	0.000	05:18
Subject Sample # 1	0.212	05:19
Breath Volume	1.847	05:19
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.210	05:22
Breath Volume	1.496	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 03/26/2025 18:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 03/10/2025 23:48:37

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.217	23:44
Breath Volume	2.160	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.224	23:47
Breath Volume	2.457	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/26/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: KIDD

DATE/TIME OF TEST: 03/21/2025 02:32:12

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.242	02:27
Breath Volume	1.210	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.247	02:30
Breath Volume	3.410	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 03/26/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CLARK,ADDISON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: LEAGLE

DATE/TIME OF TEST: 03/22/2025 08:10:42

Diagnostics Check	OK	08:04
Air Blank	0.000	08:04
Control Test	0.081	08:05
Air Blank	0.000	08:05
Subject Sample # 1	0.204	08:06
Breath Volume	1.605	08:06
Air Blank	0.000	08:06
Air Blank	0.000	08:08
Subject Sample # 2	0.188	08:09
Breath Volume	1.675	08:09
Air Blank	0.000	08:09
Control Test	0.080	08:10
Air Blank	0.000	08:10
Diagnostics Check	OK	08:10

Data Download Date/Time: 03/26/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: NEILY,STEVEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 02/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: TALAVERA

DATE/TIME OF TEST: 03/23/2025 03:56:44

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.141	03:52
Breath Volume	1.855	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.142	03:55
Breath Volume	2.066	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 03/26/2025 18:29