

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD, GLENN, B  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: LANGSTON, BENJAMIN  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MASTERSON

DATE/TIME OF TEST: 11/15/2024 21:48:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:41 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Subject Sample # 1 | 0.196 | 21:43 |
| Breath Volume      | 2.230 | 21:43 |
| Air Blank          | 0.000 | 21:44 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 2 | 0.195 | 21:46 |
| Breath Volume      | 1.535 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Control Test       | 0.078 | 21:47 |
| Air Blank          | 0.000 | 21:48 |
| Diagnostics Check  | OK    | 21:48 |

---

Data Download Date/Time: 01/03/2025 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEHEAD,BRANDON,A  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: CALLAWAY,B  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: LAMOREE

DATE/TIME OF TEST: 11/18/2024 23:13:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Control Test       | 0.080 | 23:08 |
| Air Blank          | 0.000 | 23:08 |
| Subject Sample # 1 | 0.097 | 23:09 |
| Breath Volume      | 2.054 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Subject Sample # 2 | 0.089 | 23:12 |
| Breath Volume      | 2.078 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.080 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Diagnostics Check  | OK    | 23:13 |

---

Data Download Date/Time: 01/03/2025 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,L,S  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: NORSWORTHY,B  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SNIPES

DATE/TIME OF TEST: 11/19/2024 21:35:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:32 |                      |
| Air Blank         | 0.000 | 21:33 |                      |
| Control Test      | 0.080 | 21:33 |                      |
| Air Blank         | 0.000 | 21:33 |                      |
| Subject Sample    | # 1   | 21:34 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:34 |                      |
| Air Blank         | 0.000 | 21:34 |                      |
| Control Test      | 0.080 | 21:34 |                      |
| Air Blank         | 0.000 | 21:35 |                      |
| Diagnostics Check | OK    | 21:35 |                      |

---

Data Download Date/Time: 01/03/2025 13:36

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,GLENN,B  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: WHITE,LAGUESHA  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 05:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 11/24/2024 06:43:42

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 06:32 |                    |
| Air Blank         | 0.000 | 06:32 |                    |
| Control Test      | 0.080 | 06:33 |                    |
| Air Blank         | 0.000 | 06:33 |                    |
| Subject Sample    | # 1   | 06:36 | No Sample Provided |
| Breath Volume     | 0.000 | 06:36 |                    |
| Air Blank         | 0.000 | 06:37 |                    |
| Air Blank         | 0.000 | 06:39 |                    |
| Subject Sample    | # 2   | 06:42 | No Sample Provided |
| Breath Volume     | 0.000 | 06:42 |                    |
| Air Blank         | 0.000 | 06:42 |                    |
| Control Test      | 0.080 | 06:43 |                    |
| Air Blank         | 0.000 | 06:43 |                    |
| Diagnostics Check | OK    | 06:43 |                    |

---

Data Download Date/Time: 01/03/2025 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SEELEY,CPL,V  
BREATH TEST OPERATOR AGENCY: Bradford County SO  
ARRESTING OFFICER: JONES,DEP  
ARRESTING OFFICER AGENCY: Bradford County SO  
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: CULVERSON

DATE/TIME OF TEST: 11/30/2024 07:56:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:50 |
| Air Blank          | 0.000 | 07:50 |
| Control Test       | 0.081 | 07:51 |
| Air Blank          | 0.000 | 07:51 |
| Subject Sample # 1 | 0.000 | 07:52 |
| Breath Volume      | 2.804 | 07:52 |
| Air Blank          | 0.000 | 07:52 |
| Air Blank          | 0.000 | 07:54 |
| Subject Sample # 2 | 0.000 | 07:54 |
| Breath Volume      | 2.472 | 07:54 |
| Air Blank          | 0.000 | 07:55 |
| Control Test       | 0.080 | 07:55 |
| Air Blank          | 0.000 | 07:56 |
| Diagnostics Check  | OK    | 07:56 |

---

Data Download Date/Time: 01/03/2025 13:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: COURT ORD  
ARREST TIME: 08:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/07/2024 08:56:55

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 08:50 |            |
| Air Blank          | 0.000 | 08:51 |            |
| Control Test       | 0.079 | 08:51 |            |
| Air Blank          | 0.000 | 08:51 |            |
| Subject Sample # 1 | 0.000 | 08:52 |            |
| Breath Volume      | 2.617 | 08:52 |            |
| Air Blank          | 0.000 | 08:53 |            |
| Air Blank          | 0.000 | 08:55 |            |
| Subject Sample # 2 | 0.000 | 08:55 |            |
| Breath Volume      | 1.972 | 08:55 |            |
| Air Blank          | 0.000 | 08:56 |            |
|                    |       | 08:56 | RFI Detect |
| Air Blank          | 0.000 | 08:56 |            |

---

Data Download Date/Time: 12/09/2024 17:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: COURT ORD  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 11/07/2024 09:27:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:20 |
| Air Blank          | 0.000 | 09:21 |
| Control Test       | 0.077 | 09:21 |
| Air Blank          | 0.000 | 09:22 |
| Subject Sample # 1 | 0.000 | 09:23 |
| Breath Volume      | 1.792 | 09:23 |
| Air Blank          | 0.000 | 09:23 |
| Air Blank          | 0.000 | 09:25 |
| Subject Sample # 2 | 0.000 | 09:26 |
| Breath Volume      | 2.054 | 09:26 |
| Air Blank          | 0.000 | 09:26 |
| Control Test       | 0.077 | 09:26 |
| Air Blank          | 0.000 | 09:27 |
| Diagnostics Check  | OK    | 09:27 |

---

Data Download Date/Time: 12/09/2024 17:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: COURT ORD  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 11/07/2024 10:27:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:20 |
| Air Blank          | 0.000 | 10:21 |
| Control Test       | 0.078 | 10:21 |
| Air Blank          | 0.000 | 10:21 |
| Subject Sample # 1 | 0.000 | 10:22 |
| Breath Volume      | 1.335 | 10:22 |
| Air Blank          | 0.000 | 10:23 |
| Air Blank          | 0.000 | 10:25 |
| Subject Sample # 2 | 0.000 | 10:26 |
| Breath Volume      | 1.914 | 10:26 |
| Air Blank          | 0.000 | 10:26 |
| Control Test       | 0.077 | 10:27 |
| Air Blank          | 0.000 | 10:27 |
| Diagnostics Check  | OK    | 10:27 |

---

Data Download Date/Time: 12/09/2024 17:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: EVANS,STEVEN,J

BREATH TEST OPERATOR AGENCY: Haines City PD

ARRESTING OFFICER: EVANS,STEVEN

ARRESTING OFFICER AGENCY: Haines City PD

LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402783528

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: PROB/PAROL

ARREST TIME: 11:12

GENDER: Male

OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: VELEZ LEON

DATE/TIME OF TEST: 11/07/2024 12:05:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:59 |
| Air Blank          | 0.000 | 12:00 |
| Control Test       | 0.077 | 12:00 |
| Air Blank          | 0.000 | 12:01 |
| Subject Sample # 1 | 0.000 | 12:01 |
| Breath Volume      | 1.894 | 12:01 |
| Air Blank          | 0.000 | 12:01 |
| Air Blank          | 0.000 | 12:03 |
| Subject Sample # 2 | 0.000 | 12:04 |
| Breath Volume      | 1.621 | 12:04 |
| Air Blank          | 0.000 | 12:04 |
| Control Test       | 0.078 | 12:04 |
| Air Blank          | 0.000 | 12:05 |
| Diagnostics Check  | OK    | 12:05 |

---

Data Download Date/Time: 12/09/2024 17:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 11:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:44

SUBJECT'S LAST NAME: MCGRUFF

DATE/TIME OF TEST: 11/07/2024 12:21:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:15 |
| Air Blank          | 0.000 | 12:16 |
| Control Test       | 0.077 | 12:16 |
| Air Blank          | 0.000 | 12:16 |
| Subject Sample # 1 | 0.000 | 12:17 |
| Breath Volume      | 1.535 | 12:17 |
| Air Blank          | 0.000 | 12:17 |
| Air Blank          | 0.000 | 12:19 |
| Subject Sample # 2 | 0.000 | 12:20 |
| Breath Volume      | 1.777 | 12:20 |
| Air Blank          | 0.000 | 12:20 |
| Control Test       | 0.078 | 12:20 |
| Air Blank          | 0.000 | 12:21 |
| Diagnostics Check  | OK    | 12:21 |

---

Data Download Date/Time: 12/09/2024 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 15:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 11/07/2024 16:57:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:51 |
| Air Blank          | 0.000 | 16:51 |
| Control Test       | 0.078 | 16:51 |
| Air Blank          | 0.000 | 16:52 |
| Subject Sample # 1 | 0.000 | 16:52 |
| Breath Volume      | 1.132 | 16:52 |
| Air Blank          | 0.000 | 16:53 |
| Air Blank          | 0.000 | 16:55 |
| Subject Sample # 2 | 0.000 | 16:55 |
| Breath Volume      | 1.378 | 16:55 |
| Air Blank          | 0.000 | 16:56 |
| Control Test       | 0.078 | 16:56 |
| Air Blank          | 0.000 | 16:57 |
| Diagnostics Check  | OK    | 16:57 |

---

Data Download Date/Time: 12/09/2024 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: MOREL-CAMPOS,ROMMY  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402783528  
DRY GAS EXPIRY DATE: 06/27/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WEEMS

DATE/TIME OF TEST: 11/17/2024 14:35:34

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 14:21 |                           |
| Air Blank         | 0.000 | 14:22 |                           |
| Control Test      | 0.077 | 14:22 |                           |
| Air Blank         | 0.000 | 14:23 |                           |
| Subject Sample    | # 1   | 14:26 | No Sample Provided        |
| Breath Volume     | 0.000 | 14:26 |                           |
| Air Blank         | 0.000 | 14:26 |                           |
| Air Blank         | 0.000 | 14:28 |                           |
| Subject Sample    | # 2   | 0.269 | 14:30                     |
| Breath Volume     | 2.031 | 14:30 |                           |
| Air Blank         | 0.000 | 14:31 |                           |
| Air Blank         | 0.000 | 14:33 |                           |
| Subject Sample    | # 3   | 0.283 | 14:33                     |
| Breath Volume     | 1.796 | 14:33 |                           |
| Air Blank         | 0.000 | 14:34 |                           |
| Control Test      | 0.074 | 14:34 | Control Outside Tolerance |
| Air Blank         | 0.000 | 14:35 |                           |

Data Download Date/Time: 12/09/2024 17:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: D'AGOSTINO,STEFANO,V  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: HILLIKER,COREY  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: AVANT

DATE/TIME OF TEST: 11/28/2024 00:41:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.083 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.126 | 00:35 |
| Breath Volume      | 1.230 | 00:35 |
| Air Blank          | 0.000 | 00:36 |
| Air Blank          | 0.000 | 00:38 |
| Subject Sample # 2 | 0.135 | 00:39 |
| Breath Volume      | 1.347 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.081 | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Diagnostics Check  | OK    | 00:41 |

---

Data Download Date/Time: 12/12/2024 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: UPSHAW,JASON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: KIGHT,GARRET  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 12:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 11/01/2024 02:07:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:04 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Control Test      | 0.082 | 02:05 |                      |
| Air Blank         | 0.000 | 02:05 |                      |
| Subject Sample    | # 1   | 02:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:06 |                      |
| Air Blank         | 0.000 | 02:07 |                      |
| Control Test      | 0.081 | 02:07 |                      |
| Air Blank         | 0.000 | 02:07 |                      |
| Diagnostics Check | OK    | 02:07 |                      |

---

Data Download Date/Time: 11/12/2024 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DELANCY,JARRELL,M  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: OSBORNE,OWEN  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LABELLMAN

DATE/TIME OF TEST: 11/01/2024 05:57:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:51 |
| Air Blank          | 0.000 | 05:51 |
| Control Test       | 0.082 | 05:52 |
| Air Blank          | 0.000 | 05:52 |
| Subject Sample # 1 | 0.188 | 05:52 |
| Breath Volume      | 3.312 | 05:52 |
| Air Blank          | 0.000 | 05:53 |
| Air Blank          | 0.000 | 05:55 |
| Subject Sample # 2 | 0.185 | 05:55 |
| Breath Volume      | 2.015 | 05:55 |
| Air Blank          | 0.000 | 05:56 |
| Control Test       | 0.081 | 05:57 |
| Air Blank          | 0.000 | 05:57 |
| Diagnostics Check  | OK    | 05:57 |

---

Data Download Date/Time: 11/12/2024 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRAMER,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GARRIDO,ANTHONY  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BARTOLON

DATE/TIME OF TEST: 11/02/2024 02:33:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.082 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Subject Sample # 1 | 0.240 | 02:25 |
| Breath Volume      | 2.058 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 2 | 0.214 | 02:29 |
| Breath Volume      | 1.574 | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 3 | 0.234 | 02:31 |
| Breath Volume      | 1.460 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.079 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Diagnostics Check  | OK    | 02:33 |

---

Data Download Date/Time: 11/12/2024 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KRAMER,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: STAFFORD,ROBERT  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/02/2024 04:51:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Control Test       | 0.082 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 1 | 0.200 | 04:47 |
| Breath Volume      | 2.035 | 04:47 |
| Air Blank          | 0.000 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Subject Sample # 2 | 0.201 | 04:50 |
| Breath Volume      | 2.058 | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.081 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Diagnostics Check  | OK    | 04:51 |

---

Data Download Date/Time: 11/12/2024 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVIN,TOMMY,L  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: MESHES,GREG  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MERRIWEATHER

DATE/TIME OF TEST: 11/02/2024 05:16:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:13 |                      |
| Air Blank         | 0.000 | 05:14 |                      |
| Control Test      | 0.081 | 05:14 |                      |
| Air Blank         | 0.000 | 05:15 |                      |
| Subject Sample    | # 1   | 05:15 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:15 |                      |
| Air Blank         | 0.000 | 05:15 |                      |
| Control Test      | 0.081 | 05:16 |                      |
| Air Blank         | 0.000 | 05:16 |                      |
| Diagnostics Check | OK    | 05:16 |                      |

---

Data Download Date/Time: 11/12/2024 11:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: HESKETT,KIM  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: DOBY

DATE/TIME OF TEST: 11/02/2024 23:51:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.082 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Subject Sample # 1 | 0.161 | 23:47 |
| Breath Volume      | 2.109 | 23:47 |
| Air Blank          | 0.000 | 23:47 |
| Air Blank          | 0.000 | 23:49 |
| Subject Sample # 2 | 0.160 | 23:50 |
| Breath Volume      | 2.011 | 23:50 |
| Air Blank          | 0.000 | 23:50 |
| Control Test       | 0.080 | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Diagnostics Check  | OK    | 23:51 |

---

Data Download Date/Time: 11/12/2024 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LITZKOW,CARLTON,D  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: LITZKOW,CARLTON  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: KLEEHAMMER

DATE/TIME OF TEST: 11/03/2024 22:46:09

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:34 |                    |
| Air Blank         | 0.000 | 22:35 |                    |
| Control Test      | 0.081 | 22:35 |                    |
| Air Blank         | 0.000 | 22:36 |                    |
| Subject Sample    | # 1   | 22:39 | No Sample Provided |
| Breath Volume     | 0.000 | 22:39 |                    |
| Air Blank         | 0.000 | 22:39 |                    |
| Air Blank         | 0.000 | 22:41 |                    |
| Subject Sample    | # 2   | 22:44 | No Sample Provided |
| Breath Volume     | 0.000 | 22:44 |                    |
| Air Blank         | 0.000 | 22:45 |                    |
| Control Test      | 0.081 | 22:45 |                    |
| Air Blank         | 0.000 | 22:46 |                    |
| Diagnostics Check | OK    | 22:46 |                    |

---

Data Download Date/Time: 11/12/2024 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MORRIS,ERIC  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 11/09/2024 08:54:24

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 08:43 |                    |
| Air Blank         | 0.000 | 08:43 |                    |
| Control Test      | 0.083 | 08:43 |                    |
| Air Blank         | 0.000 | 08:44 |                    |
| Subject Sample    | # 1   | 08:47 | No Sample Provided |
| Breath Volume     | 0.000 | 08:47 |                    |
| Air Blank         | 0.000 | 08:47 |                    |
| Air Blank         | 0.000 | 08:49 |                    |
| Subject Sample    | # 2   | 08:53 | No Sample Provided |
| Breath Volume     | 0.000 | 08:53 |                    |
| Air Blank         | 0.000 | 08:53 |                    |
| Control Test      | 0.083 | 08:53 |                    |
| Air Blank         | 0.000 | 08:54 |                    |
| Diagnostics Check | OK    | 08:54 |                    |

---

Data Download Date/Time: 11/12/2024 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD,ANTHONY,M  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: MORRIS,ERIC  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 07:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 11/09/2024 09:03:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:56 |
| Air Blank          | 0.000 | 08:57 |
| Control Test       | 0.083 | 08:57 |
| Air Blank          | 0.000 | 08:58 |
| Subject Sample # 1 | 0.112 | 08:58 |
| Breath Volume      | 2.917 | 08:58 |
| Air Blank          | 0.000 | 08:59 |
| Air Blank          | 0.000 | 09:01 |
| Subject Sample # 2 | 0.110 | 09:01 |
| Breath Volume      | 2.175 | 09:01 |
| Air Blank          | 0.000 | 09:02 |
| Control Test       | 0.081 | 09:02 |
| Air Blank          | 0.000 | 09:03 |
| Diagnostics Check  | OK    | 09:03 |

---

Data Download Date/Time: 11/12/2024 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STEMEN,AARON,M  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: STEMEN,AARON  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RATHBUN

DATE/TIME OF TEST: 11/10/2024 19:25:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:15 |
| Air Blank          | 0.000 | 19:15 |
| Control Test       | 0.082 | 19:16 |
| Air Blank          | 0.000 | 19:16 |
| Subject Sample # 1 | 0.254 | 19:17 |
| Breath Volume      | 2.320 | 19:17 |
| Air Blank          | 0.000 | 19:18 |
| Air Blank          | 0.000 | 19:19 |
| Subject Sample # 2 | 0.233 | 19:20 |
| Breath Volume      | 1.824 | 19:20 |
| Air Blank          | 0.000 | 19:20 |
| Air Blank          | 0.000 | 19:22 |
| Subject Sample # 3 | 0.247 | 19:23 |
| Breath Volume      | 2.257 | 19:23 |
| Air Blank          | 0.000 | 19:24 |
| Control Test       | 0.082 | 19:24 |
| Air Blank          | 0.000 | 19:25 |
| Diagnostics Check  | OK    | 19:25 |

---

Data Download Date/Time: 11/12/2024 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTE,JUSTIN,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: COTE,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 11/15/2024 02:12:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.081 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 1 | 0.170 | 02:08 |
| Breath Volume      | 9.375 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.163 | 02:11 |
| Breath Volume      | 8.937 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.082 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 12/26/2024 14:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITTINGTON,CAROLINE,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WHITTINGTON,CAROLINE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 11/16/2024 20:26:47

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:23 |                      |
| Air Blank         | 0.000 | 20:24 |                      |
| Control Test      | 0.081 | 20:24 |                      |
| Air Blank         | 0.000 | 20:24 |                      |
| Subject Sample    | # 1   | 20:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:25 |                      |
| Air Blank         | 0.000 | 20:25 |                      |
| Control Test      | 0.083 | 20:26 |                      |
| Air Blank         | 0.000 | 20:26 |                      |
| Diagnostics Check | OK    | 20:26 |                      |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: MESHES,GREGORY  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/17/2024 00:44:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.083 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 1 | 0.160 | 00:40 |
| Breath Volume      | 4.765 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 2 | 0.158 | 00:43 |
| Breath Volume      | 6.058 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.082 | 00:44 |
| Air Blank          | 0.000 | 00:44 |
| Diagnostics Check  | OK    | 00:44 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OSBOURNE,DANTRELL,S  
BREATH TEST OPERATOR AGENCY: Gainesville PD  
ARRESTING OFFICER: JIN,KEVIN  
ARRESTING OFFICER AGENCY: Gainesville PD  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 11/17/2024 02:35:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.082 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.136 | 02:31 |
| Breath Volume      | 5.042 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.142 | 02:34 |
| Breath Volume      | 7.152 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.082 | 02:35 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COTE,JUSTIN,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: COTE,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LASSITER

DATE/TIME OF TEST: 11/19/2024 03:45:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Control Test       | 0.082 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Subject Sample # 1 | 0.224 | 03:40 |
| Breath Volume      | 4.617 | 03:40 |
| Air Blank          | 0.000 | 03:41 |
| Air Blank          | 0.000 | 03:43 |
| Subject Sample # 2 | 0.218 | 03:44 |
| Breath Volume      | 4.195 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Control Test       | 0.082 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Diagnostics Check  | OK    | 03:45 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASH,EARL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MASH,EARL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/23/2024 00:46:51

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Control Test       | 0.083 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.130 | 00:42 |
| Breath Volume      | 2.429 | 00:42 |
| Air Blank          | 0.000 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 2 | 0.137 | 00:45 |
| Breath Volume      | 2.582 | 00:45 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.083 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Diagnostics Check  | OK    | 00:46 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRIS,DILLON,J  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: VARGAS,CESAR  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 11/24/2024 16:53:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:44 |
| Air Blank          | 0.000 | 16:45 |
| Control Test       | 0.082 | 16:45 |
| Air Blank          | 0.000 | 16:45 |
| Subject Sample # 1 | 0.237 | 16:47 |
| Breath Volume      | 2.304 | 16:47 |
| Air Blank          | 0.000 | 16:48 |
| Air Blank          | 0.000 | 16:50 |
| Subject Sample # 2 | 0.233 | 16:51 |
| Breath Volume      | 1.824 | 16:51 |
| Air Blank          | 0.000 | 16:52 |
| Control Test       | 0.080 | 16:52 |
| Air Blank          | 0.000 | 16:53 |
| Diagnostics Check  | OK    | 16:53 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAHARAJ, KRISHNA, H  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: GARRIDO, ANTHONY  
ARRESTING OFFICER AGENCY: Alachua PD  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 11/25/2024 21:57:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Control Test       | 0.082 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Subject Sample # 1 | 0.224 | 21:52 |
| Breath Volume      | 1.632 | 21:52 |
| Air Blank          | 0.000 | 21:53 |
| Air Blank          | 0.000 | 21:55 |
| Subject Sample # 2 | 0.228 | 21:55 |
| Breath Volume      | 1.140 | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.080 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Diagnostics Check  | OK    | 21:57 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MASH,EARL,C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MASH,EARL  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 11/29/2024 00:11:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Control Test       | 0.083 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Subject Sample # 1 | 0.000 | 00:07 |
| Breath Volume      | 2.750 | 00:07 |
| Air Blank          | 0.000 | 00:08 |
| Air Blank          | 0.000 | 00:10 |
| Subject Sample # 2 | 0.000 | 00:10 |
| Breath Volume      | 2.867 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.082 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 12/26/2024 14:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZANT,TANNER,T  
BREATH TEST OPERATOR AGENCY: Okaloosa County SO  
ARRESTING OFFICER: VANNIEUWENHOVEN,SCOTT  
ARRESTING OFFICER AGENCY: Okaloosa County SO  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 11/27/2024 05:12:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Control Test       | 0.078 | 05:07 |
| Air Blank          | 0.000 | 05:07 |
| Subject Sample # 1 | 0.161 | 05:07 |
| Breath Volume      | 2.238 | 05:07 |
| Air Blank          | 0.000 | 05:08 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 2 | 0.167 | 05:10 |
| Breath Volume      | 2.070 | 05:10 |
| Air Blank          | 0.000 | 05:11 |
| Control Test       | 0.078 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Diagnostics Check  | OK    | 05:12 |

---

Data Download Date/Time: 12/27/2024 11:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MESTEK,JOHN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: MESTEK,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CID VEGA

DATE/TIME OF TEST: 11/03/2024 02:36:50

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:25 |                    |
| Air Blank         | 0.000 | 02:25 |                    |
| Control Test      | 0.080 | 02:26 |                    |
| Air Blank         | 0.000 | 02:26 |                    |
| Subject Sample    | # 1   | 02:29 | No Sample Provided |
| Breath Volume     | 0.000 | 02:29 |                    |
| Air Blank         | 0.000 | 02:30 |                    |
| Air Blank         | 0.000 | 02:32 |                    |
| Subject Sample    | # 2   | 02:35 | No Sample Provided |
| Breath Volume     | 0.000 | 02:35 |                    |
| Air Blank         | 0.000 | 02:35 |                    |
| Control Test      | 0.080 | 02:36 |                    |
| Air Blank         | 0.000 | 02:36 |                    |
| Diagnostics Check | OK    | 02:36 |                    |

---

Data Download Date/Time: 11/05/2024 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BABEL,NICHOLAS,S  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: BABEL,NICHOLAS  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/16/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 11/03/2024 14:11:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:05 |
| Air Blank          | 0.000 | 14:05 |
| Control Test       | 0.080 | 14:05 |
| Air Blank          | 0.000 | 14:06 |
| Subject Sample # 1 | 0.198 | 14:07 |
| Breath Volume      | 1.226 | 14:07 |
| Air Blank          | 0.000 | 14:07 |
| Air Blank          | 0.000 | 14:09 |
| Subject Sample # 2 | 0.203 | 14:10 |
| Breath Volume      | 1.425 | 14:10 |
| Air Blank          | 0.000 | 14:10 |
| Control Test       | 0.079 | 14:11 |
| Air Blank          | 0.000 | 14:11 |
| Diagnostics Check  | OK    | 14:11 |

---

Data Download Date/Time: 11/05/2024 11:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORRE, ADAM, J  
BREATH TEST OPERATOR AGENCY: Escambia County SO  
ARRESTING OFFICER: LEGGETT, GARRETT  
ARRESTING OFFICER AGENCY: Escambia County SO  
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 11/08/2024 20:09:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:00 |
| Air Blank          | 0.000 | 20:01 |
| Control Test       | 0.080 | 20:01 |
| Air Blank          | 0.000 | 20:02 |
| Subject Sample # 1 | 0.138 | 20:03 |
| Breath Volume      | 1.328 | 20:03 |
| Air Blank          | 0.000 | 20:04 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 2 | 0.156 | 20:07 |
| Breath Volume      | 1.496 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Control Test       | 0.078 | 20:08 |
| Air Blank          | 0.000 | 20:09 |
| Diagnostics Check  | OK    | 20:09 |

---

Data Download Date/Time: 12/10/2024 10:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PLEASANTS,STEVEN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: PLEASANTS,STEVEN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402690226  
DRY GAS EXPIRY DATE: 03/09/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 11/23/2024 23:16:46

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 23:07 |            |
| Air Blank         | 0.000     | 23:08 |            |
| Control Test      | 0.081     | 23:08 |            |
| Air Blank         | 0.000     | 23:08 |            |
| Subject Sample    | # 1 0.179 | 23:11 |            |
| Breath Volume     | 1.464     | 23:11 |            |
| Air Blank         | 0.000     | 23:12 |            |
| Air Blank         | 0.000     | 23:13 |            |
| Subject Sample    | # 2 0.172 | 23:16 |            |
| Breath Volume     | 1.488     | 23:16 |            |
| Air Blank         |           | 23:16 | RFI Detect |
| Air Blank         | 0.000     | 23:16 |            |

---

Data Download Date/Time: 12/10/2024 10:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRISP,JESSICA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CRISP,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: S315585795510

DATE/TIME OF TEST: 11/06/2024 18:50:54

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 18:49 |            |
| Air Blank         | 0.000 | 18:49 |            |
| Control Test      | 0.078 | 18:50 |            |
| Air Blank         |       | 18:50 | RFI Detect |
| Air Blank         | 0.000 | 18:50 |            |

---

Data Download Date/Time: 11/11/2024 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRISP,JESSICA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CRISP,JESSICA  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 18:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 11/06/2024 19:00:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:53 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.078 | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Subject Sample # 1 | 0.016 | 18:55 |
| Breath Volume      | 2.898 | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.012 | 18:58 |
| Breath Volume      | 1.816 | 18:58 |
| Air Blank          | 0.000 | 18:59 |
| Control Test       | 0.077 | 18:59 |
| Air Blank          | 0.000 | 18:59 |
| Diagnostics Check  | OK    | 19:00 |

---

Data Download Date/Time: 11/11/2024 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,GARY,E  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DANIELS,GARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/02/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MCNEALY

DATE/TIME OF TEST: 11/08/2024 21:30:02

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 21:18 |                |
| Air Blank         | 0.000     | 21:18 |                |
| Control Test      | 0.078     | 21:19 |                |
| Air Blank         | 0.000     | 21:19 |                |
| Subject Sample    | # 1 0.180 | 21:22 | Volume Not Met |
| Breath Volume     | 0.335     | 21:22 |                |
| Air Blank         | 0.000     | 21:23 |                |
| Air Blank         | 0.000     | 21:25 |                |
| Subject Sample    | # 2 0.180 | 21:28 | Volume Not Met |
| Breath Volume     | 0.507     | 21:28 |                |
| Air Blank         | 0.000     | 21:29 |                |
| Control Test      | 0.078     | 21:29 |                |
| Air Blank         | 0.000     | 21:29 |                |
| Diagnostics Check | OK        | 21:30 |                |

---

Data Download Date/Time: 11/11/2024 15:51

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUSICK,ROBERT,C  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: BUSICK,ROBERT  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/13/2024 23:13:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:10 |                      |
| Air Blank         | 0.000 | 23:11 |                      |
| Control Test      | 0.080 | 23:11 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Subject Sample    | # 1   | 23:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:12 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Control Test      | 0.078 | 23:13 |                      |
| Air Blank         | 0.000 | 23:13 |                      |
| Diagnostics Check | OK    | 23:13 |                      |

---

Data Download Date/Time: 12/26/2024 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,TRUETT,L  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: CORBIN,TRUETT  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: BERNINGER

DATE/TIME OF TEST: 11/20/2024 20:51:17

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 20:45 |            |
| Air Blank          | 0.000 | 20:45 |            |
| Control Test       | 0.079 | 20:46 |            |
| Air Blank          | 0.000 | 20:46 |            |
| Subject Sample # 1 | 0.000 | 20:47 |            |
| Breath Volume      | 3.355 | 20:47 |            |
| Air Blank          | 0.000 | 20:47 |            |
| Air Blank          | 0.000 | 20:49 |            |
| Subject Sample # 2 | 0.000 | 20:50 |            |
| Breath Volume      | 2.242 | 20:50 |            |
| Air Blank          | 0.000 | 20:50 |            |
| Control Test       | 0.079 | 20:50 | RFI Detect |
| Air Blank          | 0.000 | 20:51 |            |

---

Data Download Date/Time: 12/26/2024 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DANIELS,GARY,E  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: DANIELS,GARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 09:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/23/2024 11:08:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:02 |
| Air Blank          | 0.000 | 11:02 |
| Control Test       | 0.078 | 11:02 |
| Air Blank          | 0.000 | 11:03 |
| Subject Sample # 1 | 0.207 | 11:03 |
| Breath Volume      | 2.398 | 11:03 |
| Air Blank          | 0.000 | 11:04 |
| Air Blank          | 0.000 | 11:06 |
| Subject Sample # 2 | 0.213 | 11:06 |
| Breath Volume      | 1.972 | 11:06 |
| Air Blank          | 0.000 | 11:07 |
| Control Test       | 0.079 | 11:07 |
| Air Blank          | 0.000 | 11:08 |
| Diagnostics Check  | OK    | 11:08 |

---

Data Download Date/Time: 12/26/2024 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,TRUETT,L  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: CORBIN,TRUETT  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/27/2024 03:51:45

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:51 |            |
| Air Blank         |       | 03:51 | RFI Detect |
| Air Blank         | 0.000 | 03:51 |            |

---

Data Download Date/Time: 12/26/2024 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,TRUETT,L  
BREATH TEST OPERATOR AGENCY: Jackson County SO  
ARRESTING OFFICER: CORBIN,TRUETT  
ARRESTING OFFICER AGENCY: Jackson County SO  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402843437  
DRY GAS EXPIRY DATE: 09/08/2026  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/27/2024 04:01:06

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 03:54 |            |
| Air Blank          | 0.000 | 03:55 |            |
| Control Test       | 0.079 | 03:55 |            |
| Air Blank          | 0.000 | 03:56 |            |
| Subject Sample # 1 | 0.000 | 03:57 |            |
| Breath Volume      | 3.164 | 03:57 |            |
| Air Blank          | 0.000 | 03:57 |            |
| Air Blank          | 0.000 | 03:59 |            |
| Subject Sample # 2 | 0.000 | 03:59 |            |
| Breath Volume      | 3.363 | 03:59 |            |
| Air Blank          | 0.000 | 04:00 |            |
| Control Test       | 0.063 | 04:00 | RFI Detect |
| Air Blank          | 0.000 | 04:01 |            |

---

Data Download Date/Time: 12/26/2024 14:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRICKER,JEFFREY,S  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FRICKER,JEFFREY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: ADMIN  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/06/2024 10:47:09

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 10:46 |            |
| Air Blank         | 0.000 | 10:46 |            |
| Control Test      | 0.079 | 10:46 |            |
| Air Blank         |       | 10:47 | RFI Detect |
| Air Blank         |       | 10:47 | RFI Detect |

---

Data Download Date/Time: 11/12/2024 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRICKER,JEFFREY,S  
BREATH TEST OPERATOR AGENCY: Alachua County SO  
ARRESTING OFFICER: FRICKER,JEFFREY  
ARRESTING OFFICER AGENCY: Alachua County SO  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01524080A4  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: ADMIN  
ARREST TIME: 10:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/06/2024 10:55:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:49 |
| Air Blank          | 0.000 | 10:50 |
| Control Test       | 0.080 | 10:50 |
| Air Blank          | 0.000 | 10:50 |
| Subject Sample # 1 | 0.038 | 10:51 |
| Breath Volume      | 3.605 | 10:51 |
| Air Blank          | 0.000 | 10:51 |
| Air Blank          | 0.000 | 10:53 |
| Subject Sample # 2 | 0.038 | 10:54 |
| Breath Volume      | 3.402 | 10:54 |
| Air Blank          | 0.000 | 10:54 |
| Control Test       | 0.079 | 10:55 |
| Air Blank          | 0.000 | 10:55 |
| Diagnostics Check  | OK    | 10:55 |

---

Data Download Date/Time: 11/12/2024 10:50

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOYE,JAMES,M  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WOODS,KENNETH  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 11/02/2024 01:03:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.080 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.094 | 00:58 |
| Breath Volume      | 2.429 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.092 | 01:01 |
| Breath Volume      | 2.468 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.080 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 11/06/2024 08:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: HANCOCK, HENRY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HESLER

DATE/TIME OF TEST: 11/02/2024 03:17:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:13 |                      |
| Air Blank         | 0.000 | 03:13 |                      |
| Control Test      | 0.081 | 03:13 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 1   | 03:15 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.081 | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Diagnostics Check | OK    | 03:17 |                      |

---

Data Download Date/Time: 11/06/2024 08:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPENCER, RAYMOND, L  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: ALBRECHT, SYDNEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LOVELAND

DATE/TIME OF TEST: 11/02/2024 05:24:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:15 |
| Air Blank          | 0.000 | 05:16 |
| Control Test       | 0.080 | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Subject Sample # 1 | 0.111 | 05:20 |
| Breath Volume      | 1.558 | 05:20 |
| Air Blank          | 0.000 | 05:20 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 2 | 0.105 | 05:23 |
| Breath Volume      | 1.390 | 05:23 |
| Air Blank          | 0.000 | 05:23 |
| Control Test       | 0.079 | 05:24 |
| Air Blank          | 0.000 | 05:24 |
| Diagnostics Check  | OK    | 05:24 |

---

Data Download Date/Time: 11/06/2024 08:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOWER,RACHEL,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: BOWER,RACHEL  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: BUI  
ARREST TIME: 18:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 11/02/2024 20:24:10

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:19 |                      |
| Air Blank         | 0.000 | 20:19 |                      |
| Control Test      | 0.080 | 20:19 |                      |
| Air Blank         | 0.000 | 20:20 |                      |
| Subject Sample    | # 1   | 20:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:22 |                      |
| Air Blank         | 0.000 | 20:23 |                      |
| Control Test      | 0.080 | 20:23 |                      |
| Air Blank         | 0.000 | 20:24 |                      |
| Diagnostics Check | OK    | 20:24 |                      |

---

Data Download Date/Time: 11/06/2024 08:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MACPHERSON,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: CORRMAN,KEVIN  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 11/02/2024 21:52:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.081 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 1 | 0.164 | 21:48 |
| Breath Volume      | 2.867 | 21:48 |
| Air Blank          | 0.000 | 21:48 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 2 | 0.161 | 21:51 |
| Breath Volume      | 2.769 | 21:51 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.081 | 21:52 |
| Air Blank          | 0.000 | 21:52 |
| Diagnostics Check  | OK    | 21:52 |

---

Data Download Date/Time: 11/06/2024 08:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUPPLES,ZACHARY,L  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: CUPPLES,ZACHARY  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BORDEN

DATE/TIME OF TEST: 11/03/2024 01:48:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Control Test       | 0.081 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Subject Sample # 1 | 0.157 | 01:43 |
| Breath Volume      | 2.125 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Air Blank          | 0.000 | 01:46 |
| Subject Sample # 2 | 0.156 | 01:47 |
| Breath Volume      | 1.671 | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Diagnostics Check  | OK    | 01:48 |

---

Data Download Date/Time: 11/06/2024 08:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000772

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WELCH, BLAKE, T  
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO  
ARRESTING OFFICER: WARD, STANLEY  
ARRESTING OFFICER AGENCY: Santa Rosa County SO  
LAST AGENCY INSPECTION: 10/05/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402558895  
DRY GAS EXPIRY DATE: 10/03/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PISARICH

DATE/TIME OF TEST: 11/04/2024 04:09:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Control Test       | 0.081 | 04:01 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 1 | 0.140 | 04:03 |
| Breath Volume      | 2.402 | 04:03 |
| Air Blank          | 0.000 | 04:03 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 2 | 0.145 | 04:07 |
| Breath Volume      | 2.410 | 04:07 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.080 | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Diagnostics Check  | OK    | 04:09 |

---

Data Download Date/Time: 11/06/2024 08:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000773

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STONE,KENNETH,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: STONE,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33922080A1  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/04/2024 22:51:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:49 |                      |
| Air Blank         | 0.000 | 22:49 |                      |
| Control Test      | 0.079 | 22:49 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Subject Sample    | # 1   | 22:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:50 |                      |
| Air Blank         | 0.000 | 22:50 |                      |
| Control Test      | 0.079 | 22:51 |                      |
| Air Blank         | 0.000 | 22:51 |                      |
| Diagnostics Check | OK    | 22:51 |                      |

---

Data Download Date/Time: 11/13/2024 15:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALLEN,JOSHUA,B  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: ALLEN,JOSHUA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: PEREZSANCHEZ

DATE/TIME OF TEST: 11/19/2024 21:43:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.080 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.159 | 21:38 |
| Breath Volume      | 1.589 | 21:38 |
| Air Blank          | 0.000 | 21:39 |
| Air Blank          | 0.000 | 21:41 |
| Subject Sample # 2 | 0.156 | 21:42 |
| Breath Volume      | 1.695 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Control Test       | 0.080 | 21:43 |
| Air Blank          | 0.000 | 21:43 |
| Diagnostics Check  | OK    | 21:43 |

---

Data Download Date/Time: 12/30/2024 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BASS,DANIEL,K  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: HENDERSON,VICTORIA  
ARRESTING OFFICER AGENCY: FL Dept of AG  
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SUMMERLIN

DATE/TIME OF TEST: 11/29/2024 02:36:46

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:25 |                |
| Air Blank         | 0.000     | 02:25 |                |
| Control Test      | 0.082     | 02:26 |                |
| Air Blank         | 0.000     | 02:26 |                |
| Subject Sample    | # 1 0.131 | 02:29 | Volume Not Met |
| Breath Volume     | 1.019     | 02:29 |                |
| Air Blank         | 0.000     | 02:30 |                |
| Air Blank         | 0.000     | 02:32 |                |
| Subject Sample    | # 2 0.140 | 02:35 | Volume Not Met |
| Breath Volume     | 0.234     | 02:35 |                |
| Air Blank         | 0.000     | 02:35 |                |
| Control Test      | 0.082     | 02:36 |                |
| Air Blank         | 0.000     | 02:36 |                |
| Diagnostics Check | OK        | 02:36 |                |

---

Data Download Date/Time: 12/30/2024 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RHODEN,JASON,T  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: RHODEN,JASON  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HICKERSON

DATE/TIME OF TEST: 11/02/2024 21:18:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.079 | 21:12 |
| Air Blank          | 0.000 | 21:13 |
| Subject Sample # 1 | 0.229 | 21:13 |
| Breath Volume      | 2.285 | 21:13 |
| Air Blank          | 0.000 | 21:14 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 2 | 0.232 | 21:17 |
| Breath Volume      | 2.144 | 21:17 |
| Air Blank          | 0.000 | 21:17 |
| Control Test       | 0.078 | 21:18 |
| Air Blank          | 0.000 | 21:18 |
| Diagnostics Check  | OK    | 21:18 |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,RYAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FISHER,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 17:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 11/09/2024 18:24:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Control Test       | 0.078 | 18:19 |
| Air Blank          | 0.000 | 18:19 |
| Subject Sample # 1 | 0.241 | 18:20 |
| Breath Volume      | 2.503 | 18:20 |
| Air Blank          | 0.000 | 18:21 |
| Air Blank          | 0.000 | 18:22 |
| Subject Sample # 2 | 0.223 | 18:23 |
| Breath Volume      | 1.597 | 18:23 |
| Air Blank          | 0.000 | 18:23 |
| Control Test       | 0.077 | 18:24 |
| Air Blank          | 0.000 | 18:24 |
| Diagnostics Check  | OK    | 18:24 |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURR,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DURR,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 11/14/2024 21:11:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:00 |                      |
| Air Blank         | 0.000 | 21:00 |                      |
| Control Test      | 0.079 | 21:01 |                      |
| Air Blank         | 0.000 | 21:01 |                      |
| Subject Sample    | # 1   | 21:04 | No Sample Provided   |
| Breath Volume     | 0.000 | 21:04 |                      |
| Air Blank         | 0.000 | 21:05 |                      |
| Air Blank         | 0.000 | 21:07 |                      |
| Subject Sample    | # 2   | 21:09 | Subject Test Refused |
| Breath Volume     | 0.085 | 21:09 |                      |
| Air Blank         | 0.000 | 21:10 |                      |
| Control Test      | 0.079 | 21:10 |                      |
| Air Blank         | 0.000 | 21:11 |                      |
| Diagnostics Check | OK    | 21:11 |                      |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LORTON,CORNELIUS  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: WHEDBEE

DATE/TIME OF TEST: 11/18/2024 22:59:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Control Test       | 0.078 | 22:53 |
| Air Blank          | 0.000 | 22:54 |
| Subject Sample # 1 | 0.147 | 22:55 |
| Breath Volume      | 1.218 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.141 | 22:58 |
| Breath Volume      | 1.183 | 22:58 |
| Air Blank          | 0.000 | 22:58 |
| Control Test       | 0.078 | 22:59 |
| Air Blank          | 0.000 | 22:59 |
| Diagnostics Check  | OK    | 22:59 |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TEDJESKE,JAMES,R  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: TEDJESKE,JAMES  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ROMINE

DATE/TIME OF TEST: 11/22/2024 02:22:11

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 02:15 |                           |
| Air Blank         | 0.000     | 02:16 |                           |
| Control Test      | 0.079     | 02:16 |                           |
| Air Blank         | 0.000     | 02:17 |                           |
| Subject Sample    | # 1 0.257 | 02:17 |                           |
| Breath Volume     | 2.734     | 02:17 |                           |
| Air Blank         | 0.000     | 02:18 |                           |
| Air Blank         | 0.000     | 02:20 |                           |
| Subject Sample    | # 2 0.239 | 02:20 |                           |
| Breath Volume     | 3.085     | 02:20 |                           |
| Air Blank         | 0.000     | 02:21 |                           |
| Control Test      | 0.074     | 02:21 | Control Outside Tolerance |
| Air Blank         | 0.000     | 02:22 |                           |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLE,DANDRA,M  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: MONTEMURNO,M  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 16:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: MCKIRE

DATE/TIME OF TEST: 11/22/2024 17:20:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:17 |                      |
| Air Blank         | 0.000 | 17:17 |                      |
| Control Test      | 0.079 | 17:18 |                      |
| Air Blank         | 0.000 | 17:18 |                      |
| Subject Sample    | # 1   | 17:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:18 |                      |
| Air Blank         | 0.000 | 17:19 |                      |
| Control Test      | 0.080 | 17:19 |                      |
| Air Blank         | 0.000 | 17:20 |                      |
| Diagnostics Check | OK    | 17:20 |                      |

---

Data Download Date/Time: 11/25/2024 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SUMNER,NATHAN,J  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SUMNER,NATHAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 06:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: DEAS

DATE/TIME OF TEST: 11/23/2024 07:14:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:08 |
| Air Blank          | 0.000 | 07:09 |
| Control Test       | 0.081 | 07:09 |
| Air Blank          | 0.000 | 07:09 |
| Subject Sample # 1 | 0.061 | 07:10 |
| Breath Volume      | 1.582 | 07:10 |
| Air Blank          | 0.000 | 07:10 |
| Air Blank          | 0.000 | 07:12 |
| Subject Sample # 2 | 0.059 | 07:13 |
| Breath Volume      | 2.265 | 07:13 |
| Air Blank          | 0.000 | 07:14 |
| Control Test       | 0.078 | 07:14 |
| Air Blank          | 0.000 | 07:14 |
| Diagnostics Check  | OK    | 07:14 |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LORTON,CORNELIUS  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CANADAY

DATE/TIME OF TEST: 11/24/2024 02:32:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:29 |                      |
| Air Blank         | 0.000 | 02:30 |                      |
| Control Test      | 0.079 | 02:30 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Subject Sample    | # 1   | 02:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:31 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Control Test      | 0.079 | 02:32 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Diagnostics Check | OK    | 02:32 |                      |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BINES,DYLAN,B  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: BINES,DYLAN  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TITTERINGTON

DATE/TIME OF TEST: 11/24/2024 03:14:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Control Test       | 0.079 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Subject Sample # 1 | 0.173 | 03:10 |
| Breath Volume      | 2.972 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 2 | 0.173 | 03:13 |
| Breath Volume      | 2.707 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Control Test       | 0.078 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Diagnostics Check  | OK    | 03:14 |

---

Data Download Date/Time: 11/25/2024 15:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,RYAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FISHER,RYAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 13:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 11/27/2024 14:13:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:07 |
| Air Blank          | 0.000 | 14:08 |
| Control Test       | 0.079 | 14:08 |
| Air Blank          | 0.000 | 14:09 |
| Subject Sample # 1 | 0.237 | 14:09 |
| Breath Volume      | 2.886 | 14:09 |
| Air Blank          | 0.000 | 14:10 |
| Air Blank          | 0.000 | 14:11 |
| Subject Sample # 2 | 0.220 | 14:12 |
| Breath Volume      | 2.316 | 14:12 |
| Air Blank          | 0.000 | 14:12 |
| Control Test       | 0.076 | 14:13 |
| Air Blank          | 0.000 | 14:13 |
| Diagnostics Check  | OK    | 14:13 |

---

Data Download Date/Time: 12/25/2024 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q  
BREATH TEST OPERATOR AGENCY: Columbia County SO  
ARRESTING OFFICER: LORTON,CORNELIUS  
ARRESTING OFFICER AGENCY: Columbia County SO  
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04924080A5  
DRY GAS EXPIRY DATE: 03/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 23:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 11/28/2024 23:58:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.078 | 23:52 |
| Air Blank          | 0.000 | 23:53 |
| Subject Sample # 1 | 0.255 | 23:54 |
| Breath Volume      | 3.328 | 23:54 |
| Air Blank          | 0.000 | 23:54 |
| Air Blank          | 0.000 | 23:56 |
| Subject Sample # 2 | 0.249 | 23:57 |
| Breath Volume      | 2.644 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Control Test       | 0.077 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Diagnostics Check  | OK    | 23:58 |

---

Data Download Date/Time: 12/25/2024 09:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000778

INSTRUMENT REGISTERED TO: NAS JACKSONVILLE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON,TAVARES,S  
BREATH TEST OPERATOR AGENCY: NAS Jacksonville PD  
ARRESTING OFFICER: WILSON,TAVARES  
ARRESTING OFFICER AGENCY: NAS Jacksonville PD  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A3  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MCMILON

DATE/TIME OF TEST: 11/29/2024 01:09:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:06 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Control Test      | 0.083 | 01:07 |                      |
| Air Blank         | 0.000 | 01:07 |                      |
| Subject Sample    | # 1   | 01:08 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:08 |                      |
| Air Blank         | 0.000 | 01:08 |                      |
| Control Test      | 0.081 | 01:08 |                      |
| Air Blank         | 0.000 | 01:09 |                      |
| Diagnostics Check | OK    | 01:09 |                      |

---

Data Download Date/Time: 12/10/2024 11:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ETHRIDGE,ALISA,B  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: MINCKS,ROBERT  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A2  
DRY GAS EXPIRY DATE: 05/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/23/2024 06:40:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:34 |
| Air Blank          | 0.000 | 06:35 |
| Control Test       | 0.082 | 06:35 |
| Air Blank          | 0.000 | 06:35 |
| Subject Sample # 1 | 0.029 | 06:36 |
| Breath Volume      | 2.046 | 06:36 |
| Air Blank          | 0.000 | 06:36 |
| Air Blank          | 0.000 | 06:38 |
| Subject Sample # 2 | 0.024 | 06:39 |
| Breath Volume      | 3.429 | 06:39 |
| Air Blank          | 0.000 | 06:39 |
| Control Test       | 0.078 | 06:40 |
| Air Blank          | 0.000 | 06:40 |
| Diagnostics Check  | OK    | 06:40 |

---

Data Download Date/Time: 12/30/2024 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ETHRIDGE,ALISA,B  
BREATH TEST OPERATOR AGENCY: Madison County SO  
ARRESTING OFFICER: MINCKS,ROBERT  
ARRESTING OFFICER AGENCY: Madison PD  
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 08223080A2  
DRY GAS EXPIRY DATE: 05/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 11/28/2024 06:58:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:52 |
| Air Blank          | 0.000 | 06:52 |
| Control Test       | 0.081 | 06:52 |
| Air Blank          | 0.000 | 06:53 |
| Subject Sample # 1 | 0.149 | 06:53 |
| Breath Volume      | 3.496 | 06:53 |
| Air Blank          | 0.000 | 06:54 |
| Air Blank          | 0.000 | 06:56 |
| Subject Sample # 2 | 0.144 | 06:56 |
| Breath Volume      | 3.726 | 06:56 |
| Air Blank          | 0.000 | 06:57 |
| Control Test       | 0.077 | 06:57 |
| Air Blank          | 0.000 | 06:58 |
| Diagnostics Check  | OK    | 06:58 |

---

Data Download Date/Time: 12/30/2024 14:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DURR,KYLE,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: DURR,KYLE  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 11/09/2024 20:09:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:03 |
| Air Blank          | 0.000 | 20:03 |
| Control Test       | 0.081 | 20:03 |
| Air Blank          | 0.000 | 20:04 |
| Subject Sample # 1 | 0.023 | 20:04 |
| Breath Volume      | 2.242 | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Air Blank          | 0.000 | 20:07 |
| Subject Sample # 2 | 0.019 | 20:07 |
| Breath Volume      | 1.609 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Control Test       | 0.081 | 20:08 |
| Air Blank          | 0.000 | 20:09 |
| Diagnostics Check  | OK    | 20:09 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: RECCA,MICHAEL  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 11/10/2024 03:32:24

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:27 |            |
| Air Blank         | 0.000     | 03:28 |            |
| Control Test      | 0.081     | 03:28 |            |
| Air Blank         | 0.000     | 03:29 |            |
| Subject Sample    | # 1 0.028 | 03:31 |            |
| Breath Volume     | 2.242     | 03:31 |            |
| Air Blank         |           | 03:31 | Purge Fail |
| Air Blank         | 0.000     | 03:32 |            |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WILLIAMS,KYLE  
ARRESTING OFFICER AGENCY: Live Oak PD  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ROMULO DURAN

DATE/TIME OF TEST: 11/16/2024 00:17:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Control Test       | 0.081 | 00:12 |
| Air Blank          | 0.000 | 00:12 |
| Subject Sample # 1 | 0.151 | 00:13 |
| Breath Volume      | 2.539 | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Air Blank          | 0.000 | 00:15 |
| Subject Sample # 2 | 0.149 | 00:15 |
| Breath Volume      | 2.613 | 00:15 |
| Air Blank          | 0.000 | 00:16 |
| Control Test       | 0.082 | 00:16 |
| Air Blank          | 0.000 | 00:17 |
| Diagnostics Check  | OK    | 00:17 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WAGNER,DUSTIN  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MCCORMACK

DATE/TIME OF TEST: 11/16/2024 03:33:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.081 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 1 | 0.276 | 03:28 |
| Breath Volume      | 2.097 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.271 | 03:31 |
| Breath Volume      | 2.019 | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.079 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:33 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SINGLETARY, MATHEW, C  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: SINGLETARY, MATHEW  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 13:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: BALDERAS

DATE/TIME OF TEST: 11/16/2024 14:09:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:00 |
| Air Blank          | 0.000 | 14:00 |
| Control Test       | 0.081 | 14:01 |
| Air Blank          | 0.000 | 14:01 |
| Subject Sample # 1 | 0.218 | 14:02 |
| Breath Volume      | 2.484 | 14:02 |
| Air Blank          | 0.000 | 14:02 |
| Air Blank          | 0.000 | 14:04 |
| Subject Sample # 2 | 0.195 | 14:04 |
| Breath Volume      | 1.125 | 14:04 |
| Air Blank          | 0.000 | 14:05 |
| Air Blank          | 0.000 | 14:07 |
| Subject Sample # 3 | 0.218 | 14:07 |
| Breath Volume      | 2.367 | 14:07 |
| Air Blank          | 0.000 | 14:08 |
| Control Test       | 0.080 | 14:09 |
| Air Blank          | 0.000 | 14:09 |
| Diagnostics Check  | OK    | 14:09 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WAGNER,DUSTIN  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 11/17/2024 04:22:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:19 |                      |
| Air Blank         | 0.000 | 04:20 |                      |
| Control Test      | 0.081 | 04:20 |                      |
| Air Blank         | 0.000 | 04:21 |                      |
| Subject Sample    | # 1   | 04:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:21 |                      |
| Air Blank         | 0.000 | 04:21 |                      |
| Control Test      | 0.081 | 04:22 |                      |
| Air Blank         | 0.000 | 04:22 |                      |
| Diagnostics Check | OK    | 04:22 |                      |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: AGAPITO,SAVANNAH  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WILLSON

DATE/TIME OF TEST: 11/20/2024 22:15:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:08 |
| Air Blank          | 0.000 | 22:08 |
| Control Test       | 0.081 | 22:09 |
| Air Blank          | 0.000 | 22:09 |
| Subject Sample # 1 | 0.185 | 22:10 |
| Breath Volume      | 4.015 | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 2 | 0.182 | 22:13 |
| Breath Volume      | 1.121 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Control Test       | 0.080 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Diagnostics Check  | OK    | 22:15 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIS,CLEVELAND,D  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WILLIS,CLEVELAND  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: CERMAK

DATE/TIME OF TEST: 11/22/2024 17:54:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:48 |
| Air Blank          | 0.000 | 17:49 |
| Control Test       | 0.081 | 17:49 |
| Air Blank          | 0.000 | 17:50 |
| Subject Sample # 1 | 0.242 | 17:50 |
| Breath Volume      | 3.351 | 17:50 |
| Air Blank          | 0.000 | 17:51 |
| Air Blank          | 0.000 | 17:52 |
| Subject Sample # 2 | 0.236 | 17:53 |
| Breath Volume      | 3.714 | 17:53 |
| Air Blank          | 0.000 | 17:54 |
| Control Test       | 0.079 | 17:54 |
| Air Blank          | 0.000 | 17:54 |
| Diagnostics Check  | OK    | 17:54 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: HOWARD,ROGER  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 11/22/2024 20:25:57

---

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 20:20 |                 |
| Air Blank          | 0.000 | 20:20 |                 |
| Control Test       | 0.081 | 20:20 |                 |
| Air Blank          | 0.000 | 20:21 |                 |
| Subject Sample # 1 | 0.099 | 20:22 |                 |
| Breath Volume      | 4.242 | 20:22 |                 |
| Air Blank          | 0.000 | 20:22 |                 |
| Air Blank          |       | 20:24 | Improper Sample |
| Air Blank          | 0.000 | 20:25 |                 |
| Control Test       | 0.081 | 20:25 |                 |
| Air Blank          | 0.000 | 20:25 |                 |
| Diagnostics Check  | OK    | 20:25 |                 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: HOWARD,ROGER  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 11/22/2024 20:34:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.081 | 20:29 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 1 | 0.102 | 20:30 |
| Breath Volume      | 4.917 | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 2 | 0.095 | 20:32 |
| Breath Volume      | 4.312 | 20:32 |
| Air Blank          | 0.000 | 20:33 |
| Control Test       | 0.078 | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Diagnostics Check  | OK    | 20:34 |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAGNER,DUSTIN,R  
BREATH TEST OPERATOR AGENCY: Suwannee County SO  
ARRESTING OFFICER: WAGNER,DUSTIN  
ARRESTING OFFICER AGENCY: Suwannee County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 11/27/2024 05:13:49

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:11 |                      |
| Air Blank         | 0.000 | 05:11 |                      |
| Control Test      | 0.081 | 05:11 |                      |
| Air Blank         | 0.000 | 05:12 |                      |
| Subject Sample    | # 1   | 05:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:12 |                      |
| Air Blank         | 0.000 | 05:12 |                      |
| Control Test      | 0.082 | 05:13 |                      |
| Air Blank         | 0.000 | 05:13 |                      |
| Diagnostics Check | OK    | 05:13 |                      |

---

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORGAN,DARRAN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MORGAN,DARRAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 14323080  
DRY GAS EXPIRY DATE: 09/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: SCHELAH

DATE/TIME OF TEST: 11/27/2024 17:26:28

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 17:18 |                      |
| Air Blank         | 0.000 | 17:18 |                      |
| Control Test      | 0.081 | 17:18 |                      |
| Air Blank         | 0.000 | 17:19 |                      |
| Subject Sample    | # 1   | 17:22 | No Sample Provided   |
| Breath Volume     | 0.000 | 17:22 |                      |
| Air Blank         | 0.000 | 17:22 |                      |
| Air Blank         | 0.000 | 17:24 |                      |
| Subject Sample    | # 2   | 17:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 17:25 |                      |
| Air Blank         | 0.000 | 17:25 |                      |
| Control Test      | 0.082 | 17:25 |                      |
| Air Blank         | 0.000 | 17:26 |                      |
| Diagnostics Check | OK    | 17:26 |                      |

Data Download Date/Time: 12/27/2024 16:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLS,RICHARD,J  
BREATH TEST OPERATOR AGENCY: Port St. Joe PD  
ARRESTING OFFICER: BRADSHAW,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Gulf County SO  
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402946336  
DRY GAS EXPIRY DATE: 01/17/2027  
VIOLATION CODE: DUI  
ARREST TIME: 21:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: RODRIGUEZ FLORES

DATE/TIME OF TEST: 11/20/2024 22:47:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.079 | 22:41 |
| Air Blank          | 0.000 | 22:41 |
| Subject Sample # 1 | 0.145 | 22:43 |
| Breath Volume      | 1.691 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Subject Sample # 2 | 0.145 | 22:46 |
| Breath Volume      | 1.757 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.080 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 12/28/2024 00:49

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LENDERMAN,TYLER,K  
BREATH TEST OPERATOR AGENCY: Chipley PD  
ARRESTING OFFICER: LENDERMAN,TYLER  
ARRESTING OFFICER AGENCY: Chipley PD  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 11/13/2024 03:28:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.081 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 1 | 0.155 | 03:23 |
| Breath Volume      | 3.363 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 2 | 0.155 | 03:26 |
| Breath Volume      | 1.492 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.078 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 12/26/2024 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRISP,JESSICA,M  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: FOSTER,LOGAN  
ARRESTING OFFICER AGENCY: Washington County SO  
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402477284  
DRY GAS EXPIRY DATE: 06/27/2025  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: MITCHEM

DATE/TIME OF TEST: 11/15/2024 10:08:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:02 |
| Air Blank          | 0.000 | 10:02 |
| Control Test       | 0.081 | 10:02 |
| Air Blank          | 0.000 | 10:03 |
| Subject Sample # 1 | 0.124 | 10:03 |
| Breath Volume      | 1.414 | 10:03 |
| Air Blank          | 0.000 | 10:04 |
| Air Blank          | 0.000 | 10:06 |
| Subject Sample # 2 | 0.138 | 10:07 |
| Breath Volume      | 1.566 | 10:07 |
| Air Blank          | 0.000 | 10:07 |
| Control Test       | 0.081 | 10:08 |
| Air Blank          | 0.000 | 10:08 |
| Diagnostics Check  | OK    | 10:08 |

---

Data Download Date/Time: 12/26/2024 14:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,DONALD,R  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SELLERS,DONALD  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CORLISS

DATE/TIME OF TEST: 11/09/2024 01:21:06

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 01:19 |                           |
| Air Blank         | 0.000 | 01:20 |                           |
| Control Test      | 0.069 | 01:20 | Control Outside Tolerance |
| Air Blank         | 0.000 | 01:21 |                           |

Data Download Date/Time: 11/19/2024 16:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,DONALD,R  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SELLERS,DONALD  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CORLISS

DATE/TIME OF TEST: 11/09/2024 01:24:39

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 01:23 |                           |
| Air Blank         | 0.000 | 01:23 |                           |
| Control Test      | 0.068 | 01:24 | Control Outside Tolerance |
| Air Blank         | 0.000 | 01:24 |                           |

---

Data Download Date/Time: 11/19/2024 16:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SELLERS,DONALD,R  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SELLERS,DONALD  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: CORLISS

DATE/TIME OF TEST: 11/09/2024 01:30:11

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 01:28 |                           |
| Air Blank         | 0.000 | 01:29 |                           |
| Control Test      | 0.069 | 01:29 | Control Outside Tolerance |
| Air Blank         | 0.000 | 01:30 |                           |

---

Data Download Date/Time: 11/19/2024 16:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: GOFF,BETH  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 11/10/2024 01:19:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.076 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 1 | 0.167 | 01:14 |
| Breath Volume      | 3.308 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 2 | 0.168 | 01:17 |
| Breath Volume      | 3.363 | 01:17 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.077 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 11/19/2024 16:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMILLION,LANCE,C  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: LEDBETTER,CALEB  
ARRESTING OFFICER AGENCY: Pensacola PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: INGALLS

DATE/TIME OF TEST: 11/18/2024 00:39:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.076 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Subject Sample # 1 | 0.254 | 00:33 |
| Breath Volume      | 1.710 | 00:33 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Subject Sample # 2 | 0.252 | 00:37 |
| Breath Volume      | 2.515 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.076 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Diagnostics Check  | OK    | 00:39 |

---

Data Download Date/Time: 11/19/2024 16:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALEXANDER,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD  
ARRESTING OFFICER: SELLERS,DONALD  
ARRESTING OFFICER AGENCY: Gulf Breeze PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402367487  
DRY GAS EXPIRY DATE: 04/28/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:01  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HOLBROOK

DATE/TIME OF TEST: 11/20/2024 03:38:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.076 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.166 | 03:34 |
| Breath Volume      | 2.148 | 03:34 |
| Air Blank          | 0.000 | 03:34 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.162 | 03:37 |
| Breath Volume      | 2.082 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.076 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Diagnostics Check  | OK    | 03:38 |

---

Data Download Date/Time: 12/16/2024 17:37

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 11/01/2024 02:45:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.081 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 1 | 0.082 | 02:39 |
| Breath Volume      | 1.250 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 2 | 0.085 | 02:43 |
| Breath Volume      | 1.476 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Control Test       | 0.081 | 02:44 |
| Air Blank          | 0.000 | 02:44 |
| Diagnostics Check  | OK    | 02:45 |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 11/02/2024 04:23:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:20 |                      |
| Air Blank         | 0.000 | 04:21 |                      |
| Control Test      | 0.081 | 04:21 |                      |
| Air Blank         | 0.000 | 04:22 |                      |
| Subject Sample    | # 1   | 04:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:22 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Control Test      | 0.081 | 04:23 |                      |
| Air Blank         | 0.000 | 04:23 |                      |
| Diagnostics Check | OK    | 04:23 |                      |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUICK,SHANE,C  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: BELDING,JUDSON  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 19:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: PELKY

DATE/TIME OF TEST: 11/06/2024 20:03:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:00 |                      |
| Air Blank         | 0.000 | 20:01 |                      |
| Control Test      | 0.081 | 20:01 |                      |
| Air Blank         | 0.000 | 20:02 |                      |
| Subject Sample    | # 1   | 20:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:02 |                      |
| Air Blank         | 0.000 | 20:02 |                      |
| Control Test      | 0.081 | 20:03 |                      |
| Air Blank         | 0.000 | 20:03 |                      |
| Diagnostics Check | OK    | 20:03 |                      |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/10/2024 04:46:19

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:43 |                      |
| Air Blank         | 0.000 | 04:43 |                      |
| Control Test      | 0.081 | 04:43 |                      |
| Air Blank         | 0.000 | 04:44 |                      |
| Subject Sample    | # 1   | 04:44 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:44 |                      |
| Air Blank         | 0.000 | 04:45 |                      |
| Control Test      | 0.080 | 04:45 |                      |
| Air Blank         | 0.000 | 04:46 |                      |
| Diagnostics Check | OK    | 04:46 |                      |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: SHOEMAKER

DATE/TIME OF TEST: 11/14/2024 21:52:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.080 | 21:46 |
| Air Blank          | 0.000 | 21:47 |
| Subject Sample # 1 | 0.149 | 21:47 |
| Breath Volume      | 1.828 | 21:47 |
| Air Blank          | 0.000 | 21:48 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 2 | 0.149 | 21:50 |
| Breath Volume      | 1.531 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.080 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Diagnostics Check  | OK    | 21:52 |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER, MATTHEW, D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER, MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 03:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 11/17/2024 03:52:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.081 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.146 | 03:48 |
| Breath Volume      | 2.742 | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.142 | 03:51 |
| Breath Volume      | 2.957 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.080 | 03:52 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 11/18/2024 12:30

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOSIER,MATTHEW,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GOSIER,MATTHEW  
ARRESTING OFFICER AGENCY: Largo PD  
LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 11/22/2024 22:47:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:38 |
| Air Blank          | 0.000 | 22:39 |
| Control Test       | 0.081 | 22:39 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 1 | 0.166 | 22:42 |
| Breath Volume      | 1.503 | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Air Blank          | 0.000 | 22:44 |
| Subject Sample # 2 | 0.174 | 22:45 |
| Breath Volume      | 1.617 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.079 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 12/24/2024 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: UPCHURCH,JERRY

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 23:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DE LA CRUZ MONTIEL

DATE/TIME OF TEST: 11/25/2024 00:34:35

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 00:17 |                    |
| Air Blank         | 0.000 | 00:17 |                    |
| Control Test      | 0.081 | 00:17 |                    |
| Air Blank         | 0.000 | 00:18 |                    |
| Subject Sample    | # 1   | 00:21 | No Sample Provided |
| Breath Volume     | 0.000 | 00:21 |                    |
| Air Blank         | 0.000 | 00:21 |                    |
| Air Blank         | 0.000 | 00:23 |                    |
| Subject Sample    | # 2   | 0.174 |                    |
| Breath Volume     | 1.910 | 00:27 |                    |
| Air Blank         | 0.000 | 00:27 |                    |
| Air Blank         | 0.000 | 00:29 |                    |
| Subject Sample    | # 3   | 0.169 | Volume Not Met     |
| Breath Volume     | 1.054 | 00:32 |                    |
| Air Blank         | 0.000 | 00:33 |                    |
| Control Test      | 0.079 | 00:33 |                    |
| Air Blank         | 0.000 | 00:34 |                    |
| Diagnostics Check | OK    | 00:34 |                    |

---

Data Download Date/Time: 12/24/2024 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 19:48

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/28/2024 20:35:05

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 20:21 |                |
| Air Blank         | 0.000     | 20:21 |                |
| Control Test      | 0.081     | 20:22 |                |
| Air Blank         | 0.000     | 20:22 |                |
| Subject Sample    | # 1 0.154 | 20:25 | Volume Not Met |
| Breath Volume     | 0.535     | 20:25 |                |
| Air Blank         | 0.000     | 20:26 |                |
| Air Blank         | 0.000     | 20:28 |                |
| Subject Sample    | # 2 0.189 | 20:29 |                |
| Breath Volume     | 1.664     | 20:29 |                |
| Air Blank         | 0.000     | 20:29 |                |
| Air Blank         | 0.000     | 20:31 |                |
| Subject Sample    | # 3 0.196 | 20:33 |                |
| Breath Volume     | 1.968     | 20:33 |                |
| Air Blank         | 0.000     | 20:34 |                |
| Control Test      | 0.079     | 20:34 |                |
| Air Blank         | 0.000     | 20:35 |                |
| Diagnostics Check | OK        | 20:35 |                |

---

Data Download Date/Time: 12/24/2024 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:46

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: MCCAIN

DATE/TIME OF TEST: 11/29/2024 22:47:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.081 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Subject Sample # 1 | 0.000 | 22:42 |
| Breath Volume      | 1.824 | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Air Blank          | 0.000 | 22:44 |
| Subject Sample # 2 | 0.000 | 22:45 |
| Breath Volume      | 2.445 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Control Test       | 0.081 | 22:46 |
| Air Blank          | 0.000 | 22:47 |
| Diagnostics Check  | OK    | 22:47 |

---

Data Download Date/Time: 12/24/2024 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 20:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 11/30/2024 20:58:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:52 |
| Air Blank          | 0.000 | 20:52 |
| Control Test       | 0.080 | 20:52 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 1 | 0.173 | 20:53 |
| Breath Volume      | 2.085 | 20:53 |
| Air Blank          | 0.000 | 20:54 |
| Air Blank          | 0.000 | 20:56 |
| Subject Sample # 2 | 0.182 | 20:56 |
| Breath Volume      | 1.976 | 20:56 |
| Air Blank          | 0.000 | 20:57 |
| Control Test       | 0.080 | 20:57 |
| Air Blank          | 0.000 | 20:58 |
| Diagnostics Check  | OK    | 20:58 |

---

Data Download Date/Time: 12/24/2024 10:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CERMAK,OFC,D  
BREATH TEST OPERATOR AGENCY: Largo PD  
ARRESTING OFFICER: GRIGGS,JACK  
ARRESTING OFFICER AGENCY: Indian Shores PD  
LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A3  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 11/30/2024 23:35:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:29 |
| Air Blank          | 0.000 | 23:29 |
| Control Test       | 0.080 | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Subject Sample # 1 | 0.020 | 23:31 |
| Breath Volume      | 1.859 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Air Blank          | 0.000 | 23:33 |
| Subject Sample # 2 | 0.019 | 23:33 |
| Breath Volume      | 2.046 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Control Test       | 0.081 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Diagnostics Check  | OK    | 23:35 |

---

Data Download Date/Time: 12/24/2024 10:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITES,DEJAS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WAITES,DEJAS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CARVAJAL TOLOZA

DATE/TIME OF TEST: 11/08/2024 00:19:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.081 | 00:13 |
| Air Blank          | 0.000 | 00:14 |
| Subject Sample # 1 | 0.066 | 00:14 |
| Breath Volume      | 2.085 | 00:14 |
| Air Blank          | 0.000 | 00:15 |
| Air Blank          | 0.000 | 00:17 |
| Subject Sample # 2 | 0.075 | 00:17 |
| Breath Volume      | 2.484 | 00:17 |
| Air Blank          | 0.000 | 00:18 |
| Control Test       | 0.080 | 00:18 |
| Air Blank          | 0.000 | 00:19 |
| Diagnostics Check  | OK    | 00:19 |

---

Data Download Date/Time: 11/21/2024 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILKEY,JOSHUA,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: GILKEY,JOSHUA  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: MENDEZ-CACERES

DATE/TIME OF TEST: 11/08/2024 17:04:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:57 |
| Air Blank          | 0.000 | 16:57 |
| Control Test       | 0.082 | 16:57 |
| Air Blank          | 0.000 | 16:58 |
| Subject Sample # 1 | 0.236 | 16:59 |
| Breath Volume      | 1.359 | 16:59 |
| Air Blank          | 0.000 | 16:59 |
| Air Blank          | 0.000 | 17:01 |
| Subject Sample # 2 | 0.239 | 17:02 |
| Breath Volume      | 1.644 | 17:02 |
| Air Blank          | 0.000 | 17:03 |
| Control Test       | 0.078 | 17:03 |
| Air Blank          | 0.000 | 17:04 |
| Diagnostics Check  | OK    | 17:04 |

---

Data Download Date/Time: 11/21/2024 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: NETO,JOSE  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/10/2024 03:34:18

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 03:33 |            |
| Air Blank         |       | 03:33 | RFI Detect |
| Air Blank         | 0.000 | 03:34 |            |

---

Data Download Date/Time: 11/21/2024 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: NETO,JOSE  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 11/10/2024 03:44:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.081 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.173 | 03:39 |
| Breath Volume      | 2.152 | 03:39 |
| Air Blank          | 0.000 | 03:39 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 2 | 0.159 | 03:42 |
| Breath Volume      | 2.199 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Control Test       | 0.078 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Diagnostics Check  | OK    | 03:44 |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LAPOLLA,BRANDON  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: DONOFRIO

DATE/TIME OF TEST: 11/11/2024 09:49:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:43 |
| Air Blank          | 0.000 | 09:44 |
| Control Test       | 0.082 | 09:44 |
| Air Blank          | 0.000 | 09:44 |
| Subject Sample # 1 | 0.000 | 09:45 |
| Breath Volume      | 2.957 | 09:45 |
| Air Blank          | 0.000 | 09:45 |
| Air Blank          | 0.000 | 09:47 |
| Subject Sample # 2 | 0.000 | 09:48 |
| Breath Volume      | 2.839 | 09:48 |
| Air Blank          | 0.000 | 09:48 |
| Control Test       | 0.082 | 09:49 |
| Air Blank          | 0.000 | 09:49 |
| Diagnostics Check  | OK    | 09:49 |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WAITES,DEJAS,A  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: WAITES,DEJAS  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GALEANO

DATE/TIME OF TEST: 11/12/2024 03:12:38

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 03:02 |            |
| Air Blank         | 0.000     | 03:02 |            |
| Control Test      | 0.081     | 03:03 |            |
| Air Blank         | 0.000     | 03:03 |            |
| Subject Sample    | # 1 0.224 | 03:04 |            |
| Breath Volume     | 1.199     | 03:04 |            |
| Air Blank         | 0.000     | 03:04 |            |
| Air Blank         | 0.000     | 03:06 |            |
| Subject Sample    | # 2 0.255 | 03:07 |            |
| Breath Volume     | 2.406     | 03:07 |            |
| Air Blank         | 0.000     | 03:08 |            |
| Air Blank         | 0.000     | 03:10 |            |
| Subject Sample    | # 3 0.250 | 03:11 |            |
| Breath Volume     | 2.015     | 03:11 |            |
| Air Blank         | 0.000     | 03:11 |            |
| Control Test      | 0.079     | 03:12 |            |
| Air Blank         |           | 03:12 | RFI Detect |
| Air Blank         | 0.000     | 03:12 |            |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MYERS,KIRK  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: LODAI ROCK

DATE/TIME OF TEST: 11/15/2024 01:14:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:07 |
| Air Blank          | 0.000 | 01:07 |
| Control Test       | 0.081 | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Subject Sample # 1 | 0.268 | 01:09 |
| Breath Volume      | 2.664 | 01:09 |
| Air Blank          | 0.000 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Subject Sample # 2 | 0.256 | 01:12 |
| Breath Volume      | 3.386 | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.077 | 01:13 |
| Air Blank          | 0.000 | 01:13 |
| Diagnostics Check  | OK    | 01:14 |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DIEGO,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: FHP Troop L  
ARRESTING OFFICER: DIEGO,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop L  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/15/2024 20:35:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Control Test       | 0.082 | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Subject Sample # 1 | 0.000 | 20:31 |
| Breath Volume      | 3.570 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Air Blank          | 0.000 | 20:34 |
| Subject Sample # 2 | 0.000 | 20:34 |
| Breath Volume      | 3.531 | 20:34 |
| Air Blank          | 0.000 | 20:35 |
| Control Test       | 0.080 | 20:35 |
| Air Blank          | 0.000 | 20:35 |
| Diagnostics Check  | OK    | 20:35 |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMB,HUNTER,  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: LANIER,BRANDEN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 15:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: MADRID ALMAZAN

DATE/TIME OF TEST: 11/20/2024 15:50:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:42 |
| Air Blank          | 0.000 | 15:43 |
| Control Test       | 0.081 | 15:43 |
| Air Blank          | 0.000 | 15:44 |
| Subject Sample # 1 | 0.211 | 15:45 |
| Breath Volume      | 2.289 | 15:45 |
| Air Blank          | 0.000 | 15:46 |
| Air Blank          | 0.000 | 15:48 |
| Subject Sample # 2 | 0.216 | 15:48 |
| Breath Volume      | 1.589 | 15:48 |
| Air Blank          | 0.000 | 15:49 |
| Control Test       | 0.080 | 15:49 |
| Air Blank          | 0.000 | 15:50 |
| Diagnostics Check  | OK    | 15:50 |

---

Data Download Date/Time: 11/21/2024 15:12

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: MORENO,JAIRO  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MORALES VALLE

DATE/TIME OF TEST: 11/24/2024 00:47:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:40 |
| Air Blank          | 0.000 | 00:40 |
| Control Test       | 0.082 | 00:41 |
| Air Blank          | 0.000 | 00:41 |
| Subject Sample # 1 | 0.138 | 00:42 |
| Breath Volume      | 2.113 | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 2 | 0.144 | 00:45 |
| Breath Volume      | 2.753 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.079 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Diagnostics Check  | OK    | 00:47 |

---

Data Download Date/Time: 12/30/2024 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: KICKLIGHTER,KEVIN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: JONES JR

DATE/TIME OF TEST: 11/24/2024 01:34:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:27 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.080 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.194 | 01:29 |
| Breath Volume      | 1.167 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 2 | 0.196 | 01:32 |
| Breath Volume      | 1.285 | 01:32 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.076 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 12/30/2024 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R  
BREATH TEST OPERATOR AGENCY: St. Lucie County SO  
ARRESTING OFFICER: SALERNO,JAIDEN  
ARRESTING OFFICER AGENCY: St. Lucie County SO  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 11/24/2024 03:09:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:02 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.081 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 1 | 0.105 | 03:05 |
| Breath Volume      | 1.382 | 03:05 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:07 |
| Subject Sample # 2 | 0.104 | 03:08 |
| Breath Volume      | 1.527 | 03:08 |
| Air Blank          | 0.000 | 03:09 |
| Control Test       | 0.077 | 03:09 |
| Air Blank          | 0.000 | 03:09 |
| Diagnostics Check  | OK    | 03:09 |

---

Data Download Date/Time: 12/30/2024 16:21

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEMON,MATTHEW,M  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: MEMON,MATTHEW  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: ALVARADO-SORTO

DATE/TIME OF TEST: 11/24/2024 22:31:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Control Test       | 0.082 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Subject Sample # 1 | 0.207 | 22:27 |
| Breath Volume      | 3.582 | 22:27 |
| Air Blank          | 0.000 | 22:27 |
| Air Blank          | 0.000 | 22:29 |
| Subject Sample # 2 | 0.207 | 22:29 |
| Breath Volume      | 3.519 | 22:29 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.080 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Diagnostics Check  | OK    | 22:31 |

---

Data Download Date/Time: 12/30/2024 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIGG,JACOB,M  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: PIGG,JACOB  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 11/26/2024 04:22:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:15 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.082 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 1 | 0.109 | 04:17 |
| Breath Volume      | 2.253 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Air Blank          | 0.000 | 04:20 |
| Subject Sample # 2 | 0.100 | 04:20 |
| Breath Volume      | 2.371 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.081 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Diagnostics Check  | OK    | 04:22 |

---

Data Download Date/Time: 12/30/2024 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROCATO,SHAWN,V  
BREATH TEST OPERATOR AGENCY: Fort Pierce PD  
ARRESTING OFFICER: BROCATO,SHAWN  
ARRESTING OFFICER AGENCY: Fort Pierce PD  
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A4  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 11/26/2024 21:35:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:28 |
| Air Blank          | 0.000 | 21:28 |
| Control Test       | 0.082 | 21:28 |
| Air Blank          | 0.000 | 21:29 |
| Subject Sample # 1 | 0.197 | 21:30 |
| Breath Volume      | 1.390 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Air Blank          | 0.000 | 21:33 |
| Subject Sample # 2 | 0.185 | 21:33 |
| Breath Volume      | 1.750 | 21:33 |
| Air Blank          | 0.000 | 21:34 |
| Control Test       | 0.077 | 21:34 |
| Air Blank          | 0.000 | 21:35 |
| Diagnostics Check  | OK    | 21:35 |

---

Data Download Date/Time: 12/30/2024 16:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 10/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DA SILVA

DATE/TIME OF TEST: 11/02/2024 02:22:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:19 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Control Test      | 0.082 | 02:20 |                      |
| Air Blank         | 0.000 | 02:20 |                      |
| Subject Sample    | # 1   | 02:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:21 |                      |
| Air Blank         | 0.000 | 02:21 |                      |
| Control Test      | 0.081 | 02:21 |                      |
| Air Blank         | 0.000 | 02:22 |                      |
| Diagnostics Check | OK    | 02:22 |                      |

---

Data Download Date/Time: 11/03/2024 17:16

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GILMAN,DUSTIN,D  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: NESTER,LOGAN  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 10/04/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: ACEVEDO VELAZQUEZ

DATE/TIME OF TEST: 11/02/2024 05:03:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.082 | 04:58 |
| Air Blank          | 0.000 | 04:58 |
| Subject Sample # 1 | 0.158 | 04:59 |
| Breath Volume      | 2.148 | 04:59 |
| Air Blank          | 0.000 | 04:59 |
| Air Blank          | 0.000 | 05:01 |
| Subject Sample # 2 | 0.148 | 05:02 |
| Breath Volume      | 1.687 | 05:02 |
| Air Blank          | 0.000 | 05:02 |
| Control Test       | 0.081 | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Diagnostics Check  | OK    | 05:03 |

---

Data Download Date/Time: 11/03/2024 17:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,LATORIAN,T  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TURNER,LATORIAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: VELASQUEZ GUERRERO

DATE/TIME OF TEST: 11/10/2024 06:10:25

|                   |       |             |                    |
|-------------------|-------|-------------|--------------------|
| Diagnostics Check | OK    | 05:55       |                    |
| Air Blank         | 0.000 | 05:56       |                    |
| Control Test      | 0.082 | 05:56       |                    |
| Air Blank         | 0.000 | 05:57       |                    |
| Subject Sample    | # 1   | 06:00       | No Sample Provided |
| Breath Volume     | 0.000 | 06:00       |                    |
| Air Blank         | 0.000 | 06:00       |                    |
| Air Blank         | 0.000 | 06:02       |                    |
| Subject Sample    | # 2   | 0.121 06:03 |                    |
| Breath Volume     | 1.367 | 06:03       |                    |
| Air Blank         | 0.000 | 06:04       |                    |
| Air Blank         | 0.000 | 06:05       |                    |
| Subject Sample    | # 3   | 06:09       | No Sample Provided |
| Breath Volume     | 0.000 | 06:09       |                    |
| Air Blank         | 0.000 | 06:09       |                    |
| Control Test      | 0.081 | 06:09       |                    |
| Air Blank         | 0.000 | 06:10       |                    |
| Diagnostics Check | OK    | 06:10       |                    |

Data Download Date/Time: 12/09/2024 17:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,LATORIAN,T  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TURNER,LATORIAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: VELASQUEZ GUERRERO

DATE/TIME OF TEST: 11/10/2024 06:37:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:32 |                      |
| Air Blank         | 0.000 | 06:33 |                      |
| Control Test      | 0.081 | 06:33 |                      |
| Air Blank         | 0.000 | 06:33 |                      |
| Subject Sample    | # 1   | 06:35 | Subject Test Refused |
| Breath Volume     | 0.554 | 06:35 |                      |
| Air Blank         | 0.000 | 06:36 |                      |
| Control Test      | 0.081 | 06:36 |                      |
| Air Blank         | 0.000 | 06:36 |                      |
| Diagnostics Check | OK    | 06:37 |                      |

---

Data Download Date/Time: 12/09/2024 17:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: ALCALDE,ALEXANDER  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 08:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: LEITAO

DATE/TIME OF TEST: 11/14/2024 09:50:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 09:45 |                      |
| Air Blank         | 0.000 | 09:46 |                      |
| Control Test      | 0.083 | 09:46 |                      |
| Air Blank         | 0.000 | 09:46 |                      |
| Subject Sample    | # 1   | 09:49 | Subject Test Refused |
| Breath Volume     | 0.000 | 09:49 |                      |
| Air Blank         | 0.000 | 09:49 |                      |
| Control Test      | 0.081 | 09:50 |                      |
| Air Blank         | 0.000 | 09:50 |                      |
| Diagnostics Check | OK    | 09:50 |                      |

---

Data Download Date/Time: 12/09/2024 17:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FRIAS ARCE

DATE/TIME OF TEST: 11/16/2024 03:27:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.083 | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Subject Sample # 1 | 0.199 | 03:22 |
| Breath Volume      | 1.179 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.198 | 03:26 |
| Breath Volume      | 1.460 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.082 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 12/09/2024 17:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: DOUGLAS,JORDAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/17/2024 20:32:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:26 |
| Air Blank          | 0.000 | 20:26 |
| Control Test       | 0.082 | 20:26 |
| Air Blank          | 0.000 | 20:27 |
| Subject Sample # 1 | 0.175 | 20:28 |
| Breath Volume      | 1.621 | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Air Blank          | 0.000 | 20:30 |
| Subject Sample # 2 | 0.177 | 20:31 |
| Breath Volume      | 1.726 | 20:31 |
| Air Blank          | 0.000 | 20:31 |
| Control Test       | 0.082 | 20:31 |
| Air Blank          | 0.000 | 20:32 |
| Diagnostics Check  | OK    | 20:32 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: JULES,EMMANUELLO  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: STOVER

DATE/TIME OF TEST: 11/17/2024 22:03:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:57 |
| Air Blank          | 0.000 | 21:57 |
| Control Test       | 0.082 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Subject Sample # 1 | 0.179 | 21:59 |
| Breath Volume      | 1.613 | 21:59 |
| Air Blank          | 0.000 | 21:59 |
| Air Blank          | 0.000 | 22:01 |
| Subject Sample # 2 | 0.167 | 22:01 |
| Breath Volume      | 1.593 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Control Test       | 0.081 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Diagnostics Check  | OK    | 22:03 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: HERRERA-SANTAMARIA

DATE/TIME OF TEST: 11/20/2024 22:00:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:54 |
| Air Blank          | 0.000 | 21:55 |
| Control Test       | 0.081 | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 1 | 0.083 | 21:56 |
| Breath Volume      | 1.863 | 21:56 |
| Air Blank          | 0.000 | 21:57 |
| Air Blank          | 0.000 | 21:59 |
| Subject Sample # 2 | 0.079 | 21:59 |
| Breath Volume      | 1.582 | 21:59 |
| Air Blank          | 0.000 | 22:00 |
| Control Test       | 0.082 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Diagnostics Check  | OK    | 22:00 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: COURT ORD  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 11/21/2024 09:42:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:35 |
| Air Blank          | 0.000 | 09:36 |
| Control Test       | 0.083 | 09:36 |
| Air Blank          | 0.000 | 09:36 |
| Subject Sample # 1 | 0.000 | 09:38 |
| Breath Volume      | 2.835 | 09:38 |
| Air Blank          | 0.000 | 09:38 |
| Air Blank          | 0.000 | 09:40 |
| Subject Sample # 2 | 0.000 | 09:41 |
| Breath Volume      | 3.199 | 09:41 |
| Air Blank          | 0.000 | 09:41 |
| Control Test       | 0.083 | 09:41 |
| Air Blank          | 0.000 | 09:42 |
| Diagnostics Check  | OK    | 09:42 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: N/A,N/A  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: COURT ORD  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: MANZANO MELENDEZ

DATE/TIME OF TEST: 11/21/2024 10:11:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:05 |
| Air Blank          | 0.000 | 10:05 |
| Control Test       | 0.082 | 10:06 |
| Air Blank          | 0.000 | 10:06 |
| Subject Sample # 1 | 0.000 | 10:07 |
| Breath Volume      | 3.449 | 10:07 |
| Air Blank          | 0.000 | 10:07 |
| Air Blank          | 0.000 | 10:09 |
| Subject Sample # 2 | 0.000 | 10:10 |
| Breath Volume      | 2.933 | 10:10 |
| Air Blank          | 0.000 | 10:10 |
| Control Test       | 0.083 | 10:11 |
| Air Blank          | 0.000 | 10:11 |
| Diagnostics Check  | OK    | 10:11 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,LATORIAN,T  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TURNER,LATORIAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 11/23/2024 03:28:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:21 |
| Air Blank          | 0.000 | 03:22 |
| Control Test       | 0.082 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Subject Sample # 1 | 0.128 | 03:23 |
| Breath Volume      | 1.707 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 2 | 0.140 | 03:26 |
| Breath Volume      | 1.425 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.082 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Diagnostics Check  | OK    | 03:28 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: HICKS,DANIEL  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 11/25/2024 12:07:15

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 12:05 |            |
| Air Blank         | 0.000 | 12:06 |            |
| Control Test      | 0.082 | 12:06 |            |
| Air Blank         |       | 12:06 | RFI Detect |
| Air Blank         | 0.000 | 12:07 |            |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HICKS,DANIEL,N  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: HICKS,DANIEL  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 11:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 11/25/2024 12:16:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:10 |
| Air Blank          | 0.000 | 12:10 |
| Control Test       | 0.082 | 12:11 |
| Air Blank          | 0.000 | 12:11 |
| Subject Sample # 1 | 0.045 | 12:12 |
| Breath Volume      | 2.222 | 12:12 |
| Air Blank          | 0.000 | 12:13 |
| Air Blank          | 0.000 | 12:14 |
| Subject Sample # 2 | 0.044 | 12:15 |
| Breath Volume      | 2.265 | 12:15 |
| Air Blank          | 0.000 | 12:15 |
| Control Test       | 0.082 | 12:16 |
| Air Blank          | 0.000 | 12:16 |
| Diagnostics Check  | OK    | 12:16 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EVANS,STEVEN,J  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: EVANS,STEVEN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 12:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 11/26/2024 13:59:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:53 |
| Air Blank          | 0.000 | 13:53 |
| Control Test       | 0.082 | 13:54 |
| Air Blank          | 0.000 | 13:54 |
| Subject Sample # 1 | 0.000 | 13:54 |
| Breath Volume      | 1.277 | 13:54 |
| Air Blank          | 0.000 | 13:55 |
| Air Blank          | 0.000 | 13:57 |
| Subject Sample # 2 | 0.000 | 13:57 |
| Breath Volume      | 1.226 | 13:57 |
| Air Blank          | 0.000 | 13:58 |
| Control Test       | 0.081 | 13:58 |
| Air Blank          | 0.000 | 13:59 |
| Diagnostics Check  | OK    | 13:59 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TURNER,LATORIAN,T  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: TURNER,LATORIAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ANTEMATE BARRIOS

DATE/TIME OF TEST: 11/29/2024 03:31:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.082 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.073 | 03:26 |
| Breath Volume      | 1.343 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.074 | 03:29 |
| Breath Volume      | 1.484 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.080 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRUNER,SEAN,R  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: BRUNER,SEAN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402623828  
DRY GAS EXPIRY DATE: 12/22/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 11/30/2024 02:28:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.083 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Subject Sample # 1 | 0.164 | 02:24 |
| Breath Volume      | 1.117 | 02:24 |
| Air Blank          | 0.000 | 02:24 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 2 | 0.155 | 02:26 |
| Breath Volume      | 1.648 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Control Test       | 0.082 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Diagnostics Check  | OK    | 02:28 |

---

Data Download Date/Time: 12/09/2024 17:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURRY, WILLIAM, C  
BREATH TEST OPERATOR AGENCY: FL Dept of AG  
ARRESTING OFFICER: WATSON, JOSEPH  
ARRESTING OFFICER AGENCY: Hamilton County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 10124080A2  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LAGASSE

DATE/TIME OF TEST: 11/14/2024 21:51:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Control Test       | 0.082 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 1 | 0.205 | 21:47 |
| Breath Volume      | 2.503 | 21:47 |
| Air Blank          | 0.000 | 21:47 |
| Air Blank          | 0.000 | 21:49 |
| Subject Sample # 2 | 0.201 | 21:49 |
| Breath Volume      | 2.214 | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Control Test       | 0.081 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Diagnostics Check  | OK    | 21:51 |

---

Data Download Date/Time: 11/26/2024 00:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SMITH, MATTHEW, M

BREATH TEST OPERATOR AGENCY: Perry PD

ARRESTING OFFICER: NA, NA

ARRESTING OFFICER AGENCY: Perry PD

LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: ADMIN

ARREST TIME: 17:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: WINDHAM

DATE/TIME OF TEST: 11/12/2024 17:33:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:27 |
| Air Blank          | 0.000 | 17:27 |
| Control Test       | 0.079 | 17:28 |
| Air Blank          | 0.000 | 17:28 |
| Subject Sample # 1 | 0.000 | 17:29 |
| Breath Volume      | 3.003 | 17:29 |
| Air Blank          | 0.000 | 17:29 |
| Air Blank          | 0.000 | 17:31 |
| Subject Sample # 2 | 0.000 | 17:31 |
| Breath Volume      | 2.703 | 17:31 |
| Air Blank          | 0.000 | 17:32 |
| Control Test       | 0.079 | 17:32 |
| Air Blank          | 0.000 | 17:33 |
| Diagnostics Check  | OK    | 17:33 |

---

Data Download Date/Time: 12/23/2024 14:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOWLTON,ERIC,S  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: KNOWLTON,ERIC  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GODSIL

DATE/TIME OF TEST: 11/01/2024 03:05:32

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 02:54 |                    |
| Air Blank         | 0.000 | 02:54 |                    |
| Control Test      | 0.080 | 02:54 |                    |
| Air Blank         | 0.000 | 02:55 |                    |
| Subject Sample    | # 1   | 02:58 | No Sample Provided |
| Breath Volume     | 0.000 | 02:58 |                    |
| Air Blank         | 0.000 | 02:59 |                    |
| Air Blank         | 0.000 | 03:01 |                    |
| Subject Sample    | # 2   | 03:04 | Volume Not Met     |
| Breath Volume     | 0.218 | 03:04 |                    |
| Air Blank         | 0.000 | 03:04 |                    |
| Control Test      | 0.080 | 03:04 |                    |
| Air Blank         | 0.000 | 03:05 |                    |
| Diagnostics Check | OK    | 03:05 |                    |

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KNOWLTON,ERIC,S  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: KNOWLTON,ERIC  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GODSIL

DATE/TIME OF TEST: 11/01/2024 03:18:32

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 03:07 |                      |
| Air Blank         | 0.000     | 03:08 |                      |
| Control Test      | 0.079     | 03:08 |                      |
| Air Blank         | 0.000     | 03:09 |                      |
| Subject Sample    | # 1 0.158 | 03:12 | Volume Not Met       |
| Breath Volume     | 0.261     | 03:12 |                      |
| Air Blank         | 0.000     | 03:13 |                      |
| Air Blank         | 0.000     | 03:14 |                      |
| Subject Sample    | # 2       | 03:16 | Subject Test Refused |
| Breath Volume     | 0.652     | 03:16 |                      |
| Air Blank         | 0.000     | 03:17 |                      |
| Control Test      | 0.077     | 03:17 |                      |
| Air Blank         | 0.000     | 03:18 |                      |
| Diagnostics Check | OK        | 03:18 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONEY,JUSTIN,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: LONEY,JUSTIN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 12:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: FANNIN

DATE/TIME OF TEST: 11/02/2024 14:50:49

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 14:39 |                |
| Air Blank         | 0.000     | 14:39 |                |
| Control Test      | 0.080     | 14:39 |                |
| Air Blank         | 0.000     | 14:40 |                |
| Subject Sample    | # 1 0.192 | 14:43 | Volume Not Met |
| Breath Volume     | 0.894     | 14:43 |                |
| Air Blank         | 0.000     | 14:44 |                |
| Air Blank         | 0.000     | 14:46 |                |
| Subject Sample    | # 2 0.165 | 14:49 | Volume Not Met |
| Breath Volume     | 0.578     | 14:49 |                |
| Air Blank         | 0.000     | 14:49 |                |
| Control Test      | 0.079     | 14:50 |                |
| Air Blank         | 0.000     | 14:50 |                |
| Diagnostics Check | OK        | 14:50 |                |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 11/03/2024 00:33:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:25 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.080 | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Subject Sample # 1 | 0.163 | 00:27 |
| Breath Volume      | 1.199 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.170 | 00:31 |
| Breath Volume      | 1.632 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.079 | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Diagnostics Check  | OK    | 00:33 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 11/03/2024 04:49:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:40 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.079 | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.240 | 04:44 |
| Breath Volume      | 1.187 | 04:44 |
| Air Blank          | 0.000 | 04:45 |
| Air Blank          | 0.000 | 04:47 |
| Subject Sample # 2 | 0.231 | 04:48 |
| Breath Volume      | 1.917 | 04:48 |
| Air Blank          | 0.000 | 04:49 |
| Control Test       | 0.078 | 04:49 |
| Air Blank          | 0.000 | 04:49 |
| Diagnostics Check  | OK    | 04:49 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COYNE

DATE/TIME OF TEST: 11/04/2024 03:56:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:50 |
| Air Blank          | 0.000 | 03:50 |
| Control Test       | 0.079 | 03:50 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 1 | 0.201 | 03:51 |
| Breath Volume      | 1.855 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 2 | 0.188 | 03:55 |
| Breath Volume      | 2.207 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Control Test       | 0.079 | 03:56 |
| Air Blank          | 0.000 | 03:56 |
| Diagnostics Check  | OK    | 03:56 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEFFERNAN,EDWARD,J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FITCH,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: ASHER

DATE/TIME OF TEST: 11/05/2024 23:18:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:12 |
| Air Blank          | 0.000 | 23:12 |
| Control Test       | 0.080 | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Subject Sample # 1 | 0.000 | 23:13 |
| Breath Volume      | 3.867 | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Air Blank          | 0.000 | 23:16 |
| Subject Sample # 2 | 0.000 | 23:16 |
| Breath Volume      | 3.007 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Control Test       | 0.081 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Diagnostics Check  | OK    | 23:18 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEIGH,DAVID,E  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: MEYERS,ERIC  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 11/07/2024 00:36:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:33 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Control Test      | 0.080 | 00:34 |                      |
| Air Blank         | 0.000 | 00:35 |                      |
| Subject Sample    | # 1   | 00:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:35 |                      |
| Air Blank         | 0.000 | 00:35 |                      |
| Control Test      | 0.080 | 00:36 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Diagnostics Check | OK    | 00:36 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARRINGTON,A,  
BREATH TEST OPERATOR AGENCY: Belleview PD  
ARRESTING OFFICER: HARRINGTON,A  
ARRESTING OFFICER AGENCY: Belleview PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 22:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DIAZ DOMINGUEZ

DATE/TIME OF TEST: 11/09/2024 22:59:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:52 |
| Air Blank          | 0.000 | 22:53 |
| Control Test       | 0.079 | 22:53 |
| Air Blank          | 0.000 | 22:53 |
| Subject Sample # 1 | 0.224 | 22:54 |
| Breath Volume      | 2.324 | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Air Blank          | 0.000 | 22:57 |
| Subject Sample # 2 | 0.219 | 22:57 |
| Breath Volume      | 2.464 | 22:57 |
| Air Blank          | 0.000 | 22:58 |
| Control Test       | 0.079 | 22:58 |
| Air Blank          | 0.000 | 22:59 |
| Diagnostics Check  | OK    | 22:59 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THEIS,ZACHARY,R  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: THEIS,ZACHARY  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 11/10/2024 02:15:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:12 |                      |
| Air Blank         | 0.000 | 02:13 |                      |
| Control Test      | 0.080 | 02:13 |                      |
| Air Blank         | 0.000 | 02:13 |                      |
| Subject Sample    | # 1   | 02:14 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:14 |                      |
| Air Blank         | 0.000 | 02:14 |                      |
| Control Test      | 0.080 | 02:14 |                      |
| Air Blank         | 0.000 | 02:15 |                      |
| Diagnostics Check | OK    | 02:15 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSLEY, ABIGAIL, G  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: HINSLEY, ABIGAIL  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GENTLES

DATE/TIME OF TEST: 11/10/2024 05:30:28

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:18 |                |
| Air Blank         | 0.000     | 05:18 |                |
| Control Test      | 0.081     | 05:19 |                |
| Air Blank         | 0.000     | 05:19 |                |
| Subject Sample    | # 1 0.188 | 05:22 | Volume Not Met |
| Breath Volume     | 0.039     | 05:22 |                |
| Air Blank         | 0.000     | 05:23 |                |
| Air Blank         | 0.000     | 05:25 |                |
| Subject Sample    | # 2 0.139 | 05:28 | Volume Not Met |
| Breath Volume     | 0.652     | 05:28 |                |
| Air Blank         | 0.000     | 05:29 |                |
| Control Test      | 0.078     | 05:29 |                |
| Air Blank         | 0.000     | 05:30 |                |
| Diagnostics Check | OK        | 05:30 |                |

---

Data Download Date/Time: 11/29/2024 15:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HINSLEY,ABIGAIL,G  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: HINSLEY,ABIGAIL  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GENTLES

DATE/TIME OF TEST: 11/10/2024 05:42:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:35 |
| Air Blank          | 0.000 | 05:35 |
| Control Test       | 0.076 | 05:36 |
| Air Blank          | 0.000 | 05:36 |
| Subject Sample # 1 | 0.187 | 05:37 |
| Breath Volume      | 1.656 | 05:37 |
| Air Blank          | 0.000 | 05:38 |
| Air Blank          | 0.000 | 05:40 |
| Subject Sample # 2 | 0.178 | 05:40 |
| Breath Volume      | 2.031 | 05:40 |
| Air Blank          | 0.000 | 05:41 |
| Control Test       | 0.077 | 05:41 |
| Air Blank          | 0.000 | 05:42 |
| Diagnostics Check  | OK    | 05:42 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: FUTRAL,WILLIAM  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 20:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BYRNES

DATE/TIME OF TEST: 11/11/2024 21:43:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:36 |
| Air Blank          | 0.000 | 21:36 |
| Control Test       | 0.080 | 21:37 |
| Air Blank          | 0.000 | 21:37 |
| Subject Sample # 1 | 0.000 | 21:39 |
| Breath Volume      | 1.644 | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Air Blank          | 0.000 | 21:41 |
| Subject Sample # 2 | 0.000 | 21:42 |
| Breath Volume      | 1.722 | 21:42 |
| Air Blank          | 0.000 | 21:42 |
| Control Test       | 0.080 | 21:42 |
| Air Blank          | 0.000 | 21:43 |
| Diagnostics Check  | OK    | 21:43 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOUCHER,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BOUCHER,DANIEL  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 16:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: KEE

DATE/TIME OF TEST: 11/12/2024 17:40:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:33 |
| Air Blank          | 0.000 | 17:33 |
| Control Test       | 0.080 | 17:34 |
| Air Blank          | 0.000 | 17:34 |
| Subject Sample # 1 | 0.201 | 17:35 |
| Breath Volume      | 1.648 | 17:35 |
| Air Blank          | 0.000 | 17:36 |
| Air Blank          | 0.000 | 17:37 |
| Subject Sample # 2 | 0.211 | 17:38 |
| Breath Volume      | 2.160 | 17:38 |
| Air Blank          | 0.000 | 17:39 |
| Control Test       | 0.079 | 17:39 |
| Air Blank          | 0.000 | 17:40 |
| Diagnostics Check  | OK    | 17:40 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREITAS,MARCIA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FREITAS,MARCIA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 18:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: CHINCHILLA LOPEZ

DATE/TIME OF TEST: 11/14/2024 19:34:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:27 |
| Air Blank          | 0.000 | 19:27 |
| Control Test       | 0.080 | 19:28 |
| Air Blank          | 0.000 | 19:28 |
| Subject Sample # 1 | 0.165 | 19:29 |
| Breath Volume      | 2.316 | 19:29 |
| Air Blank          | 0.000 | 19:30 |
| Air Blank          | 0.000 | 19:31 |
| Subject Sample # 2 | 0.158 | 19:32 |
| Breath Volume      | 2.761 | 19:32 |
| Air Blank          | 0.000 | 19:33 |
| Control Test       | 0.080 | 19:33 |
| Air Blank          | 0.000 | 19:34 |
| Diagnostics Check  | OK    | 19:34 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAILEY, RONALD, J  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: BAILEY, RONALD  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 20:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 11/14/2024 21:45:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:38 |
| Air Blank          | 0.000 | 21:38 |
| Control Test       | 0.080 | 21:39 |
| Air Blank          | 0.000 | 21:39 |
| Subject Sample # 1 | 0.103 | 21:40 |
| Breath Volume      | 3.062 | 21:40 |
| Air Blank          | 0.000 | 21:41 |
| Air Blank          | 0.000 | 21:43 |
| Subject Sample # 2 | 0.100 | 21:44 |
| Breath Volume      | 2.140 | 21:44 |
| Air Blank          | 0.000 | 21:44 |
| Control Test       | 0.079 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Diagnostics Check  | OK    | 21:45 |

---

Data Download Date/Time: 11/29/2024 15:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURLEIGH,DAVID,E  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: LICALZI,KYLIEE  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 11/16/2024 02:48:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:46 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.081 | 02:46 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Subject Sample    | # 1   | 02:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:47 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Control Test      | 0.080 | 02:48 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Diagnostics Check | OK    | 02:48 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA,DAVID,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SANTANA,DAVID  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ESPINAL

DATE/TIME OF TEST: 11/17/2024 00:23:39

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:20 |                      |
| Air Blank         | 0.000 | 00:21 |                      |
| Control Test      | 0.080 | 00:21 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Subject Sample    | # 1   | 00:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:22 |                      |
| Air Blank         | 0.000 | 00:22 |                      |
| Control Test      | 0.080 | 00:23 |                      |
| Air Blank         | 0.000 | 00:23 |                      |
| Diagnostics Check | OK    | 00:23 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRINGER,MICHAEL,S  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: KENNY,WESLEY  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MARLER

DATE/TIME OF TEST: 11/17/2024 01:39:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Control Test       | 0.080 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 1 | 0.215 | 01:34 |
| Breath Volume      | 1.464 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Air Blank          | 0.000 | 01:36 |
| Subject Sample # 2 | 0.218 | 01:37 |
| Breath Volume      | 2.878 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.080 | 01:38 |
| Air Blank          | 0.000 | 01:39 |
| Diagnostics Check  | OK    | 01:39 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PAIGE,ROGER,S  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PAIGE,ROGER  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 15:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 11/17/2024 16:19:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:16 |                      |
| Air Blank         | 0.000 | 16:17 |                      |
| Control Test      | 0.080 | 16:17 |                      |
| Air Blank         | 0.000 | 16:17 |                      |
| Subject Sample    | # 1   | 16:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:18 |                      |
| Air Blank         | 0.000 | 16:18 |                      |
| Control Test      | 0.081 | 16:18 |                      |
| Air Blank         | 0.000 | 16:19 |                      |
| Diagnostics Check | OK    | 16:19 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEGUERO,FEDERICO,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PEGUERO,FEDERICO  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 08:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 11/18/2024 09:22:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 09:15 |
| Air Blank          | 0.000 | 09:15 |
| Control Test       | 0.080 | 09:16 |
| Air Blank          | 0.000 | 09:16 |
| Subject Sample # 1 | 0.131 | 09:17 |
| Breath Volume      | 1.945 | 09:17 |
| Air Blank          | 0.000 | 09:18 |
| Air Blank          | 0.000 | 09:20 |
| Subject Sample # 2 | 0.126 | 09:20 |
| Breath Volume      | 1.640 | 09:20 |
| Air Blank          | 0.000 | 09:21 |
| Control Test       | 0.077 | 09:22 |
| Air Blank          | 0.000 | 09:22 |
| Diagnostics Check  | OK    | 09:22 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREITAS,MARCIA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FREITAS,MARCIA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: OTHER  
ARREST TIME: 12:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: CHINCHILLA LOPEZ

DATE/TIME OF TEST: 11/18/2024 23:26:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Control Test      | 0.081 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Subject Sample    | # 1   | 23:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:25 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.080 | 23:26 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Diagnostics Check | OK    | 23:26 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: MENDOZA,SEBASTIAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GUERRERO OSORIO

DATE/TIME OF TEST: 11/20/2024 00:32:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:26 |
| Air Blank          | 0.000 | 00:26 |
| Control Test       | 0.081 | 00:26 |
| Air Blank          | 0.000 | 00:27 |
| Subject Sample # 1 | 0.161 | 00:27 |
| Breath Volume      | 1.863 | 00:27 |
| Air Blank          | 0.000 | 00:28 |
| Air Blank          | 0.000 | 00:30 |
| Subject Sample # 2 | 0.156 | 00:30 |
| Breath Volume      | 1.792 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Control Test       | 0.079 | 00:31 |
| Air Blank          | 0.000 | 00:31 |
| Diagnostics Check  | OK    | 00:32 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,ROBERT,M  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: PEREZ,ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 19:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GAFFNEY

DATE/TIME OF TEST: 11/20/2024 20:38:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:35 |                      |
| Air Blank         | 0.000 | 20:35 |                      |
| Control Test      | 0.080 | 20:36 |                      |
| Air Blank         | 0.000 | 20:36 |                      |
| Subject Sample    | # 1   | 20:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:36 |                      |
| Air Blank         | 0.000 | 20:37 |                      |
| Control Test      | 0.079 | 20:37 |                      |
| Air Blank         | 0.000 | 20:38 |                      |
| Diagnostics Check | OK    | 20:38 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTANA,DAVID,D  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: SANTANA,DAVID  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 11/21/2024 22:23:25

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:20 |                      |
| Air Blank         | 0.000 | 22:20 |                      |
| Control Test      | 0.082 | 22:21 |                      |
| Air Blank         | 0.000 | 22:21 |                      |
| Subject Sample    | # 1   | 22:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:22 |                      |
| Air Blank         | 0.000 | 22:22 |                      |
| Control Test      | 0.081 | 22:22 |                      |
| Air Blank         | 0.000 | 22:23 |                      |
| Diagnostics Check | OK    | 22:23 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEGUERO,FEDERICO,  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PEGUERO,FEDERICO  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 12:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/23/2024 13:34:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:28 |
| Air Blank          | 0.000 | 13:29 |
| Control Test       | 0.081 | 13:29 |
| Air Blank          | 0.000 | 13:30 |
| Subject Sample # 1 | 0.000 | 13:30 |
| Breath Volume      | 2.636 | 13:30 |
| Air Blank          | 0.000 | 13:30 |
| Air Blank          | 0.000 | 13:32 |
| Subject Sample # 2 | 0.000 | 13:33 |
| Breath Volume      | 2.710 | 13:33 |
| Air Blank          | 0.000 | 13:33 |
| Control Test       | 0.080 | 13:34 |
| Air Blank          | 0.000 | 13:34 |
| Diagnostics Check  | OK    | 13:34 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DESANDRE,DAVID,H  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: DESANDRE,DAVID  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 20:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SHREWSBURY

DATE/TIME OF TEST: 11/24/2024 21:31:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:24 |
| Air Blank          | 0.000 | 21:25 |
| Control Test       | 0.081 | 21:25 |
| Air Blank          | 0.000 | 21:26 |
| Subject Sample # 1 | 0.000 | 21:26 |
| Breath Volume      | 5.699 | 21:26 |
| Air Blank          | 0.000 | 21:27 |
| Air Blank          | 0.000 | 21:29 |
| Subject Sample # 2 | 0.000 | 21:29 |
| Breath Volume      | 4.269 | 21:29 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.079 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Diagnostics Check  | OK    | 21:31 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FREITAS,MARCIA,A  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: FREITAS,MARCIA  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GARCIA CASTRO

DATE/TIME OF TEST: 11/25/2024 01:59:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.081 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 1 | 0.221 | 01:54 |
| Breath Volume      | 1.562 | 01:54 |
| Air Blank          | 0.000 | 01:55 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 2 | 0.218 | 01:57 |
| Breath Volume      | 1.550 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Control Test       | 0.079 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Diagnostics Check  | OK    | 01:59 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LINNELL,JASON,E  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: LINNELL,JASON  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 20:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/25/2024 21:45:47

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 21:34 |                    |
| Air Blank         | 0.000 | 21:34 |                    |
| Control Test      | 0.081 | 21:35 |                    |
| Air Blank         | 0.000 | 21:35 |                    |
| Subject Sample    | # 1   | 21:38 | No Sample Provided |
| Breath Volume     | 0.000 | 21:38 |                    |
| Air Blank         | 0.000 | 21:39 |                    |
| Air Blank         | 0.000 | 21:41 |                    |
| Subject Sample    | # 2   | 21:44 | No Sample Provided |
| Breath Volume     | 0.000 | 21:44 |                    |
| Air Blank         | 0.000 | 21:44 |                    |
| Control Test      | 0.081 | 21:45 |                    |
| Air Blank         | 0.000 | 21:45 |                    |
| Diagnostics Check | OK    | 21:45 |                    |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZAREM, BRANDEN, M  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: HAMMONS, VICTORIA  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 16:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: BURKWEST

DATE/TIME OF TEST: 11/27/2024 18:03:46

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:00 |                      |
| Air Blank         | 0.000 | 18:01 |                      |
| Control Test      | 0.080 | 18:01 |                      |
| Air Blank         | 0.000 | 18:02 |                      |
| Subject Sample    | # 1   | 18:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:02 |                      |
| Air Blank         | 0.000 | 18:02 |                      |
| Control Test      | 0.080 | 18:03 |                      |
| Air Blank         | 0.000 | 18:03 |                      |
| Diagnostics Check | OK    | 18:03 |                      |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PHILLIPS,V,J  
BREATH TEST OPERATOR AGENCY: Ocala PD  
ARRESTING OFFICER: COUGHLIN,M  
ARRESTING OFFICER AGENCY: Ocala PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: PROVITOLA

DATE/TIME OF TEST: 11/27/2024 23:00:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:54 |
| Air Blank          | 0.000 | 22:54 |
| Control Test       | 0.081 | 22:55 |
| Air Blank          | 0.000 | 22:55 |
| Subject Sample # 1 | 0.156 | 22:56 |
| Breath Volume      | 3.011 | 22:56 |
| Air Blank          | 0.000 | 22:56 |
| Air Blank          | 0.000 | 22:58 |
| Subject Sample # 2 | 0.148 | 22:59 |
| Breath Volume      | 2.152 | 22:59 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.080 | 23:00 |
| Air Blank          | 0.000 | 23:00 |
| Diagnostics Check  | OK    | 23:00 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANCHEZ,ASHLEY,E  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: RIOS,KENNETH  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: APARICIO GUZMAN

DATE/TIME OF TEST: 11/28/2024 00:36:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:30 |
| Air Blank          | 0.000 | 00:30 |
| Control Test       | 0.081 | 00:30 |
| Air Blank          | 0.000 | 00:31 |
| Subject Sample # 1 | 0.152 | 00:31 |
| Breath Volume      | 1.695 | 00:31 |
| Air Blank          | 0.000 | 00:32 |
| Air Blank          | 0.000 | 00:34 |
| Subject Sample # 2 | 0.147 | 00:35 |
| Breath Volume      | 1.531 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Control Test       | 0.080 | 00:36 |
| Air Blank          | 0.000 | 00:36 |
| Diagnostics Check  | OK    | 00:36 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LATHAM,JAMIE,L  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: PATTERSON,HANNAH  
ARRESTING OFFICER AGENCY: Dunnellon PD  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CINTRON

DATE/TIME OF TEST: 11/29/2024 02:38:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Subject Sample # 1 | 0.150 | 02:32 |
| Breath Volume      | 1.789 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 2 | 0.154 | 02:36 |
| Breath Volume      | 2.007 | 02:36 |
| Air Blank          | 0.000 | 02:37 |
| Control Test       | 0.079 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Diagnostics Check  | OK    | 02:38 |

---

Data Download Date/Time: 11/29/2024 15:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: PEREZ RIVERO

DATE/TIME OF TEST: 11/30/2024 22:02:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:55 |
| Air Blank          | 0.000 | 21:56 |
| Control Test       | 0.081 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 1 | 0.225 | 21:58 |
| Breath Volume      | 1.480 | 21:58 |
| Air Blank          | 0.000 | 21:58 |
| Air Blank          | 0.000 | 22:00 |
| Subject Sample # 2 | 0.221 | 22:01 |
| Breath Volume      | 1.894 | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.079 | 22:02 |
| Air Blank          | 0.000 | 22:02 |
| Diagnostics Check  | OK    | 22:02 |

---

Data Download Date/Time: 12/18/2024 18:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23324080A2  
DRY GAS EXPIRY DATE: 10/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/26/2024 23:41:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.079 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Subject Sample # 1 | 0.120 | 23:36 |
| Breath Volume      | 3.363 | 23:36 |
| Air Blank          | 0.000 | 23:37 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 2 | 0.120 | 23:39 |
| Breath Volume      | 2.500 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.079 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Diagnostics Check  | OK    | 23:41 |

---

Data Download Date/Time: 12/28/2024 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23324080A2  
DRY GAS EXPIRY DATE: 10/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/29/2024 22:15:10

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:03 |                    |
| Air Blank         | 0.000 | 22:04 |                    |
| Control Test      | 0.080 | 22:04 |                    |
| Air Blank         | 0.000 | 22:05 |                    |
| Subject Sample    | # 1   | 22:08 | No Sample Provided |
| Breath Volume     | 0.000 | 22:08 |                    |
| Air Blank         | 0.000 | 22:08 |                    |
| Air Blank         | 0.000 | 22:10 |                    |
| Subject Sample    | # 2   | 22:13 | No Sample Provided |
| Breath Volume     | 0.000 | 22:13 |                    |
| Air Blank         | 0.000 | 22:14 |                    |
| Control Test      | 0.080 | 22:14 |                    |
| Air Blank         | 0.000 | 22:15 |                    |
| Diagnostics Check | OK    | 22:15 |                    |

---

Data Download Date/Time: 12/28/2024 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAYTON,A,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23324080A2  
DRY GAS EXPIRY DATE: 10/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LUNDBERG

DATE/TIME OF TEST: 11/30/2024 01:38:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 1 | 0.173 | 01:33 |
| Breath Volume      | 1.621 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.173 | 01:36 |
| Breath Volume      | 1.679 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.079 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Diagnostics Check  | OK    | 01:38 |

---

Data Download Date/Time: 12/28/2024 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAYTON,A,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: FERRO,T  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23324080A2  
DRY GAS EXPIRY DATE: 10/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SANDERS-HOLLAND

DATE/TIME OF TEST: 11/30/2024 02:18:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:13 |                      |
| Air Blank         | 0.000 | 02:14 |                      |
| Control Test      | 0.078 | 02:14 |                      |
| Air Blank         | 0.000 | 02:15 |                      |
| Subject Sample    | # 1   | 02:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Control Test      | 0.078 | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Diagnostics Check | OK    | 02:18 |                      |

---

Data Download Date/Time: 12/28/2024 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAYTON,A,  
BREATH TEST OPERATOR AGENCY: Hernando County SO  
ARRESTING OFFICER: CLAYTON,A  
ARRESTING OFFICER AGENCY: Hernando County SO  
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 23324080A2  
DRY GAS EXPIRY DATE: 10/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 04:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: FIEDLER

DATE/TIME OF TEST: 11/30/2024 06:22:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:15 |
| Air Blank          | 0.000 | 06:15 |
| Control Test       | 0.080 | 06:16 |
| Air Blank          | 0.000 | 06:16 |
| Subject Sample # 1 | 0.120 | 06:17 |
| Breath Volume      | 2.531 | 06:17 |
| Air Blank          | 0.000 | 06:17 |
| Air Blank          | 0.000 | 06:19 |
| Subject Sample # 2 | 0.123 | 06:20 |
| Breath Volume      | 3.144 | 06:20 |
| Air Blank          | 0.000 | 06:21 |
| Control Test       | 0.078 | 06:21 |
| Air Blank          | 0.000 | 06:22 |
| Diagnostics Check  | OK    | 06:22 |

---

Data Download Date/Time: 12/28/2024 11:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L  
BREATH TEST OPERATOR AGENCY: Sumter County SO  
ARRESTING OFFICER: GOEHRING,J  
ARRESTING OFFICER AGENCY: Sumter County SO  
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231603  
DRY GAS EXPIRY DATE: 06/05/2026  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PALADANO

DATE/TIME OF TEST: 11/30/2024 01:28:36

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:25 |                      |
| Air Blank         | 0.000 | 01:26 |                      |
| Control Test      | 0.079 | 01:26 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Subject Sample    | # 1   | 01:27 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:27 |                      |
| Air Blank         | 0.000 | 01:27 |                      |
| Control Test      | 0.079 | 01:27 |                      |
| Air Blank         | 0.000 | 01:28 |                      |
| Diagnostics Check | OK    | 01:28 |                      |

---

Data Download Date/Time: 12/29/2024 18:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLOWAY, GRANT, D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: GALLOWAY, GRANT  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: FITTS

DATE/TIME OF TEST: 11/05/2024 17:26:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:20 |
| Air Blank          | 0.000 | 17:21 |
| Control Test       | 0.079 | 17:21 |
| Air Blank          | 0.000 | 17:21 |
| Subject Sample # 1 | 0.185 | 17:22 |
| Breath Volume      | 1.632 | 17:22 |
| Air Blank          | 0.000 | 17:22 |
| Air Blank          | 0.000 | 17:24 |
| Subject Sample # 2 | 0.193 | 17:25 |
| Breath Volume      | 1.882 | 17:25 |
| Air Blank          | 0.000 | 17:25 |
| Control Test       | 0.080 | 17:26 |
| Air Blank          | 0.000 | 17:26 |
| Diagnostics Check  | OK    | 17:26 |

---

Data Download Date/Time: 12/01/2024 18:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: HAMPSHIRE,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: LEACH

DATE/TIME OF TEST: 11/05/2024 23:02:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:59 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Control Test      | 0.079 | 23:00 |                      |
| Air Blank         | 0.000 | 23:00 |                      |
| Subject Sample    | # 1   | 23:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:00 |                      |
| Air Blank         | 0.000 | 23:01 |                      |
| Control Test      | 0.079 | 23:01 |                      |
| Air Blank         | 0.000 | 23:02 |                      |
| Diagnostics Check | OK    | 23:02 |                      |

---

Data Download Date/Time: 12/01/2024 19:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GASSERT

DATE/TIME OF TEST: 11/05/2024 23:21:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:13 |
| Air Blank          | 0.000 | 23:14 |
| Control Test       | 0.079 | 23:14 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 1 | 0.211 | 23:16 |
| Breath Volume      | 1.574 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Air Blank          | 0.000 | 23:19 |
| Subject Sample # 2 | 0.212 | 23:19 |
| Breath Volume      | 1.476 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.078 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Diagnostics Check  | OK    | 23:21 |

---

Data Download Date/Time: 12/01/2024 19:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: KITO,DEPUTY  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 11/07/2024 01:05:30

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:02 |                      |
| Air Blank         | 0.000 | 01:03 |                      |
| Control Test      | 0.080 | 01:03 |                      |
| Air Blank         | 0.000 | 01:03 |                      |
| Subject Sample    | # 1   | 01:04 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:04 |                      |
| Air Blank         | 0.000 | 01:04 |                      |
| Control Test      | 0.080 | 01:04 |                      |
| Air Blank         | 0.000 | 01:05 |                      |
| Diagnostics Check | OK    | 01:05 |                      |

---

Data Download Date/Time: 12/01/2024 19:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAYLOR,TROY,  
BREATH TEST OPERATOR AGENCY: Citrus County SO  
ARRESTING OFFICER: CARTER,SHADE  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: SAGAR

DATE/TIME OF TEST: 11/09/2024 11:41:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 11:32 |                |
| Air Blank         | 0.000     | 11:33 |                |
| Control Test      | 0.079     | 11:33 |                |
| Air Blank         | 0.000     | 11:34 |                |
| Subject Sample    | # 1 0.000 | 11:37 | Volume Not Met |
| Breath Volume     | 0.226     | 11:37 |                |
| Air Blank         | 0.000     | 11:37 |                |
| Air Blank         | 0.000     | 11:39 |                |
| Subject Sample    | # 2       | 11:40 | RFI Detect     |
| Breath Volume     | 0.000     | 11:40 |                |
| Air Blank         |           | 11:40 | RFI Detect     |
| Air Blank         | 0.000     | 11:41 |                |

---

Data Download Date/Time: 12/01/2024 19:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: WALKER,J  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 11/10/2024 04:10:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:08 |                      |
| Air Blank         | 0.000 | 04:08 |                      |
| Control Test      | 0.079 | 04:08 |                      |
| Air Blank         | 0.000 | 04:09 |                      |
| Subject Sample    | # 1   | 04:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:09 |                      |
| Air Blank         | 0.000 | 04:10 |                      |
| Control Test      | 0.079 | 04:10 |                      |
| Air Blank         | 0.000 | 04:10 |                      |
| Diagnostics Check | OK    | 04:10 |                      |

---

Data Download Date/Time: 12/01/2024 19:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURKE,BRITTANY,A  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: FERRER,D  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 14:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: FITTS

DATE/TIME OF TEST: 11/10/2024 16:08:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:01 |
| Air Blank          | 0.000 | 16:02 |
| Control Test       | 0.079 | 16:02 |
| Air Blank          | 0.000 | 16:03 |
| Subject Sample # 1 | 0.097 | 16:03 |
| Breath Volume      | 2.527 | 16:03 |
| Air Blank          | 0.000 | 16:04 |
| Air Blank          | 0.000 | 16:06 |
| Subject Sample # 2 | 0.086 | 16:06 |
| Breath Volume      | 1.109 | 16:06 |
| Air Blank          | 0.000 | 16:07 |
| Control Test       | 0.077 | 16:07 |
| Air Blank          | 0.000 | 16:08 |
| Diagnostics Check  | OK    | 16:08 |

---

Data Download Date/Time: 12/01/2024 19:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: STINES

DATE/TIME OF TEST: 11/11/2024 01:33:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:31 |                      |
| Air Blank         | 0.000 | 01:31 |                      |
| Control Test      | 0.079 | 01:31 |                      |
| Air Blank         | 0.000 | 01:32 |                      |
| Subject Sample    | # 1   | 01:32 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:32 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Control Test      | 0.080 | 01:33 |                      |
| Air Blank         | 0.000 | 01:33 |                      |
| Diagnostics Check | OK    | 01:33 |                      |

---

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: GARRISON,T  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: STINES

DATE/TIME OF TEST: 11/11/2024 00:49:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:46 |                      |
| Air Blank         | 0.000 | 00:46 |                      |
| Control Test      | 0.079 | 00:47 |                      |
| Air Blank         | 0.000 | 00:47 |                      |
| Subject Sample    | # 1   | 00:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:47 |                      |
| Air Blank         | 0.000 | 00:48 |                      |
| Control Test      | 0.079 | 00:48 |                      |
| Air Blank         | 0.000 | 00:49 |                      |
| Diagnostics Check | OK    | 00:49 |                      |

---

Data Download Date/Time: 12/01/2024 19:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BERLYNE,R,D  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: BERLYNE,R  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ALDRICH

DATE/TIME OF TEST: 11/13/2024 03:16:25

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 02:59 |                |
| Air Blank         | 0.000     | 02:59 |                |
| Control Test      | 0.079     | 03:00 |                |
| Air Blank         | 0.000     | 03:00 |                |
| Subject Sample    | # 1 0.148 | 03:03 | Volume Not Met |
| Breath Volume     | 0.367     | 03:03 |                |
| Air Blank         | 0.000     | 03:04 |                |
| Air Blank         | 0.000     | 03:06 |                |
| Subject Sample    | # 2 0.188 | 03:09 |                |
| Breath Volume     | 1.738     | 03:09 |                |
| Air Blank         | 0.000     | 03:09 |                |
| Air Blank         | 0.000     | 03:11 |                |
| Subject Sample    | # 3 0.179 | 03:14 | Volume Not Met |
| Breath Volume     | 0.097     | 03:14 |                |
| Air Blank         | 0.000     | 03:15 |                |
| Control Test      | 0.075     | 03:15 |                |
| Air Blank         | 0.000     | 03:16 |                |
| Diagnostics Check | OK        | 03:16 |                |

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: ROSADO,A  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 16:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 11/17/2024 19:22:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:15 |
| Air Blank          | 0.000 | 19:15 |
| Control Test       | 0.079 | 19:16 |
| Air Blank          | 0.000 | 19:16 |
| Subject Sample # 1 | 0.222 | 19:17 |
| Breath Volume      | 2.246 | 19:17 |
| Air Blank          | 0.000 | 19:18 |
| Air Blank          | 0.000 | 19:19 |
| Subject Sample # 2 | 0.214 | 19:20 |
| Breath Volume      | 1.492 | 19:20 |
| Air Blank          | 0.000 | 19:21 |
| Control Test       | 0.079 | 19:21 |
| Air Blank          | 0.000 | 19:22 |
| Diagnostics Check  | OK    | 19:22 |

---

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: FERRER,B  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MERRICK

DATE/TIME OF TEST: 11/22/2024 23:01:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:54 |
| Air Blank          | 0.000 | 22:55 |
| Control Test       | 0.079 | 22:55 |
| Air Blank          | 0.000 | 22:56 |
| Subject Sample # 1 | 0.147 | 22:56 |
| Breath Volume      | 2.234 | 22:56 |
| Air Blank          | 0.000 | 22:57 |
| Air Blank          | 0.000 | 22:59 |
| Subject Sample # 2 | 0.149 | 22:59 |
| Breath Volume      | 3.300 | 22:59 |
| Air Blank          | 0.000 | 23:00 |
| Control Test       | 0.079 | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Diagnostics Check  | OK    | 23:01 |

---

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,LENNARD,  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: KNIGHT,C  
ARRESTING OFFICER AGENCY: Citrus County SO  
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 11/27/2024 01:58:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Control Test       | 0.079 | 01:52 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 1 | 0.147 | 01:53 |
| Breath Volume      | 2.066 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Air Blank          | 0.000 | 01:56 |
| Subject Sample # 2 | 0.136 | 01:56 |
| Breath Volume      | 1.464 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.079 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Diagnostics Check  | OK    | 01:58 |

---

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROWN,JEFFERY,P  
BREATH TEST OPERATOR AGENCY: CCA Citrus  
ARRESTING OFFICER: KICKSTAD,ERIC  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 231043  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 11/30/2024 23:51:53

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:49 |                      |
| Air Blank         | 0.000 | 23:49 |                      |
| Control Test      | 0.079 | 23:49 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Subject Sample    | # 1   | 23:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:50 |                      |
| Air Blank         | 0.000 | 23:50 |                      |
| Control Test      | 0.080 | 23:51 |                      |
| Air Blank         | 0.000 | 23:51 |                      |
| Diagnostics Check | OK    | 23:51 |                      |

---

Data Download Date/Time: 12/01/2024 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROWE,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: ROMAN,ALONZO  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GERMOSEN

DATE/TIME OF TEST: 11/02/2024 00:54:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.080 | 00:49 |
| Air Blank          | 0.000 | 00:49 |
| Subject Sample # 1 | 0.196 | 00:50 |
| Breath Volume      | 1.601 | 00:50 |
| Air Blank          | 0.000 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Subject Sample # 2 | 0.211 | 00:53 |
| Breath Volume      | 1.644 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Control Test       | 0.080 | 00:54 |
| Air Blank          | 0.000 | 00:54 |
| Diagnostics Check  | OK    | 00:54 |

---

Data Download Date/Time: 11/18/2024 14:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROWE,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: NELSON,GARRETT  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: RODMAN

DATE/TIME OF TEST: 11/02/2024 01:28:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Control Test       | 0.079 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 1 | 0.098 | 01:24 |
| Breath Volume      | 2.296 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Subject Sample # 2 | 0.095 | 01:27 |
| Breath Volume      | 2.121 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.079 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Diagnostics Check  | OK    | 01:28 |

---

Data Download Date/Time: 11/18/2024 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CROWE,SCOTT,M  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: NELSON,GARRETT  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SIBILA-LOPEZ

DATE/TIME OF TEST: 11/08/2024 03:29:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.081 | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.000 | 03:25 |
| Breath Volume      | 2.312 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 2 | 0.000 | 03:27 |
| Breath Volume      | 2.347 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Control Test       | 0.080 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Diagnostics Check  | OK    | 03:29 |

---

Data Download Date/Time: 11/18/2024 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: VELEZ,JORDEN  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 08223080A5  
DRY GAS EXPIRY DATE: 06/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 11/10/2024 20:56:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:48 |
| Air Blank          | 0.000 | 20:48 |
| Control Test       | 0.075 | 20:48 |
| Air Blank          | 0.000 | 20:49 |
| Subject Sample # 1 | 0.158 | 20:50 |
| Breath Volume      | 2.113 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Air Blank          | 0.000 | 20:52 |
| Subject Sample # 2 | 0.155 | 20:54 |
| Breath Volume      | 1.765 | 20:54 |
| Air Blank          | 0.000 | 20:55 |
| Control Test       | 0.078 | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Diagnostics Check  | OK    | 20:56 |

---

Data Download Date/Time: 11/18/2024 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BLUM,SAVANNAH,K

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: NELSON,GARRETT

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A5

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 03:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 11/17/2024 04:25:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.079 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 1 | 0.168 | 04:21 |
| Breath Volume      | 2.410 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 2 | 0.169 | 04:23 |
| Breath Volume      | 2.214 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Control Test       | 0.080 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Diagnostics Check  | OK    | 04:25 |

---

Data Download Date/Time: 11/18/2024 14:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 11/02/2024 23:50:27

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 23:48 |            |
| Air Blank         | 0.000 | 23:48 |            |
| Control Test      | 0.079 | 23:49 |            |
| Air Blank         | 0.000 | 23:49 |            |
| Subject Sample    | # 1   | 23:49 | RFI Detect |
| Breath Volume     | 0.000 | 23:49 |            |
| Air Blank         |       | 23:49 | RFI Detect |
| Air Blank         | 0.000 | 23:50 |            |

---

Data Download Date/Time: 11/29/2024 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: TROUTMAN

DATE/TIME OF TEST: 11/02/2024 23:53:06

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 23:52 |            |
| Air Blank         |       | 23:52 | RFI Detect |
| Air Blank         | 0.000 | 23:53 |            |

---

Data Download Date/Time: 11/29/2024 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R  
BREATH TEST OPERATOR AGENCY: FHP Troop B  
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN  
ARRESTING OFFICER AGENCY: FHP Troop B  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BOLLES

DATE/TIME OF TEST: 11/04/2024 01:33:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:26 |
| Air Blank          | 0.000 | 01:27 |
| Control Test       | 0.080 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 1 | 0.255 | 01:28 |
| Breath Volume      | 2.238 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Air Blank          | 0.000 | 01:31 |
| Subject Sample # 2 | 0.264 | 01:31 |
| Breath Volume      | 2.484 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Diagnostics Check  | OK    | 01:33 |

---

Data Download Date/Time: 11/29/2024 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COYNE

DATE/TIME OF TEST: 11/04/2024 02:57:28

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 02:56 |            |
| Air Blank         | 0.000 | 02:56 |            |
| Control Test      | 0.080 | 02:56 |            |
| Air Blank         |       | 02:57 | RFI Detect |
| Air Blank         |       | 02:57 | RFI Detect |

---

Data Download Date/Time: 11/29/2024 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YOUNG,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Marion County SO  
ARRESTING OFFICER: YOUNG,ROBERT  
ARRESTING OFFICER AGENCY: Marion County SO  
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 403008479  
DRY GAS EXPIRY DATE: 03/28/2027  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: COYNE

DATE/TIME OF TEST: 11/04/2024 03:04:21

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 02:58 |            |
| Air Blank         | 0.000     | 02:59 |            |
| Control Test      | 0.079     | 02:59 |            |
| Air Blank         | 0.000     | 02:59 |            |
| Subject Sample    | # 1 0.240 | 03:00 |            |
| Breath Volume     | 2.390     | 03:00 |            |
| Air Blank         | 0.000     | 03:01 |            |
| Air Blank         | 0.000     | 03:02 |            |
|                   |           | 03:02 | RFI Detect |
| Air Blank         | 0.000     | 03:03 |            |
| Control Test      | 0.080     | 03:03 |            |
| Air Blank         | 0.000     | 03:04 |            |
| Diagnostics Check | OK        | 03:04 |            |

---

Data Download Date/Time: 11/29/2024 16:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: RODRIGUEZ,C  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/01/2024 02:03:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Control Test       | 0.081 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Subject Sample # 1 | 0.049 | 01:59 |
| Breath Volume      | 2.445 | 01:59 |
| Air Blank          | 0.000 | 01:59 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 2 | 0.048 | 02:02 |
| Breath Volume      | 2.164 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.080 | 02:03 |
| Air Blank          | 0.000 | 02:03 |
| Diagnostics Check  | OK    | 02:03 |

---

Data Download Date/Time: 11/15/2024 09:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAKER,J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: WERSAL

DATE/TIME OF TEST: 11/01/2024 04:48:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:41 |
| Air Blank          | 0.000 | 04:41 |
| Control Test       | 0.081 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 1 | 0.148 | 04:43 |
| Breath Volume      | 1.273 | 04:43 |
| Air Blank          | 0.000 | 04:44 |
| Air Blank          | 0.000 | 04:46 |
| Subject Sample # 2 | 0.165 | 04:47 |
| Breath Volume      | 3.406 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.079 | 04:48 |
| Air Blank          | 0.000 | 04:48 |
| Diagnostics Check  | OK    | 04:48 |

---

Data Download Date/Time: 11/15/2024 09:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LANDRY, G  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: TOWNSEND

DATE/TIME OF TEST: 11/02/2024 00:51:03

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 00:39 |                |
| Air Blank         | 0.000     | 00:39 |                |
| Control Test      | 0.080     | 00:40 |                |
| Air Blank         | 0.000     | 00:40 |                |
| Subject Sample    | # 1 0.149 | 00:43 | Volume Not Met |
| Breath Volume     | 0.480     | 00:43 |                |
| Air Blank         | 0.000     | 00:44 |                |
| Air Blank         | 0.000     | 00:46 |                |
| Subject Sample    | # 2 0.162 | 00:49 | Volume Not Met |
| Breath Volume     | 0.773     | 00:49 |                |
| Air Blank         | 0.000     | 00:50 |                |
| Control Test      | 0.078     | 00:50 |                |
| Air Blank         | 0.000     | 00:50 |                |
| Diagnostics Check | OK        | 00:51 |                |

---

Data Download Date/Time: 11/15/2024 09:56

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: BADEN,T  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BOBBITT

DATE/TIME OF TEST: 11/07/2024 02:07:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.079 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Subject Sample # 1 | 0.195 | 02:02 |
| Breath Volume      | 2.238 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 2 | 0.203 | 02:06 |
| Breath Volume      | 2.292 | 02:06 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.079 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Diagnostics Check  | OK    | 02:07 |

---

Data Download Date/Time: 11/15/2024 09:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: JACOBS, D  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: CULEBRO PEREZ

DATE/TIME OF TEST: 11/10/2024 03:48:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Control Test       | 0.081 | 03:41 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 1 | 0.106 | 03:43 |
| Breath Volume      | 2.718 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Subject Sample # 2 | 0.101 | 03:46 |
| Breath Volume      | 1.523 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.079 | 03:47 |
| Air Blank          | 0.000 | 03:47 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 11/15/2024 09:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Tampa PD  
ARRESTING OFFICER: SARRASIN,R  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 11/10/2024 06:30:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:22 |
| Air Blank          | 0.000 | 06:22 |
| Control Test       | 0.081 | 06:22 |
| Air Blank          | 0.000 | 06:23 |
| Subject Sample # 1 | 0.161 | 06:25 |
| Breath Volume      | 1.886 | 06:25 |
| Air Blank          | 0.000 | 06:25 |
| Air Blank          | 0.000 | 06:27 |
| Subject Sample # 2 | 0.168 | 06:28 |
| Breath Volume      | 1.398 | 06:28 |
| Air Blank          | 0.000 | 06:29 |
| Control Test       | 0.079 | 06:29 |
| Air Blank          | 0.000 | 06:29 |
| Diagnostics Check  | OK    | 06:30 |

---

Data Download Date/Time: 11/15/2024 09:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: VAN-TREESE, S  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: HALLORAN

DATE/TIME OF TEST: 11/10/2024 22:01:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:55 |
| Air Blank          | 0.000 | 21:55 |
| Control Test       | 0.081 | 21:56 |
| Air Blank          | 0.000 | 21:56 |
| Subject Sample # 1 | 0.231 | 21:57 |
| Breath Volume      | 2.542 | 21:57 |
| Air Blank          | 0.000 | 21:58 |
| Air Blank          | 0.000 | 21:59 |
| Subject Sample # 2 | 0.230 | 22:00 |
| Breath Volume      | 2.101 | 22:00 |
| Air Blank          | 0.000 | 22:00 |
| Control Test       | 0.078 | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Diagnostics Check  | OK    | 22:01 |

---

Data Download Date/Time: 11/15/2024 09:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TRUJILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 10/17/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: OLIVA SALGUERO

DATE/TIME OF TEST: 11/11/2024 01:19:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:08 |
| Control Test       | 0.080 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.189 | 01:10 |
| Breath Volume      | 2.117 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 2 | 0.168 | 01:15 |
| Breath Volume      | 2.324 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 3 | 0.182 | 01:18 |
| Breath Volume      | 1.179 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Control Test       | 0.077 | 01:19 |
| Air Blank          | 0.000 | 01:19 |
| Diagnostics Check  | OK    | 01:19 |

---

Data Download Date/Time: 11/15/2024 09:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RODRIGUEZ, M  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JUPITER

DATE/TIME OF TEST: 11/16/2024 02:54:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.081 | 02:49 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 1 | 0.136 | 02:50 |
| Breath Volume      | 3.230 | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Air Blank          | 0.000 | 02:52 |
| Subject Sample # 2 | 0.133 | 02:53 |
| Breath Volume      | 3.675 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Diagnostics Check  | OK    | 02:54 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: LEISCH,B  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LETANG

DATE/TIME OF TEST: 11/16/2024 03:51:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.080 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 1 | 0.192 | 03:47 |
| Breath Volume      | 1.847 | 03:47 |
| Air Blank          | 0.000 | 03:47 |
| Air Blank          | 0.000 | 03:49 |
| Subject Sample # 2 | 0.190 | 03:50 |
| Breath Volume      | 1.761 | 03:50 |
| Air Blank          | 0.000 | 03:50 |
| Control Test       | 0.079 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Diagnostics Check  | OK    | 03:51 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: ROSADO

DATE/TIME OF TEST: 11/16/2024 05:16:16

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 05:08 |                      |
| Air Blank         | 0.000     | 05:08 |                      |
| Control Test      | 0.080     | 05:08 |                      |
| Air Blank         | 0.000     | 05:09 |                      |
| Subject Sample    | # 1 0.201 | 05:10 |                      |
| Breath Volume     | 1.632     | 05:10 |                      |
| Air Blank         | 0.000     | 05:10 |                      |
| Air Blank         | 0.000     | 05:12 |                      |
| Subject Sample    | # 2       | 05:14 | Subject Test Refused |
| Breath Volume     | 0.000     | 05:14 |                      |
| Air Blank         | 0.000     | 05:15 |                      |
| Control Test      | 0.079     | 05:15 |                      |
| Air Blank         | 0.000     | 05:16 |                      |
| Diagnostics Check | OK        | 05:16 |                      |

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLEIN,ETHAN,H  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PEPEN,J  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: FALL CHAVEZ

DATE/TIME OF TEST: 11/17/2024 01:54:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:47 |
| Air Blank          | 0.000 | 01:47 |
| Control Test       | 0.081 | 01:48 |
| Air Blank          | 0.000 | 01:48 |
| Subject Sample # 1 | 0.051 | 01:49 |
| Breath Volume      | 1.656 | 01:49 |
| Air Blank          | 0.000 | 01:49 |
| Air Blank          | 0.000 | 01:51 |
| Subject Sample # 2 | 0.051 | 01:52 |
| Breath Volume      | 1.578 | 01:52 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.080 | 01:53 |
| Air Blank          | 0.000 | 01:54 |
| Diagnostics Check  | OK    | 01:54 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: BAKER, J  
ARRESTING OFFICER AGENCY: Tampa PD  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: YUHAS

DATE/TIME OF TEST: 11/20/2024 23:44:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Control Test       | 0.080 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Subject Sample # 1 | 0.156 | 23:39 |
| Breath Volume      | 1.593 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Air Blank          | 0.000 | 23:42 |
| Subject Sample # 2 | 0.136 | 23:42 |
| Breath Volume      | 1.164 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.079 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Diagnostics Check  | OK    | 23:44 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,MINDY,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: RENDON,D  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 11/23/2024 04:33:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.083 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 1 | 0.089 | 04:28 |
| Breath Volume      | 1.750 | 04:28 |
| Air Blank          | 0.000 | 04:29 |
| Air Blank          | 0.000 | 04:30 |
| Subject Sample # 2 | 0.088 | 04:31 |
| Breath Volume      | 2.250 | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Control Test       | 0.081 | 04:32 |
| Air Blank          | 0.000 | 04:32 |
| Diagnostics Check  | OK    | 04:33 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: KLIPIN, A  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: COMM VEHIC  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MCFIELD JR

DATE/TIME OF TEST: 11/24/2024 03:18:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Control Test       | 0.082 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Subject Sample # 1 | 0.145 | 03:14 |
| Breath Volume      | 2.578 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Air Blank          | 0.000 | 03:16 |
| Subject Sample # 2 | 0.145 | 03:17 |
| Breath Volume      | 2.066 | 03:17 |
| Air Blank          | 0.000 | 03:17 |
| Control Test       | 0.080 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Diagnostics Check  | OK    | 03:18 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: TRUJILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RIVERA-AYALA

DATE/TIME OF TEST: 11/24/2024 03:39:16

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.080 | 03:32 |
| Air Blank          | 0.000 | 03:33 |
| Subject Sample # 1 | 0.178 | 03:34 |
| Breath Volume      | 1.835 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Subject Sample # 2 | 0.173 | 03:37 |
| Breath Volume      | 1.941 | 03:37 |
| Air Blank          | 0.000 | 03:38 |
| Control Test       | 0.080 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Diagnostics Check  | OK    | 03:39 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop C  
ARRESTING OFFICER: CARRASCO,A  
ARRESTING OFFICER AGENCY: FHP Troop C  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/24/2024 22:18:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:11 |
| Air Blank          | 0.000 | 22:11 |
| Control Test       | 0.081 | 22:11 |
| Air Blank          | 0.000 | 22:12 |
| Subject Sample # 1 | 0.134 | 22:13 |
| Breath Volume      | 1.867 | 22:13 |
| Air Blank          | 0.000 | 22:14 |
| Air Blank          | 0.000 | 22:15 |
| Subject Sample # 2 | 0.130 | 22:16 |
| Breath Volume      | 1.687 | 22:16 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.080 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Diagnostics Check  | OK    | 22:18 |

---

Data Download Date/Time: 12/12/2024 13:40

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GANT, TIFFANY,  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: CAMPER

DATE/TIME OF TEST: 11/24/2024 22:25:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:21 |                      |
| Air Blank         | 0.000 | 22:22 |                      |
| Control Test      | 0.080 | 22:22 |                      |
| Air Blank         | 0.000 | 22:23 |                      |
| Subject Sample    | # 1   | 22:23 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:23 |                      |
| Air Blank         | 0.000 | 22:24 |                      |
| Control Test      | 0.079 | 22:24 |                      |
| Air Blank         | 0.000 | 22:25 |                      |
| Diagnostics Check | OK    | 22:25 |                      |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO, M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MCCLELLAN JR

DATE/TIME OF TEST: 11/25/2024 04:06:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:59 |
| Air Blank          | 0.000 | 04:00 |
| Control Test       | 0.081 | 04:00 |
| Air Blank          | 0.000 | 04:01 |
| Subject Sample # 1 | 0.132 | 04:02 |
| Breath Volume      | 2.480 | 04:02 |
| Air Blank          | 0.000 | 04:02 |
| Air Blank          | 0.000 | 04:04 |
| Subject Sample # 2 | 0.132 | 04:04 |
| Breath Volume      | 2.066 | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.080 | 04:05 |
| Air Blank          | 0.000 | 04:06 |
| Diagnostics Check  | OK    | 04:06 |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: HERNANDEZ, S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 11/28/2024 22:08:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:01 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.080 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.275 | 22:03 |
| Breath Volume      | 1.460 | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.282 | 22:06 |
| Breath Volume      | 1.972 | 22:06 |
| Air Blank          | 0.000 | 22:07 |
| Control Test       | 0.078 | 22:07 |
| Air Blank          | 0.000 | 22:08 |
| Diagnostics Check  | OK    | 22:08 |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: MILILLO,M  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 03:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 11/29/2024 04:16:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:08 |
| Air Blank          | 0.000 | 04:08 |
| Control Test       | 0.080 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 1 | 0.120 | 04:11 |
| Breath Volume      | 1.433 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 2 | 0.116 | 04:14 |
| Breath Volume      | 2.683 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.078 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Diagnostics Check  | OK    | 04:16 |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: WILLIAMS, P  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 11/29/2024 23:26:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.082 | 23:20 |
| Air Blank          | 0.000 | 23:21 |
| Subject Sample # 1 | 0.263 | 23:21 |
| Breath Volume      | 2.722 | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Air Blank          | 0.000 | 23:24 |
| Subject Sample # 2 | 0.272 | 23:24 |
| Breath Volume      | 1.949 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Control Test       | 0.079 | 23:25 |
| Air Blank          | 0.000 | 23:26 |
| Diagnostics Check  | OK    | 23:26 |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIERCEY,JANICE,F  
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO  
ARRESTING OFFICER: PORTER,S  
ARRESTING OFFICER AGENCY: Hillsborough Co SO  
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28423080A2  
DRY GAS EXPIRY DATE: 11/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: FIELDS JR

DATE/TIME OF TEST: 11/30/2024 02:24:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:16 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.081 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Subject Sample # 1 | 0.124 | 02:18 |
| Breath Volume      | 1.425 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.125 | 02:22 |
| Breath Volume      | 1.453 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.080 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 12/12/2024 13:41

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLO,DEREK,J  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: FARRAR,KENNA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GOMEZ MORENO

DATE/TIME OF TEST: 11/01/2024 01:49:52

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:49 |            |
| Air Blank         |       | 01:49 | RFI Detect |
| Air Blank         | 0.000 | 01:49 |            |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLO,DEREK,J  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: FARRAR,KENNA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GOMEZ MORENO

DATE/TIME OF TEST: 11/01/2024 02:01:53

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:52 |                |
| Air Blank         | 0.000     | 01:53 |                |
| Control Test      | 0.079     | 01:53 |                |
| Air Blank         | 0.000     | 01:54 |                |
| Subject Sample    | # 1 0.157 | 01:57 | Volume Not Met |
| Breath Volume     | 0.316     | 01:57 |                |
| Air Blank         | 0.000     | 01:58 |                |
| Air Blank         | 0.000     | 01:59 |                |
| Subject Sample    | # 2 0.179 | 02:00 |                |
| Breath Volume     | 2.378     | 02:00 |                |
| Air Blank         |           | 02:00 | RFI Detect     |
| Air Blank         | 0.000     | 02:00 |                |
| Control Test      | 0.078     | 02:01 |                |
| Air Blank         | 0.000     | 02:01 |                |
| Diagnostics Check | OK        | 02:01 |                |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLO,DEREK,J  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: FARRAR,KENNA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GOMEZ MORENO

DATE/TIME OF TEST: 11/01/2024 02:12:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:05 |
| Air Blank          | 0.000 | 02:06 |
| Control Test       | 0.078 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Subject Sample # 1 | 0.176 | 02:08 |
| Breath Volume      | 1.500 | 02:08 |
| Air Blank          | 0.000 | 02:08 |
| Air Blank          | 0.000 | 02:10 |
| Subject Sample # 2 | 0.166 | 02:11 |
| Breath Volume      | 1.921 | 02:11 |
| Air Blank          | 0.000 | 02:11 |
| Control Test       | 0.079 | 02:12 |
| Air Blank          | 0.000 | 02:12 |
| Diagnostics Check  | OK    | 02:12 |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LATIMER,J,A  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: LATIMER,J  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: FENELLI

DATE/TIME OF TEST: 11/01/2024 22:57:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:55 |                      |
| Air Blank         | 0.000 | 22:55 |                      |
| Control Test      | 0.079 | 22:55 |                      |
| Air Blank         | 0.000 | 22:56 |                      |
| Subject Sample    | # 1   | 22:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:56 |                      |
| Air Blank         | 0.000 | 22:57 |                      |
| Control Test      | 0.079 | 22:57 |                      |
| Air Blank         | 0.000 | 22:57 |                      |
| Diagnostics Check | OK    | 22:57 |                      |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RILEY,BRANDON,J  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: RILEY,BRANDON  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 11/03/2024 06:33:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:27 |
| Air Blank          | 0.000 | 06:27 |
| Control Test       | 0.080 | 06:28 |
| Air Blank          | 0.000 | 06:28 |
| Subject Sample # 1 | 0.129 | 06:29 |
| Breath Volume      | 1.460 | 06:29 |
| Air Blank          | 0.000 | 06:29 |
| Air Blank          | 0.000 | 06:31 |
| Subject Sample # 2 | 0.135 | 06:32 |
| Breath Volume      | 2.121 | 06:32 |
| Air Blank          | 0.000 | 06:32 |
| Control Test       | 0.079 | 06:33 |
| Air Blank          | 0.000 | 06:33 |
| Diagnostics Check  | OK    | 06:33 |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: ECKERDT,ANDRU  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CORDOVA ALFARO

DATE/TIME OF TEST: 11/04/2024 00:03:34

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:00 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Control Test      | 0.079 | 00:01 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Subject Sample    | # 1   | 00:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:02 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Control Test      | 0.079 | 00:02 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Diagnostics Check | OK    | 00:03 |                      |

---

Data Download Date/Time: 11/19/2024 10:42

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LAJEUNESSE,J,  
BREATH TEST OPERATOR AGENCY: Astatula PD  
ARRESTING OFFICER: POSTMA,M  
ARRESTING OFFICER AGENCY: Tavares PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 06:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: BEADLE

DATE/TIME OF TEST: 11/04/2024 07:41:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:38 |                      |
| Air Blank         | 0.000 | 07:38 |                      |
| Control Test      | 0.079 | 07:39 |                      |
| Air Blank         | 0.000 | 07:39 |                      |
| Subject Sample    | # 1   | 07:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:39 |                      |
| Air Blank         | 0.000 | 07:40 |                      |
| Control Test      | 0.079 | 07:40 |                      |
| Air Blank         | 0.000 | 07:41 |                      |
| Diagnostics Check | OK    | 07:41 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANNICK,COURTNEY,  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: VAZQUEZ,X  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: DOR

DATE/TIME OF TEST: 11/06/2024 23:01:13

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 22:58 |                      |
| Air Blank         | 0.000 | 22:58 |                      |
| Control Test      | 0.080 | 22:58 |                      |
| Air Blank         | 0.000 | 22:59 |                      |
| Subject Sample    | # 1   | 22:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 22:59 |                      |
| Air Blank         | 0.000 | 23:00 |                      |
| Control Test      | 0.079 | 23:00 |                      |
| Air Blank         | 0.000 | 23:01 |                      |
| Diagnostics Check | OK    | 23:01 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENTON,AMY,B  
BREATH TEST OPERATOR AGENCY: Astatula PD  
ARRESTING OFFICER: MACIAS,JANELLY  
ARRESTING OFFICER AGENCY: Astatula PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: DENNIS III

DATE/TIME OF TEST: 11/09/2024 20:00:41

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:57 |                      |
| Air Blank         | 0.000 | 19:58 |                      |
| Control Test      | 0.079 | 19:58 |                      |
| Air Blank         | 0.000 | 19:59 |                      |
| Subject Sample    | # 1   | 19:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:59 |                      |
| Air Blank         | 0.000 | 19:59 |                      |
| Control Test      | 0.079 | 20:00 |                      |
| Air Blank         | 0.000 | 20:00 |                      |
| Diagnostics Check | OK    | 20:00 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: MORRISON,K  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: CASTRO QUIROS

DATE/TIME OF TEST: 11/10/2024 00:09:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:03 |
| Air Blank          | 0.000 | 00:03 |
| Control Test       | 0.079 | 00:04 |
| Air Blank          | 0.000 | 00:04 |
| Subject Sample # 1 | 0.147 | 00:05 |
| Breath Volume      | 1.566 | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Air Blank          | 0.000 | 00:07 |
| Subject Sample # 2 | 0.144 | 00:08 |
| Breath Volume      | 1.992 | 00:08 |
| Air Blank          | 0.000 | 00:08 |
| Control Test       | 0.078 | 00:09 |
| Air Blank          | 0.000 | 00:09 |
| Diagnostics Check  | OK    | 00:09 |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: NEWBOLD,DARRIES  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/11/2024 01:42:40

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:26 |                |
| Air Blank         | 0.000     | 01:27 |                |
| Control Test      | 0.080     | 01:27 |                |
| Air Blank         | 0.000     | 01:28 |                |
| Subject Sample    | # 1 0.175 | 01:31 |                |
| Breath Volume     | 1.472     | 01:31 |                |
| Air Blank         | 0.000     | 01:32 |                |
| Air Blank         | 0.000     | 01:33 |                |
| Subject Sample    | # 2 0.141 | 01:36 | Volume Not Met |
| Breath Volume     | 0.000     | 01:36 |                |
| Air Blank         | 0.000     | 01:37 |                |
| Air Blank         | 0.000     | 01:39 |                |
| Subject Sample    | # 3 0.162 | 01:41 |                |
| Breath Volume     | 1.605     | 01:41 |                |
| Air Blank         | 0.000     | 01:41 |                |
| Control Test      | 0.078     | 01:42 |                |
| Air Blank         | 0.000     | 01:42 |                |
| Diagnostics Check | OK        | 01:42 |                |

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANNICL,COURTNEY,  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: HUMMEL,D  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 18:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: FRESON

DATE/TIME OF TEST: 11/11/2024 19:22:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:19 |                      |
| Air Blank         | 0.000 | 19:19 |                      |
| Control Test      | 0.080 | 19:20 |                      |
| Air Blank         | 0.000 | 19:20 |                      |
| Subject Sample    | # 1   | 19:21 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:21 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Control Test      | 0.079 | 19:22 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Diagnostics Check | OK    | 19:22 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BANNICK,COURTNEY,  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: DOVE,W  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: COYLE

DATE/TIME OF TEST: 11/11/2024 22:23:51

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 22:12 |                    |
| Air Blank         | 0.000 | 22:12 |                    |
| Control Test      | 0.080 | 22:13 |                    |
| Air Blank         | 0.000 | 22:13 |                    |
| Subject Sample    | # 1   | 22:16 | No Sample Provided |
| Breath Volume     | 0.000 | 22:16 |                    |
| Air Blank         | 0.000 | 22:17 |                    |
| Air Blank         | 0.000 | 22:19 |                    |
| Subject Sample    | # 2   | 22:22 | No Sample Provided |
| Breath Volume     | 0.000 | 22:22 |                    |
| Air Blank         | 0.000 | 22:22 |                    |
| Control Test      | 0.077 | 22:23 |                    |
| Air Blank         | 0.000 | 22:23 |                    |
| Diagnostics Check | OK    | 22:23 |                    |

---

Data Download Date/Time: 11/19/2024 10:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JOHNSON,D,  
BREATH TEST OPERATOR AGENCY: Groveland PD  
ARRESTING OFFICER: JOHNSON,D  
ARRESTING OFFICER AGENCY: Groveland PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 11/12/2024 20:38:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:28 |
| Air Blank          | 0.000 | 20:28 |
| Control Test       | 0.079 | 20:28 |
| Air Blank          | 0.000 | 20:29 |
| Subject Sample # 1 | 0.208 | 20:30 |
| Breath Volume      | 2.945 | 20:30 |
| Air Blank          | 0.000 | 20:30 |
| Air Blank          | 0.000 | 20:32 |
| Subject Sample # 2 | 0.240 | 20:33 |
| Breath Volume      | 3.582 | 20:33 |
| Air Blank          | 0.000 | 20:34 |
| Air Blank          | 0.000 | 20:36 |
| Subject Sample # 3 | 0.219 | 20:36 |
| Breath Volume      | 3.292 | 20:36 |
| Air Blank          | 0.000 | 20:37 |
| Control Test       | 0.079 | 20:37 |
| Air Blank          | 0.000 | 20:38 |
| Diagnostics Check  | OK    | 20:38 |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOLCHARRAN,DILON,C  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: BERRY,SHAYNE  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STARNES

DATE/TIME OF TEST: 11/14/2024 02:44:09

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:41 |                      |
| Air Blank         | 0.000 | 02:41 |                      |
| Control Test      | 0.080 | 02:42 |                      |
| Air Blank         | 0.000 | 02:42 |                      |
| Subject Sample    | # 1   | 02:42 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:42 |                      |
| Air Blank         | 0.000 | 02:43 |                      |
| Control Test      | 0.079 | 02:43 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Diagnostics Check | OK    | 02:44 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,STEPHEN,V  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: FABERLLE,ALEX  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 19:05  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/15/2024 20:13:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Control Test       | 0.080 | 20:07 |
| Air Blank          | 0.000 | 20:08 |
| Subject Sample # 1 | 0.146 | 20:08 |
| Breath Volume      | 2.175 | 20:08 |
| Air Blank          | 0.000 | 20:09 |
| Air Blank          | 0.000 | 20:11 |
| Subject Sample # 2 | 0.144 | 20:11 |
| Breath Volume      | 2.097 | 20:11 |
| Air Blank          | 0.000 | 20:12 |
| Control Test       | 0.079 | 20:12 |
| Air Blank          | 0.000 | 20:13 |
| Diagnostics Check  | OK    | 20:13 |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUSHONG,NICOLE,R  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: DESOTO,J  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/16/2024 03:18:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:15 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.081 | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Subject Sample    | # 1   | 03:17 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:17 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.080 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Diagnostics Check | OK    | 03:18 |                      |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,STEPHEN,V  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: MCDANIEL,ABBIE  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 11/16/2024 22:17:48

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 22:11 |            |
| Air Blank         | 0.000     | 22:11 |            |
| Control Test      | 0.079     | 22:12 |            |
| Air Blank         | 0.000     | 22:12 |            |
| Subject Sample    | # 1 0.129 | 22:13 |            |
| Breath Volume     | 2.710     | 22:13 |            |
| Air Blank         | 0.000     | 22:14 |            |
| Air Blank         | 0.000     | 22:15 |            |
| Subject Sample    | # 2 0.122 | 22:16 |            |
| Breath Volume     | 1.339     | 22:16 |            |
| Air Blank         | 0.000     | 22:17 |            |
|                   |           | 22:17 | RFI Detect |
| Air Blank         | 0.000     | 22:17 |            |

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,STEPHEN,V  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: DURYE, MATTHEW  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 221303  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GUZYNSKI

DATE/TIME OF TEST: 11/17/2024 23:58:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:51 |
| Air Blank          | 0.000 | 23:51 |
| Control Test       | 0.079 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.123 | 23:53 |
| Breath Volume      | 5.570 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 2 | 0.122 | 23:56 |
| Breath Volume      | 4.988 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Control Test       | 0.079 | 23:57 |
| Air Blank          | 0.000 | 23:57 |
| Diagnostics Check  | OK    | 23:58 |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,DAVID,R  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: BURNS,J  
ARRESTING OFFICER AGENCY: Lady Lake PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 10:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: TROSO

DATE/TIME OF TEST: 11/18/2024 12:04:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:58 |
| Air Blank          | 0.000 | 11:59 |
| Control Test       | 0.079 | 11:59 |
| Air Blank          | 0.000 | 12:00 |
| Subject Sample # 1 | 0.000 | 12:00 |
| Breath Volume      | 2.625 | 12:00 |
| Air Blank          | 0.000 | 12:01 |
| Air Blank          | 0.000 | 12:03 |
| Subject Sample # 2 | 0.000 | 12:03 |
| Breath Volume      | 2.515 | 12:03 |
| Air Blank          | 0.000 | 12:03 |
| Control Test       | 0.079 | 12:04 |
| Air Blank          | 0.000 | 12:04 |
| Diagnostics Check  | OK    | 12:04 |

---

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: FRANCIS,K  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 10/03/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 11/19/2024 02:37:56

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:35 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Control Test      | 0.080 | 02:35 |                      |
| Air Blank         | 0.000 | 02:36 |                      |
| Subject Sample    | # 1   | 02:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:36 |                      |
| Air Blank         | 0.000 | 02:37 |                      |
| Control Test      | 0.080 | 02:37 |                      |
| Air Blank         | 0.000 | 02:37 |                      |
| Diagnostics Check | OK    | 02:37 |                      |

Data Download Date/Time: 11/19/2024 10:43

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REBANDO,BRIAN,M  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: BENTIL,GERALD  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: LUIS HERNANDEZ

DATE/TIME OF TEST: 11/20/2024 03:50:08

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:37 |                |
| Air Blank         | 0.000     | 03:37 |                |
| Control Test      | 0.079     | 03:37 |                |
| Air Blank         | 0.000     | 03:38 |                |
| Subject Sample    | # 1 0.145 | 03:38 |                |
| Breath Volume     | 1.847     | 03:38 |                |
| Air Blank         | 0.000     | 03:39 |                |
| Air Blank         | 0.000     | 03:41 |                |
| Subject Sample    | # 2 0.117 | 03:44 | Volume Not Met |
| Breath Volume     | 1.117     | 03:44 |                |
| Air Blank         | 0.000     | 03:45 |                |
| Air Blank         | 0.000     | 03:47 |                |
| Subject Sample    | # 3 0.143 | 03:48 |                |
| Breath Volume     | 2.335     | 03:48 |                |
| Air Blank         | 0.000     | 03:49 |                |
| Control Test      | 0.075     | 03:49 |                |
| Air Blank         | 0.000     | 03:50 |                |
| Diagnostics Check | OK        | 03:50 |                |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,DAVID,R  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: POLO,DEREK  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: LING

DATE/TIME OF TEST: 11/22/2024 23:17:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.080 | 23:11 |
| Air Blank          | 0.000 | 23:11 |
| Subject Sample # 1 | 0.157 | 23:13 |
| Breath Volume      | 2.074 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 2 | 0.172 | 23:15 |
| Breath Volume      | 1.957 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Control Test       | 0.081 | 23:16 |
| Air Blank          | 0.000 | 23:17 |
| Diagnostics Check  | OK    | 23:17 |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLO,DEREK,J  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: SANDBAKKEN,PAUL  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 11/23/2024 01:40:18

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:38 |            |
| Air Blank         | 0.000 | 01:38 |            |
| Control Test      | 0.080 | 01:38 |            |
| Air Blank         | 0.000 | 01:39 |            |
| Subject Sample    | # 1   | 01:39 | RFI Detect |
| Breath Volume     | 0.000 | 01:39 |            |
| Air Blank         |       | 01:39 | RFI Detect |
| Air Blank         | 0.000 | 01:40 |            |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POLO,DEREK,J  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: SANDBAKKEN,PAUL  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 11/23/2024 01:50:17

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 01:43 |            |
| Air Blank         | 0.000     | 01:44 |            |
| Control Test      | 0.080     | 01:44 |            |
| Air Blank         | 0.000     | 01:45 |            |
| Subject Sample    | # 1 0.000 | 01:45 |            |
| Breath Volume     | 1.867     | 01:45 |            |
| Air Blank         | 0.000     | 01:46 |            |
| Air Blank         | 0.000     | 01:48 |            |
| Subject Sample    | # 2 0.000 | 01:48 |            |
| Breath Volume     | 2.656     | 01:48 |            |
| Air Blank         | 0.000     | 01:49 |            |
| Control Test      | 0.080     | 01:49 |            |
| Air Blank         |           | 01:49 | RFI Detect |
| Air Blank         | 0.000     | 01:50 |            |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: DANGELILLO,A  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GUTIERREZ DIAZ

DATE/TIME OF TEST: 11/24/2024 00:37:11

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:34 |                      |
| Air Blank         | 0.000 | 00:34 |                      |
| Control Test      | 0.080 | 00:35 |                      |
| Air Blank         | 0.000 | 00:35 |                      |
| Subject Sample    | # 1   | 00:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:35 |                      |
| Air Blank         | 0.000 | 00:36 |                      |
| Control Test      | 0.079 | 00:36 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Diagnostics Check | OK    | 00:37 |                      |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: DANGELILLO,A  
ARRESTING OFFICER AGENCY: Mount Dora PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: GUTIERREZ DIAZ

DATE/TIME OF TEST: 11/24/2024 01:06:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.080 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 1 | 0.068 | 01:02 |
| Breath Volume      | 1.367 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 2 | 0.064 | 01:05 |
| Breath Volume      | 1.128 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.078 | 01:06 |
| Air Blank          | 0.000 | 01:06 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WILDEMAN-STUART,NATASH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DEINE

DATE/TIME OF TEST: 11/27/2024 00:59:53

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Control Test      | 0.080 | 00:57 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Subject Sample    | # 1   | 00:58 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:58 |                      |
| Air Blank         | 0.000 | 00:58 |                      |
| Control Test      | 0.079 | 00:59 |                      |
| Air Blank         | 0.000 | 00:59 |                      |
| Diagnostics Check | OK    | 00:59 |                      |

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ECKERDT,ANDRU,J  
BREATH TEST OPERATOR AGENCY: Lady Lake PD  
ARRESTING OFFICER: ALLEN,BREANNA  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 23:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 11/28/2024 00:44:40

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 00:36 |                           |
| Air Blank         | 0.000     | 00:37 |                           |
| Control Test      | 0.080     | 00:37 |                           |
| Air Blank         | 0.000     | 00:37 |                           |
| Subject Sample    | # 1 0.190 | 00:38 |                           |
| Breath Volume     | 1.355     | 00:38 |                           |
| Air Blank         | 0.000     | 00:39 |                           |
| Air Blank         | 0.000     | 00:41 |                           |
| Subject Sample    | # 2 0.188 | 00:43 |                           |
| Breath Volume     | 1.136     | 00:43 |                           |
| Air Blank         | 0.000     | 00:43 |                           |
| Control Test      | 0.068     | 00:44 | Control Outside Tolerance |
| Air Blank         | 0.000     | 00:44 |                           |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POST,HAYDEN,L  
BREATH TEST OPERATOR AGENCY: Leesburg PD  
ARRESTING OFFICER: ABUSERYDZE,DAVID  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: DE LUNA GUZMAN

DATE/TIME OF TEST: 11/28/2024 23:13:36

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 23:10 |            |
| Air Blank         | 0.000     | 23:10 |            |
| Control Test      | 0.079     | 23:10 |            |
| Air Blank         | 0.000     | 23:11 |            |
| Subject Sample    | # 1 0.281 | 23:12 |            |
| Breath Volume     | 1.140     | 23:12 |            |
| Air Blank         |           | 23:13 | Purge Fail |
| Air Blank         | 0.000     | 23:13 |            |

---

Data Download Date/Time: 12/17/2024 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WILDEMAN-STUART,NATASH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 01:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SHAIN

DATE/TIME OF TEST: 11/30/2024 02:31:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.080 | 02:22 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 1 | 0.214 | 02:23 |
| Breath Volume      | 2.679 | 02:23 |
| Air Blank          | 0.000 | 02:23 |
| Air Blank          | 0.000 | 02:25 |
| Subject Sample # 2 | 0.177 | 02:26 |
| Breath Volume      | 1.726 | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Air Blank          | 0.000 | 02:28 |
| Subject Sample # 3 | 0.208 | 02:29 |
| Breath Volume      | 3.964 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.081 | 02:30 |
| Air Blank          | 0.000 | 02:30 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 12/17/2024 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARKER,STEPHEN,V  
BREATH TEST OPERATOR AGENCY: Eustis PD  
ARRESTING OFFICER: PAUNA,MITCHELL  
ARRESTING OFFICER AGENCY: Eustis PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 04:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 11/30/2024 05:13:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:04 |
| Air Blank          | 0.000 | 05:04 |
| Control Test       | 0.080 | 05:05 |
| Air Blank          | 0.000 | 05:05 |
| Subject Sample # 1 | 0.156 | 05:08 |
| Breath Volume      | 1.644 | 05:08 |
| Air Blank          | 0.000 | 05:09 |
| Air Blank          | 0.000 | 05:10 |
| Subject Sample # 2 | 0.158 | 05:11 |
| Breath Volume      | 1.914 | 05:11 |
| Air Blank          | 0.000 | 05:12 |
| Control Test       | 0.078 | 05:12 |
| Air Blank          | 0.000 | 05:13 |
| Diagnostics Check  | OK    | 05:13 |

---

Data Download Date/Time: 12/17/2024 10:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WILDEMAN-STUART,NATASH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 11/30/2024 22:59:26

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 22:53 |                |
| Air Blank         | 0.000     | 22:54 |                |
| Control Test      | 0.080     | 22:54 |                |
| Air Blank         | 0.000     | 22:55 |                |
| Subject Sample    | # 1 0.111 | 22:58 | Volume Not Met |
| Breath Volume     | 0.816     | 22:58 |                |
| Air Blank         |           | 22:58 | Purge Fail     |
| Air Blank         | 0.000     | 22:59 |                |

Data Download Date/Time: 12/17/2024 10:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: WILDEMAN-STUART,NATASH  
ARRESTING OFFICER AGENCY: Fruitland Park PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 21:58  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 11/30/2024 23:11:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:01 |
| Air Blank          | 0.000 | 23:02 |
| Control Test       | 0.079 | 23:02 |
| Air Blank          | 0.000 | 23:03 |
| Subject Sample # 1 | 0.120 | 23:06 |
| Breath Volume      | 1.390 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Air Blank          | 0.000 | 23:08 |
| Subject Sample # 2 | 0.118 | 23:09 |
| Breath Volume      | 1.250 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.079 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Diagnostics Check  | OK    | 23:11 |

---

Data Download Date/Time: 12/17/2024 10:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L  
BREATH TEST OPERATOR AGENCY: Fruitland Park PD  
ARRESTING OFFICER: HILLIKER,COREY  
ARRESTING OFFICER AGENCY: Leesburg PD  
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 33522080A5  
DRY GAS EXPIRY DATE: 01/05/2025  
VIOLATION CODE: DUI  
ARREST TIME: 22:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: DRAUGHN

DATE/TIME OF TEST: 11/30/2024 23:34:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:28 |
| Air Blank          | 0.000 | 23:29 |
| Control Test       | 0.078 | 23:29 |
| Air Blank          | 0.000 | 23:30 |
| Subject Sample # 1 | 0.157 | 23:30 |
| Breath Volume      | 2.136 | 23:30 |
| Air Blank          | 0.000 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Subject Sample # 2 | 0.155 | 23:33 |
| Breath Volume      | 2.277 | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Control Test       | 0.080 | 23:34 |
| Air Blank          | 0.000 | 23:34 |
| Diagnostics Check  | OK    | 23:34 |

---

Data Download Date/Time: 12/17/2024 10:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: PULLIAM

DATE/TIME OF TEST: 11/01/2024 21:19:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:12 |
| Air Blank          | 0.000 | 21:13 |
| Control Test       | 0.081 | 21:13 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 1 | 0.000 | 21:14 |
| Breath Volume      | 1.316 | 21:14 |
| Air Blank          | 0.000 | 21:14 |
| Air Blank          | 0.000 | 21:16 |
| Subject Sample # 2 | 0.000 | 21:17 |
| Breath Volume      | 1.753 | 21:17 |
| Air Blank          | 0.000 | 21:18 |
| Control Test       | 0.081 | 21:18 |
| Air Blank          | 0.000 | 21:19 |
| Diagnostics Check  | OK    | 21:19 |

---

Data Download Date/Time: 11/15/2024 17:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 03:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 11/02/2024 03:50:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:44 |
| Air Blank          | 0.000 | 03:45 |
| Control Test       | 0.080 | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 1 | 0.000 | 03:46 |
| Breath Volume      | 2.937 | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Air Blank          | 0.000 | 03:48 |
| Subject Sample # 2 | 0.000 | 03:49 |
| Breath Volume      | 3.921 | 03:49 |
| Air Blank          | 0.000 | 03:49 |
| Control Test       | 0.080 | 03:50 |
| Air Blank          | 0.000 | 03:50 |
| Diagnostics Check  | OK    | 03:50 |

---

Data Download Date/Time: 11/15/2024 17:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOSS,ROBERT,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOSS,ROBERT  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 402806079  
DRY GAS EXPIRY DATE: 07/26/2026  
VIOLATION CODE: DUI  
ARREST TIME: 17:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: DAIGLE

DATE/TIME OF TEST: 11/11/2024 18:02:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:55 |
| Air Blank          | 0.000 | 17:55 |
| Control Test       | 0.080 | 17:56 |
| Air Blank          | 0.000 | 17:56 |
| Subject Sample # 1 | 0.206 | 17:57 |
| Breath Volume      | 2.003 | 17:57 |
| Air Blank          | 0.000 | 17:57 |
| Air Blank          | 0.000 | 17:59 |
| Subject Sample # 2 | 0.206 | 18:00 |
| Breath Volume      | 1.804 | 18:00 |
| Air Blank          | 0.000 | 18:01 |
| Control Test       | 0.081 | 18:01 |
| Air Blank          | 0.000 | 18:02 |
| Diagnostics Check  | OK    | 18:02 |

---

Data Download Date/Time: 11/15/2024 17:07