

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: BAYLIS,B
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 10/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/08/2024 21:25:23

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.081	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.081	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 11/26/2024 20:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: ZAYAS ID53,APRIL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: MERIDA

DATE/TIME OF TEST: 11/29/2024 17:27:38

Diagnostics Check	OK	17:20
Air Blank	0.000	17:21
Control Test	0.082	17:21
Air Blank	0.000	17:22
Subject Sample # 1	0.232	17:23
Breath Volume	2.355	17:23
Air Blank	0.000	17:23
Air Blank	0.000	17:25
Subject Sample # 2	0.232	17:26
Breath Volume	1.996	17:26
Air Blank	0.000	17:26
Control Test	0.080	17:27
Air Blank	0.000	17:27
Diagnostics Check	OK	17:27

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CARROLL,STEVE
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 11/04/2024 04:27:16

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.078	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1 0.150	04:21	Volume Not Met
Breath Volume	0.128	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 2	04:25	Subject Test Refused
Breath Volume	0.000	04:25	
Air Blank	0.000	04:26	
Control Test	0.076	04:26	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 11/22/2024 03:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MCCRANN,SEAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: LOTT

DATE/TIME OF TEST: 11/09/2024 02:48:21

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:43
Breath Volume	1.261	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.000	02:46
Breath Volume	1.781	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 11/22/2024 03:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAL, ADAM, M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BEAL, ADAM
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: ONTIVEROS

DATE/TIME OF TEST: 11/09/2024 08:24:37

Diagnostics Check	OK	08:17
Air Blank	0.000	08:18
Control Test	0.078	08:18
Air Blank	0.000	08:18
Subject Sample # 1	0.144	08:19
Breath Volume	2.046	08:19
Air Blank	0.000	08:20
Air Blank	0.000	08:21
Subject Sample # 2	0.132	08:23
Breath Volume	2.082	08:23
Air Blank	0.000	08:23
Control Test	0.076	08:24
Air Blank	0.000	08:24
Diagnostics Check	OK	08:24

Data Download Date/Time: 11/22/2024 03:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: LOBB,M
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 11/17/2024 02:39:20

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.000	02:34
Breath Volume	1.558	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.000	02:37
Breath Volume	1.875	02:37
Air Blank	0.000	02:38
Control Test	0.077	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 11/22/2024 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CAPLE,CAMERON
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/17/2024 11:21:44

Diagnostics Check	OK	11:18	
Air Blank	0.000	11:18	
Control Test	0.078	11:18	
Air Blank	0.000	11:19	
Subject Sample	# 1	11:20	Interferent Detect
Breath Volume	1.808	11:20	
Air Blank	0.000	11:20	
Control Test	0.079	11:21	
Air Blank	0.000	11:21	
Diagnostics Check	OK	11:21	

Data Download Date/Time: 11/22/2024 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CAPLE,CAMERON
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/17/2024 11:29:09

Diagnostics Check	OK	11:25	
Air Blank	0.000	11:26	
Control Test	0.078	11:26	
Air Blank	0.000	11:27	
Subject Sample	# 1	11:27	Interferent Detect
Breath Volume	1.449	11:27	
Air Blank	0.000	11:28	
Control Test	0.078	11:28	
Air Blank	0.000	11:29	
Diagnostics Check	OK	11:29	

Data Download Date/Time: 11/22/2024 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANN,SEAN,M
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CAPLE,CAMERON
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 11/17/2024 12:08:17

Diagnostics Check	OK	12:05	
Air Blank	0.000	12:05	
Control Test	0.077	12:05	
Air Blank	0.000	12:06	
Subject Sample	# 1	12:06	Interferent Detect
Breath Volume	1.582	12:06	
Air Blank	0.000	12:07	
Control Test	0.078	12:07	
Air Blank	0.000	12:08	
Diagnostics Check	OK	12:08	

Data Download Date/Time: 11/22/2024 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: WHITE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:27

SUBJECT'S LAST NAME: BILBREY

DATE/TIME OF TEST: 11/19/2024 13:55:18

Diagnostics Check	OK	13:47
Air Blank	0.000	13:47
Control Test	0.078	13:48
Air Blank	0.000	13:48
Subject Sample # 1	0.000	13:49
Breath Volume	1.992	13:49
Air Blank	0.000	13:49
Air Blank	0.000	13:51
Subject Sample # 2	0.000	13:53
Breath Volume	1.496	13:53
Air Blank	0.000	13:54
Control Test	0.078	13:54
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 11/22/2024 03:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 11/02/2024 02:45:37

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:40
Breath Volume	1.449	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:44
Breath Volume	1.769	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/30/2024 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ESPOSITA

DATE/TIME OF TEST: 11/02/2024 03:05:38

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 11/13/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: EDENFIELD

DATE/TIME OF TEST: 11/12/2024 22:22:48

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:20	
Control Test	0.078	22:20	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.078	22:22	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 11/13/2024 10:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: LOGSDON

DATE/TIME OF TEST: 11/16/2024 04:47:54

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.084	04:43
Breath Volume	1.972	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.082	04:46
Breath Volume	1.464	04:46
Air Blank	0.000	04:46
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 12/11/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: LOMINACK JR

DATE/TIME OF TEST: 11/20/2024 22:27:37

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.078	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.199	22:23
Breath Volume	1.210	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.206	22:26
Breath Volume	1.250	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 12/11/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HOFF

DATE/TIME OF TEST: 11/30/2024 01:59:32

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.153	01:54
Breath Volume	1.671	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.157	01:58
Breath Volume	2.140	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 12/11/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: ROMANO JR, ANTHONY
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 11/08/2024 22:15:05

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	3.058	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	3.027	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 11/26/2024 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWAB,JOSEPH,S
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LARSON,BEN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/16/2024 03:05:52

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.163	03:01
Breath Volume	1.261	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.171	03:04
Breath Volume	1.382	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 11/26/2024 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,J,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: UMMEL,J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/17/2024 01:53:29

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.216	01:48
Breath Volume	1.257	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.206	01:51
Breath Volume	1.300	01:51
Air Blank	0.000	01:52
Control Test	0.078	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 11/26/2024 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: ROMANO, ANTHONY
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: JULIAN

DATE/TIME OF TEST: 11/19/2024 05:27:11

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.049	05:22
Breath Volume	2.968	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.048	05:25
Breath Volume	3.156	05:25
Air Blank	0.000	05:26
Control Test	0.079	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 11/26/2024 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEBER,DYLAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/23/2024 04:35:48

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.123	04:31
Breath Volume	2.156	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.129	04:34
Breath Volume	3.546	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 11/26/2024 20:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: QUILES,S
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ORSZAG

DATE/TIME OF TEST: 11/03/2024 00:43:44

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.082	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.111	00:39
Breath Volume	1.238	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.107	00:42
Breath Volume	1.441	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HENRY,A
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 11/03/2024 02:56:07

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:51
Breath Volume	1.320	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	1.625	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: GUZMAN,DAVID
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: MESIAS

DATE/TIME OF TEST: 11/03/2024 06:50:43

Diagnostics Check	OK	06:36	
Air Blank	0.000	06:36	
Control Test	0.081	06:36	
Air Blank	0.000	06:37	
Subject Sample	# 1 0.104	06:40	Volume Not Met
Breath Volume	0.335	06:40	
Air Blank	0.000	06:41	
Air Blank	0.000	06:43	
Subject Sample	# 2 0.129	06:44	
Breath Volume	1.250	06:44	
Air Blank	0.000	06:45	
Air Blank	0.000	06:47	
Subject Sample	# 3 0.125	06:49	
Breath Volume	2.003	06:49	
Air Blank	0.000	06:49	
Control Test	0.077	06:50	
Air Blank	0.000	06:50	
Diagnostics Check	OK	06:50	

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 11/03/2024 23:29:27

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:21	
Control Test	0.081	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1 0.321	23:23	
Breath Volume	2.890	23:23	
Air Blank	0.000	23:24	
Air Blank	0.000	23:25	
Subject Sample	# 2 0.356	23:26	
Breath Volume	2.628	23:26	
Air Blank	0.000	23:27	
Air Blank		23:28	Ambient Fail
Air Blank	0.000	23:29	

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 11/03/2024 23:41:16

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.081	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1 0.298	23:33	
Breath Volume	2.917	23:33	
Air Blank	0.000	23:34	
Air Blank	0.000	23:36	
Subject Sample	# 2 0.362	23:36	
Breath Volume	2.929	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 3 0.353	23:39	
Breath Volume	1.582	23:39	
Air Blank	0.000	23:40	
Control Test	0.065	23:40	Control Outside Tolerance
Air Blank	0.000	23:41	

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,BRADLEY,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: QUILES,SEBASTIAN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/07/2024 04:36:38

Diagnostics Check	OK	04:23	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1 0.173	04:26	
Breath Volume	1.828	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 2	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 3 0.186	04:35	
Breath Volume	2.691	04:35	
Air Blank	0.000	04:35	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 11/23/2024 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: LEDOUX,JACOB
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 11/09/2024 23:58:00

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.000	23:53
Breath Volume	1.585	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	2.089	23:56
Air Blank	0.000	23:57
Control Test	0.082	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: MELBOURNE PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: MELBOURNE PD
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ARBELO-VILLARINI

DATE/TIME OF TEST: 11/10/2024 04:55:16

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.115	04:48
Breath Volume	1.378	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.115	04:53
Breath Volume	1.597	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,NATHANIEL,M
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HAMM,MARSHALL
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALAMILLO

DATE/TIME OF TEST: 11/16/2024 01:39:45

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.176	01:34
Breath Volume	3.175	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.173	01:38
Breath Volume	3.160	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BREWER, KYLE
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: BEILSMITH

DATE/TIME OF TEST: 11/16/2024 16:17:22

Diagnostics Check	OK	16:09
Air Blank	0.000	16:09
Control Test	0.082	16:10
Air Blank	0.000	16:10
Subject Sample # 1	0.160	16:11
Breath Volume	2.488	16:11
Air Blank	0.000	16:12
Air Blank	0.000	16:14
Subject Sample # 2	0.160	16:15
Breath Volume	2.144	16:15
Air Blank	0.000	16:16
Control Test	0.081	16:16
Air Blank	0.000	16:17
Diagnostics Check	OK	16:17

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARR,MICHAEL,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SCHULZE,JOSEPH
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 11/16/2024 20:45:25

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.082	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.077	20:41
Breath Volume	1.500	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.078	20:43
Breath Volume	1.500	20:43
Air Blank	0.000	20:44
Control Test	0.081	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FEDRO,DARIEN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 10/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BARAJAS

DATE/TIME OF TEST: 11/23/2024 00:50:27

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.082	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.076	00:45
Breath Volume	2.847	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.076	00:48
Breath Volume	2.601	00:48
Air Blank	0.000	00:49
Control Test	0.082	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 11/23/2024 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SANCHEZ, HIBER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SANTIZO

DATE/TIME OF TEST: 11/29/2024 04:18:16

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.082	04:13	
Air Blank	0.000	04:13	
Subject Sample # 1	0.196	04:14	
Breath Volume	2.718	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample # 2	0.189	04:17	
Breath Volume	2.281	04:17	
Air Blank		04:17	RFI Detect
Air Blank	0.000	04:18	

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SANCHEZ, HIBER
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SANTIZO

DATE/TIME OF TEST: 11/29/2024 04:26:49

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.193	04:22
Breath Volume	2.449	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.190	04:25
Breath Volume	1.871	04:25
Air Blank	0.000	04:25
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SUDMAN, ALEXANDER
ARRESTING OFFICER AGENCY: TITUSVILLE PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: BERTRAND

DATE/TIME OF TEST: 11/29/2024 04:37:14

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.268	04:32
Breath Volume	2.199	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.253	04:35
Breath Volume	2.285	04:35
Air Blank	0.000	04:36
Control Test	0.079	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 11/01/2024 22:21:27

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.082	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.000	22:16
Breath Volume	2.101	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.000	22:20
Breath Volume	2.324	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 11/03/2024 07:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DETRANO,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BONGIORNO,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: CRUZ MONROY

DATE/TIME OF TEST: 11/05/2024 01:32:15

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.215	01:25
Breath Volume	1.375	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.228	01:30
Breath Volume	2.464	01:30
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: WALASON

DATE/TIME OF TEST: 11/08/2024 12:48:49

Diagnostics Check	OK	12:45	
Air Blank	0.000	12:46	
Control Test	0.080	12:46	
Air Blank	0.000	12:47	
Subject Sample	# 1	12:47	Subject Test Refused
Breath Volume	0.000	12:47	
Air Blank	0.000	12:47	
Control Test	0.079	12:48	
Air Blank	0.000	12:48	
Diagnostics Check	OK	12:48	

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,GARY,
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: TRIEGLAFF,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 11/08/2024 16:32:49

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:26	
Control Test	0.079	16:26	
Air Blank	0.000	16:27	
Subject Sample	# 1 0.292	16:28	
Breath Volume	2.589	16:28	
Air Blank	0.000	16:28	
Air Blank	0.000	16:30	
Subject Sample	# 2	16:31	Subject Test Refused
Breath Volume	0.000	16:31	
Air Blank	0.000	16:31	
Control Test	0.078	16:32	
Air Blank	0.000	16:32	
Diagnostics Check	OK	16:32	

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESSLER,GREYSON,L
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BONGIORNO,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BLANCK

DATE/TIME OF TEST: 11/10/2024 21:24:25

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.080	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KILIAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PAOAMA LARA

DATE/TIME OF TEST: 11/13/2024 05:24:00

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.082	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:22	Subject Test Refused
Breath Volume	0.000	05:22	
Air Blank	0.000	05:23	
Control Test	0.080	05:23	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,BENJAMIN,F
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KILIAN,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PALMA LARA

DATE/TIME OF TEST: 11/13/2024 05:30:02

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:28	Subject Test Refused
Breath Volume	0.000	05:28	
Air Blank	0.000	05:29	
Control Test	0.079	05:29	
Air Blank	0.000	05:29	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 10/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 11/28/2024 03:48:40

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.122	03:44
Breath Volume	1.199	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.112	03:47
Breath Volume	1.382	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 01/17/2025 13:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: SIERRA

DATE/TIME OF TEST: 11/09/2024 16:41:19

Diagnostics Check	OK	16:35
Air Blank	0.000	16:35
Control Test	0.079	16:35
Air Blank	0.000	16:36
Subject Sample # 1	0.197	16:36
Breath Volume	2.066	16:36
Air Blank	0.000	16:37
Air Blank	0.000	16:39
Subject Sample # 2	0.187	16:39
Breath Volume	1.539	16:39
Air Blank	0.000	16:40
Control Test	0.080	16:40
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 11/18/2024 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DELACAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: WANG

DATE/TIME OF TEST: 11/13/2024 21:22:54

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.157	21:18
Breath Volume	1.765	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.155	21:21
Breath Volume	1.503	21:21
Air Blank	0.000	21:22
Control Test	0.079	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 11/18/2024 14:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: DELACAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 11/16/2024 02:19:07

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.166	02:14
Breath Volume	1.109	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.166	02:17
Breath Volume	1.515	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/18/2024 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,GAVIN,O
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: HENRY,GAVIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: CASCO

DATE/TIME OF TEST: 11/29/2024 21:45:53

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.080	21:45	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 12/23/2024 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 11/01/2024 03:16:53

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.082	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.082	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 11/01/2024 20:41:58

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.081	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:40	Subject Test Refused
Breath Volume	0.000	20:40	
Air Blank	0.000	20:41	
Control Test	0.082	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CAMPOS JIMENEZ

DATE/TIME OF TEST: 11/02/2024 01:57:18

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.051	01:52
Breath Volume	1.132	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.055	01:55
Breath Volume	2.195	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 11/02/2024 03:22:25

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.182	03:17
Breath Volume	1.742	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.171	03:20
Breath Volume	1.925	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SATHER,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GEBREANENYA

DATE/TIME OF TEST: 11/02/2024 04:32:45

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.105	04:27
Breath Volume	1.410	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:29
Subject Sample # 2	0.100	04:30
Breath Volume	1.800	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: VELARDE ROMAN

DATE/TIME OF TEST: 11/02/2024 21:10:27

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.082	21:08	
Air Blank	0.000	21:08	
Subject Sample	# 1	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Control Test	0.081	21:09	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: KHAKPOUR

DATE/TIME OF TEST: 11/02/2024 22:11:05

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.082	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.134	22:06
Breath Volume	2.644	22:06
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.128	22:09
Breath Volume	1.941	22:09
Air Blank	0.000	22:10
Control Test	0.078	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 11/03/2024 01:04:49

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.148	00:58
Breath Volume	1.417	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.138	01:02
Breath Volume	1.496	01:02
Air Blank	0.000	01:03
Control Test	0.076	01:04
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 11/03/2024 01:33:34

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 11/03/2024 01:31:41

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.136	01:27
Breath Volume	2.812	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.123	01:29
Breath Volume	1.984	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CANO ORTIZ

DATE/TIME OF TEST: 11/03/2024 03:24:19

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:22	Subject Test Refused
Breath Volume	0.000	03:22	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:24	
Diagnostics Check	OK	03:24	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/03/2024 04:46:19

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.082	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.000	04:44	
Air Blank	0.000	04:45	
Control Test	0.082	04:45	
Air Blank	0.000	04:46	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 11/04/2024 00:28:54

Diagnostics Check	OK	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.053	00:22
Breath Volume	1.601	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.054	00:27
Breath Volume	1.464	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: NEWTON LIGHTBOURNE

DATE/TIME OF TEST: 11/04/2024 01:08:50

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.081	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:07	
Control Test	0.081	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 11/04/2024 02:54:55

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.164	02:49
Breath Volume	1.570	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.155	02:53
Breath Volume	1.785	02:53
Air Blank	0.000	02:53
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: HOLTSCHNEIDER

DATE/TIME OF TEST: 11/04/2024 17:57:34

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.081	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.118	17:52
Breath Volume	2.769	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.125	17:55
Breath Volume	3.519	17:55
Air Blank	0.000	17:56
Control Test	0.080	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 11/06/2024 00:23:24

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.024	00:18
Breath Volume	1.367	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.022	00:21
Breath Volume	1.484	00:21
Air Blank	0.000	00:22
Control Test	0.081	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: KALAPP

DATE/TIME OF TEST: 11/07/2024 02:00:15

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.081	01:58	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 11/07/2024 20:04:08

Diagnostics Check	OK	20:01	
Air Blank	0.000	20:01	
Control Test	0.081	20:01	
Air Blank	0.000	20:02	
Subject Sample	# 1	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.081	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 11/07/2024 22:42:54

Diagnostics Check	OK	22:36
Air Blank	0.000	22:37
Control Test	0.081	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.056	22:38
Breath Volume	1.910	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 2	0.057	22:41
Breath Volume	2.035	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PERSAUD,V
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 11/08/2024 05:21:33

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:19	
Control Test	0.081	05:19	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:20	Subject Test Refused
Breath Volume	0.000	05:20	
Air Blank	0.000	05:20	
Control Test	0.081	05:20	
Air Blank	0.000	05:21	
Diagnostics Check	OK	05:21	

Data Download Date/Time: 11/26/2024 08:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 11/08/2024 22:31:15

Diagnostics Check	OK	22:25
Air Blank	0.000	22:25
Control Test	0.081	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.103	22:26
Breath Volume	1.644	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.105	22:29
Breath Volume	1.132	22:29
Air Blank	0.000	22:30
Control Test	0.080	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 11/08/2024 23:36:10

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.082	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Diagnostics Check	OK	23:36	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 11/09/2024 03:06:57

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.118	03:02
Breath Volume	1.812	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:04
Subject Sample # 2	0.121	03:05
Breath Volume	2.285	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAEZ,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: WHELAN

DATE/TIME OF TEST: 11/09/2024 05:20:05

Diagnostics Check	OK	05:17	
Air Blank	0.000	05:17	
Control Test	0.081	05:17	
Air Blank	0.000	05:18	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:19	
Control Test	0.081	05:19	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MUNOZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 11/09/2024 06:23:57

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.082	06:18
Air Blank	0.000	06:18
Subject Sample # 1	0.191	06:19
Breath Volume	1.863	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:21
Subject Sample # 2	0.187	06:22
Breath Volume	3.191	06:22
Air Blank	0.000	06:23
Control Test	0.080	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: ROJAS CAMPOSECO

DATE/TIME OF TEST: 11/09/2024 20:08:47

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.081	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.157	20:04
Breath Volume	2.234	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.163	20:07
Breath Volume	2.070	20:07
Air Blank	0.000	20:07
Control Test	0.080	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HERNANDEZ SANCHEZ

DATE/TIME OF TEST: 11/09/2024 23:59:22

Diagnostics Check	OK	23:56	
Air Blank	0.000	23:56	
Control Test	0.081	23:57	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:57	Subject Test Refused
Breath Volume	0.000	23:57	
Air Blank	0.000	23:58	
Control Test	0.081	23:58	
Air Blank	0.000	23:59	
Diagnostics Check	OK	23:59	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: O'QUINN

DATE/TIME OF TEST: 11/10/2024 00:39:41

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.080	00:35
Breath Volume	1.980	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.079	00:37
Breath Volume	1.773	00:37
Air Blank	0.000	00:38
Control Test	0.081	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SOWERBY

DATE/TIME OF TEST: 11/10/2024 03:08:16

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.082	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.082	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GOODE,D
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: NIMMO

DATE/TIME OF TEST: 11/10/2024 04:40:39

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.119	04:35
Breath Volume	1.574	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.118	04:38
Breath Volume	1.250	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARGUDO,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 11/10/2024 19:43:49

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.192	19:39
Breath Volume	1.742	19:39
Air Blank	0.000	19:40
Air Blank	0.000	19:41
Subject Sample # 2	0.193	19:42
Breath Volume	1.816	19:42
Air Blank	0.000	19:42
Control Test	0.077	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 11/10/2024 21:52:51

Diagnostics Check	OK	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.227	21:48
Breath Volume	2.167	21:48
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.229	21:51
Breath Volume	2.382	21:51
Air Blank	0.000	21:51
Control Test	0.079	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: TARTER

DATE/TIME OF TEST: 11/10/2024 22:55:47

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.081	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.081	22:55	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville Beach
ARRESTING OFFICER: OWENS,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: WOOLEYHAN

DATE/TIME OF TEST: 11/11/2024 01:04:41

Diagnostics Check	OK	01:01	
Air Blank	0.000	01:02	
Control Test	0.081	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:03	
Control Test	0.081	01:04	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: RAYMUNDO FRANCISCO

DATE/TIME OF TEST: 11/11/2024 03:01:45

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.082	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: AMADOR ROJO

DATE/TIME OF TEST: 11/11/2024 05:12:19

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:09	
Control Test	0.082	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 11/11/2024 22:50:15

Diagnostics Check	OK	22:47	
Air Blank	0.000	22:47	
Control Test	0.082	22:48	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.082	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 11/12/2024 01:18:49

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:14
Breath Volume	1.867	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:17
Breath Volume	2.296	01:17
Air Blank	0.000	01:17
Control Test	0.081	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RINCON ROMERO

DATE/TIME OF TEST: 11/12/2024 02:24:12

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.082	02:22	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:23	
Control Test	0.082	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: KAMMERDIENER

DATE/TIME OF TEST: 11/12/2024 05:05:55

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.082	05:03	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 11/12/2024 19:38:26

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.121	19:33
Breath Volume	3.007	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.120	19:36
Breath Volume	2.425	19:36
Air Blank	0.000	19:37
Control Test	0.080	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HANSON,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: SIJAMHODZIC

DATE/TIME OF TEST: 11/13/2024 13:35:06

Diagnostics Check	OK	13:28
Air Blank	0.000	13:28
Control Test	0.082	13:29
Air Blank	0.000	13:29
Subject Sample # 1	0.074	13:30
Breath Volume	2.578	13:30
Air Blank	0.000	13:31
Air Blank	0.000	13:32
Subject Sample # 2	0.069	13:33
Breath Volume	2.234	13:33
Air Blank	0.000	13:34
Control Test	0.080	13:34
Air Blank	0.000	13:35
Diagnostics Check	OK	13:35

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 11/15/2024 00:04:14

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.082	00:02	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.082	00:03	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 11/16/2024 03:25:58

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRADY,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 11/16/2024 07:58:08

Diagnostics Check	OK	07:51
Air Blank	0.000	07:51
Control Test	0.082	07:52
Air Blank	0.000	07:52
Subject Sample # 1	0.163	07:53
Breath Volume	1.382	07:53
Air Blank	0.000	07:54
Air Blank	0.000	07:55
Subject Sample # 2	0.164	07:56
Breath Volume	1.722	07:56
Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 11/26/2024 08:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WHYTE

DATE/TIME OF TEST: 11/16/2024 20:43:22

Diagnostics Check	OK	20:40	
Air Blank	0.000	20:40	
Control Test	0.082	20:41	
Air Blank	0.000	20:41	
Subject Sample	# 1	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:42	
Control Test	0.082	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:43	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SMYRNIOS

DATE/TIME OF TEST: 11/16/2024 23:24:33

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.082	23:22	
Air Blank	0.000	23:22	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:23	
Control Test	0.082	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 11/17/2024 03:22:13

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:21	
Control Test	0.083	03:21	
Air Blank		03:21	RFI Detect
Air Blank	0.000	03:22	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROCOURT,NIA-NICOLE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BARRON

DATE/TIME OF TEST: 11/17/2024 03:31:12

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.183	03:26
Breath Volume	3.437	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.183	03:29
Breath Volume	3.289	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: RANSOM

DATE/TIME OF TEST: 11/17/2024 22:38:53

Diagnostics Check	OK	22:31
Air Blank	0.000	22:31
Control Test	0.082	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.111	22:33
Breath Volume	1.613	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.104	22:37
Breath Volume	1.429	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GRAB

DATE/TIME OF TEST: 11/17/2024 23:58:48

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.065	23:53
Breath Volume	4.148	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.065	23:57
Breath Volume	2.351	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 11/18/2024 00:48:52

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.082	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.076	00:44
Breath Volume	1.964	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.078	00:47
Breath Volume	1.796	00:47
Air Blank	0.000	00:47
Control Test	0.081	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Neptune Beach PD
ARRESTING OFFICER: RONQUILLE,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: LOMIKOVSKYI

DATE/TIME OF TEST: 11/18/2024 01:48:55

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.197	01:44
Breath Volume	2.164	01:44
Air Blank	0.000	01:45
Air Blank	0.000	01:46
Subject Sample # 2	0.200	01:47
Breath Volume	1.539	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,B,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PEREZ AGUILAR

DATE/TIME OF TEST: 11/18/2024 04:03:37

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.257	03:58
Breath Volume	3.191	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:00
Subject Sample # 2	0.249	04:01
Breath Volume	2.894	04:01
Air Blank	0.000	04:02
Control Test	0.075	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Atlantic Beach PD
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: VANDEWEERD

DATE/TIME OF TEST: 11/18/2024 23:24:34

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.082	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.109	23:19
Breath Volume	1.269	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.111	23:22
Breath Volume	1.703	23:22
Air Blank	0.000	23:23
Control Test	0.082	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BAUMANN

DATE/TIME OF TEST: 11/19/2024 22:44:39

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.082	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.166	22:39
Breath Volume	1.261	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:41
Subject Sample # 2	0.159	22:42
Breath Volume	1.417	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: KUDUZOVIC

DATE/TIME OF TEST: 11/20/2024 21:30:07

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.101	21:25
Breath Volume	1.640	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.098	21:28
Breath Volume	2.195	21:28
Air Blank	0.000	21:29
Control Test	0.081	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: PENA VILLASMIL

DATE/TIME OF TEST: 11/21/2024 00:04:33

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.081	23:58
Air Blank	0.000	23:59
Subject Sample # 2	0.208	23:59
Breath Volume	2.234	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.198	00:02
Breath Volume	2.312	00:02
Air Blank	0.000	00:03
Control Test	0.079	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HINTON

DATE/TIME OF TEST: 11/21/2024 02:08:39

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.082	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.082	02:04
Breath Volume	1.667	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.080	02:06
Breath Volume	1.644	02:06
Air Blank	0.000	02:07
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: DUGAN

DATE/TIME OF TEST: 11/22/2024 05:31:13

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.083	05:29	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Control Test	0.083	05:30	
Air Blank	0.000	05:31	
Diagnostics Check	OK	05:31	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 11/22/2024 22:53:38

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.082	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.102	22:48
Breath Volume	3.328	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.099	22:51
Breath Volume	2.839	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GETZ,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 11/23/2024 04:38:54

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:36	
Control Test	0.082	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.082	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURGUTOVIC,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GIRTMAN

DATE/TIME OF TEST: 11/24/2024 03:00:44

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.082	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.082	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 11/24/2024 22:08:56

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.082	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.148	22:03
Breath Volume	1.480	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:05
Subject Sample # 2	0.134	22:07
Breath Volume	1.691	22:07
Air Blank	0.000	22:08
Control Test	0.080	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CLINTON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 11/25/2024 15:23:21

Diagnostics Check	OK	15:20	
Air Blank	0.000	15:20	
Control Test	0.081	15:21	
Air Blank	0.000	15:21	
Subject Sample	# 1	15:21	Subject Test Refused
Breath Volume	0.000	15:21	
Air Blank	0.000	15:22	
Control Test	0.081	15:22	
Air Blank	0.000	15:23	
Diagnostics Check	OK	15:23	

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 11/26/2024 00:53:49

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.082	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:49
Breath Volume	1.523	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.000	00:52
Breath Volume	1.871	00:52
Air Blank	0.000	00:52
Control Test	0.082	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 11/26/2024 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 11/29/2024 03:57:55

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.083	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.082	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: OFFEN

DATE/TIME OF TEST: 11/29/2024 20:10:51

Diagnostics Check	OK	20:04
Air Blank	0.000	20:04
Control Test	0.082	20:04
Air Blank	0.000	20:05
Subject Sample # 1	0.265	20:06
Breath Volume	1.765	20:06
Air Blank	0.000	20:06
Air Blank	0.000	20:08
Subject Sample # 2	0.277	20:09
Breath Volume	2.023	20:09
Air Blank	0.000	20:09
Control Test	0.079	20:10
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: NALL

DATE/TIME OF TEST: 11/02/2024 02:41:49

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.126	02:37
Breath Volume	1.226	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.124	02:40
Breath Volume	1.894	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 11/02/2024 23:23:10

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.268	23:17
Breath Volume	1.445	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.262	23:21
Breath Volume	1.304	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/03/2024 01:50:35

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.135	01:46
Breath Volume	2.808	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.136	01:49
Breath Volume	2.085	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MABRY

DATE/TIME OF TEST: 11/03/2024 04:14:31

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.021	04:10
Breath Volume	2.000	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.020	04:13
Breath Volume	2.382	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: MARTINEZ-AMAYA

DATE/TIME OF TEST: 11/03/2024 20:24:49

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.286	20:20
Breath Volume	1.285	20:20
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.277	20:23
Breath Volume	1.203	20:23
Air Blank	0.000	20:23
Control Test	0.079	20:24
Air Blank	0.000	20:24
Diagnostics Check	OK	20:24

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: PAY

DATE/TIME OF TEST: 11/08/2024 04:26:55

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.064	04:22
Breath Volume	1.875	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.064	04:25
Breath Volume	2.285	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 11/09/2024 01:48:40

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.203	01:43
Breath Volume	1.609	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.195	01:47
Breath Volume	1.500	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 11/10/2024 03:34:04

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:30	No Sample Provided
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Air Blank	0.000	03:32	
Subject Sample	# 2	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 11/12/2024 22:58:21

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.187	22:53
Breath Volume	1.421	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.177	22:56
Breath Volume	1.242	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 11/16/2024 01:38:25

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:30	
Control Test	0.079	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:34	No Sample Provided
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RAFIDI

DATE/TIME OF TEST: 11/17/2024 04:27:46

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.038	04:23
Breath Volume	2.121	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.038	04:26
Breath Volume	1.886	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 11/22/2024 01:50:40

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:46	No Sample Provided
Breath Volume	0.000	01:46	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:50	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RODECK,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GERARD

DATE/TIME OF TEST: 11/23/2024 22:51:39

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:40	
Control Test	0.080	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:43	No Sample Provided
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2	0.129 22:47	
Breath Volume	1.789	22:47	
Air Blank	0.000	22:47	
Air Blank	0.000	22:49	
Subject Sample	# 3	0.123 22:50	
Breath Volume	1.457	22:50	
Air Blank	0.000	22:50	
Control Test	0.080	22:51	
Air Blank	0.000	22:51	
Diagnostics Check	OK	22:51	

Data Download Date/Time: 11/29/2024 19:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DEVER

DATE/TIME OF TEST: 11/24/2024 03:51:48

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.178	03:47
Breath Volume	1.140	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.174	03:50
Breath Volume	1.175	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SHERIDAN

DATE/TIME OF TEST: 11/24/2024 04:06:55

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.080	03:59	
Air Blank	0.000	03:59	
Subject Sample	# 1	04:02	No Sample Provided
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 2	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,R,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/26/2024 00:28:51

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.138	00:24
Breath Volume	1.738	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.137	00:27
Breath Volume	2.140	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/26/2024 03:23:01

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:18	No Sample Provided
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: CAAL

DATE/TIME OF TEST: 11/26/2024 23:21:43

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.137	23:17
Breath Volume	1.410	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.147	23:20
Breath Volume	1.949	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: YARBROUGH,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 10/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/28/2024 04:03:35

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.149	03:58
Breath Volume	3.113	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.140	04:02
Breath Volume	1.929	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/29/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 11/29/2024 20:33:42

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.320	20:28
Breath Volume	4.398	20:28
Air Blank	0.000	20:29
Air Blank	0.000	20:30
Subject Sample # 2	0.305	20:32
Breath Volume	3.902	20:32
Air Blank	0.000	20:32
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKLEY,CORDERRO,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: YUCUTE

DATE/TIME OF TEST: 11/29/2024 22:17:29

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.199	22:12
Breath Volume	2.667	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.196	22:16
Breath Volume	1.894	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRODY,A
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: MORARU,F
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:53

SUBJECT'S LAST NAME: RODRIGUEZ TORRES

DATE/TIME OF TEST: 11/16/2024 14:31:20

Diagnostics Check	OK	14:24
Air Blank	0.000	14:25
Control Test	0.079	14:25
Air Blank	0.000	14:26
Subject Sample # 1	0.214	14:26
Breath Volume	3.457	14:26
Air Blank	0.000	14:27
Air Blank	0.000	14:29
Subject Sample # 2	0.216	14:29
Breath Volume	2.972	14:29
Air Blank	0.000	14:30
Control Test	0.078	14:30
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

Data Download Date/Time: 11/22/2024 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,SHAWNDR,T
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: CARMICHAEL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MORELL HERNANDEZ

DATE/TIME OF TEST: 11/25/2024 00:40:13

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.197	00:35
Breath Volume	3.238	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.190	00:38
Breath Volume	2.546	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/13/2024 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: RAMIREZ,ANGEL
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 10/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/08/2024 00:33:28

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.100	00:28
Breath Volume	1.226	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.105	00:31
Breath Volume	2.148	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 11/12/2024 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JUDKINS

DATE/TIME OF TEST: 11/06/2024 02:29:02

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:24	
Control Test	0.079	02:24	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:27	RFI Detect
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.079	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 11/25/2024 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 10/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PATMON

DATE/TIME OF TEST: 11/09/2024 23:48:57

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.078	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.076	23:44
Breath Volume	1.109	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.074	23:47
Breath Volume	1.218	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 11/25/2024 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PINZON MOLINA

DATE/TIME OF TEST: 11/24/2024 03:17:17

Diagnostics Check	OK	03:16	
Air Blank		03:16	RFI Detect
Air Blank	0.000	03:17	

Data Download Date/Time: 12/11/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PINZON MOLINA

DATE/TIME OF TEST: 11/24/2024 03:26:55

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.206	03:22
Breath Volume	2.707	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.199	03:25
Breath Volume	2.675	03:25
Air Blank	0.000	03:26
Control Test	0.077	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 12/11/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REBOLLEDO, OSCAR, J
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: EMERENCIANO, MATHEUS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402290677
DRY GAS EXPIRY DATE: 12/07/2024
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: DEWILDE

DATE/TIME OF TEST: 11/27/2024 15:50:05

Diagnostics Check	OK	15:43
Air Blank	0.000	15:44
Control Test	0.076	15:44
Air Blank	0.000	15:45
Subject Sample # 1	0.088	15:45
Breath Volume	2.707	15:45
Air Blank	0.000	15:46
Air Blank	0.000	15:48
Subject Sample # 2	0.086	15:48
Breath Volume	1.937	15:48
Air Blank	0.000	15:49
Control Test	0.076	15:49
Air Blank	0.000	15:50
Diagnostics Check	OK	15:50

Data Download Date/Time: 12/11/2024 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,OFFICER,
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: BERRIOS,OFFICER
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:59

SUBJECT'S LAST NAME: BALDINGER

DATE/TIME OF TEST: 11/06/2024 11:26:28

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.080	11:20
Air Blank	0.000	11:21
Subject Sample # 1	0.205	11:21
Breath Volume	2.527	11:21
Air Blank	0.000	11:22
Air Blank	0.000	11:24
Subject Sample # 2	0.195	11:24
Breath Volume	2.691	11:24
Air Blank	0.000	11:25
Control Test	0.079	11:25
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

Data Download Date/Time: 11/24/2024 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 11/14/2024 18:13:17

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.080	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.270	18:08
Breath Volume	2.832	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.267	18:11
Breath Volume	2.871	18:11
Air Blank	0.000	18:12
Control Test	0.080	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 11/24/2024 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: CUBE,K
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 11/08/2024 02:08:59

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.063	02:02
Breath Volume	1.539	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.055	02:07
Breath Volume	1.371	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 11/30/2024 15:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SCHALLES,C
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 10/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: DA CONCEICAO

DATE/TIME OF TEST: 11/17/2024 04:43:27

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.099	04:38
Breath Volume	3.351	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.093	04:41
Breath Volume	3.433	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 11/30/2024 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 11/05/2024 02:15:11

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:13	Subject Test Refused
Breath Volume	0.000	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Diagnostics Check	OK	02:15	

Data Download Date/Time: 11/14/2024 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONFERRARO,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: KOZAK

DATE/TIME OF TEST: 11/13/2024 21:07:36

Diagnostics Check	OK	21:04	
Air Blank	0.000	21:05	
Control Test	0.080	21:05	
Air Blank	0.000	21:06	
Subject Sample	# 1	21:06	Subject Test Refused
Breath Volume	0.000	21:06	
Air Blank	0.000	21:06	
Control Test	0.080	21:06	
Air Blank	0.000	21:07	
Diagnostics Check	OK	21:07	

Data Download Date/Time: 11/14/2024 19:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BRACHFELD

DATE/TIME OF TEST: 11/22/2024 22:48:22

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.080	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.142	22:44
Breath Volume	2.230	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.138	22:46
Breath Volume	2.257	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: JOHNSTON-SCHMIDT

DATE/TIME OF TEST: 11/23/2024 00:56:57

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 11/24/2024 19:18:24

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.080	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:17	
Control Test	0.081	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLOWAY,DEVARIOUS,J
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: MORESCO-OCELLO,MARIA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: TOWNES

DATE/TIME OF TEST: 11/28/2024 18:34:55

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.076	18:27
Air Blank	0.000	18:27
Subject Sample # 1	0.257	18:29
Breath Volume	1.253	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.276	18:33
Breath Volume	1.589	18:33
Air Blank	0.000	18:33
Control Test	0.076	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 12/11/2024 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: FEBLES CANO

DATE/TIME OF TEST: 11/01/2024 19:38:53

Diagnostics Check	OK	19:31
Air Blank	0.000	19:31
Control Test	0.079	19:32
Air Blank	0.000	19:32
Subject Sample # 1	0.014	19:34
Breath Volume	1.664	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.011	19:37
Breath Volume	1.296	19:37
Air Blank	0.000	19:37
Control Test	0.079	19:38
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 11/29/2024 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,H,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,H
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ALFONSO MARTIN

DATE/TIME OF TEST: 11/01/2024 23:58:30

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.188	23:53
Breath Volume	1.464	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.175	23:56
Breath Volume	1.160	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 11/29/2024 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CHANDLER,R
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: FORTUNATO

DATE/TIME OF TEST: 11/02/2024 00:31:03

Diagnostics Check	OK	00:24
Air Blank	0.000	00:25
Control Test	0.079	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.226	00:26
Breath Volume	1.703	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.225	00:29
Breath Volume	1.835	00:29
Air Blank	0.000	00:30
Control Test	0.079	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:31

Data Download Date/Time: 11/29/2024 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BECH

DATE/TIME OF TEST: 11/15/2024 03:09:15

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.246	03:04
Breath Volume	2.714	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.246	03:07
Breath Volume	2.277	03:07
Air Blank	0.000	03:08
Control Test	0.079	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 11/29/2024 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTFOOT,OFC,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: LIGHTFOOT,OFC
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BADTKE

DATE/TIME OF TEST: 11/19/2024 01:06:04

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.079	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1 0.126	00:58	Volume Not Met
Breath Volume	0.019	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2 0.125	01:04	Volume Not Met
Breath Volume	0.031	01:04	
Air Blank	0.000	01:05	
Control Test	0.077	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 11/29/2024 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VILLANICENCIO

DATE/TIME OF TEST: 11/24/2024 01:07:55

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.078	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.000	01:03
Breath Volume	2.636	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.000	01:06
Breath Volume	1.875	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 11/29/2024 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 11/01/2024 03:58:45

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.082	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.080	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WEIPPERT,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MERRIWEATHER

DATE/TIME OF TEST: 11/02/2024 04:41:22

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.119	04:36
Breath Volume	2.156	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.117	04:39
Breath Volume	1.902	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 11/03/2024 01:33:53

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.081	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.036	01:29
Breath Volume	1.488	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.033	01:32
Breath Volume	1.632	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PHIMMASANE

DATE/TIME OF TEST: 11/03/2024 03:25:46

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.082	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.164	03:20
Breath Volume	1.386	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.174	03:24
Breath Volume	2.148	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HILLIARD

DATE/TIME OF TEST: 11/04/2024 01:00:27

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.082	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.250	00:55
Breath Volume	1.867	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.256	00:58
Breath Volume	2.019	00:58
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: FENDER

DATE/TIME OF TEST: 11/04/2024 03:43:21

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.064	03:38
Breath Volume	2.382	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.067	03:41
Breath Volume	1.546	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 11/04/2024 05:50:21

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.081	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.106	05:45
Breath Volume	2.683	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:48
Subject Sample # 2	0.091	05:48
Breath Volume	2.101	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 11/25/2024 08:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 11/04/2024 23:56:01

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.082	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:55	
Control Test	0.079	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 11/06/2024 01:03:41

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.082	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.099	00:59
Breath Volume	4.265	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.092	01:02
Breath Volume	1.410	01:02
Air Blank	0.000	01:02
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SCHULER

DATE/TIME OF TEST: 11/07/2024 01:28:39

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.080	01:22	
Air Blank	0.000	01:23	
Subject Sample	# 1 0.206	01:24	
Breath Volume	1.941	01:24	
Air Blank	0.000	01:25	
Air Blank	0.000	01:26	
Subject Sample	# 2	01:27	RFI Detect
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.078	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SCHULER

DATE/TIME OF TEST: 11/07/2024 01:37:38

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.194	01:32
Breath Volume	1.980	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.185	01:35
Breath Volume	1.296	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 11/07/2024 05:03:31

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.081	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.100	04:58
Breath Volume	2.453	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.102	05:01
Breath Volume	1.847	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: ALLRED

DATE/TIME OF TEST: 11/07/2024 23:00:36

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.080	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.026	22:56
Breath Volume	2.332	22:56
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.027	22:59
Breath Volume	2.359	22:59
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,L,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HART-RANDALL

DATE/TIME OF TEST: 11/09/2024 03:26:08

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.081	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GARCIA LOPEZ

DATE/TIME OF TEST: 11/11/2024 00:35:46

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:30
Air Blank	0.000	00:30
Subject Sample # 1	0.000	00:31
Breath Volume	1.281	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.000	00:34
Breath Volume	1.328	00:34
Air Blank	0.000	00:34
Control Test	0.080	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 11/12/2024 16:08:39

Diagnostics Check	OK	16:05	
Air Blank	0.000	16:06	
Control Test	0.081	16:06	
Air Blank	0.000	16:07	
Subject Sample	# 1	16:07	Subject Test Refused
Breath Volume	0.000	16:07	
Air Blank	0.000	16:07	
Control Test	0.080	16:08	
Air Blank	0.000	16:08	
Diagnostics Check	OK	16:08	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CANAVAN

DATE/TIME OF TEST: 11/13/2024 04:06:31

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.081	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.000	04:01
Breath Volume	2.402	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.000	04:05
Breath Volume	2.187	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 11/14/2024 02:34:44

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.106	02:29
Breath Volume	3.589	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.108	02:33
Breath Volume	3.093	02:33
Air Blank	0.000	02:33
Control Test	0.081	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: NOWAK

DATE/TIME OF TEST: 11/15/2024 00:27:31

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.082	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.056	00:22
Breath Volume	2.031	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.060	00:25
Breath Volume	1.757	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ORTEGA CRUZATTI

DATE/TIME OF TEST: 11/15/2024 01:36:47

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.082	01:34	
Air Blank	0.000	01:35	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.080	01:36	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SHERIFF

DATE/TIME OF TEST: 11/16/2024 03:47:39

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.132	03:42
Breath Volume	1.531	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.120	03:46
Breath Volume	1.207	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/16/2024 04:25:57

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.081	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 1	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:25	
Control Test	0.080	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 11/17/2024 00:02:22

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.082	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.235	23:57
Breath Volume	2.835	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.238	00:00
Breath Volume	3.011	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LADD,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: KELSEY

DATE/TIME OF TEST: 11/22/2024 17:23:08

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.082	17:17
Air Blank	0.000	17:17
Subject Sample # 1	0.280	17:18
Breath Volume	1.789	17:18
Air Blank	0.000	17:19
Air Blank	0.000	17:20
Subject Sample # 2	0.284	17:21
Breath Volume	1.785	17:21
Air Blank	0.000	17:22
Control Test	0.081	17:22
Air Blank	0.000	17:23
Diagnostics Check	OK	17:23

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JERNIGAN

DATE/TIME OF TEST: 11/22/2024 23:20:00

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.082	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.125	23:15
Breath Volume	1.933	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.124	23:18
Breath Volume	1.875	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:20

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 11/23/2024 03:43:07

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.083	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KRAMER,S
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 11/23/2024 05:13:47

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.082	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.143	05:08
Breath Volume	4.695	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.145	05:12
Breath Volume	4.222	05:12
Air Blank	0.000	05:12
Control Test	0.082	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: SIMO

DATE/TIME OF TEST: 11/23/2024 07:33:51

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:31	
Control Test	0.083	07:31	
Air Blank	0.000	07:32	
Subject Sample	# 1	07:32	Subject Test Refused
Breath Volume	0.000	07:32	
Air Blank	0.000	07:32	
Control Test	0.081	07:33	
Air Blank	0.000	07:33	
Diagnostics Check	OK	07:33	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BRUGAL CRUZ

DATE/TIME OF TEST: 11/23/2024 19:33:38

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:30	
Control Test	0.083	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.081	19:33	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 11/24/2024 04:37:26

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.082	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.092	04:32
Breath Volume	1.562	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.094	04:35
Breath Volume	1.757	04:35
Air Blank	0.000	04:36
Control Test	0.082	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARGO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: DAVILA FABIAN

DATE/TIME OF TEST: 11/24/2024 16:32:27

Diagnostics Check	OK	16:25
Air Blank	0.000	16:25
Control Test	0.082	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.193	16:27
Breath Volume	2.195	16:27
Air Blank	0.000	16:28
Air Blank	0.000	16:29
Subject Sample # 2	0.194	16:30
Breath Volume	1.546	16:30
Air Blank	0.000	16:31
Control Test	0.081	16:31
Air Blank	0.000	16:32
Diagnostics Check	OK	16:32

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 11/24/2024 18:33:45

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:31	
Control Test	0.081	18:31	
Air Blank	0.000	18:32	
Subject Sample	# 1	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:32	
Control Test	0.080	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 11/25/2024 08:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: AMMONS

DATE/TIME OF TEST: 11/25/2024 02:21:35

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.082	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.160	02:16
Breath Volume	1.125	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.161	02:19
Breath Volume	1.718	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 11/25/2024 08:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/26/2024 04:09:02

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.082	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.130	04:04
Breath Volume	1.398	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.137	04:07
Breath Volume	1.230	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 12/06/2024 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: WHITSETT

DATE/TIME OF TEST: 11/26/2024 05:43:14

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:31	
Control Test	0.081	05:31	
Air Blank	0.000	05:32	
Subject Sample	# 1 0.161	05:35	Volume Not Met
Breath Volume	0.914	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:37	
Subject Sample	# 2 0.151	05:38	
Breath Volume	1.390	05:38	
Air Blank	0.000	05:39	
Air Blank	0.000	05:40	
Subject Sample	# 3 0.168	05:41	
Breath Volume	1.476	05:41	
Air Blank	0.000	05:42	
Control Test	0.077	05:42	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:43	

Data Download Date/Time: 12/06/2024 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: EWTON

DATE/TIME OF TEST: 11/28/2024 00:38:59

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.082	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:38	
Control Test	0.080	00:38	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HIXSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: RAPONE

DATE/TIME OF TEST: 11/28/2024 01:57:30

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.110	01:52
Breath Volume	1.414	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.113	01:55
Breath Volume	1.324	01:55
Air Blank	0.000	01:56
Control Test	0.082	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 11/28/2024 05:14:54

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:12	
Control Test	0.082	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 1	05:13	Subject Test Refused
Breath Volume	0.000	05:13	
Air Blank	0.000	05:13	
Control Test	0.081	05:14	
Air Blank	0.000	05:14	
Diagnostics Check	OK	05:14	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: GODINEZ ANDRES

DATE/TIME OF TEST: 11/28/2024 20:43:05

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.082	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.210	20:36
Breath Volume	1.941	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:38
Subject Sample # 2	0.211	20:41
Breath Volume	2.011	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:43

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SAAFIR

DATE/TIME OF TEST: 11/28/2024 21:28:57

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.081	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1 0.191	21:21	Volume Not Met
Breath Volume	0.492	21:21	
Air Blank	0.000	21:22	
Air Blank	0.000	21:24	
Subject Sample	# 2 0.116	21:27	Volume Not Met
Breath Volume	0.074	21:27	
Air Blank	0.000	21:27	
Control Test	0.079	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SAAFIR

DATE/TIME OF TEST: 11/28/2024 21:43:53

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.079	21:40	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 11/29/2024 05:09:44

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.188	05:04
Breath Volume	2.015	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.192	05:08
Breath Volume	1.867	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO, SOLOMON,
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: BELTRAN, CHANTEL
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 10/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: CHILDE

DATE/TIME OF TEST: 11/13/2024 11:19:43

Diagnostics Check	OK	11:07	
Air Blank	0.000	11:08	
Control Test	0.076	11:08	
Air Blank	0.000	11:08	
Subject Sample # 1	0.252	11:09	
Breath Volume	3.171	11:09	
Air Blank	0.000	11:10	
Air Blank	0.000	11:11	
Subject Sample # 2	0.240	11:15	Volume Not Met
Breath Volume	0.019	11:15	
Air Blank	0.000	11:15	
Air Blank	0.000	11:17	
Subject Sample # 3	0.232	11:17	
Breath Volume	2.605	11:17	
Air Blank	0.000	11:18	
Control Test	0.072	11:19	Control Outside Tolerance
Air Blank	0.000	11:19	

Data Download Date/Time: 11/15/2024 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: FELLOWS,PAUL
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: CHINCHAR

DATE/TIME OF TEST: 11/18/2024 23:14:40

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.075	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.231	23:10
Breath Volume	1.402	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.234	23:13
Breath Volume	1.171	23:13
Air Blank	0.000	23:13
Control Test	0.076	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 12/05/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,PAUL,S
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: GONZALES,PAUL
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: KOHLMAYER

DATE/TIME OF TEST: 11/28/2024 23:03:58

Diagnostics Check	OK	22:57
Air Blank	0.000	22:57
Control Test	0.076	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.246	22:59
Breath Volume	2.015	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.241	23:02
Breath Volume	2.179	23:02
Air Blank	0.000	23:03
Control Test	0.076	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 12/05/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,OFFICER,Z
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HUTCHINSON,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 11/04/2024 22:59:29

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.079	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.212	22:55
Breath Volume	1.394	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.195	22:58
Breath Volume	1.660	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 11/11/2024 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: WAECHTER

DATE/TIME OF TEST: 11/08/2024 20:08:41

Diagnostics Check	OK	20:00	
Air Blank	0.000	20:00	
Control Test	0.081	20:00	
Air Blank	0.000	20:01	
Subject Sample	# 1	20:04	No Sample Provided
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Air Blank	0.000	20:07	
Subject Sample	# 2	20:07	Subject Test Refused
Breath Volume	0.000	20:07	
Air Blank	0.000	20:07	
Control Test	0.081	20:08	
Air Blank	0.000	20:08	
Diagnostics Check	OK	20:08	

Data Download Date/Time: 11/11/2024 22:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 11/12/2024 01:42:55

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.082	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:41	Subject Test Refused
Breath Volume	0.000	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 12/05/2024 00:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: WOMACK

DATE/TIME OF TEST: 11/15/2024 00:26:55

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.339	00:22
Breath Volume	1.980	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.344	00:25
Breath Volume	1.746	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 12/05/2024 00:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 11/15/2024 02:19:02

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.137	02:14
Breath Volume	4.425	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.142	02:17
Breath Volume	3.753	02:17
Air Blank	0.000	02:18
Control Test	0.080	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:19

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 11/15/2024 03:01:48

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 11/15/2024 21:08:39

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.255	21:02
Breath Volume	1.648	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:04
Subject Sample # 2	0.241	21:07
Breath Volume	1.226	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: IFFLA

DATE/TIME OF TEST: 11/16/2024 00:46:46

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.000	00:42
Breath Volume	1.570	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.000	00:45
Breath Volume	1.589	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: LOONEY

DATE/TIME OF TEST: 11/16/2024 22:50:14

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:43	No Sample Provided
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.080	22:49	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: FIX

DATE/TIME OF TEST: 11/17/2024 04:14:43

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.133	04:10
Breath Volume	2.949	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.128	04:13
Breath Volume	2.007	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RUFFIN

DATE/TIME OF TEST: 11/17/2024 04:58:14

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:48	
Subject Sample	# 1	04:51	No Sample Provided
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 2	04:56	No Sample Provided
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.080	04:57	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DAVID,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: DAUGHETEE,NICHOLAS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 11/18/2024 08:17:36

Diagnostics Check	OK	08:12	
Air Blank	0.000	08:13	
Control Test	0.080	08:13	
Air Blank	0.000	08:13	
Subject Sample	# 1	08:16	Subject Test Refused
Breath Volume	0.000	08:16	
Air Blank	0.000	08:16	
Control Test	0.080	08:17	
Air Blank	0.000	08:17	
Diagnostics Check	OK	08:17	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: KONSTANTINIDIS

DATE/TIME OF TEST: 11/19/2024 10:29:29

Diagnostics Check	OK	10:23
Air Blank	0.000	10:23
Control Test	0.080	10:23
Air Blank	0.000	10:24
Subject Sample # 1	0.000	10:25
Breath Volume	2.464	10:25
Air Blank	0.000	10:25
Air Blank	0.000	10:27
Subject Sample # 2	0.000	10:28
Breath Volume	2.066	10:28
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: YUTZY

DATE/TIME OF TEST: 11/20/2024 16:46:36

Diagnostics Check	OK	16:40
Air Blank	0.000	16:40
Control Test	0.080	16:41
Air Blank	0.000	16:41
Subject Sample # 1	0.000	16:42
Breath Volume	2.453	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 2	0.000	16:45
Breath Volume	2.632	16:45
Air Blank	0.000	16:45
Control Test	0.080	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: ZAMPACH

DATE/TIME OF TEST: 11/21/2024 14:31:12

Diagnostics Check	OK	14:24
Air Blank	0.000	14:25
Control Test	0.082	14:25
Air Blank	0.000	14:26
Subject Sample # 1	0.000	14:26
Breath Volume	2.003	14:26
Air Blank	0.000	14:27
Air Blank	0.000	14:29
Subject Sample # 2	0.000	14:29
Breath Volume	1.886	14:29
Air Blank	0.000	14:30
Control Test	0.081	14:30
Air Blank	0.000	14:31
Diagnostics Check	OK	14:31

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIGAN, MATTHEW, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD, MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 11/23/2024 00:33:01

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.205	00:28
Breath Volume	4.234	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.206	00:31
Breath Volume	4.046	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:33

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 11/23/2024 01:40:04

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.228	01:35
Breath Volume	2.429	01:35
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.226	01:38
Breath Volume	2.441	01:38
Air Blank	0.000	01:39
Control Test	0.081	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:40

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,ASHLEY,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 11/23/2024 03:11:46

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.081	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:04	No Sample Provided
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 2	03:10	No Sample Provided
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.081	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,DUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: CROUSE

DATE/TIME OF TEST: 11/23/2024 05:30:02

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.081	05:24
Air Blank	0.000	05:24
Subject Sample # 1	0.111	05:25
Breath Volume	1.621	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.116	05:28
Breath Volume	1.773	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 11/27/2024 15:19:22

Diagnostics Check	OK	15:13
Air Blank	0.000	15:14
Control Test	0.081	15:14
Air Blank	0.000	15:14
Subject Sample # 1	0.000	15:15
Breath Volume	1.488	15:15
Air Blank	0.000	15:15
Air Blank	0.000	15:17
Subject Sample # 2	0.000	15:18
Breath Volume	1.457	15:18
Air Blank	0.000	15:18
Control Test	0.082	15:18
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JULIO,C
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MASLE,ADAM
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DIETRICH

DATE/TIME OF TEST: 11/28/2024 04:09:04

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.108	04:01	Volume Not Met
Breath Volume	0.429	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:04	
Subject Sample	# 2 0.110	04:04	
Breath Volume	1.160	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:07	
Subject Sample	# 3 0.108	04:07	
Breath Volume	1.203	04:07	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:09	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: WANG

DATE/TIME OF TEST: 11/29/2024 18:07:56

Diagnostics Check	OK	18:01
Air Blank	0.000	18:02
Control Test	0.080	18:02
Air Blank	0.000	18:03
Subject Sample # 1	0.199	18:03
Breath Volume	1.886	18:03
Air Blank	0.000	18:04
Air Blank	0.000	18:05
Subject Sample # 2	0.197	18:06
Breath Volume	1.449	18:06
Air Blank	0.000	18:07
Control Test	0.079	18:07
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SATYAL

DATE/TIME OF TEST: 11/30/2024 02:30:54

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,L
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 11/30/2024 22:39:40

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.081	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:35	No Sample Provided
Breath Volume	0.000	22:35	
Air Blank	0.000	22:36	
Air Blank	0.000	22:38	
Subject Sample	# 2	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.082	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WAUGH CASTLE

DATE/TIME OF TEST: 11/10/2024 03:13:15

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.080	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 11/18/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,RYAN,N
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: TERRELL,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MACIAS

DATE/TIME OF TEST: 11/11/2024 02:07:31

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.147	02:03
Breath Volume	2.019	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.150	02:06
Breath Volume	1.515	02:06
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 11/18/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISA,NERITAN,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HISA,NERITAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/17/2024 07:39:28

Diagnostics Check	OK	07:38	
Air Blank		07:39	Ambient Fail
Air Blank	0.000	07:39	

Data Download Date/Time: 11/18/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISA,NERITAN,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HISA,NERITAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: W422002004560

DATE/TIME OF TEST: 11/17/2024 07:47:29

Diagnostics Check	OK	07:41	
Air Blank	0.000	07:42	
Control Test	0.080	07:42	
Air Blank	0.000	07:43	
Subject Sample	# 1 0.194	07:43	
Breath Volume	2.199	07:43	
Air Blank	0.000	07:44	
Air Blank		07:45	Improper Sample
Air Blank	0.000	07:46	
Control Test	0.081	07:46	
Air Blank	0.000	07:47	
Diagnostics Check	OK	07:47	

Data Download Date/Time: 11/18/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISA,NERITAN,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HISA,NERITAN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 10/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:17

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 11/17/2024 08:03:11

Diagnostics Check	OK	07:56
Air Blank	0.000	07:57
Control Test	0.079	07:57
Air Blank	0.000	07:58
Subject Sample # 1	0.190	07:58
Breath Volume	4.382	07:58
Air Blank	0.000	07:59
Air Blank	0.000	08:01
Subject Sample # 2	0.192	08:01
Breath Volume	4.042	08:01
Air Blank	0.000	08:02
Control Test	0.079	08:02
Air Blank	0.000	08:03
Diagnostics Check	OK	08:03

Data Download Date/Time: 11/18/2024 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CONGRAINS

DATE/TIME OF TEST: 11/01/2024 05:13:41

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.243	05:08
Breath Volume	1.140	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.229	05:12
Breath Volume	1.183	05:12
Air Blank	0.000	05:12
Control Test	0.078	05:13
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 11/02/2024 04:07:19

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.079	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:05	No Sample Provided
Breath Volume	0.000	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: GONZALEZ MARTINEZ

DATE/TIME OF TEST: 11/03/2024 01:39:12

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.101	01:34
Breath Volume	1.816	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.099	01:37
Breath Volume	1.937	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 11/03/2024 03:03:10

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:56	No Sample Provided
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Air Blank	0.000	02:58	
Subject Sample	# 2	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,KALEN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HOLMES,KALEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: SOWELL

DATE/TIME OF TEST: 11/03/2024 21:52:33

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.084	21:47
Breath Volume	1.574	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.078	21:50
Breath Volume	1.250	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 11/03/2024 23:35:19

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.050	23:30
Breath Volume	2.054	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.045	23:33
Breath Volume	2.625	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NASON,TAYLOR
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 11/07/2024 18:07:28

Diagnostics Check	OK	17:53	
Air Blank	0.000	17:54	
Control Test	0.081	17:54	
Air Blank	0.000	17:55	
Subject Sample	# 1 0.192	17:58	Volume Not Met
Breath Volume	0.156	17:58	
Air Blank	0.000	17:59	
Air Blank	0.000	18:00	
Subject Sample	# 2 0.194	18:01	
Breath Volume	1.183	18:01	
Air Blank	0.000	18:02	
Air Blank	0.000	18:04	
Subject Sample	# 3 0.190	18:05	
Breath Volume	1.187	18:05	
Air Blank	0.000	18:06	
Control Test	0.078	18:06	
Air Blank	0.000	18:07	
Diagnostics Check	OK	18:07	

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: THACKER

DATE/TIME OF TEST: 11/08/2024 09:09:34

Diagnostics Check	OK	09:03
Air Blank	0.000	09:03
Control Test	0.080	09:04
Air Blank	0.000	09:04
Subject Sample # 1	0.000	09:05
Breath Volume	1.953	09:05
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample # 2	0.000	09:08
Breath Volume	1.250	09:08
Air Blank	0.000	09:08
Control Test	0.081	09:08
Air Blank	0.000	09:09
Diagnostics Check	OK	09:09

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 11/09/2024 23:53:41

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.080	23:43	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:46	No Sample Provided
Breath Volume	0.000	23:46	
Air Blank	0.000	23:47	
Air Blank	0.000	23:49	
Subject Sample	# 2	23:52	No Sample Provided
Breath Volume	0.000	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:53	

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: LOPEZ-PERKINS

DATE/TIME OF TEST: 11/10/2024 23:01:24

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.290	22:56
Breath Volume	1.812	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.308	22:59
Breath Volume	1.910	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 11/10/2024 23:25:34

Diagnostics Check	OK	23:24	
Air Blank		23:25	Sequence Aborted
Air Blank	0.000	23:25	

Data Download Date/Time: 11/12/2024 09:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LOVETTE

DATE/TIME OF TEST: 11/10/2024 23:35:54

Diagnostics Check	OK	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.119	23:30
Breath Volume	1.203	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.102	23:34
Breath Volume	3.945	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 11/12/2024 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MARTINEZ CHAVEZ

DATE/TIME OF TEST: 11/15/2024 01:13:58

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:03	
Control Test	0.080	01:03	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:06	No Sample Provided
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 2	01:12	No Sample Provided
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:13	

Data Download Date/Time: 12/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: TORRES CANARIO

DATE/TIME OF TEST: 11/16/2024 19:33:35

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.182	19:27
Breath Volume	1.781	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.181	19:31
Breath Volume	1.933	19:31
Air Blank	0.000	19:32
Control Test	0.079	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/10/2024 09:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PERSAUD,DEVEN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BULOW

DATE/TIME OF TEST: 11/17/2024 03:32:26

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINWRIGHT,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WAINWRIGHT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CARRIZAL

DATE/TIME OF TEST: 11/18/2024 02:34:40

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.147	02:28
Breath Volume	1.476	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.137	02:33
Breath Volume	1.281	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SALDIERA CHAVEZ

DATE/TIME OF TEST: 11/18/2024 05:04:18

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.124	04:59
Breath Volume	2.175	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.119	05:02
Breath Volume	2.292	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 11/18/2024 21:22:09

Diagnostics Check	OK	21:15
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.121	21:17
Breath Volume	1.687	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.104	21:20
Breath Volume	1.566	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: STALLINGS

DATE/TIME OF TEST: 11/21/2024 01:38:23

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
		01:37	RFI Detect
Air Blank	0.000	01:38	

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: STALLINGS

DATE/TIME OF TEST: 11/21/2024 01:48:48

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.101	01:43
Breath Volume	1.492	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.103	01:47
Breath Volume	1.972	01:47
Air Blank	0.000	01:47
Control Test	0.081	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: BARRICK

DATE/TIME OF TEST: 11/21/2024 23:05:12

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.082	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.263	23:00
Breath Volume	1.750	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.263	23:03
Breath Volume	1.691	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINWRIGHT,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MOFFAT

DATE/TIME OF TEST: 11/22/2024 01:18:26

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.080	01:04	
Air Blank	0.000	01:04	
Subject Sample	# 1 0.189	01:07	Volume Not Met
Breath Volume	0.386	01:07	
Air Blank	0.000	01:08	
Air Blank	0.000	01:10	
Subject Sample	# 2 0.206	01:12	
Breath Volume	1.105	01:12	
Air Blank	0.000	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 3 0.211	01:16	
Breath Volume	1.773	01:16	
Air Blank	0.000	01:17	
Control Test	0.079	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,EDSON,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHAVEZ,EDSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 11/22/2024 12:02:24

Diagnostics Check	OK	11:56
Air Blank	0.000	11:56
Control Test	0.081	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.077	11:57
Breath Volume	2.277	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.072	12:00
Breath Volume	1.343	12:00
Air Blank	0.000	12:01
Control Test	0.080	12:01
Air Blank	0.000	12:02
Diagnostics Check	OK	12:02

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ALMONTE DE LA CRUZ

DATE/TIME OF TEST: 11/22/2024 23:48:56

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:36	
Control Test	0.081	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1 0.148	23:40	Volume Not Met
Breath Volume	0.105	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2 0.155	23:43	
Breath Volume	1.433	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 3 0.156	23:47	
Breath Volume	1.242	23:47	
Air Blank	0.000	23:48	
Control Test	0.080	23:48	
Air Blank	0.000	23:48	
Diagnostics Check	OK	23:48	

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: CEPEDA-AYALA

DATE/TIME OF TEST: 11/23/2024 03:23:28

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.093	03:19
Breath Volume	2.144	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.090	03:21
Breath Volume	2.035	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HICIANO-BENITEZ

DATE/TIME OF TEST: 11/23/2024 05:36:46

Diagnostics Check	OK	05:25	
Air Blank	0.000	05:25	
Control Test	0.080	05:26	
Air Blank	0.000	05:26	
Subject Sample	# 1	05:29	No Sample Provided
Breath Volume	0.000	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 2	05:35	No Sample Provided
Breath Volume	0.000	05:35	
Air Blank	0.000	05:35	
Control Test	0.081	05:36	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: SANON

DATE/TIME OF TEST: 11/24/2024 03:04:27

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.109	03:00
Breath Volume	2.503	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.112	03:02
Breath Volume	2.277	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: GARCIA HERNANDEZ

DATE/TIME OF TEST: 11/24/2024 16:46:50

Diagnostics Check	OK	16:36
Air Blank	0.000	16:36
Control Test	0.080	16:37
Air Blank	0.000	16:37
Subject Sample # 1	0.305	16:38
Breath Volume	1.371	16:38
Air Blank	0.000	16:39
Air Blank	0.000	16:41
Subject Sample # 2	0.282	16:42
Breath Volume	1.351	16:42
Air Blank	0.000	16:42
Air Blank	0.000	16:44
Subject Sample # 3	0.320	16:45
Breath Volume	1.355	16:45
Air Blank	0.000	16:45
Control Test	0.079	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: CORTEZ MACEDO

DATE/TIME OF TEST: 11/24/2024 20:51:40

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.081	20:46
Air Blank	0.000	20:46
Subject Sample # 1	0.141	20:47
Breath Volume	1.460	20:47
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.141	20:50
Breath Volume	1.328	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 12/10/2024 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: LOPEZ JIMENEZ

DATE/TIME OF TEST: 11/24/2024 23:43:34

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.125	23:39
Breath Volume	1.984	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.129	23:42
Breath Volume	2.601	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 11/25/2024 00:32:24

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.000	00:27
Breath Volume	1.410	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.000	00:31
Breath Volume	1.257	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: ALMAGUER LOZOYA

DATE/TIME OF TEST: 11/28/2024 00:46:02

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Subject Sample # 1	0.150	00:41
Breath Volume	2.230	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.149	00:44
Breath Volume	1.976	00:44
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 11/28/2024 01:28:18

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.081	01:24	
Air Blank	0.000	01:25	
Subject Sample # 1	0.092	01:26	
Breath Volume	1.292	01:26	
Air Blank		01:27	Purge Fail
Air Blank	0.000	01:28	

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 11/28/2024 01:30:45

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.068	01:30	Control Outside Tolerance
Air Blank	0.000	01:30	

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ALFONSO MATOS

DATE/TIME OF TEST: 11/28/2024 21:58:01

Diagnostics Check	OK	21:49	
Air Blank	0.000	21:50	
Control Test	0.078	21:50	
Air Blank	0.000	21:50	
Subject Sample # 1	0.227	21:52	
Breath Volume	2.656	21:52	
Air Blank	0.000	21:53	
Air Blank	0.000	21:54	
Subject Sample # 2	0.216	21:56	
Breath Volume	1.757	21:56	
Air Blank	0.000	21:57	
Control Test	0.074	21:57	Control Outside Tolerance
Air Blank	0.000	21:58	

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CANO PACHECO

DATE/TIME OF TEST: 11/28/2024 23:39:49

Diagnostics Check	OK	23:33
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.215	23:35
Breath Volume	1.375	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.209	23:38
Breath Volume	1.183	23:38
Air Blank	0.000	23:38
Control Test	0.077	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 11/29/2024 03:25:18

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.338	03:20
Breath Volume	1.812	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.328	03:23
Breath Volume	2.144	03:23
Air Blank	0.000	03:24
Control Test	0.078	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: Riestra

DATE/TIME OF TEST: 11/30/2024 02:46:58

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.100	02:42
Breath Volume	1.652	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.096	02:45
Breath Volume	1.664	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 11/30/2024 03:57:20

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.147	03:52
Breath Volume	3.835	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.142	03:55
Breath Volume	2.703	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MARTELL GONZALEZ

DATE/TIME OF TEST: 11/30/2024 04:26:31

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.081	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:19	No Sample Provided
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 2	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: AVALOS LINARES

DATE/TIME OF TEST: 11/01/2024 02:52:24

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:47
Breath Volume	2.000	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:51
Breath Volume	1.925	02:51
Air Blank	0.000	02:51
Control Test	0.081	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 11/02/2024 01:44:14

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.266	01:38
Breath Volume	1.996	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.262	01:42
Breath Volume	2.078	01:42
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,Z
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LAMANDIA,V
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 11/02/2024 04:50:54

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.105	04:46
Breath Volume	3.796	04:46
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.104	04:49
Breath Volume	3.410	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: YALE

DATE/TIME OF TEST: 11/03/2024 21:39:39

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.175	21:35
Breath Volume	2.167	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.172	21:38
Breath Volume	1.546	21:38
Air Blank	0.000	21:38
Control Test	0.079	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: KETNER

DATE/TIME OF TEST: 11/04/2024 21:12:52

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.230	21:08
Breath Volume	1.359	21:08
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.237	21:11
Breath Volume	1.125	21:11
Air Blank	0.000	21:11
Control Test	0.078	21:12
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOKE,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CUNDIFFDROUSE,AMANDA
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: COLLIS

DATE/TIME OF TEST: 11/06/2024 04:56:41

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.000	04:52
Breath Volume	3.503	04:52
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.000	04:55
Breath Volume	2.921	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LAMANDIA,V
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 11/08/2024 01:43:51

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.104	01:38
Breath Volume	1.683	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.100	01:42
Breath Volume	2.046	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 11/20/2024 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 11/08/2024 05:22:54

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:11	No Sample Provided
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Air Blank	0.000	05:13	
Subject Sample	# 2	0.177	05:17
Breath Volume	1.359	05:17	
Air Blank	0.000	05:17	
Air Blank	0.000	05:19	
Subject Sample	# 3	0.164	05:21
Breath Volume	1.113	05:21	
Air Blank	0.000	05:21	
Control Test	0.078	05:22	
Air Blank	0.000	05:22	
Diagnostics Check	OK	05:22	

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 11/09/2024 00:40:04

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.166	00:35
Breath Volume	2.644	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.152	00:38
Breath Volume	1.898	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:40

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,DAVID
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: CARMAN

DATE/TIME OF TEST: 11/12/2024 15:42:30

Diagnostics Check	OK	15:35
Air Blank	0.000	15:36
Control Test	0.080	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.266	15:37
Breath Volume	1.507	15:37
Air Blank	0.000	15:38
Air Blank	0.000	15:40
Subject Sample # 2	0.281	15:40
Breath Volume	2.265	15:40
Air Blank	0.000	15:41
Control Test	0.078	15:41
Air Blank	0.000	15:42
Diagnostics Check	OK	15:42

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MAYFIELD

DATE/TIME OF TEST: 11/15/2024 01:15:36

Diagnostics Check	OK	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.029	01:10
Breath Volume	1.183	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.027	01:14
Breath Volume	1.445	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KENNY, EWAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: TUTTLE

DATE/TIME OF TEST: 11/15/2024 02:04:08

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.035	01:59
Breath Volume	3.113	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.035	02:02
Breath Volume	3.683	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: COTHRAN,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 11/16/2024 05:18:41

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:12
Air Blank	0.000	05:12
Subject Sample # 1	0.156	05:13
Breath Volume	2.257	05:13
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.161	05:17
Breath Volume	3.050	05:17
Air Blank	0.000	05:17
Control Test	0.080	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ADRIANNE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LECKY

DATE/TIME OF TEST: 11/18/2024 21:20:54

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Subject Sample # 1	0.051	21:16
Breath Volume	2.890	21:16
Air Blank	0.000	21:17
Air Blank	0.000	21:19
Subject Sample # 2	0.048	21:19
Breath Volume	2.582	21:19
Air Blank	0.000	21:19
Control Test	0.080	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 11/20/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MOHAN

DATE/TIME OF TEST: 11/20/2024 18:15:04

Diagnostics Check	OK	18:08
Air Blank	0.000	18:09
Control Test	0.079	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.267	18:10
Breath Volume	2.121	18:10
Air Blank	0.000	18:11
Air Blank	0.000	18:12
Subject Sample # 2	0.274	18:13
Breath Volume	2.113	18:13
Air Blank	0.000	18:14
Control Test	0.079	18:14
Air Blank	0.000	18:14
Diagnostics Check	OK	18:15

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/22/2024 03:56:11

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.067	03:51
Breath Volume	1.539	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.063	03:54
Breath Volume	1.289	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,STEVEN,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: TURNER,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 11/22/2024 18:23:55

Diagnostics Check	OK	18:17
Air Blank	0.000	18:18
Control Test	0.079	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.235	18:19
Breath Volume	2.570	18:19
Air Blank	0.000	18:20
Air Blank	0.000	18:21
Subject Sample # 2	0.234	18:22
Breath Volume	3.277	18:22
Air Blank	0.000	18:22
Control Test	0.080	18:23
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASPER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: JASPER,ANDREW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: LABAR

DATE/TIME OF TEST: 11/23/2024 23:27:45

Diagnostics Check	OK	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.000	23:23
Breath Volume	1.578	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.000	23:26
Breath Volume	2.367	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: GOMEZ VELASCO

DATE/TIME OF TEST: 11/24/2024 01:23:55

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.100	01:19
Breath Volume	1.914	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.095	01:22
Breath Volume	1.679	01:22
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GOURLEY

DATE/TIME OF TEST: 11/25/2024 23:00:19

Diagnostics Check	OK	22:55	
Air Blank	0.000	22:56	
Control Test	0.080	22:56	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:58	Subject Test Refused
Breath Volume	0.000	22:58	
Air Blank	0.000	22:59	
Control Test	0.080	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: BURNEY

DATE/TIME OF TEST: 11/27/2024 05:48:54

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:42
Air Blank	0.000	05:43
Subject Sample # 1	0.172	05:44
Breath Volume	2.027	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.173	05:47
Breath Volume	1.800	05:47
Air Blank	0.000	05:47
Control Test	0.079	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SNYDER,ADRIANNE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 11/28/2024 12:50:11

Diagnostics Check	OK	12:44
Air Blank	0.000	12:44
Control Test	0.079	12:44
Air Blank	0.000	12:45
Subject Sample # 1	0.151	12:45
Breath Volume	2.707	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample # 2	0.161	12:48
Breath Volume	2.359	12:48
Air Blank	0.000	12:49
Control Test	0.080	12:49
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: AVERA

DATE/TIME OF TEST: 11/29/2024 05:11:04

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.080	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.197	05:06
Breath Volume	1.167	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.186	05:09
Breath Volume	1.292	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 231603
BREATH TEST OPERATOR: BEAUREGARD,JEFF,	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: BEAUREGARD,JEFF	ARREST TIME: 19:22
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Female
LAST AGENCY INSPECTION: 10/01/2024	OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PARKINS	DATE/TIME OF TEST: 11/26/2024	20:37:41
------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:31
Subject Sample # 1	0.000	20:33
Breath Volume	1.476	20:33
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.000	20:36
Breath Volume	1.574	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 11/29/2024 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 11/22/2024 17:22:35

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.081	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.347	17:18
Breath Volume	2.175	17:18
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.352	17:21
Breath Volume	1.980	17:21
Air Blank	0.000	17:21
Control Test	0.081	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 12/06/2024 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILLIS,CALIB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 11/23/2024 17:31:10

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:27	
Control Test	0.081	17:27	
Air Blank	0.000	17:28	
Subject Sample	# 1	17:29	RFI Detect
Breath Volume	0.000	17:29	
Air Blank	0.000	17:30	
Control Test	0.082	17:30	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

Data Download Date/Time: 12/06/2024 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WILLIS,CALEB
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: ZAYAS

DATE/TIME OF TEST: 11/23/2024 17:38:20

Diagnostics Check	OK	17:35	
Air Blank	0.000	17:36	
Control Test	0.082	17:36	
Air Blank	0.000	17:36	
Subject Sample	# 1	17:36	Subject Test Refused
Breath Volume	0.000	17:36	
Air Blank	0.000	17:37	
Control Test	0.082	17:37	
Air Blank	0.000	17:38	
Diagnostics Check	OK	17:38	

Data Download Date/Time: 12/06/2024 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402328118
DRY GAS EXPIRY DATE: 01/20/2025
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 11/28/2024 11:37:00

Diagnostics Check	OK	11:30
Air Blank	0.000	11:30
Control Test	0.080	11:30
Air Blank	0.000	11:31
Subject Sample # 1	0.234	11:32
Breath Volume	2.273	11:32
Air Blank	0.000	11:32
Air Blank	0.000	11:34
Subject Sample # 2	0.233	11:35
Breath Volume	2.460	11:35
Air Blank	0.000	11:36
Control Test	0.080	11:36
Air Blank	0.000	11:36
Diagnostics Check	OK	11:37

Data Download Date/Time: 12/06/2024 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: LOPEZ,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: NIGHTINGALE

DATE/TIME OF TEST: 11/16/2024 00:26:58

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:21
Subject Sample # 1	0.164	00:22
Breath Volume	2.898	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.175	00:25
Breath Volume	2.781	00:25
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 12/03/2024 01:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALBRITTON, ROY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SOLNAY

DATE/TIME OF TEST: 11/30/2024 21:13:52

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:02	
Control Test	0.082	21:03	
Air Blank	0.000	21:03	
Subject Sample	# 1 0.126	21:06	Volume Not Met
Breath Volume	0.191	21:06	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2 0.109	21:12	Volume Not Met
Breath Volume	0.078	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 12/03/2024 01:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COOPER, WILLIAM, J

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: COOPER, WILLIAM

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 02:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 11/03/2024 02:45:30

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.109	02:39
Breath Volume	2.089	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.117	02:44
Breath Volume	3.148	02:44
Air Blank	0.000	02:44
Control Test	0.085	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: VALLABHANENI

DATE/TIME OF TEST: 11/04/2024 04:39:52

Diagnostics Check	OK	04:38	
Air Blank		04:39	Ambient Fail
Air Blank	0.000	04:39	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: COOPER, WILLIAM, J

BREATH TEST OPERATOR AGENCY: Deland PD

ARRESTING OFFICER: COOPER, WILLIAM

ARRESTING OFFICER AGENCY: Deland PD

LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1433248

DRY GAS EXPIRY DATE: 01/06/2024

VIOLATION CODE: DUI

ARREST TIME: 04:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: VALLABHANENI

DATE/TIME OF TEST: 11/04/2024 04:47:38

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.201	04:43
Breath Volume	2.125	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.187	04:46
Breath Volume	2.078	04:46
Air Blank	0.000	04:46
Control Test	0.082	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KELLY,JAVON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BURBAGE

DATE/TIME OF TEST: 11/18/2024 23:49:11

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.081	23:45	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.229	23:47	
Breath Volume	1.378	23:47	
Air Blank		23:48	Purge Fail
Air Blank	0.000	23:49	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KELLY,JAVON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BURBAGE

DATE/TIME OF TEST: 11/19/2024 00:22:46

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.082	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.209	00:18	
Breath Volume	1.710	00:18	
Air Blank	0.000	00:19	
Air Blank	0.000	00:21	
Subject Sample	# 2 0.190	00:21	
Breath Volume	1.437	00:21	
Air Blank		00:22	RFI Detect
Air Blank	0.000	00:22	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KELLY,JAVON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BURBAGE

DATE/TIME OF TEST: 11/19/2024 00:34:22

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.199	00:29
Breath Volume	1.824	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.190	00:32
Breath Volume	1.906	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: FRANKLIN,JACOB
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 11/19/2024 23:20:10

Diagnostics Check	OK	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.203	23:15
Breath Volume	3.078	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.207	23:18
Breath Volume	2.898	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: GARDNER, JAVON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FELTON

DATE/TIME OF TEST: 11/26/2024 00:57:27

Diagnostics Check	OK	00:51	
Air Blank	0.000	00:51	
Control Test	0.083	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.031	00:55	
Air Blank	0.000	00:56	
Control Test	0.086	00:56	Control Outside Tolerance
Air Blank	0.000	00:57	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KELLY,JAVON
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MELENDEZ MUNOZ

DATE/TIME OF TEST: 11/28/2024 03:43:53

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.082	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: KAZEE

DATE/TIME OF TEST: 11/29/2024 23:46:02

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.082	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.082	23:45	
Air Blank	0.000	23:45	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 10/31/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SALVIA

DATE/TIME OF TEST: 11/30/2024 02:56:51

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.082	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.233	02:52
Breath Volume	2.906	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample # 2	0.247	02:55
Breath Volume	3.644	02:55
Air Blank	0.000	02:55
Control Test	0.083	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 11/30/2024 16:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: DOTSON,DAVID
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 10/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: COHEE

DATE/TIME OF TEST: 11/06/2024 23:29:08

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.079	23:28	
Air Blank		23:28	RFI Detect
Air Blank		23:29	RFI Detect

Data Download Date/Time: 11/23/2024 19:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: DOTSON,DAVID
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 10/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: COHEE

DATE/TIME OF TEST: 11/06/2024 23:42:07

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.184	23:37
Breath Volume	2.414	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.178	23:40
Breath Volume	2.003	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 11/23/2024 19:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ZECHARIAH,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CONWAY,JACOB
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 10/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: CONNOR

DATE/TIME OF TEST: 11/09/2024 07:44:35

Diagnostics Check	OK	07:38	
Air Blank	0.000	07:39	
Control Test	0.081	07:39	
Air Blank	0.000	07:40	
Subject Sample	# 1	07:43	No Sample Provided
Breath Volume	0.000	07:43	
Air Blank	0.000	07:43	
Air Blank	0.000	07:44	

Data Download Date/Time: 11/28/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRACH,S,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 11/18/2024 02:22:59

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.081	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.057	02:18
Breath Volume	2.109	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.054	02:21
Breath Volume	2.894	02:21
Air Blank	0.000	02:22
Control Test	0.081	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/28/2024 08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,KRISTOPHER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 11/20/2024 21:51:44

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.081	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:47
Breath Volume	3.515	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:50
Breath Volume	3.566	21:50
Air Blank	0.000	21:50
Control Test	0.082	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 11/28/2024 08:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ ALGARIN, NICOLE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER, LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SAMAYOA PEREZ

DATE/TIME OF TEST: 11/03/2024 01:01:45

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.187	00:57
Breath Volume	1.449	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.191	01:00
Breath Volume	1.457	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 11/12/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: AKINYOOYE,ADETOKUNBO
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WOODBURY

DATE/TIME OF TEST: 11/03/2024 03:06:09

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.189	03:01
Breath Volume	2.648	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.188	03:04
Breath Volume	2.082	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 11/12/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 11/04/2024 13:52:33

Diagnostics Check	OK	13:46
Air Blank	0.000	13:46
Control Test	0.080	13:47
Air Blank	0.000	13:47
Subject Sample # 1	0.192	13:48
Breath Volume	1.160	13:48
Air Blank	0.000	13:48
Air Blank	0.000	13:50
Subject Sample # 2	0.185	13:51
Breath Volume	1.339	13:51
Air Blank	0.000	13:51
Control Test	0.079	13:51
Air Blank	0.000	13:52
Diagnostics Check	OK	13:52

Data Download Date/Time: 11/12/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTIE,JULIE
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:34

SUBJECT'S LAST NAME: MACSHANE

DATE/TIME OF TEST: 11/06/2024 11:10:59

Diagnostics Check	OK	10:59	
Air Blank	0.000	11:00	
Control Test	0.078	11:00	
Air Blank	0.000	11:00	
Subject Sample	# 1	11:04	No Sample Provided
Breath Volume	0.000	11:04	
Air Blank	0.000	11:04	
Air Blank	0.000	11:06	
Subject Sample	# 2	11:09	No Sample Provided
Breath Volume	0.000	11:09	
Air Blank	0.000	11:10	
Control Test	0.079	11:10	
Air Blank	0.000	11:10	
Diagnostics Check	OK	11:10	

Data Download Date/Time: 11/12/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DADEY

DATE/TIME OF TEST: 11/07/2024 00:01:01

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.081	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.162	23:56
Breath Volume	1.367	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.162	23:59
Breath Volume	1.507	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:01

Data Download Date/Time: 11/12/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: MELGAR CHAVEZ

DATE/TIME OF TEST: 11/09/2024 20:41:25

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.079	20:30	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:34	No Sample Provided
Breath Volume	0.000	20:34	
Air Blank	0.000	20:34	
Air Blank	0.000	20:36	
Subject Sample	# 2	20:40	No Sample Provided
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.079	20:40	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 11/12/2024 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREWS, TYREE, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT, JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: LANG

DATE/TIME OF TEST: 11/15/2024 03:10:43

Diagnostics Check	OK	02:59	
Air Blank	0.000	02:59	
Control Test	0.080	03:00	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:03	No Sample Provided
Breath Volume	0.000	03:03	
Air Blank	0.000	03:04	
Air Blank	0.000	03:06	
Subject Sample	# 2	03:09	No Sample Provided
Breath Volume	0.000	03:09	
Air Blank	0.000	03:09	
Control Test	0.079	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 12/10/2024 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 11/16/2024 02:58:38

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.220	02:54
Breath Volume	1.632	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.209	02:57
Breath Volume	1.484	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 12/10/2024 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 11/20/2024 05:06:07

Diagnostics Check	OK	04:54	
Air Blank	0.000	04:55	
Control Test	0.081	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:59	No Sample Provided
Breath Volume	0.000	04:59	
Air Blank	0.000	04:59	
Air Blank	0.000	05:01	
Subject Sample	# 2	05:04	No Sample Provided
Breath Volume	0.000	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 12/10/2024 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARENCO, ALFREDO, E
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: PERALTA, ALLISON
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 11/20/2024 23:06:20

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:50	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Subject Sample	# 1 0.121	22:53	
Breath Volume	1.437	22:53	
Air Blank	0.000	22:54	
Air Blank	0.000	22:55	
Subject Sample	# 2 0.101	22:59	Volume Not Met
Breath Volume	0.015	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 3 0.108	23:04	Volume Not Met
Breath Volume	0.671	23:04	
Air Blank	0.000	23:05	
Control Test	0.077	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/21/2024 08:24:56

Diagnostics Check	OK	08:14	
Air Blank	0.000	08:14	
Control Test	0.080	08:15	
Air Blank	0.000	08:15	
Subject Sample	# 1 0.081	08:18	Volume Not Met
Breath Volume	0.218	08:18	
Air Blank	0.000	08:19	
Air Blank	0.000	08:21	
Subject Sample	# 2 0.101	08:24	Volume Not Met
Breath Volume	0.324	08:24	
Air Blank		08:24	RFI Detect
Air Blank	0.000	08:24	

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 11/21/2024 08:38:16

Diagnostics Check	OK	08:30
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Subject Sample # 1	0.103	08:33
Breath Volume	1.386	08:33
Air Blank	0.000	08:34
Air Blank	0.000	08:36
Subject Sample # 2	0.101	08:36
Breath Volume	1.445	08:36
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Diagnostics Check	OK	08:38

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BOVEN,AUSTIN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 11/22/2024 01:31:00

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 11/28/2024 01:39:47

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.093	01:35
Breath Volume	1.644	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.091	01:38
Breath Volume	1.253	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ALFONSO MATOS

DATE/TIME OF TEST: 11/28/2024 22:11:29

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.217	22:07
Breath Volume	1.398	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.236	22:10
Breath Volume	1.671	22:10
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKLEY, RONALD, L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BRACKLEY, RONALD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BRUDER

DATE/TIME OF TEST: 11/29/2024 01:42:41

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.107	01:38
Breath Volume	2.949	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.104	01:41
Breath Volume	3.101	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ZHENG

DATE/TIME OF TEST: 11/29/2024 04:23:33

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.118	04:19
Breath Volume	1.937	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.120	04:22
Breath Volume	2.089	04:22
Air Blank	0.000	04:22
Control Test	0.080	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FARLESS, MITCHELL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:41

SUBJECT'S LAST NAME: BUURMA

DATE/TIME OF TEST: 11/30/2024 07:24:26

Diagnostics Check	OK	07:18
Air Blank	0.000	07:18
Control Test	0.081	07:18
Air Blank	0.000	07:19
Subject Sample # 1	0.169	07:19
Breath Volume	1.769	07:19
Air Blank	0.000	07:20
Air Blank	0.000	07:22
Subject Sample # 2	0.184	07:22
Breath Volume	1.722	07:22
Air Blank	0.000	07:23
Control Test	0.080	07:23
Air Blank	0.000	07:24
Diagnostics Check	OK	07:24

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,NICKOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WELLS,KYLE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 11/11/2024 14:47:50

Diagnostics Check	OK	14:37
Air Blank	0.000	14:37
Control Test	0.080	14:37
Air Blank	0.000	14:38
Subject Sample # 1	0.262	14:39
Breath Volume	2.324	14:39
Air Blank	0.000	14:39
Air Blank	0.000	14:41
Subject Sample # 2	0.233	14:43
Breath Volume	1.988	14:43
Air Blank	0.000	14:43
Air Blank	0.000	14:45
Subject Sample # 3	0.274	14:46
Breath Volume	1.542	14:46
Air Blank	0.000	14:46
Control Test	0.079	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 11/26/2024 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BANKSTON

DATE/TIME OF TEST: 11/15/2024 01:58:12

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.204	01:53
Breath Volume	1.105	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.199	01:56
Breath Volume	1.312	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 11/26/2024 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDGEComb,STEVEN,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BABA,TANNER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 11/17/2024 06:15:18

Diagnostics Check	OK	06:03	
Air Blank	0.000	06:04	
Control Test	0.080	06:04	
Air Blank	0.000	06:04	
Subject Sample	# 1 0.146	06:08	Volume Not Met
Breath Volume	0.542	06:08	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 2 0.154	06:13	Volume Not Met
Breath Volume	0.101	06:13	
Air Blank	0.000	06:14	
Control Test	0.077	06:14	
Air Blank	0.000	06:15	
Diagnostics Check	OK	06:15	

Data Download Date/Time: 11/26/2024 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 11/18/2024 21:59:07

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.173	21:54
Breath Volume	2.203	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.170	21:57
Breath Volume	1.414	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 11/26/2024 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: MERIZIER

DATE/TIME OF TEST: 11/22/2024 19:23:05

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.080	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.000	19:18
Breath Volume	1.753	19:18
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.000	19:21
Breath Volume	1.902	19:21
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 11/26/2024 19:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTINEZ,ZACHARY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:59

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 11/27/2024 12:31:19

Diagnostics Check	OK	12:24
Air Blank	0.000	12:24
Control Test	0.080	12:25
Air Blank	0.000	12:25
Subject Sample # 1	0.243	12:26
Breath Volume	1.925	12:26
Air Blank	0.000	12:26
Air Blank	0.000	12:28
Subject Sample # 2	0.253	12:29
Breath Volume	2.382	12:29
Air Blank	0.000	12:30
Control Test	0.080	12:30
Air Blank	0.000	12:31
Diagnostics Check	OK	12:31

Data Download Date/Time: 12/16/2024 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALETTO,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MALETTO,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: MEISTER

DATE/TIME OF TEST: 11/27/2024 20:20:59

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.010	20:16
Breath Volume	2.011	20:16
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.010	20:19
Breath Volume	2.515	20:19
Air Blank	0.000	20:20
Control Test	0.080	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 12/16/2024 16:32