

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 01524080A4
BREATH TEST OPERATOR: JONES,DEVIN,K	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Bradford County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: JONES,DEVIN	ARREST TIME: 11:03
ARRESTING OFFICER AGENCY: Bradford County SO	GENDER: Male
LAST AGENCY INSPECTION: 11/08/2024	OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: WILLIAMS	DATE/TIME OF TEST: 12/01/2024 12:27:53
-------------------------------	--

Diagnostics Check	OK	12:20
Air Blank	0.000	12:21
Control Test	0.080	12:21
Air Blank	0.000	12:22
Subject Sample # 1	0.000	12:23
Breath Volume	4.324	12:23
Air Blank	0.000	12:23
Air Blank	0.000	12:25
Subject Sample # 2	0.000	12:26
Breath Volume	4.570	12:26
Air Blank	0.000	12:27
Control Test	0.080	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 01/03/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 01524080A4
BREATH TEST OPERATOR: KONKEL,SCOTT,M	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Bradford County SO	VIOLATION CODE: DUI
ARRESTING OFFICER: ESCOBAR,ANGEL	ARREST TIME: 21:32
ARRESTING OFFICER AGENCY: Bradford County SO	GENDER: Female
LAST AGENCY INSPECTION: 11/08/2024	OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LACSON	DATE/TIME OF TEST: 12/14/2024 22:24:05
-----------------------------	--

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.080	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:19	No Sample Provided
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.080	22:23	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 01/03/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDINER,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: KONKEL,GEORGE
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 12/23/2024 12:59:46

Diagnostics Check	OK	12:53
Air Blank	0.000	12:54
Control Test	0.080	12:54
Air Blank	0.000	12:55
Subject Sample # 1	0.000	12:55
Breath Volume	4.757	12:55
Air Blank	0.000	12:56
Air Blank	0.000	12:58
Subject Sample # 2	0.000	12:58
Breath Volume	4.019	12:58
Air Blank	0.000	12:58
Control Test	0.079	12:59
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 01/03/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PALMER,ALAN,A

BREATH TEST OPERATOR AGENCY: FHP Troop G

ARRESTING OFFICER: PALMER,ALAN

ARRESTING OFFICER AGENCY: FHP Troop G

LAST AGENCY INSPECTION: 11/08/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01524080A4

DRY GAS EXPIRY DATE: 03/05/2026

VIOLATION CODE: DUI

ARREST TIME: 01:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/28/2024 01:26:58

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.082	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.162	01:22
Breath Volume	1.972	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.165	01:25
Breath Volume	1.457	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/03/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 12/10/2024 20:09:41

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.078	20:03
Air Blank	0.000	20:04
Subject Sample # 1	0.195	20:04
Breath Volume	1.605	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:06
Subject Sample # 2	0.186	20:08
Breath Volume	1.992	20:08
Air Blank	0.000	20:08
Control Test	0.078	20:09
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 12/13/2024 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 12/12/2024 02:04:13

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.138	01:59
Breath Volume	2.378	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.137	02:02
Breath Volume	1.882	02:02
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/13/2024 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: ACEVEDO VELEZ

DATE/TIME OF TEST: 12/13/2024 02:09:48

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.218	02:05
Breath Volume	1.492	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.216	02:08
Breath Volume	1.453	02:08
Air Blank	0.000	02:08
Control Test	0.077	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/13/2024 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TARR,BRITNEY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: ALFORD

DATE/TIME OF TEST: 12/13/2024 07:37:52

Diagnostics Check	OK	07:30
Air Blank	0.000	07:31
Control Test	0.079	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.210	07:32
Breath Volume	1.464	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:35
Subject Sample # 2	0.205	07:36
Breath Volume	1.773	07:36
Air Blank	0.000	07:36
Control Test	0.076	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 12/13/2024 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: QUINTANA,STEPHANIE
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: TORRES OTERO

DATE/TIME OF TEST: 12/13/2024 20:03:20

Diagnostics Check	OK	19:47	
Air Blank	0.000	19:48	
Control Test	0.078	19:48	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:52	No Sample Provided
Breath Volume	0.000	19:52	
Air Blank	0.000	19:52	
Air Blank	0.000	19:54	
Subject Sample	# 2	0.150	19:57
Breath Volume	1.554	19:57	
Air Blank	0.000	19:58	
Air Blank	0.000	20:00	
Subject Sample	# 3	0.159	20:01
Breath Volume	1.187	20:01	
Air Blank	0.000	20:02	
Control Test	0.077	20:02	
Air Blank	0.000	20:03	
Diagnostics Check	OK	20:03	

Data Download Date/Time: 12/13/2024 20:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

INSTRUMENT REGISTERED TO: FDLE

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: POST,HAYDEN,L

BREATH TEST OPERATOR AGENCY: Leesburg PD

ARRESTING OFFICER: POST,HAYDEN

ARRESTING OFFICER AGENCY: Leesburg PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 08223080A5

DRY GAS EXPIRY DATE: 06/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/07/2024 22:32:50

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.083	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 12/12/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

INSTRUMENT REGISTERED TO: FDLE	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 24923080A3
BREATH TEST OPERATOR: JONES,AUSTIN,G	DRY GAS EXPIRY DATE: 10/05/2025
BREATH TEST OPERATOR AGENCY: Everglades Nat Park	VIOLATION CODE: ADMIN
ARRESTING OFFICER: JONES,AUSTIN	ARREST TIME: 19:00
ARRESTING OFFICER AGENCY: Everglades Nat Park	GENDER: Male
LAST AGENCY INSPECTION: 12/17/2024	OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: KING	DATE/TIME OF TEST: 12/25/2024 00:35:08
---------------------------	--

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.108	00:30
Breath Volume	2.570	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.090	00:33
Breath Volume	1.640	00:33
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 01/24/2025 13:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: JOHNS,NEVIN
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 12/01/2024 02:24:59

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.083	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.078	02:19
Breath Volume	1.199	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.076	02:23
Breath Volume	1.335	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PAYNE,KALEB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/08/2024 03:48:01

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.083	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.192	03:43
Breath Volume	3.066	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.198	03:46
Breath Volume	1.519	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: BAKER,WILLIAM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: EILAND

DATE/TIME OF TEST: 12/10/2024 00:45:45

Diagnostics Check	OK	00:39
Air Blank	0.000	00:40
Control Test	0.082	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:41
Breath Volume	1.347	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	1.894	00:44
Air Blank	0.000	00:44
Control Test	0.083	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO,KENNETH,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GADDY 2

DATE/TIME OF TEST: 12/10/2024 18:29:12

Diagnostics Check	OK	18:26	
Air Blank	0.000	18:26	
Control Test	0.082	18:27	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:27	Subject Test Refused
Breath Volume	0.000	18:27	
Air Blank	0.000	18:28	
Control Test	0.082	18:28	
Air Blank	0.000	18:29	
Diagnostics Check	OK	18:29	

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,LIBBEE,S
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: WETZEL,BRODY
ARRESTING OFFICER AGENCY: Alachua PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ORR

DATE/TIME OF TEST: 12/13/2024 04:08:14

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.084	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.088	04:03
Breath Volume	1.355	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.085	04:06
Breath Volume	1.839	04:06
Air Blank	0.000	04:07
Control Test	0.083	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/13/2024 18:09:17

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.082	18:03
Air Blank	0.000	18:04
Subject Sample # 1	0.239	18:04
Breath Volume	2.531	18:04
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.240	18:07
Breath Volume	2.000	18:07
Air Blank	0.000	18:08
Control Test	0.082	18:08
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TOBY, B
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ALEXANDER, DENZEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 12/13/2024 23:46:11

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.083	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.166	23:41
Breath Volume	3.375	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.158	23:44
Breath Volume	2.539	23:44
Air Blank	0.000	23:45
Control Test	0.082	23:45
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: FINCH,JEFFERY
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:52

SUBJECT'S LAST NAME: BALESTRIERI

DATE/TIME OF TEST: 12/14/2024 15:27:22

Diagnostics Check	OK	15:24	
Air Blank	0.000	15:24	
Control Test	0.082	15:25	
Air Blank	0.000	15:25	
Subject Sample	# 1	15:25	Subject Test Refused
Breath Volume	0.000	15:25	
Air Blank	0.000	15:26	
Control Test	0.083	15:26	
Air Blank	0.000	15:27	
Diagnostics Check	OK	15:27	

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/15/2024 07:12:01

Diagnostics Check	OK	07:05
Air Blank	0.000	07:05
Control Test	0.082	07:05
Air Blank	0.000	07:06
Subject Sample # 1	0.154	07:07
Breath Volume	2.156	07:07
Air Blank	0.000	07:07
Air Blank	0.000	07:09
Subject Sample # 2	0.153	07:10
Breath Volume	1.781	07:10
Air Blank	0.000	07:11
Control Test	0.082	07:11
Air Blank	0.000	07:11
Diagnostics Check	OK	07:12

Data Download Date/Time: 12/26/2024 14:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: TILLMAN,JOE
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:11

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 12/16/2024 14:38:38

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.081	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.268	14:33
Breath Volume	3.277	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.283	14:37
Breath Volume	3.433	14:37
Air Blank	0.000	14:37
Control Test	0.082	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEYES,GABRIEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SALTER,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 12/17/2024 05:23:06

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.083	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1	05:21	Subject Test Refused
Breath Volume	0.000	05:21	
Air Blank	0.000	05:22	
Control Test	0.082	05:22	
Air Blank	0.000	05:23	
Diagnostics Check	OK	05:23	

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: STAFFORD,ROBERT
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: MARTORELLA

DATE/TIME OF TEST: 12/18/2024 21:43:26

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.082	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.082	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITZKOW,CARLTON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: DELUCA,C
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: BARTSCH

DATE/TIME OF TEST: 12/19/2024 21:56:09

Diagnostics Check	OK	21:44	
Air Blank	0.000	21:45	
Control Test	0.083	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.083	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ, KRISHNA, H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK, ISAAC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEELER

DATE/TIME OF TEST: 12/20/2024 04:17:14

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.083	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.161	04:12
Breath Volume	1.277	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.158	04:15
Breath Volume	2.042	04:15
Air Blank	0.000	04:16
Control Test	0.082	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: UPSHAW,JASON
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CICARELLO

DATE/TIME OF TEST: 12/20/2024 21:24:12

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.083	21:17
Air Blank	0.000	21:18
Subject Sample # 1	0.208	21:19
Breath Volume	2.781	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.204	21:22
Breath Volume	1.511	21:22
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Diagnostics Check	OK	21:24

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 12/20/2024 23:12:52

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.082	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.238	23:08
Breath Volume	2.363	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:10
Subject Sample # 2	0.225	23:11
Breath Volume	2.062	23:11
Air Blank	0.000	23:11
Control Test	0.082	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 12/26/2024 14:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: LADNER

DATE/TIME OF TEST: 12/14/2024 02:02:52

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.078	01:53	
Air Blank	0.000	01:54	
Subject Sample	# 1	01:57	No Sample Provided
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.079	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,MATHEW,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: REEDER,GRACEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/20/2024 20:57:28

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.150	20:53
Breath Volume	2.210	20:53
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.145	20:55
Breath Volume	1.359	20:55
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUCH,DERRICK,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: COUCH

DATE/TIME OF TEST: 12/21/2024 02:05:09

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.065	02:00
Breath Volume	1.601	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.068	02:03
Breath Volume	2.976	02:03
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECK,JIMMIE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: THOMASON,TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 12/21/2024 03:12:48

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.078	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRETT,DONALD,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: G630191604270

DATE/TIME OF TEST: 12/22/2024 03:18:24

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZANT,TANNER,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ZANT,TANNER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783527
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 13:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 12/23/2024 15:10:17

Diagnostics Check	OK	15:04
Air Blank	0.000	15:04
Control Test	0.079	15:05
Air Blank	0.000	15:05
Subject Sample # 1	0.053	15:05
Breath Volume	2.234	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.050	15:08
Breath Volume	2.394	15:08
Air Blank	0.000	15:09
Control Test	0.078	15:09
Air Blank	0.000	15:10
Diagnostics Check	OK	15:10

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ATKINS,JAMES,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ATKINS,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/24/2024 01:21:45

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:17
Breath Volume	1.941	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.000	01:20
Breath Volume	2.988	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,SHANE,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/26/2024 02:52:53

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.078	02:50	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.078	02:52	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTSON,XAVIER,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEDER,GRACEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 12/26/2024 21:07:08

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:00	
Control Test	0.078	21:01	
Air Blank	0.000	21:01	
Subject Sample	# 1 0.000	21:02	
Breath Volume	2.738	21:02	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2 0.000	21:05	
Breath Volume	1.808	21:05	
Air Blank	0.000	21:06	
Control Test	0.079	21:06	
Air Blank		21:07	RFI Detect
Air Blank		21:07	RFI Detect

Data Download Date/Time: 12/27/2024 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,TREVOR,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: THOMASON,TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/31/2024 18:03:40

Diagnostics Check	OK	18:00	
Air Blank	0.000	18:01	
Control Test	0.078	18:01	
Air Blank	0.000	18:01	
Subject Sample	# 1	18:02	Subject Test Refused
Breath Volume	0.000	18:02	
Air Blank	0.000	18:02	
Control Test	0.077	18:03	
Air Blank	0.000	18:03	
Diagnostics Check	OK	18:03	

Data Download Date/Time: 01/30/2025 17:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: AYALA ANZURES

DATE/TIME OF TEST: 12/19/2024 23:51:01

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.080	23:38	
Air Blank	0.000	23:39	
Subject Sample	# 1 0.161	23:42	Volume Not Met
Breath Volume	0.507	23:42	
Air Blank	0.000	23:42	
Air Blank	0.000	23:44	
Subject Sample	# 2 0.174	23:45	
Breath Volume	4.082	23:45	
Air Blank	0.000	23:46	
Air Blank	0.000	23:48	
Subject Sample	# 3 0.165	23:49	
Breath Volume	1.851	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:50	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 01/17/2025 19:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDRANO-UGALDE,VANESSA,Y
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: CRYER,ABBIE
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MEDRANO-UGALDE

DATE/TIME OF TEST: 12/20/2024 04:18:23

Diagnostics Check	OK	04:11
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.081	04:13
Breath Volume	2.632	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.080	04:16
Breath Volume	1.800	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 01/17/2025 19:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GLENN

DATE/TIME OF TEST: 12/22/2024 00:20:03

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.298	00:15
Breath Volume	2.132	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:17
Subject Sample # 2	0.299	00:18
Breath Volume	2.492	00:18
Air Blank	0.000	00:19
Control Test	0.080	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:20

Data Download Date/Time: 01/17/2025 19:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: STUPAR

DATE/TIME OF TEST: 12/23/2024 01:45:06

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.081	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 01/17/2025 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORCE,TAYLOR,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VAN PELT,COLTON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JOSEY

DATE/TIME OF TEST: 12/04/2024 02:17:18

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.133	02:12
Breath Volume	1.671	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.124	02:15
Breath Volume	1.742	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/10/2024 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRE, ADAM, J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: FULTON, DAKOTA
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/07/2024 00:21:20

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.000	00:17
Breath Volume	1.785	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.000	00:19
Breath Volume	1.390	00:19
Air Blank	0.000	00:20
Control Test	0.081	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/10/2024 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,CHAD,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LYNCH,CHAD
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 12/09/2024 04:57:11

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.204	04:52
Breath Volume	2.039	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.206	04:55
Breath Volume	2.031	04:55
Air Blank	0.000	04:56
Control Test	0.078	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 12/10/2024 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLENDER, MONICA, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MILLENDER, MONICA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: FLOWERS

DATE/TIME OF TEST: 12/18/2024 16:05:32

Diagnostics Check	OK	15:59
Air Blank	0.000	16:00
Control Test	0.080	16:00
Air Blank	0.000	16:00
Subject Sample # 1	0.000	16:01
Breath Volume	3.054	16:01
Air Blank	0.000	16:01
Air Blank	0.000	16:03
Subject Sample # 2	0.000	16:04
Breath Volume	3.609	16:04
Air Blank	0.000	16:04
Control Test	0.081	16:04
Air Blank	0.000	16:05
Diagnostics Check	OK	16:05

Data Download Date/Time: 02/13/2025 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MABEE

DATE/TIME OF TEST: 12/19/2024 01:07:21

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.064	01:02
Breath Volume	1.144	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.069	01:05
Breath Volume	1.582	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 02/13/2025 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 12/24/2024 21:05:19

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.183	20:59
Breath Volume	1.523	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.187	21:03
Breath Volume	1.367	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 02/13/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402690226
DRY GAS EXPIRY DATE: 03/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: LANEHART

DATE/TIME OF TEST: 12/25/2024 19:42:16

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.075	19:37
Breath Volume	1.906	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.075	19:40
Breath Volume	2.234	19:40
Air Blank	0.000	19:41
Control Test	0.081	19:41
Air Blank	0.000	19:42
Diagnostics Check	OK	19:42

Data Download Date/Time: 02/13/2025 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: LEBER,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/28/2024 00:41:59

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.081	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.138	00:37
Breath Volume	1.179	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.124	00:40
Breath Volume	1.574	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 01/07/2025 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAMER,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINCK,ISAAC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 12/29/2024 03:58:28

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.103	03:54
Breath Volume	2.433	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.107	03:56
Breath Volume	2.539	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/07/2025 09:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000760

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UPSHAW,JASON,D
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MORRIS,ERIC
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: EMMONS

DATE/TIME OF TEST: 12/31/2024 04:06:05

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.190	03:58
Breath Volume	1.488	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.218	04:01
Breath Volume	1.183	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 3	0.204	04:04
Breath Volume	2.214	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/07/2025 09:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGUIRE,KATHLEEN,T
BREATH TEST OPERATOR AGENCY: FL Dept of AG
ARRESTING OFFICER: HOWELL,DANIEL
ARRESTING OFFICER AGENCY: FL Dept of AG
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: CABRERA GARCIA

DATE/TIME OF TEST: 12/01/2024 06:56:58

Diagnostics Check	OK	06:50
Air Blank	0.000	06:50
Control Test	0.079	06:51
Air Blank	0.000	06:51
Subject Sample # 1	0.204	06:52
Breath Volume	2.496	06:52
Air Blank	0.000	06:53
Air Blank	0.000	06:54
Subject Sample # 2	0.203	06:55
Breath Volume	2.054	06:55
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:56
Diagnostics Check	OK	06:56

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL SASSO, MICHAEL, W
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: RECCA, MICHAEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 12/02/2024 17:37:09

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:34	
Control Test	0.080	17:34	
Air Blank	0.000	17:34	
Subject Sample # 1	0.385	17:35	
Breath Volume	2.132	17:35	
Air Blank		17:36	Purge Fail
Air Blank	0.000	17:37	

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL SASSO, MICHAEL, W
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: RECCA, MICHAEL
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: STEFAN

DATE/TIME OF TEST: 12/02/2024 18:00:53

Diagnostics Check	OK	17:56	
Air Blank	0.000	17:57	
Control Test	0.080	17:57	
Air Blank	0.000	17:58	
Subject Sample # 1	0.372	17:59	
Breath Volume	2.941	17:59	
Air Blank		18:00	Purge Fail
Air Blank	0.000	18:00	

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSKO,KRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: KOSKO,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/06/2024 02:02:23

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.115	01:57
Breath Volume	1.773	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.111	02:00
Breath Volume	1.507	02:00
Air Blank	0.000	02:01
Control Test	0.077	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARAFA,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TARAFA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/09/2024 02:42:12

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.180	02:37
Breath Volume	3.875	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.174	02:40
Breath Volume	3.832	02:40
Air Blank	0.000	02:41
Control Test	0.077	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURR,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DURR,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/11/2024 20:49:23

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.079	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.198	20:44
Breath Volume	2.230	20:44
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.194	20:47
Breath Volume	2.554	20:47
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: LAW,A
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 12/13/2024 01:48:17

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.227	01:41
Breath Volume	4.078	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:43
Subject Sample # 2	0.229	01:46
Breath Volume	3.351	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: WETHY

DATE/TIME OF TEST: 12/15/2024 00:54:19

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.079	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.136	00:49
Breath Volume	4.871	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.136	00:52
Breath Volume	4.031	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MACFADDEN,R
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: STACEY

DATE/TIME OF TEST: 12/23/2024 05:08:48

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.157	05:04
Breath Volume	4.359	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.155	05:07
Breath Volume	3.933	05:07
Air Blank	0.000	05:07
Control Test	0.078	05:08
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000768

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 12/25/2024 02:52:27

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:48
Breath Volume	1.671	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:51
Breath Volume	1.800	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 12/27/2024 14:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 12/03/2024 20:49:06

Diagnostics Check	OK	20:48	
Air Blank		20:48	RFI Detect
Air Blank	0.000	20:49	

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 12/03/2024 20:52:06

Diagnostics Check	OK	20:51	
Air Blank		20:51	RFI Detect
Air Blank	0.000	20:52	

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 12/03/2024 20:55:07

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:54	
Control Test	0.079	20:54	RFI Detect
Air Blank	0.000	20:55	

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 12/03/2024 21:00:08

Diagnostics Check	OK	20:57	
Air Blank	0.000	20:57	
Control Test	0.080	20:57	
Air Blank	0.000	20:58	
Subject Sample	# 1	20:58	RFI Detect
Breath Volume	1.574	20:58	
Air Blank	0.000	20:59	
Control Test	0.078	20:59	RFI Detect
Air Blank	0.000	21:00	

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: HEATON

DATE/TIME OF TEST: 12/03/2024 21:31:42

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.080	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.171	21:26
Breath Volume	1.589	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.158	21:30
Breath Volume	1.964	21:30
Air Blank	0.000	21:30
Control Test	0.076	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: WENTWORTH

DATE/TIME OF TEST: 12/04/2024 00:39:57

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 12/25/2024 09:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRYDEN,JOHN,E
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRYDEN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: MONTAGUE

DATE/TIME OF TEST: 12/04/2024 11:46:44

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:41
Air Blank	0.000	11:41
Subject Sample # 1	0.072	11:42
Breath Volume	1.523	11:42
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.061	11:45
Breath Volume	1.503	11:45
Air Blank	0.000	11:45
Control Test	0.078	11:46
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 12/25/2024 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARAFA,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TARAFA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STALVEY

DATE/TIME OF TEST: 12/08/2024 01:13:35

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.126	01:09
Breath Volume	3.566	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.130	01:12
Breath Volume	2.894	01:12
Air Blank	0.000	01:12
Control Test	0.075	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/25/2024 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORTON,CORNELIUS,Q
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: LORTON,CORNELIUS
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 12/08/2024 03:13:54

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.203	03:09
Breath Volume	1.660	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.195	03:12
Breath Volume	2.109	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/25/2024 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEMMONS,L,M
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PLEMMONS,L
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 12/08/2024 23:38:19

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:35	
Control Test	0.079	23:36	
Air Blank	0.000	23:36	
Subject Sample	# 1	23:36	Subject Test Refused
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 12/25/2024 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARAFA,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TARAFA,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 12/13/2024 01:29:21

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.108	01:25
Breath Volume	2.773	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.100	01:27
Breath Volume	1.640	01:27
Air Blank	0.000	01:28
Control Test	0.077	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/25/2024 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,DYLAN,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BINES,DYLAN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/26/2024 04:10:09

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1 0.103	03:57	
Breath Volume	1.136	03:57	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2 0.088	04:02	Volume Not Met
Breath Volume	0.410	04:02	
Air Blank	0.000	04:03	
Air Blank	0.000	04:05	
Subject Sample	# 3 0.094	04:08	Volume Not Met
Breath Volume	0.855	04:08	
Air Blank	0.000	04:09	
Control Test	0.077	04:09	
Air Blank	0.000	04:10	
Diagnostics Check	OK	04:10	

Data Download Date/Time: 01/28/2025 15:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 12/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/29/2024 03:13:45

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.212	03:09
Breath Volume	3.062	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.205	03:12
Breath Volume	2.414	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 01/25/2025 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARX,CHRISTOPHER,P
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 12/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MAHURIN

DATE/TIME OF TEST: 12/30/2024 21:39:29

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.081	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.250	21:31
Breath Volume	3.382	21:31
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.225	21:34
Breath Volume	1.671	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 3	0.233	21:37
Breath Volume	2.710	21:37
Air Blank	0.000	21:38
Control Test	0.079	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 01/25/2025 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: RANDIG,MATTHEW
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 12/05/2024 01:13:29

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.081	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.174	01:08
Breath Volume	1.402	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.179	01:11
Breath Volume	2.359	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 12/28/2024 00:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,JERRY,W
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: PEASE,OLIVER
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/07/2024 19:32:20

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.104	19:27
Breath Volume	1.199	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.119	19:30
Breath Volume	2.343	19:30
Air Blank	0.000	19:31
Control Test	0.081	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 12/28/2024 00:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,PAUL,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: YOUNG,PAUL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 12/08/2024 04:38:59

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Subject Sample	# 1	04:37	Subject Test Refused
Breath Volume	0.000	04:37	
Air Blank	0.000	04:38	
Control Test	0.080	04:38	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 12/28/2024 00:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMMER,DALTON,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: BRADSHAW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402946336
DRY GAS EXPIRY DATE: 01/17/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MUHLENKAMP

DATE/TIME OF TEST: 12/24/2024 00:18:08

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.192	00:13
Breath Volume	3.378	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.194	00:16
Breath Volume	3.359	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 12/28/2024 00:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ACKER,AARON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: ENNEN

DATE/TIME OF TEST: 12/10/2024 17:19:05

Diagnostics Check	OK	17:12
Air Blank	0.000	17:12
Control Test	0.081	17:12
Air Blank	0.000	17:13
Subject Sample # 1	0.251	17:14
Breath Volume	2.757	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:16
Subject Sample # 2	0.251	17:17
Breath Volume	2.726	17:17
Air Blank	0.000	17:18
Control Test	0.080	17:18
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 12/26/2024 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: ROULHAC

DATE/TIME OF TEST: 12/23/2024 06:26:19

Diagnostics Check	OK	06:20
Air Blank	0.000	06:20
Control Test	0.083	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.144	06:21
Breath Volume	1.199	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.142	06:24
Breath Volume	1.304	06:24
Air Blank	0.000	06:25
Control Test	0.082	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 12/26/2024 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000782

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Washington County SO
ARRESTING OFFICER: THOMPSON,JOSEPH
ARRESTING OFFICER AGENCY: Washington County SO
LAST AGENCY INSPECTION: 12/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 12/30/2024 13:15:31

Diagnostics Check	OK	13:09
Air Blank	0.000	13:10
Control Test	0.081	13:10
Air Blank	0.000	13:10
Subject Sample # 1	0.000	13:11
Breath Volume	2.027	13:11
Air Blank	0.000	13:11
Air Blank	0.000	13:13
Subject Sample # 2	0.000	13:14
Breath Volume	2.722	13:14
Air Blank	0.000	13:14
Control Test	0.081	13:14
Air Blank	0.000	13:15
Diagnostics Check	OK	13:15

Data Download Date/Time: 01/25/2025 11:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 12/02/2024 19:14:06

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.059	19:09
Breath Volume	3.648	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:11
Subject Sample # 2	0.060	19:12
Breath Volume	3.140	19:12
Air Blank	0.000	19:13
Control Test	0.080	19:13
Air Blank	0.000	19:14
Diagnostics Check	OK	19:14

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/07/2024 08:28:20

Diagnostics Check	OK	08:17	
Air Blank	0.000	08:18	
Control Test	0.080	08:18	
Air Blank	0.000	08:19	
Subject Sample	# 1 0.207	08:19	
Breath Volume	4.234	08:19	
Air Blank	0.000	08:20	
Air Blank	0.000	08:22	
Subject Sample	# 2 0.231	08:23	
Breath Volume	4.781	08:23	
Air Blank	0.000	08:24	
Air Blank	0.000	08:25	
Subject Sample	# 3 0.227	08:26	
Breath Volume	4.296	08:26	
Air Blank	0.000	08:27	
Control Test	0.074	08:27	Control Outside Tolerance
Air Blank	0.000	08:28	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/07/2024 08:44:27

Diagnostics Check	OK	08:37
Air Blank	0.000	08:38
Control Test	0.076	08:38
Air Blank	0.000	08:39
Subject Sample # 1	0.203	08:39
Breath Volume	2.500	08:39
Air Blank	0.000	08:40
Air Blank	0.000	08:42
Subject Sample # 2	0.196	08:42
Breath Volume	2.480	08:42
Air Blank	0.000	08:43
Control Test	0.075	08:43
Air Blank	0.000	08:44
Diagnostics Check	OK	08:44

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 12/09/2024 22:26:03

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.081	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1 0.278	22:13	
Breath Volume	1.312	22:13	
Air Blank	0.000	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 2	22:18	No Sample Provided
Breath Volume	0.000	22:18	
Air Blank	0.000	22:19	
Air Blank	0.000	22:21	
Subject Sample	# 3	22:24	No Sample Provided
Breath Volume	0.000	22:24	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 12/15/2024 21:19:22

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:14	
Control Test	0.080	21:14	
Air Blank	0.000	21:14	
Subject Sample	# 1 0.232	21:18	Volume Not Met
Breath Volume	1.035	21:18	
Air Blank		21:18	Purge Fail
Air Blank	0.000	21:19	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF,MICHAEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 12/15/2024 21:22:52

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.073	21:22	Control Outside Tolerance
Air Blank	0.000	21:22	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADCLIFF, MICHAEL, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RADCLIFF, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 12/15/2024 21:36:29

Diagnostics Check	OK	21:24	
Air Blank	0.000	21:24	
Control Test	0.077	21:25	
Air Blank	0.000	21:25	
Subject Sample	# 1 0.226	21:26	
Breath Volume	1.671	21:26	
Air Blank	0.000	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 2 0.247	21:30	
Breath Volume	1.417	21:30	
Air Blank	0.000	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 3 0.231	21:34	
Breath Volume	1.453	21:34	
Air Blank	0.000	21:35	
Control Test	0.074	21:35	Control Outside Tolerance
Air Blank	0.000	21:36	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000784

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREEMAN,AUSTIN,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FREEMAN,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LIEURANCE

DATE/TIME OF TEST: 12/20/2024 00:26:01

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:22	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:24	Subject Test Refused
Breath Volume	0.000	00:24	
Air Blank	0.000	00:25	
Control Test	0.081	00:25	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:26	

Data Download Date/Time: 01/03/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: SELLERS,DONALD
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: VINING

DATE/TIME OF TEST: 12/07/2024 00:58:33

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.077	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.166	00:54
Breath Volume	4.261	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.170	00:57
Breath Volume	2.589	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/16/2024 17:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,DONALD,R
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: DAVIDSON,MICHAEL
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402367487
DRY GAS EXPIRY DATE: 04/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/21/2024 23:02:07

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.077	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.067	22:57
Breath Volume	1.144	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.067	23:00
Breath Volume	1.363	23:00
Air Blank	0.000	23:01
Control Test	0.077	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 01/27/2025 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,SHANE,C
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SMITH,JACOB
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: ASARO

DATE/TIME OF TEST: 12/03/2024 18:21:23

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.082	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.141	18:16
Breath Volume	2.085	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.139	18:19
Breath Volume	1.507	18:19
Air Blank	0.000	18:20
Control Test	0.081	18:20
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 12/24/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 04:53

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: AHUMADA

DATE/TIME OF TEST: 12/22/2024 05:34:19

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.081	05:28
Air Blank	0.000	05:28
Subject Sample # 1	0.149	05:29
Breath Volume	1.976	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.143	05:32
Breath Volume	1.140	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 12/24/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: EARLY,DAVID

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 22:07

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 12/22/2024 22:50:52

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.080	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.000	22:46
Breath Volume	2.019	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.000	22:49
Breath Volume	1.898	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 12/24/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GOSIER,MATTHEW,D

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: GOSIER,MATTHEW

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 12/24/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:48

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/27/2024 22:36:29

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.082	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.189	22:31
Breath Volume	1.468	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.194	22:34
Breath Volume	1.660	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/20/2025 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/02/2024 23:16:34

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.082	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1 0.282	23:11	
Breath Volume	1.816	23:11	
Air Blank	0.000	23:12	
Air Blank	0.000	23:14	
Subject Sample	# 2 0.280	23:15	
Breath Volume	2.640	23:15	
Air Blank	0.000	23:16	
Control Test	0.074	23:16	Control Outside Tolerance
Air Blank		23:16	RFI Detect

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCATO,SHAWN,V
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: BROCATO,SHAWN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/02/2024 23:55:01

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.285	23:50
Breath Volume	2.394	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.282	23:53
Breath Volume	2.335	23:53
Air Blank	0.000	23:54
Control Test	0.075	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:55

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/03/2024 03:43:15

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.082	03:40	
Air Blank	0.000	03:40	
Subject Sample # 1	0.146	03:41	
Breath Volume	3.628	03:41	
Air Blank		03:41	RFI Detect
Air Blank	0.000	03:42	
Control Test	0.083	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/03/2024 03:53:06

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.149	03:48
Breath Volume	4.382	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.141	03:51
Breath Volume	4.355	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIGG,JACOB,M
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: MAURANCY,FRANNY
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: SANTIZO

DATE/TIME OF TEST: 12/04/2024 05:37:15

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.082	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.190	05:32
Breath Volume	2.046	05:32
Air Blank	0.000	05:33
Air Blank	0.000	05:35
Subject Sample # 2	0.176	05:35
Breath Volume	2.703	05:35
Air Blank	0.000	05:36
Control Test	0.080	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITES,DEJAS,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: FORMELIO,JONATHAN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/05/2024 21:38:40

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.151	21:34
Breath Volume	1.742	21:34
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.154	21:37
Breath Volume	1.535	21:37
Air Blank	0.000	21:37
Control Test	0.081	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEROY,TAD,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BUCHKO,B
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: ADMIN
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TUFTE

DATE/TIME OF TEST: 12/07/2024 21:11:45

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.082	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.098	21:07
Breath Volume	2.394	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.100	21:10
Breath Volume	2.382	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERTIN,SKYLAR,M
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: ARTECHE,ROBERT
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: MARINEZ-NAVARRO

DATE/TIME OF TEST: 12/11/2024 22:31:56

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.082	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	Subject Test Refused
Breath Volume	0.000	22:30	
Air Blank	0.000	22:31	
Control Test	0.082	22:31	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MOSNER

DATE/TIME OF TEST: 12/17/2024 00:24:45

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:09	
Control Test	0.082	00:10	
Air Blank	0.000	00:10	
Subject Sample # 1	0.117	00:12	
Breath Volume	1.437	00:12	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample # 2	0.114	00:17	Volume Not Met
Breath Volume	0.238	00:17	
Air Blank	0.000	00:18	
Air Blank	0.000	00:20	
Subject Sample # 3	0.107	00:22	
Breath Volume	1.113	00:22	
Air Blank	0.000	00:23	
Control Test	0.077	00:24	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSERLY,STEPHEN,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BENNING,CHEYENNE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: LEVANAS

DATE/TIME OF TEST: 12/17/2024 03:05:19

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.126	03:00
Breath Volume	1.359	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.137	03:03
Breath Volume	2.558	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 12/19/2024 00:28:49

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.038	00:24
Breath Volume	2.320	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.036	00:27
Breath Volume	1.882	00:27
Air Blank	0.000	00:27
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILKEY,JOSHUA,A
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LAPOLLA,BRANDEN
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: CORDONERO CASTELLON

DATE/TIME OF TEST: 12/22/2024 15:59:25

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.082	15:52
Air Blank	0.000	15:53
Subject Sample # 1	0.207	15:54
Breath Volume	2.660	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.192	15:57
Breath Volume	1.488	15:57
Air Blank	0.000	15:58
Control Test	0.077	15:58
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: DAMAS CRUZ

DATE/TIME OF TEST: 12/23/2024 18:07:00

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.082	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.120	18:02
Breath Volume	3.781	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.112	18:05
Breath Volume	3.453	18:05
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:07

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PERKINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ELKASMI

DATE/TIME OF TEST: 12/28/2024 20:31:34

Diagnostics Check	OK	20:30	
Air Blank		20:31	RFI Detect
Air Blank	0.000	20:31	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PERKINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ELKASMI

DATE/TIME OF TEST: 12/28/2024 20:41:36

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.082	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1 0.000	20:37	
Breath Volume	2.285	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:40	
Subject Sample	# 2	20:40	RFI Detect
Breath Volume	0.000	20:40	
Air Blank	0.000	20:40	
Control Test	0.081	20:41	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PERKINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: ELKASMI

DATE/TIME OF TEST: 12/28/2024 20:51:58

Diagnostics Check	OK	20:45
Air Blank	0.000	20:46
Control Test	0.082	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.000	20:47
Breath Volume	1.523	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.000	20:50
Breath Volume	2.082	20:50
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,LAURIE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LANIER,BRANDEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 12/29/2024 16:22:41

Diagnostics Check	OK	16:20	
Air Blank	0.000	16:20	
Control Test	0.082	16:20	
Air Blank	0.000	16:21	
Subject Sample	# 1	16:22	Sequence Aborted
Breath Volume	0.000	16:22	
Air Blank	0.000	16:22	

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,LAURIE,B
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: LANIER,BRANDEN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 11/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 12/29/2024 16:44:40

Diagnostics Check	OK	16:38
Air Blank	0.000	16:38
Control Test	0.082	16:39
Air Blank	0.000	16:39
Subject Sample # 1	0.144	16:40
Breath Volume	2.324	16:40
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.127	16:43
Breath Volume	1.714	16:43
Air Blank	0.000	16:43
Control Test	0.081	16:44
Air Blank	0.000	16:44
Diagnostics Check	OK	16:44

Data Download Date/Time: 12/30/2024 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEMON,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Fort Pierce PD
ARRESTING OFFICER: MEMON,MATTHEW
ARRESTING OFFICER AGENCY: Fort Pierce PD
LAST AGENCY INSPECTION: 12/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PASCUAL-JIMENEZ

DATE/TIME OF TEST: 12/31/2024 01:04:02

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.082	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.205	00:59
Breath Volume	3.453	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.186	01:02
Breath Volume	2.187	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 01/22/2025 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICHOLS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Port St. Joe PD
ARRESTING OFFICER: RANDIG,MATTHEW
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 12/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402607186
DRY GAS EXPIRY DATE: 12/02/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/28/2024 22:57:29

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	2.402	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	1.855	22:56
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 02/13/2025 21:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 12/05/2024 10:04:41

Diagnostics Check	OK	09:59	
Air Blank	0.000	09:59	
Control Test	0.082	10:00	
Air Blank	0.000	10:00	
Subject Sample # 1	0.000	10:01	
Breath Volume	1.851	10:01	
Air Blank	0.000	10:01	
Air Blank	0.000	10:03	
Subject Sample # 2		10:04	RFI Detect
Breath Volume	0.000	10:04	
Air Blank		10:04	RFI Detect
Air Blank	0.000	10:04	

Data Download Date/Time: 12/09/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 09:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 12/05/2024 10:14:30

Diagnostics Check	OK	10:08
Air Blank	0.000	10:08
Control Test	0.084	10:09
Air Blank	0.000	10:09
Subject Sample # 1	0.000	10:10
Breath Volume	2.164	10:10
Air Blank	0.000	10:10
Air Blank	0.000	10:12
Subject Sample # 2	0.000	10:13
Breath Volume	2.164	10:13
Air Blank	0.000	10:13
Control Test	0.082	10:13
Air Blank	0.000	10:14
Diagnostics Check	OK	10:14

Data Download Date/Time: 12/09/2024 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/05/2024 10:36:23

Diagnostics Check	OK	10:30
Air Blank	0.000	10:30
Control Test	0.082	10:30
Air Blank	0.000	10:31
Subject Sample # 1	0.000	10:32
Breath Volume	1.921	10:32
Air Blank	0.000	10:32
Air Blank	0.000	10:34
Subject Sample # 2	0.000	10:35
Breath Volume	2.007	10:35
Air Blank	0.000	10:35
Control Test	0.080	10:35
Air Blank	0.000	10:36
Diagnostics Check	OK	10:36

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/05/2024 14:55:51

Diagnostics Check	OK	14:49
Air Blank	0.000	14:50
Control Test	0.083	14:50
Air Blank	0.000	14:51
Subject Sample # 1	0.000	14:51
Breath Volume	1.398	14:51
Air Blank	0.000	14:52
Air Blank	0.000	14:54
Subject Sample # 2	0.000	14:54
Breath Volume	1.664	14:54
Air Blank	0.000	14:54
Control Test	0.081	14:55
Air Blank	0.000	14:55
Diagnostics Check	OK	14:55

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:36

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/06/2024 09:10:48

Diagnostics Check	OK	09:02	
Air Blank	0.000	09:02	
Control Test	0.081	09:03	
Air Blank	0.000	09:03	
Subject Sample	# 1	09:06	No Sample Provided
Breath Volume	0.000	09:06	
Air Blank	0.000	09:07	
Air Blank	0.000	09:09	
Subject Sample	# 2	09:09	Subject Test Refused
Breath Volume	0.000	09:09	
Air Blank	0.000	09:09	
Control Test	0.082	09:10	
Air Blank	0.000	09:10	
Diagnostics Check	OK	09:10	

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 12/06/2024 09:37:24

Diagnostics Check	OK	09:31
Air Blank	0.000	09:32
Control Test	0.082	09:32
Air Blank	0.000	09:32
Subject Sample # 1	0.000	09:33
Breath Volume	1.714	09:33
Air Blank	0.000	09:33
Air Blank	0.000	09:35
Subject Sample # 2	0.000	09:36
Breath Volume	2.093	09:36
Air Blank	0.000	09:36
Control Test	0.082	09:36
Air Blank	0.000	09:37
Diagnostics Check	OK	09:37

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/06/2024 09:55:12

Diagnostics Check	OK	09:49
Air Blank	0.000	09:49
Control Test	0.082	09:50
Air Blank	0.000	09:50
Subject Sample # 1	0.000	09:51
Breath Volume	1.691	09:51
Air Blank	0.000	09:51
Air Blank	0.000	09:53
Subject Sample # 2	0.000	09:53
Breath Volume	1.703	09:53
Air Blank	0.000	09:54
Control Test	0.081	09:54
Air Blank	0.000	09:55
Diagnostics Check	OK	09:55

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 12/06/2024 10:46:57

Diagnostics Check	OK	10:41
Air Blank	0.000	10:41
Control Test	0.081	10:41
Air Blank	0.000	10:42
Subject Sample # 1	0.000	10:42
Breath Volume	1.964	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.000	10:45
Breath Volume	1.976	10:45
Air Blank	0.000	10:46
Control Test	0.081	10:46
Air Blank	0.000	10:46
Diagnostics Check	OK	10:46

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:27

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/06/2024 10:59:48

Diagnostics Check	OK	10:53
Air Blank	0.000	10:54
Control Test	0.082	10:54
Air Blank	0.000	10:55
Subject Sample # 1	0.000	10:55
Breath Volume	3.050	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:58
Breath Volume	2.605	10:58
Air Blank	0.000	10:58
Control Test	0.081	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 11:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/06/2024 12:01:54

Diagnostics Check	OK	11:55
Air Blank	0.000	11:56
Control Test	0.082	11:56
Air Blank	0.000	11:57
Subject Sample # 1	0.000	11:57
Breath Volume	1.457	11:57
Air Blank	0.000	11:58
Air Blank	0.000	12:00
Subject Sample # 2	0.000	12:00
Breath Volume	1.742	12:00
Air Blank	0.000	12:01
Control Test	0.081	12:01
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MCLAUGHLIN,JUSTIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: D'AMOUR

DATE/TIME OF TEST: 12/06/2024 14:13:21

Diagnostics Check	OK	14:07
Air Blank	0.000	14:07
Control Test	0.082	14:08
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:09
Breath Volume	2.351	14:09
Air Blank	0.000	14:09
Air Blank	0.000	14:11
Subject Sample # 2	0.000	14:12
Breath Volume	3.027	14:12
Air Blank	0.000	14:12
Control Test	0.082	14:12
Air Blank	0.000	14:13
Diagnostics Check	OK	14:13

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/07/2024 09:47:34

Diagnostics Check	OK	09:35	
Air Blank	0.000	09:36	
Control Test	0.082	09:36	
Air Blank	0.000	09:37	
Subject Sample	# 1 0.193	09:40	Volume Not Met
Breath Volume	0.050	09:40	
Air Blank	0.000	09:41	
Air Blank	0.000	09:42	
Subject Sample	# 2 0.000	09:45	Volume Not Met
Breath Volume	0.031	09:45	
Air Blank	0.000	09:46	
Control Test	0.080	09:46	
Air Blank	0.000	09:47	
Diagnostics Check	OK	09:47	

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:59

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/07/2024 10:07:51

Diagnostics Check	OK	09:56	
Air Blank	0.000	09:56	
Control Test	0.081	09:56	
Air Blank	0.000	09:57	
Subject Sample	# 1 0.233	10:00	Volume Not Met
Breath Volume	0.070	10:00	
Air Blank	0.000	10:01	
Air Blank	0.000	10:03	
Subject Sample	# 2 0.241	10:06	Volume Not Met
Breath Volume	0.152	10:06	
Air Blank	0.000	10:06	
Control Test	0.080	10:07	
Air Blank	0.000	10:07	
Diagnostics Check	OK	10:07	

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: VELEZ LEON

DATE/TIME OF TEST: 12/09/2024 14:23:15

Diagnostics Check	OK	14:20	
Air Blank	0.000	14:20	
Control Test	0.082	14:20	
Air Blank	0.000	14:21	
Subject Sample	# 1	14:21	RFI Detect
Breath Volume	0.000	14:21	
Air Blank	0.000	14:22	
Control Test	0.081	14:22	
Air Blank		14:22	RFI Detect
Air Blank	0.000	14:23	

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: COURT ORD
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: VELEZ LEON

DATE/TIME OF TEST: 12/09/2024 14:32:06

Diagnostics Check	OK	14:26
Air Blank	0.000	14:26
Control Test	0.082	14:26
Air Blank	0.000	14:27
Subject Sample # 1	0.000	14:27
Breath Volume	2.136	14:27
Air Blank	0.000	14:28
Air Blank	0.000	14:30
Subject Sample # 2	0.000	14:30
Breath Volume	1.195	14:30
Air Blank	0.000	14:31
Control Test	0.081	14:31
Air Blank	0.000	14:32
Diagnostics Check	OK	14:32

Data Download Date/Time: 12/09/2024 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MCLAUGHLIN,JUSTIN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 12/12/2024 15:52:25

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.084	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.000	15:48
Breath Volume	2.226	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.000	15:51
Breath Volume	2.226	15:51
Air Blank	0.000	15:51
Control Test	0.083	15:51
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMAN,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 12/15/2024 03:58:49

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.083	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.140	03:53
Breath Volume	2.074	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.143	03:57
Breath Volume	1.539	03:57
Air Blank	0.000	03:57
Control Test	0.082	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HICKS,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: CARRION RODRIGUEZ

DATE/TIME OF TEST: 12/15/2024 17:44:17

Diagnostics Check	OK	17:35
Air Blank	0.000	17:36
Control Test	0.083	17:36
Air Blank	0.000	17:37
Subject Sample # 1	0.000	17:39
Breath Volume	1.914	17:39
Air Blank	0.000	17:40
Air Blank	0.000	17:42
Subject Sample # 2	0.000	17:42
Breath Volume	2.718	17:42
Air Blank	0.000	17:43
Control Test	0.081	17:43
Air Blank	0.000	17:44
Diagnostics Check	OK	17:44

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: DOUGLAS,JORDAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GROLL

DATE/TIME OF TEST: 12/16/2024 01:34:09

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.083	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.000	01:29
Breath Volume	1.378	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.000	01:32
Breath Volume	1.175	01:32
Air Blank	0.000	01:33
Control Test	0.082	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/16/2024 23:23:34

Diagnostics Check	OK	23:17
Air Blank	0.000	23:17
Control Test	0.082	23:18
Air Blank	0.000	23:18
Subject Sample # 1	0.000	23:19
Breath Volume	2.214	23:19
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.000	23:22
Breath Volume	2.675	23:22
Air Blank	0.000	23:22
Control Test	0.081	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MANZANO MELENDEZ

DATE/TIME OF TEST: 12/19/2024 09:16:22

Diagnostics Check	OK	09:10
Air Blank	0.000	09:10
Control Test	0.082	09:11
Air Blank	0.000	09:11
Subject Sample # 1	0.000	09:12
Breath Volume	2.132	09:12
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.000	09:15
Breath Volume	1.652	09:15
Air Blank	0.000	09:15
Control Test	0.080	09:15
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 12/19/2024 09:46:19

Diagnostics Check	OK	09:40
Air Blank	0.000	09:40
Control Test	0.082	09:41
Air Blank	0.000	09:41
Subject Sample # 1	0.000	09:41
Breath Volume	2.339	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.000	09:44
Breath Volume	2.515	09:44
Air Blank	0.000	09:45
Control Test	0.081	09:45
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 12/19/2024 10:12:59

Diagnostics Check	OK	10:10	
Air Blank	0.000	10:11	
Control Test	0.082	10:11	
Air Blank	0.000	10:12	
Subject Sample	# 1	10:12	RFI Detect
Breath Volume	0.816	10:12	
Air Blank		10:12	RFI Detect
Air Blank	0.000	10:12	

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: STEIN

DATE/TIME OF TEST: 12/19/2024 10:21:37

Diagnostics Check	OK	10:14
Air Blank	0.000	10:15
Control Test	0.082	10:15
Air Blank	0.000	10:16
Subject Sample # 1	0.000	10:16
Breath Volume	1.960	10:16
Air Blank	0.000	10:17
Air Blank	0.000	10:19
Subject Sample # 2	0.000	10:20
Breath Volume	1.132	10:20
Air Blank	0.000	10:20
Control Test	0.081	10:21
Air Blank	0.000	10:21
Diagnostics Check	OK	10:21

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: TOSADO MERCADO

DATE/TIME OF TEST: 12/19/2024 13:55:44

Diagnostics Check	OK	13:49
Air Blank	0.000	13:49
Control Test	0.082	13:50
Air Blank	0.000	13:50
Subject Sample # 1	0.000	13:51
Breath Volume	1.210	13:51
Air Blank	0.000	13:51
Air Blank	0.000	13:53
Subject Sample # 2	0.000	13:54
Breath Volume	1.332	13:54
Air Blank	0.000	13:54
Control Test	0.081	13:55
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS, EVANS, J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS, STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/19/2024 14:11:52

Diagnostics Check	Fail	14:11	Internal Printer Test
Air Blank	0.000	14:11	

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/19/2024 14:16:00

Diagnostics Check	Fail	14:15	Internal Printer Test
Air Blank	0.000	14:16	

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/19/2024 14:25:39

Diagnostics Check	OK	14:19
Air Blank	0.000	14:19
Control Test	0.082	14:20
Air Blank	0.000	14:20
Subject Sample # 1	0.000	14:21
Breath Volume	1.605	14:21
Air Blank	0.000	14:21
Air Blank	0.000	14:23
Subject Sample # 2	0.000	14:24
Breath Volume	1.167	14:24
Air Blank	0.000	14:24
Control Test	0.081	14:25
Air Blank	0.000	14:25
Diagnostics Check	OK	14:25

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 14:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: TOULOUTE

DATE/TIME OF TEST: 12/19/2024 15:04:53

Diagnostics Check	OK	14:57
Air Blank	0.000	14:58
Control Test	0.082	14:58
Air Blank	0.000	14:59
Subject Sample # 1	0.000	15:00
Breath Volume	1.425	15:00
Air Blank	0.000	15:00
Air Blank	0.000	15:02
Subject Sample # 2	0.000	15:03
Breath Volume	1.207	15:03
Air Blank	0.000	15:03
Control Test	0.080	15:04
Air Blank	0.000	15:04
Diagnostics Check	OK	15:04

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/19/2024 16:03:52

Diagnostics Check	OK	15:58
Air Blank	0.000	15:58
Control Test	0.082	15:58
Air Blank	0.000	15:59
Subject Sample # 1	0.000	15:59
Breath Volume	1.664	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:02
Subject Sample # 2	0.000	16:02
Breath Volume	2.476	16:02
Air Blank	0.000	16:02
Control Test	0.082	16:03
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 15:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 12/19/2024 16:18:21

Diagnostics Check	OK	16:12
Air Blank	0.000	16:12
Control Test	0.081	16:13
Air Blank	0.000	16:13
Subject Sample # 1	0.000	16:14
Breath Volume	2.023	16:14
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.000	16:16
Breath Volume	1.636	16:16
Air Blank	0.000	16:17
Control Test	0.081	16:17
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 01/02/2025 06:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: HUGHEY

DATE/TIME OF TEST: 12/20/2024 13:38:14

Diagnostics Check	OK	13:31
Air Blank	0.000	13:32
Control Test	0.082	13:32
Air Blank	0.000	13:32
Subject Sample # 1	0.000	13:33
Breath Volume	1.316	13:33
Air Blank	0.000	13:33
Air Blank	0.000	13:35
Subject Sample # 2	0.000	13:36
Breath Volume	1.589	13:36
Air Blank	0.000	13:37
Control Test	0.082	13:37
Air Blank	0.000	13:38
Diagnostics Check	OK	13:38

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/22/2024 03:06:04

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:02	
Control Test	0.083	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.082	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAUGHLIN,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/25/2024 09:43:50

Diagnostics Check	OK	09:37
Air Blank	0.000	09:38
Control Test	0.083	09:38
Air Blank	0.000	09:38
Subject Sample # 1	0.160	09:39
Breath Volume	2.152	09:39
Air Blank	0.000	09:40
Air Blank	0.000	09:41
Subject Sample # 2	0.164	09:42
Breath Volume	2.246	09:42
Air Blank	0.000	09:42
Control Test	0.082	09:43
Air Blank	0.000	09:43
Diagnostics Check	OK	09:43

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/26/2024 20:52:57

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:51	
Control Test	0.083	20:52	
Air Blank		20:52	RFI Detect
Air Blank	0.000	20:52	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/26/2024 20:56:24

Diagnostics Check	Fail	20:55	Internal Printer Test
Air Blank	0.000	20:56	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/26/2024 20:59:33

Diagnostics Check	Fail	20:59	Internal Printer Test
Air Blank	0.000	20:59	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/26/2024 21:13:06

Diagnostics Check	OK	21:01	
Air Blank	0.000	21:02	
Control Test	0.083	21:02	
Air Blank	0.000	21:03	
Subject Sample	# 1 0.232	21:06	Volume Not Met
Breath Volume	0.917	21:06	
Air Blank	0.000	21:07	
Air Blank	0.000	21:08	
Subject Sample	# 2	21:11	Subject Test Refused
Breath Volume	0.320	21:11	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HICKS,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: RODRIGUEZ CINTRON

DATE/TIME OF TEST: 12/27/2024 15:31:49

Diagnostics Check	OK	15:30	
Air Blank	0.000	15:31	
Control Test	0.080	15:31	RFI Detect
Air Blank	0.000	15:31	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HICKS,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: RODRIGUEZ CINTRON

DATE/TIME OF TEST: 12/27/2024 15:48:05

Diagnostics Check	OK	15:34	
Air Blank	0.000	15:34	
Control Test	0.082	15:35	
Air Blank	0.000	15:35	
Subject Sample	# 1 0.180	15:38	Volume Not Met
Breath Volume	1.402	15:38	
Air Blank	0.000	15:39	
Air Blank	0.000	15:41	
Subject Sample	# 2 0.196	15:41	
Breath Volume	2.085	15:41	
Air Blank	0.000	15:42	
Air Blank	0.000	15:44	
Subject Sample	# 3 0.177	15:46	
Breath Volume	3.246	15:46	
Air Blank	0.000	15:47	
Control Test	0.079	15:47	
Air Blank	0.000	15:48	
Diagnostics Check	OK	15:48	

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 06/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BAHENA

DATE/TIME OF TEST: 12/29/2024 04:14:40

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.098	04:10
Breath Volume	1.296	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:12
Subject Sample # 2	0.096	04:13
Breath Volume	1.105	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 01/02/2025 06:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CRAIG,J
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: ANDERSON,CRAIG
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 12/15/2024 00:59:15

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.213	00:53
Breath Volume	1.417	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.212	00:57
Breath Volume	2.625	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/23/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000811

INSTRUMENT REGISTERED TO: PERRY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BETHEA,COLBY,R

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BETHEA,COLBY

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 20:30

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: PENDELL

DATE/TIME OF TEST: 12/16/2024 21:21:10

Diagnostics Check	OK	21:14
Air Blank	0.000	21:15
Control Test	0.077	21:15
Air Blank	0.000	21:15
Subject Sample # 1	0.226	21:16
Breath Volume	2.660	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.213	21:19
Breath Volume	1.691	21:19
Air Blank	0.000	21:20
Control Test	0.078	21:20
Air Blank	0.000	21:21
Diagnostics Check	OK	21:21

Data Download Date/Time: 12/23/2024 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: LOPEZ MOLINA

DATE/TIME OF TEST: 12/29/2024 01:37:08

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.082	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.168	01:32
Breath Volume	2.417	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.155	01:35
Breath Volume	1.136	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 01/30/2025 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 12/29/2024 04:06:05

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.083	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.153	04:01
Breath Volume	1.855	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.148	04:04
Breath Volume	1.250	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 01/30/2025 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREMBLAY,F,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: TREMBLAY,F
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 12/29/2024 20:41:03

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.082	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 01/30/2025 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GREER,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: CAGLE

DATE/TIME OF TEST: 12/29/2024 23:42:09

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:38	
Control Test	0.082	23:39	
Air Blank	0.000	23:39	
Subject Sample	# 1	23:40	Subject Test Refused
Breath Volume	0.000	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Diagnostics Check	OK	23:42	

Data Download Date/Time: 01/30/2025 02:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CLAYTON,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: RADEMACHER

DATE/TIME OF TEST: 12/31/2024 22:37:15

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.143	22:32
Breath Volume	1.921	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.149	22:35
Breath Volume	1.636	22:35
Air Blank	0.000	22:36
Control Test	0.082	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 01/30/2025 02:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KAROWICH

DATE/TIME OF TEST: 12/01/2024 21:43:17

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:40	
Control Test	0.081	21:41	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:41	Subject Test Refused
Breath Volume	0.000	21:41	
Air Blank	0.000	21:42	
Control Test	0.081	21:42	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GILGEN

DATE/TIME OF TEST: 12/01/2024 21:59:54

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.149	21:55
Breath Volume	2.464	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:57
Subject Sample # 2	0.146	21:58
Breath Volume	1.984	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:46

SUBJECT'S LAST NAME: DUARTE DAMAS

DATE/TIME OF TEST: 12/03/2024 12:12:59

Diagnostics Check	OK	12:06
Air Blank	0.000	12:06
Control Test	0.081	12:07
Air Blank	0.000	12:07
Subject Sample # 1	0.220	12:08
Breath Volume	1.750	12:08
Air Blank	0.000	12:08
Air Blank	0.000	12:10
Subject Sample # 2	0.214	12:11
Breath Volume	1.765	12:11
Air Blank	0.000	12:12
Control Test	0.080	12:12
Air Blank	0.000	12:12
Diagnostics Check	OK	12:12

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THEIS,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: WHALEY

DATE/TIME OF TEST: 12/07/2024 03:37:19

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:24	
Subject Sample # 1	0.069	03:27	
Breath Volume	1.156	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample # 2	0.065	03:32	Volume Not Met
Breath Volume	0.765	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample # 3	0.075	03:35	
Breath Volume	1.515	03:35	
Air Blank	0.000	03:36	
Control Test	0.079	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZAREM,BRANDEN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZAREM,BRANDEN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GUZMAN GARCIA

DATE/TIME OF TEST: 12/09/2024 03:01:30

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.212	02:56
Breath Volume	1.636	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.231	02:59
Breath Volume	2.250	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSNYAK,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BOSNYAK,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: DRUMHELLER

DATE/TIME OF TEST: 12/10/2024 12:34:49

Diagnostics Check	OK	12:28
Air Blank	0.000	12:28
Control Test	0.081	12:29
Air Blank	0.000	12:29
Subject Sample # 1	0.261	12:30
Breath Volume	1.816	12:30
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.256	12:33
Breath Volume	1.585	12:33
Air Blank	0.000	12:33
Control Test	0.077	12:34
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: DEMARY,CASEY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 12/10/2024 20:03:14

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.081	19:57
Air Blank	0.000	19:57
Subject Sample # 1	0.156	19:58
Breath Volume	3.445	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:00
Subject Sample # 2	0.151	20:01
Breath Volume	3.351	20:01
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMES,JASON,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: SIMONSEN,ERIK
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 12/14/2024 02:29:50

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.206	02:25
Breath Volume	3.652	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.198	02:28
Breath Volume	1.460	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUTRAL,WILLIAM,Z
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: FUTRAL,W
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/14/2024 02:34:59

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MARIO,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: RODRIGUEZ,MARIO
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 12/14/2024 04:31:42

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.082	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.169	04:27
Breath Volume	2.621	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.164	04:30
Breath Volume	3.730	04:30
Air Blank	0.000	04:30
Control Test	0.080	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000813

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,CHRISTINA,T
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ESQUIVEL,CHRISTINA
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: KEILING

DATE/TIME OF TEST: 12/17/2024 21:55:00

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.080	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:54	
Control Test	0.080	21:54	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 12/18/2024 18:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAYTON,A,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MADDEN,D
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23324080A2
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BUTTON

DATE/TIME OF TEST: 12/24/2024 05:29:55

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.079	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.280	05:25
Breath Volume	2.355	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.282	05:28
Breath Volume	2.652	05:28
Air Blank	0.000	05:28
Control Test	0.077	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 12/28/2024 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: LYONS,C
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CZAPOR

DATE/TIME OF TEST: 12/02/2024 03:37:00

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 12/29/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: BENJUMEA ZULUAGA

DATE/TIME OF TEST: 12/02/2024 20:41:30

Diagnostics Check	OK	20:34
Air Blank	0.000	20:35
Control Test	0.080	20:35
Air Blank	0.000	20:35
Subject Sample # 1	0.181	20:37
Breath Volume	1.277	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.171	20:39
Breath Volume	1.328	20:39
Air Blank	0.000	20:40
Control Test	0.080	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 12/29/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HILLMAN,E
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: MARTINEZ CONTRERAS

DATE/TIME OF TEST: 12/07/2024 04:49:46

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.192	04:45
Breath Volume	2.046	04:45
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.192	04:48
Breath Volume	1.898	04:48
Air Blank	0.000	04:48
Control Test	0.079	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 12/29/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 12/07/2024 18:31:01

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.080	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.232	18:25
Breath Volume	1.890	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:28
Subject Sample # 2	0.227	18:29
Breath Volume	1.835	18:29
Air Blank	0.000	18:30
Control Test	0.080	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:31

Data Download Date/Time: 12/29/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WHITE,SHAWN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MARTINEZ-BELTRAN

DATE/TIME OF TEST: 12/08/2024 00:47:28

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.080	00:41	
Air Blank	0.000	00:41	
Subject Sample # 1	0.165	00:42	
Breath Volume	2.730	00:42	
Air Blank	0.000	00:43	
Air Blank	0.000	00:45	
Subject Sample # 2	0.162	00:46	
Breath Volume	2.511	00:46	
Air Blank	0.000	00:46	
Control Test	0.000	00:46	RFI Detect
Air Blank	0.000	00:47	

Data Download Date/Time: 12/29/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: FLOYD,D
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: FRISH

DATE/TIME OF TEST: 12/10/2024 19:06:46

Diagnostics Check	OK	18:58
Air Blank	0.000	18:58
Control Test	0.077	18:59
Air Blank	0.000	18:59
Subject Sample # 1	0.195	19:01
Breath Volume	1.523	19:01
Air Blank	0.000	19:02
Air Blank	0.000	19:04
Subject Sample # 2	0.200	19:05
Breath Volume	1.675	19:05
Air Blank	0.000	19:05
Control Test	0.078	19:06
Air Blank	0.000	19:06
Diagnostics Check	OK	19:06

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVARADO,RODERICK,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: WHITE,S
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: REIDER

DATE/TIME OF TEST: 12/12/2024 03:16:03

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.000	03:11
Breath Volume	2.281	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.000	03:14
Breath Volume	2.226	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:16

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRICKLAND,KAVEN,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: HRADECKY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/15/2024 01:35:01

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENNICK,SARAH,A
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CALLAWAY,C
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: MCKEE

DATE/TIME OF TEST: 12/15/2024 10:16:33

Diagnostics Check	OK	10:13	
Air Blank	0.000	10:14	
Control Test	0.078	10:14	
Air Blank	0.000	10:15	
Subject Sample	# 1	10:15	Subject Test Refused
Breath Volume	0.000	10:15	
Air Blank	0.000	10:15	
Control Test	0.078	10:15	
Air Blank	0.000	10:16	
Diagnostics Check	OK	10:16	

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CIERRA,L
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GOEHRING,JONATHAN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: SCHILLER

DATE/TIME OF TEST: 12/20/2024 19:07:33

Diagnostics Check	OK	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Subject Sample # 1	0.000	19:03
Breath Volume	2.066	19:03
Air Blank	0.000	19:03
Air Blank	0.000	19:05
Subject Sample # 2	0.000	19:06
Breath Volume	1.773	19:06
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Diagnostics Check	OK	19:07

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSSELL,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: RILEY,JAMES
ARRESTING OFFICER AGENCY: Wildwood PD
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LONGA

DATE/TIME OF TEST: 12/27/2024 01:19:36

Diagnostics Check	OK	01:12
Air Blank	0.000	01:13
Control Test	0.077	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.116	01:14
Breath Volume	3.367	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:17
Subject Sample # 2	0.119	01:17
Breath Volume	3.539	01:17
Air Blank	0.000	01:18
Control Test	0.076	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,REBECCA,G
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: MOORE,REBECCA
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 12/27/2024 02:01:39

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.000	01:57
Breath Volume	2.367	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.000	02:00
Breath Volume	2.367	02:00
Air Blank	0.000	02:00
Control Test	0.078	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIEFEL,JAMLS
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: GEOHRING,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 12/28/2024 22:27:58

Diagnostics Check	OK	22:16	
Air Blank	0.000	22:16	
Control Test	0.078	22:17	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:20	No Sample Provided
Breath Volume	0.000	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample	# 2	22:26	No Sample Provided
Breath Volume	0.000	22:26	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 12/29/2024 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BLOTZ,G
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 12/04/2024 22:43:55

Diagnostics Check	OK	22:41	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.000	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: PERMENTER,N
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/04/2024 23:31:38

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.162	23:26
Breath Volume	1.992	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.146	23:30
Breath Volume	1.847	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/07/2024 22:08:19

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Subject Sample	# 1	22:06	Subject Test Refused
Breath Volume	0.000	22:06	
Air Blank	0.000	22:07	
Control Test	0.080	22:07	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GETZ,K
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: GOODPASTER

DATE/TIME OF TEST: 12/08/2024 16:29:32

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:27	
Control Test	0.079	16:27	
Air Blank	0.000	16:27	
Subject Sample	# 1	16:28	Subject Test Refused
Breath Volume	0.000	16:28	
Air Blank	0.000	16:28	
Control Test	0.079	16:28	
Air Blank	0.000	16:29	
Diagnostics Check	OK	16:29	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 12/14/2024 03:43:50

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.079	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.079	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BEETOW,CHARLIE
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 12/14/2024 03:58:06

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.109	03:53
Breath Volume	1.496	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.108	03:56
Breath Volume	1.574	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,BRITTANY,A
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: ALEXANDER III

DATE/TIME OF TEST: 12/14/2024 15:36:28

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:27	
Control Test	0.078	15:27	
Air Blank	0.000	15:28	
Subject Sample	# 1 0.196	15:31	Volume Not Met
Breath Volume	0.167	15:31	
Air Blank	0.000	15:32	
Air Blank	0.000	15:33	
Subject Sample	# 2	15:34	Subject Test Refused
Breath Volume	0.140	15:34	
Air Blank	0.000	15:35	
Control Test	0.078	15:35	
Air Blank	0.000	15:36	
Diagnostics Check	OK	15:36	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: LYONS,T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SEFFERN

DATE/TIME OF TEST: 12/15/2024 03:22:24

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.079	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:21	
Control Test	0.079	03:21	
Air Blank	0.000	03:22	
Diagnostics Check	OK	03:22	

Data Download Date/Time: 12/30/2024 05:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JEFFERY,P
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: VANSKOY,E
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRIGGS

DATE/TIME OF TEST: 12/15/2024 03:55:51

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.079	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.148	03:51
Breath Volume	1.402	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.143	03:54
Breath Volume	1.199	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DAVIS,J
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 12/16/2024 21:13:51

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.079	21:13	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KJEKSTAD,ERIC,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: SUTTLES

DATE/TIME OF TEST: 12/18/2024 18:32:20

Diagnostics Check	OK	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.322	18:26
Breath Volume	1.257	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:28
Subject Sample # 2	0.322	18:30
Breath Volume	1.148	18:30
Air Blank	0.000	18:31
Control Test	0.078	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FERRER,F
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PRATTON

DATE/TIME OF TEST: 12/20/2024 23:43:00

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.208	23:38
Breath Volume	3.937	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.204	23:41
Breath Volume	3.410	23:41
Air Blank	0.000	23:42
Control Test	0.079	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:43

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,VICTORIA,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: WASYLIA,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: BABCOCK

DATE/TIME OF TEST: 12/21/2024 02:51:28

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.170	02:46
Breath Volume	1.660	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.175	02:49
Breath Volume	2.593	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TROY,E
BREATH TEST OPERATOR AGENCY: Citrus County SO
ARRESTING OFFICER: DAVIS,JAMES
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/21/2024 17:09:50

Diagnostics Check	OK	17:09	
Air Blank		17:09	Ambient Fail
Air Blank		17:09	Improper Sample

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOEHM, ADAM, J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: GARRISON, T
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: RAMEY

DATE/TIME OF TEST: 12/23/2024 15:44:29

Diagnostics Check	OK	15:41	
Air Blank	0.000	15:42	
Control Test	0.079	15:42	
Air Blank	0.000	15:42	
Subject Sample	# 1	15:43	Subject Test Refused
Breath Volume	0.000	15:43	
Air Blank	0.000	15:43	
Control Test	0.079	15:43	
Air Blank	0.000	15:44	
Diagnostics Check	OK	15:44	

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: JACKSON,N
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: LALLIER

DATE/TIME OF TEST: 12/23/2024 20:31:14

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.080	20:24
Air Blank	0.000	20:25
Subject Sample # 1	0.000	20:26
Breath Volume	1.269	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.000	20:29
Breath Volume	1.871	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: SAKKAB,D
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 12/23/2024 23:57:02

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.079	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.186	23:52
Breath Volume	2.417	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.189	23:55
Breath Volume	2.527	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:57

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ROSADO,A
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: KUSMAUL

DATE/TIME OF TEST: 12/24/2024 20:49:24

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.080	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1	20:48	Subject Test Refused
Breath Volume	0.000	20:48	
Air Blank	0.000	20:48	
Control Test	0.079	20:48	
Air Blank	0.000	20:49	
Diagnostics Check	OK	20:49	

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: VARGULISH

DATE/TIME OF TEST: 12/27/2024 22:36:11

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.168	22:31
Breath Volume	1.484	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.163	22:34
Breath Volume	1.562	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,LENNARD,
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: BELCHER,C
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: DEMOREST

DATE/TIME OF TEST: 12/29/2024 00:07:38

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.102	00:02
Breath Volume	1.785	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.110	00:06
Breath Volume	3.222	00:06
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 12/30/2024 05:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GILGEN

DATE/TIME OF TEST: 12/01/2024 21:09:33

Diagnostics Check	OK	21:08	
Air Blank	0.000	21:08	
		21:09	RFI Detect
Air Blank	0.000	21:09	

Data Download Date/Time: 12/18/2024 18:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: GILGEN

DATE/TIME OF TEST: 12/01/2024 21:11:40

Diagnostics Check	OK	21:11	
Air Blank		21:11	RFI Detect
Air Blank		21:11	RFI Detect

Data Download Date/Time: 12/18/2024 18:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: COLOMB

DATE/TIME OF TEST: 12/25/2024 04:41:55

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.080	04:39	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:40	RFI Detect
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Control Test	0.079	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 01/24/2025 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: COLOMB

DATE/TIME OF TEST: 12/25/2024 04:45:28

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.080	04:44	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:44	RFI Detect
Breath Volume	0.000	04:44	
Air Blank		04:44	RFI Detect
Air Blank	0.000	04:45	

Data Download Date/Time: 01/24/2025 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 12/29/2024 19:55:29

Diagnostics Check	OK	19:49
Air Blank	0.000	19:49
Control Test	0.079	19:49
Air Blank	0.000	19:50
Subject Sample # 1	0.316	19:50
Breath Volume	3.015	19:50
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.307	19:53
Breath Volume	2.929	19:53
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:55
Diagnostics Check	OK	19:55

Data Download Date/Time: 01/24/2025 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ, S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BUSTAMANTE

DATE/TIME OF TEST: 12/04/2024 05:30:08

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.195	05:24
Breath Volume	1.281	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.204	05:28
Breath Volume	1.769	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/07/2024 23:55:30

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.082	23:48
Air Blank	0.000	23:49
Subject Sample # 1	0.097	23:50
Breath Volume	1.246	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.100	23:53
Breath Volume	1.554	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: QUIROS

DATE/TIME OF TEST: 12/08/2024 03:05:16

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.178	03:00
Breath Volume	1.867	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.178	03:03
Breath Volume	1.296	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG, Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 12/08/2024 03:49:39

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.171	03:44
Breath Volume	1.535	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.168	03:47
Breath Volume	1.800	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT, TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: TORRES VERDECIA

DATE/TIME OF TEST: 12/09/2024 00:55:29

Diagnostics Check	OK	00:49
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.169	00:50
Breath Volume	2.101	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.171	00:53
Breath Volume	1.839	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: NIZ PEREZ

DATE/TIME OF TEST: 12/09/2024 01:18:51

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.189	01:14
Breath Volume	1.324	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.194	01:17
Breath Volume	1.324	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: VALDES LOPEZ

DATE/TIME OF TEST: 12/09/2024 04:49:44

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.176	04:44
Breath Volume	1.257	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:47
Subject Sample # 2	0.172	04:48
Breath Volume	1.406	04:48
Air Blank	0.000	04:48
Control Test	0.078	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:49

Data Download Date/Time: 12/12/2024 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MENDENHALL

DATE/TIME OF TEST: 12/13/2024 01:54:02

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.082	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.229	01:49
Breath Volume	1.539	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.218	01:52
Breath Volume	1.433	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/13/2024 03:55:00

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.133	03:50
Breath Volume	2.191	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.127	03:53
Breath Volume	1.808	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:54
Diagnostics Check	OK	03:55

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: RODRIGUEZ CHAVIANO

DATE/TIME OF TEST: 12/14/2024 01:21:17

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.170	01:16
Breath Volume	1.761	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.165	01:19
Breath Volume	1.523	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LOPEZ CARABALLO

DATE/TIME OF TEST: 12/14/2024 04:27:30

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.116	04:22
Breath Volume	2.234	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.117	04:25
Breath Volume	2.550	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: SHILON SANTIZ

DATE/TIME OF TEST: 12/14/2024 05:18:34

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.162	05:13
Breath Volume	1.472	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.162	05:16
Breath Volume	1.199	05:16
Air Blank	0.000	05:17
Control Test	0.080	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 12/17/2024 23:45:54

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.137	23:38
Breath Volume	1.289	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.159	23:41
Breath Volume	2.437	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 3	0.154	23:44
Breath Volume	2.679	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/18/2024 02:20:10

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.000	02:15
Breath Volume	2.687	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.000	02:18
Breath Volume	2.750	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/22/2024 02:10:24

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.169	02:05
Breath Volume	1.234	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.185	02:08
Breath Volume	1.933	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Temple Terrace PD
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: VEIGA GARCIA

DATE/TIME OF TEST: 12/22/2024 03:35:26

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.158	03:29
Breath Volume	1.835	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.150	03:33
Breath Volume	1.375	03:33
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 12/23/2024 00:41:10

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.076	00:36
Breath Volume	3.125	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.075	00:39
Breath Volume	3.089	00:39
Air Blank	0.000	00:40
Control Test	0.081	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: YARVA

DATE/TIME OF TEST: 12/23/2024 01:42:32

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.082	01:38
Breath Volume	1.503	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.087	01:40
Breath Volume	1.410	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CORKERY

DATE/TIME OF TEST: 12/23/2024 03:44:06

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.081	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.124	03:39
Breath Volume	2.980	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.127	03:42
Breath Volume	2.835	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO, ANGIE, A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT, J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: VASQUEZ PEREZ

DATE/TIME OF TEST: 12/27/2024 22:59:51

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.248	22:55
Breath Volume	2.960	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.238	22:58
Breath Volume	1.894	22:58
Air Blank	0.000	22:58
Control Test	0.080	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCEY,JANICE,F
BREATH TEST OPERATOR AGENCY: Tampa PD
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: GLATT

DATE/TIME OF TEST: 12/31/2024 22:15:03

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:10
Breath Volume	3.457	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	2.039	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:14
Diagnostics Check	OK	22:15

Data Download Date/Time: 01/28/2025 12:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TORRES,DANIEL
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: RHOM

DATE/TIME OF TEST: 12/01/2024 09:34:30

Diagnostics Check	OK	09:29	
Air Blank	0.000	09:30	
Control Test	0.081	09:30	
Air Blank	0.000	09:31	
Subject Sample	# 1 0.293	09:31	
Breath Volume	1.371	09:31	
Air Blank	0.000	09:32	
Air Blank		09:34	RFI Detect
Air Blank		09:34	RFI Detect
Air Blank		09:34	RFI Detect

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULTRON,CARLOS,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TORRES,DANIEL
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: R500805718790

DATE/TIME OF TEST: 12/01/2024 09:49:19

Diagnostics Check	OK	09:46	
Air Blank	0.000	09:46	
Control Test	0.080	09:47	
Air Blank	0.000	09:47	
Subject Sample	# 1	09:47	Subject Test Refused
Breath Volume	0.000	09:47	
Air Blank	0.000	09:48	
Control Test	0.080	09:48	
Air Blank	0.000	09:49	
Diagnostics Check	OK	09:49	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BALLADARES,A
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CZARNECKE

DATE/TIME OF TEST: 12/03/2024 22:51:08

Diagnostics Check	OK	22:50	
Air Blank		22:50	RFI Detect
Air Blank		22:51	RFI Detect

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: BALLADARES,A
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: CZARNECKE

DATE/TIME OF TEST: 12/03/2024 22:55:14

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:52	
Control Test	0.080	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:53	Subject Test Refused
Breath Volume	0.000	22:53	
Air Blank	0.000	22:54	
Control Test	0.080	22:54	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DORAN,R
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 12/03/2024 23:29:46

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.080	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: FRANCIS,KENO
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: HENDRICK

DATE/TIME OF TEST: 12/07/2024 19:25:51

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:23	
Control Test	0.080	19:23	
Air Blank	0.000	19:24	
Subject Sample	# 1	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:24	
Control Test	0.081	19:25	
Air Blank	0.000	19:25	
Diagnostics Check	OK	19:25	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: EVANS,JEFFREY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 12/09/2024 02:54:20

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank		02:54	RFI Detect
Air Blank		02:54	RFI Detect

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: EVANS,JEFFREY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 12/09/2024 02:58:22

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.079	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: BURNS,JEFFERY
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 221303
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ESPINAL LAMAR

DATE/TIME OF TEST: 12/14/2024 02:12:39

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.127	02:08
Breath Volume	1.691	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.122	02:10
Breath Volume	1.726	02:10
Air Blank	0.000	02:11
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,PHILIP,J
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HARRICHARAN

DATE/TIME OF TEST: 12/15/2024 00:22:28

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.080	00:21	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POST,HAYDEN,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: TORRES,DANIEL
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ELMORE

DATE/TIME OF TEST: 12/17/2024 04:00:24

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.188	03:55
Breath Volume	3.535	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.189	03:58
Breath Volume	2.593	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 12/17/2024 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,DAVID,R
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: MCCLAIN,DAVID
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CARLBERG

DATE/TIME OF TEST: 12/19/2024 00:56:04

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/21/2024 03:07:47

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.201	03:06	Volume Not Met
Breath Volume	0.289	03:06	
Air Blank		03:07	Purge Fail
Air Blank	0.000	03:07	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: WILDEMAN-STUART,NATASH
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/21/2024 03:13:26

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.078	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.238	03:11	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: EVANS,J
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: TRUE

DATE/TIME OF TEST: 12/22/2024 03:42:56

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.176	03:38
Breath Volume	1.136	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.196	03:41
Breath Volume	2.359	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: VARELA,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: EADY JR

DATE/TIME OF TEST: 12/22/2024 04:57:58

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.080	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1 0.133	04:53	
Breath Volume	1.410	04:53	
Air Blank	0.000	04:54	
Air Blank	0.000	04:56	
Subject Sample	# 2	04:56	Subject Test Refused
Breath Volume	0.000	04:56	
Air Blank	0.000	04:57	
Control Test	0.077	04:57	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,NATASHA,L
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: VARELA,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/22/2024 22:32:55

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:30	
Control Test	0.080	22:30	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:31	Subject Test Refused
Breath Volume	0.000	22:31	
Air Blank	0.000	22:32	
Control Test	0.080	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: NEWBOLD,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 12/26/2024 01:27:45

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:26	
Control Test	0.079	01:26	
Air Blank		01:27	RFI Detect
Air Blank	0.000	01:27	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: NEWBOLD,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 12/26/2024 01:41:48

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.079	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1 0.143	01:34	Volume Not Met
Breath Volume	0.074	01:34	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2 0.140	01:40	Volume Not Met
Breath Volume	0.082	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LUNA,ROMAN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/29/2024 03:19:11

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.160	03:14
Breath Volume	2.449	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.156	03:17
Breath Volume	1.679	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: LUTFDI,FADEE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ISBILL

DATE/TIME OF TEST: 12/30/2024 01:58:39

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.155	01:50
Breath Volume	1.972	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.181	01:53
Breath Volume	3.886	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 3	0.171	01:57
Breath Volume	3.062	01:57
Air Blank	0.000	01:57
Control Test	0.077	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,STEPHEN,V
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: OWENS,DENNIS
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/30/2024 03:38:33

Diagnostics Check	OK	03:26	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1 0.197	03:28	
Breath Volume	2.015	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:31	
Subject Sample	# 2 0.195	03:34	Volume Not Met
Breath Volume	0.343	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:37	
Subject Sample	# 3	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.076	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

INSTRUMENT REGISTERED TO: UMATILLA PD(LAKE SO)
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: COLON,D
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 33522080A5
DRY GAS EXPIRY DATE: 01/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/31/2024 02:33:34

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.295	02:27
Breath Volume	1.273	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.291	02:31
Breath Volume	1.355	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 01/16/2025 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SHOOK

DATE/TIME OF TEST: 12/01/2024 03:11:08

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:06
Breath Volume	2.441	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.000	03:09
Breath Volume	2.421	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 12/19/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 12/02/2024 01:35:32

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.154	01:30
Breath Volume	1.125	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.147	01:34
Breath Volume	1.441	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 12/19/2024 23:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ARMOUR

DATE/TIME OF TEST: 12/28/2024 21:00:11

Diagnostics Check	OK	20:48	
Air Blank	0.000	20:48	
Control Test	0.080	20:49	
Air Blank	0.000	20:49	
Subject Sample	# 1 0.158	20:52	Volume Not Met
Breath Volume	0.246	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 2 0.152	20:58	Volume Not Met
Breath Volume	0.835	20:58	
Air Blank	0.000	20:59	
Control Test	0.080	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 01/07/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: ARMOUR

DATE/TIME OF TEST: 12/28/2024 21:10:05

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.151	21:05
Breath Volume	1.910	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:07
Subject Sample # 2	0.146	21:08
Breath Volume	1.593	21:08
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:10

Data Download Date/Time: 01/07/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWLEY,MASON,C
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LAWLEY,MASON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PIPPIN

DATE/TIME OF TEST: 12/29/2024 01:23:45

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.080	01:18
Air Blank	0.000	01:18
Subject Sample # 1	0.211	01:19
Breath Volume	1.542	01:19
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.212	01:22
Breath Volume	1.652	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 01/07/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FRYAR

DATE/TIME OF TEST: 12/30/2024 00:10:03

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:05
Breath Volume	1.863	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:08
Breath Volume	2.437	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:10

Data Download Date/Time: 01/07/2025 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402806079
DRY GAS EXPIRY DATE: 07/26/2026
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 12/30/2024 04:09:11

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.122	04:03
Breath Volume	1.339	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.105	04:07
Breath Volume	1.699	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 01/07/2025 20:39