

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: PYNE,C
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 12/11/2024 23:54:10

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.082	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SCHALLON,C
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: HUMPHREY

DATE/TIME OF TEST: 12/18/2024 19:13:48

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.081	19:06	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:09	No Sample Provided
Breath Volume	0.000	19:09	
Air Blank	0.000	19:10	
Air Blank	0.000	19:12	
Subject Sample	# 2	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:12	
Control Test	0.081	19:13	
Air Blank	0.000	19:13	
Diagnostics Check	OK	19:13	

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ-SIFONTES,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,K
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/19/2024 18:26:54

Diagnostics Check	OK	18:24	
Air Blank	0.000	18:24	
Control Test	0.081	18:24	
Air Blank	0.000	18:25	
Subject Sample	# 1	18:25	Subject Test Refused
Breath Volume	0.000	18:25	
Air Blank	0.000	18:26	
Control Test	0.082	18:26	
Air Blank	0.000	18:26	
Diagnostics Check	OK	18:26	

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: FAY,T
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/20/2024 22:55:31

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.082	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.274	22:50
Breath Volume	1.359	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.273	22:53
Breath Volume	2.222	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYNY,E,
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: NOLAND,P
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: RHYNES

DATE/TIME OF TEST: 12/24/2024 16:50:00

Diagnostics Check	OK	16:47	
Air Blank	0.000	16:47	
Control Test	0.082	16:47	
Air Blank	0.000	16:48	
Subject Sample	# 1	16:48	Subject Test Refused
Breath Volume	0.000	16:48	
Air Blank	0.000	16:49	
Control Test	0.082	16:49	
Air Blank	0.000	16:49	
Diagnostics Check	OK	16:50	

Data Download Date/Time: 12/24/2024 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYLIS,BRYAN,T
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LYNCH,DANIEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 12/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/30/2024 18:25:51

Diagnostics Check	OK	18:23	
Air Blank	0.000	18:23	
Control Test	0.082	18:23	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:24	
Control Test	0.082	18:25	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 01/09/2025 20:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BUNKER

DATE/TIME OF TEST: 12/06/2024 21:59:26

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:54	
Control Test	0.079	21:54	
Air Blank	0.000	21:55	
Subject Sample	# 1 0.089	21:55	
Breath Volume	1.429	21:55	
Air Blank	0.000	21:56	
Air Blank		21:58	RFI Detect
Air Blank	0.000	21:58	
Control Test	0.079	21:58	
Air Blank	0.000	21:59	
Diagnostics Check	OK	21:59	

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BUNKER

DATE/TIME OF TEST: 12/06/2024 22:08:18

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.093	22:03
Breath Volume	1.613	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.083	22:06
Breath Volume	1.527	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CIPULLY

DATE/TIME OF TEST: 12/08/2024 01:41:51

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.195	01:37
Breath Volume	2.023	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.189	01:40
Breath Volume	1.335	01:40
Air Blank	0.000	01:40
Control Test	0.076	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,JOHN,A
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: SMITH,BEN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MARINELLI

DATE/TIME OF TEST: 12/11/2024 17:42:02

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.078	17:35
Air Blank	0.000	17:36
Subject Sample # 1	0.127	17:37
Breath Volume	1.562	17:37
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.120	17:40
Breath Volume	1.738	17:40
Air Blank	0.000	17:41
Control Test	0.076	17:41
Air Blank	0.000	17:41
Diagnostics Check	OK	17:42

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: ABRAHAMSEN,ERIK
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 12/14/2024 05:40:51

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:36	
Control Test	0.077	05:36	
Air Blank	0.000	05:37	
Subject Sample	# 1	05:39	Subject Test Refused
Breath Volume	0.000	05:39	
Air Blank	0.000	05:39	
Control Test	0.078	05:40	
Air Blank	0.000	05:40	
Diagnostics Check	OK	05:40	

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: RENKEN,EMILY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: BENEDITH CHIRINOS

DATE/TIME OF TEST: 12/14/2024 23:57:31

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:43	
Subject Sample	# 1 0.185	23:44	
Breath Volume	1.679	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2	23:50	No Sample Provided
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 3	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Control Test	0.069	23:56	Control Outside Tolerance
Air Blank	0.000	23:57	

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEBELL,DON,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: RENKEN,EMILY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BENEDITH CHIRINOS

DATE/TIME OF TEST: 12/15/2024 00:34:23

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.075	00:26	
Air Blank	0.000	00:27	
Subject Sample # 1	0.156	00:29	
Breath Volume	1.527	00:29	
Air Blank	0.000	00:30	
Air Blank	0.000	00:32	
Subject Sample # 2	0.161	00:32	
Breath Volume	1.453	00:32	
Air Blank	0.000	00:33	
Control Test	0.072	00:33	Control Outside Tolerance
Air Blank	0.000	00:34	

Data Download Date/Time: 12/18/2024 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERS,JOHN,A
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: RIVERS,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:48

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/20/2024 12:30:08

Diagnostics Check	OK	12:23
Air Blank	0.000	12:24
Control Test	0.078	12:24
Air Blank	0.000	12:24
Subject Sample # 1	0.000	12:25
Breath Volume	2.207	12:25
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample # 2	0.000	12:28
Breath Volume	1.574	12:28
Air Blank	0.000	12:29
Control Test	0.077	12:29
Air Blank	0.000	12:30
Diagnostics Check	OK	12:30

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREW

DATE/TIME OF TEST: 12/22/2024 02:55:11

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.078	02:54	
Air Blank		02:54	RFI Detect
Air Blank		02:55	RFI Detect

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/22/2024 02:57:46

Diagnostics Check	OK	02:57	
Air Blank		02:57	RFI Detect
Air Blank		02:57	RFI Detect

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/22/2024 03:00:07

Diagnostics Check	Fail	02:59	Voltage/Current Test
Air Blank		03:00	RFI Detect

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/22/2024 03:04:39

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
		03:04	RFI Detect
Air Blank		03:04	RFI Detect

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/22/2024 03:25:20

Diagnostics Check	OK	03:24	
Air Blank		03:25	RFI Detect
Air Blank		03:25	RFI Detect

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES',ABIGAIL,L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: JONES,ABIGAIL
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/22/2024 03:35:04

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.078	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.078	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: ROSENTHAL,LEXI
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 15:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/26/2024 15:59:51

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.078	15:52
Air Blank	0.000	15:53
Subject Sample # 1	0.250	15:54
Breath Volume	1.453	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.241	15:58
Breath Volume	1.515	15:58
Air Blank	0.000	15:58
Control Test	0.075	15:59
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 01/17/2025 17:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/15/2024 00:56:47

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.050	00:52
Breath Volume	1.851	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.048	00:55
Breath Volume	2.484	00:55
Air Blank	0.000	00:55
Control Test	0.077	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 01/31/2025 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 12/23/2024 02:13:19

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.269	02:08
Breath Volume	3.222	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.250	02:11
Breath Volume	3.125	02:11
Air Blank	0.000	02:12
Control Test	0.076	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 01/31/2025 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTATITA, CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MENDEZ VASQUEZ

DATE/TIME OF TEST: 12/23/2024 20:48:11

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.078	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:39	No Sample Provided
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Air Blank	0.000	20:42	
Subject Sample	# 2	0.179	20:43
Breath Volume	1.511	20:43	
Air Blank	0.000	20:44	
Air Blank	0.000	20:45	
Subject Sample	# 3	0.181	20:46
Breath Volume	2.433	20:46	
Air Blank	0.000	20:47	
Control Test	0.076	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 01/31/2025 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAYNALDO,
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SUSTAITA,CRISTIAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/25/2024 01:57:57

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.155	01:53
Breath Volume	3.632	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.144	01:56
Breath Volume	2.285	01:56
Air Blank	0.000	01:57
Control Test	0.076	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/31/2025 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GUILLERMO

DATE/TIME OF TEST: 12/26/2024 00:26:38

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.077	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.198	00:21
Breath Volume	1.710	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.185	00:25
Breath Volume	1.917	00:25
Air Blank	0.000	00:25
Control Test	0.075	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 01/31/2025 23:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 12/14/2024 20:38:38

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.080	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.080	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 01/16/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARRITT,KEVIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ASECIO BARRERA

DATE/TIME OF TEST: 12/16/2024 23:46:58

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:44	
Control Test	0.081	23:44	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:46	
Control Test	0.081	23:46	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 01/16/2025 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LAMBORGHINI,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 12/17/2024 00:29:27

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.081	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 01/16/2025 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: LUCKE

DATE/TIME OF TEST: 12/21/2024 02:04:17

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.082	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.139	01:59
Breath Volume	1.371	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.143	02:02
Breath Volume	1.320	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 01/16/2025 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLSON,BRENDAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GOOD

DATE/TIME OF TEST: 12/26/2024 03:14:46

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.081	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 01/16/2025 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,ALEXANDRA,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG304401
DRY GAS EXPIRY DATE: 02/13/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PEREZ ZAPOTE

DATE/TIME OF TEST: 12/27/2024 03:12:24

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank	0.000	03:11	
Control Test	0.080	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 01/16/2025 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 12/28/2024 21:38:28

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.079	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:37	Subject Test Refused
Breath Volume	0.000	21:37	
Air Blank	0.000	21:37	
Control Test	0.078	21:37	
Air Blank	0.000	21:38	
Diagnostics Check	OK	21:38	

Data Download Date/Time: 01/16/2025 10:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMERI,DOMINIC,J
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: BROWN,BRITTANY
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: PRESLEY

DATE/TIME OF TEST: 12/21/2024 22:46:03

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.206	22:41
Breath Volume	3.113	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.202	22:44
Breath Volume	2.789	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:46

Data Download Date/Time: 01/25/2025 01:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWAB,JOSEPH,S
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: JOSEPH,DILLON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/02/2024 02:09:47

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.119	02:04
Breath Volume	1.226	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.119	02:08
Breath Volume	1.671	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: SHUMATE,TODD
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:22

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/02/2024 10:49:43

Diagnostics Check	OK	10:43
Air Blank	0.000	10:43
Control Test	0.081	10:44
Air Blank	0.000	10:44
Subject Sample # 1	0.047	10:44
Breath Volume	3.093	10:44
Air Blank	0.000	10:45
Air Blank	0.000	10:47
Subject Sample # 2	0.048	10:48
Breath Volume	3.132	10:48
Air Blank	0.000	10:48
Control Test	0.080	10:49
Air Blank	0.000	10:49
Diagnostics Check	OK	10:49

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: UMMEL,JOSHUA
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: MILAN

DATE/TIME OF TEST: 12/06/2024 02:26:43

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.082	02:22
Breath Volume	2.429	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.084	02:25
Breath Volume	1.937	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEBER,DYLAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 12/08/2024 06:09:42

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.081	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.155	06:02
Breath Volume	1.914	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.147	06:08
Breath Volume	1.371	06:08
Air Blank	0.000	06:08
Control Test	0.080	06:09
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: CLAUSEN, CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: KISWANI

DATE/TIME OF TEST: 12/08/2024 21:37:58

Diagnostics Check	OK	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.000	21:33
Breath Volume	4.300	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.000	21:36
Breath Volume	4.046	21:36
Air Blank	0.000	21:37
Control Test	0.080	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,J,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: STANTON,E
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LATULIPPE

DATE/TIME OF TEST: 12/14/2024 01:44:54

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.149	01:39
Breath Volume	1.175	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.156	01:43
Breath Volume	2.109	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,J,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: UMMEL,J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/15/2024 05:53:02

Diagnostics Check	OK	05:46
Air Blank	0.000	05:46
Control Test	0.081	05:47
Air Blank	0.000	05:47
Subject Sample # 1	0.062	05:48
Breath Volume	1.253	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.061	05:51
Breath Volume	1.410	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:53

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEILL,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: NEILL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/18/2024 03:14:53

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.081	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.156	03:10
Breath Volume	1.476	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:12
Subject Sample # 2	0.150	03:13
Breath Volume	1.203	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,DYLAN,T
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MALDONADO,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/20/2024 22:10:25

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.080	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:03	No Sample Provided
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 2	22:09	No Sample Provided
Breath Volume	0.000	22:09	
Air Blank	0.000	22:09	
Control Test	0.079	22:09	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:10	

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL, WILLIAM, R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: SMITH, KEVIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: MUIR

DATE/TIME OF TEST: 12/25/2024 19:24:08

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.081	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.207	19:19
Breath Volume	2.406	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.210	19:22
Breath Volume	2.164	19:22
Air Blank	0.000	19:23
Control Test	0.079	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 12/28/2024 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWAB,JOSEPH,S
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: SCHWAB,JOSEPH
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/28/2024 22:02:22

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.080	21:49	
Air Blank	0.000	21:50	
Subject Sample	# 1 0.317	21:51	No .020 agreement
Breath Volume	1.496	21:51	
Air Blank	0.000	21:52	
Air Blank	0.000	21:54	
Subject Sample	# 2 0.318	21:57	Volume Not Met
Breath Volume	0.425	21:57	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample	# 3 0.287	22:00	No .020 agreement
Breath Volume	2.085	22:00	
Air Blank	0.000	22:01	
Control Test	0.078	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 01/25/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINSKER,JOSEPH,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: REYNOLDS,J
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: MAYORGA

DATE/TIME OF TEST: 12/29/2024 06:46:57

Diagnostics Check	OK	06:40
Air Blank	0.000	06:41
Control Test	0.080	06:41
Air Blank	0.000	06:42
Subject Sample # 1	0.143	06:42
Breath Volume	2.648	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:45
Subject Sample # 2	0.140	06:45
Breath Volume	2.562	06:45
Air Blank	0.000	06:46
Control Test	0.078	06:46
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

Data Download Date/Time: 01/25/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,JASON,B
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RAMOS,EVGARDO
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: KISSELL

DATE/TIME OF TEST: 12/30/2024 14:05:52

Diagnostics Check	OK	13:59
Air Blank	0.000	13:59
Control Test	0.081	14:00
Air Blank	0.000	14:00
Subject Sample # 1	0.275	14:01
Breath Volume	1.250	14:01
Air Blank	0.000	14:02
Air Blank	0.000	14:03
Subject Sample # 2	0.286	14:04
Breath Volume	2.074	14:04
Air Blank	0.000	14:04
Control Test	0.076	14:05
Air Blank	0.000	14:05
Diagnostics Check	OK	14:05

Data Download Date/Time: 01/25/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LIVINGSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 12/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: CLOW

DATE/TIME OF TEST: 12/31/2024 13:24:27

Diagnostics Check	OK	13:18
Air Blank	0.000	13:18
Control Test	0.081	13:18
Air Blank	0.000	13:19
Subject Sample # 1	0.370	13:19
Breath Volume	1.277	13:19
Air Blank	0.000	13:20
Air Blank	0.000	13:22
Subject Sample # 2	0.365	13:22
Breath Volume	1.773	13:22
Air Blank	0.000	13:23
Control Test	0.079	13:23
Air Blank	0.000	13:24
Diagnostics Check	OK	13:24

Data Download Date/Time: 01/25/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HODGDON,TODD,S

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HODGDON,TODD

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 12/25/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24923080A4

DRY GAS EXPIRY DATE: 10/05/2025

VIOLATION CODE: DUI

ARREST TIME: 21:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 12/31/2024 22:18:48

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.080	22:13
Air Blank	0.000	22:13
Subject Sample # 1	0.229	22:14
Breath Volume	3.000	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.227	22:17
Breath Volume	3.621	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 01/28/2025 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/05/2024 20:27:41

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.082	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.150	20:23
Breath Volume	1.457	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.159	20:26
Breath Volume	1.773	20:26
Air Blank	0.000	20:26
Control Test	0.082	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SHINDO,TADASHI
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WASZMER

DATE/TIME OF TEST: 12/07/2024 03:48:49

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.082	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:47	No Sample Provided
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.082	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,AUSTIN,D
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: HARDGROVE,A
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DOBROSIELSKI

DATE/TIME OF TEST: 12/11/2024 00:07:31

Diagnostics Check	OK	23:57
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:58
Subject Sample # 3	0.175	23:59
Breath Volume	1.312	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.153	00:02
Breath Volume	1.406	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.147	00:05
Breath Volume	1.757	00:05
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: SHINDO,TADASHI
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: EHKO

DATE/TIME OF TEST: 12/12/2024 01:01:12

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.082	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.011	00:56
Breath Volume	1.769	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.008	00:59
Breath Volume	2.363	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCARLEY, MATTHEW, L
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: LEBRETON, QUINN
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: MCGRAW

DATE/TIME OF TEST: 12/12/2024 22:18:53

Diagnostics Check	OK	22:12
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.151	22:14
Breath Volume	2.664	22:14
Air Blank	0.000	22:15
Air Blank	0.000	22:16
Subject Sample # 2	0.147	22:17
Breath Volume	2.789	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:18
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: MELBOURNE PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: MELBOURNE PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MOMBELLY

DATE/TIME OF TEST: 12/21/2024 03:28:57

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.081	03:24
Breath Volume	1.421	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.081	03:27
Breath Volume	1.324	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, TYLER, J
BREATH TEST OPERATOR AGENCY: BREVARD COUNTY SO
ARRESTING OFFICER: BREWER, KYLE
ARRESTING OFFICER AGENCY: BREVARD COUNTY SO
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: HUMES

DATE/TIME OF TEST: 12/23/2024 13:58:29

Diagnostics Check	OK	13:46	
Air Blank	0.000	13:47	
Control Test	0.082	13:47	
Air Blank	0.000	13:48	
Subject Sample	# 1	13:51	No Sample Provided
Breath Volume	0.000	13:51	
Air Blank	0.000	13:51	
Air Blank	0.000	13:53	
Subject Sample	# 2	13:56	Volume Not Met
Breath Volume	0.031	13:56	
Air Blank	0.000	13:57	
Control Test	0.081	13:57	
Air Blank	0.000	13:58	
Diagnostics Check	OK	13:58	

Data Download Date/Time: 12/26/2024 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/12/2024 21:13:44

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.135	21:09
Breath Volume	2.066	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.127	21:12
Breath Volume	1.527	21:12
Air Blank	0.000	21:12
Control Test	0.080	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 12/23/2024 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DEJESUS CERVANTES

DATE/TIME OF TEST: 12/18/2024 03:09:57

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.158	03:05
Breath Volume	2.890	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.160	03:08
Breath Volume	3.050	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 12/23/2024 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/19/2024 01:39:47

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.080	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 12/23/2024 09:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEENAN,K
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/01/2024 02:49:38

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.083	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.185	02:44
Breath Volume	1.605	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.183	02:47
Breath Volume	1.761	02:47
Air Blank	0.000	02:48
Control Test	0.081	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ARANGO

DATE/TIME OF TEST: 12/01/2024 04:40:49

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.091	04:35
Breath Volume	1.339	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.091	04:39
Breath Volume	1.445	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,C,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BERNARD,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 12/01/2024 07:38:56

Diagnostics Check	OK	07:31
Air Blank	0.000	07:31
Control Test	0.083	07:31
Air Blank	0.000	07:32
Subject Sample # 1	0.182	07:33
Breath Volume	3.230	07:33
Air Blank	0.000	07:34
Air Blank	0.000	07:36
Subject Sample # 2	0.183	07:37
Breath Volume	1.234	07:37
Air Blank	0.000	07:38
Control Test	0.080	07:38
Air Blank	0.000	07:38
Diagnostics Check	OK	07:38

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,C,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 12/01/2024 17:36:30

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:34	
Control Test	0.083	17:34	
Air Blank	0.000	17:34	
Subject Sample	# 1	17:35	Subject Test Refused
Breath Volume	0.000	17:35	
Air Blank	0.000	17:35	
Control Test	0.082	17:35	
Air Blank	0.000	17:36	
Diagnostics Check	OK	17:36	

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: SHELLINBURG

DATE/TIME OF TEST: 12/01/2024 19:48:16

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.083	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.333	19:43
Breath Volume	2.023	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:45
Subject Sample # 2	0.324	19:46
Breath Volume	1.503	19:46
Air Blank	0.000	19:47
Control Test	0.078	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: SAIKE

DATE/TIME OF TEST: 12/02/2024 03:54:34

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:52	
Control Test	0.083	03:52	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.082	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FARE

DATE/TIME OF TEST: 12/02/2024 04:40:03

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.082	04:34
Air Blank	0.000	04:34
Subject Sample # 1	0.117	04:35
Breath Volume	3.597	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:37
Subject Sample # 2	0.118	04:38
Breath Volume	3.609	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 12/06/2024 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: RUBINO

DATE/TIME OF TEST: 12/06/2024 19:46:17

Diagnostics Check	OK	19:39
Air Blank	0.000	19:40
Control Test	0.082	19:40
Air Blank	0.000	19:40
Subject Sample # 1	0.219	19:41
Breath Volume	1.808	19:41
Air Blank	0.000	19:42
Air Blank	0.000	19:43
Subject Sample # 2	0.200	19:44
Breath Volume	2.035	19:44
Air Blank	0.000	19:45
Control Test	0.080	19:45
Air Blank	0.000	19:46
Diagnostics Check	OK	19:46

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: JONES-RORIE

DATE/TIME OF TEST: 12/07/2024 00:28:22

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.151	00:23
Breath Volume	1.378	00:23
Air Blank	0.000	00:24
Air Blank	0.000	00:25
Subject Sample # 2	0.159	00:26
Breath Volume	1.414	00:26
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICHARDSON,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 12/07/2024 03:26:56

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.083	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/07/2024 21:23:04

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.082	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.168	21:18
Breath Volume	1.707	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.170	21:21
Breath Volume	1.875	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:23

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MUNOZ,I
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 12/08/2024 03:03:32

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.083	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.193	02:58
Breath Volume	2.222	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.186	03:01
Breath Volume	1.308	03:01
Air Blank	0.000	03:02
Control Test	0.079	03:02
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 12/08/2024 04:21:30

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:19	
Control Test	0.083	04:19	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:20	Subject Test Refused
Breath Volume	0.000	04:20	
Air Blank	0.000	04:20	
Control Test	0.083	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MAKUEI

DATE/TIME OF TEST: 12/09/2024 03:36:50

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:33	
Control Test	0.082	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 1 0.289	03:35	
Breath Volume	1.285	03:35	
Air Blank		03:36	Purge Fail
Air Blank	0.000	03:36	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HERRMANN

DATE/TIME OF TEST: 12/10/2024 22:17:56

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:15	
Control Test	0.081	22:15	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:16	
Control Test	0.081	22:17	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PORTIS

DATE/TIME OF TEST: 12/11/2024 04:05:54

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:03	
Air Blank	0.000	04:04	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.082	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: MACKTO

DATE/TIME OF TEST: 12/12/2024 05:16:52

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:14	
Control Test	0.083	05:14	
Air Blank	0.000	05:15	
Subject Sample	# 1	05:15	Subject Test Refused
Breath Volume	0.000	05:15	
Air Blank	0.000	05:15	
Control Test	0.082	05:16	
Air Blank	0.000	05:16	
Diagnostics Check	OK	05:16	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/13/2024 00:33:21

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.083	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.056	00:28
Breath Volume	1.140	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.054	00:31
Breath Volume	1.246	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MORENO FLEITES

DATE/TIME OF TEST: 12/13/2024 03:58:45

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.100	03:53
Breath Volume	2.316	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.097	03:57
Breath Volume	2.199	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 12/13/2024 20:52:41

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:50	
Control Test	0.083	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 1	20:51	Subject Test Refused
Breath Volume	0.000	20:51	
Air Blank	0.000	20:51	
Control Test	0.082	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRECO,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: NICKS

DATE/TIME OF TEST: 12/14/2024 04:21:23

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.082	04:19	
Air Blank	0.000	04:19	
Subject Sample	# 1	04:19	Subject Test Refused
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Control Test	0.082	04:20	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KOCH,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: BACHTOLD

DATE/TIME OF TEST: 12/14/2024 05:44:05

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.082	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.125	05:39
Breath Volume	2.960	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:41
Subject Sample # 2	0.122	05:42
Breath Volume	3.164	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PARMLEY

DATE/TIME OF TEST: 12/15/2024 01:58:33

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.114	01:53
Breath Volume	1.339	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.110	01:56
Breath Volume	1.421	01:56
Air Blank	0.000	01:57
Control Test	0.079	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/16/2024 04:52:43

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.082	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.099	04:47
Breath Volume	3.019	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.102	04:50
Breath Volume	3.187	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 01/09/2025 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILDEY,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LANTZ

DATE/TIME OF TEST: 12/17/2024 00:13:43

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.082	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.013	00:09
Breath Volume	1.289	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.014	00:12
Breath Volume	1.523	00:12
Air Blank	0.000	00:12
Control Test	0.082	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRUBOR,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/18/2024 00:01:33

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:59	
Control Test	0.082	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:00	
Control Test	0.082	00:00	
Air Blank	0.000	00:01	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRELL,J,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/18/2024 00:40:53

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.082	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.260	00:35
Breath Volume	1.511	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.255	00:39
Breath Volume	2.000	00:39
Air Blank	0.000	00:39
Control Test	0.078	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/18/2024 06:00:53

Diagnostics Check	OK	05:54	
Air Blank	0.000	05:55	
Control Test	0.081	05:55	
Air Blank	0.000	05:55	
Subject Sample	# 1 0.259	05:59	Volume Not Met
Breath Volume	0.746	05:59	
Air Blank		05:59	Purge Fail
Air Blank	0.000	06:00	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROUCH,L,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 12/18/2024 20:56:45

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1 0.108	20:49	Volume Not Met
Breath Volume	0.152	20:49	
Air Blank	0.000	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 2 0.118	20:54	Volume Not Met
Breath Volume	0.660	20:54	
Air Blank	0.000	20:55	
Control Test	0.080	20:56	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANFORD,B,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOLMAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/19/2024 03:44:48

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.159	03:39
Breath Volume	2.277	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.151	03:43
Breath Volume	1.554	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 12/21/2024 02:11:01

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.082	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.039	02:06
Breath Volume	2.433	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.038	02:09
Breath Volume	1.789	02:09
Air Blank	0.000	02:10
Control Test	0.082	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 12/21/2024 03:34:38

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.082	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.173	03:28
Breath Volume	2.070	03:28
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.165	03:32
Breath Volume	2.683	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BULJUBASIC

DATE/TIME OF TEST: 12/21/2024 05:53:20

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.082	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.100	05:48
Breath Volume	1.687	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.093	05:51
Breath Volume	1.425	05:51
Air Blank	0.000	05:52
Control Test	0.082	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DICKENS,S
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 12/21/2024 20:39:05

Diagnostics Check	OK	20:36	
Air Blank	0.000	20:36	
Control Test	0.083	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:38	
Control Test	0.082	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPINA,C,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GUILLEN RAMIREZ

DATE/TIME OF TEST: 12/22/2024 01:52:42

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.083	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KLING,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: WILLARD

DATE/TIME OF TEST: 12/22/2024 03:44:25

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.083	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.143	03:39
Breath Volume	1.343	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.143	03:42
Breath Volume	1.699	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: ORAMA VEGA

DATE/TIME OF TEST: 12/23/2024 02:38:51

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.144	02:34
Breath Volume	3.609	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.144	02:37
Breath Volume	2.699	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEMPF,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 12/23/2024 19:34:01

Diagnostics Check	OK	19:27
Air Blank	0.000	19:28
Control Test	0.082	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.000	19:29
Breath Volume	2.382	19:29
Air Blank	0.000	19:29
Air Blank	0.000	19:32
Subject Sample # 2	0.000	19:32
Breath Volume	2.593	19:32
Air Blank	0.000	19:33
Control Test	0.082	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:34

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BOYCHEW

DATE/TIME OF TEST: 12/26/2024 02:26:22

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.082	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:22
Breath Volume	1.304	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:24
Breath Volume	1.140	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/26/2024 03:53:14

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.082	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.082	03:52	
Air Blank	0.000	03:53	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LEON PONCE

DATE/TIME OF TEST: 12/27/2024 03:53:16

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.107	03:48
Breath Volume	2.546	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.108	03:51
Breath Volume	1.789	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARRY,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: WHITENER

DATE/TIME OF TEST: 12/27/2024 05:30:25

Diagnostics Check	OK	05:27	
Air Blank	0.000	05:27	
Control Test	0.082	05:28	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:29	Subject Test Refused
Breath Volume	0.000	05:29	
Air Blank	0.000	05:29	
Control Test	0.081	05:29	
Air Blank	0.000	05:30	
Diagnostics Check	OK	05:30	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MCCARTY

DATE/TIME OF TEST: 12/27/2024 21:13:34

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:11	
Control Test	0.081	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.080	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: OMLOR

DATE/TIME OF TEST: 12/28/2024 03:13:25

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ANDRINGA

DATE/TIME OF TEST: 12/29/2024 02:23:03

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.081	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 1	02:21	Subject Test Refused
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Control Test	0.081	02:22	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PRADON,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: JIMINEZ

DATE/TIME OF TEST: 12/29/2024 06:03:22

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:00	
Control Test	0.082	06:01	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:01	Subject Test Refused
Breath Volume	0.000	06:01	
Air Blank	0.000	06:02	
Control Test	0.081	06:02	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

Data Download Date/Time: 01/09/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 12/30/2024 04:08:43

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.169	04:02
Breath Volume	1.183	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.174	04:06
Breath Volume	2.230	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/09/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 12/31/2024 01:31:46

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.075	01:27
Breath Volume	1.710	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.072	01:30
Breath Volume	1.785	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/09/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BALASSAITIS

DATE/TIME OF TEST: 12/06/2024 02:55:30

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.089	02:50
Breath Volume	1.476	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.085	02:54
Breath Volume	2.242	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JORDAN,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: CACCAMO

DATE/TIME OF TEST: 12/07/2024 19:33:01

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:27
Subject Sample # 1	0.209	19:28
Breath Volume	2.164	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.218	19:31
Breath Volume	2.222	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HARLEY

DATE/TIME OF TEST: 12/07/2024 23:05:44

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.158	23:01
Breath Volume	1.312	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.163	23:04
Breath Volume	1.265	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 12/09/2024 21:39:44

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.171	21:35
Breath Volume	2.816	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.175	21:38
Breath Volume	3.375	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TROUTT,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 12/10/2024 06:29:06

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:20	
Control Test	0.080	06:20	
Air Blank	0.000	06:21	
Subject Sample	# 1	06:24	No Sample Provided
Breath Volume	0.000	06:24	
Air Blank	0.000	06:24	
Air Blank	0.000	06:27	
Subject Sample	# 2	06:27	Subject Test Refused
Breath Volume	0.000	06:27	
Air Blank	0.000	06:28	
Control Test	0.081	06:28	
Air Blank	0.000	06:29	
Diagnostics Check	OK	06:29	

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RASIKIENE

DATE/TIME OF TEST: 12/14/2024 01:57:36

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:49	
Control Test	0.081	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:53	No Sample Provided
Breath Volume	0.000	01:53	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Control Test	0.080	01:57	
Air Blank	0.000	01:57	
Diagnostics Check	OK	01:57	

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/14/2024 03:13:54

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.209	03:09
Breath Volume	1.941	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.211	03:12
Breath Volume	2.035	03:12
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 12/19/2024 19:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SUMMERS

DATE/TIME OF TEST: 12/14/2024 04:00:16

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BASHFORD, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WOOLARD

DATE/TIME OF TEST: 12/14/2024 04:25:43

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.079	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:21	No Sample Provided
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 2	04:24	Subject Test Refused
Breath Volume	0.000	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Diagnostics Check	OK	04:25	

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDER,S,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 12/14/2024 07:00:40

Diagnostics Check	OK	06:53
Air Blank	0.000	06:54
Control Test	0.080	06:54
Air Blank	0.000	06:54
Subject Sample # 1	0.000	06:55
Breath Volume	1.109	06:55
Air Blank	0.000	06:56
Air Blank	0.000	06:58
Subject Sample # 2	0.000	06:59
Breath Volume	1.125	06:59
Air Blank	0.000	06:59
Control Test	0.080	07:00
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIDDERS,R,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: SHUCKERS

DATE/TIME OF TEST: 12/14/2024 08:46:20

Diagnostics Check	OK	08:38
Air Blank	0.000	08:38
Control Test	0.080	08:39
Air Blank	0.000	08:39
Subject Sample # 1	0.145	08:41
Breath Volume	2.628	08:41
Air Blank	0.000	08:41
Air Blank	0.000	08:43
Subject Sample # 2	0.151	08:44
Breath Volume	2.847	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:46
Diagnostics Check	OK	08:46

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DAZEY

DATE/TIME OF TEST: 12/14/2024 23:33:02

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.276	23:27
Breath Volume	2.984	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.266	23:31
Breath Volume	1.417	23:31
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:33

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 12/15/2024 00:42:44

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.080	00:41	
Air Blank		00:42	RFI Detect
Air Blank	0.000	00:42	

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: SORRELLS

DATE/TIME OF TEST: 12/15/2024 00:51:14

Diagnostics Check	OK	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.183	00:46
Breath Volume	1.453	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.179	00:49
Breath Volume	2.652	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BENDA

DATE/TIME OF TEST: 12/15/2024 03:43:39

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.109	03:39
Breath Volume	1.234	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.113	03:42
Breath Volume	1.414	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JONES, N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/15/2024 22:02:27

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.175	21:57
Breath Volume	2.558	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.182	22:00
Breath Volume	3.269	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/15/2024 23:07:19

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.230	23:02
Breath Volume	3.691	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.227	23:05
Breath Volume	2.984	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: RIDAUGHT

DATE/TIME OF TEST: 12/18/2024 19:49:32

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.080	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.285	19:45
Breath Volume	2.917	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.278	19:48
Breath Volume	2.695	19:48
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

Data Download Date/Time: 12/19/2024 19:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: PREDLEY

DATE/TIME OF TEST: 12/20/2024 20:04:17

Diagnostics Check	OK	19:57	
Air Blank	0.000	19:58	
Control Test	0.080	19:58	
Air Blank	0.000	19:59	
Subject Sample	# 1 0.392	19:59	
Breath Volume	1.882	19:59	
Air Blank	0.000	20:00	
Air Blank	0.000	20:02	
Subject Sample	# 2	20:02	Subject Test Refused
Breath Volume	0.000	20:02	
Air Blank	0.000	20:03	
Control Test	0.079	20:03	
Air Blank	0.000	20:04	
Diagnostics Check	OK	20:04	

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 12/23/2024 02:07:56

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.174	02:01
Breath Volume	1.867	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.170	02:06
Breath Volume	1.265	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: EPPS

DATE/TIME OF TEST: 12/25/2024 02:50:44

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.208	02:45
Breath Volume	1.492	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.211	02:49
Breath Volume	1.679	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BUI,L
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 12/27/2024 03:54:23

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	No Sample Provided
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Air Blank	0.000	03:52	
Subject Sample	# 2	03:53	Subject Test Refused
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Control Test	0.080	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 12/27/2024 21:45:16

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:37	
Control Test	0.080	21:37	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:41	No Sample Provided
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 2	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Control Test	0.080	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DOANE

DATE/TIME OF TEST: 12/29/2024 00:16:40

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.155	00:11
Breath Volume	2.468	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.142	00:15
Breath Volume	1.269	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 01/27/2025 07:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KABAN,SEAN,W
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: SOKOLOHORSKY,ANDREW
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RIVAS DAVILA

DATE/TIME OF TEST: 12/18/2024 02:39:46

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.164	02:35
Breath Volume	1.437	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.165	02:38
Breath Volume	2.984	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 02/28/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JOSE,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: KENNY,MICHAEL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: VAN CORA

DATE/TIME OF TEST: 12/22/2024 22:38:48

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:32
Subject Sample # 1	0.403	22:34
Breath Volume	2.714	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:36
Subject Sample # 2	0.405	22:37
Breath Volume	2.445	22:37
Air Blank	0.000	22:37
Control Test	0.079	22:38
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 02/28/2025 18:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: RAMIREZ,ANGEL
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GALICIA GONZALEZ

DATE/TIME OF TEST: 12/06/2024 02:22:59

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.228	02:18
Breath Volume	1.480	02:18
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.240	02:21
Breath Volume	2.832	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 12/31/2024 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOSHUA,E
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: AYALA,NATALIA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 12/27/2024 02:00:41

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.194	01:56
Breath Volume	3.855	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.201	01:59
Breath Volume	3.835	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 12/31/2024 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 12/07/2024 03:37:14

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.080	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.148	03:32
Breath Volume	2.050	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.146	03:35
Breath Volume	1.429	03:35
Air Blank	0.000	03:36
Control Test	0.080	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 12/19/2024 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: CRUZ,EDGARDO
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HARDIN

DATE/TIME OF TEST: 12/07/2024 04:26:32

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.080	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.098	04:22
Breath Volume	3.542	04:22
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.100	04:25
Breath Volume	3.527	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 12/19/2024 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DONOVAN

DATE/TIME OF TEST: 12/08/2024 00:16:34

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.132	00:12
Breath Volume	2.800	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.128	00:15
Breath Volume	3.054	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:15
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 12/19/2024 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,KALEB,R
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/31/2024 23:50:25

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.079	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.169	23:46
Breath Volume	1.753	23:46
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.180	23:48
Breath Volume	1.972	23:48
Air Blank	0.000	23:49
Control Test	0.078	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 02/05/2025 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVIN, MITCHELL, A
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: MITCHELL, JARRELL
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:27

SUBJECT'S LAST NAME: POCOCK

DATE/TIME OF TEST: 12/08/2024 12:58:29

Diagnostics Check	OK	12:55	
Air Blank	0.000	12:56	
Control Test	0.080	12:56	
Air Blank	0.000	12:56	
Subject Sample	# 1	12:57	Subject Test Refused
Breath Volume	0.000	12:57	
Air Blank	0.000	12:57	
		12:58	RFI Detect
Air Blank	0.000	12:58	

Data Download Date/Time: 01/09/2025 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINCIN,WESLEY,P
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CZAJAK,MAGGIE
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WHITEMORE

DATE/TIME OF TEST: 12/06/2024 22:42:00

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Control Test	0.082	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 12/31/2024 03:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ZACHARY,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: MILLS,ZACHARY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231043
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/22/2024 04:39:10

Diagnostics Check	OK	04:29	
Air Blank	0.000	04:30	
Control Test	0.082	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.193	04:34	Volume Not Met
Breath Volume	0.238	04:34	
Air Blank	0.000	04:34	
Air Blank	0.000	04:36	
Subject Sample	# 2	04:37	Subject Test Refused
Breath Volume	0.218	04:37	
Air Blank	0.000	04:38	
Control Test	0.081	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 12/31/2024 03:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASSEN,STEVE,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GASSEN,S
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: KUKULSKI

DATE/TIME OF TEST: 12/11/2024 15:54:47

Diagnostics Check	OK	15:46	
Air Blank	0.000	15:46	
Control Test	0.078	15:47	
Air Blank	0.000	15:47	
Subject Sample	# 1	15:50	No Sample Provided
Breath Volume	0.000	15:50	
Air Blank	0.000	15:51	
Air Blank	0.000	15:53	
Subject Sample	# 2	15:53	Subject Test Refused
Breath Volume	0.000	15:53	
Air Blank	0.000	15:53	
Control Test	0.078	15:54	
Air Blank	0.000	15:54	
Diagnostics Check	OK	15:54	

Data Download Date/Time: 02/14/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,NICHOLAS,W
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SCHALLES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 12/15/2024 02:46:24

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.191	02:41
Breath Volume	1.468	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.196	02:44
Breath Volume	1.976	02:44
Air Blank	0.000	02:45
Control Test	0.078	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 02/14/2025 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,F
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: SERBATY

DATE/TIME OF TEST: 12/19/2024 16:46:49

Diagnostics Check	OK	16:40
Air Blank	0.000	16:41
Control Test	0.078	16:41
Air Blank	0.000	16:42
Subject Sample # 1	0.019	16:42
Breath Volume	3.132	16:42
Air Blank	0.000	16:43
Air Blank	0.000	16:44
Subject Sample # 2	0.018	16:45
Breath Volume	3.234	16:45
Air Blank	0.000	16:45
Control Test	0.079	16:46
Air Blank	0.000	16:46
Diagnostics Check	OK	16:46

Data Download Date/Time: 02/14/2025 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: KUBISIAK

DATE/TIME OF TEST: 12/07/2024 02:11:58

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/08/2024 03:35:21

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.082	03:33	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:34	
Control Test	0.081	03:34	
Air Blank	0.000	03:35	
Diagnostics Check	OK	03:35	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONFERRARO,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: RODRIGUEZ AGUIRRE

DATE/TIME OF TEST: 12/08/2024 19:43:50

Diagnostics Check	OK	19:41	
Air Blank	0.000	19:41	
Control Test	0.080	19:41	
Air Blank	0.000	19:42	
Subject Sample	# 1	19:42	Subject Test Refused
Breath Volume	0.000	19:42	
Air Blank	0.000	19:42	
Control Test	0.081	19:43	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HAND

DATE/TIME OF TEST: 12/09/2024 03:23:07

Diagnostics Check	OK	03:20	
Air Blank	0.000	03:20	
Control Test	0.081	03:21	
Air Blank	0.000	03:21	
Subject Sample	# 1	03:21	Subject Test Refused
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Control Test	0.080	03:22	
Air Blank	0.000	03:23	
Diagnostics Check	OK	03:23	

Data Download Date/Time: 12/11/2024 11:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TOWRISS

DATE/TIME OF TEST: 12/12/2024 00:11:30

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.082	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: SPOTO

DATE/TIME OF TEST: 12/17/2024 23:23:59

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.255	23:19
Breath Volume	2.386	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.259	23:22
Breath Volume	1.367	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: OLIVIERI

DATE/TIME OF TEST: 12/18/2024 04:03:25

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HERRERA ESTEVEZ

DATE/TIME OF TEST: 12/19/2024 02:45:08

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: MOONEY

DATE/TIME OF TEST: 12/20/2024 19:27:05

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.184	19:22
Breath Volume	2.125	19:22
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.170	19:25
Breath Volume	2.550	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:27

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ANDERSON,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 12/20/2024 23:08:11

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:07	
Control Test	0.081	23:07	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MONTELONGO,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: PEREZ GUEVARA

DATE/TIME OF TEST: 12/21/2024 00:03:34

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.081	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.221	23:59
Breath Volume	1.730	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.218	00:02
Breath Volume	1.207	00:02
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SARINE

DATE/TIME OF TEST: 12/21/2024 01:09:12

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.160	01:04
Breath Volume	2.121	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.160	01:07
Breath Volume	2.054	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: STRUCK

DATE/TIME OF TEST: 12/21/2024 20:15:56

Diagnostics Check	OK	20:09
Air Blank	0.000	20:09
Control Test	0.082	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.166	20:11
Breath Volume	2.136	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.170	20:14
Breath Volume	2.179	20:14
Air Blank	0.000	20:14
Control Test	0.081	20:15
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 12/22/2024 03:19:55

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.111	03:15
Breath Volume	1.683	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.107	03:18
Breath Volume	1.464	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 01/16/2025 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ADSIT,M
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/25/2024 22:36:55

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.080	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.080	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONFERRARO,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WINSTED

DATE/TIME OF TEST: 12/26/2024 01:00:20

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.206	00:55
Breath Volume	2.035	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.198	00:58
Breath Volume	1.183	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEATTY,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ONKKA

DATE/TIME OF TEST: 12/26/2024 01:33:30

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.080	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.214	01:28
Breath Volume	1.675	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.204	01:31
Breath Volume	1.472	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BONFERRARO,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: PLAINES

DATE/TIME OF TEST: 12/27/2024 01:51:28

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.158	01:46
Breath Volume	1.519	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.156	01:49
Breath Volume	1.632	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: CHADDUCK

DATE/TIME OF TEST: 12/29/2024 01:31:47

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:29	
Control Test	0.079	01:29	
Air Blank	0.000	01:30	
Subject Sample	# 1	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.078	01:31	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HIRSHMAN,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: PROSSER

DATE/TIME OF TEST: 12/30/2024 21:18:57

Diagnostics Check	OK	21:11
Air Blank	0.000	21:12
Control Test	0.080	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.241	21:13
Breath Volume	1.898	21:13
Air Blank	0.000	21:14
Air Blank	0.000	21:15
Subject Sample # 2	0.239	21:17
Breath Volume	1.304	21:17
Air Blank	0.000	21:18
Control Test	0.079	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOODY,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: NEVES

DATE/TIME OF TEST: 12/31/2024 19:43:20

Diagnostics Check	OK	19:40	
Air Blank	0.000	19:40	
Control Test	0.079	19:41	
Air Blank	0.000	19:41	
Subject Sample	# 1	19:41	Subject Test Refused
Breath Volume	0.000	19:41	
Air Blank	0.000	19:42	
Control Test	0.079	19:42	
Air Blank	0.000	19:43	
Diagnostics Check	OK	19:43	

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRSHMAN,BRIAN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAEDKE,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG409204
DRY GAS EXPIRY DATE: 04/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HARRIER

DATE/TIME OF TEST: 12/31/2024 23:01:58

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.200	22:57
Breath Volume	2.195	22:57
Air Blank	0.000	22:58
Air Blank	0.000	22:59
Subject Sample # 2	0.181	23:00
Breath Volume	1.429	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 01/16/2025 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OMORUYI, TYREE,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: DUQUE, STIVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HEERALALL

DATE/TIME OF TEST: 12/22/2024 02:48:44

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.077	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.091	02:43
Breath Volume	2.683	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.088	02:47
Breath Volume	2.687	02:47
Air Blank	0.000	02:47
Control Test	0.077	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/08/2025 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: HERNANDEZ-HUAPILLA

DATE/TIME OF TEST: 12/01/2024 01:12:14

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.151	01:07
Breath Volume	3.042	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:09
Subject Sample # 2	0.142	01:10
Breath Volume	2.492	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 02/13/2025 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,OFFICER,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: SMITH,OFFICER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BANUSH

DATE/TIME OF TEST: 12/04/2024 21:38:03

Diagnostics Check	OK	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.111	21:33
Breath Volume	2.144	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.113	21:36
Breath Volume	1.929	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:37
Diagnostics Check	OK	21:38

Data Download Date/Time: 02/13/2025 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: REBOLLEDO PEREZ

DATE/TIME OF TEST: 12/07/2024 18:39:59

Diagnostics Check	OK	18:30
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.077	18:35
Breath Volume	1.253	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.079	18:38
Breath Volume	1.875	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 02/13/2025 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/14/2024 18:25:04

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:08	
Control Test	0.078	18:09	
Air Blank	0.000	18:09	
Subject Sample	# 1	18:12	No Sample Provided
Breath Volume	0.000	18:12	
Air Blank	0.000	18:13	
Air Blank	0.000	18:15	
Subject Sample	# 2	0.168	18:17
Breath Volume	1.390	18:17	
Air Blank	0.000	18:18	
Air Blank	0.000	18:20	
Subject Sample	# 3	0.160	18:23
Breath Volume	0.359	18:23	Volume Not Met
Air Blank		18:24	Purge Fail
Air Blank		18:25	Purge Fail

Data Download Date/Time: 02/13/2025 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A4
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/14/2024 18:42:08

Diagnostics Check	OK	18:40	
Air Blank		18:41	Ambient Fail
Air Blank	0.000	18:42	

Data Download Date/Time: 02/13/2025 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: NEGLIA

DATE/TIME OF TEST: 12/01/2024 00:43:10

Diagnostics Check	OK	00:40	
Air Blank	0.000	00:40	
Control Test	0.084	00:41	
Air Blank	0.000	00:41	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.082	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: HTOO

DATE/TIME OF TEST: 12/01/2024 04:04:02

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.253	03:58
Breath Volume	1.695	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.258	04:02
Breath Volume	1.894	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: KROKE

DATE/TIME OF TEST: 12/01/2024 06:43:07

Diagnostics Check	OK	06:36
Air Blank	0.000	06:36
Control Test	0.083	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.197	06:38
Breath Volume	1.398	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.211	06:41
Breath Volume	1.421	06:41
Air Blank	0.000	06:42
Control Test	0.082	06:42
Air Blank	0.000	06:43
Diagnostics Check	OK	06:43

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 12/02/2024 22:04:33

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.084	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.168	21:59
Breath Volume	2.082	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.170	22:02
Breath Volume	1.851	22:02
Air Blank	0.000	22:03
Control Test	0.081	22:03
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: ABIDOGUN

DATE/TIME OF TEST: 12/03/2024 03:10:42

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.083	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.157	03:05
Breath Volume	1.371	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.140	03:09
Breath Volume	1.593	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ESTRELLA,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MCMILLIN

DATE/TIME OF TEST: 12/03/2024 17:10:32

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:08	
Control Test	0.083	17:08	
Air Blank	0.000	17:08	
Subject Sample	# 1	17:09	Subject Test Refused
Breath Volume	0.000	17:09	
Air Blank	0.000	17:09	
Control Test	0.081	17:09	
Air Blank	0.000	17:10	
Diagnostics Check	OK	17:10	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 12/04/2024 01:29:52

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.084	01:22
Air Blank	0.000	01:22
Subject Sample # 1	0.096	01:24
Breath Volume	1.480	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.092	01:28
Breath Volume	1.179	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 12/04/2024 02:49:04

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.082	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.144	02:43
Breath Volume	1.175	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.138	02:47
Breath Volume	2.914	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:49

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: ERGOTT

DATE/TIME OF TEST: 12/04/2024 16:30:24

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.083	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.175	16:25
Breath Volume	1.933	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:27
Subject Sample # 2	0.169	16:28
Breath Volume	1.960	16:28
Air Blank	0.000	16:29
Control Test	0.080	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: KINZEY

DATE/TIME OF TEST: 12/04/2024 21:00:45

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.082	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.260	20:55
Breath Volume	1.695	20:55
Air Blank	0.000	20:56
Air Blank	0.000	20:57
Subject Sample # 2	0.240	20:59
Breath Volume	1.535	20:59
Air Blank	0.000	20:59
Control Test	0.079	21:00
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/25/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: VAZQUEZ AGUILAR

DATE/TIME OF TEST: 12/04/2024 22:36:43

Diagnostics Check	OK	22:33	
Air Blank	0.000	22:34	
Control Test	0.082	22:34	
Air Blank	0.000	22:35	
Subject Sample	# 1	22:35	Subject Test Refused
Breath Volume	0.000	22:35	
Air Blank	0.000	22:35	
Control Test	0.083	22:36	
Air Blank	0.000	22:36	
Diagnostics Check	OK	22:36	

Data Download Date/Time: 12/06/2024 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 12/06/2024 23:59:18

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.082	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.133	23:54
Breath Volume	1.390	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.128	23:57
Breath Volume	1.640	23:57
Air Blank	0.000	23:58
Control Test	0.082	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/09/2025 12:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: NICKELS

DATE/TIME OF TEST: 12/07/2024 02:58:20

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.083	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.158	02:53
Breath Volume	1.597	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.149	02:56
Breath Volume	1.382	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 01/09/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORCHARDT,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NORELIA

DATE/TIME OF TEST: 12/07/2024 03:41:13

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.082	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.082	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 01/09/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,J,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 12/07/2024 21:26:20

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:23	
Control Test	0.082	21:24	
Air Blank	0.000	21:24	
Subject Sample	# 1	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:25	
Control Test	0.082	21:25	
Air Blank	0.000	21:26	
Diagnostics Check	OK	21:26	

Data Download Date/Time: 01/09/2025 12:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 12/07/2024 23:52:52

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.082	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.181	23:47
Breath Volume	2.421	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.172	23:51
Breath Volume	2.546	23:51
Air Blank	0.000	23:51
Control Test	0.082	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: GHRIR

DATE/TIME OF TEST: 12/08/2024 03:26:43

Diagnostics Check	OK	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.178	03:21
Breath Volume	3.363	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.179	03:25
Breath Volume	3.003	03:25
Air Blank	0.000	03:25
Control Test	0.082	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 12/08/2024 22:23:57

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.082	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:23	
Control Test	0.082	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MAKUEI

DATE/TIME OF TEST: 12/09/2024 03:56:49

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.312	03:47
Breath Volume	1.152	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.287	03:51
Breath Volume	1.320	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 3	0.303	03:55
Breath Volume	1.261	03:55
Air Blank	0.000	03:55
Control Test	0.077	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FITZJOHN

DATE/TIME OF TEST: 12/10/2024 22:57:08

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.081	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.000	22:52
Breath Volume	1.437	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:55
Breath Volume	1.203	22:55
Air Blank	0.000	22:56
Control Test	0.080	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 12/11/2024 03:07:28

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.082	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.068	03:01
Breath Volume	1.335	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.069	03:05
Breath Volume	2.304	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/11/2024 04:17:09

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:14	
Control Test	0.080	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GODINEZ SALES

DATE/TIME OF TEST: 12/12/2024 04:02:35

Diagnostics Check	OK	03:59	
Air Blank	0.000	04:00	
Control Test	0.083	04:00	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:01	
Control Test	0.082	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,N,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: GOMEZ ROLDAN

DATE/TIME OF TEST: 12/13/2024 00:19:22

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.083	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.026	00:14
Breath Volume	1.718	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.028	00:17
Breath Volume	1.109	00:17
Air Blank	0.000	00:18
Control Test	0.083	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/13/2024 04:07:28

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.083	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.160	04:02
Breath Volume	2.062	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.150	04:05
Breath Volume	1.988	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FALCO,R
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 12/14/2024 04:45:05

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:39	
Control Test	0.082	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:45	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P]
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: CEU

DATE/TIME OF TEST: 12/14/2024 21:29:18

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:26	
Control Test	0.082	21:27	
Air Blank	0.000	21:27	
Subject Sample	# 1	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNS,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: SHEARIN

DATE/TIME OF TEST: 12/14/2024 23:51:24

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.035	23:46
Breath Volume	1.437	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.034	23:49
Breath Volume	1.644	23:49
Air Blank	0.000	23:50
Control Test	0.081	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALERIO,ERNESTO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,ERNESTO
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PONTAZA ESPANA

DATE/TIME OF TEST: 12/15/2024 01:12:25

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.098	01:07
Breath Volume	1.460	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.102	01:10
Breath Volume	1.617	01:10
Air Blank	0.000	01:11
Control Test	0.080	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 12/15/2024 03:58:04

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.081	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.194	03:52
Breath Volume	1.148	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.201	03:56
Breath Volume	1.125	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 12/15/2024 19:48:48

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.069	19:44
Breath Volume	1.160	19:44
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.069	19:47
Breath Volume	1.210	19:47
Air Blank	0.000	19:47
Control Test	0.079	19:48
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CINOTTI,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: AKIN

DATE/TIME OF TEST: 12/15/2024 21:02:15

Diagnostics Check	OK	20:59	
Air Blank	0.000	20:59	
Control Test	0.081	21:00	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:00	Subject Test Refused
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOLMAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MAYHEW

DATE/TIME OF TEST: 12/16/2024 03:27:12

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/18/2024 06:38:02

Diagnostics Check	OK	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.291	06:32
Breath Volume	1.175	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.275	06:36
Breath Volume	1.210	06:36
Air Blank	0.000	06:37
Control Test	0.080	06:37
Air Blank	0.000	06:37
Diagnostics Check	OK	06:38

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: INNACI

DATE/TIME OF TEST: 12/18/2024 20:15:54

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:13	
Control Test	0.082	20:13	
Air Blank	0.000	20:14	
Subject Sample	# 1	20:14	Subject Test Refused
Breath Volume	0.000	20:14	
Air Blank	0.000	20:14	
Control Test	0.081	20:15	
Air Blank	0.000	20:15	
Diagnostics Check	OK	20:15	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GLAWSON

DATE/TIME OF TEST: 12/19/2024 03:27:28

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.081	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: RAPPEPORT

DATE/TIME OF TEST: 12/20/2024 00:50:53

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.082	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.291	00:46
Breath Volume	2.531	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.273	00:49
Breath Volume	1.621	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/20/2024 01:43:51

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.082	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Sequence Aborted
Breath Volume	0.000	01:43	
Air Blank	0.000	01:43	

Data Download Date/Time: 01/09/2025 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WILLAIMS

DATE/TIME OF TEST: 12/20/2024 01:49:24

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:47	
Control Test	0.081	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:48	Subject Test Refused
Breath Volume	0.000	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/21/2024 04:52:30

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.083	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Control Test	0.082	04:51	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DURHAM,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: PEREZ CORTES

DATE/TIME OF TEST: 12/22/2024 02:48:59

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.082	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.178	02:43
Breath Volume	1.472	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.187	02:47
Breath Volume	1.675	02:47
Air Blank	0.000	02:48
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHISMAN,N
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 12/22/2024 05:03:08

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.083	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.082	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,P,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: SAINTIL

DATE/TIME OF TEST: 12/22/2024 07:37:04

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.083	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.098	07:32
Breath Volume	1.304	07:32
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.107	07:35
Breath Volume	1.773	07:35
Air Blank	0.000	07:36
Control Test	0.083	07:36
Air Blank	0.000	07:36
Diagnostics Check	OK	07:37

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,E,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARRAR,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: SEAWRIGHT

DATE/TIME OF TEST: 12/22/2024 09:40:04

Diagnostics Check	OK	09:37	
Air Blank	0.000	09:37	
Control Test	0.082	09:37	
Air Blank	0.000	09:38	
Subject Sample	# 1	09:38	Subject Test Refused
Breath Volume	0.000	09:38	
Air Blank	0.000	09:39	
Control Test	0.082	09:39	
Air Blank	0.000	09:39	
Diagnostics Check	OK	09:40	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 12/22/2024 21:21:21

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:18	
Control Test	0.082	21:19	
Air Blank	0.000	21:19	
Subject Sample	# 1	21:19	Subject Test Refused
Breath Volume	0.000	21:19	
Air Blank	0.000	21:20	
Control Test	0.082	21:20	
Air Blank	0.000	21:21	
Diagnostics Check	OK	21:21	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,PETER,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/24/2024 03:27:32

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.143	03:22
Breath Volume	4.066	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.135	03:25
Breath Volume	2.316	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: WILMOT

DATE/TIME OF TEST: 12/24/2024 05:28:48

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.082	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.136	05:24
Breath Volume	3.222	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.131	05:27
Breath Volume	3.871	05:27
Air Blank	0.000	05:27
Control Test	0.080	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/25/2024 02:20:35

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.082	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.099	02:13
Breath Volume	1.300	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.104	02:18
Breath Volume	1.671	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ODVINA

DATE/TIME OF TEST: 12/26/2024 01:31:45

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.082	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.046	01:27
Breath Volume	2.101	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.046	01:30
Breath Volume	2.164	01:30
Air Blank	0.000	01:30
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: BRATULIN

DATE/TIME OF TEST: 12/26/2024 07:13:21

Diagnostics Check	OK	07:05
Air Blank	0.000	07:06
Control Test	0.082	07:06
Air Blank	0.000	07:07
Subject Sample # 1	0.224	07:08
Breath Volume	1.761	07:08
Air Blank	0.000	07:09
Air Blank	0.000	07:10
Subject Sample # 2	0.230	07:11
Breath Volume	2.519	07:11
Air Blank	0.000	07:12
Control Test	0.078	07:12
Air Blank	0.000	07:13
Diagnostics Check	OK	07:13

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ECHEVERRIAVIDAR,X
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 12/27/2024 01:31:17

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.082	01:29	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SPANN

DATE/TIME OF TEST: 12/27/2024 04:28:59

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.082	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.110	04:24
Breath Volume	1.363	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.115	04:27
Breath Volume	1.273	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TROTIER,PAUL,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BRECHT

DATE/TIME OF TEST: 12/28/2024 03:39:51

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.121	03:35
Breath Volume	1.480	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.120	03:38
Breath Volume	1.796	03:38
Air Blank	0.000	03:38
Control Test	0.080	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DESOUZA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: PIERRE-LOUIS

DATE/TIME OF TEST: 12/29/2024 03:28:16

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.220	03:23
Breath Volume	1.484	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.224	03:26
Breath Volume	1.636	03:26
Air Blank	0.000	03:27
Control Test	0.080	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAXWELL,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 12/29/2024 05:11:03

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.081	05:08	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Control Test	0.080	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: GODINEZ SALES

DATE/TIME OF TEST: 12/29/2024 19:23:31

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.081	19:16
Air Blank	0.000	19:17
Subject Sample # 1	0.178	19:18
Breath Volume	1.335	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.170	19:21
Breath Volume	1.234	19:21
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/30/2024 04:08:26

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.200	04:03
Breath Volume	1.449	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.208	04:06
Breath Volume	1.980	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIOPRYNA,GAGE,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: IEZZI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LUKA

DATE/TIME OF TEST: 12/31/2024 02:21:26

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:19	
Control Test	0.082	02:19	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:20	
Control Test	0.078	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: JEFFERS

DATE/TIME OF TEST: 12/31/2024 22:40:54

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.081	22:38	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 01/09/2025 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIELE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HARVEY,CAROLINE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 12/03/2024 19:51:11

Diagnostics Check	OK	19:44
Air Blank	0.000	19:44
Control Test	0.079	19:45
Air Blank	0.000	19:45
Subject Sample # 1	0.233	19:46
Breath Volume	1.578	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:48
Subject Sample # 2	0.252	19:49
Breath Volume	1.351	19:49
Air Blank	0.000	19:50
Control Test	0.078	19:50
Air Blank	0.000	19:51
Diagnostics Check	OK	19:51

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HELLER,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: LUTH

DATE/TIME OF TEST: 12/05/2024 19:02:08

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:57
Breath Volume	4.625	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.000	19:00
Breath Volume	5.117	19:00
Air Blank	0.000	19:01
Control Test	0.079	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BIL,BART
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BARAJAS

DATE/TIME OF TEST: 12/07/2024 02:41:15

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:36
Breath Volume	1.468	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.000	02:39
Breath Volume	1.683	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ARATA

DATE/TIME OF TEST: 12/09/2024 23:02:40

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.240	22:54
Breath Volume	1.667	22:54
Air Blank	0.000	22:55
Air Blank	0.000	22:56
Subject Sample # 2	0.210	22:57
Breath Volume	1.347	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 3	0.234	23:01
Breath Volume	2.617	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BERRIOS

DATE/TIME OF TEST: 12/12/2024 02:52:33

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.143	02:47
Breath Volume	1.621	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.150	02:51
Breath Volume	1.941	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERLYNE,R,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BERLYNE,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 12/13/2024 21:03:39

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.145	20:58
Breath Volume	1.593	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.149	21:02
Breath Volume	1.953	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ROMERO MENDOZA

DATE/TIME OF TEST: 12/14/2024 04:27:37

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1 0.136	04:17	Volume Not Met
Breath Volume	0.054	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample	# 2 0.161	04:22	
Breath Volume	1.218	04:22	
Air Blank	0.000	04:23	
Air Blank	0.000	04:24	
Subject Sample	# 3 0.161	04:26	
Breath Volume	1.867	04:26	
Air Blank	0.000	04:26	
Control Test	0.077	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,JORDAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROBERTS,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: PAYTON

DATE/TIME OF TEST: 12/14/2024 23:29:48

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:16	
Control Test	0.079	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1 0.190	23:19	
Breath Volume	3.214	23:19	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample	# 2	23:24	No Sample Provided
Breath Volume	0.222	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 3 0.170	23:28	
Breath Volume	1.824	23:28	
Air Blank	0.000	23:28	
Control Test	0.079	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GERARDI,JOHN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: VALLE GONZALEZ

DATE/TIME OF TEST: 12/15/2024 04:25:08

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.183	04:19
Breath Volume	2.371	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.195	04:23
Breath Volume	2.722	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,JORDAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ROBERTS,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LEBARON

DATE/TIME OF TEST: 12/20/2024 03:40:31

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.193	03:35
Breath Volume	4.578	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.200	03:39
Breath Volume	3.679	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIELE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: TRIGLIANOS,TIMOTHY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: POCKRANDT

DATE/TIME OF TEST: 12/22/2024 22:20:27

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.094	22:15
Breath Volume	2.859	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.099	22:18
Breath Volume	2.730	22:18
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,DEP.,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WILLIS,DEP.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/24/2024 03:11:33

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.079	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAKELAND,KATLYN,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WAKELAND,KATLYN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: MUELLER

DATE/TIME OF TEST: 12/24/2024 04:20:03

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.168	04:15
Breath Volume	2.003	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample # 2	0.166	04:18
Breath Volume	1.273	04:18
Air Blank	0.000	04:19
Control Test	0.080	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:20

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/24/2024 22:21:47

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.163	22:17
Breath Volume	2.753	22:17
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.174	22:20
Breath Volume	3.714	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CANDELARIA,JULIA
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: RABENSTEIN

DATE/TIME OF TEST: 12/31/2024 07:31:15

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.080	07:24
Air Blank	0.000	07:24
Subject Sample # 1	0.093	07:26
Breath Volume	1.191	07:26
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.087	07:29
Breath Volume	1.183	07:29
Air Blank	0.000	07:30
Control Test	0.080	07:30
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

Data Download Date/Time: 01/17/2025 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINRIGHT, DAVID, T
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: MASSILLON, KEVIN
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 11/15/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 12/02/2024 03:28:37

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.187	03:24
Breath Volume	2.832	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.192	03:27
Breath Volume	2.347	03:27
Air Blank	0.000	03:27
Control Test	0.077	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/05/2024 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO, SOLOMON,
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: BOYLE, M
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GALEANO

DATE/TIME OF TEST: 12/22/2024 06:08:34

Diagnostics Check	OK	06:02
Air Blank	0.000	06:02
Control Test	0.078	06:03
Air Blank	0.000	06:03
Subject Sample # 1	0.137	06:04
Breath Volume	2.066	06:04
Air Blank	0.000	06:04
Air Blank	0.000	06:06
Subject Sample # 2	0.133	06:07
Breath Volume	1.585	06:07
Air Blank	0.000	06:07
Control Test	0.077	06:07
Air Blank	0.000	06:08
Diagnostics Check	OK	06:08

Data Download Date/Time: 01/29/2025 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINRIGHT, DAVID, T
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: GARAFOLLO, F
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 12/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SWENDSON

DATE/TIME OF TEST: 12/28/2024 02:49:19

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.076	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.231	02:44
Breath Volume	2.222	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.245	02:47
Breath Volume	2.503	02:47
Air Blank	0.000	02:48
Control Test	0.076	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 01/29/2025 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HALL,D
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/01/2024 01:51:24

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.082	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.091	01:46
Breath Volume	2.527	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.086	01:49
Breath Volume	2.156	01:49
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 12/05/2024 00:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: CRUMBLEY

DATE/TIME OF TEST: 12/01/2024 15:27:30

Diagnostics Check	OK	15:20
Air Blank	0.000	15:20
Control Test	0.080	15:21
Air Blank	0.000	15:21
Subject Sample # 1	0.350	15:22
Breath Volume	1.878	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:25
Subject Sample # 2	0.359	15:25
Breath Volume	1.789	15:25
Air Blank	0.000	15:26
Control Test	0.075	15:26
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 12/05/2024 00:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 12/01/2024 19:08:08

Diagnostics Check	OK	19:01
Air Blank	0.000	19:02
Control Test	0.081	19:02
Air Blank	0.000	19:03
Subject Sample # 1	0.132	19:03
Breath Volume	1.734	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.135	19:06
Breath Volume	2.074	19:06
Air Blank	0.000	19:07
Control Test	0.080	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 12/05/2024 00:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MELADZE

DATE/TIME OF TEST: 12/03/2024 20:10:13

Diagnostics Check	OK	20:07	
Air Blank	0.000	20:07	
Control Test	0.082	20:07	
Air Blank	0.000	20:08	
Subject Sample	# 1	20:08	Subject Test Refused
Breath Volume	0.000	20:08	
Air Blank	0.000	20:09	
Control Test	0.082	20:09	
Air Blank	0.000	20:10	
Diagnostics Check	OK	20:10	

Data Download Date/Time: 12/05/2024 00:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 12/04/2024 01:28:57

Diagnostics Check	OK	01:22
Air Blank	0.000	01:23
Control Test	0.082	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.167	01:24
Breath Volume	2.718	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.165	01:27
Breath Volume	2.714	01:27
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 12/05/2024 00:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/05/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MERLINO

DATE/TIME OF TEST: 12/04/2024 04:48:47

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.081	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.081	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 12/05/2024 00:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: SOUTH

DATE/TIME OF TEST: 12/06/2024 03:43:24

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.310	03:37
Breath Volume	1.562	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.301	03:41
Breath Volume	1.539	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/16/2025 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FUTCH

DATE/TIME OF TEST: 12/08/2024 00:56:20

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.082	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:49	No Sample Provided
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 2	00:54	No Sample Provided
Breath Volume	0.000	00:54	
Air Blank	0.000	00:55	
Control Test	0.082	00:55	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 01/16/2025 01:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CORMAGI

DATE/TIME OF TEST: 12/12/2024 02:28:17

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.081	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 12/12/2024 03:44:28

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.082	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.081	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: PROSSER

DATE/TIME OF TEST: 12/12/2024 04:52:11

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:49	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:50	Subject Test Refused
Breath Volume	0.000	04:50	
Air Blank	0.000	04:51	
Control Test	0.081	04:51	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: IVANOV

DATE/TIME OF TEST: 12/12/2024 23:14:55

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.154	23:10
Breath Volume	1.652	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.153	23:13
Breath Volume	1.703	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GULI

DATE/TIME OF TEST: 12/13/2024 02:46:26

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.122	02:41
Breath Volume	1.843	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.120	02:44
Breath Volume	1.800	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 12/13/2024 04:15:28

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:08	No Sample Provided
Breath Volume	0.000	04:08	
Air Blank	0.000	04:08	
Air Blank	0.000	04:10	
Subject Sample	# 2	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.082	04:14	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUNINARD,MITCHELL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/14/2024 02:54:46

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/14/2024 04:05:35

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.207	04:01
Breath Volume	1.105	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.215	04:04
Breath Volume	1.339	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CAULFIELD

DATE/TIME OF TEST: 12/14/2024 05:18:48

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:14
Subject Sample # 1	0.137	05:14
Breath Volume	1.968	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:17
Subject Sample # 2	0.130	05:17
Breath Volume	1.398	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAY, III,KENNETH,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SEAY, III,KENNETH
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: OLIPHANT

DATE/TIME OF TEST: 12/14/2024 08:34:25

Diagnostics Check	OK	08:26
Air Blank	0.000	08:27
Control Test	0.081	08:27
Air Blank	0.000	08:27
Subject Sample # 1	0.143	08:29
Breath Volume	1.523	08:29
Air Blank	0.000	08:30
Air Blank	0.000	08:31
Subject Sample # 2	0.157	08:32
Breath Volume	2.062	08:32
Air Blank	0.000	08:33
Control Test	0.080	08:33
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 12/15/2024 00:46:57

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.145	00:42
Breath Volume	1.542	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.144	00:45
Breath Volume	1.839	00:45
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MYRICK

DATE/TIME OF TEST: 12/15/2024 04:19:29

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.079	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Air Blank	0.000	04:17	
Subject Sample	# 2	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: XMCLIN

DATE/TIME OF TEST: 12/18/2024 13:12:58

Diagnostics Check	OK	13:06
Air Blank	0.000	13:07
Control Test	0.080	13:07
Air Blank	0.000	13:08
Subject Sample # 1	0.000	13:08
Breath Volume	1.859	13:08
Air Blank	0.000	13:09
Air Blank	0.000	13:11
Subject Sample # 2	0.000	13:11
Breath Volume	1.773	13:11
Air Blank	0.000	13:12
Control Test	0.080	13:12
Air Blank	0.000	13:12
Diagnostics Check	OK	13:12

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 12/18/2024 18:11:26

Diagnostics Check	OK	18:04
Air Blank	0.000	18:05
Control Test	0.080	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.000	18:06
Breath Volume	2.687	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.000	18:10
Breath Volume	2.535	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: QUALLS,P
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BARNHAM

DATE/TIME OF TEST: 12/20/2024 01:25:44

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.177	01:21
Breath Volume	2.296	01:21
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.180	01:24
Breath Volume	1.593	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ZOBEL

DATE/TIME OF TEST: 12/20/2024 04:05:55

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.171	04:01
Breath Volume	1.191	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.191	04:04
Breath Volume	1.824	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 12/20/2024 04:20:20

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.131	04:15
Breath Volume	1.882	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.127	04:18
Breath Volume	2.105	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ACOSTA DIAZ

DATE/TIME OF TEST: 12/21/2024 03:55:36

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.111	03:51
Breath Volume	2.457	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.108	03:54
Breath Volume	3.199	03:54
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: GODERICH,K
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: CALUNA MOROCHO

DATE/TIME OF TEST: 12/22/2024 00:56:13

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.142	00:51
Breath Volume	2.773	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.140	00:54
Breath Volume	1.566	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 12/26/2024 19:35:22

Diagnostics Check	OK	19:28
Air Blank	0.000	19:29
Control Test	0.080	19:29
Air Blank	0.000	19:30
Subject Sample # 1	0.153	19:30
Breath Volume	1.628	19:30
Air Blank	0.000	19:31
Air Blank	0.000	19:33
Subject Sample # 2	0.154	19:33
Breath Volume	1.890	19:33
Air Blank	0.000	19:34
Control Test	0.079	19:34
Air Blank	0.000	19:35
Diagnostics Check	OK	19:35

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SCHULZ

DATE/TIME OF TEST: 12/27/2024 03:32:17

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.080	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 01/16/2025 01:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/28/2024 02:51:13

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.000	02:46
Breath Volume	3.460	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.000	02:49
Breath Volume	3.468	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,DUSTIN,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,ETHAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/28/2024 03:10:41

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.182	03:01	Volume Not Met
Breath Volume	0.761	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:03	
Subject Sample	# 2 0.180	03:04	
Breath Volume	1.246	03:04	
Air Blank	0.000	03:05	
Air Blank	0.000	03:07	
Subject Sample	# 3 0.186	03:09	
Breath Volume	2.062	03:09	
Air Blank	0.000	03:09	
Control Test	0.078	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:10	

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 12/28/2024 23:55:17

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:52	
Air Blank	0.000	23:52	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:54	
Control Test	0.080	23:54	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/29/2024 02:19:09

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.212	02:14
Breath Volume	2.058	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.207	02:17
Breath Volume	1.230	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEAU,ARIC,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:11

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 12/31/2024 15:38:20

Diagnostics Check	OK	15:31
Air Blank	0.000	15:31
Control Test	0.081	15:31
Air Blank	0.000	15:32
Subject Sample # 1	0.000	15:33
Breath Volume	1.480	15:33
Air Blank	0.000	15:34
Air Blank	0.000	15:36
Subject Sample # 2	0.000	15:36
Breath Volume	1.207	15:36
Air Blank	0.000	15:37
Control Test	0.080	15:37
Air Blank	0.000	15:38
Diagnostics Check	OK	15:38

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,E
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A1
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: FARLEY

DATE/TIME OF TEST: 12/31/2024 22:47:38

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.153	22:43
Breath Volume	2.757	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.149	22:46
Breath Volume	2.644	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 01/16/2025 01:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: UPCHURCH,JERRY,S

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: EARLY,D

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 20:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 12/10/2024 21:32:14

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:25	No Sample Provided
Breath Volume	0.000	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:30	No Sample Provided
Breath Volume	0.000	21:30	
Air Blank	0.000	21:31	
Control Test	0.080	21:31	
Air Blank	0.000	21:32	
Diagnostics Check	OK	21:32	

Data Download Date/Time: 12/24/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 21:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 12/11/2024 22:24:13

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.189	22:19
Breath Volume	2.144	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.184	22:22
Breath Volume	1.980	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 12/24/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERMAK,OFC,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: LOUX,SAGE
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: WILDMAN

DATE/TIME OF TEST: 12/14/2024 02:02:31

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1 0.219	01:51	
Breath Volume	1.464	01:51	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample	# 2 0.207	01:57	Volume Not Met
Breath Volume	0.050	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 3 0.223	02:01	
Breath Volume	1.300	02:01	
Air Blank	0.000	02:01	
Control Test	0.078	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 12/24/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 22:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: GREVE

DATE/TIME OF TEST: 12/14/2024 22:56:00

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.180	22:51
Breath Volume	2.164	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.182	22:54
Breath Volume	2.871	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:56

Data Download Date/Time: 12/24/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 11/18/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 00:58

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RACAK

DATE/TIME OF TEST: 12/21/2024 01:49:50

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.058	01:45
Breath Volume	1.449	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.058	01:48
Breath Volume	1.351	01:48
Air Blank	0.000	01:48
Control Test	0.080	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 12/24/2024 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: SINGSON,ROLAND

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 12/24/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 01:13

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/28/2024 01:57:19

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.075	01:52
Breath Volume	1.812	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.072	01:55
Breath Volume	1.566	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/20/2025 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SINGSON,ROLAND,A

BREATH TEST OPERATOR AGENCY: Largo PD

ARRESTING OFFICER: LOUX,SAGE

ARRESTING OFFICER AGENCY: Largo PD

LAST AGENCY INSPECTION: 12/24/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 01:11

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CONVERSI

DATE/TIME OF TEST: 12/28/2024 02:13:46

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:11	
Control Test	0.081	02:11	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:12	Subject Test Refused
Breath Volume	0.000	02:12	
Air Blank	0.000	02:12	
Control Test	0.079	02:13	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 01/20/2025 09:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PEREZ HERRERA

DATE/TIME OF TEST: 12/01/2024 01:39:00

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.158	01:34
Breath Volume	1.246	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.162	01:37
Breath Volume	1.441	01:37
Air Blank	0.000	01:38
Control Test	0.077	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINWRIGHT,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BRACKLEY,RONALD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: KNECHT III

DATE/TIME OF TEST: 12/01/2024 19:02:30

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.080	18:56
Air Blank	0.000	18:56
Subject Sample # 1	0.000	18:57
Breath Volume	1.132	18:57
Air Blank	0.000	18:57
Air Blank	0.000	18:59
Subject Sample # 2	0.000	19:01
Breath Volume	1.152	19:01
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:02
Diagnostics Check	OK	19:02

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAINWRIGHT,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/01/2024 20:02:28

Diagnostics Check	OK	19:55
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.284	19:57
Breath Volume	1.566	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.288	20:00
Breath Volume	1.921	20:00
Air Blank	0.000	20:01
Control Test	0.078	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/03/2024 22:47:36

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.081	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.000	22:43
Breath Volume	1.234	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.000	22:46
Breath Volume	1.914	22:46
Air Blank	0.000	22:46
Control Test	0.080	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 12/10/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER, LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 12/04/2024 19:43:23

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.081	19:35
Air Blank	0.000	19:35
Subject Sample # 1	0.159	19:37
Breath Volume	1.277	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.141	19:41
Breath Volume	1.429	19:41
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: BAUTISTA FERIA

DATE/TIME OF TEST: 12/06/2024 22:48:10

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.183	22:43
Breath Volume	2.242	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.179	22:46
Breath Volume	2.308	22:46
Air Blank	0.000	22:47
Control Test	0.080	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX, ANDREW, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIVERA, CHRISTIAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: RIVERA GUZMAN

DATE/TIME OF TEST: 12/06/2024 23:16:57

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.169	23:12
Breath Volume	1.570	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.178	23:15
Breath Volume	1.421	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DRISKELL

DATE/TIME OF TEST: 12/06/2024 23:47:35

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1 0.211	23:43	
Breath Volume	2.160	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 2	23:46	RFI Detect
Breath Volume	0.000	23:46	
Air Blank	0.000	23:46	
Control Test	0.080	23:46	
Air Blank	0.000	23:47	
Diagnostics Check	OK	23:47	

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DRISKELL

DATE/TIME OF TEST: 12/06/2024 23:54:57

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.208	23:50
Breath Volume	2.019	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.195	23:53
Breath Volume	1.375	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,ARIK,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JONES,ARIK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SEAY

DATE/TIME OF TEST: 12/07/2024 03:54:28

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.098	03:48
Breath Volume	1.460	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.088	03:52
Breath Volume	1.417	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/08/2024 03:33:57

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.081	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:32	No Sample Provided
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.081	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: WASELUES

DATE/TIME OF TEST: 12/09/2024 04:32:11

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.202	04:27
Breath Volume	1.523	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.201	04:30
Breath Volume	1.640	04:30
Air Blank	0.000	04:31
Control Test	0.080	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 12/10/2024 09:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/11/2024 21:58:57

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.082	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.000	21:54
Breath Volume	1.886	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.000	21:57
Breath Volume	2.093	21:57
Air Blank	0.000	21:58
Control Test	0.081	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 01/15/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: CANEL

DATE/TIME OF TEST: 12/13/2024 23:24:50

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Subject Sample # 1	0.162	23:20
Breath Volume	1.667	23:20
Air Blank	0.000	23:21
Air Blank	0.000	23:22
Subject Sample # 2	0.160	23:23
Breath Volume	3.304	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/14/2024 19:17:43

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.230	19:13
Breath Volume	1.234	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.242	19:16
Breath Volume	1.167	19:16
Air Blank	0.000	19:16
Control Test	0.078	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FLORES,MARIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 12/15/2024 02:50:14

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.179	02:45
Breath Volume	1.414	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.183	02:48
Breath Volume	1.636	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PERALTA,ALLISON
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ORTIZ MEDINA

DATE/TIME OF TEST: 12/15/2024 04:04:27

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.250	03:59
Breath Volume	2.574	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.257	04:02
Breath Volume	2.378	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/15/2024 10:31:03

Diagnostics Check	OK	10:25
Air Blank	0.000	10:25
Control Test	0.081	10:25
Air Blank	0.000	10:26
Subject Sample # 1	0.000	10:26
Breath Volume	2.578	10:26
Air Blank	0.000	10:27
Air Blank	0.000	10:29
Subject Sample # 2	0.000	10:29
Breath Volume	2.894	10:29
Air Blank	0.000	10:30
Control Test	0.081	10:30
Air Blank	0.000	10:30
Diagnostics Check	OK	10:31

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH, MALACHI, D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: QUEEN, ANDRE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: BYRON

DATE/TIME OF TEST: 12/15/2024 14:19:01

Diagnostics Check	OK	14:11
Air Blank	0.000	14:11
Control Test	0.081	14:11
Air Blank	0.000	14:12
Subject Sample # 1	0.011	14:14
Breath Volume	1.550	14:14
Air Blank	0.000	14:15
Air Blank	0.000	14:17
Subject Sample # 2	0.011	14:17
Breath Volume	1.941	14:17
Air Blank	0.000	14:18
Control Test	0.080	14:18
Air Blank	0.000	14:18
Diagnostics Check	OK	14:19

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: RESENDIZ

DATE/TIME OF TEST: 12/16/2024 22:28:18

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.000	22:24
Breath Volume	1.378	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.000	22:26
Breath Volume	1.628	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 12/17/2024 20:47:27

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.081	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.000	20:43
Breath Volume	1.996	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.000	20:46
Breath Volume	1.917	20:46
Air Blank	0.000	20:46
Control Test	0.081	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GOMEZ PEREZ

DATE/TIME OF TEST: 12/17/2024 21:02:15

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.033	20:57
Breath Volume	2.218	20:57
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.033	21:00
Breath Volume	1.707	21:00
Air Blank	0.000	21:01
Control Test	0.080	21:01
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: JERRY

DATE/TIME OF TEST: 12/17/2024 23:15:26

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.081	23:09
Air Blank	0.000	23:10
Subject Sample # 1	0.316	23:10
Breath Volume	2.300	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.311	23:13
Breath Volume	2.382	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MARAZZI

DATE/TIME OF TEST: 12/18/2024 00:50:07

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:45
Breath Volume	2.578	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:48
Breath Volume	2.375	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,JAMES,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DANIEL,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: DARLEY

DATE/TIME OF TEST: 12/20/2024 05:32:51

Diagnostics Check	OK	05:21	
Air Blank	0.000	05:21	
Control Test	0.080	05:22	
Air Blank	0.000	05:22	
Subject Sample	# 1	05:25	No Sample Provided
Breath Volume	0.000	05:25	
Air Blank	0.000	05:26	
Air Blank	0.000	05:28	
Subject Sample	# 2	05:31	No Sample Provided
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Control Test	0.080	05:32	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: AUDEL

DATE/TIME OF TEST: 12/20/2024 21:03:28

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.081	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.000	20:59
Breath Volume	1.855	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:02
Breath Volume	1.773	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,JAMES,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DANIEL,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: GREATHOUSE

DATE/TIME OF TEST: 12/21/2024 03:42:56

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:41	RFI Detect
Breath Volume	1.011	03:41	
Air Blank	0.000	03:42	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 01/15/2025 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: TEZEN

DATE/TIME OF TEST: 12/21/2024 04:28:56

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:18	
Control Test	0.080	04:18	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:21	No Sample Provided
Breath Volume	0.000	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample	# 2	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:28	
Control Test	0.081	04:28	
Air Blank	0.000	04:28	
Diagnostics Check	OK	04:28	

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: MARIO

DATE/TIME OF TEST: 12/22/2024 06:27:43

Diagnostics Check	OK	06:16	
Air Blank	0.000	06:16	
Control Test	0.081	06:17	
Air Blank	0.000	06:17	
Subject Sample	# 1	06:20	No Sample Provided
Breath Volume	0.000	06:20	
Air Blank	0.000	06:21	
Air Blank	0.000	06:23	
Subject Sample	# 2	06:26	No Sample Provided
Breath Volume	0.000	06:26	
Air Blank	0.000	06:26	
Control Test	0.082	06:27	
Air Blank	0.000	06:27	
Diagnostics Check	OK	06:27	

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: DIAZ GUZMAN

DATE/TIME OF TEST: 12/22/2024 08:15:27

Diagnostics Check	OK	08:09
Air Blank	0.000	08:09
Control Test	0.081	08:10
Air Blank	0.000	08:10
Subject Sample # 1	0.109	08:11
Breath Volume	2.113	08:11
Air Blank	0.000	08:11
Air Blank	0.000	08:13
Subject Sample # 2	0.097	08:13
Breath Volume	1.367	08:13
Air Blank	0.000	08:14
Control Test	0.081	08:14
Air Blank	0.000	08:15
Diagnostics Check	OK	08:15

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 12/22/2024 22:00:17

Diagnostics Check	OK	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:52
Subject Sample # 1	0.218	21:53
Breath Volume	1.640	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.199	21:58
Breath Volume	1.867	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,CARL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PAMPIN

DATE/TIME OF TEST: 12/22/2024 23:11:34

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.222	23:07
Breath Volume	1.585	23:07
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.233	23:10
Breath Volume	1.464	23:10
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 12/23/2024 03:47:51

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.256	03:43
Breath Volume	1.535	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.251	03:46
Breath Volume	2.019	03:46
Air Blank	0.000	03:46
Control Test	0.080	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: MATLA RANGEL

DATE/TIME OF TEST: 12/25/2024 06:17:06

Diagnostics Check	OK	06:06
Air Blank	0.000	06:07
Control Test	0.080	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.140	06:10
Breath Volume	1.187	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.141	06:15
Breath Volume	1.117	06:15
Air Blank	0.000	06:16
Control Test	0.078	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,JAMES,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DANIEL,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/25/2024 06:30:48

Diagnostics Check	OK	06:24
Air Blank	0.000	06:24
Control Test	0.078	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.256	06:26
Breath Volume	1.207	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:28
Subject Sample # 2	0.256	06:29
Breath Volume	1.476	06:29
Air Blank	0.000	06:29
Control Test	0.079	06:30
Air Blank	0.000	06:30
Diagnostics Check	OK	06:30

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GLASER,MATTHEW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: PEREZ SEQUEIRA

DATE/TIME OF TEST: 12/25/2024 07:52:46

Diagnostics Check	OK	07:45
Air Blank	0.000	07:45
Control Test	0.080	07:46
Air Blank	0.000	07:46
Subject Sample # 1	0.182	07:48
Breath Volume	1.277	07:48
Air Blank	0.000	07:48
Air Blank	0.000	07:50
Subject Sample # 2	0.196	07:51
Breath Volume	1.351	07:51
Air Blank	0.000	07:51
Control Test	0.079	07:52
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALCALDE,ALEXANDER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: PALACIOS VASQUEZ

DATE/TIME OF TEST: 12/25/2024 13:26:16

Diagnostics Check	OK	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Subject Sample # 1	0.162	13:21
Breath Volume	1.722	13:21
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.172	13:24
Breath Volume	1.953	13:24
Air Blank	0.000	13:25
Control Test	0.078	13:25
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: LEON LUCAS

DATE/TIME OF TEST: 12/25/2024 20:29:05

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.080	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.189	20:24
Breath Volume	1.300	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.195	20:27
Breath Volume	1.750	20:27
Air Blank	0.000	20:28
Control Test	0.080	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,AHNIYA,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/26/2024 00:18:05

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2	00:16	No Sample Provided
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.081	00:17	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/26/2024 02:06:35

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.158	01:56	Volume Not Met
Breath Volume	0.085	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2 0.162	01:59	
Breath Volume	1.214	01:59	
Air Blank	0.000	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 3 0.147	02:04	Volume Not Met
Breath Volume	0.324	02:04	
Air Blank	0.000	02:05	
Control Test	0.079	02:05	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:06	

Data Download Date/Time: 01/15/2025 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/27/2024 01:06:05

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.000	01:01
Breath Volume	2.031	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.000	01:04
Breath Volume	1.675	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:06

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/27/2024 04:00:06

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.080	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 12/28/2024 00:52:39

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.000	00:48
Breath Volume	2.566	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.000	00:51
Breath Volume	2.242	00:51
Air Blank	0.000	00:51
Control Test	0.081	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISSOM,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GRISSOM,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: PREO

DATE/TIME OF TEST: 12/28/2024 18:45:42

Diagnostics Check	OK	18:33	
Air Blank	0.000	18:34	
Control Test	0.080	18:34	
Air Blank	0.000	18:35	
Subject Sample	# 1 0.066	18:38	Volume Not Met
Breath Volume	0.351	18:38	
Air Blank	0.000	18:39	
Air Blank	0.000	18:40	
Subject Sample	# 2 0.079	18:41	
Breath Volume	1.593	18:41	
Air Blank	0.000	18:42	
Air Blank	0.000	18:43	
Subject Sample	# 3 0.080	18:44	
Breath Volume	1.371	18:44	
Air Blank	0.000	18:44	
Control Test	0.074	18:45	Control Outside Tolerance
Air Blank	0.000	18:45	

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO, DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: GARGUS

DATE/TIME OF TEST: 12/29/2024 00:38:29

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.081	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.174	00:33
Breath Volume	3.109	00:33
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.171	00:36
Breath Volume	1.851	00:36
Air Blank	0.000	00:37
Control Test	0.079	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/29/2024 01:11:47

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.155	01:05
Breath Volume	1.191	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.155	01:10
Breath Volume	1.179	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CRUZ-BARRERA

DATE/TIME OF TEST: 12/30/2024 21:33:30

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.079	21:28
Air Blank	0.000	21:28
Subject Sample # 1	0.000	21:29
Breath Volume	1.722	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:32
Breath Volume	1.164	21:32
Air Blank	0.000	21:32
Control Test	0.080	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,J
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/30/2024 21:58:35

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.000	21:54
Breath Volume	1.695	21:54
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.000	21:57
Breath Volume	2.128	21:57
Air Blank	0.000	21:57
Control Test	0.081	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REGISTER,JUSTIN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: REGISTER,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: WILSON JR

DATE/TIME OF TEST: 12/31/2024 03:21:28

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.000	03:17
Breath Volume	1.593	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.000	03:20
Breath Volume	1.210	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 01/15/2025 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FANIZZA

DATE/TIME OF TEST: 12/03/2024 01:37:31

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.117	01:33
Breath Volume	1.632	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.118	01:36
Breath Volume	1.472	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMELMAN, TYLER,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GUTIERREZ, R
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 12/03/2024 08:56:28

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.081	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.000	08:51
Breath Volume	1.257	08:51
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.000	08:55
Breath Volume	2.203	08:55
Air Blank	0.000	08:55
Control Test	0.081	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 12/03/2024 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIOT,XAVIER,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: CAPOTA,J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SALDIVA

DATE/TIME OF TEST: 12/14/2024 23:57:05

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.079	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 1 0.264	23:52	
Breath Volume	1.906	23:52	
Air Blank	0.000	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 2 0.247	23:55	
Breath Volume	1.312	23:55	
Air Blank		23:56	Purge Fail
Air Blank	0.000	23:57	

Data Download Date/Time: 12/31/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIOT,XAVIER,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: CAPOTA,J
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SALDIVA

DATE/TIME OF TEST: 12/15/2024 00:08:46

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.076	00:02
Air Blank	0.000	00:03
Subject Sample # 1	0.264	00:04
Breath Volume	2.074	00:04
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.264	00:07
Breath Volume	2.304	00:07
Air Blank	0.000	00:07
Control Test	0.077	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/31/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231603
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MAHONEY

DATE/TIME OF TEST: 12/28/2024 20:41:01

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.078	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	RFI Detect
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Control Test	0.079	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 12/31/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 231603
BREATH TEST OPERATOR: GOMEZ,DESTINEE,R	DRY GAS EXPIRY DATE: 03/05/2026
BREATH TEST OPERATOR AGENCY: Plantation PD	VIOLATION CODE: DUI
ARRESTING OFFICER: GOMEZ,DESTINEE	ARREST TIME: 20:07
ARRESTING OFFICER AGENCY: Plantation PD	GENDER: Male
LAST AGENCY INSPECTION: 11/29/2024	OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MAHONEY	DATE/TIME OF TEST: 12/28/2024	20:50:29
------------------------------	-------------------------------	----------

Diagnostics Check	OK	20:43
Air Blank	0.000	20:43
Control Test	0.078	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.209	20:45
Breath Volume	2.414	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.202	20:48
Breath Volume	1.246	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 12/31/2024 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ALBRITTON, ROY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: REILLY

DATE/TIME OF TEST: 12/01/2024 20:16:44

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.082	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.226	20:12
Breath Volume	2.261	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.223	20:15
Breath Volume	2.757	20:15
Air Blank	0.000	20:15
Control Test	0.082	20:16
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 12/03/2024 01:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PATTON 2

DATE/TIME OF TEST: 12/23/2024 00:10:10

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.082	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 01/28/2025 01:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 12/23/2024 00:40:07

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.082	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.257	00:35
Breath Volume	1.167	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:37
Subject Sample # 2	0.271	00:38
Breath Volume	1.152	00:38
Air Blank	0.000	00:39
Control Test	0.082	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 01/28/2025 01:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER, WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WINSHIP

DATE/TIME OF TEST: 12/14/2024 02:29:26

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Diagnostics Check	OK	02:29	

Data Download Date/Time: 12/30/2024 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIEDINGER,R,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WELLS,K
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/30/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1433248
DRY GAS EXPIRY DATE: 01/06/2024
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/28/2024 17:13:29

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.082	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.081	17:09
Breath Volume	1.656	17:09
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.076	17:11
Breath Volume	1.347	17:11
Air Blank	0.000	17:12
Control Test	0.080	17:12
Air Blank	0.000	17:13
Diagnostics Check	OK	17:13

Data Download Date/Time: 12/30/2024 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: DEES,BENJAMIN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/20/2024 07:53:23

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.000	07:49
Breath Volume	3.527	07:49
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.000	07:52
Breath Volume	3.769	07:52
Air Blank	0.000	07:52
Control Test	0.079	07:52
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 01/17/2025 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MCCORMICK,JOSEPH
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402635303
DRY GAS EXPIRY DATE: 01/25/2026
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: BELLES

DATE/TIME OF TEST: 12/23/2024 10:44:09

Diagnostics Check	OK	10:37
Air Blank	0.000	10:38
Control Test	0.081	10:38
Air Blank	0.000	10:38
Subject Sample # 1	0.096	10:39
Breath Volume	2.039	10:39
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.095	10:42
Breath Volume	1.734	10:42
Air Blank	0.000	10:43
Control Test	0.080	10:43
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 01/17/2025 19:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRAGUE,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SPRAGUE,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 12/13/2024 21:17:21

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.081	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.192	21:13
Breath Volume	1.523	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.198	21:15
Breath Volume	1.328	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 12/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402372586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: ELIAS PEREZ

DATE/TIME OF TEST: 12/20/2024 00:54:41

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.081	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.093	00:49
Breath Volume	3.332	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.095	00:53
Breath Volume	2.421	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 12/26/2024 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER, LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ALINDER

DATE/TIME OF TEST: 12/01/2024 06:37:38

Diagnostics Check	OK	06:21	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:22	
Subject Sample	# 1 0.149	06:25	Volume Not Met
Breath Volume	0.484	06:25	
Air Blank	0.000	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 2 0.153	06:30	
Breath Volume	1.199	06:30	
Air Blank	0.000	06:31	
Air Blank	0.000	06:32	
Subject Sample	# 3 0.145	06:35	Volume Not Met
Breath Volume	0.718	06:35	
Air Blank	0.000	06:36	
Control Test	0.075	06:37	
Air Blank	0.000	06:37	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER, LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: ALINDOR

DATE/TIME OF TEST: 12/01/2024 06:53:10

Diagnostics Check		OK	06:39	
Air Blank		0.000	06:39	
Control Test		0.077	06:40	
Air Blank		0.000	06:40	
Subject Sample	# 1	0.147	06:43	Volume Not Met
Breath Volume		0.882	06:43	
Air Blank		0.000	06:44	
Air Blank		0.000	06:46	
Subject Sample	# 2	0.147	06:47	
Breath Volume		1.179	06:47	
Air Blank		0.000	06:48	
Air Blank		0.000	06:50	
Subject Sample	# 3	0.157	06:51	
Breath Volume		1.269	06:51	
Air Blank		0.000	06:52	
Control Test		0.077	06:52	
Air Blank		0.000	06:53	
Diagnostics Check		OK	06:53	

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 12/03/2024 03:07:10

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.235	03:02
Breath Volume	1.734	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.231	03:05
Breath Volume	1.726	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,REBECCA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 12/05/2024 12:01:23

Diagnostics Check	OK	11:54
Air Blank	0.000	11:55
Control Test	0.079	11:55
Air Blank	0.000	11:55
Subject Sample # 1	0.263	11:57
Breath Volume	2.082	11:57
Air Blank	0.000	11:57
Air Blank	0.000	11:59
Subject Sample # 2	0.267	11:59
Breath Volume	1.738	11:59
Air Blank	0.000	12:00
Control Test	0.079	12:00
Air Blank	0.000	12:01
Diagnostics Check	OK	12:01

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METELLUS,SAM,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 12/06/2024 21:33:43

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.036	21:29
Breath Volume	1.277	21:29
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.038	21:32
Breath Volume	1.656	21:32
Air Blank	0.000	21:32
Control Test	0.079	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,NIKOLAUS,S
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 12/07/2024 05:30:31

Diagnostics Check	OK	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.151	05:25
Breath Volume	4.062	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.150	05:29
Breath Volume	3.503	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: GAINES

DATE/TIME OF TEST: 12/08/2024 08:38:18

Diagnostics Check	OK	08:31
Air Blank	0.000	08:32
Control Test	0.081	08:32
Air Blank	0.000	08:32
Subject Sample # 1	0.149	08:33
Breath Volume	2.644	08:33
Air Blank	0.000	08:34
Air Blank	0.000	08:36
Subject Sample # 2	0.140	08:36
Breath Volume	2.050	08:36
Air Blank	0.000	08:37
Control Test	0.080	08:37
Air Blank	0.000	08:38
Diagnostics Check	OK	08:38

Data Download Date/Time: 12/10/2024 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04924080A5
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: DUVERNO

DATE/TIME OF TEST: 12/09/2024 00:08:47

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.080	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.172	00:03
Breath Volume	1.144	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.186	00:07
Breath Volume	1.546	00:07
Air Blank	0.000	00:07
Control Test	0.078	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 12/10/2024 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 284240080A
DRY GAS EXPIRY DATE: 11/05/2020
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/10/2024 16:53:58

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:43	
Control Test	0.081	16:43	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:47	No Sample Provided
Breath Volume	0.000	16:47	
Air Blank	0.000	16:47	
Air Blank	0.000	16:49	
Subject Sample	# 2	16:52	No Sample Provided
Breath Volume	0.000	16:52	
Air Blank	0.000	16:53	
Control Test	0.080	16:53	
Air Blank	0.000	16:53	
Diagnostics Check	OK	16:53	

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,KIMBERLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DYE,BRANDON
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 284240080A
DRY GAS EXPIRY DATE: 11/05/2020
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:03

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 12/13/2024 09:35:45

Diagnostics Check	OK	09:29
Air Blank	0.000	09:30
Control Test	0.082	09:30
Air Blank	0.000	09:31
Subject Sample # 1	0.007	09:31
Breath Volume	3.398	09:31
Air Blank	0.000	09:31
Air Blank	0.000	09:33
Subject Sample # 2	0.000	09:34
Breath Volume	4.339	09:34
Air Blank	0.000	09:34
Control Test	0.080	09:35
Air Blank	0.000	09:35
Diagnostics Check	OK	09:35

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLISER,SHAUN,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WALLISER,SHAUN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 284240080A
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 12/13/2024 18:53:57

Diagnostics Check	OK	18:46
Air Blank	0.000	18:46
Control Test	0.081	18:47
Air Blank	0.000	18:47
Subject Sample # 1	0.277	18:48
Breath Volume	1.453	18:48
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.262	18:52
Breath Volume	1.144	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,MELISSA,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WILLIAMS,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 284240080A
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 12/13/2024 22:02:27

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.081	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:55	No Sample Provided
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2	22:01	No Sample Provided
Breath Volume	0.000	22:01	
Air Blank	0.000	22:01	
Control Test	0.080	22:01	
Air Blank	0.000	22:02	
Diagnostics Check	OK	22:02	

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,MALACHI,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RANDOLPH,MALACHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 284240080A
DRY GAS EXPIRY DATE: 11/05/2020
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: RIAN

DATE/TIME OF TEST: 12/15/2024 12:27:39

Diagnostics Check	OK	12:21
Air Blank	0.000	12:21
Control Test	0.081	12:21
Air Blank	0.000	12:22
Subject Sample # 1	0.000	12:22
Breath Volume	2.097	12:22
Air Blank	0.000	12:23
Air Blank	0.000	12:25
Subject Sample # 2	0.000	12:26
Breath Volume	1.640	12:26
Air Blank	0.000	12:26
Control Test	0.081	12:27
Air Blank	0.000	12:27
Diagnostics Check	OK	12:27

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIEL,JAMES,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DANIEL,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 12/19/2024 02:09:09

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.081	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:02	No Sample Provided
Breath Volume	0.000	02:02	
Air Blank	0.000	02:02	
Air Blank	0.000	02:04	
Subject Sample	# 2	02:07	No Sample Provided
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.081	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON,KIARA,F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VILLEGAS

DATE/TIME OF TEST: 12/21/2024 00:58:20

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:53
Breath Volume	1.304	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:56
Breath Volume	1.605	00:56
Air Blank	0.000	00:57
Control Test	0.080	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLON, KIARA, F
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS, WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/21/2024 23:28:34

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.082	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.138	23:24
Breath Volume	3.289	23:24
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.145	23:27
Breath Volume	3.468	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: FRANTZ

DATE/TIME OF TEST: 12/22/2024 06:24:01

Diagnostics Check	OK	06:17
Air Blank	0.000	06:17
Control Test	0.081	06:18
Air Blank	0.000	06:18
Subject Sample # 1	0.000	06:19
Breath Volume	1.863	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.000	06:22
Breath Volume	1.710	06:22
Air Blank	0.000	06:23
Control Test	0.082	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:24

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDRUP,STANLEY,O
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/22/2024 06:40:49

Diagnostics Check	OK	06:29	
Air Blank	0.000	06:29	
Control Test	0.081	06:30	
Air Blank	0.000	06:30	
Subject Sample	# 1	06:33	No Sample Provided
Breath Volume	0.000	06:33	
Air Blank	0.000	06:34	
Air Blank	0.000	06:36	
Subject Sample	# 2	06:39	No Sample Provided
Breath Volume	0.000	06:39	
Air Blank	0.000	06:39	
Control Test	0.082	06:40	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWEN, WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/27/2024 04:24:32

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.000	04:20
Breath Volume	1.199	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:23
Breath Volume	1.457	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAWSON JR, MICHAEL, E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JOHNSON, MARCUS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: STRIGLE II

DATE/TIME OF TEST: 12/28/2024 04:30:49

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.241	04:26
Breath Volume	2.890	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:28
Subject Sample # 2	0.260	04:29
Breath Volume	2.207	04:29
Air Blank	0.000	04:29
Control Test	0.079	04:30
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: FLORES CHAVES

DATE/TIME OF TEST: 12/29/2024 04:33:17

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.135	04:27
Breath Volume	2.109	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.132	04:31
Breath Volume	1.714	04:31
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 01/14/2025 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,COREY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ANGULO,DALLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/30/2024 04:25:17

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.168	04:20
Breath Volume	2.050	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.172	04:23
Breath Volume	2.753	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 01/14/2025 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX, ANDREW, R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: EVERINGHAM, JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ALLENDES DONOSO

DATE/TIME OF TEST: 12/31/2024 05:48:43

Diagnostics Check	OK	05:42
Air Blank	0.000	05:42
Control Test	0.081	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.164	05:44
Breath Volume	1.796	05:44
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.162	05:47
Breath Volume	1.937	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:48
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 01/14/2025 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAVRANEK, ADAM, J
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: HAVRANEK, ADAM
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 12/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ARIAS ESCALONA

DATE/TIME OF TEST: 12/31/2024 22:16:59

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:15	Subject Test Refused
Breath Volume	0.000	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:16	
Diagnostics Check	OK	22:16	

Data Download Date/Time: 01/14/2025 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Palatka PD
ARRESTING OFFICER: WILLIAMS,SHELTON
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/07/2024 02:21:45

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.081	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.139	02:17
Breath Volume	3.121	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.140	02:20
Breath Volume	4.199	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 12/30/2024 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,KEATON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: RAY,KEATON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:02

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/14/2024 09:32:10

Diagnostics Check	OK	09:25
Air Blank	0.000	09:25
Control Test	0.080	09:26
Air Blank	0.000	09:26
Subject Sample # 1	0.134	09:27
Breath Volume	1.191	09:27
Air Blank	0.000	09:28
Air Blank	0.000	09:30
Subject Sample # 2	0.127	09:30
Breath Volume	1.351	09:30
Air Blank	0.000	09:31
Control Test	0.080	09:31
Air Blank	0.000	09:32
Diagnostics Check	OK	09:32

Data Download Date/Time: 12/30/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001128

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELDER,BRIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ELDER,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/21/2024 22:34:48

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.081	22:29
Air Blank	0.000	22:29
Subject Sample # 1	0.239	22:30
Breath Volume	3.585	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.243	22:33
Breath Volume	3.289	22:33
Air Blank	0.000	22:33
Control Test	0.080	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 12/30/2024 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: SANTOS,KEELAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: STALLING

DATE/TIME OF TEST: 12/01/2024 23:42:46

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.080	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.044	23:38
Breath Volume	1.671	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.045	23:41
Breath Volume	1.359	23:41
Air Blank	0.000	23:41
Control Test	0.080	23:42
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 12/16/2024 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKRZYPINSKI,RICHARD,C
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATTERSON,ASHLEY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4022872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SOSTENES

DATE/TIME OF TEST: 12/20/2024 22:29:17

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.089	22:24
Breath Volume	1.238	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.095	22:27
Breath Volume	2.089	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 01/29/2025 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITKE,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: CZAJKOWSKI

DATE/TIME OF TEST: 12/04/2024 03:34:31

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.082	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.048	03:30
Breath Volume	3.308	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.048	03:33
Breath Volume	3.679	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CLARK,ADDISON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/04/2024 20:44:30

Diagnostics Check	OK	20:38
Air Blank	0.000	20:39
Control Test	0.080	20:39
Air Blank	0.000	20:40
Subject Sample # 1	0.000	20:40
Breath Volume	1.101	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.000	20:43
Breath Volume	2.304	20:43
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BUETTNER,CADEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/06/2024 02:36:32

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:32	
Subject Sample	# 1 0.195	02:32	
Breath Volume	3.203	02:32	
Air Blank	0.000	02:33	
Air Blank	0.000	02:35	
Subject Sample	# 2 0.194	02:35	
Breath Volume	2.925	02:35	
Air Blank		02:36	RFI Detect
Air Blank	0.000	02:36	

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BARRY,KEVIN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 12/08/2024 02:04:22

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.093	02:00
Breath Volume	1.464	02:00
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.090	02:02
Breath Volume	2.847	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BABA,TANNER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: SEIFERTERT

DATE/TIME OF TEST: 12/11/2024 06:14:27

Diagnostics Check	OK	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.209	06:09
Breath Volume	1.363	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.220	06:12
Breath Volume	2.179	06:12
Air Blank	0.000	06:13
Control Test	0.079	06:13
Air Blank	0.000	06:14
Diagnostics Check	OK	06:14

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/13/2024 23:01:37

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.081	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.179	22:57
Breath Volume	2.226	22:57
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.193	23:00
Breath Volume	1.855	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 12/16/2024 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERFEL,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WERFEL,JEFFREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 12/20/2024 22:03:53

Diagnostics Check	OK	22:02	
Air Blank		22:03	Ambient Fail
Air Blank	0.000	22:03	

Data Download Date/Time: 01/29/2025 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERFEL,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WERFEL,JEFFREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 12/20/2024 22:19:37

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.080	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1 0.114	22:10	Volume Not Met
Breath Volume	0.632	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2 0.171	22:14	
Breath Volume	2.589	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample	# 3 0.166	22:18	
Breath Volume	2.113	22:18	
Air Blank	0.000	22:18	
Control Test	0.078	22:19	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 01/29/2025 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: VILLALOBOS MORA

DATE/TIME OF TEST: 12/23/2024 19:48:30

Diagnostics Check	OK	19:40
Air Blank	0.000	19:40
Control Test	0.079	19:41
Air Blank	0.000	19:41
Subject Sample # 1	0.162	19:43
Breath Volume	2.531	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.162	19:47
Breath Volume	3.433	19:47
Air Blank	0.000	19:47
Control Test	0.080	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 01/29/2025 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ZARNOWIEC

DATE/TIME OF TEST: 12/24/2024 02:11:20

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.120	02:06
Breath Volume	2.625	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.117	02:09
Breath Volume	2.679	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/29/2025 16:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CLARK,ADDISON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ZALAVARRIA REYES

DATE/TIME OF TEST: 12/29/2024 03:43:57

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.175	03:39
Breath Volume	1.855	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.170	03:42
Breath Volume	1.429	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 01/29/2025 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODIN,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WOODIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: PANASCI

DATE/TIME OF TEST: 12/30/2024 01:34:38

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.079	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.161	01:30
Breath Volume	3.628	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.159	01:33
Breath Volume	3.992	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 01/29/2025 16:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,NICKOLAS,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MARTIN,CODY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 12/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: AMOROSI

DATE/TIME OF TEST: 12/31/2024 20:33:37

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.080	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.122	20:29
Breath Volume	2.593	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.118	20:32
Breath Volume	2.292	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 01/29/2025 16:21