

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/05/2024 22:31:45

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.079	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.100	22:27
Breath Volume	1.410	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.098	22:30
Breath Volume	1.796	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 12/06/2024 01:10:01

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.079	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.042	01:05
Breath Volume	3.238	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.041	01:08
Breath Volume	1.679	01:08
Air Blank	0.000	01:09
Control Test	0.080	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:10

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 12/20/2024 00:58:34

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.000	00:54
Breath Volume	3.207	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.000	00:57
Breath Volume	3.261	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/20/2024 23:24:22

Diagnostics Check	OK	23:23	
Air Blank		23:24	RFI Detect
Air Blank		23:24	RFI Detect

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/20/2024 23:31:48

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.125	23:27
Breath Volume	2.949	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.126	23:30
Breath Volume	3.367	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/21/2024 00:58:50

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1 0.167	00:56	
Breath Volume	2.894	00:56	
Air Blank	0.000	00:56	
Air Blank		00:58	RFI Detect
Air Blank		00:58	RFI Detect
Air Blank	0.000	00:58	

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCANDLESS,MATTHEW,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCANDLESS,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/21/2024 01:06:55

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.081	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.166	01:02
Breath Volume	2.105	01:02
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.163	01:05
Breath Volume	2.015	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:06
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 12/27/2024 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MCCULLOUGH,JAMES
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ROZWADOWSKI

DATE/TIME OF TEST: 12/01/2024 01:11:01

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.012	01:06
Breath Volume	1.480	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.013	01:09
Breath Volume	1.648	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:10
Diagnostics Check	OK	01:11

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MCCULLOUGH,JAMES
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 12/01/2024 04:16:52

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.239	04:12
Breath Volume	2.117	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.239	04:15
Breath Volume	3.289	04:15
Air Blank	0.000	04:15
Control Test	0.078	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 12/03/2024 00:55:33

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.078	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.050	00:54	
Breath Volume	3.433	00:54	
Air Blank		00:54	RFI Detect
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HO

DATE/TIME OF TEST: 12/03/2024 01:03:23

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.078	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.048	00:58
Breath Volume	3.472	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.046	01:01
Breath Volume	3.375	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORDLE,ABIGAIL,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORDLE,ABIGAIL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: ZEIDAN

DATE/TIME OF TEST: 12/03/2024 12:55:27

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.079	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.000	12:50
Breath Volume	1.371	12:50
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.000	12:54
Breath Volume	1.625	12:54
Air Blank	0.000	12:54
Control Test	0.078	12:54
Air Blank	0.000	12:55
Diagnostics Check	OK	12:55

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMANIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: ALIMI

DATE/TIME OF TEST: 12/07/2024 02:11:36

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.078	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1 0.082	01:58	
Breath Volume	2.027	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2 0.085	02:04	Volume Not Met
Breath Volume	1.128	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 3 0.083	02:10	Volume Not Met
Breath Volume	0.050	02:10	
Air Blank	0.000	02:10	
Control Test	0.078	02:10	
Air Blank	0.000	02:11	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMANIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: ALIMI

DATE/TIME OF TEST: 12/07/2024 02:19:16

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.081	02:14
Breath Volume	2.285	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.078	02:17
Breath Volume	2.191	02:17
Air Blank	0.000	02:18
Control Test	0.077	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: COLIN,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 12/07/2024 21:35:50

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.078	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.181	21:30
Breath Volume	1.710	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.175	21:34
Breath Volume	1.277	21:34
Air Blank	0.000	21:34
Control Test	0.077	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/08/2024 03:08:07

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.087	03:03
Breath Volume	1.242	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.083	03:06
Breath Volume	1.285	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JERRY,F
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CARROLL,JERRY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: PEA

DATE/TIME OF TEST: 12/09/2024 00:24:08

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.000	00:19
Breath Volume	2.628	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.000	00:22
Breath Volume	3.109	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: KIRKSEY

DATE/TIME OF TEST: 12/10/2024 03:28:28

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.158	03:23
Breath Volume	1.132	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.157	03:26
Breath Volume	1.925	03:26
Air Blank	0.000	03:27
Control Test	0.077	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK',G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: BADOLATO

DATE/TIME OF TEST: 12/11/2024 22:07:03

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.076	22:01	
Air Blank	0.000	22:01	
Subject Sample # 1	0.265	22:02	
Breath Volume	2.222	22:02	
Air Blank	0.000	22:02	
Air Blank	0.000	22:04	
Subject Sample # 2	0.273	22:05	
Breath Volume	1.863	22:05	
Air Blank	0.000	22:06	
Control Test	0.073	22:06	Control Outside Tolerance
Air Blank	0.000	22:07	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LYNTH

DATE/TIME OF TEST: 12/11/2024 22:16:15

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.075	22:10
Air Blank	0.000	22:10
Subject Sample # 1	0.155	22:11
Breath Volume	1.277	22:11
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.160	22:14
Breath Volume	2.039	22:14
Air Blank	0.000	22:15
Control Test	0.075	22:15
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALIJAGIC,SAMIR,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ALIJAGIC,SAMIR
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 12/12/2024 23:38:13

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.078	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.200	23:33
Breath Volume	2.910	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.195	23:36
Breath Volume	3.539	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AUCH,DOMANIK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: XIUHTECUTLI

DATE/TIME OF TEST: 12/13/2024 22:13:21

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.052	22:08
Breath Volume	1.753	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.048	22:11
Breath Volume	1.121	22:11
Air Blank	0.000	22:12
Control Test	0.078	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: KEYS

DATE/TIME OF TEST: 12/14/2024 03:09:28

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:58	
Control Test	0.078	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:02	No Sample Provided
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2	03:08	No Sample Provided
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Control Test	0.078	03:08	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MACKEY

DATE/TIME OF TEST: 12/14/2024 03:33:51

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.078	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.147	03:29
Breath Volume	3.464	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.148	03:32
Breath Volume	3.414	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JERRY,F
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CARROLL,JERRY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: PAASCHE

DATE/TIME OF TEST: 12/16/2024 18:30:01

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.077	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.242	18:25
Breath Volume	2.531	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.243	18:28
Breath Volume	2.582	18:28
Air Blank	0.000	18:29
Control Test	0.075	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:30

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CLELAND,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 12/18/2024 04:16:08

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.111	04:05	Volume Not Met
Breath Volume	0.058	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 2 0.103	04:10	
Breath Volume	2.320	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 3 0.116	04:14	
Breath Volume	2.664	04:14	
Air Blank	0.000	04:15	
Control Test	0.077	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/19/2024 01:39:05

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.078	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.165	01:34
Breath Volume	1.457	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.160	01:37
Breath Volume	1.921	01:37
Air Blank	0.000	01:38
Control Test	0.076	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 12/20/2024 02:34:38

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.077	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.130	02:30
Breath Volume	1.269	02:30
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.125	02:33
Breath Volume	2.746	02:33
Air Blank	0.000	02:33
Control Test	0.078	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SAAD

DATE/TIME OF TEST: 12/21/2024 02:18:00

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.185	02:13
Breath Volume	1.316	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.175	02:16
Breath Volume	1.175	02:16
Air Blank	0.000	02:17
Control Test	0.076	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:18

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: REECE,STEPHEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: VANAIIRSDALE

DATE/TIME OF TEST: 12/24/2024 20:35:15

Diagnostics Check	OK	20:28
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Subject Sample # 1	0.232	20:30
Breath Volume	1.792	20:30
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.226	20:33
Breath Volume	1.859	20:33
Air Blank	0.000	20:34
Control Test	0.078	20:34
Air Blank	0.000	20:35
Diagnostics Check	OK	20:35

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/24/2024 20:44:24

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.078	20:38	
Air Blank	0.000	20:38	
Subject Sample	# 1 0.201	20:39	
Breath Volume	2.070	20:39	
Air Blank	0.000	20:40	
Air Blank	0.000	20:42	
Subject Sample	# 2 0.194	20:42	
Breath Volume	1.898	20:42	
Air Blank	0.000	20:43	
Control Test	0.072	20:43	Control Outside Tolerance
Air Blank	0.000	20:44	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/24/2024 20:52:41

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.077	20:46	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.207	20:47	
Breath Volume	2.195	20:47	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 2 0.196	20:51	
Breath Volume	2.406	20:51	
Air Blank	0.000	20:51	
Control Test	0.074	20:52	Control Outside Tolerance
Air Blank	0.000	20:52	

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: CHAPPELL

DATE/TIME OF TEST: 12/26/2024 01:47:53

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.077	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.167	01:43
Breath Volume	2.101	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.156	01:46
Breath Volume	2.214	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 12/27/2024 23:53:17

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.078	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.328	23:48
Breath Volume	1.679	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.315	23:51
Breath Volume	1.730	23:51
Air Blank	0.000	23:52
Control Test	0.077	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: KALYANASUNDARAM

DATE/TIME OF TEST: 12/29/2024 00:58:17

Diagnostics Check	OK	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.099	00:53
Breath Volume	2.027	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.091	00:56
Breath Volume	1.125	00:56
Air Blank	0.000	00:57
Control Test	0.077	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UTLEY,BRETT,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KING,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 09:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 12/29/2024 10:01:17

Diagnostics Check	OK	09:55
Air Blank	0.000	09:55
Control Test	0.078	09:56
Air Blank	0.000	09:56
Subject Sample # 1	0.000	09:57
Breath Volume	3.464	09:57
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.000	09:59
Breath Volume	2.468	09:59
Air Blank	0.000	10:00
Control Test	0.077	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: NOUMAN

DATE/TIME OF TEST: 12/30/2024 01:08:49

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.204	01:04
Breath Volume	2.769	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.203	01:07
Breath Volume	3.351	01:07
Air Blank	0.000	01:07
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SMITH,TRISTAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MUYRES

DATE/TIME OF TEST: 12/30/2024 03:33:09

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.241	03:28
Breath Volume	3.636	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.249	03:31
Breath Volume	2.636	03:31
Air Blank	0.000	03:32
Control Test	0.077	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/30/2024 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERICKSON,NATHAN,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LUTTRELL,EMMETT
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: DONATO

DATE/TIME OF TEST: 12/06/2024 20:58:25

Diagnostics Check	OK	20:51
Air Blank	0.000	20:52
Control Test	0.078	20:52
Air Blank	0.000	20:53
Subject Sample # 1	0.171	20:53
Breath Volume	1.933	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.186	20:56
Breath Volume	3.140	20:56
Air Blank	0.000	20:57
Control Test	0.077	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 12/13/2024 08:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,J,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MURPHY,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/07/2024 20:46:30

Diagnostics Check	OK	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.274	20:39
Breath Volume	1.207	20:39
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.256	20:44
Breath Volume	1.320	20:44
Air Blank	0.000	20:45
Control Test	0.076	20:45
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 12/13/2024 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CORD

DATE/TIME OF TEST: 12/09/2024 02:02:46

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.162	01:58
Breath Volume	2.039	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.159	02:01
Breath Volume	2.343	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/13/2024 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: DOWDY

DATE/TIME OF TEST: 12/10/2024 00:37:44

Diagnostics Check	OK	00:31
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:32
Subject Sample # 1	0.000	00:33
Breath Volume	3.488	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.000	00:36
Breath Volume	2.972	00:36
Air Blank	0.000	00:36
Control Test	0.081	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 12/13/2024 08:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REXFORD,MARK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: REXFORD,MARK
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: INMAN

DATE/TIME OF TEST: 12/15/2024 03:16:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.197	03:12
Breath Volume	2.183	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:14
Subject Sample # 2	0.202	03:15
Breath Volume	1.988	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERICKSON,NATHAN,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BARRETT,JOSHUA
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: SZIRANKO

DATE/TIME OF TEST: 12/17/2024 01:57:26

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.155	01:52
Breath Volume	3.003	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.163	01:55
Breath Volume	2.347	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERICKSON,NATHAN,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MOORHEAD,JAKE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MARCEY

DATE/TIME OF TEST: 12/21/2024 02:46:28

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.244	02:41
Breath Volume	2.890	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.244	02:44
Breath Volume	2.355	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: CUCCO

DATE/TIME OF TEST: 12/21/2024 07:01:48

Diagnostics Check	OK	06:55
Air Blank	0.000	06:56
Control Test	0.080	06:56
Air Blank	0.000	06:57
Subject Sample # 1	0.000	06:57
Breath Volume	3.062	06:57
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.000	07:00
Breath Volume	2.292	07:00
Air Blank	0.000	07:00
Control Test	0.080	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINSON,RYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WILKINSON,RYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RECKART

DATE/TIME OF TEST: 12/22/2024 03:11:40

Diagnostics Check	OK	03:07	
Air Blank	0.000	03:08	
Control Test	0.080	03:08	
Air Blank	0.000	03:08	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:10	
Control Test	0.080	03:11	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,DARIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARPER,DARIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JR OLIVIO

DATE/TIME OF TEST: 12/25/2024 02:06:07

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.112	01:59
Breath Volume	1.593	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.109	02:04
Breath Volume	2.429	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 12/25/2024 22:01:50

Diagnostics Check	OK	21:58	
Air Blank	0.000	21:59	
Control Test	0.079	21:59	
Air Blank	0.000	22:00	
Subject Sample	# 1	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: PAVLICHEK

DATE/TIME OF TEST: 12/26/2024 11:37:48

Diagnostics Check	OK	11:30
Air Blank	0.000	11:31
Control Test	0.080	11:31
Air Blank	0.000	11:32
Subject Sample # 1	0.135	11:32
Breath Volume	1.183	11:32
Air Blank	0.000	11:33
Air Blank	0.000	11:35
Subject Sample # 2	0.123	11:36
Breath Volume	1.375	11:36
Air Blank	0.000	11:36
Control Test	0.079	11:37
Air Blank	0.000	11:37
Diagnostics Check	OK	11:37

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON, TYLER, S
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TORRES, BRENDON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/13/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GUZMAN-BEYDOYA

DATE/TIME OF TEST: 12/28/2024 03:20:33

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.166	03:15
Breath Volume	1.316	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.160	03:19
Breath Volume	1.535	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 01/14/2025 08:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OMORUYI,TYREE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 12/01/2024 02:39:00

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.149	02:34
Breath Volume	1.613	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.145	02:37
Breath Volume	1.226	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/01/2024 03:04:36

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.264	02:54
Breath Volume	2.054	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.236	02:59
Breath Volume	1.593	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:01
Subject Sample # 3	0.256	03:03
Breath Volume	2.750	03:03
Air Blank	0.000	03:03
Control Test	0.080	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HERNANDEZ SANCHEZ

DATE/TIME OF TEST: 12/01/2024 05:10:50

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.081	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.256	05:06
Breath Volume	1.812	05:06
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.268	05:09
Breath Volume	2.527	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 06:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: TELESKA

DATE/TIME OF TEST: 12/07/2024 08:23:00

Diagnostics Check	OK	08:14
Air Blank	0.000	08:15
Control Test	0.082	08:15
Air Blank	0.000	08:16
Subject Sample # 1	0.131	08:18
Breath Volume	1.636	08:18
Air Blank	0.000	08:18
Air Blank	0.000	08:20
Subject Sample # 2	0.138	08:21
Breath Volume	2.035	08:21
Air Blank	0.000	08:22
Control Test	0.081	08:22
Air Blank	0.000	08:22
Diagnostics Check	OK	08:23

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CORVIL

DATE/TIME OF TEST: 12/07/2024 23:27:37

Diagnostics Check	OK	23:20
Air Blank	0.000	23:21
Control Test	0.081	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.269	23:23
Breath Volume	2.476	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.274	23:26
Breath Volume	1.425	23:26
Air Blank	0.000	23:26
Control Test	0.080	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANTONINI,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RUEDA PENA

DATE/TIME OF TEST: 12/08/2024 00:34:02

Diagnostics Check	OK	00:27
Air Blank	0.000	00:28
Control Test	0.081	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.254	00:29
Breath Volume	2.464	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.248	00:32
Breath Volume	2.578	00:32
Air Blank	0.000	00:33
Control Test	0.081	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:34

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IDARRAGA,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: ALVAREZ RODRIGUEZ

DATE/TIME OF TEST: 12/08/2024 16:30:08

Diagnostics Check	OK	16:23
Air Blank	0.000	16:24
Control Test	0.081	16:24
Air Blank	0.000	16:25
Subject Sample # 1	0.241	16:25
Breath Volume	3.328	16:25
Air Blank	0.000	16:26
Air Blank	0.000	16:28
Subject Sample # 2	0.240	16:28
Breath Volume	3.292	16:28
Air Blank	0.000	16:29
Control Test	0.081	16:29
Air Blank	0.000	16:30
Diagnostics Check	OK	16:30

Data Download Date/Time: 12/11/2024 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEONE,DANIEL
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 11/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 16:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: COMISKEY JR.

DATE/TIME OF TEST: 12/08/2024 17:59:16

Diagnostics Check	OK	17:52
Air Blank	0.000	17:53
Control Test	0.081	17:53
Air Blank	0.000	17:54
Subject Sample # 1	0.119	17:54
Breath Volume	1.496	17:54
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.123	17:57
Breath Volume	1.890	17:57
Air Blank	0.000	17:58
Control Test	0.080	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 12/11/2024 09:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REVOREDO,JULIO
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MONCRIEFFE

DATE/TIME OF TEST: 12/13/2024 23:46:31

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.081	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:45	RFI Detect
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 01/08/2025 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: REVOREDO,JULIO
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MONCRIEFFE

DATE/TIME OF TEST: 12/13/2024 23:54:33

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.000	23:50
Breath Volume	2.199	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.000	23:53
Breath Volume	2.503	23:53
Air Blank	0.000	23:53
Control Test	0.080	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 01/08/2025 10:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRAZANA,DANIEL,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: MOWBRAY,MARK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 12/08/2024 23:44:58

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.080	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.167	23:40
Breath Volume	3.023	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.172	23:43
Breath Volume	2.832	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/31/2024 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRAZANA,DANIEL,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: DINGUS,DAVID
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BECKWITH

DATE/TIME OF TEST: 12/17/2024 02:01:52

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.140	01:56
Breath Volume	1.726	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.144	02:00
Breath Volume	1.535	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 12/31/2024 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRAZANA,DANIEL,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: LEWIS,JORDAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: MCCONKEY

DATE/TIME OF TEST: 12/18/2024 23:57:55

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.104	23:53
Breath Volume	1.585	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.112	23:56
Breath Volume	3.281	23:56
Air Blank	0.000	23:56
Control Test	0.078	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 12/31/2024 07:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRAZANA,DANIEL,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: MUNIZ,JONATHAN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 11/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A4
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MILLIGAN

DATE/TIME OF TEST: 12/19/2024 04:08:54

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.027	04:00	Volume Not Met
Breath Volume	0.085	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.023	04:04	
Breath Volume	2.082	04:04	
Air Blank	0.000	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 3 0.024	04:07	
Breath Volume	2.281	04:07	
Air Blank	0.000	04:07	
Control Test	0.078	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 12/31/2024 07:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/04/2024 05:44:44

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.083	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.152	05:40
Breath Volume	1.871	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.157	05:43
Breath Volume	1.632	05:43
Air Blank	0.000	05:43
Control Test	0.082	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: AWI

DATE/TIME OF TEST: 12/06/2024 00:57:43

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.081	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1 0.239	00:50	Volume Not Met
Breath Volume	0.507	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:56	Subject Test Refused
Breath Volume	0.117	00:56	
Air Blank	0.000	00:56	
Control Test	0.077	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILLS,JOHN,T
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MCCABE,PATRICK
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: FEATHERSTON

DATE/TIME OF TEST: 12/08/2024 03:57:34

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.083	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.237	03:52
Breath Volume	3.496	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.220	03:55
Breath Volume	3.281	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILLS,JOHN,T
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KING,CASEY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/09/2024 04:11:09

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.082	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.033	04:06
Breath Volume	2.269	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:08
Subject Sample # 2	0.028	04:09
Breath Volume	1.792	04:09
Air Blank	0.000	04:10
Control Test	0.082	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS, WILLIAM, D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS, WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CRISWELL SHIPLEY

DATE/TIME OF TEST: 12/13/2024 04:13:02

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.083	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.092	04:08
Breath Volume	1.761	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.088	04:11
Breath Volume	3.121	04:11
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CLARK,AARON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: VAUSE

DATE/TIME OF TEST: 12/14/2024 00:58:42

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.082	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.163	00:54
Breath Volume	2.496	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.172	00:57
Breath Volume	3.746	00:57
Air Blank	0.000	00:57
Control Test	0.081	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/16/2024 05:35:24

Diagnostics Check	OK	05:34	
Air Blank		05:34	RFI Detect
Air Blank	0.000	05:35	

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/16/2024 05:46:47

Diagnostics Check	OK	05:37
Air Blank	0.000	05:37
Control Test	0.082	05:37
Air Blank	0.000	05:38
Subject Sample # 1	0.090	05:39
Breath Volume	1.136	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.083	05:43
Breath Volume	1.121	05:43
Air Blank	0.000	05:44
Control Test	0.080	05:44
Air Blank	0.000	05:44
Diagnostics Check	OK	05:46

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 11/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MATIS

DATE/TIME OF TEST: 12/16/2024 06:02:10

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:52	
Control Test	0.081	05:52	
Air Blank	0.000	05:53	
Subject Sample	# 1	05:56	No Sample Provided
Breath Volume	0.000	05:56	
Air Blank	0.000	05:56	
Air Blank	0.000	05:58	
Subject Sample	# 2	06:00	Subject Test Refused
Breath Volume	0.113	06:00	
Air Blank	0.000	06:01	
Control Test	0.082	06:01	
Air Blank	0.000	06:02	
Diagnostics Check	OK	06:02	

Data Download Date/Time: 12/18/2024 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: TRUMAN,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/21/2024 23:12:44

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.083	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.196	23:08
Breath Volume	2.863	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.191	23:11
Breath Volume	1.949	23:11
Air Blank	0.000	23:11
Control Test	0.082	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 01/15/2025 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001153

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: JOYNER, J
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ROZELLE

DATE/TIME OF TEST: 12/22/2024 22:36:14

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.081	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.201	22:31
Breath Volume	2.144	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.201	22:34
Breath Volume	2.808	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 01/15/2025 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: DEES,BENJAMIN
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/20/2024 07:49:06

Diagnostics Check	OK	07:43
Air Blank	0.000	07:43
Control Test	0.081	07:43
Air Blank	0.000	07:44
Subject Sample # 1	0.000	07:44
Breath Volume	3.628	07:44
Air Blank	0.000	07:45
Air Blank	0.000	07:47
Subject Sample # 2	0.000	07:47
Breath Volume	3.718	07:47
Air Blank	0.000	07:48
Control Test	0.080	07:48
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 01/17/2025 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MCCORMICK,JOSEPH
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: BELLES

DATE/TIME OF TEST: 12/23/2024 10:29:05

Diagnostics Check	OK	10:28	
Air Blank		10:28	RFI Detect
Air Blank		10:29	RFI Detect

Data Download Date/Time: 01/17/2025 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEES,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: MCCORMICK,JOSEPH
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 11/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: ADMIN
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: BELLES

DATE/TIME OF TEST: 12/23/2024 10:32:55

Diagnostics Check	OK	10:31	
Air Blank	0.000	10:31	
Control Test	0.080	10:31	
Air Blank		10:32	RFI Detect
Air Blank	0.000	10:32	

Data Download Date/Time: 01/17/2025 20:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEELEY,JAMES,E
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: FEELEY,JAMES
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: PINGITORE

DATE/TIME OF TEST: 12/20/2024 04:15:25

Diagnostics Check	OK	04:09
Air Blank	0.000	04:09
Control Test	0.081	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.165	04:11
Breath Volume	2.386	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.163	04:13
Breath Volume	1.960	04:13
Air Blank	0.000	04:14
Control Test	0.081	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 01/23/2025 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGNER, THOMAS, W
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MORRIS, JOSHUA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: LAMM

DATE/TIME OF TEST: 12/27/2024 17:36:44

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.081	17:31	
Air Blank	0.000	17:32	
Subject Sample # 1	0.105	17:33	
Breath Volume	2.281	17:33	
Air Blank	0.000	17:33	
Air Blank	0.000	17:35	
Subject Sample # 2	0.104	17:36	
Breath Volume	2.277	17:36	
Air Blank		17:36	RFI Detect
Air Blank	0.000	17:36	

Data Download Date/Time: 01/23/2025 17:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAGNER, THOMAS, W
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MORRIS, JOSHUA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: LAMM

DATE/TIME OF TEST: 12/27/2024 17:52:34

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.081	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.105	17:48
Breath Volume	2.445	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.099	17:51
Breath Volume	2.253	17:51
Air Blank	0.000	17:51
Control Test	0.080	17:51
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 01/23/2025 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: BERNAL,ANTHONY
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402477283
DRY GAS EXPIRY DATE: 06/24/2025
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/31/2024 21:25:58

Diagnostics Check	OK	21:19
Air Blank	0.000	21:19
Control Test	0.082	21:19
Air Blank	0.000	21:20
Subject Sample # 1	0.194	21:21
Breath Volume	1.687	21:21
Air Blank	0.000	21:22
Air Blank	0.000	21:23
Subject Sample # 2	0.203	21:24
Breath Volume	1.246	21:24
Air Blank	0.000	21:25
Control Test	0.081	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 01/23/2025 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,J,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LATIMER,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/01/2024 21:50:19

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.082	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.082	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LATIMER,J,A

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: HERNANDEZ,D

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402357111

DRY GAS EXPIRY DATE: 02/10/2025

VIOLATION CODE: DUI

ARREST TIME: 02:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 12/05/2024 03:32:45

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:30	
Control Test	0.082	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:31	Subject Test Refused
Breath Volume	0.000	03:31	
Air Blank	0.000	03:31	
Control Test	0.082	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATIMER,J,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LATIMER,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CHIRINO

DATE/TIME OF TEST: 12/21/2024 01:45:44

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.179	01:41
Breath Volume	1.960	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.177	01:44
Breath Volume	2.121	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: RAGHUNANDAN,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 12/21/2024 05:13:23

Diagnostics Check	OK	05:10	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:11	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.082	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ZOLLWEG,B
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 12/22/2024 02:22:56

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.082	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.183	02:12
Breath Volume	1.414	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.156	02:18
Breath Volume	2.039	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 3	0.168	02:21
Breath Volume	2.050	02:21
Air Blank	0.000	02:22
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: MENCARELLI,D
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/22/2024 03:04:57

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.151	03:00
Breath Volume	1.792	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.146	03:03
Breath Volume	1.769	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 12/23/2024 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: ZOLLWEG,B
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 12/23/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402357111
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/27/2024 00:18:17

Diagnostics Check	OK	00:17	
Air Blank		00:18	RFI Detect
Air Blank		00:18	RFI Detect

Data Download Date/Time: 01/22/2025 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LUNA,R
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: RODRIGUEZ-GARCIA

DATE/TIME OF TEST: 12/10/2024 23:55:53

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.168	23:51
Breath Volume	2.976	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.168	23:54
Breath Volume	2.871	23:54
Air Blank	0.000	23:54
Control Test	0.081	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 01/21/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LUNA,R
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/24/2024 00:48:13

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:45	
Control Test	0.081	00:45	
Air Blank	0.000	00:46	
Subject Sample	# 1	00:46	Subject Test Refused
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Control Test	0.081	00:47	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 01/21/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 12/25/2024 03:18:20

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.134	03:13
Breath Volume	2.593	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.128	03:16
Breath Volume	2.496	03:16
Air Blank	0.000	03:17
Control Test	0.081	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 01/21/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: JOHNSON,D
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01923080A3
DRY GAS EXPIRY DATE: 02/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 12/30/2024 23:51:10

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:48	
Control Test	0.081	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 1	23:49	Subject Test Refused
Breath Volume	0.000	23:49	
Air Blank	0.000	23:50	
Control Test	0.081	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 01/21/2025 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/08/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A5
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: LUTZ

DATE/TIME OF TEST: 12/08/2024 20:55:48

Diagnostics Check	OK	20:49
Air Blank	0.000	20:50
Control Test	0.081	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.291	20:51
Breath Volume	1.519	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:53
Subject Sample # 2	0.287	20:54
Breath Volume	1.167	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 01/27/2025 17:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: DORFMAN

DATE/TIME OF TEST: 12/07/2024 01:08:39

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.082	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.200	01:04
Breath Volume	1.878	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.192	01:07
Breath Volume	1.152	01:07
Air Blank	0.000	01:07
Control Test	0.082	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 12/18/2024 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: DORSAINVILLE

DATE/TIME OF TEST: 12/12/2024 20:31:20

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:28	
Control Test	0.081	20:28	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Control Test	0.082	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 12/18/2024 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/13/2024 22:16:56

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.083	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.356	22:12
Breath Volume	3.070	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.347	22:15
Breath Volume	1.785	22:15
Air Blank	0.000	22:16
Control Test	0.081	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 12/18/2024 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 12/15/2024 00:58:01

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.082	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.275	00:53
Breath Volume	1.449	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.277	00:56
Breath Volume	1.738	00:56
Air Blank	0.000	00:57
Control Test	0.081	00:57
Air Blank	0.000	00:57
Diagnostics Check	OK	00:58

Data Download Date/Time: 12/18/2024 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: MAROTO,VALERY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/22/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: VARGAS GARCIA

DATE/TIME OF TEST: 12/15/2024 01:17:54

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.132	01:12
Breath Volume	2.160	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.127	01:16
Breath Volume	1.984	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 12/18/2024 20:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALEZI

DATE/TIME OF TEST: 12/20/2024 03:14:37

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.082	03:08
Air Blank	0.000	03:08
Subject Sample # 1	0.237	03:09
Breath Volume	2.320	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.237	03:13
Breath Volume	2.632	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 01/17/2025 22:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 12/18/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 220213
DRY GAS EXPIRY DATE: 02/10/2025
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 12/22/2024 01:36:47

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.083	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.153	01:32
Breath Volume	1.410	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.157	01:35
Breath Volume	1.437	01:35
Air Blank	0.000	01:35
Control Test	0.082	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 01/17/2025 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/02/2024 17:21:41

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.081	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.000	17:17
Breath Volume	2.406	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.000	17:20
Breath Volume	1.941	17:20
Air Blank	0.000	17:20
Control Test	0.081	17:21
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

Data Download Date/Time: 12/19/2024 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DARYL,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFERY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/04/2024 04:13:54

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.137	04:09
Breath Volume	2.816	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.137	04:12
Breath Volume	2.628	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 12/19/2024 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWIFT,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: LERNER,MATTHEW
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/06/2024 00:32:41

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.159	00:27
Breath Volume	1.578	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.177	00:31
Breath Volume	1.464	00:31
Air Blank	0.000	00:31
Control Test	0.079	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 12/19/2024 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELCHER,ANDREW,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BENNETT,BRANDY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: AUBERT

DATE/TIME OF TEST: 12/08/2024 03:36:02

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.010	03:31
Breath Volume	3.726	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.009	03:34
Breath Volume	3.578	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:36

Data Download Date/Time: 12/19/2024 22:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: SHOWALTER

DATE/TIME OF TEST: 12/21/2024 01:49:21

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1 0.200	01:42	Volume Not Met
Breath Volume	0.031	01:42	
Air Blank	0.000	01:42	
Air Blank	0.000	01:44	
Subject Sample	# 2 0.225	01:47	Volume Not Met
Breath Volume	0.136	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 01/07/2025 19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIGHT,BRANDON,T
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: KIGHT,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: MCCORD JR.

DATE/TIME OF TEST: 12/25/2024 15:55:39

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.080	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:51
Breath Volume	3.617	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:54
Breath Volume	3.917	15:54
Air Blank	0.000	15:54
Control Test	0.080	15:54
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 01/07/2025 19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,DARRELL,C
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: NORRIS,DARRELL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403005538
DRY GAS EXPIRY DATE: 03/25/2027
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 12/31/2024 22:58:08

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.139	22:53
Breath Volume	2.281	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.137	22:56
Breath Volume	2.320	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/07/2025 19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUTH,DAMIAN,W
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LUCAS,KOBI'
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: TREJO-ALVAREZ

DATE/TIME OF TEST: 12/06/2024 21:50:31

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.080	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.121	21:46
Breath Volume	2.878	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.120	21:49
Breath Volume	2.214	21:49
Air Blank	0.000	21:49
Control Test	0.080	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 12/09/2024 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAY,JON,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: GAY,JON
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 12/07/2024 02:30:56

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:15	
Control Test	0.080	02:15	
Air Blank	0.000	02:16	
Subject Sample	# 1 0.227	02:19	No .020 agreement
Breath Volume	1.218	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 2 0.248	02:23	No .020 agreement
Breath Volume	1.472	02:23	
Air Blank	0.000	02:24	
Air Blank	0.000	02:26	
Subject Sample	# 3 0.211	02:29	Volume Not Met
Breath Volume	0.570	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 12/09/2024 02:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINGRY,HOBIE,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: KINGRY,HOBIE
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: AYALA HERNANDEZ

DATE/TIME OF TEST: 12/07/2024 02:54:26

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.218	02:49
Breath Volume	2.457	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.224	02:52
Breath Volume	2.933	02:52
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 12/09/2024 02:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUTH,DAMIAN,W
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FOSTER,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 12/08/2024 01:43:39

Diagnostics Check	OK	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.159	01:36
Breath Volume	1.285	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.156	01:42
Breath Volume	1.867	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 12/09/2024 02:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY, ELISE, N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/02/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 07:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: LOPEZ JIMENEZ

DATE/TIME OF TEST: 12/08/2024 08:00:21

Diagnostics Check	OK	07:52
Air Blank	0.000	07:53
Control Test	0.080	07:53
Air Blank	0.000	07:53
Subject Sample # 1	0.126	07:55
Breath Volume	1.996	07:55
Air Blank	0.000	07:56
Air Blank	0.000	07:58
Subject Sample # 2	0.128	07:58
Breath Volume	2.750	07:58
Air Blank	0.000	07:59
Control Test	0.080	07:59
Air Blank	0.000	08:00
Diagnostics Check	OK	08:00

Data Download Date/Time: 12/09/2024 02:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SCOTT,DYLAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 12/10/2024 02:11:37

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.172	02:07
Breath Volume	1.886	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.159	02:10
Breath Volume	1.691	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SCOTT,DYLAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 12/11/2024 03:05:25

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.183	03:00
Breath Volume	3.683	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.175	03:03
Breath Volume	2.933	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER,DILLON,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: LEE,NICHOLAS
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WARFIELD

DATE/TIME OF TEST: 12/13/2024 01:28:13

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.189	01:23
Breath Volume	2.593	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.201	01:26
Breath Volume	2.300	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER,DILLON,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BUNCH,BRADFORD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/13/2024 23:27:44

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.182	23:22
Breath Volume	1.457	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.184	23:26
Breath Volume	2.222	23:26
Air Blank	0.000	23:26
Control Test	0.078	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FORTUNE,CALVIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BOND

DATE/TIME OF TEST: 12/14/2024 02:39:44

Diagnostics Check	OK	02:30	
Air Blank	0.000	02:30	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.139	02:34	Volume Not Met
Breath Volume	0.894	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2	02:39	Sequence Aborted
Breath Volume	0.000	02:39	
Air Blank	0.000	02:39	

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: HOGAN,JEREMY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: INFINGER

DATE/TIME OF TEST: 12/14/2024 03:38:41

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.090	03:34
Breath Volume	1.792	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.092	03:37
Breath Volume	2.304	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/16/2024 00:22:41

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	1.539	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	1.789	00:21
Air Blank	0.000	00:21
Control Test	0.080	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER, TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/17/2024 23:18:51

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.145	23:14
Breath Volume	2.019	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.139	23:17
Breath Volume	1.847	23:17
Air Blank	0.000	23:17
Control Test	0.080	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SCOTT,DYLAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: TRIMBLE

DATE/TIME OF TEST: 12/18/2024 22:27:04

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.078	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.248	22:18
Breath Volume	2.136	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.284	22:22
Breath Volume	2.816	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 3	0.243	22:25
Breath Volume	1.531	22:25
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:27

Data Download Date/Time: 01/11/2025 22:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CADY, ELISE, N
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ZAYAGO BONILLA

DATE/TIME OF TEST: 12/19/2024 21:34:37

Diagnostics Check	OK	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:29
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:30
Breath Volume	2.800	21:30
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:33
Breath Volume	2.382	21:33
Air Blank	0.000	21:33
Control Test	0.080	21:34
Air Blank	0.000	21:34
Diagnostics Check	OK	21:34

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINN,CAMERON,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FORTUNE,CALVIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/20/2024 03:57:34

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:55	
Control Test	0.080	03:55	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Diagnostics Check	OK	03:57	

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINN,CAMERON,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: FORTUNE,CALVIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/20/2024 04:15:24

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.155	04:10
Breath Volume	1.367	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.145	04:13
Breath Volume	1.347	04:13
Air Blank	0.000	04:14
Control Test	0.077	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BUCKALEW,TYLER
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: THURMAN JR

DATE/TIME OF TEST: 12/24/2024 01:27:16

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.204	01:22
Breath Volume	2.320	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:25
Subject Sample # 2	0.207	01:25
Breath Volume	2.527	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF, RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MALOY

DATE/TIME OF TEST: 12/26/2024 19:32:25

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.079	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.189	19:27
Breath Volume	3.000	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.190	19:30
Breath Volume	3.785	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,GARRETT,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 12/27/2024 23:59:22

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.079	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.000	23:55
Breath Volume	1.847	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	1.226	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:58
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY, TYLER, J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: PHILLIPS, KHALIL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 12/28/2024 04:59:29

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.079	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.045	04:55
Breath Volume	1.730	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.042	04:58
Breath Volume	1.566	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: STALNAKER, TEDDY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 12/28/2024 18:37:03

Diagnostics Check	OK	18:30
Air Blank	0.000	18:31
Control Test	0.081	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.213	18:32
Breath Volume	2.273	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.219	18:35
Breath Volume	2.734	18:35
Air Blank	0.000	18:36
Control Test	0.079	18:36
Air Blank	0.000	18:36
Diagnostics Check	OK	18:37

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DANIEL,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: JONES,DANIEL
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: VONSTEIN

DATE/TIME OF TEST: 12/29/2024 19:17:05

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Subject Sample # 1	0.116	19:12
Breath Volume	3.250	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:14
Subject Sample # 2	0.109	19:15
Breath Volume	2.164	19:15
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLYN,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MIDYETTE,COREY
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/09/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024030567
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LESNER

DATE/TIME OF TEST: 12/30/2024 22:09:59

Diagnostics Check	OK	22:07	
Air Blank	0.000	22:07	
Control Test	0.080	22:07	
Air Blank	0.000	22:08	
Subject Sample	# 1	22:08	Subject Test Refused
Breath Volume	0.000	22:08	
Air Blank	0.000	22:09	
Control Test	0.079	22:09	
Air Blank	0.000	22:09	
Diagnostics Check	OK	22:09	

Data Download Date/Time: 01/11/2025 22:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,WILLIAM,W
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: MILLER,WILLIAM
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CANO

DATE/TIME OF TEST: 12/07/2024 02:55:33

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.081	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.215	02:51
Breath Volume	1.304	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.213	02:53
Breath Volume	2.007	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 01/05/2025 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,WILLIAM,W
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: DAY,GARRETT
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 12/03/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06723080A5
DRY GAS EXPIRY DATE: 04/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 12/22/2024 20:20:21

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.081	20:18	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:19	Subject Test Refused
Breath Volume	0.000	20:19	
Air Blank	0.000	20:19	
Control Test	0.081	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 01/05/2025 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH,LOGAN,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: ARNAUTU,CRISTIAN
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ARGUELLO

DATE/TIME OF TEST: 12/06/2024 21:49:18

Diagnostics Check	OK	21:39
Air Blank	0.000	21:40
Control Test	0.081	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.205	21:43
Breath Volume	1.742	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.210	21:47
Breath Volume	2.648	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 01/08/2025 09:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,EDWARD,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: RODRIGUEZ,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: RAYMUNDO MATOM

DATE/TIME OF TEST: 12/20/2024 00:20:19

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.080	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.130	00:15
Breath Volume	2.015	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.129	00:18
Breath Volume	3.320	00:18
Air Blank	0.000	00:19
Control Test	0.079	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 12/31/2024 20:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: BAEZ,A
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: FELIPE

DATE/TIME OF TEST: 12/29/2024 11:57:59

Diagnostics Check	OK	11:50
Air Blank	0.000	11:50
Control Test	0.079	11:51
Air Blank	0.000	11:51
Subject Sample # 1	0.159	11:53
Breath Volume	2.441	11:53
Air Blank	0.000	11:53
Air Blank	0.000	11:55
Subject Sample # 2	0.164	11:56
Breath Volume	1.449	11:56
Air Blank	0.000	11:57
Control Test	0.079	11:57
Air Blank	0.000	11:57
Diagnostics Check	OK	11:57

Data Download Date/Time: 12/31/2024 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,JACOB,A
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: BELL,KENTAVIOUS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 11/29/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: LUPICA

DATE/TIME OF TEST: 12/30/2024 23:21:23

Diagnostics Check	OK	23:09	
Air Blank	0.000	23:09	
Control Test	0.079	23:10	
Air Blank	0.000	23:10	
Subject Sample	# 1 0.207	23:11	No .020 agreement
Breath Volume	2.093	23:11	
Air Blank	0.000	23:11	
Air Blank	0.000	23:13	
Subject Sample	# 2 0.179	23:14	No .020 agreement
Breath Volume	1.261	23:14	
Air Blank	0.000	23:15	
Air Blank	0.000	23:16	
Subject Sample	# 3	23:20	No Sample Provided
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.079	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 12/31/2024 20:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: DIMINI,NICHOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: FAVORS

DATE/TIME OF TEST: 12/05/2024 23:49:16

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.182	23:41
Breath Volume	1.285	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.178	23:47
Breath Volume	1.394	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: SIENA,MICHAEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HALLENBECK

DATE/TIME OF TEST: 12/11/2024 04:02:23

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.103	03:57
Breath Volume	2.222	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.103	04:00
Breath Volume	2.320	04:00
Air Blank	0.000	04:01
Control Test	0.078	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMIN, PEYTON, C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: RAMIREZ, KEVIN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 12/12/2024 22:26:50

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.220	22:22
Breath Volume	1.441	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.205	22:25
Breath Volume	1.660	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,ISAAC,A
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WILLEMS,BENJAMIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/16/2024 02:19:15

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:12
Subject Sample # 1	0.189	02:13
Breath Volume	1.136	02:13
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.197	02:17
Breath Volume	1.949	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOENBROD,JESSICA,
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MCKINNEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MALVOISIN

DATE/TIME OF TEST: 12/22/2024 04:07:32

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.225	04:02
Breath Volume	4.074	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.227	04:05
Breath Volume	3.621	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,ISAAC,A
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: PIERCE,NATHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 12/23/2024 23:11:36

Diagnostics Check	OK	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.132	23:06
Breath Volume	1.910	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.132	23:09
Breath Volume	2.054	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SETH,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: PIERCE,NATHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 12/01/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402477284
DRY GAS EXPIRY DATE: 06/27/2025
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: NICKERSON

DATE/TIME OF TEST: 12/29/2024 22:48:12

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:42
Subject Sample # 1	0.098	22:43
Breath Volume	2.210	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:45
Subject Sample # 2	0.100	22:46
Breath Volume	1.386	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/08/2025 05:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,PARKER,L
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 12/01/2024 23:24:15

Diagnostics Check	OK	23:14	
Air Blank	0.000	23:14	
Control Test	0.078	23:14	
Air Blank	0.000	23:15	
Subject Sample	# 1 0.185	23:17	
Breath Volume	2.406	23:17	
Air Blank	0.000	23:18	
Air Blank	0.000	23:19	
Subject Sample	# 2 0.168	23:22	Volume Not Met
Breath Volume	0.093	23:22	
Air Blank		23:23	Purge Fail
Air Blank	0.000	23:24	

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,PARKER,L
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY,PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ARCHER

DATE/TIME OF TEST: 12/01/2024 23:57:17

Diagnostics Check	OK	23:45	
Air Blank	0.000	23:45	
Control Test	0.076	23:46	
Air Blank	0.000	23:46	
Subject Sample	# 1 0.123	23:49	Volume Not Met
Breath Volume	0.460	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 2 0.158	23:55	Volume Not Met
Breath Volume	0.093	23:55	
Air Blank	0.000	23:56	
Control Test	0.077	23:56	
Air Blank	0.000	23:57	
Diagnostics Check	OK	23:57	

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: EDWARDS, BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 12/09/2024 19:01:06

Diagnostics Check	OK	18:52
Air Blank	0.000	18:52
Control Test	0.078	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.190	18:55
Breath Volume	1.820	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.191	18:59
Breath Volume	1.933	18:59
Air Blank	0.000	19:00
Control Test	0.076	19:00
Air Blank	0.000	19:01
Diagnostics Check	OK	19:01

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY, PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: LIBBY

DATE/TIME OF TEST: 12/14/2024 21:20:23

Diagnostics Check	OK	21:12	
Air Blank	0.000	21:13	
Control Test	0.078	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1 0.145	21:17	Volume Not Met
Breath Volume	0.542	21:17	
Air Blank	0.000	21:17	
Air Blank		21:19	Ambient Fail
Air Blank	0.000	21:20	

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: SWEENEY, PARKER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: LIBBY

DATE/TIME OF TEST: 12/14/2024 21:27:11

Diagnostics Check	OK	21:22	
Air Blank	0.000	21:22	
Control Test	0.077	21:23	
Air Blank	0.000	21:23	
Subject Sample	# 1	21:26	Sequence Aborted
Breath Volume	0.015	21:26	
Air Blank	0.000	21:27	

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: NOLAN,NICOLETTE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: KINNETT

DATE/TIME OF TEST: 12/17/2024 01:21:38

Diagnostics Check	OK	01:14
Air Blank	0.000	01:15
Control Test	0.077	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.064	01:16
Breath Volume	4.347	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.061	01:19
Breath Volume	3.253	01:19
Air Blank	0.000	01:20
Control Test	0.076	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTHRIE, ADAM, D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GUTHRIE, ADAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: ALLEMAND

DATE/TIME OF TEST: 12/20/2024 03:06:54

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.078	03:01
Air Blank	0.000	03:01
Subject Sample # 1	0.234	03:02
Breath Volume	2.695	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.223	03:05
Breath Volume	4.664	03:05
Air Blank	0.000	03:05
Control Test	0.077	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, DAVID, B
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: CASTELLANO, VINCENT
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 11/17/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 4028433437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MATT

DATE/TIME OF TEST: 12/21/2024 20:13:06

Diagnostics Check	OK	20:05
Air Blank	0.000	20:06
Control Test	0.078	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.124	20:08
Breath Volume	1.773	20:08
Air Blank	0.000	20:09
Air Blank	0.000	20:11
Subject Sample # 2	0.127	20:11
Breath Volume	2.097	20:11
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:13
Diagnostics Check	OK	20:13

Data Download Date/Time: 12/31/2024 07:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRUMMOND,ALEX,V
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: SHELKEY,DAVID
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 12/08/2024 00:02:07

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.082	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.170	23:57
Breath Volume	2.621	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.173	00:00
Breath Volume	2.406	00:00
Air Blank	0.000	00:01
Control Test	0.083	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 12/30/2024 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAUENHOFER,TIMOTHY,H
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FRAUENHOFER,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MARTIN JIMENEZ

DATE/TIME OF TEST: 12/15/2024 20:27:50

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.081	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.198	20:23
Breath Volume	2.789	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:25
Subject Sample # 2	0.208	20:26
Breath Volume	4.019	20:26
Air Blank	0.000	20:26
Control Test	0.079	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/30/2024 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: BUTLER,MATTHEW
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A2
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: PETRONE

DATE/TIME OF TEST: 12/16/2024 02:13:11

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.240	02:08
Breath Volume	2.546	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.229	02:11
Breath Volume	2.242	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 12/30/2024 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JONATHAN,F
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MARCO,EMMA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 11/10/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096062
DRY GAS EXPIRY DATE: 07/15/2027
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 12/20/2024 17:24:59

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.081	17:19
Air Blank	0.000	17:19
Subject Sample # 1	0.188	17:20
Breath Volume	2.546	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.182	17:23
Breath Volume	2.957	17:23
Air Blank	0.000	17:24
Control Test	0.082	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

Data Download Date/Time: 12/22/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS, ABIGAYLE, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: EDWARDS, ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/10/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096062
DRY GAS EXPIRY DATE: 07/15/2027
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:49

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 12/22/2024 07:16:15

Diagnostics Check	OK	07:09
Air Blank	0.000	07:10
Control Test	0.082	07:10
Air Blank	0.000	07:10
Subject Sample # 1	0.157	07:11
Breath Volume	2.746	07:11
Air Blank	0.000	07:11
Air Blank	0.000	07:13
Subject Sample # 2	0.154	07:14
Breath Volume	2.292	07:14
Air Blank	0.000	07:15
Control Test	0.082	07:15
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 12/22/2024 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: AMADORZAMBRANA

DATE/TIME OF TEST: 12/02/2024 04:11:10

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.147	04:06
Breath Volume	2.468	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.143	04:09
Breath Volume	1.925	04:09
Air Blank	0.000	04:10
Control Test	0.079	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,HENRY,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GARCIA,HENRY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MARTINEZ RAMOS

DATE/TIME OF TEST: 12/02/2024 18:32:21

Diagnostics Check	OK	18:28	
Air Blank	0.000	18:28	
Control Test	0.079	18:29	
Air Blank	0.000	18:29	
Subject Sample	# 1	18:30	Subject Test Refused
Breath Volume	0.000	18:30	
Air Blank	0.000	18:31	
Control Test	0.079	18:31	
Air Blank	0.000	18:32	
Diagnostics Check	OK	18:32	

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RODRIGUEZ,GIOVANY
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GONZALEZ PEREZ

DATE/TIME OF TEST: 12/02/2024 23:51:24

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.157	23:46
Breath Volume	1.273	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.160	23:49
Breath Volume	1.539	23:49
Air Blank	0.000	23:50
Control Test	0.078	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 12/05/2024 23:49:32

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.000	23:45
Breath Volume	2.484	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.000	23:48
Breath Volume	1.542	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: NUNEZ ECHENIQUE

DATE/TIME OF TEST: 12/06/2024 02:41:24

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.227	02:36
Breath Volume	4.855	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.237	02:39
Breath Volume	3.734	02:39
Air Blank	0.000	02:40
Control Test	0.077	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: FERREIRA

DATE/TIME OF TEST: 12/06/2024 02:51:55

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.180	02:47
Breath Volume	1.660	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.170	02:50
Breath Volume	1.367	02:50
Air Blank	0.000	02:51
Control Test	0.076	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 12/06/2024 03:16:02

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:04	
Control Test	0.079	03:05	
Air Blank	0.000	03:05	
Subject Sample	# 1 0.141	03:08	Volume Not Met
Breath Volume	0.273	03:08	
Air Blank	0.000	03:09	
Air Blank	0.000	03:11	
Subject Sample	# 2 0.136	03:14	Volume Not Met
Breath Volume	0.312	03:14	
Air Blank	0.000	03:15	
Control Test	0.076	03:15	
Air Blank	0.000	03:15	
Diagnostics Check	OK	03:16	

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRO,JEFFREY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 12/06/2024 03:30:52

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.078	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1 0.122	03:23	Volume Not Met
Breath Volume	0.335	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 2 0.134	03:29	Volume Not Met
Breath Volume	0.726	03:29	
Air Blank	0.000	03:29	
Control Test	0.077	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:30	

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY'
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 12/07/2024 06:02:13

Diagnostics Check	OK	06:01	
Air Blank		06:01	Sequence Aborted
Air Blank	0.000	06:02	

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/16/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 12/07/2024 06:11:53

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.080	06:04
Air Blank	0.000	06:05
Subject Sample # 1	0.160	06:07
Breath Volume	1.679	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.167	06:10
Breath Volume	2.683	06:10
Air Blank	0.000	06:10
Control Test	0.076	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 12/07/2024 19:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ALONSOCALDERON

DATE/TIME OF TEST: 12/08/2024 01:34:17

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.207	01:29
Breath Volume	2.167	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.205	01:32
Breath Volume	1.941	01:32
Air Blank	0.000	01:33
Control Test	0.080	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SARTIPI

DATE/TIME OF TEST: 12/15/2024 02:36:17

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.115	02:31
Breath Volume	2.496	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.105	02:34
Breath Volume	1.917	02:34
Air Blank	0.000	02:35
Control Test	0.079	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: RIVERA BALASQUIDE

DATE/TIME OF TEST: 12/18/2024 03:41:10

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.079	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.143	03:36
Breath Volume	2.492	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.136	03:39
Breath Volume	2.074	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PETIT-HOMME

DATE/TIME OF TEST: 12/21/2024 06:00:40

Diagnostics Check	OK	05:52
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Subject Sample # 1	0.166	05:55
Breath Volume	2.105	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.171	05:59
Breath Volume	2.363	05:59
Air Blank	0.000	05:59
Control Test	0.078	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HODGSON SHARP

DATE/TIME OF TEST: 12/22/2024 01:02:59

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.080	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1 0.159	00:58	Volume Not Met
Breath Volume	0.046	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank		01:02	RFI Detect
Air Blank	0.000	01:02	

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HODGSON SHARP

DATE/TIME OF TEST: 12/22/2024 01:16:31

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1 0.214	01:09	
Breath Volume	1.730	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2 0.201	01:15	Volume Not Met
Breath Volume	0.109	01:15	
Air Blank		01:16	Purge Fail
Air Blank	0.000	01:16	

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HODGSON SHARP

DATE/TIME OF TEST: 12/22/2024 01:23:41

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.075	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:21	Subject Test Refused
Breath Volume	0.316	01:21	
Air Blank	0.000	01:22	
Control Test	0.077	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: PRESENDIEU

DATE/TIME OF TEST: 12/25/2024 22:50:49

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:38	
Control Test	0.081	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1 0.203	22:42	Volume Not Met
Breath Volume	0.785	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:45	
Subject Sample	# 2 0.270	22:45	
Breath Volume	2.187	22:45	
Air Blank	0.000	22:46	
Air Blank	0.000	22:48	
Subject Sample	# 3 0.271	22:49	
Breath Volume	2.175	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:50	
Air Blank	0.000	22:50	
Diagnostics Check	OK	22:50	

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,TREVOR,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: NEWBOLD

DATE/TIME OF TEST: 12/27/2024 01:48:05

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.184	01:42
Breath Volume	1.804	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.184	01:46
Breath Volume	1.996	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BODDEN,DANIEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BODDEN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 12/27/2024 02:50:41

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.113	02:44
Breath Volume	2.039	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.127	02:49
Breath Volume	1.796	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:50
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: RAYMUNDO RAYMUNDO

DATE/TIME OF TEST: 12/27/2024 23:37:22

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.191	23:32
Breath Volume	1.285	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.200	23:35
Breath Volume	1.847	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 02/24/2025 19:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANUEVA, HECTOR, A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VILLANUEVA, HECTOR
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 12/28/2024 02:40:16

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.080	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.158	02:30	Volume Not Met
Breath Volume	0.125	02:30	
Air Blank	0.000	02:31	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.164	02:33	No .020 agreement
Breath Volume	1.457	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 3 0.129	02:38	No .020 agreement
Breath Volume	2.253	02:38	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Diagnostics Check	OK	02:40	

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACEVEDO,ANGELA,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SEMIDEY

DATE/TIME OF TEST: 12/28/2024 03:29:06

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: OSORIO URBINA

DATE/TIME OF TEST: 12/29/2024 07:55:02

Diagnostics Check	OK	07:47
Air Blank	0.000	07:47
Control Test	0.080	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.172	07:50
Breath Volume	1.808	07:50
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.164	07:53
Breath Volume	2.265	07:53
Air Blank	0.000	07:54
Control Test	0.078	07:54
Air Blank	0.000	07:54
Diagnostics Check	OK	07:55

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JONES,KEVINE
ARRESTING OFFICER AGENCY: Miami Springs PD
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: VERA ZAMBRANO

DATE/TIME OF TEST: 12/29/2024 10:41:29

Diagnostics Check	OK	10:36	
Air Blank	0.000	10:37	
Control Test	0.080	10:37	
Air Blank	0.000	10:38	
Subject Sample	# 1	10:40	Subject Test Refused
Breath Volume	0.000	10:40	
Air Blank	0.000	10:40	
Control Test	0.080	10:40	
Air Blank	0.000	10:41	
Diagnostics Check	OK	10:41	

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ANDREW,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HERNANDEZ,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CHOW

DATE/TIME OF TEST: 12/29/2024 23:34:12

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.080	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1 0.222	23:26	Volume Not Met
Breath Volume	0.546	23:26	
Air Blank	0.000	23:27	
Air Blank	0.000	23:29	
Subject Sample	# 2	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULVEY,KEVIN,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MULVEY,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/07/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 12/31/2024 21:43:52

Diagnostics Check	OK	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Subject Sample # 1	0.167	21:38
Breath Volume	1.511	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.170	21:42
Breath Volume	1.507	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 02/24/2025 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 11/27/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HERNANDEZ CRUZ

DATE/TIME OF TEST: 12/18/2024 22:37:18

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:26	
Control Test	0.081	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1 0.000	22:30	Volume Not Met
Breath Volume	0.957	22:30	
Air Blank	0.000	22:31	
Air Blank	0.000	22:33	
Subject Sample	# 2	22:35	Subject Test Refused
Breath Volume	0.480	22:35	
Air Blank	0.000	22:36	
Control Test	0.083	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 01/04/2025 02:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILLIAMS,ERIC
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/26/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MINIACI

DATE/TIME OF TEST: 12/31/2024 00:43:10

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.081	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.067	00:33	
Breath Volume	2.386	00:33	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2 0.039	00:38	Volume Not Met
Breath Volume	0.617	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 3 0.068	00:41	
Breath Volume	1.363	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 01/22/2025 10:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSE, JESUS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CUZ TOX

DATE/TIME OF TEST: 12/01/2024 00:21:22

Diagnostics Check	OK	00:14
Air Blank	0.000	00:15
Control Test	0.083	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.083	00:16
Breath Volume	2.496	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.080	00:19
Breath Volume	2.222	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONCALVES, GUSTAVO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SEADER

DATE/TIME OF TEST: 12/01/2024 03:55:08

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.083	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.100	03:50
Breath Volume	1.753	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.096	03:53
Breath Volume	1.597	03:53
Air Blank	0.000	03:54
Control Test	0.083	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MASON,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: NESTON

DATE/TIME OF TEST: 12/01/2024 05:02:51

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.083	04:56
Air Blank	0.000	04:57
Subject Sample # 1	0.025	04:57
Breath Volume	1.304	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.025	05:01
Breath Volume	2.789	05:01
Air Blank	0.000	05:01
Control Test	0.083	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOMACK, ANTIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MERCADO

DATE/TIME OF TEST: 12/02/2024 04:18:50

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.083	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.127	04:14
Breath Volume	2.699	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.131	04:17
Breath Volume	2.171	04:17
Air Blank	0.000	04:17
Control Test	0.082	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER, THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: STOCKDALE

DATE/TIME OF TEST: 12/03/2024 01:50:46

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.083	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.132	01:45
Breath Volume	2.054	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.122	01:49
Breath Volume	1.472	01:49
Air Blank	0.000	01:49
Control Test	0.082	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DE OLIVEIRA,EDUARDO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 11/19/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: MENDEZ MIRANDA

DATE/TIME OF TEST: 12/04/2024 00:21:54

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.083	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.193	00:17
Breath Volume	1.507	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.188	00:20
Breath Volume	1.132	00:20
Air Blank	0.000	00:20
Control Test	0.083	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 12/04/2024 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/04/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CUADRA JR

DATE/TIME OF TEST: 12/21/2024 02:13:38

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.083	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.217	02:08
Breath Volume	2.226	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.219	02:12
Breath Volume	1.300	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 01/14/2025 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYLIS,BRYAN,T
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LYNCH,DANIEL
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 12/24/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: DUBLIN

DATE/TIME OF TEST: 12/25/2024 21:00:42

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.080	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.081	21:00	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 01/09/2025 20:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, TYRELL, J
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: MANIS, AMANDA
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 11/20/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 12/08/2024 08:04:26

Diagnostics Check	OK	07:59	
Air Blank	0.000	08:00	
Control Test	0.080	08:00	
Air Blank	0.000	08:00	
Subject Sample	# 1	08:03	Subject Test Refused
Breath Volume	0.253	08:03	
Air Blank	0.000	08:03	
Control Test	0.080	08:03	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 12/21/2024 18:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JOCELYN,L
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: INGRAM,RYAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 12/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: REMSEN

DATE/TIME OF TEST: 12/24/2024 16:41:46

Diagnostics Check	OK	16:26	
Air Blank	0.000	16:27	
Control Test	0.079	16:27	
Air Blank	0.000	16:28	
Subject Sample	# 1 0.229	16:31	Volume Not Met
Breath Volume	0.859	16:31	
Air Blank	0.000	16:32	
Air Blank	0.000	16:33	
Subject Sample	# 2 0.267	16:34	
Breath Volume	1.808	16:34	
Air Blank	0.000	16:35	
Air Blank	0.000	16:36	
Subject Sample	# 3 0.266	16:39	
Breath Volume	2.218	16:39	
Air Blank	0.000	16:40	
Control Test	0.077	16:41	
Air Blank	0.000	16:41	
Diagnostics Check	OK	16:41	

Data Download Date/Time: 02/01/2025 00:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON,J,
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: NUSL,G
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/21/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: MURRELL

DATE/TIME OF TEST: 12/25/2024 13:32:10

Diagnostics Check	OK	13:29	
Air Blank	0.000	13:29	
Control Test	0.080	13:30	
Air Blank	0.000	13:30	
Subject Sample	# 1	13:30	Subject Test Refused
Breath Volume	0.000	13:30	
Air Blank	0.000	13:31	
Control Test	0.079	13:31	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:32	

Data Download Date/Time: 02/01/2025 00:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELORBE,DOMINICK
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BARALA

DATE/TIME OF TEST: 12/01/2024 02:35:46

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.082	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	RFI Detect
Breath Volume	0.000	02:35	
Air Blank		02:35	RFI Detect
Air Blank	0.000	02:35	

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELORBE,DOMINICK
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BARALA

DATE/TIME OF TEST: 12/01/2024 02:44:27

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.137	02:40
Breath Volume	2.152	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.134	02:42
Breath Volume	2.046	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: THOMAS,ASTON
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/01/2024 03:12:11

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:04	
Control Test	0.081	03:04	
Air Blank	0.000	03:05	
Subject Sample	# 1 0.000	03:08	Volume Not Met
Breath Volume	0.000	03:08	
Air Blank	0.000	03:08	
Air Blank	0.000	03:10	
Subject Sample	# 2	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.077	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GILL,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/03/2024 03:00:59

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.207	02:56
Breath Volume	2.453	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.207	02:59
Breath Volume	2.589	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS, ADAM, J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PITTA, CEASER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 12/04/2024 06:37:15

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.082	06:31
Air Blank	0.000	06:32
Subject Sample # 1	0.135	06:32
Breath Volume	2.660	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.140	06:35
Breath Volume	3.167	06:35
Air Blank	0.000	06:36
Control Test	0.080	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELROSSO,DOMINIC
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: FRIGON

DATE/TIME OF TEST: 12/06/2024 01:12:38

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.082	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.209	01:08
Breath Volume	2.527	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.200	01:11
Breath Volume	2.562	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BYRD,RYAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/06/2024 01:50:29

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.083	01:46
Breath Volume	3.300	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.083	01:48
Breath Volume	2.570	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOLAND,THOMAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/07/2024 00:50:56

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.179	00:46
Breath Volume	2.378	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:48
Subject Sample # 2	0.167	00:49
Breath Volume	1.960	00:49
Air Blank	0.000	00:50
Control Test	0.082	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOLKOV,ISABELLA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: BALARI

DATE/TIME OF TEST: 12/07/2024 04:57:11

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.145	04:52
Breath Volume	2.687	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.150	04:55
Breath Volume	2.898	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NOLAND,T
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CALHOUN

DATE/TIME OF TEST: 12/09/2024 00:10:06

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.082	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.183	00:05
Breath Volume	1.734	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.173	00:08
Breath Volume	2.140	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BYRD,R
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: PEREZ LOPEZ

DATE/TIME OF TEST: 12/10/2024 05:26:26

Diagnostics Check	OK	05:19
Air Blank	0.000	05:19
Control Test	0.082	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.159	05:21
Breath Volume	2.089	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.153	05:24
Breath Volume	1.796	05:24
Air Blank	0.000	05:25
Control Test	0.079	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CAEZ-ROJAS,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: REYNOSO LOPEZ

DATE/TIME OF TEST: 12/11/2024 03:33:29

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.238	03:28
Breath Volume	1.867	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.248	03:31
Breath Volume	1.480	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/17/2024 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,CHASE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FRANKLIN

DATE/TIME OF TEST: 12/11/2024 04:18:21

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.000	04:13
Breath Volume	1.632	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.000	04:16
Breath Volume	1.699	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RODRIGUEZ,JOSHUA
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: O'TOOLE

DATE/TIME OF TEST: 12/13/2024 02:11:01

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.153	02:06
Breath Volume	3.871	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.153	02:09
Breath Volume	3.796	02:09
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:11

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MATEO,YORDAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/13/2024 23:00:54

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.082	22:55
Air Blank	0.000	22:55
Subject Sample # 1	0.173	22:56
Breath Volume	1.574	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:58
Subject Sample # 2	0.187	22:59
Breath Volume	1.683	22:59
Air Blank	0.000	23:00
Control Test	0.081	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VAN ROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TULLY

DATE/TIME OF TEST: 12/13/2024 23:10:05

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.237	23:05
Breath Volume	3.820	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.225	23:08
Breath Volume	3.367	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:10

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,JUSTIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 12/14/2024 22:28:58

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.082	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.116	22:24
Breath Volume	2.660	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.112	22:27
Breath Volume	2.398	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NORTHCUTT,ERIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 12/15/2024 02:11:38

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.082	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.192	02:07
Breath Volume	2.652	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.195	02:10
Breath Volume	2.792	02:10
Air Blank	0.000	02:10
Control Test	0.082	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: THOMAS,A
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: BARILLAS GARCIA

DATE/TIME OF TEST: 12/15/2024 21:55:52

Diagnostics Check	OK	21:49
Air Blank	0.000	21:49
Control Test	0.081	21:49
Air Blank	0.000	21:50
Subject Sample # 1	0.230	21:51
Breath Volume	1.851	21:51
Air Blank	0.000	21:51
Air Blank	0.000	21:53
Subject Sample # 2	0.228	21:54
Breath Volume	2.050	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:55

Data Download Date/Time: 12/17/2024 05:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARMSTRONG,I
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 12/17/2024 22:39:43

Diagnostics Check	OK	22:32
Air Blank	0.000	22:33
Control Test	0.082	22:33
Air Blank	0.000	22:33
Subject Sample # 1	0.080	22:35
Breath Volume	1.746	22:35
Air Blank	0.000	22:35
Air Blank	0.000	22:37
Subject Sample # 2	0.077	22:38
Breath Volume	1.687	22:38
Air Blank	0.000	22:38
Control Test	0.081	22:39
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOUROUMOUSIS,GEORGE,T
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEVERNOE,MATTHEW
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: VRLJANVIC

DATE/TIME OF TEST: 12/19/2024 21:48:25

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.165	21:43
Breath Volume	1.914	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.161	21:46
Breath Volume	1.472	21:46
Air Blank	0.000	21:47
Control Test	0.081	21:47
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CAEZ-ROJAS,J
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LA ROSA

DATE/TIME OF TEST: 12/20/2024 01:43:18

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.082	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.123	01:38
Breath Volume	1.855	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.122	01:41
Breath Volume	3.023	01:41
Air Blank	0.000	01:42
Control Test	0.080	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEVERNOE,M
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HOLLIDAY

DATE/TIME OF TEST: 12/20/2024 04:38:34

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.183	04:33
Breath Volume	2.500	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.187	04:37
Breath Volume	2.855	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CANADY,CHRISTINA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 12/20/2024 21:12:29

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:59	
Control Test	0.081	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1 0.189	21:03	Volume Not Met
Breath Volume	0.335	21:03	
Air Blank	0.000	21:03	
Air Blank	0.000	21:05	
Subject Sample	# 2 0.196	21:07	
Breath Volume	2.289	21:07	
Air Blank	0.000	21:08	
Air Blank	0.000	21:10	
Subject Sample	# 3 0.183	21:10	
Breath Volume	1.347	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank	0.000	21:12	
Diagnostics Check	OK	21:12	

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHOANAOVER,S
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 12/21/2024 03:37:10

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.082	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1 0.000	03:33	
Breath Volume	1.316	03:33	
Air Blank	0.000	03:34	
Air Blank	0.000	03:36	
Subject Sample	# 2	03:36	RFI Detect
Breath Volume	0.214	03:36	
Air Blank		03:36	RFI Detect
Air Blank	0.000	03:37	

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,LEONARD
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SABOURIN

DATE/TIME OF TEST: 12/21/2024 03:53:34

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.208	03:49
Breath Volume	6.132	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.208	03:51
Breath Volume	5.695	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,MICHAEL
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/22/2024 04:34:10

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.081	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.097	04:29
Breath Volume	2.320	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.095	04:32
Breath Volume	2.218	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,JAMES,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: VELAZQUEZ,RODERICK
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: BOYER

DATE/TIME OF TEST: 12/22/2024 07:48:59

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:36	
Control Test	0.082	07:36	
Air Blank	0.000	07:37	
Subject Sample	# 1 0.108	07:40	Volume Not Met
Breath Volume	0.085	07:40	
Air Blank	0.000	07:41	
Air Blank	0.000	07:42	
Subject Sample	# 2 0.094	07:43	
Breath Volume	2.937	07:43	
Air Blank	0.000	07:44	
Air Blank	0.000	07:46	
Subject Sample	# 3 0.101	07:47	
Breath Volume	2.156	07:47	
Air Blank	0.000	07:48	
Control Test	0.080	07:48	
Air Blank	0.000	07:48	
Diagnostics Check	OK	07:48	

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHOONOVER,S
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/23/2024 05:31:26

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.082	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.146	05:26
Breath Volume	2.648	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.142	05:29
Breath Volume	2.144	05:29
Air Blank	0.000	05:30
Control Test	0.082	05:30
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALLEY,B
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: DELUCIAN

DATE/TIME OF TEST: 12/24/2024 21:44:35

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.082	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.000	21:39
Breath Volume	1.718	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.000	21:43
Breath Volume	1.390	21:43
Air Blank	0.000	21:43
Control Test	0.082	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,L
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: IRISH-FARIAS

DATE/TIME OF TEST: 12/25/2024 23:36:56

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.082	23:28
Air Blank	0.000	23:29
Subject Sample # 1	0.229	23:31
Breath Volume	1.585	23:31
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.241	23:35
Breath Volume	1.128	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 01/02/2025 00:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ALBRECHT,LEONARD
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: BROWING

DATE/TIME OF TEST: 12/28/2024 00:45:38

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.082	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.211	00:38
Breath Volume	3.007	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.183	00:40
Breath Volume	2.238	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 3	0.183	00:44
Breath Volume	2.312	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 01/02/2025 00:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,ALEXANDER
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/28/2024 03:52:23

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.082	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:48	No Sample Provided
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2	03:51	Sequence Aborted
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	

Data Download Date/Time: 01/02/2025 00:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UTZIG,NATHANIEL
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 12/28/2024 04:21:12

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.240	04:16
Breath Volume	3.937	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.242	04:19
Breath Volume	3.359	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 01/02/2025 00:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ESTRADA,B
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LA LUZ

DATE/TIME OF TEST: 12/30/2024 02:53:19

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.081	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1 0.153	02:46	Volume Not Met
Breath Volume	0.042	02:46	
Air Blank	0.000	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 2	02:51	Subject Test Refused
Breath Volume	0.210	02:51	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 01/02/2025 00:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ESTRADA,B
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 11/14/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623830
DRY GAS EXPIRY DATE: 12/28/2025
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LA LUZ

DATE/TIME OF TEST: 12/30/2024 03:03:04

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.171	02:58
Breath Volume	1.851	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.160	03:01
Breath Volume	2.171	03:01
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 01/02/2025 00:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLSON,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG326805
DRY GAS EXPIRY DATE: 09/25/2025
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: DEL VALLE

DATE/TIME OF TEST: 12/12/2024 00:00:00

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.079	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.079	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 02/04/2025 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,DAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG326805
DRY GAS EXPIRY DATE: 09/25/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: FIELD

DATE/TIME OF TEST: 12/26/2024 03:34:24

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.202	03:29
Breath Volume	3.195	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.196	03:32
Breath Volume	2.757	03:32
Air Blank	0.000	03:33
Control Test	0.078	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 02/04/2025 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG326805
DRY GAS EXPIRY DATE: 09/25/2025
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/26/2024 22:15:21

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:12	
Control Test	0.079	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.078	22:14	
Air Blank	0.000	22:15	
Diagnostics Check	OK	22:15	

Data Download Date/Time: 02/04/2025 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLE,ELIZABETH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG326805
DRY GAS EXPIRY DATE: 09/25/2025
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PARRAVANI

DATE/TIME OF TEST: 12/28/2024 02:58:52

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.165	02:54
Breath Volume	1.789	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.162	02:57
Breath Volume	1.105	02:57
Air Blank	0.000	02:57
Control Test	0.077	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 02/04/2025 14:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/11/2024

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG326805
DRY GAS EXPIRY DATE: 09/25/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BOBA

DATE/TIME OF TEST: 12/31/2024 22:42:23

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.078	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:41	
Control Test	0.079	22:41	
Air Blank	0.000	22:42	
Diagnostics Check	OK	22:42	

Data Download Date/Time: 02/04/2025 14:12