

Alcohol Testing Program - Instrument Processing Sheet

Agency: PONCE INLET PD

Instrument Serial Number: 80-001158

Date In: 6/24/2026

DI Completion Date: 7/9/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/24/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>6/30/2026</u>		Flow Adjustment By: _____																																									
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Amended Form 40 regarding upload issue was shipped with instrument; placed copy in correspondence folder. SLH 6/24/2026		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>218</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.535</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1008</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																									
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 03/03/2028	0.080	MP5089	202603B 03/03/2028	0.200	MP5090	202603D 03/04/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																										
Simulator	Serial #	Lot#/Exp																																											
0.050	MP5088	202603A 03/03/2028																																											
0.080	MP5089	202603B 03/03/2028																																											
0.200	MP5090	202603D 03/04/2028																																											
0.080 DGS	N/A	23825080A1 10/5/2027																																											
Optical Bench Adjustment By: <u>SLH</u> Barometric Pressure Gauge: <u>1014</u> ID#: <u>28662</u>		Department Inspection By: <u>SLH</u> Barometric Pressure ID#: <u>34419</u>																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25120 50 <u>SLH</u></td> <td>5/6/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>09/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25120 50 <u>SLH</u>	5/6/2027	0.080 DGS	N/A	AG526603	09/23/2027	Gauge: <u>1016</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																										
0.000	MP6294	N/A	N/A																																										
0.040	MP6295	25090	3/11/2027																																										
0.100	MP6296	24390	10/29/2026																																										
0.200	MP6297	25020	1/14/2027																																										
0.300	MP6298	25120 50 <u>SLH</u>	5/6/2027																																										
0.080 DGS	N/A	AG526603	09/23/2027																																										
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP5088																																												
0.080	MP5089																																												
0.200	MP5090																																												
☒ Post Optical Bench Adjustment Stability Checks		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: _____																							
Simulator	Serial #	Lot #	Expiration																																										
0.050	MP5088	202603A	03/03/2028																																										
0.080	MP5089	202603B	03/03/2028																																										
0.200	MP5090	202603D	03/04/2028																																										
0.080 DGS	N/A	23825080A1	10/5/2027																																										
Gauge ID #: <u>34419</u> Gauge: <u>1014</u> Instrument: <u>1014</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
Notes/Suggested Service: Stabilities performed 7/1/2026. OBA=optical bench adjustment. SLH 7/1/2026 Post stabilities performed 2 separate days. SLH 7/7/2026 Corrected OBA 0.300 lot # to read 25150. SLH 7/9/2026		<table border="0" style="width: 100%;"> <tr> <td style="width: 50%; vertical-align: bottom;"> Digitally signed by Wen-Chi Pierson Date: 2026.07.10 09:53:12 -04'00' </td> <td style="width: 50%; vertical-align: bottom;"> Digitally signed by Kaitlyn Spearin Date: 2026.07.14 07:58:26 -04'00' </td> </tr> </table>				Digitally signed by Wen-Chi Pierson Date: 2026.07.10 09:53:12 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.07.14 07:58:26 -04'00'																																						
Digitally signed by Wen-Chi Pierson Date: 2026.07.10 09:53:12 -04'00'	Digitally signed by Kaitlyn Spearin Date: 2026.07.14 07:58:26 -04'00'																																												
Tech Review		Admin Review																																											

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PONCE INLET PD
Time of Inspection: 09:19

Date of Inspection: 07/01/2026

Serial Number: 80-001158
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/01/2026
Date

Stability Checks

80-001158

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet
PONCE INLET PD Intoxilyzer - Alcotest Analyzer Model 8000 07/01/2026 SN 80-001158 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:42 Control Test 0.049 09:42 Air Blank 0.000 09:43 Control Test 0.049 09:44 Air Blank 0.000 09:44 Control Test 0.049 09:45 Air Blank 0.000 09:45 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>M. Sigala</i>	PONCE INLET PD Intoxilyzer - Alcotest Analyzer Model 8000 07/01/2026 SN 80-001158 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:30 Control Test 0.077 09:30 Air Blank 0.000 09:31 Control Test 0.077 09:32 Air Blank 0.000 09:32 Control Test 0.077 09:33 Air Blank 0.000 09:33 Control Test Stats Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>M. Sigala</i>	PONCE INLET PD Intoxilyzer - Alcotest Analyzer Model 8000 07/01/2026 SN 80-001158 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:55 Control Test 0.198 09:55 Air Blank 0.000 09:56 Control Test 0.199 09:57 Air Blank 0.000 09:57 Control Test 0.198 09:58 Air Blank 0.000 09:58 Control Test Stats Average 0.1993 Std Dev 0.0006 Rel Std Dev(%) 0.2911 Operator's Signature <i>M. Sigala</i>	PONCE INLET PD Intoxilyzer - Alcotest Analyzer Model 8000 07/01/2026 SN 80-001158 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:36 Control Test 0.080 09:36 Air Blank 0.000 09:36 Control Test 0.081 09:37 Air Blank 0.000 09:37 Control Test 0.081 09:37 Air Blank 0.000 09:38 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157 Operator's Signature <i>M. Sigala</i>

Will need OBA

Verified instrument was able to read gas pressure.



PONCE INLET PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001158
07/07/2026 13:53:25

SUB

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0770	(-0.0060)	0.0820	(-0.0200)
Sample #2	0.0770	(0.0020)	0.0760	(-0.0170)
Sample #3	0.0760	(0.0080)	0.0740	(-0.0120)
Sample #4	0.1210	(-0.0020)	0.0830	(-0.0210)
Avg % Abs	0.0913	(0.0027)	0.0777	(-0.0167)
STD DEV	0.0257	(0.0050)	0.0047	(0.0045)
REL STD DEV	28.135	(188.746)	6.085	(27.055)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8580	(-0.0140)	1.4960	(-0.0200)
Sample #2	0.8520	(0.0000)	1.4860	(0.0020)
Sample #3	0.8440	(0.0080)	1.4730	(0.0030)
Sample #4	0.8200	(0.0320)	1.4570	(0.0190)
Avg % Abs	0.8387	(0.0133)	1.4720	(0.0080)
STD DEV	0.0167	(0.0167)	0.0145	(0.0095)
REL STD DEV	1.986	(124.900)	0.987	(119.242)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.9170	(-0.0120)	3.5050	(0.0060)
Sample #2	1.9030	(0.0100)	3.5030	(0.0170)
Sample #3	1.9300	(0.0070)	3.5290	(0.0030)
Sample #4	1.8850	(0.0300)	3.5130	(0.0160)
Avg % Abs	1.9060	(0.0157)	3.5150	(0.0120)
STD DEV	0.0226	(0.0125)	0.0131	(0.0078)
REL STD DEV	1.188	(79.809)	0.373	(65.085)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.7070	(-0.0060)	6.8270	(-0.0130)
Sample #2	3.6920	(0.0030)	6.7980	(0.0160)
Sample #3	3.6970	(0.0090)	6.8010	(0.0130)
Sample #4	3.6980	(0.0180)	6.8140	(0.0000)
Avg % Abs	3.6957	(0.0100)	6.8043	(0.0097)
STD DEV	0.0032	(0.0075)	0.0085	(0.0085)
REL STD DEV	0.087	(75.498)	0.125	(87.982)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.3650	(-0.0030)	9.9090	(-0.0210)
Sample #2	5.3840	(-0.0040)	9.8800	(-0.0080)
Sample #3	5.3750	(0.0190)	9.8710	(0.0130)
Sample #4	5.3760	(0.0180)	9.8780	(-0.0080)
Avg % Abs	5.3783	(0.0110)	9.8763	(-0.0010)
STD DEV	0.0049	(0.0130)	0.0047	(0.0121)
REL STD DEV	0.092	(118.182)	0.048	(1212.435)

PONCE INLET PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001158 *SUB*
07/07/2026 13:53:25

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-230.47			-91.86	
First Order Coef	2542.47			1332.11	
Second Order Coef	28.71			12.42	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	0.000	-0.0000	0.000	0.000	-0.0002
0.040	0.040	-0.0004	0.040	0.040	0.0002
0.100	0.099	0.0009	0.100	0.100	0.0004
0.200	0.201	-0.0007	0.200	0.200	-0.0005
0.300	0.300	0.0002	0.300	0.300	0.0002
-----			-----		

<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3144.00		3490.00
Sample #2	3130.00		3476.00
Sample #3	3203.00		3516.00
Sample #4	3156.00		3492.00
Avg	3163.0000		3494.6667
STD DEV	37.0000		20.1329
REL STD DEV	1.170		0.576
H2O adjust (mg/l*10k)	646		315

Barometric Pressure = 1014

*****CALIBRATION SUCCESSFUL*****

Post Optical Bench Adjustment Stability Checks 80-001158

SAT

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L
50.003 g/210L of Wet			
PONCE INLET PC Intoxilyzer - Alcohol Analyzer Model 8100 SN 81-001158 07/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:19 Control Test 0.050 09:20 Air Blank 0.000 09:21 Control Test 0.050 09:22 Air Blank 0.000 09:22 Control Test 0.050 09:23 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Thyiscal</i> Operator's Signature	PONCE INLET PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001158 07/07/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:52 Control Test 0.079 14:53 Air Blank 0.000 14:53 Control Test 0.078 14:54 Air Blank 0.000 14:55 Control Test 0.078 14:55 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370 <i>Thyiscal</i> Operator's Signature	PONCE INLET PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001158 07/09/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:26 Control Test 0.198 09:27 Air Blank 0.000 09:27 Control Test 0.197 09:28 Air Blank 0.000 09:29 Control Test 0.198 09:29 Air Blank 0.000 09:30 Control Test Stats Average 0.1977 Std Dev 0.0006 Rel Std Dev(%) 0.2921 <i>Thyiscal</i> Operator's Signature	PONCE INLET PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001158 07/07/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:15 Control Test 0.079 15:15 Air Blank 0.000 15:16 Control Test 0.079 15:16 Air Blank 0.000 15:16 Control Test 0.079 15:17 Air Blank 0.000 15:17 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Thyiscal</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PONCE INLET PD

Time of Inspection: 13:03

Date of Inspection: 07/09/2026

Serial Number: 80-001158

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.196	0.080
0.000	0.050	0.077	0.196	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.195	0.081
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.195	0.081


Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:


The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/09/2026
Date



ANAB
ANSI National Accredited Body
ACCREDITED
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001158, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001158</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PONCE INLET PD</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/09/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>13:03</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/09/2026

Date

Myra Lott
LEANDRA HIGGINBOTHAM,
Department Inspector