

# Instrument Processing Sheet

Agency: FHP Instrument Serial Number: 80-006628  
 Date In: 11/24/2025 DI Completion Date: 12/2/2025  Ship  P/U  H/D  CMI  EE

<b>Intake</b> By: <u>SLH</u> Date: <u>11/24/25</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Was missing DGS o-ring	<b>Quality Checks</b> By: <u>SLH</u> Date: <u>12/1/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.472</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1017</u> Instrument: <u>1016</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Adjustment</b> By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
---	--	--

Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:
0.050	MP6291	202406K	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:	SLH	
		06/19/2026			
0.080	MP6292	202406L		SLH	
		06/19/2026			
0.200	MP6293	202406N		SLH	
		06/20/2026			
0.080 DGS	N/A	AG510701		SLH	
		04/17/2027			

<b>Optical Bench Adjustment</b> By: _____	<b>Department Inspection</b> By: <u>SLH</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28421</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1007</u> Instrument: <u>1005</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP6292																																								
0.200	MP6293																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks and addl DGS stability <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other: Note from Agency																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;">                     Digitally signed by Taylor Gutschow                      Date: 2025.12.03 14:38:54 -05'00'                 </div> <div style="text-align: center;"> <b>Shayla Platt</b>                      Admin Review                 </div> <div style="text-align: center;">                     Digitally signed by Shayla Platt                      Date: 2025.12.04 14:59:11 -05'00'                 </div> </div>
--------------------------	---

11/20/2025

80-006628 – Intoxilyzer 8000

The instrument gave multiple Ambient Failures during Inspection. Room was clear was of Electronics and there was nothing present that would have triggered it. May be a tank purge issue.

Please return to:

Sergeant Maykoll Souza  
11059 International Drive  
Orlando, Florida 32821

# Stability Checks

80-006628 Sub  
12/1/2025

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet																																																																																																																																																
<p>Performed Root Cause Analysis</p> <p>FHP Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006628 12/01/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.010</td><td>12:24</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.010</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.010</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.010</td><td>12:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.047</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.010	12:24	Control Test	0.049	12:24	Air Blank	0.010	12:25	Control Test	0.048	12:25	Air Blank	0.010	12:26	Control Test	0.049	12:27	Air Blank	0.010	12:27	Control Test Stats			Average	0.047		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>Performed Root Cause Analysis</p> <p>FHP Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006628 12/01/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:14	Control Test	0.080	12:14	Air Blank	0.000	12:15	Control Test	0.079	12:15	Air Blank	0.000	12:16	Control Test	0.079	12:17	Air Blank	0.000	12:17	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>Performed Root Cause Analysis</p> <p>FHP Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006628 12/01/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.200</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.200</td><td>12:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.200</td><td>12:21</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:21</td></tr> <tr><td>Air Blank</td><td>0.200</td><td>12:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.200	12:18	Control Test	0.200	12:19	Air Blank	0.200	12:19	Control Test	0.200	12:20	Air Blank	0.200	12:21	Control Test	0.200	12:21	Air Blank	0.200	12:22	Control Test Stats			Average	0.2000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>Performed Root Cause Analysis</p> <p>FHP Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006628 12/01/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:09	Control Test	0.080	12:09	Air Blank	0.000	12:10	Control Test	0.081	12:10	Air Blank	0.000	12:10	Control Test	0.081	12:11	Air Blank	0.000	12:11	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.010	12:24																																																																																																																																																	
Control Test	0.049	12:24																																																																																																																																																	
Air Blank	0.010	12:25																																																																																																																																																	
Control Test	0.048	12:25																																																																																																																																																	
Air Blank	0.010	12:26																																																																																																																																																	
Control Test	0.049	12:27																																																																																																																																																	
Air Blank	0.010	12:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.047																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.080	12:14																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.079	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.079	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.200	12:18																																																																																																																																																	
Control Test	0.200	12:19																																																																																																																																																	
Air Blank	0.200	12:19																																																																																																																																																	
Control Test	0.200	12:20																																																																																																																																																	
Air Blank	0.200	12:21																																																																																																																																																	
Control Test	0.200	12:21																																																																																																																																																	
Air Blank	0.200	12:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.080	12:09																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.081	12:10																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.081	12:11																																																																																																																																																	
Air Blank	0.000	12:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		
<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 																																																																																																																																																

# Additional Stabilities

FHP  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
12/01/2025  
Software: 8100.27

**DQS**  
SN 80-006628

12/1/2025 80-006628 set

Due to note from agency,  
ran additional 10-DQS  
stabilities after quality checks  
completed. No ambient failures.  
(same lot/expiration) **SUT**

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.081	12:31
Air Blank	0.000	12:31
Control Test	0.080	12:32
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:33
Air Blank	0.000	12:34
Control Test	0.080	12:34
Air Blank	0.000	12:35
Control Test	0.081	12:35
Air Blank	0.000	12:35
Control Test	0.081	12:36
Air Blank	0.000	12:36
Control Test	0.081	12:37
Air Blank	0.000	12:37
Control Test	0.080	12:37
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0807	
Std Dev	0.0005	
Rel Std Dev(%)	0.5986	

  
Operator Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP  
Time of Inspection: 12:14

Date of Inspection: 12/02/2025

Serial Number: 80-006628  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.081
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.079	0.200	0.080


Standard Deviations	0.0005	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

12/02/2025  
Date



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2331 Phillips Road  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006628, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006628</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/02/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:14</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3)  
 The instrument results before and after any adjustment are found in the associated pre and post stability checks

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Mejlo*

12/02/2025

Date

LEANDRA HIGGINBOTHAM,  
 Department Inspector

FDLE/ATP Form 69 October 2024  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality