

Alcohol Testing Program - Instrument Processing Sheet

Agency: ESCAMBIA COUNTY SO

Instrument Serial Number: 80-006435

Date In: 6/19/2026

DI Completion Date: 7/15/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>6/19/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>6/26/2026</u>		Flow Adjustment By: <u>WKP</u>																																	
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>220</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.082</u> (.139-.169) 36 mm <u>0.093</u> (.156-.190) 53 mm <u>0.125</u> (.228-.278) 103 mm <u>0.281</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1016</u> Instrument: <u>1018</u> ☒ Stability Checks		ATP 102 / ATP 103 / ATP Flow Column #: <u>102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>223 / 224 / 225</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP 105 / ATP 105 / ATP 105</u> 32 mm <u>0.246/0.253/0.246</u> (.139-.169) 36 mm <u>0.250/0.269/0.250</u> (.156-.190) 53 mm <u>0.300/0.347/0.308</u> (.228-.278) 103 mm <u>0.488/0.488/0.496</u> (.447-.547)																																	
						<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____														
						Simulator	Serial #	Lot#/Exp																													
						0.050	MP5088	202603A 3/3/2028																													
						0.080	MP5089	202603B 3/3/2028																													
0.200	MP5090	202603D 3/4/2028																																			
0.080 DGS	N/A	23825080A1 10/5/2027																																			
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																																			
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34420</u>																																			
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1016</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>							
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Simulator	Serial Number																																				
0.000	MP5086																																				
Interferent	MP5087																																				
0.050	MP5088																																				
0.080	MP5089																																				
0.200	MP5090																																				
Gauge ID #: _____ Gauge: _____ Instrument: _____		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Adjustment ☒ Calibration Certificate ☐ Form 40 ☐ Optical Bench Adjustment ☒ Other: <u>AI Note</u>																																			
Notes/Suggested Service: Agency requested deregistration of instrument after being contacted about the instrument needing repairs with regards to the failed flow adjustments. The department inspection was subsequently performed to ensure the instrument was in compliance prior to deregistration after reviewing the breath testing history of the instrument. WKP 7/15/2026 Instrument was de-registered on 7/15/2026. Instrument settings were updated to put the instrument into 'disabled mode'. WKP 7/15/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																			
		<table style="width:100%;"> <tr> <td style="width:50%;"> Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.16 10:00:43 -04'00' </td> <td style="width:50%;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.16 10:18:24 -04'00' </td> </tr> <tr> <td>Tech Review</td> <td>Admin Review</td> </tr> </table>		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.16 10:00:43 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.16 10:18:24 -04'00'	Tech Review	Admin Review																														
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.07.16 10:00:43 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.16 10:18:24 -04'00'																																				
Tech Review	Admin Review																																				

NOT RECOGNISING BREATH TEST

mempr 6/19/20 80-000435

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 08:49

Date of Inspection: 06/26/2026

Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

06/26/2026

Date

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

#1

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 2.000

2: Rate (Liters/min) = 15

SQRT(Diff)) = 7.547

3: Rate (Liters/min) = 30

SQRT(Diff)) = 11.660

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 994

Rounded Intercept = -131457

Correlation = 0.97998

nenpi

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

#12

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 2.000

2: Rate (Liters/min) = 15

SQRT(Diff)) = 7.742

3: Rate (Liters/min) = 30

SQRT(Diff)) = 11.746

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 980

Rounded Intercept = -129778

Correlation = 0.97642

ment

#3

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 2.449
2: Rate (Liters/min) = 15
SQRT(Diff)) = 7.480
3: Rate (Liters/min) = 30
SQRT(Diff)) = 11.703
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 1046
Rounded Intercept = -265147
Correlation = 0.98636

merpi

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:48
Control Test	0.050	09:48
Air Blank	0.000	09:49
Control Test	0.049	09:50
Air Blank	0.000	09:50
Control Test	0.049	09:51
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

met

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.078	10:01
Air Blank	0.000	10:01
Control Test	0.078	10:02
Air Blank	0.000	10:02
Control Test	0.078	10:03
Air Blank	0.000	10:04
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.199	10:29
Air Blank	0.000	10:30
Control Test	0.199	10:30
Air Blank	0.000	10:31
Control Test	0.199	10:31
Air Blank	0.000	10:32
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

DGS #1

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:58
Control Test	0.078	12:58
Air Blank	0.000	12:58
Control Test	0.078	12:59
Air Blank	0.000	12:59
Control Test	INT*	13:00
Air Blank	PUR**	13:00
Air Blank	PUR**	13:01

*Interferent Detect

**Purge Fail

menye

Operator's Signature

was unhooking a simulator on a nearby
instrument containing acetone. Removed
simulator from proximity, turned on fan, and
repeated. WCP 6/26/2026

DGS #2

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
06/26/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:03
Control Test	0.078	13:04
Air Blank	0.000	13:04
Control Test	0.078	13:05
Air Blank	0.000	13:05
Control Test	0.078	13:05
Air Blank	0.000	13:06
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 09:44

Date of Inspection: 07/15/2026

Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23725080A1 Exp: 10/05/2027
0.000	0.048	0.077	0.197	0.077
0.000	0.049	0.077	0.198	0.077
0.000	0.049	0.077	0.198	0.077
0.000	0.048	0.077	0.198	0.076
0.000	0.048	0.077	0.198	0.076
0.000	0.048	0.077	0.198	0.076
0.000	0.048	0.077	0.198	0.076
0.000	0.048	0.076	0.198	0.076
0.000	0.048	0.077	0.198	0.076

Standard Deviations	0.0004	0.0003	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON
Signature and Printed Name

07/15/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006435, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006435</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/15/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:44</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Chi Pierson
Date: 2026.07.15
12:59:57 -04'00'

Pierson

07/15/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality