

Alcohol Testing Program - Instrument Processing Sheet

Agency: Crestview PD Instrument Serial Number: 80-005058

Date In: 5/8/2026 DI Completion Date: 5/15/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>IAS</u> Date: <u>5/8/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>5/8/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>211</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.164</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1011</u> Instrument: <u>1002</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																															
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:25%;">Lot#/Exp</th> <th style="width:40%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603D</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>202603B</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23725080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603D	☐ Tank Sensor Tare	202603B	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23725080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603D	☐ Tank Sensor Tare																																
		202603B	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23725080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>WKP</u>		Department Inspection By: <u>WKP</u>																																	
Barometric Pressure Gauge: <u>1015</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>34419</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1015</u> Instrument: <u>1015</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																															
0.000	MP6294	N/A	N/A																																
0.040	MP6295	25090	3/11/2027																																
0.100	MP6296	25282	8/12/2027																																
0.200	MP6297	25020	1/14/2027																																
0.300	MP6298	25150	5/6/2027																																
0.080 DGS	N/A	AG510701	4/17/2027																																
☒ Post Optical Bench Adjustment Stability Checks				0.080 MP5089 0.200 MP5090																															
Simulator	Serial #	Lot #	Expiration	Attachments																															
0.050	MP5088	202603A	3/3/2028	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																															
0.080	MP5089	202603B	3/3/2028																																
0.200	MP5090	202603D	3/4/2028																																
0.080 DGS	N/A	23725080A1	10/5/2027																																
Gauge ID #: <u>28427</u> Gauge: <u>1013</u> Instrument: <u>1015</u>																																			
Notes/Suggested Service: *Tech Correction: Corrected the 0.080 ARS lot number. WKP 5/21/2026 **Admin Correction: Corrected the 0.080 ARS lot number. WKP 5/22/2026				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Digitally signed by Taylor Gutschow Date: 2026.05.21 15:16:40 -04'00'																															
				Digitally signed by Kaitlyn Spearin Date: 2026.05.22 13:28:42 -04'00'																															
Tech Review				Admin Review																															

0.08 DAS Stability #1

Tightened dry gas due to it being loose
and repeated.

WKP 5/8/2026

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005056
05/08/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:47
Control Test	0.091	13:47
Air Blank	0.000	13:48
Control Test	0.082	13:48
Air Blank	0.000	13:49
Control Test	0.072	13:49
Air Blank	0.000	13:49
Control Test Stats		
Average	0.0817	
Std Dev	0.0095	
Rel Std Dev(%)	11.6380	

WKP

Operator's Signature

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053 CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 15/08/2026 Software: 8100.27 SN 80-005058 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:33	Control Test	0.049	13:33	Air Blank	0.000	13:34	Control Test	0.050	13:35	Air Blank	0.000	13:35	Control Test	0.049	13:36	Air Blank	0.000	13:36	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		0.077 to 0.083 CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 05/08/2026 Software: 8100.27 SN 80-005058 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:37	Control Test	0.077	13:38	Air Blank	0.000	13:39	Control Test	0.078	13:39	Air Blank	0.000	13:40	Control Test	0.078	13:40	Air Blank	0.000	13:41	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		0.194 to 0.206 CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 05/08/2026 Software: 8100.27 SN 80-005058 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:42	Control Test	0.197	13:43	Air Blank	0.000	13:43	Control Test	0.196	13:44	Air Blank	0.000	13:44	Control Test	0.196	13:45	Air Blank	0.000	13:46	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		0.077 to 0.083 CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 15/08/2026 Software: 8100.27 SN 80-005058 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.082</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test</td><td>0.084</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0823</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.8553</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	13:51	Control Test	0.082	13:51	Air Blank	0.000	13:52	Control Test	0.081	13:52	Air Blank	0.000	13:53	Control Test	0.084	13:53	Air Blank	0.000	13:53	Control Test Stats			Average	0.0823		Std Dev	0.0015		Rel Std Dev(%)	1.8553	
Air Blank	0.000	13:33																																																																																																																																					
Control Test	0.049	13:33																																																																																																																																					
Air Blank	0.000	13:34																																																																																																																																					
Control Test	0.050	13:35																																																																																																																																					
Air Blank	0.000	13:35																																																																																																																																					
Control Test	0.049	13:36																																																																																																																																					
Air Blank	0.000	13:36																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0493																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1703																																																																																																																																						
Air Blank	0.000	13:37																																																																																																																																					
Control Test	0.077	13:38																																																																																																																																					
Air Blank	0.000	13:39																																																																																																																																					
Control Test	0.078	13:39																																																																																																																																					
Air Blank	0.000	13:40																																																																																																																																					
Control Test	0.078	13:40																																																																																																																																					
Air Blank	0.000	13:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0777																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7434																																																																																																																																						
Air Blank	0.000	13:42																																																																																																																																					
Control Test	0.197	13:43																																																																																																																																					
Air Blank	0.000	13:43																																																																																																																																					
Control Test	0.196	13:44																																																																																																																																					
Air Blank	0.000	13:44																																																																																																																																					
Control Test	0.196	13:45																																																																																																																																					
Air Blank	0.000	13:46																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1963																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2941																																																																																																																																						
Air Blank	0.000	13:51																																																																																																																																					
Control Test	0.082	13:51																																																																																																																																					
Air Blank	0.000	13:52																																																																																																																																					
Control Test	0.081	13:52																																																																																																																																					
Air Blank	0.000	13:53																																																																																																																																					
Control Test	0.084	13:53																																																																																																																																					
Air Blank	0.000	13:53																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0823																																																																																																																																						
Std Dev	0.0015																																																																																																																																						
Rel Std Dev(%)	1.8553																																																																																																																																						

CRESTVIEW PD
Intoxilizer - Alcohol Analyzer
Model 8000
05/15/2026
Auto Calibration
Max Power Res Value = 36
Auto Range Res Value = 19

SN 80-005058
10:19:10

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.063
Std Dev = 0.02 Rel Std Dev = 35.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.824
Std Dev = 0.03 Rel Std Dev = 4.09
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.937
Std Dev = 0.13 Rel Std Dev = 1.64
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.712
Std Dev = 0.03 Rel Std Dev = 0.85
Sol Val = 1.4266 mg/l or 0.300 g/210L
% Abs = 5.433
Std Dev = 0.03 Rel Std Dev = 0.57
Zero Order Coef = -162.75
First Order Coef = 2487.27
Second Order Coef = 31.85
Standard Deviation = 10.534757

Sol Value = 0.100 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12605, Sum Io = 13446

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0420 (-0.0033)

Sample #2 = 0.0570 (-0.0053)

Sample #3 = 0.0440 (-0.0130)

Sample #4 = 0.0880 (-0.0280)

Avg % Abs = 0.0630 (-0.0120)

STD DEV = 0.0226 (-0.0165)

REL STD DEV = 35.881 (137.699)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.063
Std Dev = 0.02 Rel Std Dev = 35.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.824
Std Dev = 0.03 Rel Std Dev = 4.09
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.937
Std Dev = 0.13 Rel Std Dev = 1.64
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.712
Std Dev = 0.03 Rel Std Dev = 0.85
Sol Val = 1.4266 mg/l or 0.300 g/210L
% Abs = 5.433
Std Dev = 0.03 Rel Std Dev = 0.57
Zero Order Coef = -162.75
First Order Coef = 2487.27
Second Order Coef = 31.85
Standard Deviation = 10.534757

Sol Value = 0.100 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12605, Sum Io = 13446

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0420 (-0.0033)

Sample #2 = 0.0570 (-0.0053)

Sample #3 = 0.0440 (-0.0130)

Sample #4 = 0.0880 (-0.0280)

Avg % Abs = 0.0630 (-0.0120)

STD DEV = 0.0226 (-0.0165)

REL STD DEV = 35.881 (137.699)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.063
Std Dev = 0.02 Rel Std Dev = 35.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.824
Std Dev = 0.03 Rel Std Dev = 4.09
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.937
Std Dev = 0.13 Rel Std Dev = 1.64
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.712
Std Dev = 0.03 Rel Std Dev = 0.85
Sol Val = 1.4266 mg/l or 0.300 g/210L
% Abs = 5.433
Std Dev = 0.03 Rel Std Dev = 0.57
Zero Order Coef = -162.75
First Order Coef = 2487.27
Second Order Coef = 31.85
Standard Deviation = 10.534757

Sol Value = 0.100 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12605, Sum Io = 13446

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0420 (-0.0033)

Sample #2 = 0.0570 (-0.0053)

Sample #3 = 0.0440 (-0.0130)

Sample #4 = 0.0880 (-0.0280)

Avg % Abs = 0.0630 (-0.0120)

STD DEV = 0.0226 (-0.0165)

REL STD DEV = 35.881 (137.699)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

***** CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1310 (-0.0120)

Sample #2 = 0.1220 (-0.0161)

Sample #3 = 0.1500 (-0.0311)

Sample #4 = 0.1470 (-0.0141)

Avg % Abs = 0.1397 (-0.0203)

STD DEV = 0.0154 (-0.0093)

REL STD DEV = 11.007 (45.696)

***** CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8100 (-0.0177)

Sample #2 = 0.8100 (-0.0047)

Sample #3 = 0.8620 (-0.0131)

Sample #4 = 0.7990 (-0.0180)

Avg % Abs = 0.8237 (-0.0013)

STD DEV = 0.0337 (-0.0159)

REL STD DEV = 4.085 (4784.355)

Optical Bench
calibration
adjustment
5/15/20 WKP

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																												
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																												
✓	✓	✓	✓																																																																																																												
CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 05/15/2026 Software: 8100.27 SN 80-005058 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:20	Control Test	0.050	11:21	Air Blank	0.000	11:21	Control Test	0.050	11:22	Air Blank	0.000	11:22	Control Test	0.050	11:23	Air Blank	0.000	11:24	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 05/15/2026 Software: 8100.27 SN 80-005058 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:15	Control Test	0.198	11:15	Air Blank	0.000	11:16	Control Test	0.198	11:17	Air Blank	0.000	11:17	Control Test	0.197	11:18	Air Blank	0.000	11:18	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		CRESTVIEW PD Intoxilyzer - Alcohol Analyzer Model 8000 05/15/2026 Software: 8100.27 SN 80-005058 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:31	Control Test	0.080	11:31	Air Blank	0.000	11:31	Control Test	0.080	11:32	Air Blank	0.000	11:32	Control Test	0.080	11:33	Air Blank	0.000	11:33	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		0.077 to 0.083
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:20																																																																																																													
Control Test	0.050	11:21																																																																																																													
Air Blank	0.000	11:21																																																																																																													
Control Test	0.050	11:22																																																																																																													
Air Blank	0.000	11:22																																																																																																													
Control Test	0.050	11:23																																																																																																													
Air Blank	0.000	11:24																																																																																																													
Control Test Stats																																																																																																															
Average	0.0500																																																																																																														
Std Dev	0.0000																																																																																																														
Rel Std Dev(%)	0.0000																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:15																																																																																																													
Control Test	0.198	11:15																																																																																																													
Air Blank	0.000	11:16																																																																																																													
Control Test	0.198	11:17																																																																																																													
Air Blank	0.000	11:17																																																																																																													
Control Test	0.197	11:18																																																																																																													
Air Blank	0.000	11:18																																																																																																													
Control Test Stats																																																																																																															
Average	0.1977																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2921																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:31																																																																																																													
Control Test	0.080	11:31																																																																																																													
Air Blank	0.000	11:31																																																																																																													
Control Test	0.080	11:32																																																																																																													
Air Blank	0.000	11:32																																																																																																													
Control Test	0.080	11:33																																																																																																													
Air Blank	0.000	11:33																																																																																																													
Control Test Stats																																																																																																															
Average	0.0800																																																																																																														
Std Dev	0.0000																																																																																																														
Rel Std Dev(%)	0.0000																																																																																																														
Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>																																																																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CRESTVIEW PD

Time of Inspection: 13:54

Date of Inspection: 05/15/2026

Serial Number: 80-005058

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.077	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.079

Standard Deviations	0.0000	0.0003	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

05/15/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005058, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005058</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CRESTVIEW PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/15/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:54</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.05.19 08:53:21
-04'00'

05/15/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality