

Alcohol Testing Program - Instrument Processing Sheet

Agency: WALTON COUNTY SO

Instrument Serial Number: 80-001311

Date In: 7/23/2026

DI Completion Date: 7/30/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>7/23/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>7/23/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>157</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1016</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202603D</td> <td>☐ Other: _____</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6291	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6292	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP6293	202603D	☐ Other: _____	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP6291	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP6292	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP6293	202603D	☐ Other: _____																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>WKP</u>																																	
Barometric Pressure Gauge: <u>1008</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>28427</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1009</u> Instrument: <u>1010</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																															
0.000	MP6294	N/A	N/A																																
0.040	MP6295	25090	3/11/2027																																
0.100	MP6296	24390	10/29/2026																																
0.200	MP6297	25020	1/14/2027																																
0.300	MP6298	25150	5/6/2027																																
0.080 DGS	N/A	AG526603	9/23/2027																																
☒ Post Optical Bench Adjustment Stability Checks				0.080 MP5089																															
Simulator	Serial #	Lot #	Expiration	0.200 MP5090																															
0.050	MP6291	202603A	03/03/2028																																
0.080	MP6292	202603B	03/03/2028																																
0.200	MP6293	202603D	03/04/2028																																
0.080 DGS	N/A	23825080A1	10/5/2027																																
Gauge ID #: <u>34419</u> Gauge: <u>1008</u> Instrument: <u>1008</u>				Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Adjustment ☒ Calibration Certificate ☐ Form 40 ☒ Optical Bench Adjustment ☐ Other: _____																															
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Digitally signed by Taylor Gutschow Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.02 20:46:52 -04'00' Date: 2026.08.03 13:01:24 -04'00'																															
Tech Review				Admin Review																															

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001311 07/23/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Wentz</i>	Test	g/210L	Time	Air Blank	0.000	11:53	Control Test	0.050	11:54	Air Blank	0.000	11:54	Control Test	0.049	11:55	Air Blank	0.000	11:55	Control Test	0.049	11:56	Air Blank	0.000	11:57	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001311 07/23/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Wentz</i>	Test	g/210L	Time	Air Blank	0.000	12:06	Control Test	0.078	12:07	Air Blank	0.000	12:08	Control Test	0.077	12:08	Air Blank	0.000	12:09	Control Test	0.078	12:10	Air Blank	0.000	12:10	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001311 07/23/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1960</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Wentz</i>	Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.196	12:26	Air Blank	0.000	12:27	Control Test	0.196	12:27	Air Blank	0.000	12:28	Control Test	0.196	12:29	Air Blank	0.000	12:29	Control Test Stats			Average	0.1960		Std Dev	0.0000		Rel Std Dev(%)	0.0000		WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001311 07/23/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Wentz</i>	Test	g/210L	Time	Air Blank	0.000	10:01	Control Test	0.081	10:02	Air Blank	0.000	10:02	Control Test	0.081	10:02	Air Blank	0.000	10:03	Control Test	0.081	10:03	Air Blank	0.000	10:04	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.050	11:54																																																																																																																																																	
Air Blank	0.000	11:54																																																																																																																																																	
Control Test	0.049	11:55																																																																																																																																																	
Air Blank	0.000	11:55																																																																																																																																																	
Control Test	0.049	11:56																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.078	12:07																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test	0.077	12:08																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.078	12:10																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:25																																																																																																																																																	
Control Test	0.196	12:26																																																																																																																																																	
Air Blank	0.000	12:27																																																																																																																																																	
Control Test	0.196	12:27																																																																																																																																																	
Air Blank	0.000	12:28																																																																																																																																																	
Control Test	0.196	12:29																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1960																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test	0.081	10:02																																																																																																																																																	
Air Blank	0.000	10:02																																																																																																																																																	
Control Test	0.081	10:02																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.081	10:03																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311
07/28/2026 09:00:41

SMA

Auto Calibration

pg 1 of 2

<<<< 3um >>>>

<<<< 9um >>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0580	(-0.0110)	0.1100	(-0.0110)
Sample #2	0.1090	(-0.0460)	0.1250	(-0.0260)
Sample #3	0.1240	(-0.0460)	0.1450	(-0.0330)
Sample #4	0.1210	(-0.0460)	0.1450	(-0.0380)
Avg % Abs	0.1180	(-0.0460)	0.1383	(-0.0323)
STD DEV	0.0079	(0.0000)	0.0115	(0.0060)
REL STD DEV	6.726	(0.000)	8.347	(18.642)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.7880	(-0.0060)	1.5290	(-0.0170)
Sample #2	0.8160	(-0.0050)	1.5450	(-0.0170)
Sample #3	0.8230	(-0.0040)	1.5450	(-0.0040)
Sample #4	0.8060	(0.0210)	1.5200	(0.0130)
Avg % Abs	0.8150	(0.0040)	1.5367	(-0.0027)
STD DEV	0.0085	(0.0147)	0.0144	(0.0150)
REL STD DEV	1.048	(368.273)	0.939	(564.164)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.8630	(-0.0100)	3.5990	(-0.0200)
Sample #2	1.8520	(0.0090)	3.5750	(0.0190)
Sample #3	1.8280	(0.0170)	3.6020	(0.0000)
Sample #4	1.8410	(0.0040)	3.5820	(0.0000)
Avg % Abs	1.8403	(0.0100)	3.5863	(0.0063)
STD DEV	0.0120	(0.0066)	0.0140	(0.0110)
REL STD DEV	0.653	(65.574)	0.391	(173.205)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.5390	(-0.0200)	6.8410	(-0.0060)
Sample #2	3.5140	(-0.0040)	6.8220	(0.0180)
Sample #3	3.5370	(-0.0020)	6.8110	(0.0180)
Sample #4	3.5250	(0.0160)	6.8490	(0.0190)
Avg % Abs	3.5253	(0.0033)	6.8273	(0.0183)
STD DEV	0.0115	(0.0110)	0.0196	(0.0006)
REL STD DEV	0.326	(330.454)	0.286	(3.149)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.1570	(-0.0210)	9.9280	(-0.0130)
Sample #2	5.1990	(-0.0240)	9.9210	(0.0150)
Sample #3	5.1650	(-0.0060)	9.9010	(0.0350)
Sample #4	5.1760	(-0.0040)	9.9170	(0.0230)
Avg % Abs	5.1800	(-0.0113)	9.9130	(0.0243)
STD DEV	0.0173	(0.0110)	0.0106	(0.0101)
REL STD DEV	0.335	(97.192)	0.107	(41.369)

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001311 ✓
07/28/2026 09:00:41

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-327.61			-183.38	
First Order Coef	2734.24			1336.07	
Second Order Coef	16.82			12.47	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	-0.000	0.0001	0.000	0.000	-0.0000
0.040	0.040	-0.0002	0.040	0.040	0.0001
0.100	0.100	0.0000	0.100	0.100	-0.0001
0.200	0.200	0.0001	0.200	0.200	0.0001
0.300	0.300	-0.0000	0.300	0.300	-0.0000
-----			-----		



<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	2972.00		3384.00
Sample #2	2941.00		3363.00
Sample #3	2919.00		3364.00
Sample #4	2903.00		3380.00
Avg	2921.0000		3369.0000
STD DEV	19.0788		9.5394
REL STD DEV	0.653		0.283
H2O adjust (mg/l*10k)	888		440

Barometric Pressure = 1008

*****CALIBRATION SUCCESSFUL***** ✓ SUT

sub

Post Optical Bench Adjustment Stability Checks 80-001311

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
☒ 0.047 to 0.053 g/210L	☒ 0.077 to 0.083 g/210L	☒ 0.194 to 0.206 g/210L	☒ 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001311	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001311	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001311	WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001311																																																																																																																																																
<table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr><tr><td>Control Test</td><td>0.050</td><td>10:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:55	Control Test	0.049	10:56	Air Blank	0.000	10:56	Control Test	0.050	10:57	Air Blank	0.000	10:58	Control Test	0.049	10:58	Air Blank	0.000	10:59	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:39	Control Test	0.079	10:40	Air Blank	0.000	10:40	Control Test	0.079	10:41	Air Blank	0.000	10:42	Control Test	0.079	10:42	Air Blank	0.000	10:43	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test</td><td>0.199</td><td>10:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test</td><td>0.198</td><td>10:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1983</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:50	Control Test	0.198	10:51	Air Blank	0.000	10:52	Control Test	0.199	10:52	Air Blank	0.000	10:53	Control Test	0.198	10:54	Air Blank	0.000	10:54	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		<table border="1"><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:48</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	10:45	Control Test	0.081	10:46	Air Blank	0.000	10:46	Control Test	0.080	10:47	Air Blank	0.000	10:47	Control Test	0.080	10:47	Air Blank	0.000	10:48	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.049	10:56																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.050	10:57																																																																																																																																																	
Air Blank	0.000	10:58																																																																																																																																																	
Control Test	0.049	10:58																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.079	10:40																																																																																																																																																	
Air Blank	0.000	10:40																																																																																																																																																	
Control Test	0.079	10:41																																																																																																																																																	
Air Blank	0.000	10:42																																																																																																																																																	
Control Test	0.079	10:42																																																																																																																																																	
Air Blank	0.000	10:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:50																																																																																																																																																	
Control Test	0.198	10:51																																																																																																																																																	
Air Blank	0.000	10:52																																																																																																																																																	
Control Test	0.199	10:52																																																																																																																																																	
Air Blank	0.000	10:53																																																																																																																																																	
Control Test	0.198	10:54																																																																																																																																																	
Air Blank	0.000	10:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:45																																																																																																																																																	
Control Test	0.081	10:46																																																																																																																																																	
Air Blank	0.000	10:46																																																																																																																																																	
Control Test	0.080	10:47																																																																																																																																																	
Air Blank	0.000	10:47																																																																																																																																																	
Control Test	0.080	10:47																																																																																																																																																	
Air Blank	0.000	10:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
 Operator Signature	 Operator Signature	 Operator Signature	 Operator Signature																																																																																																																																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 11:15

Date of Inspection: 07/30/2026

Serial Number: 80-001311
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.078	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.051	0.079	0.199	0.081
0.000	0.051	0.078	0.198	0.081
0.000	0.051	0.079	0.199	0.080

Standard Deviations	0.0006	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson WEN-CHI K PIERSON
Signature and Printed Name

07/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001311 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001311</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>WALTON COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/30/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:15</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.07.31 09:02:32 -04'00'

07/30/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality