

Alcohol Testing Program - Instrument Processing Sheet

Agency: ALACHUA COUNTY SO Instrument Serial Number: 80-000737
 Date In: 3/5/2026 DI Completion Date: 3/6/2026 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>3/5/2026</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>WKP</u> Date: <u>3/5/2026</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.511</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>34418</u> Gauge: <u>1018</u> Instrument: <u>1017</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:
0.050	MP5088	202406K	<input type="checkbox"/> Battery Replacement	
		6/19/2026		
0.080	MP5089	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement	
		6/19/2026		
0.200	MP5090	202406N	<input type="checkbox"/> Tank Sensor Tare	
		6/20/2026		
0.080 DGS	N/A	28424080A3	<input type="checkbox"/> Breath Tube Replacement	
		11/5/2026		

Optical Bench Adjustment By: WKP **Department Inspection** By: WKP

Barometric Pressure Gauge: <u>1018</u> ID#: <u>28421</u>				Barometric Pressure ID#: <u>28427</u>	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1019</u>	Instrument: <u>1020</u>
0.000	MP6294	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/2027</u>
0.040	MP6295	25090	3/11/2027	Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>9/22/2027</u>
0.100	MP6296	25282	8/12/2027	Simulator	Serial Number
0.200	MP6297	25020	1/14/2027	0.000	MP6289
0.300	MP6298	25150	5/6/2027	Interferent	MP6290
0.080 DGS	N/A	AG510701	4/17/2027	0.050	MP6291

Post Optical Bench Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050	MP5088	202406K	6/19/2026
0.080	MP5089	202406L	6/19/2026
0.200	MP5090	202406N	6/20/2026
0.080 DGS	N/A	28424080A3	11/5/2026

Gauge ID #: <u>34418</u> Gauge: <u>1017</u> Instrument: <u>1018</u>	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Optical Bench Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other:
------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------





Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Shayla Platt Tech Review	Shayla Platt Admin Review

Digitally signed by Shayla Platt Date: 2026.03.07 14:18:51 -05'00' Digitally signed by Shayla Platt Date: 2026.03.07 14:19:26 -05'00'

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✗																																																																																																																																																
<p>ALACHUA COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test</td><td>0.046</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.045</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.046</td><td>14:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0457</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2643</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	14:19	Control Test	0.046	14:20	Air Blank	0.000	14:21	Control Test	0.045	14:21	Air Blank	0.000	14:22	Control Test	0.046	14:23	Air Blank	0.000	14:23	Control Test Stats			Average	0.0457		Std Dev	0.0016		Rel Std Dev(%)	1.2643		<p>ALACHUA COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.076</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr> <tr><td>Control Test</td><td>0.076</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0763</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	14:24	Control Test	0.076	14:25	Air Blank	0.000	14:25	Control Test	0.077	14:26	Air Blank	0.000	14:26	Control Test	0.076	14:27	Air Blank	0.000	14:28	Control Test Stats			Average	0.0763		Std Dev	0.0016		Rel Std Dev(%)	0.7564		<p>ALACHUA COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5842</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	14:32	Control Test	0.197	14:32	Air Blank	0.000	14:33	Control Test	0.199	14:33	Air Blank	0.000	14:34	Control Test	0.197	14:35	Air Blank	0.000	14:35	Control Test Stats			Average	0.1977		Std Dev	0.0012		Rel Std Dev(%)	0.5842		<p>ALACHUA COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:37</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p>Operator's Signature _____</p>	Test	g/210L	Time	Air Blank	0.000	14:37	Control Test	0.080	14:37	Air Blank	0.000	14:38	Control Test	0.080	14:38	Air Blank	0.000	14:38	Control Test	0.081	14:39	Air Blank	0.000	14:39	Control Test Stats			Average	0.0803		Std Dev	0.0016		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:19																																																																																																																																																	
Control Test	0.046	14:20																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.045	14:21																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.046	14:23																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0457																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	1.2643																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test	0.076	14:25																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test	0.077	14:26																																																																																																																																																	
Air Blank	0.000	14:26																																																																																																																																																	
Control Test	0.076	14:27																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:32																																																																																																																																																	
Control Test	0.197	14:32																																																																																																																																																	
Air Blank	0.000	14:33																																																																																																																																																	
Control Test	0.199	14:33																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																	
Control Test	0.197	14:35																																																																																																																																																	
Air Blank	0.000	14:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5842																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:37																																																																																																																																																	
Control Test	0.080	14:37																																																																																																																																																	
Air Blank	0.000	14:38																																																																																																																																																	
Control Test	0.080	14:38																																																																																																																																																	
Air Blank	0.000	14:38																																																																																																																																																	
Control Test	0.081	14:39																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet																																																																																																																																																
<p>ALACHUA COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:49</td></tr> <tr><td>Control Test</td><td>0.150</td><td>15:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:51</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	15:49	Control Test	0.150	15:50	Air Blank	0.000	15:50	Control Test	0.049	15:51	Air Blank	0.000	15:51	Control Test	0.050	15:52	Air Blank	0.000	15:52	Control Test Stats			Average	0.1497		Std Dev	0.0016		Rel. Std Dev(%)	1.1625		<p>ALACHUA COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:55</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:56</td></tr> <tr><td>Control Test</td><td>0.081</td><td>15:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0011</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2500</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	15:54	Control Test	0.079	15:55	Air Blank	0.000	15:55	Control Test	0.080	15:56	Air Blank	0.000	15:56	Control Test	0.081	15:57	Air Blank	0.000	15:57	Control Test Stats			Average	0.0800		Std Dev	0.0011		Rel. Std Dev(%)	1.2500		<p>ALACHUA COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:59</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:00</td></tr> <tr><td>Control Test</td><td>0.199</td><td>16:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:01</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2916</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	15:59	Control Test	0.199	15:59	Air Blank	0.000	16:00	Control Test	0.199	16:00	Air Blank	0.000	16:01	Control Test	0.198	16:02	Air Blank	0.000	16:02	Control Test Stats			Average	0.1997		Std Dev	0.0016		Rel. Std Dev(%)	0.2916		<p>ALACHUA COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000737 03/05/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:05</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	16:03	Control Test	0.078	16:04	Air Blank	0.000	16:04	Control Test	0.079	16:05	Air Blank	0.000	16:05	Control Test	0.079	16:05	Air Blank	0.000	16:06	Control Test Stats			Average	0.0787		Std Dev	0.0016		Rel. Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:49																																																																																																																																																	
Control Test	0.150	15:50																																																																																																																																																	
Air Blank	0.000	15:50																																																																																																																																																	
Control Test	0.049	15:51																																																																																																																																																	
Air Blank	0.000	15:51																																																																																																																																																	
Control Test	0.050	15:52																																																																																																																																																	
Air Blank	0.000	15:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1497																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel. Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:54																																																																																																																																																	
Control Test	0.079	15:55																																																																																																																																																	
Air Blank	0.000	15:55																																																																																																																																																	
Control Test	0.080	15:56																																																																																																																																																	
Air Blank	0.000	15:56																																																																																																																																																	
Control Test	0.081	15:57																																																																																																																																																	
Air Blank	0.000	15:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0011																																																																																																																																																		
Rel. Std Dev(%)	1.2500																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:59																																																																																																																																																	
Control Test	0.199	15:59																																																																																																																																																	
Air Blank	0.000	16:00																																																																																																																																																	
Control Test	0.199	16:00																																																																																																																																																	
Air Blank	0.000	16:01																																																																																																																																																	
Control Test	0.198	16:02																																																																																																																																																	
Air Blank	0.000	16:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel. Std Dev(%)	0.2916																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:03																																																																																																																																																	
Control Test	0.078	16:04																																																																																																																																																	
Air Blank	0.000	16:04																																																																																																																																																	
Control Test	0.079	16:05																																																																																																																																																	
Air Blank	0.000	16:05																																																																																																																																																	
Control Test	0.079	16:05																																																																																																																																																	
Air Blank	0.000	16:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA COUNTY SO
Time of Inspection: 11:39

Date of Inspection: 03/06/2026

Serial Number: 80-000737
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.049	0.080	0.000 / 0.200	0.079
0.000	0.049	0.081	0.000 / 0.200	0.079
0.000	0.049	0.081	0.000 / 0.201	0.078
0.000	0.048	0.081	0.000 / 0.201	0.079
0.000	0.049	0.081	0.000 / 0.201	0.079
0.000	0.049	0.082	0.000 / 0.200	0.078
0.000	0.049	0.081	0.000 / 0.201	0.079
0.000	0.049	0.082	0.000 / 0.201	0.079
0.000	0.050	0.082	0.000 / 0.201	0.079
0.000	0.049	0.083	0.000 / 0.201	0.079

Standard Deviations	0.0004	0.0008	0.0000 / 0.0004	0.0004
---------------------	--------	--------	-----------------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

0.20g/210L wet bath simulator was not properly connected to instrument. corrected connection and repeated. WKP 3/6/2026

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wenchi K Pierson WEN-CHI K PIERSON
Signature and Printed Name

03/06/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000737, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000737</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>ALACHUA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/06/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:39</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Wen-Chi Pierson
Date: 2026.03.06 14:36:56 -05'00'

Wen-Chi Pierson

03/06/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality