

Alcohol Testing Program - Instrument Processing Sheet

Agency: Palm Beach County Sheriff's Office Instrument Serial Number: 80-001741  
 Date In: 2/9/2026 DI Completion Date: 2/9/2026  Ship  P/U  H/D  CMI  EE

<b>Intake By:</b> <u>TDG</u> <b>Date:</b> <u>2/9/2026</u>	<b>Quality Checks By:</b> <u>TDG</u> <b>Date:</b> <u>2/9/2026</u>	<b>Flow Adjustment By:</b> _____																																
<input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: No box.	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>217</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1026</u> Instrument: <u>1021</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																
	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	<b>Maintenance By:</b> _____ <b>Date:</b> _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____																	
Simulator	Serial #	Lot#/Exp																																
0.050	MP6286	202406K 6/19/2026																																
0.080	MP6287	202406L 6/19/2026																																
0.200	MP6288	202406N 6/20/2026																																
0.080 DGS	N/A	AG526603 9/23/2027																																
<b>Optical Bench Adjustment</b> By: _____	<b>Department Inspection</b> By: <u>TDG</u>																																	
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1026</u> Instrument: <u>1022</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>					
Simulator	Serial #	Lot #	Expiration																															
0.000		N/A	N/A																															
0.040																																		
0.100																																		
0.200																																		
0.300																																		
0.080 DGS	N/A																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	
Simulator	Serial #	Lot #	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
Simulator	Serial Number																																	
0.000	MP6284																																	
Interferent	MP6285																																	
0.050	MP6286																																	
0.080	MP6287																																	
0.200	MP6288																																	
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other: _____												
Simulator	Serial #	Lot #	Expiration																															
0.050																																		
0.080																																		
0.200																																		
0.080 DGS	N/A																																	
Gauge ID #: _____ Gauge: _____ Instrument: _____																																		
Notes/Suggested Service:	<input checked="" type="checkbox"/> <b>Instrument Complies with Chapter 11D-8, FAC</b> <input type="checkbox"/> <b>Instrument Does Not Comply with Chapter 11D-8, FAC</b> <input checked="" type="checkbox"/> <b>Return to/Place into Evidentiary Use</b> <input type="checkbox"/> <b>Remain Out of Evidentiary Use</b> <input checked="" type="checkbox"/> <b>Conduct an Agency Inspection Before Evidentiary Use</b>																																	
	LeAndra Higginbotham <small>Digitally signed by LeAndra Higginbotham Date: 2026.02.09 14:16:39 -05'00'</small>	Wen-Chi Pierson <small>Digitally signed by Wen-Chi Pierson Date: 2026.02.09 15:01:52 -05'00'</small>																																
	<b>Tech Review</b>	<b>Admin Review</b>																																

# Stability Checks

<b>0.05g/210L</b> 0.047 to 0.053 <input checked="" type="checkbox"/>	<b>0.08g/210L</b> 0.077 to 0.083 <input checked="" type="checkbox"/>	<b>0.20g/210L</b> 0.194 to 0.206 <input checked="" type="checkbox"/>	<b>DGS 0.08g/210L</b> 0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet <input checked="" type="checkbox"/>																																																																																																																																																
<p>PALM BEACH COUNTY SO                      Intoxilyzer - Alcohol Analyzer                      Model 8000 SN 80-001741                      02/09/2026                      Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test State</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	11:03	Control Test	0.050	11:03	Air Blank	0.000	11:04	Control Test	0.049	11:05	Air Blank	0.000	11:05	Control Test	0.050	11:06	Air Blank	0.000	11:06	Control Test State			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<p>PALM BEACH COUNTY SO                      Intoxilyzer - Alcohol Analyzer                      Model 8000 SN 80-001741                      02/09/2026                      Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test State</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	11:09	Control Test	0.080	11:10	Air Blank	0.000	11:11	Control Test	0.080	11:11	Air Blank	0.000	11:12	Control Test	0.080	11:13	Air Blank	0.000	11:13	Control Test State			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PALM BEACH COUNTY SO                      Intoxilyzer - Alcohol Analyzer                      Model 8000 SN 80-001741                      02/09/2026                      Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test State</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	11:15	Control Test	0.199	11:15	Air Blank	0.000	11:16	Control Test	0.199	11:17	Air Blank	0.000	11:17	Control Test	0.199	11:18	Air Blank	0.000	11:18	Control Test State			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PALM BEACH COUNTY SO                      Intoxilyzer - Alcohol Analyzer                      Model 8000 SN 80-001741                      02/09/2026                      Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test State</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>MC</i></p>	Test	g/210L	Time	Air Blank	0.000	10:23	Control Test	0.081	10:23	Air Blank	0.000	10:24	Control Test	0.081	10:24	Air Blank	0.000	10:25	Control Test	0.082	10:25	Air Blank	0.000	10:25	Control Test State			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.050	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.049	11:05																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test	0.050	11:06																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test State																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:09																																																																																																																																																	
Control Test	0.080	11:10																																																																																																																																																	
Air Blank	0.000	11:11																																																																																																																																																	
Control Test	0.080	11:11																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test	0.080	11:13																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test State																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.199	11:15																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.199	11:17																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.199	11:18																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test State																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:23																																																																																																																																																	
Control Test	0.081	10:23																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test	0.081	10:24																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test	0.082	10:25																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test State																																																																																																																																																			
Average	0.0813																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7099																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH COUNTY SO  
Time of Inspection: 13:59

Date of Inspection: 02/09/2026

Serial Number: 80-001741  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.080	0.198	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.198	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.080	0.198	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.080	0.199	0.081


Standard Deviations	0.0004	0.0000	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/09/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001741 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001741</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PALM BEACH COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/09/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:59</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/09/2026

Date

Taylor  
Gutschow

Digitally signed by Taylor  
Gutschow  
Date: 2026.02.09 14:04:02  
-05'00'

TAYLOR D GUTSCHOW,  
Department Inspector