

Alcohol Testing Program - Instrument Processing Sheet

Agency: GULF BREEZE PD

Instrument Serial Number: 80-004771

Date In: 5/18/2026

DI Completion Date: 6/1/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>5/18/2026</u>		Quality Checks By: <u>IAS</u> Date: <u>5/22/2026</u>		Flow Adjustment By: <u>SLH</u>																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Partial speaker foam present.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>157</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP105</u> 32 mm .167 (.139-.169) 36 mm .187 (.156-.190) 53 mm .253 (.228-.278) 103 mm .550 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: <u>ATP102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>158</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP105</u> 32 mm 0.156 (.139-.169) 36 mm 0.167 (.156-.190) 53 mm 0.234 (.228-.278) 103 mm 0.535 (.447-.547)																
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>20603A 202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>20603B 202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>20603D 202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	20603A 202603A 3/3/2028	0.080	MP5089	20603B 202603B 3/3/2028	0.200	MP5090	20603D 202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	20603A 202603A 3/3/2028																		
0.080	MP5089	20603B 202603B 3/3/2028																		
0.200	MP5090	20603D 202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28662</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1012</u> Instrument: <u>1011</u>																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																
0.100				Simulator	Serial Number															
0.200				0.000	MP5086															
0.300				Interferent	MP5087															
0.080 DGS	N/A			0.050	MP5088															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089															
Simulator	Serial #	Lot #	Expiration	0.200	MP5090															
0.050				Attachments																
0.080				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____ Gauge: _____ Instrument: _____																				
Notes/Suggested Service: Tech Review: Corrected lot numbers for 3 wet stabilities. IAS 6/4/2026				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Digitally signed by Taylor Gutschow Date: 2026.06.04 16:05:29 -04'00'																
				Digitally signed by Shayla Platt Date: 2026.06.05 07:49:01 -04'00'																
				Tech Review	Admin Review															

Sut

GULF BREEDER 40
Innovalyzer - Alcohol Analyzer
Model: 8001 SN: 80-004071
05/22/2020
Software: 8.00.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SORT(DIFF) = 6.707

2: Rate (Liters/min) = 15
SORT(DIFF) = 11.817

3: Rate (Liters/min) = 30
SORT(DIFF) = 22.781

Dependent Data Scale Factor = 100000 Liters

Independent Data Scale Factor = 256

Rounded Slope = 593

Rounded Intercept = -414616

Correlation = 0.99439

80-004771

no equipment or user
error so performed
adjustment.
Sut

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet
0.050 g/210L 0.047 to 0.053 g/210L GULF SPEEDS PD Investigator = Manual Manager Model 8000 05/22/2026 Software: 8111.27 Test g/210L Time ----- Air Blank 0.000 15:42 Control Test 0.050 15:42 Air Blank 0.000 15:43 Control Test 0.050 15:44 Air Blank 0.000 15:44 Control Test 0.049 15:45 Air Blank 0.000 15:45 Control Test Stats Average 0.0497 Std Dev 0.0016 Rel Std Dev(%) 1.1555	0.080 g/210L 0.077 to 0.083 g/210L GULF SPEEDS PD Investigator = Manual Manager Model 8000 05/22/2026 Software: 8111.27 Test g/210L Time ----- Air Blank 0.000 15:42 Control Test 0.079 15:42 Air Blank 0.000 15:43 Control Test 0.079 15:44 Air Blank 0.000 15:45 Control Test 0.079 15:45 Air Blank 0.000 15:45 Control Test Stats Average 0.0791 Std Dev 0.0010 Rel Std Dev(%) 1.0000	0.200 g/210L 0.194 to 0.206 g/210L GULF SPEEDS PD Investigator = Manual Manager Model 8000 05/22/2026 Software: 8111.27 Test g/210L Time ----- Air Blank 0.000 15:49 Control Test 0.199 15:49 Air Blank 0.000 15:50 Control Test 0.199 15:50 Air Blank 0.000 15:51 Control Test 0.198 15:51 Air Blank 0.000 15:52 Control Test Stats Average 0.1990 Std Dev 0.0000 Rel Std Dev(%) 0.0010	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet GULF SPEEDS PD Investigator = Manual Manager Model 8000 05/22/2026 Software: 8111.27 Test g/210L Time ----- Air Blank 0.000 15:50 Control Test 0.080 15:50 Air Blank 0.000 15:51 Control Test 0.080 15:51 Air Blank 0.000 15:52 Control Test 0.081 15:52 Air Blank 0.000 15:53 Control Test Stats Average 0.0807 Std Dev 0.0005 Rel Std Dev(%) 0.7157
<i>Isabella Szymanski</i> Operator's Signature	<i>Isabella Szymanski</i> Operator's Signature	<i>Isabella Szymanski</i> Operator's Signature	<i>Isabella Szymanski</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 10:53

Date of Inspection: 06/01/2026

Serial Number: 80-004771
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A <i>2028</i> Exp: 03/03/ 2027 <i>WLP 6/11/27</i>	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson

WEN-CHI K PIERSON
Signature and Printed Name

06/01/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-004771, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-004771</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GULF BREEZE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/01/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:53</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson

Digitally signed by Wen-Chi Pierson
Date: 2026.06.01 14:42:06
+04'00'

06/01/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality