

Alcohol Testing Program - Instrument Processing Sheet

Agency: MADISON COUNTY SO Instrument Serial Number: 80-000779
 Date In: 1/8/2026 DI Completion Date: 01/08/2026 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>1/8/2026</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>WKP</u> Date: <u>01/08/2026</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>230</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.531</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1015</u> Instrument: <u>1015</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
--	--	--

Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:
0.050	MP5088	202406K	<input type="checkbox"/> Battery Replacement	
		06/19/2026		
0.080	MP5089	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement	
		06/19/2026		
0.200	MP5090	202406N	<input type="checkbox"/> Tank Sensor Tare	
		06/20/2026		
0.080 DGS	N/A	AG510701	<input type="checkbox"/> Breath Tube Replacement	
		04/17/2027		

Optical Bench Adjustment By: _____ **Department Inspection** By: WKP
 Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: 28662

Simulator	Serial #	Lot #	Expiration	Gauge: <u>1015</u> Instrument: <u>1014</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>
0.100				Simulator Serial Number
0.200				0.000 MP6289
0.300				Interferent MP6290
0.080 DGS	N/A			0.050 MP5088
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080 MP5089
0.050				0.200 MP5090

Simulator Expiration 0.050 0.080 0.200 0.080 DGS N/A Gauge ID #: _____ Gauge: _____ Instrument: _____	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other:
---	--

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Taylor Gutschow Date: 2026.01.27 12:25:50 -05'00' Digitally signed by Kaitlyn Spearin Date: 2026.01.28 12:51:21 -05'00'
--------------------------	---

Taylor Gutschow Tech Review
 Kaitlyn Spearin Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>MEDISON COUNTY SO Intoxilyzer - Alcolon Analyzer Model: 8000 SN 60-000779 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.045</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:56</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.049</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>0.000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.00	12:53	Control Test	0.045	12:54	Air Blank	0.00	12:54	Control Test	0.049	12:55	Air Blank	0.00	12:56	Control Test	0.049	12:56	Air Blank	0.00	12:57	Control Test Stats			Average	0.049		Std Dev	0.000		Rel. Std Dev (%)	0.000		<p>MEDISON COUNTY SO Intoxilyzer - Alcolon Analyzer Model: 8100 SN 80-000779 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>13:29</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.079</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>1.2658</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.00	13:29	Control Test	0.080	13:29	Air Blank	0.00	13:30	Control Test	0.079	13:31	Air Blank	0.00	13:31	Control Test	0.078	13:32	Air Blank	0.00	13:32	Control Test Stats			Average	0.079		Std Dev	0.000		Rel. Std Dev (%)	1.2658		<p>MEDISON COUNTY SO Intoxilyzer - Alcolon Analyzer Model: 8000 SN 60-000779 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.195</td><td>13:34</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.195</td><td>13:37</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>13:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.197</td><td></td></tr> <tr><td>Std Dev</td><td>0.006</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>0.2956</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.00	13:34	Control Test	0.195	13:34	Air Blank	0.00	13:35	Control Test	0.198	13:36	Air Blank	0.00	13:36	Control Test	0.195	13:37	Air Blank	0.00	13:37	Control Test Stats			Average	0.197		Std Dev	0.006		Rel. Std Dev (%)	0.2956		<p>MEDISON COUNTY SO Intoxilyzer - Alcolon Analyzer Model: 8000 SN 80-000779 01/08/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>12:23</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:25</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>12:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.080</td><td></td></tr> <tr><td>Std Dev</td><td>0.005</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>1.5015</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.00	12:23	Control Test	0.082	12:24	Air Blank	0.00	12:24	Control Test	0.080	12:24	Air Blank	0.00	12:25	Control Test	0.079	12:25	Air Blank	0.00	12:26	Control Test Stats			Average	0.080		Std Dev	0.005		Rel. Std Dev (%)	1.5015	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.00	12:53																																																																																																																																																	
Control Test	0.045	12:54																																																																																																																																																	
Air Blank	0.00	12:54																																																																																																																																																	
Control Test	0.049	12:55																																																																																																																																																	
Air Blank	0.00	12:56																																																																																																																																																	
Control Test	0.049	12:56																																																																																																																																																	
Air Blank	0.00	12:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.049																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Rel. Std Dev (%)	0.000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.00	13:29																																																																																																																																																	
Control Test	0.080	13:29																																																																																																																																																	
Air Blank	0.00	13:30																																																																																																																																																	
Control Test	0.079	13:31																																																																																																																																																	
Air Blank	0.00	13:31																																																																																																																																																	
Control Test	0.078	13:32																																																																																																																																																	
Air Blank	0.00	13:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.079																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Rel. Std Dev (%)	1.2658																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.00	13:34																																																																																																																																																	
Control Test	0.195	13:34																																																																																																																																																	
Air Blank	0.00	13:35																																																																																																																																																	
Control Test	0.198	13:36																																																																																																																																																	
Air Blank	0.00	13:36																																																																																																																																																	
Control Test	0.195	13:37																																																																																																																																																	
Air Blank	0.00	13:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.197																																																																																																																																																		
Std Dev	0.006																																																																																																																																																		
Rel. Std Dev (%)	0.2956																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.00	12:23																																																																																																																																																	
Control Test	0.082	12:24																																																																																																																																																	
Air Blank	0.00	12:24																																																																																																																																																	
Control Test	0.080	12:24																																																																																																																																																	
Air Blank	0.00	12:25																																																																																																																																																	
Control Test	0.079	12:25																																																																																																																																																	
Air Blank	0.00	12:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.080																																																																																																																																																		
Std Dev	0.005																																																																																																																																																		
Rel. Std Dev (%)	1.5015																																																																																																																																																		
<p><i>ment</i> Operator's Signature</p>	<p><i>ment</i> Operator's Signature</p>	<p><i>ment</i> Operator's Signature</p>	<p><i>ment</i> Operator's Signature</p>																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 15:08

Date of Inspection: 01/08/2026

Serial Number: 80-000779
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701- Exp: 04/17/2027
/ 0.000	0.049	0.078	0.195	0.080
/ 0.000	0.049	0.077	0.194	0.080
/ 0.000	0.049	0.077	0.195	0.080
/ 0.000	0.049	0.078	0.195	0.080
/ 0.000	0.049	0.077	0.195	0.080
/ 0.000	0.049	0.077	0.195	0.081
/ 0.000	0.049	0.077	0.195	0.081
/ 0.000	0.048	0.078	0.195	0.080
/ 0.000	0.049	0.077	0.195	0.080
/ 0.000	0.049	0.078	0.195	0.080

wkp 01/08/2026

Standard Deviations	0.0003	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

00: Ambient Fail. *Turned fan on near instrument to clear ambient air. Repeated. wkp 01/08/2026*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

wkp _____ WEN-CHI K PIERSON
Signature and Printed Name

01/08/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000779, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000779</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MADISON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/08/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:08</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.01.26 10:46:29 -05'00'

01/08/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality