

Alcohol Testing Program - Instrument Processing Sheet

Agency: FLAGLER COUNTY SO Instrument Serial Number: 80-001138  
 Date In: 1/21/2026 DI Completion Date: 1/27/2026  Ship  P/U  H/D  CMI  EE

<b>Intake By:</b> <u>WKP</u> <b>Date:</b> <u>1/21/2026</u>	<b>Quality Checks By:</b> <u>SLH</u> <b>Date:</b> <u>1/27/2026</u>	<b>Flow Adjustment By:</b> _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>194</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1029</u> Instrument: <u>1029</u>	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

<input checked="" type="checkbox"/> Stability Checks			<b>Maintenance By:</b> _____	<b>Date:</b> _____
Simulator	Serial #	Lot#/Exp		
0.050	MP5088	202406K 06/19/2026	<input type="checkbox"/> Battery Replacement	
0.080	MP5089	202406L 06/19/2026	<input type="checkbox"/> Dry Gas Regulator Replacement	
0.200	MP5090	202406N 06/20/2026	<input type="checkbox"/> Tank Sensor Tare	
0.080 DGS	N/A	AG510701 04/17/2027	<input type="checkbox"/> Breath Tube Replacement	
			<input type="checkbox"/> Other:	

**Optical Bench Adjustment** By: \_\_\_\_\_ **Department Inspection** By: SLH

Barometric Pressure Gauge: _____ ID#: _____				Barometric Pressure ID#: <u>34419</u>	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1025</u>	Instrument: <u>1024</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>
0.100				Simulator	Serial Number
0.200				0.000	MP5086
0.300				Interferent	MP5087
0.080 DGS	N/A			0.050	MP5088
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	MP5089
Simulator	Serial #	Lot #	Expiration	0.200	MP5090
0.050				<b>Attachments</b>	
0.080				<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
0.200				<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Adjustment
0.080 DGS	N/A			<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
Gauge ID #: _____				<input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Other:
Gauge: _____ Instrument: _____					

Notes/Suggested Service:	<input checked="" type="checkbox"/> <b>Instrument Complies with Chapter 11D-8, FAC</b> <input type="checkbox"/> <b>Instrument Does Not Comply with Chapter 11D-8, FAC</b> <input checked="" type="checkbox"/> <b>Return to/Place into Evidentiary Use</b> <input type="checkbox"/> <b>Remain Out of Evidentiary Use</b> <input checked="" type="checkbox"/> <b>Conduct an Agency Inspection Before Evidentiary Use</b>
	Digitally signed by Taylor Gutschow Date: 2026.01.29 13:24:58 -05'00'
	Digitally signed by Wen-Chi Pierson Date: 2026.01.29 15:07:57 -05'00'
<b>Tech Review</b>	<b>Admin Review</b>

# Stability Checks

1/27/2026  
80-001138  
Swt

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>FLAGLER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2026 Software: 8100.27</p> <p>SN 80-001138</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigler</i></p>	Test	g/210L	Time	Air Blank	0.000	12:50	Control Test	0.050	12:51	Air Blank	0.000	12:51	Control Test	0.050	12:52	Air Blank	0.000	12:53	Control Test	0.050	12:53	Air Blank	0.000	12:54	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FLAGLER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2026 Software: 8100.27</p> <p>SN 80-001138</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigler</i></p>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.080	13:01	Air Blank	0.000	13:02	Control Test	0.079	13:03	Air Blank	0.000	13:03	Control Test	0.081	13:04	Air Blank	0.000	13:04	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500		<p>FLAGLER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2026 Software: 8100.27</p> <p>SN 80-001138</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigler</i></p>	Test	g/210L	Time	Air Blank	0.000	13:06	Control Test	0.199	13:07	Air Blank	0.000	13:07	Control Test	0.199	13:08	Air Blank	0.000	13:09	Control Test	0.199	13:09	Air Blank	0.000	13:10	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FLAGLER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/27/2026 Software: 8100.27</p> <p>SN 80-001138</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>M. Sigler</i></p>	Test	g/210L	Time	Air Blank	0.000	13:12	Control Test	0.081	13:12	Air Blank	0.000	13:13	Control Test	0.081	13:13	Air Blank	0.000	13:14	Control Test	0.081	13:14	Air Blank	0.000	13:14	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test	0.050	12:51																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.050	12:52																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.050	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.080	13:01																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.079	13:03																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.081	13:04																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.199	13:07																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test	0.199	13:08																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.199	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.081	13:12																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.081	13:13																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test	0.081	13:14																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO  
Time of Inspection: 16:33

Date of Inspection: 01/27/2026

Serial Number: 80-001138  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.079	0.199	0.080
0.000	0.051	0.079	0.199	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.079	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.081
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.080


Standard Deviations	0.0005	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 LEANDRA HIGGINBOTHAM  
 Signature and Printed Name

01/27/2026  
 Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001138, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001138</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FLAGLER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/27/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:33</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Date

  
LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 January 2026  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality