

Alcohol Testing Program - Instrument Processing Sheet

Agency: OKEECHOBEE COUNTY SO

Instrument Serial Number: 80-001321

Date In: 4/30/2026

DI Completion Date: 5/4/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: WKP Date: 4/30/2026		Quality Checks By: IAS Date: 5/1/2026		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 220 ☒ Flow Verification (L/s) Flow Column #: ATP102 32 mm .152 (.139-.169) 36 mm .164 (.156-.190) 53 mm .250 (.228-.278) 103 mm .527 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 34419 Gauge: 1010 Instrument: 1010 ☒ Stability Checks <table><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr><tr><td>0.050</td><td>MP5088</td><td>202603A 3/3/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B 3/3/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D 3/4/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/2027</td></tr></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547) Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP5088	202603A 3/3/2028																		
0.080	MP5089	202603B 3/3/2028																		
0.200	MP5090	202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: KTS																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28427																		
Simulator	Serial #	Lot #	Expiration	Gauge: 1013	Instrument: 1015															
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: 2025-D	Exp: 9/25/2026 2027*															
0.040				Acetone Stock Solution Lot #: 2025-B	Exp: 9/22/2026 2027*															
0.100				Simulator	Serial Number															
0.200				0.000	MP6289															
0.300				Interferent	MP6290															
0.080 DGS	N/A			0.050	MP6291															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6292															
Simulator	Serial #	Lot #	Expiration	0.200	MP6293															
0.050				Attachments																
0.080				☒ Form 41	☐ Post-Stability Checks															
0.200				☒ Stability Checks	☐ Flow Adjustment															
0.080 DGS	N/A			☒ Calibration Certificate	☒ Form 40															
Gauge ID #: _____				☐ Optical Bench Adjustment	☐ Other:															
Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
Tech Review: Corrected dates for Mouth Alcohol and Acetone Solutions. KTS 5/18/26				LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.05.18 16:44:17 -04'00' Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.05.19 10:19:12 -04'00'																
Tech Review				Admin Review																

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO

Serial Number: 80-001321

Time of Inspection: 10:14

Date of Inspection: 05/01/2026

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

Bypassed AI to operate instrument. Compliance undetermined. IAS 5/5/26
Reprinted form. IAS 5/5/26 - Printer not accessible. IAS 5/5/26

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Isabella Smykowski

ISABELLA A SMYKOWSKI

Signature and Printed Name

05/01/2026

Date

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet																																																																																																																																																
OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001321 05/01/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:42</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:45</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Ashtella Smythawski</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:42	Control Test	0.048	15:43	Air Blank	0.000	15:43	Control Test	0.048	15:44	Air Blank	0.000	15:45	Control Test	0.048	15:45	Air Blank	0.000	15:46	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001321 05/01/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:34</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <i>Ashtella Smythawski</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:31	Control Test	0.078	15:32	Air Blank	0.000	15:32	Control Test	0.078	15:33	Air Blank	0.000	15:34	Control Test	0.077	15:34	Air Blank	0.000	15:35	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-00121 05/01/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test</td><td>0.203</td><td>15:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:38</td></tr> <tr><td>Control Test</td><td>0.202</td><td>15:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:39</td></tr> <tr><td>Control Test</td><td>0.201</td><td>15:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2020</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.4950</td><td></td></tr> </tbody> </table> <i>Ashtella Smythawski</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:37	Control Test	0.203	15:38	Air Blank	0.000	15:38	Control Test	0.202	15:39	Air Blank	0.000	15:39	Control Test	0.201	15:40	Air Blank	0.000	15:41	Control Test Stats			Average	0.2020		Std Dev	0.0010		Rel Std Dev(%)	0.4950		OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001321 05/01/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Ashtella Smythawski</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	15:25	Control Test	0.079	15:26	Air Blank	0.000	15:26	Control Test	0.079	15:26	Air Blank	0.000	15:27	Control Test	0.079	15:27	Air Blank	0.000	15:28	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:42																																																																																																																																																	
Control Test	0.048	15:43																																																																																																																																																	
Air Blank	0.000	15:43																																																																																																																																																	
Control Test	0.048	15:44																																																																																																																																																	
Air Blank	0.000	15:45																																																																																																																																																	
Control Test	0.048	15:45																																																																																																																																																	
Air Blank	0.000	15:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:31																																																																																																																																																	
Control Test	0.078	15:32																																																																																																																																																	
Air Blank	0.000	15:32																																																																																																																																																	
Control Test	0.078	15:33																																																																																																																																																	
Air Blank	0.000	15:34																																																																																																																																																	
Control Test	0.077	15:34																																																																																																																																																	
Air Blank	0.000	15:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:37																																																																																																																																																	
Control Test	0.203	15:38																																																																																																																																																	
Air Blank	0.000	15:38																																																																																																																																																	
Control Test	0.202	15:39																																																																																																																																																	
Air Blank	0.000	15:39																																																																																																																																																	
Control Test	0.201	15:40																																																																																																																																																	
Air Blank	0.000	15:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2020																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.4950																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test	0.079	15:26																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.079	15:26																																																																																																																																																	
Air Blank	0.000	15:27																																																																																																																																																	
Control Test	0.079	15:27																																																																																																																																																	
Air Blank	0.000	15:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 13:11

Date of Inspection: 05/04/2026

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.047	0.077	0.203	0.079
0.000	0.048	0.078	0.203	0.080
0.000	0.048	0.078	0.203	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.078	0.204	0.079
0.000	0.048	0.079	0.203	0.079


Standard Deviations	0.0003	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

05/04/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001321, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001321</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>OKEECHOBEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/04/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:11</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/04/2026

Date


KATIE T SPEARIN,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1