

Instrument Processing Sheet

Agency: Orange County SO Instrument Serial Number: 80-001256
 Date In: 11/4/2025 DI Completion Date: 11/18/2025 Ship P/U H/D CMI EE

Intake By: <u>SP</u> Date: <u>11/4/25</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>WKP</u> Date: <u>11/14/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>278</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 103</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.523</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1018.9</u> Instrument: <u>1019</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
---	---	--

Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:
0.050	MP6291	202406K	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:		
		06/19/2026			
0.080	MP6292	202406L			
		06/19/2026			
0.200	MP6293	202406N			
		06/20/2026			
0.080 DGS	N/A	AG510701			
		04/17/2027			

Optical Bench Adjustment By: _____	Department Inspection By: <u>SLH</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28427</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1017</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP6292	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6289																																								
Interferent	MP6290																																								
0.050	MP5088																																								
0.080	MP6292																																								
0.200	MP5090																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Barometric Pressure Gauge: _____ ID#: _____																																									

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: right;"> Digitally signed by Shayla Platt Date: 2025.11.20 17:19:57 -05'00' </div>
	Digitally signed by Taylor Gutschow Date: 2025.11.19 11:39:36 -05'00'
Tech Review	Admin Review

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																						
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																						
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																						
<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001256 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.049	11:14	Air Blank	0.000	11:15	Control Test	0.048	11:16	Air Blank	0.000	11:16	Control Test	0.049	11:17	Air Blank	0.000	11:17	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001256 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	11:19	Control Test	0.080	11:20	Air Blank	0.000	11:20	Control Test	0.079	11:21	Air Blank	0.000	11:22	Control Test	0.079	11:22	Air Blank	0.000	11:23	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001256 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test</td><td>0.202</td><td>11:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test</td><td>0.202</td><td>11:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2017</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2863</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	11:25	Control Test	0.202	11:26	Air Blank	0.000	11:26	Control Test	0.202	11:26	Air Blank	0.000	11:27	Control Test	0.201	11:27	Air Blank	0.000	11:28	Control Test	0.201	11:28	Air Blank	0.000	11:29	Control Test Stats			Average	0.2017		Std Dev	0.0006		Rel Std Dev(%)	0.2863		<p>ORANGE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001256 11/14/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	11:30	Control Test	0.079	11:30	Air Blank	0.000	11:31	Control Test	0.079	11:31	Air Blank	0.000	11:32	Control Test	0.080	11:32	Air Blank	0.000	11:33	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	11:14																																																																																																																																																							
Control Test	0.049	11:14																																																																																																																																																							
Air Blank	0.000	11:15																																																																																																																																																							
Control Test	0.048	11:16																																																																																																																																																							
Air Blank	0.000	11:16																																																																																																																																																							
Control Test	0.049	11:17																																																																																																																																																							
Air Blank	0.000	11:17																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0487																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	1.1863																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	11:19																																																																																																																																																							
Control Test	0.080	11:20																																																																																																																																																							
Air Blank	0.000	11:20																																																																																																																																																							
Control Test	0.079	11:21																																																																																																																																																							
Air Blank	0.000	11:22																																																																																																																																																							
Control Test	0.079	11:22																																																																																																																																																							
Air Blank	0.000	11:23																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0793																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.7277																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	11:25																																																																																																																																																							
Control Test	0.202	11:26																																																																																																																																																							
Air Blank	0.000	11:26																																																																																																																																																							
Control Test	0.202	11:26																																																																																																																																																							
Air Blank	0.000	11:27																																																																																																																																																							
Control Test	0.201	11:27																																																																																																																																																							
Air Blank	0.000	11:28																																																																																																																																																							
Control Test	0.201	11:28																																																																																																																																																							
Air Blank	0.000	11:29																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.2017																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.2863																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	11:30																																																																																																																																																							
Control Test	0.079	11:30																																																																																																																																																							
Air Blank	0.000	11:31																																																																																																																																																							
Control Test	0.079	11:31																																																																																																																																																							
Air Blank	0.000	11:32																																																																																																																																																							
Control Test	0.080	11:32																																																																																																																																																							
Air Blank	0.000	11:33																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0793																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.7277																																																																																																																																																								

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 14:01

Date of Inspection: 11/18/2025

Serial Number: 80-001256
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.078	0.198	0.000 / 0.079
0.000	0.049	0.078	0.198	0.000 / 0.079
0.000	0.049	0.078	0.198	0.000 / 0.079
0.000	0.049	0.078	0.198	0.000 / 0.079
0.000	0.049	0.078	0.198	0.000 / 0.079
0.000	0.050	0.079	0.199	0.000 / 0.079
0.000	0.049	0.078	0.199	0.000 / 0.079
0.000	0.050	0.079	0.199	0.000 / 0.079
0.000	0.049	0.078	0.199	0.000 / 0.078
0.000	0.050	0.078	0.199	0.000 / 0.078
Standard Deviations	0.0004	0.0004	0.0005	0.0000 / 0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

08: Control Outside ^{*}GAS TUBE NOT CONNECTED, CONNECTED

** Control Outside Tolerance -
inadvertently typed over sub*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

11/18/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001256, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001256</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/18/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:01</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


LEANDRA HIGGINBOTHAM,
Date 11/18/2025
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality