

Taylor,

80-005549

needs it

2015 Dept Inspection

Σ

MDSO Registration

Thy 1

R. Coors

10-01-2025
ML

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO
Time of Inspection: 09:45

Date of Inspection: 11/13/2025

Serial Number: 80-005549
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

11/13/2025
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>MIAMI DAGE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-05549 10/08/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:33	Control Test	0.049	12:33	Air Blank	0.000	12:34	Control Test	0.049	12:35	Air Blank	0.000	12:35	Control Test	0.049	12:35	Air Blank	0.000	12:36	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>MIAMI DAGE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-05549 10/08/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:39	Control Test	0.080	12:40	Air Blank	0.000	12:41	Control Test	0.079	12:41	Air Blank	0.000	12:42	Control Test	0.079	12:43	Air Blank	0.000	12:43	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel. Std Dev(%)	0.7277		<p>MIAMI DAGE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-05549 10/08/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2926</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:48	Control Test	0.198	12:48	Air Blank	0.000	12:49	Control Test	0.197	12:50	Air Blank	0.000	12:50	Control Test	0.197	12:51	Air Blank	0.000	12:52	Control Test Stats			Average	0.1973		Std Dev	0.0005		Rel. Std Dev(%)	0.2926		<p>MIAMI DAGE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-05549 10/08/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:55</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:55	Control Test	0.077	11:55	Air Blank	0.000	11:56	Control Test	0.077	11:56	Air Blank	0.000	11:57	Control Test	0.077	11:57	Air Blank	0.000	11:58	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test	0.049	12:33																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.049	12:35																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test	0.049	12:35																																																																																																																																																	
Air Blank	0.000	12:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.080	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test	0.079	12:41																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test	0.079	12:43																																																																																																																																																	
Air Blank	0.000	12:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.198	12:48																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.197	12:50																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test	0.197	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev(%)	0.2926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:55																																																																																																																																																	
Control Test	0.077	11:55																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.077	11:56																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test	0.077	11:57																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
<p>Operator's Signature <i>MC</i></p>	<p>Operator's Signature <i>MC</i></p>	<p>Operator's Signature <i>MC</i></p>	<p>Operator's Signature <i>MC</i></p>																																																																																																																																																

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤ 0.003 of Wet ✓																																																																																																																																																
<p>MIAMI DPE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005549 11/13/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.051</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:04	Control Test	0.051	14:05	Air Blank	0.000	14:06	Control Test	0.050	14:06	Air Blank	0.000	14:07	Control Test	0.050	14:08	Air Blank	0.000	14:08	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel. Std Dev(%)	1.1471		<p>MIAMI DPE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005549 11/13/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:20	Control Test	0.081	14:21	Air Blank	0.000	14:22	Control Test	0.080	14:22	Air Blank	0.000	14:23	Control Test	0.080	14:24	Air Blank	0.000	14:24	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel. Std Dev(%)	0.7187		<p>MIAMI DPE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005549 11/13/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:11	Control Test	0.200	14:12	Air Blank	0.000	14:12	Control Test	0.200	14:13	Air Blank	0.000	14:14	Control Test	0.199	14:14	Air Blank	0.000	14:15	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel. Std Dev(%)	0.2892		<p>MIAMI DPE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005549 11/13/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:46</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:45	Control Test	0.079	13:45	Air Blank	0.000	13:46	Control Test	0.079	13:46	Air Blank	0.000	13:47	Control Test	0.078	13:47	Air Blank	0.000	13:48	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel. Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test	0.051	14:05																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test	0.050	14:06																																																																																																																																																	
Air Blank	0.000	14:07																																																																																																																																																	
Control Test	0.050	14:08																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test	0.081	14:21																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.080	14:22																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.080	14:24																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.200	14:12																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test	0.200	14:13																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test	0.199	14:14																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2892																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test	0.079	13:45																																																																																																																																																	
Air Blank	0.000	13:46																																																																																																																																																	
Control Test	0.079	13:46																																																																																																																																																	
Air Blank	0.000	13:47																																																																																																																																																	
Control Test	0.078	13:47																																																																																																																																																	
Air Blank	0.000	13:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO
Time of Inspection: 14:31

Date of Inspection: 11/14/2025

Serial Number: 80-005549
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.080	0.199	0.078
0.000	0.048	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.077
0.000	0.048	0.080	0.198	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078

Standard Deviations	0.0004	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005549, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005549</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI DADE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/14/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:31</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor Gutschow
Digitally signed by Taylor Gutschow
Date: 2025.11.17 13:00:32 -05'00'

11/14/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality