

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 14:27

Date of Inspection: 11/07/2022

Serial Number: 80-007494

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15922080A4 Exp: 08/05/2024
0.000	0.050	0.079	0.198	0.000 / 0.079
0.000	0.050	0.080	0.198	0.079 / 0.079
0.000	0.050	0.079	0.199	0.079 / 0.080

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. / 08: Control Outside Tolerance.

TOO MUCH MOUTH ALCOHOL / TUBE WAS DISCONNECTED FROM CALIBRATION INLET.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAMON J PULASKI

Signature and Printed Name

11/07/2022
Date