

Alcohol Testing Program - Instrument Processing Sheet

Agency: Lake County SO

Instrument Serial Number: 80-000825

Date In: 6/1/2026

DI Completion Date: 6/11/26

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>KTS</u> Date: <u>6/1/2026</u>		Quality Checks By: <u>IAS</u> Date: <u>6/5/2026</u>		Flow Adjustment By: _____																											
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>217</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP105</u> 32 mm <u>.167</u> (.139-.169) 36 mm <u>.183</u> (.156-.190) 53 mm <u>.250</u> (.228-.278) 103 mm <u>.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1016</u> Instrument: <u>1015</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																											
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202603A 3/3/2028</td></tr><tr><td>0.080</td><td>MP6292</td><td>202603B 3/3/2028</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 6/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 3/3/2028	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202406N 6/20/2026	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:												
Simulator	Serial #	Lot#/Exp																													
0.050	MP6291	202603A 3/3/2028																													
0.080	MP6292	202603B 3/3/2028																													
0.200	MP6293	202406N 6/20/2026																													
0.080 DGS	N/A	23825080A1 10/5/2027																													
Optical Bench Adjustment By: _____		Department Inspection By: <u>KTS</u>																													
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1011</u> Instrument: <u>1011</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>	
Simulator	Serial #	Lot #	Expiration																												
0.000		N/A	N/A																												
0.040																															
0.100																															
0.200																															
0.300																															
0.080 DGS	N/A																														
☐ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293	Attachments															
Simulator	Serial Number																														
0.000	MP6289																														
Interferent	MP6290																														
0.050	MP6291																														
0.080	MP6292																														
0.200	MP6293																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment									
Simulator	Serial #	Lot #	Expiration																												
0.050																															
0.080																															
0.200																															
0.080 DGS	N/A																														
Gauge ID #: _____ Gauge: _____ Instrument: _____		☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																													
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																													
		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.16 08:45:21 -04'00'																													
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.06.19 14:55:26 -04'00'																													
		Tech Review Admin Review																													

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 12:11

Date of Inspection: 06/05/2026

Serial Number: 80-000825
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE UNDETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Isabella Smykowski

ISABELLA A SMYKOWSKI

Signature and Printed Name

06/05/2026
Date

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																				
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																				
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000825 06/05/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1853</td><td></td></tr> </table>	Air Blank	0.000	13:30	Control Test	0.049	13:31	Air Blank	0.000	13:32	Control Test	0.048	13:32	Air Blank	0.000	13:33	Control Test	0.049	13:34	Air Blank	0.000	13:34	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1853		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000825 06/05/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </table>	Air Blank	0.000	13:24	Control Test	0.078	13:24	Air Blank	0.000	13:25	Control Test	0.077	13:26	Air Blank	0.000	13:26	Control Test	0.077	13:27	Air Blank	0.000	13:28	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000825 06/05/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5852</td><td></td></tr> </table>	Air Blank	0.000	13:17	Control Test	0.198	13:18	Air Blank	0.000	13:19	Control Test	0.198	13:19	Air Blank	0.000	13:20	Control Test	0.196	13:21	Air Blank	0.000	13:21	Control Test Stats			Average	0.1973		Std Dev	0.0012		Rel Std Dev(%)	0.5852		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000825 06/05/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table>	Air Blank	0.000	13:12	Control Test	0.080	13:12	Air Blank	0.000	13:13	Control Test	0.079	13:13	Air Blank	0.000	13:14	Control Test	0.079	13:14	Air Blank	0.000	13:14	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Air Blank	0.000	13:30																																																																																																																																					
Control Test	0.049	13:31																																																																																																																																					
Air Blank	0.000	13:32																																																																																																																																					
Control Test	0.048	13:32																																																																																																																																					
Air Blank	0.000	13:33																																																																																																																																					
Control Test	0.049	13:34																																																																																																																																					
Air Blank	0.000	13:34																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0487																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1853																																																																																																																																						
Air Blank	0.000	13:24																																																																																																																																					
Control Test	0.078	13:24																																																																																																																																					
Air Blank	0.000	13:25																																																																																																																																					
Control Test	0.077	13:26																																																																																																																																					
Air Blank	0.000	13:26																																																																																																																																					
Control Test	0.077	13:27																																																																																																																																					
Air Blank	0.000	13:28																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0773																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7466																																																																																																																																						
Air Blank	0.000	13:17																																																																																																																																					
Control Test	0.198	13:18																																																																																																																																					
Air Blank	0.000	13:19																																																																																																																																					
Control Test	0.198	13:19																																																																																																																																					
Air Blank	0.000	13:20																																																																																																																																					
Control Test	0.196	13:21																																																																																																																																					
Air Blank	0.000	13:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1973																																																																																																																																						
Std Dev	0.0012																																																																																																																																						
Rel Std Dev(%)	0.5852																																																																																																																																						
Air Blank	0.000	13:12																																																																																																																																					
Control Test	0.080	13:12																																																																																																																																					
Air Blank	0.000	13:13																																																																																																																																					
Control Test	0.079	13:13																																																																																																																																					
Air Blank	0.000	13:14																																																																																																																																					
Control Test	0.079	13:14																																																																																																																																					
Air Blank	0.000	13:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						
<i>Isabella Snyders</i> Operator's Signature	<i>Isabella Snyders</i> Operator's Signature	<i>Isabella Snyders</i> Operator's Signature	<i>Isabella Snyders</i> Operator's Signature																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 12:43

Date of Inspection: 06/11/2026

Serial Number: 80-000825
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.079	0.198	0.079


Standard Deviations	0.0005	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

06/11/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000825, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000825</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAKE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/11/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:43</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Date

06/11/2026


KATIE T SPEARIN,
Department Inspector

Service • Integrity • Respect • Quality