

Alcohol Testing Program - Instrument Processing Sheet

Agency: HARDEE COUNTY SO

Instrument Serial Number: 80-001077

Date In: 6/25/2026

DI Completion Date: 7/21/26

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>6/25/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>7/14/26</u>		Flow Adjustment By: _____																																										
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>208</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm 0.160 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.242 (.228-.278) 103 mm 0.500 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1017</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																										
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/28</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/28</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/28</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/27</td> </tr> <tr> <td colspan="2"></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A 3/3/28	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP5089	202603B 3/3/28	0.200	MP5090	202603D 3/4/28	0.080 DGS	N/A	23825080A1 10/5/27																						
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																					
						0.050	MP5088	202603A 3/3/28	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																																					
						0.080	MP5089	202603B 3/3/28																																						
0.200	MP5090	202603D 3/4/28																																												
0.080 DGS	N/A	23825080A1 10/5/27																																												
Optical Bench Adjustment By: _____ Department Inspection By: <u>KTS</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>28427</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 15%;">Serial #</th> <th style="width: 15%;">Lot #</th> <th style="width: 15%;">Expiration</th> <th style="width: 40%;">Gauge: <u>1007</u> Instrument: <u>1009</u></th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> <td>Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> <td>Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> <td>Simulator Serial Number</td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> <td>0.000 MP5086</td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> <td>Interferent MP5087</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> <td>0.050 MP5088</td> </tr> <tr> <td colspan="4"> ☐ Post Optical Bench Adjustment Stability Checks </td> <td>0.080 MP5089</td> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td>0.200 MP5090</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	Gauge: <u>1007</u> Instrument: <u>1009</u>	0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u>	0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>	0.100				Simulator Serial Number	0.200				0.000 MP5086	0.300				Interferent MP5087	0.080 DGS	N/A			0.050 MP5088	☐ Post Optical Bench Adjustment Stability Checks				0.080 MP5089	0.050				0.200 MP5090
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1007</u> Instrument: <u>1009</u>																																										
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u>																																										
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																																										
0.100				Simulator Serial Number																																										
0.200				0.000 MP5086																																										
0.300				Interferent MP5087																																										
0.080 DGS	N/A			0.050 MP5088																																										
☐ Post Optical Bench Adjustment Stability Checks				0.080 MP5089																																										
0.050				0.200 MP5090																																										
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 15%;">Serial #</th> <th style="width: 15%;">Lot #</th> <th style="width: 15%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Adjustment ☒ Calibration Certificate ☒ Form 40 ☐ Optical Bench Adjustment ☒ Other: AI Note																								
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Gauge ID #: _____ Gauge: _____ Instrument: _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																												
Notes/Suggested Service: Stability Checks conducted on 7/15/26. KTS		Digitally signed by Taylor Gutschow Date: 2026.08.02 20:44:00 -04'00'																																												
		Digitally signed by LeAndra Higginbotham Date: 2026.08.03 17:32:59 -04'00'																																												
		Tech Review Admin Review																																												

Having issue getting it to complete the monthly inspection. If you can not get it to work.
Please pull the breath test that are on it and ship back. We are not going to repair it.

Thank You

Captain Cartwright

80-001077
WKP 6/25/2026

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 12:16

Date of Inspection: 07/14/2026

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE COMPLIANCE NOT DETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.





KATIE T SPEARIN

Signature and Printed Name

07/14/2026
Date

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:17	Control Test	0.050	13:17	Air Blank	0.000	13:18	Control Test	0.050	13:19	Air Blank	0.000	13:19	Control Test	0.049	13:20	Air Blank	0.000	13:20	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.078	13:23	Air Blank	0.000	13:24	Control Test	0.078	13:24	Air Blank	0.000	13:25	Control Test	0.078	13:25	Air Blank	0.000	13:25	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test</td><td>0.195</td><td>13:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:28</td></tr> <tr><td>Control Test</td><td>0.195</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr> <tr><td>Control Test</td><td>0.195</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2956</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:27	Control Test	0.195	13:28	Air Blank	0.000	13:28	Control Test	0.195	13:29	Air Blank	0.000	13:29	Control Test	0.195	13:30	Air Blank	0.000	13:31	Control Test Stats			Average	0.1953		Std Dev	0.0006		Rel Std Dev(%)	0.2956		HARDEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001077 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:13	Control Test	0.081	13:13	Air Blank	0.000	13:14	Control Test	0.081	13:14	Air Blank	0.000	13:15	Control Test	0.081	13:15	Air Blank	0.000	13:15	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.050	13:17																																																																																																																																																	
Air Blank	0.000	13:18																																																																																																																																																	
Control Test	0.050	13:19																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.049	13:20																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test	0.078	13:23																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.078	13:24																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.078	13:25																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test	0.195	13:28																																																																																																																																																	
Air Blank	0.000	13:28																																																																																																																																																	
Control Test	0.195	13:29																																																																																																																																																	
Air Blank	0.000	13:29																																																																																																																																																	
Control Test	0.195	13:30																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1953																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2956																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.081	13:13																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test	0.081	13:14																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.081	13:15																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 09:55

Date of Inspection: 07/21/2026

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.051	0.079	0.198	0.081
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.079	0.198	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.199	0.080


Standard Deviations	0.0000	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


KATIE T SPEARIN
Signature and Printed Name

07/21/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001077</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HARDEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/21/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:55</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/21/2026

Date

KATIE T SPEARIN,
Department Inspector