

Alcohol Testing Program - Instrument Processing Sheet

Agency: Jacksonville Sheriff Office Instrument Serial Number: 80-001278

Date In: 7/24/2026 DI Completion Date: 7/31/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>7/24/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>7/24/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>191</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.253</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1011</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>SLH</u>																																	
Barometric Pressure Gauge: <u>1008</u> ID#: <u>34419</u>		Barometric Pressure ID#: <u>28427</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1010</u> Instrument: <u>1012</u>																															
0.000	MP6294	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u>																															
0.040	MP6295	25090	03/11/2027	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																															
0.100	MP6296	24390	10/29/2026	Simulator	Serial Number																														
0.200	MP6297	25020	01/14/2027	0.000	MP5086																														
0.300	MP6298	25150	05/06/2027	Interferent	MP5087																														
0.080 DGS	N/A	AG526603	09/23/2027	0.050	MP5088																														
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																														
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																														
0.050	MP6291	202603A	03/03/2028	Attachments																															
0.080	MP6292	202603B	03/03/2028	☒ Form 41	☒ Post-Stability Checks																														
0.200	MP6293	202603D	03/04/2028	☒ Stability Checks	☐ Flow Adjustment																														
0.080 DGS	N/A	23825080A1	10/5/2027	☒ Calibration Certificate	☐ Form 40																														
Gauge ID #: <u>28427</u>				☒ Optical Bench Adjustment	☐ Other:																														
Gauge: <u>1007</u> Instrument: <u>1008</u>																																			
Notes/Suggested Service: Performed flow verification 7/28/2026. SLH 7/28/2026				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
				Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.03 11:42:41 04:00' Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.03 12:23:09 -04:00'																															
				Tech Review	Admin Review																														

JACKSONVILLE SC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001278
07/24/2026
Software: 9100.27

Test	g/210L	Time
Air Blank	0.000	16:23
Control Test	0.075 —	16:23
Air Blank	0.000	16:24
Control Test	0.077	16:24
Air Blank	0.000	16:25
Control Test	0.077	16:26
Air Blank	0.000	16:26
Control Test Stats		
Average	0.0763	
Std Dev	0.0012	
Rel Std Dev(%)	1.5127	


Operator's Signature

as stability ran 1st
replicate, secured
fit. see 2nd run
for 0.080. ~~sub~~

Stability Checks

80-00 1278

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
JACKSONVILLE SD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001278 07/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>16:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:38</td></tr> <tr><td>Control Test</td><td>0.048</td><td>16:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:39</td></tr> <tr><td>Control Test</td><td>0.049</td><td>16:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:36	Control Test	0.049	16:37	Air Blank	0.000	16:38	Control Test	0.048	16:38	Air Blank	0.000	16:39	Control Test	0.049	16:39	Air Blank	0.000	16:40	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		#2 JACKSONVILLE SD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001278 07/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:46</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:47</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:48</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:46	Control Test	0.077	16:47	Air Blank	0.000	16:47	Control Test	0.077	16:48	Air Blank	0.000	16:48	Control Test	0.077	16:49	Air Blank	0.000	16:50	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		JACKSONVILLE SD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001278 07/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:00</td></tr> <tr><td>Control Test</td><td>0.195</td><td>17:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:01</td></tr> <tr><td>Control Test</td><td>0.195</td><td>17:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:03</td></tr> <tr><td>Control Test</td><td>0.195</td><td>17:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1950</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	17:00	Control Test	0.195	17:01	Air Blank	0.000	17:01	Control Test	0.195	17:02	Air Blank	0.000	17:03	Control Test	0.195	17:03	Air Blank	0.000	17:04	Control Test Stats			Average	0.1950		Std Dev	0.0000		Rel Std Dev(%)	0.0000		DGS JACKSONVILLE SD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001278 07/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:28</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:29</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:30</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:28	Control Test	0.081	16:28	Air Blank	0.000	16:29	Control Test	0.081	16:29	Air Blank	0.000	16:30	Control Test	0.081	16:30	Air Blank	0.000	16:31	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:36																																																																																																																																																	
Control Test	0.049	16:37																																																																																																																																																	
Air Blank	0.000	16:38																																																																																																																																																	
Control Test	0.048	16:38																																																																																																																																																	
Air Blank	0.000	16:39																																																																																																																																																	
Control Test	0.049	16:39																																																																																																																																																	
Air Blank	0.000	16:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:46																																																																																																																																																	
Control Test	0.077	16:47																																																																																																																																																	
Air Blank	0.000	16:47																																																																																																																																																	
Control Test	0.077	16:48																																																																																																																																																	
Air Blank	0.000	16:48																																																																																																																																																	
Control Test	0.077	16:49																																																																																																																																																	
Air Blank	0.000	16:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:00																																																																																																																																																	
Control Test	0.195	17:01																																																																																																																																																	
Air Blank	0.000	17:01																																																																																																																																																	
Control Test	0.195	17:02																																																																																																																																																	
Air Blank	0.000	17:03																																																																																																																																																	
Control Test	0.195	17:03																																																																																																																																																	
Air Blank	0.000	17:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1950																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:28																																																																																																																																																	
Control Test	0.081	16:28																																																																																																																																																	
Air Blank	0.000	16:29																																																																																																																																																	
Control Test	0.081	16:29																																																																																																																																																	
Air Blank	0.000	16:30																																																																																																																																																	
Control Test	0.081	16:30																																																																																																																																																	
Air Blank	0.000	16:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

needs optical Bench
adjustment

Verified instrument was
able to read gas pressure. ☒

JACKSONVILLE SO ✓
Intoxilyzer - Alcohol Analyzer
Model 8000
07/28/2026

SN 80-001278 ✓
10:11:48

SUB

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0600	(-0.0130)	0.0780	(0.0000)
Sample #2	0.1180	(-0.0270)	0.1340	(-0.0390)
Sample #3	0.1270	(-0.0310)	0.1350	(-0.0420)
Sample #4	0.1020	(-0.0160)	0.1340	(-0.0330)
Avg % Abs	0.1157	(-0.0247)	0.1343	(-0.0380)
STD DEV	0.0127	(0.0078)	0.0006	(0.0046)
REL STD DEV	10.947	(31.490)	0.430	(12.059)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8110	(-0.0240)	1.4820	(-0.0130)
Sample #2	0.7980	(-0.0430)	1.4770	(-0.0160)
Sample #3	0.8330	(-0.0540)	1.4940	(-0.0150)
Sample #4	0.8060	(-0.0430)	1.4530	(-0.0080)
Avg % Abs	0.8123	(-0.0467)	1.4747	(-0.0130)
STD DEV	0.0183	(0.0064)	0.0206	(0.0044)
REL STD DEV	2.258	(13.609)	1.397	(33.530)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.8250	(-0.0060)	3.4780	(-0.0150)
Sample #2	1.8110	(-0.0240)	3.4320	(0.0170)
Sample #3	1.8320	(-0.0450)	3.4550	(-0.0080)
Sample #4	1.8180	(-0.0310)	3.4440	(0.0000)
Avg % Abs	1.8203	(-0.0333)	3.4437	(0.0030)
STD DEV	0.0107	(0.0107)	0.0115	(0.0128)
REL STD DEV	0.587	(32.078)	0.334	(425.572)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.5580	(-0.0300)	6.6660	(-0.0130)
Sample #2	3.5420	(0.0300)	6.6250	(0.0670)
Sample #3	3.5180	(0.0230)	6.5940	(0.0610)
Sample #4	3.5350	(0.0400)	6.5940	(0.0660)
Avg % Abs	3.5317	(0.0310)	6.6043	(0.0647)
STD DEV	0.0123	(0.0085)	0.0179	(0.0032)
REL STD DEV	0.349	(27.561)	0.271	(4.971)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.1170	(-0.0130)	9.6300	(-0.0110)
Sample #2	5.1710	(-0.0620)	9.6150	(0.0130)
Sample #3	5.1900	(-0.0780)	9.6260	(0.0000)
Sample #4	5.1570	(-0.0610)	9.6010	(0.0120)
Avg % Abs	5.1727	(-0.0670)	9.6140	(0.0083)
STD DEV	0.0166	(0.0095)	0.0125	(0.0072)
REL STD DEV	0.320	(14.238)	0.130	(86.810)

JACKSONVILLE SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001278 ✓
 07/28/2026 10:11:48

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-314.78			-183.45	
First Order Coef	2738.65			1396.19	
Second Order Coef	15.85			11.29	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	0.000	-0.0000	0.000	0.000	-0.0001
0.040	0.040	-0.0003	0.040	0.040	0.0001
0.100	0.099	0.0008	0.100	0.100	0.0001
0.200	0.201	-0.0007	0.200	0.200	-0.0001
0.300	0.300	0.0002	0.300	0.300	0.0001
-----			-----		

<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3143.00		3421.00
Sample #2	3105.00		3387.00
Sample #3	3118.00		3418.00
Sample #4	3002.00		3396.00
Avg	3075.0000		3400.3333
STD DEV	63.5531		15.9478
REL STD DEV	2.067		0.469
H2O adjust (mg/l*10k)	734		409

Barometric Pressure = 1007

*****CALIBRATION SUCCESSFUL*****

✓ SNA

Post Optical Bench Adjustment Stability Checks 80-001278

Sut

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
JACKSONVILLE SO ✓ Intoxilyzer - Alconol Analyzer ✓ Model 8000 SN 80-001278 ✓ 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	11:06	Control Test	0.050	11:07	Air Blank	0.000	11:07	Control Test	0.050	11:08	Air Blank	0.000	11:08	Control Test	0.050	11:09	Air Blank	0.000	11:10	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		JACKSONVILLE SO ✓ Intoxilyzer - Alconol Analyzer ✓ Model 8000 SN 80-001278 ✓ 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	11:00	Control Test	0.080	11:01	Air Blank	0.000	11:01	Control Test	0.079	11:02	Air Blank	0.000	11:02	Control Test	0.079	11:03	Air Blank	0.000	11:04	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		JACKSONVILLE SO ✓ Intoxilyzer - Alconol Analyzer ✓ Model 8000 SN 80-001278 ✓ 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5783</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	11:16	Control Test	0.201	11:16	Air Blank	0.000	11:17	Control Test	0.199	11:18	Air Blank	0.000	11:18	Control Test	0.199	11:19	Air Blank	0.000	11:20	Control Test Stats			Average	0.1997		Std Dev	0.0012		Rel Std Dev(%)	0.5783		JACKSONVILLE SO ✓ Intoxilyzer - Alconol Analyzer ✓ Model 8000 SN 80-001278 ✓ 07/28/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Miguel</i>	Test	g/210L	Time	Air Blank	0.000	11:23	Control Test	0.080	11:23	Air Blank	0.000	11:24	Control Test	0.080	11:24	Air Blank	0.000	11:24	Control Test	0.081	11:25	Air Blank	0.000	11:25	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test	0.050	11:07																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.050	11:08																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test	0.050	11:09																																																																																																																																																	
Air Blank	0.000	11:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test	0.080	11:01																																																																																																																																																	
Air Blank	0.000	11:01																																																																																																																																																	
Control Test	0.079	11:02																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.079	11:03																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.201	11:16																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.199	11:18																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.199	11:19																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5783																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.080	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.080	11:24																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.081	11:25																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 15:26

Date of Inspection: 07/31/2026

Serial Number: 80-001278
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.051	0.078	0.200	0.080
0.000	0.051	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080


Standard Deviations	0.0004	0.0003	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/31/2026
Date



ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001278, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001278</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>07/31/2026</u>	0.080 g/210 L	0.004
Calibration Time:	<u>15:26</u>	0.200 g/210 L	0.008
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/31/2026
Date

Stephane
LEANDRA HIGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality