

Alcohol Testing Program - Instrument Processing Sheet

Agency: Miami Dade Sheriff's Office Instrument Serial Number: 80-000880
 Date In: 11/14/2025 DI Completion Date: 2/5/2026 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>1/16/2026</u>	Quality Checks By: <u>TDG</u> Date: <u>2/4/2026</u>	Flow Adjustment By: _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>233</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1022</u> Instrument: <u>1021</u>	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

Notes: Agency reports the instrument is giving Interferent Detect, Ambient Fail, and Diagnostic Fail (Analytical Stability Test) messages.

Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____
0.050	MP6286	202406K 6/19/2026	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____
0.080	MP6287	202406L 6/19/2026	
0.200	MP6288	202406N 6/20/2026	
0.080 DGS	N/A	AG429602 10/22/2026	

Optical Bench Adjustment By: _____	Department Inspection By: <u>TDG</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1015</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks																																									
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input checked="" type="checkbox"/> Other: Return Material Authorization																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Gauge ID #: _____ Gauge: _____ Instrument: _____																																									

Notes/Suggested Service: Monitored the DVM screen. The 9-micron channel signal fluctuates from 0 to ~16. (TDG 2/4/26)
 Could not reproduce exception messages described by the agency. Obtained out-of-range results during the Department Inspection that were not due to simulator/user error. Sending instrument to CMI for evaluation. (TDG 2/5/26)

<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
LeAndra Higginbotham <small>Digitally signed by LeAndra Higginbotham Date: 2026.02.13 13:06:43 -05'00'</small> Wen-Chi Pierson <small>Digitally signed by Wen-Chi Pierson Date: 2026.02.13 13:46:57 -05'00'</small>
Tech Review Admin Review

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO
Time of Inspection: 12:52

Date of Inspection: 02/04/2026

Serial Number: 80-000880
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.







TAYLOR D GUTSCHOW

Signature and Printed Name

02/04/2026
Date

Stability Checks

0.05g/210L 0.047 to 0.053 ✓			0.08g/210L 0.077 to 0.083 ✓			0.20g/210L 0.194 to 0.206 ✓			DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓		
MIAMI DADE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000880 02/04/2026 Software: 8100.27			MIAMI DADE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000880 02/04/2026 Software: 8100.27			MIAMI DADE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000880 02/04/2026 Software: 8100.27			MIAMI DADE SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000880 02/04/2026 Software: 8100.27		
Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	13:01	Air Blank	0.000	13:08	Air Blank	0.000	12:54	Air Blank	0.000	13:13
Control Test	0.051	13:01	Control Test	0.081	13:09	Control Test	0.198	12:54	Control Test	0.078	13:13
Air Blank	0.000	13:02	Air Blank	0.000	13:09	Air Blank	0.000	12:55	Air Blank	0.000	13:14
Control Test	0.050	13:03	Control Test	0.080	13:10	Control Test	0.201	12:55	Control Test	0.080	13:14
Air Blank	0.000	13:03	Air Blank	0.000	13:10	Air Blank	0.000	12:56	Air Blank	0.000	13:15
Control Test	0.049	13:04	Control Test	0.080	13:11	Control Test	0.202	12:56	Control Test	0.079	13:15
Air Blank	0.000	13:04	Air Blank	0.000	13:12	Air Blank	0.000	12:57	Air Blank	0.000	13:16
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0500		Average	0.0803		Average	0.2003		Average	0.0790	
Std Dev	0.0010		Std Dev	0.0006		Std Dev	0.0021		Std Dev	0.0010	
Rel Std Dev(%)	2.0000		Rel Std Dev(%)	0.7187		Rel Std Dev(%)	1.0391		Rel Std Dev(%)	1.2658	
 Operator's Signature			 Operator's Signature			 Operator's Signature			 Operator's Signature		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO
Time of Inspection: 11:34

Date of Inspection: 02/05/2026

Serial Number: 80-000880
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000 / 0.000	0.049	0.079	0.202 / 0.200	0.084
0.000 / 0.000	0.050	0.081	0.202 / 0.202	0.077
0.000 / 0.000	0.049	0.081	0.203 / 0.198	0.070
0.007 / 0.000	0.050	0.081	0.202 / 0.203	0.068
0.000 / 0.000	0.049	0.080	0.202 / 0.201	0.069
0.000 / 0.000	0.048	0.083	0.214 / 0.199	0.068
0.000 / 0.000	0.050	0.081	0.206 / 0.199	0.069
0.007 / 0.000	0.049	0.084	0.206 / 0.200	0.068
0.000 / 0.000	0.049	0.083	0.206 / 0.203	0.069
0.000 / 0.000	0.049	0.084	0.203 / 0.199	0.070
Standard Deviations	0.0006	0.0017	0.0037 / 0.0017	0.0052

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0023 Number of Simulators Used: 5

Remarks:

00: Control Outside Tolerance. 20: Control Outside Tolerance. 08: Control Outside Tolerance. Non-compliance: .

Suspect the signal at the 9-micron channel was responsible for failed tests (see note on IPS). Resolved to send the instrument to repair after the initial Alcohol-Free Test failed and repeated the test to generate more data to send to the repair facility. Did the same with the 0.20 Test. Did not repeat the 0.08 DGS Test due to sufficient data gathered at that point. Department Inspection failed and was marked as noncompliant. (TDG 2/5/26)

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

02/05/2026
Date

Return Material Authorization

Ship to:

CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Richard Closius on 2/6/2026

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000880

Bill To Address: Miami Dade Sheriff's Office Attn: Richard Closius	Ship to Address: Florida Department of Law Enforcement Fort Myers Regional Operations Center Attn: Taylor Gutschow 4700 Terminal Drive, Suite 1 Fort Myers, FL 33907
--	---

Reason for Return:

The signal at the 9-micron channel is fluctuating from 0-16. The agency reports receiving Ambient Fails, Interferent Detects, and Diagnostic Fails (Analytical Stability Test). During the Department Inspection at FDLE, the instrument gave out-of-range results that were not due to simulator or user error. See attached Department Inspection paperwork.

Agency wishes CMI to know they may have to pay with fiscal budget starting after 9/30/2026.

I require an estimate **BEFORE** any repairs will be authorized and/or conducted

Please contact: Richard Closius Phone #: 305-458-3991

Email: rlclosius@mdso.com

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us