

Alcohol Testing Program - Instrument Processing Sheet

Agency: Orange County SO

Instrument Serial Number: 80-001258

Date In: 5/4/2026

DI Completion Date: 05/18/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: SLH Date: 5/4/2026		Quality Checks By: SLH Date: 5/5/2026		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 156 ☒ Flow Verification (L/s) Flow Column #: ATP 103 32 mm0.140(.139-.169) 36 mm0.160(.156-.190) 53 mm0.234(.228-.278) 103 mm0.511(.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1011Instrument: 1013 ☒ Stability Checks <table><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202603A 03/03/2028</td></tr><tr><td>0.080</td><td>MP6292</td><td>202603B 03/03/2028</td></tr><tr><td>0.200</td><td>MP6293</td><td>202603D 03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23725080A1 10/05/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 03/03/2028	0.080	MP6292	202603B 03/03/2028	0.200	MP6293	202603D 03/04/2028	0.080 DGS	N/A	23725080A1 10/05/2027	Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm(.139-.169) 36 mm(.156-.190) 53 mm(.228-.278) 103 mm(.447-.547)	
Simulator	Serial #	Lot#/Exp																		
0.050	MP6291	202603A 03/03/2028																		
0.080	MP6292	202603B 03/03/2028																		
0.200	MP6293	202603D 03/04/2028																		
0.080 DGS	N/A	23725080A1 10/05/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: SLH																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28662																		
SimulatorSerial #Lot #Expiration		Gauge: 1019Instrument: 1020																		
0.000N/AN/A		Mouth Alcohol Solution Lot #: 2025-DExp: 09/25/2027																		
0.040		Acetone Stock Solution Lot #: 2025-BExp: 09/22/2027																		
0.100		Simulator		Serial Number																
0.200		0.000		MP5086																
0.300		Interferent		MP5087																
0.080 DGSN/A		0.050		MP5088																
☐ Post Optical Bench Adjustment Stability Checks		0.080		MP5089																
SimulatorSerial #Lot #Expiration		0.200		MP5090																
0.050		Attachments																		
0.080		☒ Form 41		☐ Post-Stability Checks																
0.200		☒ Stability Checks		☐ Flow Adjustment																
0.080 DGSN/A		☒ Calibration Certificate		☐ Form 40																
Gauge ID #: _____		☐ Optical Bench Adjustment		☐ Other:																
Gauge: _____Instrument: _____																				
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																		
		Taylor GutschowDigitally signed by Taylor Gutschow Date: 2026.05.21 13:51:40 -04'00'		Wen-Chi PiersonDigitally signed by Wen-Chi Pierson Date: 2026.05.22 08:48:32 -04'00'																
		Tech Review		Admin Review																

Stability Checks

80-001258 sub

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
ORANGE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001258 05/05/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test</td><td>0.047</td><td>14:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr> <tr><td>Control Test</td><td>0.047</td><td>14:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0473</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2198</td><td></td></tr> </tbody> </table> <i>Mijjal</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:32	Control Test	0.048	14:33	Air Blank	0.000	14:33	Control Test	0.047	14:34	Air Blank	0.000	14:34	Control Test	0.047	14:35	Air Blank	0.000	14:35	Control Test Stats			Average	0.0473		Std Dev	0.0006		Rel Std Dev(%)	1.2198		ORANGE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001258 05/05/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Mijjal</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:46	Control Test	0.077	14:47	Air Blank	0.000	14:48	Control Test	0.077	14:48	Air Blank	0.000	14:49	Control Test	0.077	14:49	Air Blank	0.000	14:50	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ORANGE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001258 05/05/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.194</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr> <tr><td>Control Test</td><td>0.194</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1947</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5932</td><td></td></tr> </tbody> </table> <i>Mijjal</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:24	Control Test	0.194	14:25	Air Blank	0.000	14:25	Control Test	0.196	14:26	Air Blank	0.000	14:26	Control Test	0.194	14:27	Air Blank	0.000	14:28	Control Test Stats			Average	0.1947		Std Dev	0.0012		Rel Std Dev(%)	0.5932		ORANGE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001258 05/05/2026 Software: 8100.27 DGS <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <i>Mijjal</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:42	Control Test	0.078	14:42	Air Blank	0.000	14:43	Control Test	0.079	14:43	Air Blank	0.000	14:44	Control Test	0.079	14:44	Air Blank	0.000	14:44	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:32																																																																																																																																																	
Control Test	0.048	14:33																																																																																																																																																	
Air Blank	0.000	14:33																																																																																																																																																	
Control Test	0.047	14:34																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																	
Control Test	0.047	14:35																																																																																																																																																	
Air Blank	0.000	14:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0473																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2198																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																	
Control Test	0.077	14:47																																																																																																																																																	
Air Blank	0.000	14:48																																																																																																																																																	
Control Test	0.077	14:48																																																																																																																																																	
Air Blank	0.000	14:49																																																																																																																																																	
Control Test	0.077	14:49																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test	0.194	14:25																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test	0.196	14:26																																																																																																																																																	
Air Blank	0.000	14:26																																																																																																																																																	
Control Test	0.194	14:27																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1947																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5932																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.078	14:42																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.079	14:43																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test	0.079	14:44																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

instrument reading tank pressure ✓ sub

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 13:39

Date of Inspection: 05/18/2026

Serial Number: 80-001258
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.076	0.199	0.079
0.000	0.049	0.077	0.200	0.079
0.000	0.049	0.076	0.199	0.079
0.000	0.049	0.077	0.198	0.079
0.000	0.049	0.077	0.198	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.049	0.077	0.198	0.079
0.000	0.049	0.077	0.198	0.079

Standard Deviations	0.0003	0.0004	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

05/18/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001258, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001258</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/18/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:39</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.


Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/18/2026

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1