

Instrument Processing Sheet

Agency: St. Petersburg Police Department Instrument Serial Number: 80-001653
 Date In: 10/30/2025 DI Completion Date: N/A Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>10/30/25</u> <input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input checked="" type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	Quality Checks By: _____ Date: _____ <input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) <input type="checkbox"/> Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ <input type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Notes: Dropped off by Agency Inspector. No box. Breath test data entry appears to be disabled. Instrument has already had a calibration this year.

Simulator	Serial #	Lot#/Exp	Maintenance
0.050			<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
0.080			
0.200			
0.080 DGS	N/A		

Optical Bench Adjustment By: _____	Department Inspection By: _____			
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: _____			
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____
0.040				Acetone Stock Solution Lot #: _____ Exp: _____
0.100				Simulator _____ Serial Number _____
0.200				0.000
0.300				Interferent
0.080 DGS	N/A			0.050
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080
0.050				0.200
0.080				Attachments
0.200				<input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:
0.080 DGS	N/A			

Notes/Suggested Service: Verified setup on 10/30. Data entry was disabled. Enabled data entry. Will monitor over a period of time to ensure the setting remains unchanged. (TDG 10/30/25) Verified setup again on 12/9. Data entry is still enabled. Shipping back to the agency in a loaner box without another Department Inspection. (TDG 12/9/25)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

 Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by
Shayla Platt
 Date: 2025.12.15 13:57:30 -05'00'

Digitally signed by LeAndra Higginbotham
 Date: 2025.12.17 10:10:23 -05'00'

Tech Review
Admin Review

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001653
03/20/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:05
Control Test	0.044	14:06
Air Blank	0.000	14:07
Control Test	0.042	14:07
Air Blank	0.000	14:08
Control Test	0.042	14:08
Air Blank	0.000	14:09
Control Test Stats		
Average	0.0427	
Std Dev	0.0012	
Rel. Std Dev(%)	2.7163	

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001653
03/20/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:54
Control Test	0.084	13:55
Air Blank	0.000	13:56
Control Test	0.076	13:56
Air Blank	0.000	13:57
Control Test	0.083	13:57
Air Blank	0.000	13:58
Control Test Stats		
Average	0.0810	
Std Dev	0.0044	
Rel. Std Dev(%)	5.3814	

Wet

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001653
03/20/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:27
Control Test	0.201	14:27
Air Blank	0.000	14:28
Control Test	0.198	14:29
Air Blank	0.000	14:29
Control Test	0.197	14:30
Air Blank	0.000	14:30
Control Test Stats		
Average	0.1987	
Std Dev	0.0021	
Rel. Std Dev(%)	1.0478	

Operator's Signature

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001653
03/20/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:01
Control Test	0.087	14:01
Air Blank	0.000	14:01
Control Test	0.081	14:02
Air Blank	0.000	14:02
Control Test	0.082	14:03
Air Blank	0.000	14:03
Control Test Stats		
Average	0.0833	
Std Dev	0.0032	
Rel. Std Dev(%)	3.8575	

DGS

Operator's Signature

Optical Bench Calibration Adjustment 80-001653 DA 3/27/25

ST PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001653
03/27/2025 13:45:47

Auto Calibration
Max Power Res Value = 98
Auto Range Res Value = 85

Sol Value = 0.000 g/210L ***
Fit Value = 0.0100 mg/L ****
Samples Taken = 4, Discarded = 1
Sum Io = 12624, Sum Ie = 13223

Sample % Abs (% Abs Ref)
Sample #1 = 0.1030 (-0.0060)
Sample #2 = 0.1270 (0.0150)
Sample #3 = 0.1170 (0.0190)
Sample #4 = 0.1310 (0.0430)
Avg % Abs = 0.1250 (0.0257)
STD DEV = 0.0072 (0.0151)
REL STD DEV = 5.769 (59.002)

Sample % Abs (% Abs Ref)
Sample #1 = 0.2940 (0.0710)
Sample #2 = 0.3000 (0.3030)
Sample #3 = 0.1420 (0.2460)
Sample #4 = -0.3940 (0.7820)
Avg % Abs = 0.0160 (0.4437)
STD DEV = 0.3638 (0.2944)
REL STD DEV = 2273.454 (66.353)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/L ****
Samples Taken = 4, Discarded = 1
Sum Io = 12608, Sum Ie = 13108

Sample % Abs (% Abs Ref)
Sample #1 = 0.8590 (-0.0220)
Sample #2 = 0.8400 (0.0020)
Sample #3 = 0.8500 (0.0060)
Sample #4 = 0.8560 (0.0020)
Avg % Abs = 0.8487 (0.0033)
STD DEV = 0.0081 (0.0023)
REL STD DEV = 0.952 (69.282)

Sample % Abs (% Abs Ref)
Sample #1 = 2.0250 (0.2490)
Sample #2 = 1.6360 (-0.4040)
Sample #3 = 2.5820 (-0.1440)
Sample #4 = 1.2650 (0.6040)
Avg % Abs = 1.8210 (0.0167)
STD DEV = 0.6680 (0.5233)
REL STD DEV = 36.683 (2803.489)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/L ****
Samples Taken = 4, Discarded = 1
Sum Io = 12605, Sum Ie = 13061

Sample % Abs (% Abs Ref)
Sample #1 = 0.8480 (-0.0060)
Sample #2 = 0.8570 (-0.0010)
Sample #3 = 0.8630 (0.0190)
Sample #4 = 0.8590 (0.0190)
Avg % Abs = 0.8630 (0.0090)
STD DEV = 0.0060 (0.0100)
REL STD DEV = 0.695 (0.1110)

Sample % Abs (% Abs Ref)
Sample #1 = 1.5050 (-0.0540)
Sample #2 = 1.6430 (-0.4720)
Sample #3 = 1.5020 (-0.0320)
Sample #4 = 1.3070 (-0.1290)
Avg % Abs = 1.4840 (-0.2110)
STD DEV = 0.1667 (0.2312)
REL STD DEV = 11.369 (109.563)

**** AUTO CAL FAIL

Optical bench calibration adjustment wouldn't progress at the 0.04g/210L concentration. I checked the simulator for proper seal and connection to the instrument and repeated analysis of the 0.04g/210L solution. The instrument failed the calibration adjustment after second analysis attempt. I plan to repeat the optical bench calibration adjustment on 03/28/25. DA 3/27/25.

Optical Bench Calibration Adjustment 80-001453 3/28/25 DA

ST PETERSBURG PD
 Intoxilyzer - Alcohol Analyser
 Model 8000 SN 60-001653
 03/28/2025 18:35:18

Auto Calibration
 Max Power Res Value = 98
 Auto Range Res Value = 87

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12611, Sum Io = 12979

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1930	(-0.0220)
Sample #2 =	0.1030	(-0.0160)
Sample #3 =	0.1200	(-0.0050)
Sample #4 =	0.1030	(0.0000)
Avg % Abs =	0.1087	(-0.0070)
STD DEU =	0.0098	(0.0082)
REL STD DEU =	9.032	(116.934)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	-0.1270	(-0.1190)
Sample #2 =	0.4470	(-0.4470)
Sample #3 =	-0.1650	(-0.4280)
Sample #4 =	0.4390	(-0.5150)
Avg % Abs =	0.2737	(-0.4633)
STD DEU =	0.2933	(0.0457)
REL STD DEU =	107.192	(9.872)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12610, Sum Io = 13049

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8451	(-0.0370)
Sample #2 =	0.8590	(-0.0190)
Sample #3 =	0.8600	(-0.0200)
Sample #4 =	0.8360	(-0.0100)
Avg % Abs =	0.8517	(-0.0163)
STD DEU =	0.0136	(0.0055)
REL STD DEU =	1.594	(33.720)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.4860	(-0.1260)
Sample #2 =	1.7070	(-0.5990)
Sample #3 =	1.5960	(-0.3680)
Sample #4 =	1.6230	(-0.5570)
Avg % Abs =	1.6420	(-0.5180)
STD DEU =	0.0579	(0.1137)
REL STD DEU =	3.525	(21.853)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l ****
 Samples Taken = 4, Discarded = 1
 Sum Io = 12614, Sum Io = 13079

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8020	(-0.0180)
Sample #2 =	0.8360	(-0.0280)
Sample #3 =	0.8310	(-0.0050)
Sample #4 =	0.8330	(-0.0170)
Avg % Abs =	0.8333	(-0.0167)
STD DEU =	0.0025	(0.0115)
REL STD DEU =	0.302	(39.022)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5690	(-0.1200)
Sample #2 =	1.5060	(0.0030)
Sample #3 =	1.7260	(-0.0050)
Sample #4 =	1.3230	(0.2640)
Avg % Abs =	1.5190	(0.0870)
STD DEU =	0.2017	(0.1534)
REL STD DEU =	13.280	(176.267)

**** AUTO CAL FAIL

I tested all alcohol reference solutions used within this calibration adjustment using a training instrument after the 0.04g/210L concentration failed. All alcohol reference solutions were within range. I repeated the analysis of the 0.04g/210L concentration and the calibration adjustment failed. DA 3/28/25

Return Material Authorization

Ship to: CMI, Inc.
 Enforcement Electronics

Shipment to repair facility authorized by: Dustin Hall on 03/28/2025

Items Returned: Instrument Supplies Other Describe: _____
Instrument Model: Intoxilyzer 8000 Serial Number: 80-001653

Bill To Address: <u>St Petersburg PD</u> Attn: <u>Dustin Hall</u>	Ship to Address: <u>Florida Department of Law Enforcement</u> <u>FMROC</u> <u>Alcohol Testing Program</u> <u>4700 Terminal Drive, Suite 1</u> <u>Fort Myers, FL 33907</u>
-------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Reason for Return:
Instrument will not pass the 0.04g/210L concentration during optical bench calibration adjustment. I repeated the adjustment twice.

Please choose one of the following options:

- 1. I _____, authorize all repairs.
- 2. I _____, authorize repairs up to \$_____.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Dustin Hall
Phone #: 941-448-8755 Email: DLHALL@stpete.org

ATP Contact Name: Taylor Gutschow ATP Email: taylorgutschow@fdle.state.fl.us