

Alcohol Testing Program - Instrument Processing Sheet

Agency: INDIAN RIVER CO SO

Instrument Serial Number: 80-001328

Date In: 6/8/2026

DI Completion Date: 6/10/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>KTS</u> Date: <u>6/8/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>6/10/2026</u>		Flow Adjustment By: _____																				
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Came in a regular box. AI note in correspondence.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>127</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1009</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																				
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: <u>KTS</u> Date: <u>6/10/2026</u> ☐ Battery Replacement ☒ Dry Gas Regulator Replacement ☒ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	
						Simulator	Serial #	Lot#/Exp																
						0.050	MP5088	202603A 3/3/2028																
						0.080	MP5089	202603B 3/3/2028																
0.200	MP5090	202603D 3/4/2028																						
0.080 DGS	N/A	23825080A1 10/5/2027																						
Optical Bench Adjustment By: _____		Department Inspection By: <u>KTS</u>																						
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																						
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1013</u> Instrument: <u>1009</u>																				
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																				
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																				
0.100				Simulator	Serial Number																			
0.200				0.000	MP5086																			
0.300				Interferent	MP5087																			
0.080 DGS	N/A			0.050	MP5088																			
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																			
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																			
0.050				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____																				
0.080																								
0.200																								
0.080 DGS	N/A																							
Gauge ID #: _____ Gauge: _____ Instrument: _____																								
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																				
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.06.16 09:14:49 -04'00'				LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.18 18:54:03 -04'00'																				
Tech Review				Admin Review																				

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤ 0.003 g/210L of Wet																																																																																																																																																
INDIAN RIVER CSO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001328 06/10/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:53</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:54</td></tr> <tr><td>Control Test</td><td>0.048</td><td>09:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	09:53	Control Test	0.048	09:53	Air Blank	0.000	09:54	Control Test	0.048	09:55	Air Blank	0.000	09:55	Control Test	0.049	09:56	Air Blank	0.000	09:56	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel. Std Dev(%)	1.1945		INDIAN RIVER CSO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001328 06/10/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	09:57	Control Test	0.077	09:58	Air Blank	0.000	09:59	Control Test	0.077	09:59	Air Blank	0.000	10:00	Control Test	0.077	10:01	Air Blank	0.000	10:01	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		INDIAN RIVER CSO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001328 06/10/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	10:03	Control Test	0.198	10:04	Air Blank	0.000	10:05	Control Test	0.198	10:05	Air Blank	0.000	10:06	Control Test	0.198	10:07	Air Blank	0.000	10:07	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		INDIAN RIVER CSO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001328 06/10/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	09:49	Control Test	0.079	09:49	Air Blank	0.000	09:50	Control Test	0.079	09:50	Air Blank	0.000	09:50	Control Test	0.079	09:51	Air Blank	0.000	09:51	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:53																																																																																																																																																	
Control Test	0.048	09:53																																																																																																																																																	
Air Blank	0.000	09:54																																																																																																																																																	
Control Test	0.048	09:55																																																																																																																																																	
Air Blank	0.000	09:55																																																																																																																																																	
Control Test	0.049	09:56																																																																																																																																																	
Air Blank	0.000	09:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:57																																																																																																																																																	
Control Test	0.077	09:58																																																																																																																																																	
Air Blank	0.000	09:59																																																																																																																																																	
Control Test	0.077	09:59																																																																																																																																																	
Air Blank	0.000	10:00																																																																																																																																																	
Control Test	0.077	10:01																																																																																																																																																	
Air Blank	0.000	10:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.198	10:04																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test	0.198	10:05																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.198	10:07																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:49																																																																																																																																																	
Control Test	0.079	09:49																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.079	09:50																																																																																																																																																	
Air Blank	0.000	09:50																																																																																																																																																	
Control Test	0.079	09:51																																																																																																																																																	
Air Blank	0.000	09:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CSO
Time of Inspection: 14:13

Date of Inspection: 06/10/2026

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.077	0.199	0.079
0.000	0.049	0.077	0.199	0.079


Standard Deviations	0.0003	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

06/10/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001328, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001328</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>INDIAN RIVER CSO</u>	0.050 g/210 L 0.004
Calibration Date:	<u>06/10/2026</u>	0.080 g/210 L 0.004
Calibration Time:	<u>14:13</u>	0.200 g/210 L 0.008
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program


KATIE T SPEARIN,
Department Inspector

Date

06/10/2026

Service • Integrity • Respect • Quality

Alcohol Testing Program - Instrument Processing Sheet

Agency: INDIAN RIVER CO SO

Instrument Serial Number: 80-001328

Date In: 1/29/2026

DI Completion Date: 1/30/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>KTS</u> Date: <u>1/29/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>1/30/2026</u>		Flow Adjustment By: _____																																		
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>125</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1014</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																		
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 6/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 6/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 6/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG510701 4/17/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202406K 6/19/2026	0.080	MP6292	202406L 6/19/2026	0.200	MP6293	202406N 6/20/2026	0.080 DGS	N/A	AG510701 4/17/2027	<table border="1"><thead><tr><th>Maintenance</th><th>By:</th><th>Date:</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Tank Sensor Tare</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other:</td><td></td><td></td></tr></tbody></table>		Maintenance	By:	Date:	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Tank Sensor Tare			☐ Breath Tube Replacement			☐ Other:		
Simulator	Serial #	Lot#/Exp																																				
0.050	MP6291	202406K 6/19/2026																																				
0.080	MP6292	202406L 6/19/2026																																				
0.200	MP6293	202406N 6/20/2026																																				
0.080 DGS	N/A	AG510701 4/17/2027																																				
Maintenance	By:	Date:																																				
☐ Battery Replacement																																						
☐ Dry Gas Regulator Replacement																																						
☐ Tank Sensor Tare																																						
☐ Breath Tube Replacement																																						
☐ Other:																																						
Optical Bench Adjustment By: _____		Department Inspection By: <u>KTS</u>																																				
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1014</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>								
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
☐ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
Simulator	Serial Number																																					
0.000	MP6294																																					
Interferent	MP5087																																					
0.050	MP5088																																					
0.080	MP5089																																					
0.200	MP5090																																					
Gauge ID #: _____ Gauge: _____ Instrument: _____		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment		☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																		
Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																				
		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.02.05 10:07:08 -05'00'		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.02.05 15:01:29 -05'00'																																		
		Tech Review		Admin Review																																		

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
INDIAN RIVER CSO Intoxilyzer - Alconol Analyzer Model: 8000 01/30/2026 Software: 8100.27 SN 80-001328 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>07:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	07:02	Control Test	0.049	07:03	Air Blank	0.000	07:03	Control Test	0.050	07:04	Air Blank	0.000	07:05	Control Test	0.049	07:05	Air Blank	0.000	07:06	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		INDIAN RIVER CSO Intoxilyzer - Alconol Analyzer Model: 8000 01/30/2026 Software: 8100.27 SN 80-001328 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:07</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	07:07	Control Test	0.080	07:08	Air Blank	0.000	07:09	Control Test	0.079	07:09	Air Blank	0.000	07:10	Control Test	0.079	07:11	Air Blank	0.000	07:11	Control Test Stats			Average	0.0793		Std Dev	0.0016		Rel Std Dev(%)	0.7277		INDIAN RIVER CSO Intoxilyzer - Alconol Analyzer Model: 8000 01/30/2026 Software: 8100.27 SN 80-001328 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:12</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:14</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2010</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	07:12	Control Test	0.201	07:13	Air Blank	0.000	07:14	Control Test	0.201	07:14	Air Blank	0.000	07:15	Control Test	0.201	07:16	Air Blank	0.000	07:16	Control Test Stats			Average	0.2010		Std Dev	0.0000		Rel Std Dev(%)	0.0000		INDIAN RIVER CSO Intoxilyzer - Alconol Analyzer Model: 8000 01/30/2026 Software: 8100.27 SN 80-001328 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	06:58	Control Test	0.080	06:58	Air Blank	0.000	06:59	Control Test	0.080	06:59	Air Blank	0.000	07:00	Control Test	0.080	07:00	Air Blank	0.000	07:01	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.049	07:03																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test	0.050	07:04																																																																																																																																																	
Air Blank	0.000	07:05																																																																																																																																																	
Control Test	0.049	07:05																																																																																																																																																	
Air Blank	0.000	07:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:07																																																																																																																																																	
Control Test	0.080	07:08																																																																																																																																																	
Air Blank	0.000	07:09																																																																																																																																																	
Control Test	0.079	07:09																																																																																																																																																	
Air Blank	0.000	07:10																																																																																																																																																	
Control Test	0.079	07:11																																																																																																																																																	
Air Blank	0.000	07:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:12																																																																																																																																																	
Control Test	0.201	07:13																																																																																																																																																	
Air Blank	0.000	07:14																																																																																																																																																	
Control Test	0.201	07:14																																																																																																																																																	
Air Blank	0.000	07:15																																																																																																																																																	
Control Test	0.201	07:16																																																																																																																																																	
Air Blank	0.000	07:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.080	06:58																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.080	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.080	07:00																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CSO
Time of Inspection: 09:56

Date of Inspection: 01/30/2026

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.080	0.200	0.080


Standard Deviations	0.0005	0.0004	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

01/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001328, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001328</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>INDIAN RIVER CSO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/30/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:56</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/30/2026

Date

KATIE T SPEARIN,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality