

Alcohol Testing Program - Instrument Processing Sheet

Agency: Florida Department of Law Enforcement Instrument Serial Number: 80-005403

Date In: N/A DI Completion Date: 6/18/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/16/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/16/2026</u>		Flow Adjustment By: _____																											
☒ Annual ☐ Dropped Off ☒ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Missing two back feet. Instrument arrived at Tallahassee FDLE laboratory on 6/17/2026 in same visual condition marked above. WKP 6/17/2026		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>220</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1015</u> Instrument: <u>1013</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																											
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____ </td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG526603</td> </tr> <tr> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603	9/23/2027						
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																						
						0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____																						
						0.080	MP6287	202603B 3/3/2028																							
0.200	MP6288	202406N 6/20/2026																													
0.080 DGS	N/A	AG526603																													
		9/23/2027																													
Optical Bench Adjustment By: <u>WKP</u>		Department Inspection By: <u>WKP</u>																													
Barometric Pressure Gauge: <u>1011</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>34419 (x2)</u>																													
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>5/6/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25150	5/6/2027	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: <u>1012 / 1012</u> Instrument: <u>1011 / 1011</u> Mouth Alcohol Solution Lot #: <u>2025-D (x2)</u> Exp: <u>9/25/2027 (x2)</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>	
Simulator	Serial #	Lot #	Expiration																												
0.000	MP6294	N/A	N/A																												
0.040	MP6295	25090	3/11/2027																												
0.100	MP6296	24390	10/29/2026																												
0.200	MP6297	25020	1/14/2027																												
0.300	MP6298	25150	5/6/2027																												
0.080 DGS	N/A	AG526603	9/23/2027																												
☒ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A</td> <td>3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B</td> <td>3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D</td> <td>3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027								
Simulator	Serial #	Lot #	Expiration																												
0.050	MP5088	202603A	3/3/2028																												
0.080	MP5089	202603B	3/3/2028																												
0.200	MP5090	202603D	3/4/2028																												
0.080 DGS	N/A	23825080A1	10/5/2027																												
Gauge ID #: <u>28662</u> Gauge: <u>1011</u> Instrument: <u>1011</u>		0.000 MP6289 0.050 MP5087 0.080 MP5088 0.200 MP5089 0.080 DGS MP5090																													
Notes/Suggested Service: Loaner will be retained by FDLE until requested by an agency. (TDG 6/16/26)		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																													
		☒ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☒ Other: Request for Registration																													
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																													
Digitally signed by Shayla Platt Date: 2026.06.18 12:45:58 -04'00'		Digitally signed by Shayla Platt Date: 2026.06.18 12:45:58 -04'00'																													
		Digitally signed by Shayla Platt Date: 2026.06.18 12:45:58 -04'00'																													

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 14:28

Date of Inspection: 06/16/2026

Serial Number: 80-005403

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/16/2026
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
FOLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:33 Control Test 0.049 14:34 Air Blank 0.000 14:35 Control Test 0.049 14:35 Air Blank 0.000 14:36 Control Test 0.049 14:36 Air Blank 0.000 14:37 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000	FOLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:41 Control Test 0.078 14:42 Air Blank 0.000 14:42 Control Test 0.079 14:43 Air Blank 0.000 14:43 Control Test 0.079 14:44 Air Blank 0.000 14:45 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7335	FOLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:56 Control Test 0.197 14:57 Air Blank 0.000 14:57 Control Test 0.197 14:58 Air Blank 0.000 14:59 Control Test 0.198 14:59 Air Blank 0.000 15:00 Control Test Stats Average 0.1973 Std Dev 0.0016 Rel Std Dev(%) 0.2926	FOLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/16/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:30 Control Test 0.076 14:31 Air Blank 0.000 14:31 Control Test 0.077 14:31 Air Blank 0.000 14:32 Control Test 0.077 14:32 Air Blank 0.000 14:33 Control Test Stats Average 0.0767 Std Dev 0.0006 Rel Std Dev(%) 0.7531
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
FDLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/18/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:04 Control Test 0.051 10:05 Air Blank 0.000 10:05 Control Test 0.051 10:06 Air Blank 0.000 10:06 Control Test 0.051 10:07 Air Blank 0.000 10:08 Control Test Stats Average 0.0510 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	FDLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/18/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:09 Control Test 0.079 10:10 Air Blank 0.000 10:10 Control Test 0.079 10:11 Air Blank 0.000 10:11 Control Test 0.079 10:12 Air Blank 0.000 10:12 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	FDLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/18/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:14 Control Test 0.198 10:14 Air Blank 0.000 10:15 Control Test 0.198 10:15 Air Blank 0.000 10:16 Control Test 0.198 10:17 Air Blank 0.000 10:17 Control Test Stats Average 0.1980 Std Dev 0.0000 Rel. Std Dev(%) 0.0000	FDLE Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005403 06/18/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:19 Control Test 0.081 10:19 Air Blank 0.000 10:20 Control Test 0.081 10:20 Air Blank 0.000 10:21 Control Test 0.081 10:21 Air Blank 0.000 10:22 Control Test 0.081 10:22 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel. Std Dev(%) 0.0000
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 12:05

Date of Inspection: 06/18/2026

Serial Number: 80-005403

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.051	0.078	0.199	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.052	0.080	0.199	0.079
0.000	0.052	0.080	0.198	0.080
0.000	0.052	0.080	0.199	0.080
0.000	0.051	0.080	0.200	0.080

Standard Deviations	0.0008	0.0006	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON
Signature and Printed Name

06/18/2026
Date

Request for Registration

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-005403

OWNING AGENCY: Florida Department of Law Enforcement

DATE OF DEPARTMENT INSPECTION: _____

AGENCY INSPECTOR: N/A

ADDRESS: 2331 Phillips Road

CITY, STATE, ZIP: Tallahassee, FL 32308

TELEPHONE NUMBER: 850-617-1290

EMAIL ADDRESS (if available): alcoholtestingprogram@fdle.state.fl.us

For Program Office Use Only:

- ☐ Registration Issued
- ☒ Instrument Added to COBRA
- ☒ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Master Instrument List



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005403, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005403</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/18/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:05</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Digitally signed by
Wen-Chi Pierson

Date: 2026.06.18
12:35:51 -04'00'

Pierson

06/18/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality