

Alcohol Testing Program - Instrument Processing Sheet

Agency: Osceola County SO

Instrument Serial Number: 80-001715

Date In: 7/8/2026

DI Completion Date: 7/24/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>SLH</u> Date: <u>7/8/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>7/9/2026</u>		Flow Adjustment By: _____																																					
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>217</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.498</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1016</u> Instrument: <u>1016</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																					
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A 03/03/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B 03/03/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D 03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 03/03/2028	0.080	MP5089	202603B 03/03/2028	0.200	MP5090	202603D 03/04/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																						
Simulator	Serial #	Lot#/Exp																																							
0.050	MP5088	202603A 03/03/2028																																							
0.080	MP5089	202603B 03/03/2028																																							
0.200	MP5090	202603D 03/04/2028																																							
0.080 DGS	N/A	23825080A1 10/5/2027																																							
Optical Bench Adjustment By: <u>WKP</u>		Department Inspection By: <u>SLH</u>																																							
Barometric Pressure Gauge: <u>1010</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>28427</u>																																							
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>3/11/2027</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>1/14/2027</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>9/23/2027</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	AG526603	9/23/2027	Gauge: <u>1014</u> Instrument: <u>1015</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>											
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6294	N/A	N/A																																						
0.040	MP6295	25090	3/11/2027																																						
0.100	MP6296	24390	10/29/2026																																						
0.200	MP6297	25020	1/14/2027																																						
0.300	MP6298	24430	12/10/2026																																						
0.080 DGS	N/A	AG526603	9/23/2027																																						
☒ Post Optical Bench Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>3/3/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>3/3/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>3/4/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1</td><td>10/5/2027</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	3/3/2028	0.080	MP5089	202603B	3/3/2028	0.200	MP5090	202603D	3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP5088	202603A	3/3/2028																																						
0.080	MP5089	202603B	3/3/2028																																						
0.200	MP5090	202603D	3/4/2028																																						
0.080 DGS	N/A	23825080A1	10/5/2027																																						
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
Gauge ID #: <u>34420</u> Gauge: <u>1010</u> Instrument: <u>1009</u>		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																																							
Notes/Suggested Service:		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																																							
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																							
		Digitally signed by Taylor Gutschow Date: 2026.07.27 15:33:26 -04'00'																																							
		Taylor Gutschow Shayla Platt Tech Review Admin Review 12:53:58 -04'00'																																							

Stability Checks

80-00 1715

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test</td><td>0.047</td><td>11:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0833</td><td></td></tr> </tbody> </table> <i>Mijl</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:24	Control Test	0.049	11:25	Air Blank	0.000	11:25	Control Test	0.047	11:26	Air Blank	0.000	11:26	Control Test	0.048	11:27	Air Blank	0.000	11:28	Control Test Stats			Average	0.0480		Std Dev	0.0010		Rel Std Dev(%)	2.0833		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0763</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr> </tbody> </table> <i>Mijl</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:59	Control Test	0.076	12:59	Air Blank	0.000	13:00	Control Test	0.076	13:00	Air Blank	0.000	13:01	Control Test	0.077	13:02	Air Blank	0.000	13:02	Control Test Stats			Average	0.0763		Std Dev	0.0006		Rel Std Dev(%)	0.7564		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </tbody> </table> <i>Mijl</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:03	Control Test	0.197	13:04	Air Blank	0.000	13:05	Control Test	0.197	13:05	Air Blank	0.000	13:06	Control Test	0.198	13:07	Air Blank	0.000	13:07	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/09/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <i>Mijl</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:11	Control Test	0.080	13:12	Air Blank	0.000	13:12	Control Test	0.081	13:13	Air Blank	0.000	13:13	Control Test	0.080	13:13	Air Blank	0.000	13:14	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.049	11:25																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test	0.047	11:26																																																																																																																																																	
Air Blank	0.000	11:26																																																																																																																																																	
Control Test	0.048	11:27																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0833																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.076	12:59																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test	0.076	13:00																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.077	13:02																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.197	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.197	13:05																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.198	13:07																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.080	13:12																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.081	13:13																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.080	13:13																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

needs optical bench
adjustment. sub

Verified instrument was
able to read gas pressure. ☒

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
07/21/2026 12:10:07

Auto Calibration
Max Power Res Value = 19
Auto Range Res Value = 13

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12601, 9um lo = 12866

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.2710 (-0.0210)
Sample #2 = 0.2710 (0.0320)
Sample #3 = 0.2610 (0.0580)
Sample #4 = 0.2720 (0.0610)
Avg % Abs = 0.2680 (0.0503)
STD DEV = 0.0061 (0.0159)
REL STD DEV = 2.270 (31.684)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.3300 (-0.0120)
Sample #2 = 0.3180 (0.0130)
Sample #3 = 0.3060 (0.0040)
Sample #4 = 0.3360 (0.0160)
Avg % Abs = 0.3200 (0.0110)
STD DEV = 0.0151 (0.0062)
REL STD DEV = 4.719 (56.773)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12591, 9um lo = 12862

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.0200 (-0.0200)
Sample #2 = 0.9730 (0.0110)
Sample #3 = 0.9740 (0.0520)
Sample #4 = 0.9880 (0.0600)
Avg % Abs = 0.9783 (0.0410)
STD DEV = 0.0084 (0.0263)
REL STD DEV = 0.857 (64.114)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.6410 (-0.0090)
Sample #2 = 1.6160 (0.0080)
Sample #3 = 1.6120 (0.0330)
Sample #4 = 1.6210 (0.0470)
Avg % Abs = 1.6163 (0.0293)
STD DEV = 0.0045 (0.0198)
REL STD DEV = 0.279 (67.353)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12579, 9um lo = 12854

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 2.0360 (-0.0200)
Sample #2 = 2.0300 (0.0050)
Sample #3 = 2.0200 (0.0380)
Sample #4 = 2.0230 (0.0280)
Avg % Abs = 2.0243 (0.0237)
STD DEV = 0.0051 (0.0169)
REL STD DEV = 0.253 (71.499)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.5920 (-0.0250)
Sample #2 = 3.5160 (0.0330)
Sample #3 = 3.5550 (0.0170)
Sample #4 = 3.5640 (0.0220)
Avg % Abs = 3.5450 (0.0240)
STD DEV = 0.0255 (0.0082)
REL STD DEV = 0.720 (34.106)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12573, 9um lo = 12848

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7520 (0.0020)
Sample #2 = 3.7290 (0.0090)
Sample #3 = 3.7150 (0.0470)
Sample #4 = 3.7190 (0.0430)
Avg % Abs = 3.7210 (0.0330)
STD DEV = 0.0072 (0.0209)
REL STD DEV = 0.194 (63.275)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.6690 (-0.0150)
Sample #2 = 6.5970 (0.0360)
Sample #3 = 6.6240 (0.0380)
Sample #4 = 6.6120 (0.0390)
Avg % Abs = 6.6110 (0.0377)
STD DEV = 0.0135 (0.0015)
REL STD DEV = 0.205 (4.055)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12569, 9um lo = 12845

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.4380 (-0.0040)
Sample #2 = 5.3860 (0.0320)
Sample #3 = 5.3990 (0.0520)
Sample #4 = 5.4080 (0.0410)
Avg % Abs = 5.3977 (0.0417)
STD DEV = 0.0111 (0.0100)
REL STD DEV = 0.205 (24.040)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.6200 (-0.0090)
Sample #2 = 9.5610 (0.0530)
Sample #3 = 9.5870 (0.0500)
Sample #4 = 9.5230 (0.0620)
Avg % Abs = 9.5570 (0.0550)
STD DEV = 0.0322 (0.0062)
REL STD DEV = 0.337 (11.355)

Sol Value = 0.400 g/210L ***
Fit value = 1.905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12573, 9um lo = 12848

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7520 (0.0020)
Sample #2 = 3.7290 (0.0090)
Sample #3 = 3.7150 (0.0470)
Sample #4 = 3.7190 (0.0430)
Avg % Abs = 3.7210 (0.0330)
STD DEV = 0.0072 (0.0209)
REL STD DEV = 0.194 (63.275)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.268
Std Dev = 0.01 Rel Std Dev = 2.27
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.978
Std Dev = 0.01 Rel Std Dev = 0.86
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.024
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.721
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.398
Std Dev = 0.01 Rel Std Dev = 0.20
Zero Order Coef = -728.35
First Order Coef = 2679.81
Second Order Coef = 19.05
Standard Deviation = 12.241229

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.320
Std Dev = 0.02 Rel Std Dev = 4.72
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.616
Std Dev = 0.00 Rel Std Dev = 0.28
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.545
Std Dev = 0.03 Rel Std Dev = 0.72
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.611
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.557
Std Dev = 0.03 Rel Std Dev = 0.34
Zero Order Coef = -454.60
First Order Coef = 1433.82
Second Order Coef = 11.36
Standard Deviation = 8.201324

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0002
0.040 0.040 -0.0001
0.100 0.100 -0.0003
0.200 0.200 0.0004
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 0.0003
0.100 0.100 -0.0002
0.200 0.200 0.0001
0.300 0.300 -0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2949.00
Sample #2 = 2635.00
Sample #3 = 2569.00
Sample #4 = 2580.00
Average Result = 2594.6667
STD DEV = 35.3601
REL STD DEV = 1.363

***** CHANNEL 2 *****
Sample #1 = 3000.00
Sample #2 = 2988.00
Sample #3 = 2980.00
Sample #4 = 2985.00
Average Result = 2984.3333
STD DEV = 4.0415
REL STD DEV = 0.135

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1009
3 um H2O Adjust (mg/l*10,000) = 1215
9 um H2O Adjust (mg/l*10,000) = 825
***** AUTO CAL PASS

80-001715
Optical Bench
Calibration Adjustment
WLP 7/21/2026

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:48</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:49</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menp</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:48	Control Test	0.050	13:49	Air Blank	0.000	13:49	Control Test	0.050	13:50	Air Blank	0.000	13:50	Control Test	0.050	13:51	Air Blank	0.000	13:52	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:03</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4555</td><td></td></tr> </tbody> </table> <i>menp</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:01	Control Test	0.080	14:02	Air Blank	0.000	14:03	Control Test	0.080	14:03	Air Blank	0.000	14:04	Control Test	0.079	14:04	Air Blank	0.000	14:05	Control Test Stats			Average	0.0793		Std Dev	0.0012		Rel Std Dev(%)	1.4555		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table> <i>menp</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:20	Control Test	0.199	14:21	Air Blank	0.000	14:22	Control Test	0.199	14:22	Air Blank	0.000	14:23	Control Test	0.198	14:24	Air Blank	0.000	14:24	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		OSCEOLA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001715 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <i>menp</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:11	Control Test	0.080	13:11	Air Blank	0.000	13:12	Control Test	0.080	13:12	Air Blank	0.000	13:13	Control Test	0.081	13:13	Air Blank	0.000	13:14	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:48																																																																																																																																																	
Control Test	0.050	13:49																																																																																																																																																	
Air Blank	0.000	13:49																																																																																																																																																	
Control Test	0.050	13:50																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test	0.050	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:01																																																																																																																																																	
Control Test	0.080	14:02																																																																																																																																																	
Air Blank	0.000	14:03																																																																																																																																																	
Control Test	0.080	14:03																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test	0.079	14:04																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4555																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test	0.199	14:21																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.199	14:22																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.198	14:24																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.080	13:11																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.080	13:12																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.081	13:13																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 14:52

Date of Inspection: 07/24/2026

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.051	0.081	0.199	0.079

Standard Deviations	0.0006	0.0006	0.0007	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/24/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001715, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001715</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OSCEOLA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/24/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:52</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION


This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program


LEANDRA HIGGINSBOTHAM,
Department Inspector

07/24/2026

Date

Service • Integrity • Respect • Quality

Alcohol Testing Program - Instrument Processing Sheet

Agency: OSCEOLA COUNTY SO **Instrument Serial Number:** 80-001715

Date In: 1/8/2026 **DI Completion Date:** N/A ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: <u>KTS</u> Date: <u>1/8/2026</u>		Quality Checks By: _____ Date: _____		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI note. Right keyboard screw is broken off. KTS 1/8/26		☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value: _____ ☐ Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) ☐ Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ ☐ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	
Simulator	Serial #	Lot#/Exp																		
0.050																				
0.080																				
0.200																				
0.080 DGS	N/A																			
Optical Bench Adjustment By: _____		Department Inspection By: _____																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: _____																		
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____																
0.040				Acetone Stock Solution Lot #: _____ Exp: _____																
0.100				Simulator	Serial Number															
0.200				0.000																
0.300				Interferent																
0.080 DGS	N/A			0.050																
☐ Post Optical Bench Adjustment Stability Checks				0.080																
Simulator	Serial #	Lot #	Expiration	0.200																
0.050				Attachments																
0.080				☐ Form 41 ☐ Post-Stability Checks ☐ Stability Checks ☐ Flow Adjustment ☐ Calibration Certificate ☒ Form 40 ☐ Optical Bench Adjustment ☒ Other: _____																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____ Gauge: _____ Instrument: _____																				
Notes/Suggested Service: Additional forms in packet are 2 Form 40's showing DSP failure, a CMI service form and a FDLE return authorization sheet. SLH 1/26/26				☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																
				<table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;"> Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.02.05 14:39:54 -05'00' </td> <td style="width: 50%; text-align: center;"> Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.02.05 15:15:15 -05'00' </td> </tr> </table>		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.02.05 14:39:54 -05'00'	Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.02.05 15:15:15 -05'00'													
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.02.05 14:39:54 -05'00'	Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.02.05 15:15:15 -05'00'															
<table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;">Tech Review</td> <td style="width: 50%; text-align: center;">Admin Review</td> </tr> </table>		Tech Review	Admin Review																	
Tech Review	Admin Review																			

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 09:28

Date of Inspection: 12/24/2025

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

DSP FAILNon-compliance:DIAGNOSTIC CHECK SHOWED DSP FAIL.

DSP - Failure during diagnostic check

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL/SHAWN E LYONS/DANIEL

Signature and Printed Name

12/24/2025
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 09:28

Date of Inspection: 12/24/2025

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

DSP FAILNon-compliance:DIAGNOSTIC CHECK SHOWED DSP FAIL.

SENT OUT FOR MAINT/REPAIRS

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons

SD
Signature and Printed Name

DANIEL/SHAWN E LYONS/DANIEL

12/24/2025
Date



Service Evaluation Form

*** Hazardous Material Warning! – DO NOT return gas cylinder with instrument! ***

Contact information:

Name: DEP. DAN LYONS Phone: (407) 799-1711
Email: DANIEL.LYONS@OSCEOLA.SHERIFF.ORG Customer # 347441 (contact Customer Service)
866-835-0690

Your Billing Address

Agency Name: OSCEOLA CO SHERIFF'S OFFICE
Address: 2601 E. HWY 192
City: KISSIMMEE State: FL Zip: 34744
Credit Card/PO #:
Name on Card:
Expiration Date: CVV

Your Shipping Address

Agency Name: OSCEOLA CO SHERIFF'S OFFICE
Address: 2601 E. HWY 192
City: KISSIMMEE State: FL Zip: 34744
Contact Person: DANIEL LYONS
Contact Phone: 407 799 1711
Email: DANIEL.LYONS@OSCEOLA.SHERIFF.ORG

3% processing fee charged for payment made with credit card.

Instrument Serial Number: 80-001715

Detailed Description of Problem:

DSP FAILURE DURING DIAGNOSTIC CHECK

Note: For instruments not under warranty, an evaluation fee of \$90.⁰⁰ (infrared) or \$48 (fuel cell), plus return shipping cost, will apply to all service items.

End of Life (EOL) – Customers who send in instrument(s) which CMI considers end-of-life (S-D2, SD5, I-200, I-300, "motor-based" I-240, I-400, or I-5000), will be provided options for either upgrade or replacement opportunities, as well as choices for the disposal of end-of-life instrument(s).

☒ **Calibration-Only Service (Alcoblow, I-500, I-800 only) - \$48**
Any instrument which cannot be calibrated will receive a repair estimate for approval.

☒ I authorize all repairs to be performed. ☒ Please send an estimate for approval

☒ Certificate of Calibration (\$25 for I-500 and I-800) (\$43 for I-240, I-400, and I-900)

Authorized By:

DANIEL LYONS DEPUTY
Name (Please Print) Title
D Lyons 445 12/29/25
Signature Date

Ship item to:

CMI, Inc.
Attn: Service Dept.
316 East Ninth Street
Owensboro, KY 42303

Note: Please ship items in their original shipping container

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Daniel Lyons on 12/24/2025Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-001715

Bill To Address:

Osceola County Sheriff Office
2601 East Irlo Bronson Memorial Hwy
Kissimmee, FL 34744

Ship to Address:

FDLE Offsite Mail Facility
c/o FDLE HQ
813 B Lake Bradford Rd
Tallahassee, FL 32304

Reason for Return:

DSP failure on initialization and DVM monitoring were zeros at the 3 and 9 filters.

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Daniel Lyons Phone #: 407-799-1711Email: Daniel.Lyons@osceolasheriff.orgATP Contact Name: LeAndra Higginbotham ATP Email: leandrahigginbotham@fdle.state.fl.us