

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 21:00

Serial Number: 80-000831
Date of Inspection: 10/19/2023 Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

While doing the Agency Inspection and during the initial Diagnostic Check to put the instrument into Ready Mode the instrument stated Modem Test Fail. Talked with Shayla Platt with FDLE and she advised to return the instrument to FDLE for Department Inspection.

The above instrument complies (____) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cpl. C. Thomas 514/1602

Christopher L THOMAS

Signature and Printed Name

10/19/2023
Date