

Alcohol Testing Program - Instrument Processing Sheet

Agency: BARTOW PD Instrument Serial Number: 80-001350
 Date In: 12/30/2025 DI Completion Date: 1/15/26 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/30/2025</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: AI note. KTS 1/12/26	Quality Checks By: <u>KTS</u> Date: <u>1/12/26</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>117</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.136</u> (.139-.169) 36 mm <u>0.152</u> (.156-.190) 53 mm <u>0.222</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1024</u> Instrument: <u>1019</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: <u>KTS</u> Flow Column #: <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>117 / 118</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm <u>0.140 / 0.144</u> (.139-.169) 36 mm <u>0.156 / 0.156</u> (.156-.190) 53 mm <u>0.230 / 0.238</u> (.228-.278) 103 mm <u>0.500 / 0.492</u> (.447-.547)
--	---	---

			Maintenance By: <u>KTS</u>	Date: <u>1/6/26</u>
			<input checked="" type="checkbox"/> Battery Replacement	
			<input type="checkbox"/> Dry Gas Regulator Replacement	
			<input type="checkbox"/> Tank Sensor Tare	
			<input type="checkbox"/> Breath Tube Replacement	
			<input type="checkbox"/> Other:	

Optical Bench Adjustment By: _____ **Department Inspection** By: KTS
 Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: 28427

Simulator	Serial #	Lot #	Expiration	Gauge: 1011	Instrument: 1007
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/27</u>
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>9/22/27</u>
0.100				Simulator	Serial Number
0.200				0.000	<u>MP6289</u>
0.300				Interferent	<u>MP6290</u>
0.080 DGS	N/A			0.050	<u>MP5088</u>
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	<u>MP5089</u>
Simulator	Serial #	Lot #	Expiration	0.200	<u>MP5090</u>
0.050				Attachments	
0.080				<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
0.200				<input checked="" type="checkbox"/> Stability Checks	<input checked="" type="checkbox"/> Flow Adjustment
0.080 DGS	N/A			<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
Gauge ID #: _____				<input type="checkbox"/> Optical Bench Adjustment	<input checked="" type="checkbox"/> Other: Agency Memo
Gauge: _____ Instrument: _____					

Notes/Suggested Service: Flow adjustment conducted on 1/12/26 did not print due to paper jam. Second flow adjustment conducted on 1/14/26. KTS 1/14/26
 Opened instrument back up on 1/13/26 to ensure all connections were secured during battery change after stability results on 1/12/26. Tube was unattached. Attached tube, closed instrument and conducted additional stabilities on 1/14/26. KTS 1/14/26

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.01.23 11:39:06 -05'00' Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.01.26 09:02:05 -05'00'

Tech Review **Admin Review**



Stephen P. Walker
Chief of Police

BARTOW POLICE DEPARTMENT

Good morning, Kaitlyn Spearin,

The Intoxilyzer 8000 shipped in this box needs the annual 2026 inspection and an internal battery replacement. If you have any questions, email me at msills.pd@cityofbartow.net or contact me via cell phone at 863-800-7217. Either way is fine and I will respond to either method.

Officer Sills #1116

BARTOW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001350
01/12/2026
Software: 8100.27

#1: DGS 1/12/26
KTS

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	PUR*	09:44
Air Blank	PUR*	09:44

*Purge Fail



Operator's Signature

BARTOW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001350
01/12/2026
Software: 8100.27

#2: DGS
1/12/26 KTS

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.000	09:52
Air Blank	0.000	09:53
Control Test	0.000	09:53
Air Blank	0.000	09:53
Control Test	0.000	09:54
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature

BARTOW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001350
01/12/2026
Software: 8100.27

#3: DGS
1/12/26 KTS

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.017	10:15
Air Blank	0.000	10:15
Control Test	0.000	10:16
Air Blank	0.000	10:16
Control Test	0.000	10:16
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0057	
Std Dev	0.0098	
Rel Std Dev(%)	173.2051	




Operator's Signature

BARTOW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001350
01/12/2026
Software: 8100.27

KTS
~~0.05g/l~~
0.050 Stability
1/12/26 KTS

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.000	10:19
Air Blank	0.000	10:19
Control Test	0.000	10:20
Air Blank	0.000	10:20
Control Test	0.000	10:21
Air Blank	0.000	10:22
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

BARTOW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001350
01/14/2026
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.000

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.043

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.949

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 645

Rounded Intercept = -423379

Correlation = 0.99750

Flow Adjustment
1/14/26 *Kat*

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
<p>BARTON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001350 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:42</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:43</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:44</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0408</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	08:42	Control Test	0.050	08:42	Air Blank	0.000	08:43	Control Test	0.048	08:44	Air Blank	0.000	08:44	Control Test	0.049	08:45	Air Blank	0.000	08:46	Control Test Stats			Average	0.0490		Std Dev	0.0010		Rel Std Dev(%)	2.0408		<p>BARTON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001350 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:39</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>08:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	08:38	Control Test	0.080	08:38	Air Blank	0.000	08:39	Control Test	0.081	08:39	Air Blank	0.000	08:40	Control Test	0.080	08:40	Air Blank	0.000	08:40	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<p>BARTON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001350 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr> <tr><td>Control Test</td><td>0.196</td><td>08:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr> <tr><td>Control Test</td><td>0.196</td><td>08:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test</td><td>0.195</td><td>08:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	08:52	Control Test	0.196	08:53	Air Blank	0.000	08:54	Control Test	0.196	08:54	Air Blank	0.000	08:55	Control Test	0.195	08:56	Air Blank	0.000	08:56	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		<p>BARTON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001350 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:47</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	08:47	Control Test	0.079	08:48	Air Blank	0.000	08:49	Control Test	0.079	08:49	Air Blank	0.000	08:50	Control Test	0.079	08:51	Air Blank	0.000	08:51	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:42																																																																																																																																																	
Control Test	0.050	08:42																																																																																																																																																	
Air Blank	0.000	08:43																																																																																																																																																	
Control Test	0.048	08:44																																																																																																																																																	
Air Blank	0.000	08:44																																																																																																																																																	
Control Test	0.049	08:45																																																																																																																																																	
Air Blank	0.000	08:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.0408																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:38																																																																																																																																																	
Control Test	0.080	08:38																																																																																																																																																	
Air Blank	0.000	08:39																																																																																																																																																	
Control Test	0.081	08:39																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test	0.080	08:40																																																																																																																																																	
Air Blank	0.000	08:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test	0.196	08:53																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test	0.196	08:54																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.195	08:56																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:47																																																																																																																																																	
Control Test	0.079	08:48																																																																																																																																																	
Air Blank	0.000	08:49																																																																																																																																																	
Control Test	0.079	08:49																																																																																																																																																	
Air Blank	0.000	08:50																																																																																																																																																	
Control Test	0.079	08:51																																																																																																																																																	
Air Blank	0.000	08:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BARTOW PD

Time of Inspection: 12:27

Date of Inspection: 01/15/2026

Serial Number: 80-001350

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.051	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.051	0.079	0.197	0.079

Standard Deviations	0.0004	0.0004	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

01/15/2026
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001350, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001350</u>	UNCERTAINTY* ±
Owning Agency:	<u>BARTOW PD</u>	0.050 g/ 210 L
Calibration Date:	<u>01/15/2026</u>	0.080 g/ 210 L
Calibration Time:	<u>12:27</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/15/2026

Date


 KATIE T SPEARIN,
 Department Inspector

FDLE/ATP Form 69 January 2026
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality